



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 15, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: OCOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/09/26 11:31 through 04/24/26 11:33.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
OCOK	6040914-02	Air	04/02/26 23:59	04/09/26 11:31
OCOK	6042411-01	Air	04/14/26 23:59	04/24/26 11:33
OCOK	6042411-02	Air	04/20/26 23:59	04/24/26 11:33



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQS SITE CODE: 40 109 1037
SITE CODE: OCOK

Description: OCOK
Pressure @ Receipt: 5.5
Comments:

Lab ID: 6040914-02
Canister #: 43311

Sampled: 04/02/26 23:59
Received: 04/09/26 11:31
Analyzed: 04/23/26 20:05

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.505	2.50		0.0237
Chloromethane	0.522	1.08		0.0213
Dichlorotetrafluoroethane	0.0181	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	ND	ND	CF, LL, U	0.0215
Ethylene oxide	0.117	0.21		0.0276
Bromomethane	0.00930	0.04	U	0.0104
Chloroethane	0.0183	0.05		0.0136
Acetonitrile	245	412.00	E, QB-01, QB-03 CF, LK, QB-01	0.184
Acrolein	0.255	0.59		0.0977
Trichlorofluoromethane	0.235	1.32		0.00950
Acrylonitrile	0.0248	0.05		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.109	0.38		0.0489
Carbon Disulfide	0.0155	0.05	U	0.0473
Trichlorotrifluoroethane	0.0785	0.60		0.0210
trans-1,2-Dichloroethylene	8.00E-4	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	0.0280	0.11		0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0132	0.06	U	0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0112	0.05	U	0.0116
1,1,1-Trichloroethane	8.00E-4	0.00	U	0.00952
Benzene	0.0865	0.28		0.0158
Carbon Tetrachloride	0.0774	0.49		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	ND	ND	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0196	0.08	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6040914-02

Sampled: 04/02/26 23:59

Pressure @ Receipt: 5.5

Canister #: 43311

Received: 04/09/26 11:31

Comments:

Analyzed: 04/23/26 20:05

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.110	0.42		0.0380
Dibromochloromethane	2.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0227	0.11	U	0.0258
Tetrachloroethylene	0.00510	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0119	0.05	U	0.0130
m,p-Xylene	0.0277	0.12	U	0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0108	0.05	U	0.0235
1,3,5-Trimethylbenzene	0.00160	0.01	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00170	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	3.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQS SITE CODE: 40 109 1037
SITE CODE: OCOK

Description: OCOK
Pressure @ Receipt: 4.5
Comments:

Lab ID: 6042411-01
Canister #: 912401

Sampled: 04/14/26 23:59
Received: 04/24/26 11:33
Analyzed: 05/01/26 22:23

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.466	2.31		0.0237
Chloromethane	0.518	1.07		0.0213
Dichlorotetrafluoroethane	0.0150	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	ND	ND	U	0.0215
Ethylene oxide	0.0744	0.13		0.0276
Bromomethane	0.00930	0.04	U	0.0104
Chloroethane	0.00750	0.02	U	0.0136
Acetonitrile	590	993.00	E	0.184
Acrolein	0.163	0.37	CF, LK	0.0977
Trichlorofluoromethane	0.217	1.22		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.121	0.42		0.0489
Carbon Disulfide	0.0198	0.06	U	0.0473
Trichlorotrifluoroethane	0.0703	0.54		0.0210
trans-1,2-Dichloroethylene	0.00420	0.02	U	0.0171
1,1-Dichloroethane	6.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0168	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0130	0.05		0.0116
1,1,1-Trichloroethane	9.00E-4	0.00	U	0.00952
Carbon Tetrachloride	0.0741	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00270	0.01	U	0.0170
Methyl Methacrylate	ND	ND	LK, U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0262	0.11	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6042411-01

Sampled: 04/14/26 23:59

Pressure @ Receipt: 4.5

Canister #: 912401

Received: 04/24/26 11:33

Comments:

Analyzed: 05/01/26 22:23

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.157	0.59		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0264	0.12		0.0258
Tetrachloroethylene	0.00630	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0120	0.05	U	0.0130
m,p-Xylene	0.0308	0.13	U	0.0418
Bromoform	0.00290	0.03	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0115	0.05	U	0.0235
1,3,5-Trimethylbenzene	0.00200	0.01	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00210	0.01	U	0.0287
o-Dichlorobenzene	4.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00100	0.01	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQ5 SITE CODE: 40 109 1037
SITE CODE: OCOK

Description: OCOK
Pressure @ Receipt: 4.5
Comments:

Lab ID: 6042411-02
Canister #: 40468

Sampled: 04/20/26 23:59
Received: 04/24/26 11:33
Analyzed: 05/02/26 02:17

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.473	2.34		0.0237
Chloromethane	0.510	1.06		0.0213
Dichlorotetrafluoroethane	0.0182	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00410	0.01	CF, LL, U	0.0215
Ethylene oxide	0.0353	0.06		0.0276
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.00830	0.02	U	0.0136
Acetonitrile	244	411.00	E	0.184
Acrolein	0.118	0.27	CF, LK	0.0977
Trichlorofluoromethane	0.216	1.22		0.00950
Acrylonitrile	0.0231	0.05		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.164	0.57		0.0489
Carbon Disulfide	0.0150	0.05	U	0.0473
Trichlorotrifluoroethane	0.0752	0.58		0.0210
trans-1,2-Dichloroethylene	0.00130	0.01	U	0.0171
1,1-Dichloroethane	0.00150	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	2.00E-4	0.00	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	6.00E-4	0.00	U	0.00905
Chloroform	0.0176	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0191	0.08		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.124	0.40		0.0158
Carbon Tetrachloride	0.0683	0.43		0.00750
tert-Amyl Methyl Ether	2.00E-4	0.00	LL, U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	8.00E-4	0.01	U	0.00889
Trichloroethylene	0.00250	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0130	0.05	U	0.0578



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6042411-02

Sampled: 04/20/26 23:59

Pressure @ Receipt: 4.5

Canister #: 40468

Received: 04/24/26 11:33

Comments:

Analyzed: 05/02/26 02:17

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.124	0.47		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0269	0.13		0.0258
Tetrachloroethylene	0.00670	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0124	0.05	U	0.0130
m,p-Xylene	0.0314	0.14	U	0.0418
Bromoform	0.00100	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0110	0.05	U	0.0235
1,3,5-Trimethylbenzene	0.00200	0.01	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00170	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	2.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2302 - Summa Canister Prep

Blank (B6D2302-BLK1)

Prepared: 04/17/26 Analyzed: 04/23/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, QB-01, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2302 - Summa Canister Prep

Blank (B6D2302-BLK1) Continued

Prepared: 04/17/26 Analyzed: 04/23/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1)

Prepared: 04/29/26 Analyzed: 05/01/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1) Continued

Prepared: 04/29/26 Analyzed: 05/01/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				LL, U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E1210 - Summa Canister Prep

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1210 - Summa Canister Prep

Blank (B6E1210-BLK1)

Prepared: 04/30/26 Analyzed: 05/01/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				LK, U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQS SITE CODE: 40 109 1037
SITE CODE: OCOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1210 - Summa Canister Prep

Blank (B6E1210-BLK1) Continued

Prepared: 04/30/26 Analyzed: 05/01/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604073

Calibration Check (2604073-CCV1)

Prepared & Analyzed: 04/23/26

Dichlorodifluoromethane	1.19	ppbv	15.7	30.00
Chloromethane	1.07	ppbv	5.2	30.00
Dichlorotetrafluoroethane	1.19	ppbv	14.7	30.00
Vinyl chloride	1.06	ppbv	2.8	30.00
1,3-Butadiene	0.924	ppbv	-10.6	30.00
Ethylene oxide	0.971	ppbv	1.6	30.00
Bromomethane	1.22	ppbv	18.0	30.00
Chloroethane	1.19	ppbv	14.5	30.00
Acetonitrile	1.07	ppbv	3.2	30.00
Acrolein	1.09	ppbv	6.5	30.00
Trichlorofluoromethane	1.23	ppbv	21.9	30.00
Acrylonitrile	1.11	ppbv	0.2	30.00
1,1-Dichloroethene	1.21	ppbv	16.2	30.00
Dichloromethane	1.11	ppbv	6.9	30.00
Carbon Disulfide	1.27	ppbv	16.7	30.00
Trichlorotrifluoroethane	1.27	ppbv	23.3	30.00
trans-1,2-Dichloroethylene	1.14	ppbv	9.5	30.00
1,1-Dichloroethane	1.15	ppbv	11.1	30.00
Methyl tert-Butyl Ether	0.950	ppbv	-8.3	30.00
Chloroprene	0.951	ppbv	-8.4	30.00
cis-1,2-Dichloroethylene	1.09	ppbv	5.2	30.00
Bromochloromethane	1.22	ppbv	18.8	30.00
Chloroform	1.19	ppbv	13.7	30.00
Ethyl tert-Butyl Ether	0.901	ppbv	-13.0	30.00
1,2-Dichloroethane	1.15	ppbv	11.4	30.00
1,1,1-Trichloroethane	1.17	ppbv	13.2	30.00
Benzene	1.10	ppbv	5.5	30.00
Carbon Tetrachloride	1.19	ppbv	16.5	30.00
tert-Amyl Methyl Ether	0.866	ppbv	-16.2	30.00
1,2-Dichloropropane	1.10	ppbv	5.9	30.00
Ethyl Acrylate	0.881	ppbv	-15.7	30.00
Bromodichloromethane	1.08	ppbv	3.8	30.00
Trichloroethylene	1.08	ppbv	4.9	30.00
Methyl Methacrylate	0.863	ppbv	-17.2	30.00
cis-1,3-Dichloropropene	0.972	ppbv	-6.9	30.00
Methyl Isobutyl Ketone	0.867	ppbv	-16.6	30.00

QB-01
LK, QB-01

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604073

Calibration Check (2604073-CCV1) Continued

Prepared & Analyzed: 04/23/26

trans-1,3-Dichloropropene	1.06	ppbv	1.0	30.00
1,1,2-Trichloroethane	1.12	ppbv	7.5	30.00
Toluene	1.01	ppbv	-3.2	30.00
Dibromochloromethane	1.09	ppbv	5.8	30.00
1,2-Dibromoethane	1.08	ppbv	3.7	30.00
n-Octane	0.882	ppbv	-15.1	30.00
Tetrachloroethylene	1.15	ppbv	10.4	30.00
Chlorobenzene	1.13	ppbv	7.0	30.00
Ethylbenzene	0.987	ppbv	-5.7	30.00
m,p-Xylene	2.09	ppbv	-0.1	30.00
Bromoform	1.21	ppbv	15.3	30.00
Styrene	1.08	ppbv	2.1	30.00
1,1,2,2-Tetrachloroethane	1.13	ppbv	6.8	30.00
o-Xylene	0.977	ppbv	-7.5	30.00
1,3,5-Trimethylbenzene	0.919	ppbv	-12.6	30.00
1,2,4-Trimethylbenzene	0.901	ppbv	-14.5	30.00
m-Dichlorobenzene	1.09	ppbv	5.1	30.00
p-Dichlorobenzene	1.07	ppbv	2.6	30.00
o-Dichlorobenzene	1.08	ppbv	3.0	30.00
1,2,4-Trichlorobenzene	0.953	ppbv	-8.2	30.00
Hexachloro-1,3-butadiene	1.08	ppbv	2.1	30.00

Calibration Check (2604073-CCV2)

Prepared: 04/23/26 Analyzed: 04/24/26

Dichlorodifluoromethane	1.17	ppbv	13.5	30.00
Chloromethane	1.07	ppbv	4.8	30.00
Dichlorotetrafluoroethane	1.18	ppbv	13.7	30.00
Vinyl chloride	1.01	ppbv	-2.4	30.00
1,3-Butadiene	0.816	ppbv	-21.1	30.00
Ethylene oxide	0.960	ppbv	0.4	30.00
Bromomethane	1.21	ppbv	17.5	30.00
Chloroethane	1.17	ppbv	13.0	30.00
Acetonitrile	1.12	ppbv	7.7	30.00
Acrolein	1.07	ppbv	4.8	30.00
Trichlorofluoromethane	1.20	ppbv	18.7	30.00
Acrylonitrile	1.07	ppbv	-3.0	30.00
1,1-Dichloroethene	1.21	ppbv	16.0	30.00

QB-01
LK, QB-01

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604073

Calibration Check (2604073-CCV2) Continued

Prepared: 04/23/26 Analyzed: 04/24/26

Dichloromethane	1.11	ppbv	6.9	30.00
Carbon Disulfide	1.28	ppbv	17.2	30.00
Trichlorotrifluoroethane	1.27	ppbv	23.5	30.00
trans-1,2-Dichloroethylene	1.11	ppbv	7.0	30.00
1,1-Dichloroethane	1.14	ppbv	10.0	30.00
Methyl tert-Butyl Ether	0.874	ppbv	-15.6	30.00
Chloroprene	0.894	ppbv	-13.8	30.00
cis-1,2-Dichloroethylene	1.06	ppbv	2.7	30.00
Bromochloromethane	1.21	ppbv	17.7	30.00
Chloroform	1.17	ppbv	11.8	30.00
Ethyl tert-Butyl Ether	0.832	ppbv	-19.7	30.00
1,2-Dichloroethane	1.12	ppbv	8.6	30.00
1,1,1-Trichloroethane	1.13	ppbv	9.7	30.00
Benzene	1.06	ppbv	1.8	30.00
Carbon Tetrachloride	1.15	ppbv	12.7	30.00
tert-Amyl Methyl Ether	0.795	ppbv	-23.1	30.00
1,2-Dichloropropane	0.994	ppbv	-4.2	30.00
Ethyl Acrylate	0.858	ppbv	-18.0	30.00
Bromodichloromethane	1.06	ppbv	2.6	30.00
Trichloroethylene	1.06	ppbv	2.9	30.00
Methyl Methacrylate	0.836	ppbv	-19.7	30.00
cis-1,3-Dichloropropene	0.931	ppbv	-10.8	30.00
Methyl Isobutyl Ketone	0.842	ppbv	-19.0	30.00
trans-1,3-Dichloropropene	1.04	ppbv	-0.8	30.00
1,1,2-Trichloroethane	1.13	ppbv	8.2	30.00
Toluene	0.979	ppbv	-5.9	30.00
Dibromochloromethane	1.08	ppbv	4.7	30.00
1,2-Dibromoethane	1.07	ppbv	2.4	30.00
n-Octane	0.861	ppbv	-17.1	30.00
Tetrachloroethylene	1.13	ppbv	8.2	30.00
Chlorobenzene	1.11	ppbv	5.8	30.00
Ethylbenzene	0.954	ppbv	-8.8	30.00
m,p-Xylene	2.04	ppbv	-2.4	30.00
Bromoform	1.20	ppbv	14.9	30.00
Styrene	1.06	ppbv	0.5	30.00
1,1,2,2-Tetrachloroethane	1.14	ppbv	7.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQS SITE CODE: 40 109 1037
SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604073

Calibration Check (2604073-CCV2) Continued

Prepared: 04/23/26 Analyzed: 04/24/26

o-Xylene	0.937	ppbv	-11.3	30.00
1,3,5-Trimethylbenzene	0.852	ppbv	-19.0	30.00
1,2,4-Trimethylbenzene	0.837	ppbv	-20.6	30.00
m-Dichlorobenzene	1.07	ppbv	2.5	30.00
p-Dichlorobenzene	1.04	ppbv	-0.6	30.00
o-Dichlorobenzene	1.04	ppbv	-0.4	30.00
1,2,4-Trichlorobenzene	0.892	ppbv	-14.0	30.00
Hexachloro-1,3-butadiene	1.04	ppbv	-1.5	30.00

Sequence 2605002

Calibration Check (2605002-CCV1)

Prepared & Analyzed: 05/01/26

Dichlorodifluoromethane	1.07	ppbv	3.9	30.00
Chloromethane	1.04	ppbv	1.9	30.00
Dichlorotetrafluoroethane	1.11	ppbv	7.6	30.00
Vinyl chloride	0.944	ppbv	-8.5	30.00
1,3-Butadiene	0.755	ppbv	-27.0	30.00
Ethylene oxide	0.934	ppbv	-2.3	30.00
Bromomethane	1.15	ppbv	11.1	30.00
Chloroethane	1.16	ppbv	11.7	30.00
Acetonitrile	1.05	ppbv	1.5	30.00
Acrolein	1.06	ppbv	4.2	30.00
Trichlorofluoromethane	1.09	ppbv	8.3	30.00
Acrylonitrile	1.07	ppbv	-3.4	30.00
1,1-Dichloroethene	1.18	ppbv	13.1	30.00
Dichloromethane	1.09	ppbv	5.5	30.00
Carbon Disulfide	1.26	ppbv	15.5	30.00
Trichlorotrifluoroethane	1.21	ppbv	17.3	30.00
trans-1,2-Dichloroethylene	1.03	ppbv	-0.7	30.00
1,1-Dichloroethane	1.11	ppbv	7.2	30.00
Methyl tert-Butyl Ether	0.802	ppbv	-22.6	30.00
Chloroprene	0.823	ppbv	-20.7	30.00
cis-1,2-Dichloroethylene	1.02	ppbv	-1.5	30.00
Bromochloromethane	1.12	ppbv	9.3	30.00
Chloroform	1.11	ppbv	6.1	30.00
Ethyl tert-Butyl Ether	0.766	ppbv	-26.1	30.00
1,2-Dichloroethane	1.05	ppbv	1.2	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV1) Continued

Prepared & Analyzed: 05/01/26

1,1,1-Trichloroethane	1.03	ppbv	-0.3	30.00
Benzene	0.998	ppbv	-3.8	30.00
Carbon Tetrachloride	1.03	ppbv	0.3	30.00
tert-Amyl Methyl Ether	0.735	ppbv	-28.9	30.00
1,2-Dichloropropane	1.04	ppbv	0.6	30.00
Ethyl Acrylate	0.865	ppbv	-17.3	30.00
Bromodichloromethane	1.08	ppbv	4.2	30.00
Trichloroethylene	1.03	ppbv	-0.07	30.00
Methyl Methacrylate	0.832	ppbv	-20.2	30.00
cis-1,3-Dichloropropene	0.902	ppbv	-13.6	30.00
Methyl Isobutyl Ketone	0.842	ppbv	-19.0	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-2.1	30.00
1,1,2-Trichloroethane	1.14	ppbv	9.1	30.00
Toluene	0.952	ppbv	-8.5	30.00
Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.02	ppbv	-2.4	30.00
n-Octane	0.875	ppbv	-15.7	30.00
Tetrachloroethylene	1.07	ppbv	2.7	30.00
Chlorobenzene	1.12	ppbv	6.5	30.00
Ethylbenzene	0.940	ppbv	-10.1	30.00
m,p-Xylene	2.04	ppbv	-2.4	30.00
Bromoform	1.18	ppbv	12.9	30.00
Styrene	1.06	ppbv	0.4	30.00
1,1,2,2-Tetrachloroethane	1.22	ppbv	15.6	30.00
o-Xylene	0.905	ppbv	-14.3	30.00
1,3,5-Trimethylbenzene	0.801	ppbv	-23.8	30.00
1,2,4-Trimethylbenzene	0.776	ppbv	-26.3	30.00
m-Dichlorobenzene	1.04	ppbv	-0.3	30.00
p-Dichlorobenzene	1.01	ppbv	-3.9	30.00
o-Dichlorobenzene	0.996	ppbv	-4.8	30.00
1,2,4-Trichlorobenzene	0.826	ppbv	-20.5	30.00
Hexachloro-1,3-butadiene	0.982	ppbv	-6.8	30.00

Calibration Check (2605002-CCV2)

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorodifluoromethane	1.07	ppbv	3.8	30.00
Chloromethane	1.05	ppbv	3.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorotetrafluoroethane	1.12	ppbv	7.7	30.00	
Vinyl chloride	0.945	ppbv	-8.5	30.00	
1,3-Butadiene	0.737	ppbv	-28.7	30.00	
Ethylene oxide	0.934	ppbv	-2.3	30.00	
Bromomethane	1.15	ppbv	11.0	30.00	
Chloroethane	1.16	ppbv	12.1	30.00	
Acetonitrile	1.10	ppbv	5.6	30.00	
Acrolein	1.05	ppbv	3.4	30.00	LK
Trichlorofluoromethane	1.09	ppbv	7.4	30.00	
Acrylonitrile	1.06	ppbv	-4.2	30.00	
1,1-Dichloroethene	1.17	ppbv	12.9	30.00	
Dichloromethane	1.11	ppbv	7.1	30.00	
Carbon Disulfide	1.25	ppbv	14.8	30.00	
Trichlorotrifluoroethane	1.20	ppbv	16.1	30.00	
trans-1,2-Dichloroethylene	1.03	ppbv	-1.2	30.00	
1,1-Dichloroethane	1.12	ppbv	7.8	30.00	
Methyl tert-Butyl Ether	0.789	ppbv	-23.8	30.00	
Chloroprene	0.800	ppbv	-22.9	30.00	
cis-1,2-Dichloroethylene	1.02	ppbv	-2.0	30.00	
Bromochloromethane	1.12	ppbv	8.5	30.00	
Chloroform	1.11	ppbv	5.9	30.00	
Ethyl tert-Butyl Ether	0.753	ppbv	-27.3	30.00	
1,2-Dichloroethane	1.04	ppbv	0.5	30.00	
1,1,1-Trichloroethane	1.02	ppbv	-1.2	30.00	
Benzene	0.991	ppbv	-4.5	30.00	
Carbon Tetrachloride	1.01	ppbv	-1.3	30.00	
tert-Amyl Methyl Ether	0.721	ppbv	-30.2	30.00	LL
1,2-Dichloropropane	1.06	ppbv	2.3	30.00	
Ethyl Acrylate	0.869	ppbv	-16.9	30.00	
Bromodichloromethane	1.09	ppbv	5.1	30.00	
Trichloroethylene	1.03	ppbv	-0.3	30.00	
Methyl Methacrylate	0.825	ppbv	-20.8	30.00	
cis-1,3-Dichloropropene	0.901	ppbv	-13.7	30.00	
Methyl Isobutyl Ketone	0.857	ppbv	-17.6	30.00	
trans-1,3-Dichloropropene	1.03	ppbv	-1.6	30.00	
1,1,2-Trichloroethane	1.16	ppbv	10.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Toluene	0.954	ppbv	-8.3	30.00
Dibromochloromethane	1.06	ppbv	2.1	30.00
1,2-Dibromoethane	1.02	ppbv	-1.8	30.00
n-Octane	0.879	ppbv	-15.3	30.00
Tetrachloroethylene	1.08	ppbv	3.2	30.00
Chlorobenzene	1.12	ppbv	6.1	30.00
Ethylbenzene	0.924	ppbv	-11.7	30.00
m,p-Xylene	2.01	ppbv	-3.6	30.00
Bromoform	1.18	ppbv	12.9	30.00
Styrene	1.05	ppbv	-0.8	30.00
1,1,2,2-Tetrachloroethane	1.23	ppbv	17.0	30.00
o-Xylene	0.884	ppbv	-16.3	30.00
1,3,5-Trimethylbenzene	0.768	ppbv	-27.0	30.00
1,2,4-Trimethylbenzene	0.744	ppbv	-29.4	30.00
m-Dichlorobenzene	1.02	ppbv	-2.0	30.00
p-Dichlorobenzene	0.986	ppbv	-5.8	30.00
o-Dichlorobenzene	0.976	ppbv	-6.7	30.00
1,2,4-Trichlorobenzene	0.805	ppbv	-22.4	30.00
Hexachloro-1,3-butadiene	0.972	ppbv	-7.7	30.00

Sequence 2605034

Calibration Check (2605034-CCV1)

Prepared & Analyzed: 05/01/26

Dichlorodifluoromethane	0.961	ppbv	-6.5	30.00
Chloromethane	0.910	ppbv	-10.8	30.00
Dichlorotetrafluoroethane	0.937	ppbv	-9.5	30.00
Vinyl chloride	0.839	ppbv	-18.7	30.00
1,3-Butadiene	0.724	ppbv	-30.0	30.00
Ethylene oxide	0.953	ppbv	-0.3	30.00
Bromomethane	1.03	ppbv	0.2	30.00
Chloroethane	1.04	ppbv	0.5	30.00
Acetonitrile	1.06	ppbv	2.2	30.00
Acrolein	0.996	ppbv	-2.4	30.00
Trichlorofluoromethane	1.01	ppbv	0.02	30.00
Acrylonitrile	1.09	ppbv	-1.7	30.00
1,1-Dichloroethene	1.04	ppbv	0.2	30.00
Dichloromethane	1.00	ppbv	-3.3	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV1) Continued

Prepared & Analyzed: 05/01/26

Carbon Disulfide	1.15	ppbv	5.4	30.00
Trichlorotrifluoroethane	1.08	ppbv	4.4	30.00
trans-1,2-Dichloroethylene	0.980	ppbv	-5.8	30.00
1,1-Dichloroethane	1.15	ppbv	11.4	30.00
Methyl tert-Butyl Ether	0.880	ppbv	-15.0	30.00
Chloroprene	0.879	ppbv	-15.3	30.00
cis-1,2-Dichloroethylene	1.09	ppbv	5.2	30.00
Bromochloromethane	1.14	ppbv	11.3	30.00
Chloroform	1.19	ppbv	13.9	30.00
Ethyl tert-Butyl Ether	0.923	ppbv	-10.9	30.00
1,2-Dichloroethane	0.988	ppbv	-4.5	30.00
1,1,1-Trichloroethane	1.08	ppbv	4.9	30.00
Benzene	0.963	ppbv	-7.2	30.00
Carbon Tetrachloride	1.05	ppbv	2.4	30.00
tert-Amyl Methyl Ether	0.747	ppbv	-27.7	30.00
1,2-Dichloropropane	1.05	ppbv	0.7	30.00
Ethyl Acrylate	0.927	ppbv	-11.4	30.00
Bromodichloromethane	1.20	ppbv	15.9	30.00
Trichloroethylene	1.08	ppbv	4.9	30.00
Methyl Methacrylate	0.922	ppbv	-11.5	30.00
cis-1,3-Dichloropropene	0.888	ppbv	-14.9	30.00
Methyl Isobutyl Ketone	0.996	ppbv	-4.2	30.00
trans-1,3-Dichloropropene	0.983	ppbv	-6.4	30.00
1,1,2-Trichloroethane	1.16	ppbv	11.0	30.00
Toluene	0.997	ppbv	-4.2	30.00
Dibromochloromethane	1.17	ppbv	13.0	30.00
1,2-Dibromoethane	1.07	ppbv	2.9	30.00
n-Octane	0.924	ppbv	-11.0	30.00
Tetrachloroethylene	1.09	ppbv	4.0	30.00
Chlorobenzene	1.15	ppbv	9.2	30.00
Ethylbenzene	0.953	ppbv	-8.9	30.00
m,p-Xylene	2.00	ppbv	-4.4	30.00
Bromoform	1.15	ppbv	10.0	30.00
Styrene	1.00	ppbv	-5.0	30.00
1,1,2,2-Tetrachloroethane	1.21	ppbv	15.1	30.00
o-Xylene	0.907	ppbv	-14.1	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:27
SUBMITTED: 04/09/26 to 04/24/26
AQS SITE CODE: 40 109 1037
SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV1) Continued

Prepared & Analyzed: 05/01/26

1,3,5-Trimethylbenzene	0.826	ppbv	-21.4	30.00
1,2,4-Trimethylbenzene	0.815	ppbv	-22.7	30.00
m-Dichlorobenzene	1.07	ppbv	2.3	30.00
p-Dichlorobenzene	1.03	ppbv	-1.8	30.00
o-Dichlorobenzene	1.05	ppbv	0.6	30.00
1,2,4-Trichlorobenzene	1.00	ppbv	-3.2	30.00
Hexachloro-1,3-butadiene	1.23	ppbv	16.9	30.00

Calibration Check (2605034-CCV2)

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorodifluoromethane	1.07	ppbv	4.3	30.00
Chloromethane	1.03	ppbv	1.0	30.00
Dichlorotetrafluoroethane	1.06	ppbv	2.6	30.00
Vinyl chloride	0.925	ppbv	-10.4	30.00
1,3-Butadiene	0.745	ppbv	-27.9	30.00
Ethylene oxide	1.07	ppbv	12.3	30.00
Bromomethane	1.16	ppbv	12.3	30.00
Chloroethane	1.15	ppbv	11.3	30.00
Acetonitrile	1.25	ppbv	20.5	30.00
Acrolein	1.06	ppbv	3.8	30.00
Trichlorofluoromethane	1.08	ppbv	6.6	30.00
Acrylonitrile	1.24	ppbv	12.8	30.00
1,1-Dichloroethene	1.10	ppbv	6.0	30.00
Dichloromethane	1.16	ppbv	11.8	30.00
Carbon Disulfide	1.25	ppbv	15.3	30.00
Trichlorotrifluoroethane	1.18	ppbv	14.7	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.4	30.00
1,1-Dichloroethane	1.10	ppbv	6.5	30.00
Methyl tert-Butyl Ether	0.918	ppbv	-11.4	30.00
Chloroprene	0.801	ppbv	-22.9	30.00
cis-1,2-Dichloroethylene	0.966	ppbv	-6.8	30.00
Bromochloromethane	1.13	ppbv	9.9	30.00
Chloroform	1.21	ppbv	15.4	30.00
Ethyl tert-Butyl Ether	0.766	ppbv	-26.1	30.00
1,2-Dichloroethane	1.11	ppbv	7.3	30.00
1,1,1-Trichloroethane	1.12	ppbv	8.5	30.00
Benzene	1.03	ppbv	-1.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Carbon Tetrachloride	1.14	ppbv	10.9	30.00	
tert-Amyl Methyl Ether	0.752	ppbv	-27.3	30.00	
1,2-Dichloropropane	1.13	ppbv	8.7	30.00	
Ethyl Acrylate	0.931	ppbv	-11.0	30.00	
Bromodichloromethane	1.25	ppbv	20.6	30.00	
Trichloroethylene	1.10	ppbv	6.3	30.00	
Methyl Methacrylate	0.918	ppbv	-11.9	30.00	LK
cis-1,3-Dichloropropene	0.872	ppbv	-16.5	30.00	
Methyl Isobutyl Ketone	1.01	ppbv	-3.2	30.00	
trans-1,3-Dichloropropene	1.00	ppbv	-4.3	30.00	
1,1,2-Trichloroethane	1.21	ppbv	15.6	30.00	
Toluene	0.990	ppbv	-4.8	30.00	
Dibromochloromethane	1.18	ppbv	14.5	30.00	
1,2-Dibromoethane	1.07	ppbv	2.7	30.00	
n-Octane	0.930	ppbv	-10.4	30.00	
Tetrachloroethylene	1.10	ppbv	5.0	30.00	
Chlorobenzene	1.17	ppbv	11.0	30.00	
Ethylbenzene	0.936	ppbv	-10.5	30.00	
m,p-Xylene	1.98	ppbv	-5.3	30.00	
Bromoform	1.18	ppbv	12.5	30.00	
Styrene	0.983	ppbv	-7.1	30.00	
1,1,2,2-Tetrachloroethane	1.30	ppbv	23.2	30.00	
o-Xylene	0.876	ppbv	-17.0	30.00	
1,3,5-Trimethylbenzene	0.778	ppbv	-26.1	30.00	
1,2,4-Trimethylbenzene	0.768	ppbv	-27.1	30.00	
m-Dichlorobenzene	1.07	ppbv	2.8	30.00	
p-Dichlorobenzene	1.02	ppbv	-2.3	30.00	
o-Dichlorobenzene	1.04	ppbv	-0.6	30.00	
1,2,4-Trichlorobenzene	0.961	ppbv	-7.4	30.00	
Hexachloro-1,3-butadiene	1.26	ppbv	19.7	30.00	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:27

SUBMITTED: 04/09/26 to 04/24/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.