

LabNumber	Analysis	Sampled	Status	Project	SampleName	SampleType	SampleComments
6052726-01	Metals Analysis- HiVol 2026	4/2/2026 23:59	Invalid	OCOK	OCOK Q9551800 C1 VOID	Collocated - C1	Col 1 - Void values too low
6052726-08	Metals Analysis- HiVol 2026	4/2/2026 23:59	Invalid	OCOK	OCOK Q9551799 C2 VOID	Collocated - C2	Col 2 - Void values too low
6052726-02	Metals Analysis- HiVol 2026	4/8/2026 23:59	Invalid	OCOK	OCOK Q9551798 C1 VOID	Collocated - C1	Col 1 - Void failed cal after motor change
6052726-03	Metals Analysis- HiVol 2026	4/14/2026 23:59	Invalid	OCOK	OCOK Q9551796 C1 VOID	Collocated - C1	Col 1 - Void failed cal
6041536-01	TO-15 2026	4/2/2026 23:59	Invalid	OCOK	OCOK VOID	Field Sample	Void-Pre-Pressure issue
6041536-02	TO-15 2026	4/8/2026 23:59	Invalid	OCOK	OCOK VOID	Field Sample	Void-Pre-Pressure issue
6041536-03	TO-15 2026	4/8/2026 23:59	Invalid	OCOK	OCOK VOID	Field Sample	Void-Post-Flow rate zero error