



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

April 17, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: OCOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 03/04/26 12:07.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
OCOK	6030452-09	Air	02/27/26 23:59	03/04/26 12:07
OCOK FB	6030452-10	Air	02/27/26 23:59	03/04/26 12:07



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6030452-09

Sampled: 02/27/26 23:59

Temp Rec'd (°C): 16.7

Sample Volume: 1157.8 L

Received: 03/04/26 12:07

Cartridge Pouch #: 26G2640

Analyzed: 03/27/26 13:45

Comments: Reschedule from 2-25-26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.10	TT	0.0547
Acetaldehyde	2.57	TT	0.0226
Acetone	6.78	TT	0.0838
Propionaldehyde	0.692	TT	0.179
Crotonaldehyde	ND	TT, U	0.0164
Butyraldehyde	0.225	TT	0.00967
2-Butanone	1.13	TT	0.0364
Benzaldehyde	0.136	TT	0.0374
Valeraldehyde	ND	CE, TT, U	0.0105
Hexaldehyde	0.153	TT	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK FB

Lab ID: 6030452-10

Sampled: 02/27/26 23:59

Temp Rec'd (°C): 16.7

Sample Volume: 1157.8 L

Received: 03/04/26 12:07

Cartridge Pouch #: 26G2641

Analyzed: 03/27/26 14:17

Comments: Reschedule from 2-25-26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0944	TT	0.0547
Acetaldehyde	0.0310	TT	0.0226
Acetone	0.126	TT	0.0838
Propionaldehyde	0.0452	TT, U	0.179
Crotonaldehyde	ND	TT, U	0.0164
Butyraldehyde	ND	TT, U	0.00967
2-Butanone	0.0433	TT	0.0364
Benzaldehyde	0.0646	TT	0.0374
Valeraldehyde	ND	TT, U	0.0105
Hexaldehyde	ND	TT, U	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603100 - B6C0409

Calibration Blank (2603100-CCB1)

Prepared & Analyzed: 03/27/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2603100-CCB2)

Prepared & Analyzed: 03/27/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2603100-CCV1)

Prepared & Analyzed: 03/27/26

Formaldehyde	0.144	ug/ml	0.15000	96.0	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Acetone	0.151	ug/ml	0.15000	100	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.152	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.160	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1

Calibration Check (2603100-CCV2)

Prepared & Analyzed: 03/27/26

Formaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603100 - B6C0409

Calibration Check (2603100-CCV2) Contin

Prepared & Analyzed: 03/27/26

Acetone	0.149	ug/ml	0.15000		99.6	84.9-115.1
Propionaldehyde	0.152	ug/ml	0.15000		102	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000		101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000		97.3	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000		100	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000		102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000		100	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000		99.4	84.9-115.1

Calibration Check (2603100-CCV3)

Prepared & Analyzed: 03/27/26

Formaldehyde	0.145	ug/ml	0.15000		96.8	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000		99.4	84.9-115.1
Acetone	0.149	ug/ml	0.15000		99.5	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Crotonaldehyde	0.148	ug/ml	0.15000		98.8	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000		97.6	84.9-115.1
2-Butanone	0.152	ug/ml	0.15000		102	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000		102	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000		97.9	84.9-115.1

Batch B6C0409 - TO-11A extraction

Blank (B6C0409-BLK1)

Prepared: 03/05/26 Analyzed: 03/27/26

Formaldehyde	0.120	ug/m ³ Air				
Formaldehyde	0.120	ug/m ³ Air				
Acetaldehyde	0.0439	ug/m ³ Air				
Acetaldehyde	ND	ug/m ³ Air				U
Acetone	0.153	ug/m ³ Air				
Acetone	ND	ug/m ³ Air				U
Propionaldehyde	ND	ug/m ³ Air				U
Propionaldehyde	ND	ug/m ³ Air				U
Crotonaldehyde	ND	ug/m ³ Air				U
Crotonaldehyde	ND	ug/m ³ Air				U
Butyraldehyde	ND	ug/m ³ Air				U
Butyraldehyde	ND	ug/m ³ Air				U
2-Butanone	ND	ug/m ³ Air				U
2-Butanone	0.0501	ug/m ³ Air				
Benzaldehyde	0.0915	ug/m ³ Air				
Benzaldehyde	0.0915	ug/m ³ Air				

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C0409 - TO-11A extraction

Blank (B6C0409-BLK1) Continued

Prepared: 03/05/26 Analyzed: 03/27/26

Valeraldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6C0409-BLK3)

Prepared: 03/05/26 Analyzed: 03/27/26

Formaldehyde	ND	ug/m ³ Air							U
Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6C0409-BS1)

Prepared: 03/05/26 Analyzed: 03/27/26

Formaldehyde	0.304	ug/ml	0.30000	101	79.9-120.1
Formaldehyde	0.304	ug/ml	0.30000	101	79.9-120.1
Acetaldehyde	0.304	ug/ml	0.30000	101	79.9-120.1
Acetaldehyde	0.304	ug/ml	0.30000	101	79.9-120.1
Acetone	0.321	ug/ml	0.30000	107	69.9-130.1
Acetone	0.321	ug/ml	0.30000	107	69.9-130.1
Propionaldehyde	0.309	ug/ml	0.30000	103	69.9-130.1
Propionaldehyde	0.309	ug/ml	0.30000	103	69.9-130.1
Crotonaldehyde	0.296	ug/ml	0.30000	98.7	69.9-130.1
Crotonaldehyde	0.296	ug/ml	0.30000	98.7	69.9-130.1
Butyraldehyde	0.265	ug/ml	0.30000	88.2	69.9-130.1
Butyraldehyde	0.265	ug/ml	0.30000	88.2	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C0409 - TO-11A extraction

LCS (B6C0409-BS1) Continued

Prepared: 03/05/26 Analyzed: 03/27/26

2-Butanone	0.251	ug/ml	0.30000	83.7	69.9-130.1
2-Butanone	0.251	ug/ml	0.30000	83.7	69.9-130.1
Benzaldehyde	0.332	ug/ml	0.30000	111	69.9-130.1
Benzaldehyde	0.332	ug/ml	0.30000	111	69.9-130.1
Valeraldehyde	0.267	ug/ml	0.30000	89.0	69.9-130.1
Valeraldehyde	0.267	ug/ml	0.30000	89.0	69.9-130.1
Hexaldehyde	0.264	ug/ml	0.30000	88.2	69.9-130.1
Hexaldehyde	0.264	ug/ml	0.30000	88.2	69.9-130.1

LCS Dup (B6C0409-BSD1)

Prepared: 03/05/26 Analyzed: 03/27/26

Formaldehyde	0.325	ug/ml	0.30000	108	79.9-120.1	6.57	20
Formaldehyde	0.325	ug/ml	0.30000	108	79.9-120.1	6.57	20
Acetaldehyde	0.315	ug/ml	0.30000	105	79.9-120.1	3.51	20
Acetaldehyde	0.315	ug/ml	0.30000	105	79.9-120.1	3.51	20
Acetone	0.341	ug/ml	0.30000	114	69.9-130.1	6.02	20
Acetone	0.341	ug/ml	0.30000	114	69.9-130.1	6.02	20
Propionaldehyde	0.323	ug/ml	0.30000	108	69.9-130.1	4.56	20
Propionaldehyde	0.323	ug/ml	0.30000	108	69.9-130.1	4.56	20
Crotonaldehyde	0.305	ug/ml	0.30000	102	69.9-130.1	2.90	20
Crotonaldehyde	0.305	ug/ml	0.30000	102	69.9-130.1	2.90	20
Butyraldehyde	0.274	ug/ml	0.30000	91.4	69.9-130.1	3.60	20
Butyraldehyde	0.274	ug/ml	0.30000	91.4	69.9-130.1	3.60	20
2-Butanone	0.263	ug/ml	0.30000	87.6	69.9-130.1	4.60	20
2-Butanone	0.263	ug/ml	0.30000	87.6	69.9-130.1	4.60	20
Benzaldehyde	0.330	ug/ml	0.30000	110	69.9-130.1	0.505	20
Benzaldehyde	0.330	ug/ml	0.30000	110	69.9-130.1	0.505	20
Valeraldehyde	0.275	ug/ml	0.30000	91.7	69.9-130.1	3.00	20
Valeraldehyde	0.275	ug/ml	0.30000	91.7	69.9-130.1	3.00	20
Hexaldehyde	0.269	ug/ml	0.30000	89.6	69.9-130.1	1.67	20
Hexaldehyde	0.269	ug/ml	0.30000	89.6	69.9-130.1	1.67	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/17/26 09:28

SUBMITTED: 03/04/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.