



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

May 14, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: OCOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 03/05/26 13:13 through 04/01/26 15:01.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
OCOK	6030530-03	Air	03/03/26 23:59	03/05/26 13:13
OCOK D2	6031819-06	Air	03/15/26 23:59	03/18/26 11:37
OCOK	6031819-07	Air	03/09/26 23:59	03/18/26 11:37
OCOK	6032527-02	Air	03/21/26 23:59	03/25/26 15:09
OCOK	6040135-02	Air	03/27/26 23:59	04/01/26 15:01
OCOK FB	6040135-03	Air	03/27/26 23:59	04/01/26 15:01



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6030530-03

Sampled: 03/03/26 23:59

Temp Rec'd (°C): 6.4

Sample Volume: 1155.6 L

Received: 03/05/26 13:13

Cartridge Pouch #: 26G2642

Analyzed: 03/23/26 14:21

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.14		0.0548
Acetaldehyde	1.61		0.0227
Acetone	1.83		0.0839
Propionaldehyde	0.515		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.158		0.00969
2-Butanone	0.488		0.0365
Benzaldehyde	0.0757		0.0375
Valeraldehyde	0.150		0.0105
Hexaldehyde	0.116		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D2	Lab ID: 6031819-06	Sampled: 03/15/26 23:59
Temp Rec'd (°C): 1.0	Sample Volume: 1152.5 L	Received: 03/18/26 11:37
	Cartridge Pouch #: 26G2645	Analyzed: 03/26/26 14:21

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.70		0.0549
Acetaldehyde	1.18		0.0227
Acetone	2.17		0.0842
Propionaldehyde	0.336	QB-01	0.180
Crotonaldehyde	ND	U	0.0165
Butyraldehyde	0.0949		0.00972
2-Butanone	0.425		0.0366
Benzaldehyde	0.0860		0.0376
Valeraldehyde	0.0968		0.0105
Hexaldehyde	0.0545		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK	Lab ID: 6031819-07	Sampled: 03/09/26 23:59
Temp Rec'd (°C): 2.7	Sample Volume: 1158.3 L	Received: 03/18/26 11:37
	Cartridge Pouch #: 26G2643	Analyzed: 04/03/26 18:31

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.53		0.0546
Acetaldehyde	1.72		0.0226
Acetone	1.43		0.0837
Propionaldehyde	0.502	QB-01	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.176		0.00967
2-Butanone	0.463		0.0364
Benzaldehyde	0.111		0.0374
Valeraldehyde	0.201		0.0104
Hexaldehyde	0.130		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6032527-02

Sampled: 03/21/26 23:59

Temp Rec'd (°C): 7.6

Sample Volume: 1143.6 L

Received: 03/25/26 15:09

Cartridge Pouch #: 26G2646

Analyzed: 04/01/26 14:55

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	4.49	QB-01	0.0554
Acetaldehyde	4.00		0.0229
Acetone	5.13		0.0848
Propionaldehyde	1.25		0.181
Crotonaldehyde	0.0313		0.0166
Butyraldehyde	0.559		0.00979
2-Butanone	1.45		0.0369
Benzaldehyde	0.238		0.0379
Valeraldehyde	0.352		0.0106
Hexaldehyde	0.420		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6040135-02

Sampled: 03/27/26 23:59

Temp Rec'd (°C): 16.7

Sample Volume: 1143.8 L

Received: 04/01/26 15:01

Cartridge Pouch #: 26G2647

Analyzed: 04/03/26 03:24

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.57	TT	0.0553
Acetaldehyde	1.18	TT	0.0229
Acetone	2.58	TT	0.0848
Propionaldehyde	0.515	QB-01, TT	0.181
Crotonaldehyde	0.0295	TT	0.0166
Butyraldehyde	0.190	TT	0.00979
2-Butanone	0.410	TT	0.0369
Benzaldehyde	0.0983	TT	0.0379
Valeraldehyde	0.129	TT	0.0106
Hexaldehyde	0.129	TT	0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK FB

Lab ID: 6040135-03

Sampled: 03/27/26 23:59

Temp Rec'd (°C): 16.7

Sample Volume: 1143.8 L

Received: 04/01/26 15:01

Cartridge Pouch #: 26G2648

Analyzed: 04/03/26 03:41

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0409	TT, U	0.0553
Acetaldehyde	0.0261	TT	0.0229
Acetone	0.0456	TT, U	0.0848
Propionaldehyde	0.153	QB-01, TT, U	0.181
Crotonaldehyde	ND	TT, U	0.0166
Butyraldehyde	ND	TT, U	0.00979
2-Butanone	ND	TT, U	0.0369
Benzaldehyde	ND	TT, U	0.0379
Valeraldehyde	ND	TT, U	0.0106
Hexaldehyde	ND	TT, U	0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603077 - B6C1203

Calibration Blank (2603077-CCB1)

Prepared & Analyzed: 03/23/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2603077-CCB2)

Prepared & Analyzed: 03/23/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2603077-CCV1)

Prepared & Analyzed: 03/23/26

Formaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.151	ug/ml	0.15000	101	84.9-115.1
Propionaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	97.0	84.9-115.1

Calibration Check (2603077-CCV2)

Prepared & Analyzed: 03/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603077 - B6C1203

Calibration Check (2603077-CCV2) Contin

Prepared & Analyzed: 03/23/26

Acetone	0.150	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.4	84.9-115.1
Benzaldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1

Calibration Check (2603077-CCV3)

Prepared & Analyzed: 03/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Batch 2603096 - B6C2604

Calibration Blank (2603096-CCB1)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2603096-CCB2)

Prepared: 03/26/26 Analyzed: 03/27/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603096 - B6C2604

Calibration Blank (2603096-CCB2) Contin

Prepared: 03/26/26 Analyzed: 03/27/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2603096-CCV1)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Benzaldehyde	0.162	ug/ml	0.15000	108	84.9-115.1
Valeraldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2603096-CCV2)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.145	ug/ml	0.15000	97.0	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.1	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.8	84.9-115.1
Benzaldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2603096-CCV3)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.2	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603096 - B6C2604

Calibration Check (2603096-CCV3) Contin

Prepared & Analyzed: 03/26/26

Hexaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
-------------	-------	-------	---------	------	------------

Calibration Check (2603096-CCV4)

Prepared: 03/26/26 Analyzed: 03/27/26

Formaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.5	84.9-115.1
Benzaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1

Batch 2604004 - B6C3112

Calibration Blank (2604004-CCB1)

Prepared & Analyzed: 04/01/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2604004-CCB2)

Prepared & Analyzed: 04/01/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2604004-CCV1)

Prepared & Analyzed: 04/01/26

Formaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604004 - B6C3112

Calibration Check (2604004-CCV1) Contin

Prepared & Analyzed: 04/01/26

Acetaldehyde	0.146	ug/ml	0.15000		97.6	84.9-115.1
Acetone	0.150	ug/ml	0.15000		99.9	84.9-115.1
Propionaldehyde	0.154	ug/ml	0.15000		103	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000		102	84.9-115.1
Butyraldehyde	0.149	ug/ml	0.15000		99.4	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000		98.7	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000		102	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000		97.5	84.9-115.1

Calibration Check (2604004-CCV2)

Prepared & Analyzed: 04/01/26

Formaldehyde	0.146	ug/ml	0.15000		97.5	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000		98.6	84.9-115.1
Acetone	0.147	ug/ml	0.15000		98.0	84.9-115.1
Propionaldehyde	0.155	ug/ml	0.15000		103	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000		103	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000		98.4	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000		99.9	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000		104	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000		99.5	84.9-115.1

Calibration Check (2604004-CCV3)

Prepared & Analyzed: 04/01/26

Formaldehyde	0.146	ug/ml	0.15000		97.6	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000		97.9	84.9-115.1
Acetone	0.149	ug/ml	0.15000		99.4	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000		101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000		97.5	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000		98.8	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1
Valeraldehyde	0.154	ug/ml	0.15000		102	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1

Batch 2604008 - B6D0204

Calibration Blank (2604008-CCB1)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604008 - B6D0204

Calibration Blank (2604008-CCB1) Contin

Prepared & Analyzed: 04/02/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2604008-CCB2)

Prepared: 04/02/26 Analyzed: 04/03/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604008-CCV1)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.151	ug/ml	0.15000		100	84.9-115.1			
Acetaldehyde	0.151	ug/ml	0.15000		100	84.9-115.1			
Acetone	0.152	ug/ml	0.15000		102	84.9-115.1			
Propionaldehyde	0.159	ug/ml	0.15000		106	84.9-115.1			
Crotonaldehyde	0.155	ug/ml	0.15000		103	84.9-115.1			
Butyraldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			
2-Butanone	0.151	ug/ml	0.15000		101	84.9-115.1			
Benzaldehyde	0.163	ug/ml	0.15000		108	84.9-115.1			
Valeraldehyde	0.158	ug/ml	0.15000		106	84.9-115.1			
Hexaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			

Calibration Check (2604008-CCV2)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.150	ug/ml	0.15000		99.9	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			
Acetone	0.152	ug/ml	0.15000		102	84.9-115.1			
Propionaldehyde	0.156	ug/ml	0.15000		104	84.9-115.1			
Crotonaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1			
Butyraldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			
2-Butanone	0.151	ug/ml	0.15000		101	84.9-115.1			
Benzaldehyde	0.161	ug/ml	0.15000		107	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604008 - B6D0204

Calibration Check (2604008-CCV2) Contin

Prepared & Analyzed: 04/02/26

Valeraldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1

Calibration Check (2604008-CCV3)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Acetaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Acetone	0.152	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.157	ug/ml	0.15000	104	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.8	84.9-115.1
Benzaldehyde	0.161	ug/ml	0.15000	107	84.9-115.1
Valeraldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Hexaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1

Calibration Check (2604008-CCV4)

Prepared: 04/02/26 Analyzed: 04/03/26

Formaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Acetaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetone	0.152	ug/ml	0.15000	101	84.9-115.1
Propionaldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
2-Butanone	0.153	ug/ml	0.15000	102	84.9-115.1
Benzaldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.160	ug/ml	0.15000	107	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1

Calibration Check (2604008-CCV5)

Prepared: 04/02/26 Analyzed: 04/03/26

Formaldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Acetone	0.152	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1

Batch 2604013 - B6C1904

Calibration Blank (2604013-CCB1)

Prepared & Analyzed: 04/03/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604013 - B6C1904

Calibration Blank (2604013-CCB1) Contin

Prepared & Analyzed: 04/03/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2604013-CCB2)

Prepared & Analyzed: 04/03/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604013-CCV1)

Prepared & Analyzed: 04/03/26

Formaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Acetone	0.153	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Crotonaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.5	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1

Calibration Check (2604013-CCV2)

Prepared & Analyzed: 04/03/26

Formaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetone	0.153	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604013 - B6C1904

Calibration Check (2604013-CCV2) Contin

Prepared & Analyzed: 04/03/26

Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.161	ug/ml	0.15000	107	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Hexaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1

Calibration Check (2604013-CCV3)

Prepared & Analyzed: 04/03/26

Formaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Acetone	0.155	ug/ml	0.15000	103	84.9-115.1
Propionaldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.163	ug/ml	0.15000	109	84.9-115.1
Valeraldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1

Batch B6C1203 - TO-11A extraction

Blank (B6C1203-BLK1)

Prepared: 03/12/26 Analyzed: 03/23/26

Formaldehyde	ND	ug/m ³ Air			U
Acetaldehyde	0.0384	ug/m ³ Air			
Acetone	ND	ug/m ³ Air			U
Propionaldehyde	ND	ug/m ³ Air			U
Crotonaldehyde	ND	ug/m ³ Air			U
Butyraldehyde	ND	ug/m ³ Air			U
2-Butanone	ND	ug/m ³ Air			U
Benzaldehyde	0.0513	ug/m ³ Air			
Valeraldehyde	ND	ug/m ³ Air			U
Hexaldehyde	ND	ug/m ³ Air			U

LCS (B6C1203-BS1)

Prepared: 03/12/26 Analyzed: 03/23/26

Formaldehyde	0.295	ug/ml	0.30000	98.2	79.9-120.1
Acetaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1
Acetone	0.304	ug/ml	0.30000	101	69.9-130.1
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1
Crotonaldehyde	0.293	ug/ml	0.30000	97.8	69.9-130.1
Butyraldehyde	0.258	ug/ml	0.30000	86.0	69.9-130.1
2-Butanone	0.239	ug/ml	0.30000	79.7	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C1203 - TO-11A extraction

LCS (B6C1203-BS1) Continued

Prepared: 03/12/26 Analyzed: 03/23/26

Benzaldehyde	0.327	ug/ml	0.30000	109	69.9-130.1
Valeraldehyde	0.265	ug/ml	0.30000	88.3	69.9-130.1
Hexaldehyde	0.256	ug/ml	0.30000	85.3	69.9-130.1

LCS Dup (B6C1203-BSD1)

Prepared: 03/12/26 Analyzed: 03/23/26

Formaldehyde	0.298	ug/ml	0.30000	99.5	79.9-120.1	1.23	20
Acetaldehyde	0.298	ug/ml	0.30000	99.4	79.9-120.1	1.46	20
Acetone	0.307	ug/ml	0.30000	102	69.9-130.1	1.13	20
Propionaldehyde	0.354	ug/ml	0.30000	118	69.9-130.1	0.806	20
Crotonaldehyde	0.294	ug/ml	0.30000	98.1	69.9-130.1	0.337	20
Butyraldehyde	0.264	ug/ml	0.30000	88.0	69.9-130.1	2.29	20
2-Butanone	0.243	ug/ml	0.30000	80.9	69.9-130.1	1.52	20
Benzaldehyde	0.327	ug/ml	0.30000	109	69.9-130.1	0.233	20
Valeraldehyde	0.271	ug/ml	0.30000	90.3	69.9-130.1	2.22	20
Hexaldehyde	0.261	ug/ml	0.30000	86.9	69.9-130.1	1.81	20

Batch B6C1904 - TO-11A extraction

Blank (B6C1904-BLK1)

Prepared: 03/19/26 Analyzed: 04/03/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0302	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	0.0624	ug/m ³ Air					
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6C1904-BS1)

Prepared: 03/19/26 Analyzed: 04/03/26

Formaldehyde	0.293	ug/ml	0.30000	97.7	79.9-120.1
Acetaldehyde	0.293	ug/ml	0.30000	97.7	79.9-120.1
Acetone	0.305	ug/ml	0.30000	102	69.9-130.1
Propionaldehyde	0.356	ug/ml	0.30000	119	69.9-130.1
Crotonaldehyde	0.287	ug/ml	0.30000	95.8	69.9-130.1
Butyraldehyde	0.262	ug/ml	0.30000	87.3	69.9-130.1
2-Butanone	0.243	ug/ml	0.30000	80.9	69.9-130.1
Benzaldehyde	0.325	ug/ml	0.30000	108	69.9-130.1
Valeraldehyde	0.269	ug/ml	0.30000	89.8	69.9-130.1
Hexaldehyde	0.256	ug/ml	0.30000	85.3	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C1904 - TO-11A extraction

LCS Dup (B6C1904-BSD1)

Prepared: 03/19/26 Analyzed: 04/03/26

Formaldehyde	0.305	ug/ml	0.30000	102	79.9-120.1	3.88	20
Acetaldehyde	0.303	ug/ml	0.30000	101	79.9-120.1	3.36	20
Acetone	0.316	ug/ml	0.30000	105	69.9-130.1	3.53	20
Propionaldehyde	0.366	ug/ml	0.30000	122	69.9-130.1	2.76	20
Crotonaldehyde	0.294	ug/ml	0.30000	98.2	69.9-130.1	2.41	20
Butyraldehyde	0.273	ug/ml	0.30000	91.2	69.9-130.1	4.34	20
2-Butanone	0.246	ug/ml	0.30000	81.9	69.9-130.1	1.23	20
Benzaldehyde	0.334	ug/ml	0.30000	111	69.9-130.1	2.69	20
Valeraldehyde	0.275	ug/ml	0.30000	91.8	69.9-130.1	2.21	20
Hexaldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1	1.95	20

Batch B6C2604 - TO-11A extraction

Blank (B6C2604-BLK1)

Prepared & Analyzed: 03/26/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0303	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6C2604-BS1)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.292	ug/ml	0.30000	97.3	79.9-120.1
Acetaldehyde	0.290	ug/ml	0.30000	96.8	79.9-120.1
Acetone	0.295	ug/ml	0.30000	98.3	69.9-130.1
Propionaldehyde	0.365	ug/ml	0.30000	122	69.9-130.1
Crotonaldehyde	0.286	ug/ml	0.30000	95.2	69.9-130.1
Butyraldehyde	0.258	ug/ml	0.30000	85.8	69.9-130.1
2-Butanone	0.234	ug/ml	0.30000	78.1	69.9-130.1
Benzaldehyde	0.317	ug/ml	0.30000	106	69.9-130.1
Valeraldehyde	0.266	ug/ml	0.30000	88.5	69.9-130.1
Hexaldehyde	0.250	ug/ml	0.30000	83.4	69.9-130.1

LCS Dup (B6C2604-BSD1)

Prepared & Analyzed: 03/26/26

Formaldehyde	0.292	ug/ml	0.30000	97.3	79.9-120.1	0.0241	20
Acetaldehyde	0.292	ug/ml	0.30000	97.4	79.9-120.1	0.694	20
Acetone	0.296	ug/ml	0.30000	98.5	69.9-130.1	0.222	20

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C2604 - TO-11A extraction

LCS Dup (B6C2604-BSD1) Continued

Prepared & Analyzed: 03/26/26

Propionaldehyde	0.372	ug/ml	0.30000	124	69.9-130.1	1.75	20	
Crotonaldehyde	0.288	ug/ml	0.30000	96.0	69.9-130.1	0.855	20	
Butyraldehyde	0.262	ug/ml	0.30000	87.2	69.9-130.1	1.55	20	
2-Butanone	0.236	ug/ml	0.30000	78.8	69.9-130.1	0.870	20	
Benzaldehyde	0.319	ug/ml	0.30000	106	69.9-130.1	0.743	20	
Valeraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1	1.65	20	
Hexaldehyde	0.254	ug/ml	0.30000	84.6	69.9-130.1	1.47	20	

Duplicate (B6C2604-DUP1)

Source: 6031819-06 Prepared & Analyzed: 03/26/26

Formaldehyde	1.70	ug/m ³ Air	1.70	0.312	10.1	
Acetaldehyde	1.18	ug/m ³ Air	1.18	0.0916	10.1	
Acetone	2.18	ug/m ³ Air	2.17	0.215	10.1	
Propionaldehyde	0.342	ug/m ³ Air	0.336	1.94	10.1	QB-01
Crotonaldehyde	ND	ug/m ³ Air	ND		10.1	U
Butyraldehyde	0.0927	ug/m ³ Air	0.0949	2.33	10.1	
2-Butanone	0.427	ug/m ³ Air	0.425	0.604	10.1	
Benzaldehyde	0.0653	ug/m ³ Air	0.0860	27.4	10.1	
Valeraldehyde	0.0841	ug/m ³ Air	0.0968	14.1	10.1	
Hexaldehyde	0.0568	ug/m ³ Air	0.0545	4.23	10.1	

Batch B6C3112 - TO-11A extraction

Blank (B6C3112-BLK1)

Prepared: 03/31/26 Analyzed: 04/01/26

Formaldehyde	ND	ug/m ³ Air				U
Acetaldehyde	0.0313	ug/m ³ Air				
Acetone	ND	ug/m ³ Air				U
Propionaldehyde	ND	ug/m ³ Air				QB-01, U
Crotonaldehyde	ND	ug/m ³ Air				U
Butyraldehyde	ND	ug/m ³ Air				U
2-Butanone	ND	ug/m ³ Air				U
Benzaldehyde	ND	ug/m ³ Air				U
Valeraldehyde	0.0257	ug/m ³ Air				
Hexaldehyde	ND	ug/m ³ Air				U

LCS (B6C3112-BS1)

Prepared: 03/31/26 Analyzed: 04/01/26

Formaldehyde	0.294	ug/ml	0.30000	98.0	79.9-120.1
Acetaldehyde	0.293	ug/ml	0.30000	97.8	79.9-120.1
Acetone	0.300	ug/ml	0.30000	100	69.9-130.1
Propionaldehyde	0.365	ug/ml	0.30000	122	69.9-130.1
Crotonaldehyde	0.288	ug/ml	0.30000	95.9	69.9-130.1
Butyraldehyde	0.260	ug/ml	0.30000	86.7	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C3112 - TO-11A extraction

LCS (B6C3112-BS1) Continued

Prepared: 03/31/26 Analyzed: 04/01/26

2-Butanone	0.238	ug/ml	0.30000	79.3	69.9-130.1
Benzaldehyde	0.321	ug/ml	0.30000	107	69.9-130.1
Valeraldehyde	0.265	ug/ml	0.30000	88.3	69.9-130.1
Hexaldehyde	0.253	ug/ml	0.30000	84.4	69.9-130.1

LCS Dup (B6C3112-BSD1)

Prepared: 03/31/26 Analyzed: 04/01/26

Formaldehyde	0.293	ug/ml	0.30000	97.7	79.9-120.1	0.293	20
Acetaldehyde	0.293	ug/ml	0.30000	97.8	79.9-120.1	0.00409	20
Acetone	0.301	ug/ml	0.30000	100	69.9-130.1	0.0993	20
Propionaldehyde	0.365	ug/ml	0.30000	122	69.9-130.1	0.197	20
Crotonaldehyde	0.288	ug/ml	0.30000	95.9	69.9-130.1	0.00854	20
Butyraldehyde	0.260	ug/ml	0.30000	86.8	69.9-130.1	0.0560	20
2-Butanone	0.239	ug/ml	0.30000	79.7	69.9-130.1	0.432	20
Benzaldehyde	0.316	ug/ml	0.30000	105	69.9-130.1	1.39	20
Valeraldehyde	0.268	ug/ml	0.30000	89.2	69.9-130.1	0.991	20
Hexaldehyde	0.253	ug/ml	0.30000	84.4	69.9-130.1	0.102	20

Batch B6D0204 - TO-11A extraction

Blank (B6D0204-BLK1)

Prepared & Analyzed: 04/02/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6D0204-BS1)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.299	ug/ml	0.30000	99.6	79.9-120.1
Acetaldehyde	0.299	ug/ml	0.30000	99.7	79.9-120.1
Acetone	0.302	ug/ml	0.30000	101	69.9-130.1
Propionaldehyde	0.359	ug/ml	0.30000	120	69.9-130.1
Crotonaldehyde	0.295	ug/ml	0.30000	98.2	69.9-130.1
Butyraldehyde	0.264	ug/ml	0.30000	87.9	69.9-130.1
2-Butanone	0.237	ug/ml	0.30000	78.9	69.9-130.1
Benzaldehyde	0.323	ug/ml	0.30000	108	69.9-130.1
Valeraldehyde	0.269	ug/ml	0.30000	89.6	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D0204 - TO-11A extraction

LCS (B6D0204-BS1) Continued

Prepared & Analyzed: 04/02/26

Hexaldehyde	0.257	ug/ml	0.30000	85.7	69.9-130.1		
-------------	-------	-------	---------	------	------------	--	--

LCS Dup (B6D0204-BSD1)

Prepared & Analyzed: 04/02/26

Formaldehyde	0.300	ug/ml	0.30000	100	79.9-120.1	0.410	20
Acetaldehyde	0.298	ug/ml	0.30000	99.4	79.9-120.1	0.295	20
Acetone	0.305	ug/ml	0.30000	102	69.9-130.1	0.986	20
Propionaldehyde	0.367	ug/ml	0.30000	122	69.9-130.1	2.10	20
Crotonaldehyde	0.293	ug/ml	0.30000	97.8	69.9-130.1	0.408	20
Butyraldehyde	0.269	ug/ml	0.30000	89.6	69.9-130.1	1.89	20
2-Butanone	0.241	ug/ml	0.30000	80.5	69.9-130.1	1.95	20
Benzaldehyde	0.321	ug/ml	0.30000	107	69.9-130.1	0.390	20
Valeraldehyde	0.270	ug/ml	0.30000	90.1	69.9-130.1	0.642	20
Hexaldehyde	0.259	ug/ml	0.30000	86.2	69.9-130.1	0.562	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 09:58

SUBMITTED: 03/05/26 to 04/01/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
QB-01	Analyte exceeds method blank criteria
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.