



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 08, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: OCOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/06/26 12:20 through 06/03/26 13:39.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
OCOK D1	6050635-03	Air	05/02/26 23:59	05/06/26 12:20
OCOK D2	6050635-04	Air	05/02/26 23:59	05/06/26 12:20
OCOK	6051940-03	Air	05/08/26 23:59	05/19/26 10:56
OCOK D1	6051940-04	Air	05/14/26 23:59	05/19/26 10:56
OCOK D2	6051940-05	Air	05/14/26 23:59	05/19/26 10:56
OCOK	6060309-01	Air	05/20/26 23:59	06/03/26 13:39
OCOK	6060309-02	Air	05/26/26 23:59	06/03/26 13:39
OCOK FB	6060309-03	Air	05/26/26 23:59	06/03/26 13:39



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D1	Lab ID: 6050635-03	Sampled: 05/02/26 23:59
Temp Rec'd (°C): 0.5	Sample Volume: 1142.2 L	Received: 05/06/26 12:20
	Cartridge Pouch #: 26G2655	Analyzed: 05/11/26 15:16

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.32		0.0554
Acetaldehyde	1.82		0.0229
Acetone	4.70		0.0849
Propionaldehyde	0.686		0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.248		0.00981
2-Butanone	0.691		0.0369
Benzaldehyde	0.107		0.0379
Valeraldehyde	0.122		0.0106
Hexaldehyde	0.127		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D2	Lab ID: 6050635-04	Sampled: 05/02/26 23:59
Temp Rec'd (°C): 0.5	Sample Volume: 1141.1 L	Received: 05/06/26 12:20
	Cartridge Pouch #: 26M0033	Analyzed: 05/11/26 16:05
Comments: Duplicate Makeup from 3-15-26		

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.30		0.0555
Acetaldehyde	1.81		0.0230
Acetone	4.60		0.0850
Propionaldehyde	0.670		0.181
Crotonaldehyde	ND	U	0.0167
Butyraldehyde	0.249		0.00982
2-Butanone	0.691		0.0370
Benzaldehyde	0.102		0.0379
Valeraldehyde	0.127		0.0106
Hexaldehyde	0.130		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK	Lab ID: 6051940-03	Sampled: 05/08/26 23:59
Temp Rec'd (°C): 7.6	Sample Volume: 1142.7 L	Received: 05/19/26 10:56
	Cartridge Pouch #: 26G2656	Analyzed: 05/20/26 19:01

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.55		0.0554
Acetaldehyde	2.12		0.0229
Acetone	1.70		0.0849
Propionaldehyde	0.573		0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.268		0.00980
2-Butanone	0.337		0.0369
Benzaldehyde	0.136		0.0379
Valeraldehyde	0.299		0.0106
Hexaldehyde	0.251		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D1	Lab ID: 6051940-04	Sampled: 05/14/26 23:59
Temp Rec'd (°C): 7.6	Sample Volume: 1140.2 L	Received: 05/19/26 10:56
	Cartridge Pouch #: 26G2657	Analyzed: 05/22/26 18:00

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	6.25		0.0555
Acetaldehyde	2.27		0.0230
Acetone	2.48		0.0851
Propionaldehyde	0.548		0.182
Crotonaldehyde	ND	U	0.0167
Butyraldehyde	0.231		0.00982
2-Butanone	0.445		0.0370
Benzaldehyde	0.115		0.0380
Valeraldehyde	0.243		0.0106
Hexaldehyde	0.154		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D2	Lab ID: 6051940-05	Sampled: 05/14/26 23:59
Temp Rec'd (°C): 7.6	Sample Volume: 1139.4 L	Received: 05/19/26 10:56
	Cartridge Pouch #: 26G2658	Analyzed: 05/22/26 18:49

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	6.34		0.0556
Acetaldehyde	2.31		0.0230
Acetone	2.53		0.0851
Propionaldehyde	0.564		0.182
Crotonaldehyde	ND	U	0.0167
Butyraldehyde	0.242		0.00983
2-Butanone	0.453		0.0370
Benzaldehyde	0.105		0.0380
Valeraldehyde	0.257		0.0106
Hexaldehyde	0.171		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK	Lab ID: 6060309-01	Sampled: 05/20/26 23:59
Temp Rec'd (°C): 9.0	Sample Volume: 1145.8 L	Received: 06/03/26 13:39
	Cartridge Pouch #: 26G2659	Analyzed: 06/04/26 21:11

Comments: Recd past extraction HT

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.01	S-Rec	0.0552
Acetaldehyde	1.01	S-Rec	0.0229
Acetone	0.854	S-Rec	0.0847
Propionaldehyde	0.164	U, S-Rec	0.181
Crotonaldehyde	ND	U, S-Rec	0.0166
Butyraldehyde	0.0795	S-Rec	0.00977
2-Butanone	0.121	S-Rec	0.0368
Benzaldehyde	0.0762	S-Rec	0.0378
Valeraldehyde	0.0925	S-Rec	0.0106
Hexaldehyde	0.0668	S-Rec	0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6060309-02

Sampled: 05/26/26 23:59

Temp Rec'd (°C): 9.0

Sample Volume: 1144.2 L

Received: 06/03/26 13:39

Cartridge Pouch #: 26G2660

Analyzed: 06/04/26 21:28

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	4.18		0.0553
Acetaldehyde	1.43		0.0229
Acetone	1.09		0.0848
Propionaldehyde	0.315		0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.172		0.00979
2-Butanone	0.231		0.0369
Benzaldehyde	0.125		0.0378
Valeraldehyde	0.167		0.0106
Hexaldehyde	0.135		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK FB

Lab ID: 6060309-03

Sampled: 05/26/26 23:59

Temp Rec'd (°C): 9.0

Sample Volume: 1144.2 L

Received: 06/03/26 13:39

Cartridge Pouch #: 26G2661

Analyzed: 06/04/26 21:44

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0336	U	0.0553
Acetaldehyde	0.0305		0.0229
Acetone	0.0649	U	0.0848
Propionaldehyde	0.0445	U	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	ND	U	0.00979
2-Butanone	ND	U	0.0369
Benzaldehyde	0.0418		0.0378
Valeraldehyde	ND	U	0.0106
Hexaldehyde	ND	U	0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Blank (2605031-CCB1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2605031-CCB2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605031-CCV1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.143	ug/ml	0.15000		95.4	84.9-115.1			
Acetaldehyde	0.146	ug/ml	0.15000		97.3	84.9-115.1			
Acetone	0.147	ug/ml	0.15000		97.7	84.9-115.1			
Propionaldehyde	0.149	ug/ml	0.15000		99.4	84.9-115.1			
Crotonaldehyde	0.149	ug/ml	0.15000		99.6	84.9-115.1			
Butyraldehyde	0.140	ug/ml	0.15000		93.4	84.9-115.1			
2-Butanone	0.145	ug/ml	0.15000		96.7	84.9-115.1			
Benzaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1			
Valeraldehyde	0.146	ug/ml	0.15000		97.1	84.9-115.1			
Hexaldehyde	0.145	ug/ml	0.15000		96.5	84.9-115.1			

Calibration Check (2605031-CCV2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000		96.1	84.9-115.1			
Acetaldehyde	0.146	ug/ml	0.15000		97.5	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Check (2605031-CCV2) Contin

Prepared & Analyzed: 05/11/26

Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1

Calibration Check (2605031-CCV3)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.6	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	95.7	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605069-CCB2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB2) Contin

Prepared & Analyzed: 05/20/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605069-CCV1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.9	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
2-Butanone	0.154	ug/ml	0.15000	102	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1

Calibration Check (2605069-CCV2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Calibration Check (2605069-CCV3)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Check (2605069-CCV3) Contin

Prepared & Analyzed: 05/20/26

Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
-------------	-------	-------	---------	------	------------

Batch 2605076 - B6E2205

Calibration Blank (2605076-CCB1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605076-CCB2)

Prepared: 05/22/26 Analyzed: 05/23/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605076-CCV1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.4	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.8	84.9-115.1
Crotonaldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.8	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.3	84.9-115.1
Benzaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Valeraldehyde	0.142	ug/ml	0.15000	94.8	84.9-115.1
Hexaldehyde	0.140	ug/ml	0.15000	93.4	84.9-115.1

Calibration Check (2605076-CCV2)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605076 - B6E2205

Calibration Check (2605076-CCV2) Contin

Prepared & Analyzed: 05/22/26

Acetaldehyde	0.146	ug/ml	0.15000	97.0	84.9-115.1
Acetone	0.144	ug/ml	0.15000	96.1	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.139	ug/ml	0.15000	92.7	84.9-115.1
2-Butanone	0.143	ug/ml	0.15000	95.0	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1

Calibration Check (2605076-CCV3)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.142	ug/ml	0.15000	95.0	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Propionaldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.9	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	96.0	84.9-115.1
Benzaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Hexaldehyde	0.139	ug/ml	0.15000	92.8	84.9-115.1

Calibration Check (2605076-CCV4)

Prepared: 05/22/26 Analyzed: 05/23/26

Formaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Propionaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	94.1	84.9-115.1
2-Butanone	0.142	ug/ml	0.15000	94.9	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Valeraldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1
Hexaldehyde	0.141	ug/ml	0.15000	94.0	84.9-115.1

Batch 2606019 - B6F0403

Calibration Blank (2606019-CCB1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606019 - B6F0403

Calibration Blank (2606019-CCB1) Contin

Prepared & Analyzed: 06/04/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2606019-CCB2)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2606019-CCV1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.6	84.9-115.1
Benzaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1

Calibration Check (2606019-CCV2)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.8	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.8	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606019 - B6F0403

Calibration Check (2606019-CCV2) Contin

Prepared & Analyzed: 06/04/26

Valeraldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1

Calibration Check (2606019-CCV3)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1

Batch B6E1108 - TO-11A extraction

Blank (B6E1108-BLK1)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air				U
Acetaldehyde	0.0325	ug/m ³ Air				
Acetone	ND	ug/m ³ Air				U
Propionaldehyde	ND	ug/m ³ Air				CE, U
Crotonaldehyde	ND	ug/m ³ Air				U
Butyraldehyde	ND	ug/m ³ Air				U
2-Butanone	ND	ug/m ³ Air				U
Benzaldehyde	0.0448	ug/m ³ Air				
Valeraldehyde	ND	ug/m ³ Air				U
Hexaldehyde	ND	ug/m ³ Air				U

LCS (B6E1108-BS1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.288	ug/ml	0.30000	95.9	79.9-120.1
Acetaldehyde	0.293	ug/ml	0.30000	97.6	79.9-120.1
Acetone	0.294	ug/ml	0.30000	98.0	69.9-130.1
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1
Crotonaldehyde	0.276	ug/ml	0.30000	92.0	69.9-130.1
Butyraldehyde	0.263	ug/ml	0.30000	87.6	69.9-130.1
2-Butanone	0.238	ug/ml	0.30000	79.4	69.9-130.1
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1
Valeraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1
Hexaldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1

LCS Dup (B6E1108-BSD1)

Prepared & Analyzed: 05/11/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1108 - TO-11A extraction

LCS Dup (B6E1108-BSD1) Continued

Prepared & Analyzed: 05/11/26

Formaldehyde	0.290	ug/ml	0.30000		96.6	79.9-120.1	0.660	20	
Acetaldehyde	0.294	ug/ml	0.30000		97.9	79.9-120.1	0.305	20	
Acetone	0.293	ug/ml	0.30000		97.5	69.9-130.1	0.439	20	
Propionaldehyde	0.348	ug/ml	0.30000		116	69.9-130.1	0.877	20	
Crotonaldehyde	0.271	ug/ml	0.30000		90.4	69.9-130.1	1.74	20	
Butyraldehyde	0.259	ug/ml	0.30000		86.2	69.9-130.1	1.60	20	
2-Butanone	0.240	ug/ml	0.30000		80.1	69.9-130.1	0.778	20	
Benzaldehyde	0.315	ug/ml	0.30000		105	69.9-130.1	0.358	20	
Valeraldehyde	0.255	ug/ml	0.30000		84.9	69.9-130.1	2.43	20	
Hexaldehyde	0.249	ug/ml	0.30000		83.1	69.9-130.1	2.06	20	

Duplicate (B6E1108-DUP1)

Source: 6050635-03 Prepared & Analyzed: 05/11/26

Formaldehyde	3.32	ug/m ³ Air		3.32			0.0115	10.1	
Acetaldehyde	1.82	ug/m ³ Air		1.82			0.289	10.1	
Acetone	4.69	ug/m ³ Air		4.70			0.113	10.1	
Propionaldehyde	0.693	ug/m ³ Air		0.686			0.942	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.250	ug/m ³ Air		0.248			1.03	10.1	
2-Butanone	0.698	ug/m ³ Air		0.691			0.955	10.1	
Benzaldehyde	0.109	ug/m ³ Air		0.107			2.36	10.1	
Valeraldehyde	0.126	ug/m ³ Air		0.122			2.82	10.1	
Hexaldehyde	0.120	ug/m ³ Air		0.127			5.82	10.1	

Duplicate (B6E1108-DUP2)

Source: 6050635-04 Prepared & Analyzed: 05/11/26

Formaldehyde	3.29	ug/m ³ Air		3.30			0.451	10.1	
Acetaldehyde	1.80	ug/m ³ Air		1.81			0.303	10.1	
Acetone	4.60	ug/m ³ Air		4.60			0.0396	10.1	
Propionaldehyde	0.681	ug/m ³ Air		0.670			1.62	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.249	ug/m ³ Air		0.249			0.119	10.1	
2-Butanone	0.691	ug/m ³ Air		0.691			0.00172	10.1	
Benzaldehyde	0.111	ug/m ³ Air		0.102			8.51	10.1	
Valeraldehyde	0.117	ug/m ³ Air		0.127			7.57	10.1	
Hexaldehyde	0.119	ug/m ³ Air		0.130			8.82	10.1	

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1)

Prepared: 05/20/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1) Continued

Prepared: 05/20/26 Analyzed: 05/22/26

Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E2006-BS1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.298	ug/ml	0.30000	99.2	79.9-120.1		
Acetaldehyde	0.289	ug/ml	0.30000	96.2	79.9-120.1		
Acetone	0.296	ug/ml	0.30000	98.8	69.9-130.1		
Propionaldehyde	0.269	ug/ml	0.30000	89.6	69.9-130.1		
Crotonaldehyde	0.291	ug/ml	0.30000	97.1	69.9-130.1		
Butyraldehyde	0.257	ug/ml	0.30000	85.5	69.9-130.1		
2-Butanone	0.240	ug/ml	0.30000	79.9	69.9-130.1		
Benzaldehyde	0.310	ug/ml	0.30000	103	69.9-130.1		
Valeraldehyde	0.255	ug/ml	0.30000	85.0	69.9-130.1		
Hexaldehyde	0.244	ug/ml	0.30000	81.4	69.9-130.1		

LCS Dup (B6E2006-BSD1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.297	ug/ml	0.30000	98.9	79.9-120.1	0.349	20
Acetaldehyde	0.290	ug/ml	0.30000	96.7	79.9-120.1	0.499	20
Acetone	0.294	ug/ml	0.30000	98.1	69.9-130.1	0.653	20
Propionaldehyde	0.267	ug/ml	0.30000	88.9	69.9-130.1	0.876	20
Crotonaldehyde	0.281	ug/ml	0.30000	93.7	69.9-130.1	3.50	20
Butyraldehyde	0.253	ug/ml	0.30000	84.3	69.9-130.1	1.36	20
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.186	20
Benzaldehyde	0.307	ug/ml	0.30000	102	69.9-130.1	0.986	20
Valeraldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	0.252	20
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1	0.678	20

Duplicate (B6E2006-DUP2)

Source: 6051940-03 Prepared & Analyzed: 05/20/26

Formaldehyde	3.55	ug/m ³ Air	3.55			0.0818	10.1
Acetaldehyde	2.12	ug/m ³ Air	2.12			0.177	10.1
Acetone	1.70	ug/m ³ Air	1.70			0.00661	10.1
Propionaldehyde	0.571	ug/m ³ Air	0.573			0.340	10.1
Crotonaldehyde	ND	ug/m ³ Air	ND				10.1
Butyraldehyde	0.269	ug/m ³ Air	0.268			0.274	10.1
2-Butanone	0.321	ug/m ³ Air	0.337			4.96	10.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2006 - TO-11A extraction

Duplicate (B6E2006-DUP2) Continued Source: 6051940-03 Prepared & Analyzed: 05/20/26

Benzaldehyde	0.135	ug/m ³ Air		0.136			1.01	10.1	
Valeraldehyde	0.295	ug/m ³ Air		0.299			1.27	10.1	
Hexaldehyde	0.253	ug/m ³ Air		0.251			1.01	10.1	

Batch B6E2205 - TO-11A extraction

Blank (B6E2205-BLK1) Prepared & Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0284	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E2205-BS1) Prepared & Analyzed: 05/22/26

Formaldehyde	0.274	ug/ml	0.30000	91.5	79.9-120.1		
Acetaldehyde	0.288	ug/ml	0.30000	96.1	79.9-120.1		
Acetone	0.289	ug/ml	0.30000	96.2	69.9-130.1		
Propionaldehyde	0.302	ug/ml	0.30000	101	69.9-130.1		
Crotonaldehyde	0.281	ug/ml	0.30000	93.5	69.9-130.1		
Butyraldehyde	0.245	ug/ml	0.30000	81.8	69.9-130.1		
2-Butanone	0.230	ug/ml	0.30000	76.5	69.9-130.1		
Benzaldehyde	0.298	ug/ml	0.30000	99.3	69.9-130.1		
Valeraldehyde	0.248	ug/ml	0.30000	82.7	69.9-130.1		
Hexaldehyde	0.242	ug/ml	0.30000	80.8	69.9-130.1		

LCS Dup (B6E2205-BSD1) Prepared & Analyzed: 05/22/26

Formaldehyde	0.274	ug/ml	0.30000	91.2	79.9-120.1	0.300	20
Acetaldehyde	0.285	ug/ml	0.30000	95.0	79.9-120.1	1.19	20
Acetone	0.286	ug/ml	0.30000	95.3	69.9-130.1	0.879	20
Propionaldehyde	0.298	ug/ml	0.30000	99.4	69.9-130.1	1.45	20
Crotonaldehyde	0.277	ug/ml	0.30000	92.4	69.9-130.1	1.27	20
Butyraldehyde	0.246	ug/ml	0.30000	81.9	69.9-130.1	0.192	20
2-Butanone	0.228	ug/ml	0.30000	75.9	69.9-130.1	0.879	20
Benzaldehyde	0.296	ug/ml	0.30000	98.7	69.9-130.1	0.681	20
Valeraldehyde	0.245	ug/ml	0.30000	81.6	69.9-130.1	1.42	20
Hexaldehyde	0.243	ug/ml	0.30000	81.0	69.9-130.1	0.229	20

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2205 - TO-11A extraction

Duplicate (B6E2205-DUP5)

Source: 6051940-04 Prepared & Analyzed: 05/22/26

Formaldehyde	6.23	ug/m ³ Air		6.25			0.249	10.1	
Acetaldehyde	2.27	ug/m ³ Air		2.27			0.0696	10.1	
Acetone	2.50	ug/m ³ Air		2.48			0.522	10.1	
Propionaldehyde	0.553	ug/m ³ Air		0.548			0.789	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.231	ug/m ³ Air		0.231			0.0222	10.1	
2-Butanone	0.460	ug/m ³ Air		0.445			3.19	10.1	
Benzaldehyde	0.110	ug/m ³ Air		0.115			4.29	10.1	
Valeraldehyde	0.256	ug/m ³ Air		0.243			5.27	10.1	
Hexaldehyde	0.151	ug/m ³ Air		0.154			1.47	10.1	

Duplicate (B6E2205-DUP6)

Source: 6051940-05 Prepared & Analyzed: 05/22/26

Formaldehyde	6.35	ug/m ³ Air		6.34			0.0994	10.1	
Acetaldehyde	2.31	ug/m ³ Air		2.31			0.135	10.1	
Acetone	2.53	ug/m ³ Air		2.53			0.0888	10.1	
Propionaldehyde	0.558	ug/m ³ Air		0.564			1.10	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.247	ug/m ³ Air		0.242			2.08	10.1	
2-Butanone	0.454	ug/m ³ Air		0.453			0.344	10.1	
Benzaldehyde	0.114	ug/m ³ Air		0.105			8.12	10.1	
Valeraldehyde	0.267	ug/m ³ Air		0.257			3.91	10.1	
Hexaldehyde	0.174	ug/m ³ Air		0.171			1.50	10.1	

Batch B6F0403 - TO-11A extraction

Blank (B6F0403-BLK1)

Prepared & Analyzed: 06/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0265	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0473	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6F0403-BS1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.278	ug/ml	0.30000	92.7	79.9-120.1	
Acetaldehyde	0.281	ug/ml	0.30000	93.7	79.9-120.1	
Acetone	0.279	ug/ml	0.30000	93.2	69.9-130.1	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6F0403 - TO-11A extraction

LCS (B6F0403-BS1) Continued

Prepared & Analyzed: 06/04/26

Propionaldehyde	0.325	ug/ml	0.30000	108	69.9-130.1
Crotonaldehyde	0.253	ug/ml	0.30000	84.4	69.9-130.1
Butyraldehyde	0.240	ug/ml	0.30000	80.2	69.9-130.1
2-Butanone	0.223	ug/ml	0.30000	74.4	69.9-130.1
Benzaldehyde	0.295	ug/ml	0.30000	98.3	69.9-130.1
Valeraldehyde	0.243	ug/ml	0.30000	80.9	69.9-130.1
Hexaldehyde	0.237	ug/ml	0.30000	79.1	69.9-130.1

LCS Dup (B6F0403-BSD1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.280	ug/ml	0.30000	93.2	79.9-120.1	0.576	20
Acetaldehyde	0.283	ug/ml	0.30000	94.4	79.9-120.1	0.758	20
Acetone	0.283	ug/ml	0.30000	94.3	69.9-130.1	1.21	20
Propionaldehyde	0.322	ug/ml	0.30000	107	69.9-130.1	0.892	20
Crotonaldehyde	0.257	ug/ml	0.30000	85.5	69.9-130.1	1.33	20
Butyraldehyde	0.242	ug/ml	0.30000	80.6	69.9-130.1	0.556	20
2-Butanone	0.224	ug/ml	0.30000	74.7	69.9-130.1	0.333	20
Benzaldehyde	0.290	ug/ml	0.30000	96.8	69.9-130.1	1.55	20
Valeraldehyde	0.244	ug/ml	0.30000	81.3	69.9-130.1	0.500	20
Hexaldehyde	0.239	ug/ml	0.30000	79.7	69.9-130.1	0.759	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/08/26 09:16

SUBMITTED: 05/06/26 to 06/03/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Notes and Definitions

U	Under Detection Limit
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
O-05	This sample was extracted outside of the EPA recommended holding time.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
D-F	Duplicate exceeds DQO criteria.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.