



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 13, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/07/26 11:07 through 05/19/26 11:54.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6040775-01	Air	04/02/26 23:59	04/07/26 11:07
TMOK	6041615-03	Air	04/07/26 23:59	04/16/26 12:39
TMOK	6041615-04	Air	04/08/26 23:59	04/16/26 12:39
TMOK	6042228-01	Air	04/14/26 23:59	04/22/26 14:05
TMOK	6042228-02	Air	04/17/26 23:59	04/22/26 14:05
TMOK	6042228-03	Air	04/20/26 23:59	04/22/26 14:05
TMOK D1	6043007-01	Air	04/28/26 23:59	04/30/26 11:12
TMOK D2	6043007-02	Air	04/28/26 23:59	04/30/26 11:12
TMOK	6050702-02	Air	05/02/26 23:59	05/07/26 10:54
TMOK	6051320-02	Air	05/08/26 23:59	05/13/26 12:55
TMOK	6051950-01	Air	04/26/26 23:59	05/19/26 11:54
TMOK FB	6051950-02	Air	04/26/26 23:59	05/19/26 11:54



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040775-01

Sampled: 04/02/26 23:59

Temp Rec'd (°C): 6.5

Sample Volume: 1156.6 L

Received: 04/07/26 11:07

Cartridge Pouch #: 26G0957

Analyzed: 04/14/26 18:46

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.01		0.0547
Acetaldehyde	1.51		0.0227
Acetone	1.56		0.0839
Propionaldehyde	0.386		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.159		0.00968
2-Butanone	0.371		0.0365
Benzaldehyde	0.103		0.0374
Valeraldehyde	0.121		0.0105
Hexaldehyde	0.139		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6041615-03	Sampled: 04/07/26 23:59
Temp Rec'd (°C): 2.7	Sample Volume: 1155.2 L	Received: 04/16/26 12:39
	Cartridge Pouch #: 26M0017	Analyzed: 04/20/26 22:09

Comments: Make up for 1/2/26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.49		0.0548
Acetaldehyde	1.88		0.0227
Acetone	3.56		0.0840
Propionaldehyde	0.587	QB-01	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.209		0.00970
2-Butanone	0.822		0.0365
Benzaldehyde	0.142		0.0375
Valeraldehyde	0.0955		0.0105
Hexaldehyde	0.240		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6041615-04

Sampled: 04/08/26 23:59

Temp Rec'd (°C): 2.7

Sample Volume: 1156.2 L

Received: 04/16/26 12:39

Cartridge Pouch #: 26G0958

Analyzed: 04/20/26 22:25

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.52		0.0547
Acetaldehyde	1.79		0.0227
Acetone	4.95		0.0839
Propionaldehyde	0.568	QB-01	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.192		0.00969
2-Butanone	0.789		0.0365
Benzaldehyde	0.123		0.0375
Valeraldehyde	0.0841		0.0105
Hexaldehyde	0.176		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042228-01

Sampled: 04/14/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1157 L

Received: 04/22/26 14:05

Cartridge Pouch #: 26G0959

Analyzed: 04/23/26 20:14

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.99		0.0547
Acetaldehyde	1.14		0.0226
Acetone	0.650		0.0838
Propionaldehyde	0.318		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.119		0.00968
2-Butanone	0.214		0.0365
Benzaldehyde	0.119		0.0374
Valeraldehyde	0.0694		0.0105
Hexaldehyde	0.100		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042228-02

Sampled: 04/17/26 23:59

Temp Rec'd (°C): -0.5

Sample Volume: 1154.9 L

Received: 04/22/26 14:05

Cartridge Pouch #: 26M0020

Analyzed: 04/27/26 20:45

Comments: makeup for 1-20-26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.33		0.0548
Acetaldehyde	1.16		0.0227
Acetone	0.775		0.0840
Propionaldehyde	0.343	QB-01	0.179
Crotonaldehyde	ND	U	0.0165
Butyraldehyde	0.138		0.00970
2-Butanone	0.230		0.0365
Benzaldehyde	0.0858		0.0375
Valeraldehyde	0.0740		0.0105
Hexaldehyde	0.111		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042228-03

Sampled: 04/20/26 23:59

Temp Rec'd (°C): -0.5

Sample Volume: 1156.5 L

Received: 04/22/26 14:05

Cartridge Pouch #: 26G0960

Analyzed: 05/04/26 15:10

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.80		0.0547
Acetaldehyde	1.51		0.0227
Acetone	4.93		0.0839
Propionaldehyde	0.578	QB-01	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.176		0.00968
2-Butanone	1.61		0.0365
Benzaldehyde	0.123		0.0374
Valeraldehyde	0.0736		0.0105
Hexaldehyde	0.0834		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1	Lab ID: 6043007-01	Sampled: 04/28/26 23:59
Temp Rec'd (°C): 8.6	Sample Volume: 1150 L	Received: 04/30/26 11:12
	Cartridge Pouch #: 26M0025	Analyzed: 05/07/26 19:41

Comments: Makeup from 2/13/26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.64		0.0550
Acetaldehyde	0.852		0.0228
Acetone	0.886		0.0843
Propionaldehyde	0.238		0.180
Crotonaldehyde	ND	U	0.0165
Butyraldehyde	0.0755		0.00974
2-Butanone	0.187		0.0367
Benzaldehyde	0.0754		0.0377
Valeraldehyde	0.0505		0.0105
Hexaldehyde	0.0740		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2	Lab ID: 6043007-02	Sampled: 04/28/26 23:59
Temp Rec'd (°C): 8.6	Sample Volume: 1149.3 L	Received: 04/30/26 11:12
	Cartridge Pouch #: 26M0026	Analyzed: 05/07/26 20:13

Comments: Makeup from 2/13/26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.69		0.0551
Acetaldehyde	0.842		0.0228
Acetone	0.890		0.0844
Propionaldehyde	0.233		0.180
Crotonaldehyde	ND	U	0.0165
Butyraldehyde	0.0739		0.00975
2-Butanone	0.181		0.0367
Benzaldehyde	0.0881		0.0377
Valeraldehyde	0.0427		0.0105
Hexaldehyde	0.0741		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6050702-02

Sampled: 05/02/26 23:59

Temp Rec'd (°C): 3.4

Sample Volume: 1156.1 L

Received: 05/07/26 10:54

Cartridge Pouch #: 26G0963

Analyzed: 05/11/26 16:37

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.84		0.0548
Acetaldehyde	1.33		0.0227
Acetone	4.01		0.0839
Propionaldehyde	0.476	QB-01	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.149		0.00969
2-Butanone	0.535		0.0365
Benzaldehyde	0.107		0.0375
Valeraldehyde	0.0687		0.0105
Hexaldehyde	0.142		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6051320-02	Sampled: 05/08/26 23:59
Temp Rec'd (°C): 4.9	Sample Volume: 1156.1 L	Received: 05/13/26 12:55
	Cartridge Pouch #: 26G0964	Analyzed: 05/19/26 19:52

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	4.11		0.0548
Acetaldehyde	1.81		0.0227
Acetone	1.86		0.0839
Propionaldehyde	0.400		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.192		0.00969
2-Butanone	0.555		0.0365
Benzaldehyde	0.138		0.0375
Valeraldehyde	0.234		0.0105
Hexaldehyde	0.228		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6051950-01	Sampled: 04/26/26 23:59
Temp Rec'd (°C): 4.5	Sample Volume: 1156.6 L	Received: 05/19/26 11:54
	Cartridge Pouch #: 26G0961	Analyzed: 05/20/26 20:06
Comments: Recd past extraction HT		

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.78	DNPH, S-Rec	0.0547
Acetaldehyde	0.849	DNPH, S-Rec	0.0227
Acetone	0.540	DNPH, S-Rec	0.0839
Propionaldehyde	0.217	DNPH, S-Rec	0.179
Crotonaldehyde	ND	DNPH, S-Rec, U	0.0164
Butyraldehyde	0.129	DNPH, S-Rec	0.00968
2-Butanone	0.700	DNPH, S-Rec	0.0365
Benzaldehyde	0.124	DNPH, S-Rec	0.0374
Valeraldehyde	0.116	DNPH, S-Rec	0.0105
Hexaldehyde	0.130	DNPH, S-Rec	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK FB	Lab ID: 6051950-02	Sampled: 04/26/26 23:59
Temp Rec'd (°C): 4.5	Sample Volume: 1156.6 L	Received: 05/19/26 11:54
	Cartridge Pouch #: 26G0962	Analyzed: 05/20/26 20:22
Comments: Recd past extraction HT		

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0310	S-Rec, U	0.0547
Acetaldehyde	0.0231	S-Rec	0.0227
Acetone	0.0311	S-Rec, U	0.0839
Propionaldehyde	0.0514	S-Rec, U	0.179
Crotonaldehyde	ND	S-Rec, U	0.0164
Butyraldehyde	ND	S-Rec, U	0.00968
2-Butanone	ND	S-Rec, U	0.0365
Benzaldehyde	0.0360	S-Rec, U	0.0374
Valeraldehyde	ND	S-Rec, U	0.0105
Hexaldehyde	ND	S-Rec, U	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604044 - B6D0902

Calibration Blank (2604044-CCB1)

Prepared & Analyzed: 04/14/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2604044-CCB2)

Prepared & Analyzed: 04/14/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604044-CCV1)

Prepared & Analyzed: 04/14/26

Formaldehyde	0.143	ug/ml	0.15000		95.6	84.9-115.1			
Acetaldehyde	0.148	ug/ml	0.15000		98.7	84.9-115.1			
Acetone	0.149	ug/ml	0.15000		99.2	84.9-115.1			
Propionaldehyde	0.149	ug/ml	0.15000		99.6	84.9-115.1			
Crotonaldehyde	0.151	ug/ml	0.15000		101	84.9-115.1			
Butyraldehyde	0.145	ug/ml	0.15000		96.8	84.9-115.1			
2-Butanone	0.150	ug/ml	0.15000		99.8	84.9-115.1			
Benzaldehyde	0.154	ug/ml	0.15000		102	84.9-115.1			
Valeraldehyde	0.152	ug/ml	0.15000		102	84.9-115.1			
Hexaldehyde	0.146	ug/ml	0.15000		97.1	84.9-115.1			

Calibration Check (2604044-CCV2)

Prepared & Analyzed: 04/14/26

Formaldehyde	0.144	ug/ml	0.15000		96.0	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		99.7	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604044 - B6D0902

Calibration Check (2604044-CCV2) Contin

Prepared & Analyzed: 04/14/26

Acetone	0.150	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1

Calibration Check (2604044-CCV3)

Prepared & Analyzed: 04/14/26

Formaldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.5	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.7	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1

Batch 2604065 - B6D1701

Calibration Blank (2604065-CCB1)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2604065-CCB2)

Prepared: 04/20/26 Analyzed: 04/21/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604065 - B6D1701

Calibration Blank (2604065-CCB2) Contin

Prepared: 04/20/26 Analyzed: 04/21/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604065-CCV1)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.146	ug/ml	0.15000	97.7	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.6	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2604065-CCV2)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.2	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1

Calibration Check (2604065-CCV3)

Prepared: 04/20/26 Analyzed: 04/21/26

Formaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Acetaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Acetone	0.150	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.3	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604065 - B6D1701

Calibration Check (2604065-CCV3) Contin

Prepared: 04/20/26 Analyzed: 04/21/26

Hexaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
-------------	-------	-------	---------	------	------------

Batch 2604076 - B6D2305

Calibration Blank (2604076-CCB1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2604076-CCB2)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2604076-CCV1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1

Calibration Check (2604076-CCV2)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604076 - B6D2305

Calibration Check (2604076-CCV2) Contin

Prepared & Analyzed: 04/23/26

Acetaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1

Calibration Check (2604076-CCV3)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.4	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1

Batch 2604088 - B6D2709

Calibration Blank (2604088-CCB1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2604088-CCB2)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604088 - B6D2709

Calibration Blank (2604088-CCB2) Contin

Prepared & Analyzed: 04/27/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604088-CCV1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.143	ug/ml	0.15000	95.0	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.6	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1

Calibration Check (2604088-CCV2)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.2	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.5	84.9-115.1

Calibration Check (2604088-CCV3)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.0	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.0	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	104	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604088 - B6D2709

Calibration Check (2604088-CCV3) Contin

Prepared & Analyzed: 04/27/26

Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1

Batch 2605011 - B6E0103

Calibration Blank (2605011-CCB1)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605011-CCB2)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605011-CCV1)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.0	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Crotonaldehyde	0.148	ug/ml	0.15000	98.4	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1

Calibration Check (2605011-CCV2)

Prepared & Analyzed: 05/04/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605011 - B6E0103

Calibration Check (2605011-CCV2) Contin

Prepared & Analyzed: 05/04/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Butyraldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.4	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Hexaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1

Calibration Check (2605011-CCV3)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.6	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Crotonaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
2-Butanone	0.140	ug/ml	0.15000	93.2	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1

Batch 2605024 - B6E0702

Calibration Blank (2605024-CCB1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2605024-CCB2)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605024 - B6E0702

Calibration Blank (2605024-CCB2) Contin

Prepared & Analyzed: 05/07/26

Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605024-CCV1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.4	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1

Calibration Check (2605024-CCV2)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2605024-CCV3)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	94.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.1	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605024 - B6E0702

Calibration Check (2605024-CCV3) Contin

Prepared & Analyzed: 05/07/26

Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1

Batch 2605031 - B6E1108

Calibration Blank (2605031-CCB1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605031-CCB2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605031-CCV1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Butyraldehyde	0.140	ug/ml	0.15000	93.4	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Check (2605031-CCV2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1

Calibration Check (2605031-CCV3)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.6	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	95.7	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch 2605066 - B6E1503

Calibration Blank (2605066-CCB1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2605066-CCB2)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605066 - B6E1503

Calibration Blank (2605066-CCB2) Contin

Prepared & Analyzed: 05/19/26

Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605066-CCV1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.0	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Hexaldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1

Calibration Check (2605066-CCV2)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1

Calibration Check (2605066-CCV3)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.5	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.5	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605066 - B6E1503

Calibration Check (2605066-CCV3) Contin

Prepared & Analyzed: 05/19/26

Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.5	84.9-115.1

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605069-CCB2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605069-CCV1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.9	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
2-Butanone	0.154	ug/ml	0.15000	102	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Check (2605069-CCV2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Calibration Check (2605069-CCV3)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch B6D0902 - TO-11A extraction

Blank (B6D0902-BLK1)

Prepared: 04/09/26 Analyzed: 04/14/26

Formaldehyde	ND	ug/m ³ Air	U
Acetaldehyde	ND	ug/m ³ Air	U
Acetone	ND	ug/m ³ Air	U
Propionaldehyde	ND	ug/m ³ Air	U
Crotonaldehyde	ND	ug/m ³ Air	U
Butyraldehyde	ND	ug/m ³ Air	U
2-Butanone	ND	ug/m ³ Air	U
Benzaldehyde	0.0454	ug/m ³ Air	
Valeraldehyde	ND	ug/m ³ Air	U
Hexaldehyde	0.0221	ug/m ³ Air	

Blank (B6D0902-BLK3)

Prepared: 04/09/26 Analyzed: 04/14/26

Formaldehyde	ND	ug/m ³ Air	U
Acetaldehyde	ND	ug/m ³ Air	U
Acetone	ND	ug/m ³ Air	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D0902 - TO-11A extraction

Blank (B6D0902-BLK3) Continued

Prepared: 04/09/26 Analyzed: 04/14/26

Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D0902-BS1)

Prepared: 04/09/26 Analyzed: 04/14/26

Formaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1		
Acetaldehyde	0.298	ug/ml	0.30000	99.2	79.9-120.1		
Acetone	0.304	ug/ml	0.30000	101	69.9-130.1		
Propionaldehyde	0.331	ug/ml	0.30000	110	69.9-130.1		
Crotonaldehyde	0.283	ug/ml	0.30000	94.4	69.9-130.1		
Butyraldehyde	0.262	ug/ml	0.30000	87.4	69.9-130.1		
2-Butanone	0.250	ug/ml	0.30000	83.3	69.9-130.1		
Benzaldehyde	0.322	ug/ml	0.30000	107	69.9-130.1		
Valeraldehyde	0.262	ug/ml	0.30000	87.4	69.9-130.1		
Hexaldehyde	0.256	ug/ml	0.30000	85.2	69.9-130.1		

LCS Dup (B6D0902-BSD1)

Prepared: 04/09/26 Analyzed: 04/14/26

Formaldehyde	0.290	ug/ml	0.30000	96.6	79.9-120.1	1.29	20
Acetaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1	1.33	20
Acetone	0.305	ug/ml	0.30000	102	69.9-130.1	0.128	20
Propionaldehyde	0.328	ug/ml	0.30000	109	69.9-130.1	0.906	20
Crotonaldehyde	0.282	ug/ml	0.30000	94.1	69.9-130.1	0.235	20
Butyraldehyde	0.260	ug/ml	0.30000	86.6	69.9-130.1	0.911	20
2-Butanone	0.249	ug/ml	0.30000	83.0	69.9-130.1	0.411	20
Benzaldehyde	0.317	ug/ml	0.30000	106	69.9-130.1	1.64	20
Valeraldehyde	0.260	ug/ml	0.30000	86.8	69.9-130.1	0.651	20
Hexaldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	0.432	20

Batch B6D1701 - TO-11A extraction

Blank (B6D1701-BLK1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	ND	ug/m ³ Air						U
Acetaldehyde	0.0326	ug/m ³ Air						
Acetone	ND	ug/m ³ Air						U
Propionaldehyde	ND	ug/m ³ Air						QB-01, U
Crotonaldehyde	ND	ug/m ³ Air						U
Butyraldehyde	ND	ug/m ³ Air						U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1701 - TO-11A extraction

Blank (B6D1701-BLK1) Continued

Prepared: 04/17/26 Analyzed: 04/20/26

2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0685	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6D1701-BLK3)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D1701-BS1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	0.298	ug/ml	0.30000	99.2	79.9-120.1				
Acetaldehyde	0.308	ug/ml	0.30000	103	79.9-120.1				
Acetone	0.315	ug/ml	0.30000	105	69.9-130.1				
Propionaldehyde	0.343	ug/ml	0.30000	114	69.9-130.1				
Crotonaldehyde	0.294	ug/ml	0.30000	97.9	69.9-130.1				
Butyraldehyde	0.273	ug/ml	0.30000	91.0	69.9-130.1				
2-Butanone	0.250	ug/ml	0.30000	83.3	69.9-130.1				
Benzaldehyde	0.331	ug/ml	0.30000	110	69.9-130.1				
Valeraldehyde	0.271	ug/ml	0.30000	90.3	69.9-130.1				
Hexaldehyde	0.265	ug/ml	0.30000	88.2	69.9-130.1				

LCS Dup (B6D1701-BSD1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	0.286	ug/ml	0.30000	95.2	79.9-120.1	4.09	20		
Acetaldehyde	0.294	ug/ml	0.30000	98.1	79.9-120.1	4.44	20		
Acetone	0.299	ug/ml	0.30000	99.8	69.9-130.1	4.98	20		
Propionaldehyde	0.328	ug/ml	0.30000	109	69.9-130.1	4.29	20		
Crotonaldehyde	0.282	ug/ml	0.30000	93.9	69.9-130.1	4.18	20		
Butyraldehyde	0.259	ug/ml	0.30000	86.2	69.9-130.1	5.34	20		
2-Butanone	0.239	ug/ml	0.30000	79.8	69.9-130.1	4.31	20		
Benzaldehyde	0.320	ug/ml	0.30000	107	69.9-130.1	3.30	20		
Valeraldehyde	0.265	ug/ml	0.30000	88.3	69.9-130.1	2.23	20		
Hexaldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	3.84	20		

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1701 - TO-11A extraction

Duplicate (B6D1701-DUP2)

Source: 6041615-04 Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	3.51	ug/m ³ Air		3.52			0.206	10.1	
Acetaldehyde	1.79	ug/m ³ Air		1.79			0.258	10.1	
Acetone	4.94	ug/m ³ Air		4.95			0.122	10.1	
Propionaldehyde	0.565	ug/m ³ Air		0.568			0.458	10.1	QB-01
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.185	ug/m ³ Air		0.192			3.52	10.1	
2-Butanone	0.780	ug/m ³ Air		0.789			1.19	10.1	
Benzaldehyde	0.118	ug/m ³ Air		0.123			4.30	10.1	
Valeraldehyde	0.0921	ug/m ³ Air		0.0841			9.09	10.1	
Hexaldehyde	0.175	ug/m ³ Air		0.176			0.466	10.1	

Batch B6D2305 - TO-11A extraction

Blank (B6D2305-BLK1)

Prepared & Analyzed: 04/23/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0281	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0441	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6D2305-BLK3)

Prepared & Analyzed: 04/23/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D2305-BS1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.283	ug/ml	0.30000	94.3	79.9-120.1
Acetaldehyde	0.294	ug/ml	0.30000	98.1	79.9-120.1
Acetone	0.297	ug/ml	0.30000	98.9	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D2305 - TO-11A extraction

LCS (B6D2305-BS1) Continued

Prepared & Analyzed: 04/23/26

Propionaldehyde	0.329	ug/ml	0.30000	110	69.9-130.1
Crotonaldehyde	0.289	ug/ml	0.30000	96.3	69.9-130.1
Butyraldehyde	0.262	ug/ml	0.30000	87.2	69.9-130.1
2-Butanone	0.240	ug/ml	0.30000	79.9	69.9-130.1
Benzaldehyde	0.317	ug/ml	0.30000	106	69.9-130.1
Valeraldehyde	0.263	ug/ml	0.30000	87.7	69.9-130.1
Hexaldehyde	0.256	ug/ml	0.30000	85.4	69.9-130.1

LCS Dup (B6D2305-BSD1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.284	ug/ml	0.30000	94.6	79.9-120.1	0.236	20
Acetaldehyde	0.296	ug/ml	0.30000	98.8	79.9-120.1	0.747	20
Acetone	0.298	ug/ml	0.30000	99.2	69.9-130.1	0.274	20
Propionaldehyde	0.330	ug/ml	0.30000	110	69.9-130.1	0.258	20
Crotonaldehyde	0.288	ug/ml	0.30000	96.1	69.9-130.1	0.144	20
Butyraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1	0.274	20
2-Butanone	0.237	ug/ml	0.30000	79.0	69.9-130.1	1.25	20
Benzaldehyde	0.319	ug/ml	0.30000	106	69.9-130.1	0.707	20
Valeraldehyde	0.260	ug/ml	0.30000	86.6	69.9-130.1	1.32	20
Hexaldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	0.592	20

Batch B6D2709 - TO-11A extraction

Blank (B6D2709-BLK1)

Prepared & Analyzed: 04/27/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0322	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6D2709-BLK3)

Prepared & Analyzed: 04/27/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D2709 - TO-11A extraction

Blank (B6D2709-BLK3) Continued

Prepared & Analyzed: 04/27/26

2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D2709-BS1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.279	ug/ml	0.30000		93.0	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.287	ug/ml	0.30000		95.7	69.9-130.1			
Propionaldehyde	0.332	ug/ml	0.30000		111	69.9-130.1			
Crotonaldehyde	0.271	ug/ml	0.30000		90.3	69.9-130.1			
Butyraldehyde	0.255	ug/ml	0.30000		85.0	69.9-130.1			
2-Butanone	0.234	ug/ml	0.30000		77.9	69.9-130.1			
Benzaldehyde	0.306	ug/ml	0.30000		102	69.9-130.1			
Valeraldehyde	0.256	ug/ml	0.30000		85.2	69.9-130.1			
Hexaldehyde	0.247	ug/ml	0.30000		82.2	69.9-130.1			

LCS Dup (B6D2709-BSD1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.277	ug/ml	0.30000		92.3	79.9-120.1	0.699	20	
Acetaldehyde	0.289	ug/ml	0.30000		96.5	79.9-120.1	0.298	20	
Acetone	0.288	ug/ml	0.30000		95.9	69.9-130.1	0.165	20	
Propionaldehyde	0.331	ug/ml	0.30000		110	69.9-130.1	0.398	20	
Crotonaldehyde	0.274	ug/ml	0.30000		91.2	69.9-130.1	1.08	20	
Butyraldehyde	0.252	ug/ml	0.30000		84.0	69.9-130.1	1.22	20	
2-Butanone	0.229	ug/ml	0.30000		76.4	69.9-130.1	2.02	20	
Benzaldehyde	0.308	ug/ml	0.30000		103	69.9-130.1	0.824	20	
Valeraldehyde	0.253	ug/ml	0.30000		84.3	69.9-130.1	1.01	20	
Hexaldehyde	0.248	ug/ml	0.30000		82.6	69.9-130.1	0.420	20	

Batch B6E0103 - TO-11A extraction

Blank (B6E0103-BLK1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0369	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0103 - TO-11A extraction

Blank (B6E0103-BLK1) Continued

Prepared: 05/01/26 Analyzed: 05/04/26

Hexaldehyde	ND	ug/m ³ Air							U
-------------	----	-----------------------	--	--	--	--	--	--	---

Blank (B6E0103-BLK3)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E0103-BS1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	0.290	ug/ml	0.30000	96.6	79.9-120.1		
Acetaldehyde	0.289	ug/ml	0.30000	96.2	79.9-120.1		
Acetone	0.298	ug/ml	0.30000	99.3	69.9-130.1		
Propionaldehyde	0.370	ug/ml	0.30000	123	69.9-130.1		
Crotonaldehyde	0.286	ug/ml	0.30000	95.4	69.9-130.1		
Butyraldehyde	0.259	ug/ml	0.30000	86.4	69.9-130.1		
2-Butanone	0.237	ug/ml	0.30000	79.2	69.9-130.1		
Benzaldehyde	0.312	ug/ml	0.30000	104	69.9-130.1		
Valeraldehyde	0.261	ug/ml	0.30000	86.9	69.9-130.1		
Hexaldehyde	0.249	ug/ml	0.30000	83.0	69.9-130.1		

LCS Dup (B6E0103-BSD1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	0.288	ug/ml	0.30000	96.0	79.9-120.1	0.640	20
Acetaldehyde	0.287	ug/ml	0.30000	95.7	79.9-120.1	0.479	20
Acetone	0.293	ug/ml	0.30000	97.5	69.9-130.1	1.79	20
Propionaldehyde	0.376	ug/ml	0.30000	125	69.9-130.1	1.60	20
Crotonaldehyde	0.280	ug/ml	0.30000	93.3	69.9-130.1	2.16	20
Butyraldehyde	0.256	ug/ml	0.30000	85.2	69.9-130.1	1.45	20
2-Butanone	0.236	ug/ml	0.30000	78.5	69.9-130.1	0.807	20
Benzaldehyde	0.312	ug/ml	0.30000	104	69.9-130.1	0.119	20
Valeraldehyde	0.258	ug/ml	0.30000	86.1	69.9-130.1	0.976	20
Hexaldehyde	0.250	ug/ml	0.30000	83.3	69.9-130.1	0.401	20

Batch B6E0702 - TO-11A extraction

Blank (B6E0702-BLK1)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air							U
--------------	----	-----------------------	--	--	--	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0702 - TO-11A extraction

Blank (B6E0702-BLK1) Continued

Prepared & Analyzed: 05/07/26

Acetaldehyde	0.0303	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0517	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E0702-BLK3)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E0702-BS1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.288	ug/ml	0.30000		96.0	79.9-120.1			
Acetaldehyde	0.294	ug/ml	0.30000		98.1	79.9-120.1			
Acetone	0.294	ug/ml	0.30000		98.1	69.9-130.1			
Propionaldehyde	0.335	ug/ml	0.30000		112	69.9-130.1			
Crotonaldehyde	0.277	ug/ml	0.30000		92.2	69.9-130.1			
Butyraldehyde	0.260	ug/ml	0.30000		86.8	69.9-130.1			
2-Butanone	0.242	ug/ml	0.30000		80.7	69.9-130.1			
Benzaldehyde	0.310	ug/ml	0.30000		103	69.9-130.1			
Valeraldehyde	0.259	ug/ml	0.30000		86.4	69.9-130.1			
Hexaldehyde	0.251	ug/ml	0.30000		83.6	69.9-130.1			

LCS Dup (B6E0702-BSD1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.288	ug/ml	0.30000		95.9	79.9-120.1	0.125	20	
Acetaldehyde	0.294	ug/ml	0.30000		97.9	79.9-120.1	0.145	20	
Acetone	0.293	ug/ml	0.30000		97.6	69.9-130.1	0.495	20	
Propionaldehyde	0.334	ug/ml	0.30000		111	69.9-130.1	0.412	20	
Crotonaldehyde	0.277	ug/ml	0.30000		92.4	69.9-130.1	0.162	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0702 - TO-11A extraction

LCS Dup (B6E0702-BSD1) Continued

Prepared & Analyzed: 05/07/26

Butyraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1	0.315	20	
2-Butanone	0.244	ug/ml	0.30000	81.4	69.9-130.1	0.906	20	
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1	1.30	20	
Valeraldehyde	0.259	ug/ml	0.30000	86.5	69.9-130.1	0.0858	20	
Hexaldehyde	0.251	ug/ml	0.30000	83.6	69.9-130.1	0.102	20	

Duplicate (B6E0702-DUP2)

Source: 6043007-01 Prepared & Analyzed: 05/07/26

Formaldehyde	1.65	ug/m ³ Air	1.64			0.394	10.1	
Acetaldehyde	0.855	ug/m ³ Air	0.852			0.443	10.1	
Acetone	0.891	ug/m ³ Air	0.886			0.555	10.1	
Propionaldehyde	0.231	ug/m ³ Air	0.238			2.96	10.1	
Crotonaldehyde	ND	ug/m ³ Air	ND				10.1	U
Butyraldehyde	0.0731	ug/m ³ Air	0.0755			3.22	10.1	
2-Butanone	0.195	ug/m ³ Air	0.187			3.91	10.1	
Benzaldehyde	0.0810	ug/m ³ Air	0.0754			7.16	10.1	
Valeraldehyde	0.0500	ug/m ³ Air	0.0505			0.976	10.1	
Hexaldehyde	0.0728	ug/m ³ Air	0.0740			1.61	10.1	

Duplicate (B6E0702-DUP3)

Source: 6043007-02 Prepared & Analyzed: 05/07/26

Formaldehyde	1.69	ug/m ³ Air	1.69			0.00231	10.1	
Acetaldehyde	0.847	ug/m ³ Air	0.842			0.624	10.1	
Acetone	0.880	ug/m ³ Air	0.890			1.09	10.1	
Propionaldehyde	0.232	ug/m ³ Air	0.233			0.578	10.1	
Crotonaldehyde	ND	ug/m ³ Air	ND				10.1	U
Butyraldehyde	0.0697	ug/m ³ Air	0.0739			5.83	10.1	
2-Butanone	0.177	ug/m ³ Air	0.181			2.34	10.1	
Benzaldehyde	0.0667	ug/m ³ Air	0.0881			27.7	10.1	
Valeraldehyde	0.0419	ug/m ³ Air	0.0427			1.92	10.1	
Hexaldehyde	0.0759	ug/m ³ Air	0.0741			2.44	10.1	

Batch B6E1108 - TO-11A extraction

Blank (B6E1108-BLK1)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air						U
Acetaldehyde	0.0325	ug/m ³ Air						
Acetone	ND	ug/m ³ Air						U
Propionaldehyde	ND	ug/m ³ Air						QB-01, U
Crotonaldehyde	ND	ug/m ³ Air						U
Butyraldehyde	ND	ug/m ³ Air						U
2-Butanone	ND	ug/m ³ Air						U
Benzaldehyde	0.0448	ug/m ³ Air						

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1108 - TO-11A extraction

Blank (B6E1108-BLK1) Continued

Prepared & Analyzed: 05/11/26

Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E1108-BLK3)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E1108-BS1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.288	ug/ml	0.30000	95.9	79.9-120.1		
Acetaldehyde	0.293	ug/ml	0.30000	97.6	79.9-120.1		
Acetone	0.294	ug/ml	0.30000	98.0	69.9-130.1		
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1		
Crotonaldehyde	0.276	ug/ml	0.30000	92.0	69.9-130.1		
Butyraldehyde	0.263	ug/ml	0.30000	87.6	69.9-130.1		
2-Butanone	0.238	ug/ml	0.30000	79.4	69.9-130.1		
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1		
Valeraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1		
Hexaldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1		

LCS Dup (B6E1108-BSD1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.290	ug/ml	0.30000	96.6	79.9-120.1	0.660	20
Acetaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1	0.305	20
Acetone	0.293	ug/ml	0.30000	97.5	69.9-130.1	0.439	20
Propionaldehyde	0.348	ug/ml	0.30000	116	69.9-130.1	0.877	20
Crotonaldehyde	0.271	ug/ml	0.30000	90.4	69.9-130.1	1.74	20
Butyraldehyde	0.259	ug/ml	0.30000	86.2	69.9-130.1	1.60	20
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.778	20
Benzaldehyde	0.315	ug/ml	0.30000	105	69.9-130.1	0.358	20
Valeraldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	2.43	20
Hexaldehyde	0.249	ug/ml	0.30000	83.1	69.9-130.1	2.06	20

Batch B6E1503 - TO-11A extraction

Blank (B6E1503-BLK1)

Prepared: 05/15/26 Analyzed: 05/22/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1503 - TO-11A extraction

Blank (B6E1503-BLK1) Continued

Prepared: 05/15/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E1503-BLK3)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E1503-BS1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.325	ug/ml	0.30000	108	79.9-120.1			
Acetaldehyde	0.285	ug/ml	0.30000	95.0	79.9-120.1			
Acetone	0.295	ug/ml	0.30000	98.5	69.9-130.1			
Propionaldehyde	0.266	ug/ml	0.30000	88.6	69.9-130.1			
Crotonaldehyde	0.278	ug/ml	0.30000	92.6	69.9-130.1			
Butyraldehyde	0.250	ug/ml	0.30000	83.3	69.9-130.1			
2-Butanone	0.241	ug/ml	0.30000	80.4	69.9-130.1			
Benzaldehyde	0.308	ug/ml	0.30000	103	69.9-130.1			
Valeraldehyde	0.249	ug/ml	0.30000	83.0	69.9-130.1			
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1			

LCS Dup (B6E1503-BSD1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.332	ug/ml	0.30000	111	79.9-120.1	2.31	20	
Acetaldehyde	0.289	ug/ml	0.30000	96.4	79.9-120.1	1.46	20	
Acetone	0.301	ug/ml	0.30000	100	69.9-130.1	2.03	20	
Propionaldehyde	0.267	ug/ml	0.30000	89.1	69.9-130.1	0.612	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1503 - TO-11A extraction

LCS Dup (B6E1503-BSD1) Continued

Prepared: 05/15/26 Analyzed: 05/19/26

Crotonaldehyde	0.283	ug/ml	0.30000		94.4	69.9-130.1	1.90	20	
Butyraldehyde	0.255	ug/ml	0.30000		85.1	69.9-130.1	2.14	20	
2-Butanone	0.243	ug/ml	0.30000		81.0	69.9-130.1	0.708	20	
Benzaldehyde	0.311	ug/ml	0.30000		104	69.9-130.1	1.04	20	
Valeraldehyde	0.254	ug/ml	0.30000		84.8	69.9-130.1	2.11	20	
Hexaldehyde	0.243	ug/ml	0.30000		81.0	69.9-130.1	1.12	20	

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1)

Prepared: 05/20/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E2006-BLK3)

Prepared & Analyzed: 05/20/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E2006-BS1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.298	ug/ml	0.30000		99.2	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.296	ug/ml	0.30000		98.8	69.9-130.1			
Propionaldehyde	0.269	ug/ml	0.30000		89.6	69.9-130.1			
Crotonaldehyde	0.291	ug/ml	0.30000		97.1	69.9-130.1			
Butyraldehyde	0.257	ug/ml	0.30000		85.5	69.9-130.1			
2-Butanone	0.240	ug/ml	0.30000		79.9	69.9-130.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2006 - TO-11A extraction

LCS (B6E2006-BS1) Continued

Prepared & Analyzed: 05/20/26

Benzaldehyde	0.310	ug/ml	0.30000	103	69.9-130.1
Valeraldehyde	0.255	ug/ml	0.30000	85.0	69.9-130.1
Hexaldehyde	0.244	ug/ml	0.30000	81.4	69.9-130.1

LCS Dup (B6E2006-BSD1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.297	ug/ml	0.30000	98.9	79.9-120.1	0.349	20
Acetaldehyde	0.290	ug/ml	0.30000	96.7	79.9-120.1	0.499	20
Acetone	0.294	ug/ml	0.30000	98.1	69.9-130.1	0.653	20
Propionaldehyde	0.267	ug/ml	0.30000	88.9	69.9-130.1	0.876	20
Crotonaldehyde	0.281	ug/ml	0.30000	93.7	69.9-130.1	3.50	20
Butyraldehyde	0.253	ug/ml	0.30000	84.3	69.9-130.1	1.36	20
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.186	20
Benzaldehyde	0.307	ug/ml	0.30000	102	69.9-130.1	0.986	20
Valeraldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	0.252	20
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1	0.678	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/13/26 10:19

SUBMITTED: 04/07/26 to 05/19/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
QB-01	Analyte exceeds method blank criteria
O-05	This sample was extracted outside of the EPA recommended holding time.
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
D-F	Duplicate exceeds DQO criteria.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.