



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 18, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/06/26 10:13 through 05/01/26 11:28.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6040617-02	Air	04/02/26 23:59	04/06/26 10:13
TMOK	6041615-01	Air	04/08/26 23:59	04/16/26 12:39
TMOK	6041615-02	Air	04/14/26 23:59	04/16/26 12:39
TMOK	6042062-01	Air	04/07/26 23:59	04/20/26 11:07
TMOK	6042320-01	Air	04/17/26 23:59	04/23/26 11:15
TMOK	6042320-02	Air	04/20/26 23:59	04/23/26 11:15
TMOK	6050126-01	Air	04/26/26 23:59	05/01/26 11:28



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Description: TMOK
Pressure @ Receipt: 4.0
Comments:

Lab ID: 6040617-02
Canister #: 33266

Sampled: 04/02/26 23:59
Received: 04/06/26 10:13
Analyzed: 04/21/26 13:52

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.515	2.55		0.0237
Chloromethane	0.483	1.00		0.0213
Dichlorotetrafluoroethane	0.0174	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00260	0.01	CF, LL, U	0.0215
Bromomethane	0.00970	0.04	U	0.0104
Chloroethane	0.00520	0.01	U	0.0136
Acetonitrile	199	335.00	E, QB-03	0.184
Acrolein	0.114	0.26		0.0977
Trichlorofluoromethane	0.240	1.35		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.174	0.61		0.0489
Carbon Disulfide	0.00630	0.02	U	0.0473
Trichlorotrifluoroethane	0.0567	0.44		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	5.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	1.00E-4	0.00	U	0.00905
Chloroform	0.0164	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0114	0.05	U	0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.113	0.36		0.0158
Carbon Tetrachloride	0.0702	0.44		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00290	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0117	0.05	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040617-02

Sampled: 04/02/26 23:59

Pressure @ Receipt: 4.0

Canister #: 33266

Received: 04/06/26 10:13

Comments:

Analyzed: 04/21/26 13:52

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.242	0.91		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0374	0.18		0.0258
Tetrachloroethylene	0.00350	0.02	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0319	0.14		0.0130
m,p-Xylene	0.0988	0.43		0.0418
Bromoform	0.00210	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	ND	ND	U	0.0235
1,3,5-Trimethylbenzene	0.00510	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00200	0.01	U	0.0287
o-Dichlorobenzene	4.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00310	0.02	U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6041615-01

Sampled: 04/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 912427

Received: 04/16/26 12:39

Comments:

Analyzed: 04/27/26 21:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.490	2.43		0.0237
Chloromethane	0.473	0.98		0.0213
Dichlorotetrafluoroethane	0.0178	0.13		0.0112
Vinyl chloride	5.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0130	0.03	CF, LL, U	0.0215
Bromomethane	0.00850	0.03	U	0.0104
Chloroethane	0.00800	0.02	U	0.0136
Acetonitrile	184	310.00	E, QB-01	0.184
Acrolein	0.115	0.26	LK	0.0977
Trichlorofluoromethane	0.228	1.28		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.179	0.62		0.0489
Carbon Disulfide	0.00660	0.02	U	0.0473
Trichlorotrifluoroethane	0.0771	0.59		0.0210
trans-1,2-Dichloroethylene	0.00160	0.01	U	0.0171
1,1-Dichloroethane	0.00170	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0251	0.12		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0293	0.12		0.0116
1,1,1-Trichloroethane	0.00160	0.01	U	0.00952
Benzene	0.217	0.70		0.0158
Carbon Tetrachloride	0.0748	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00160	0.01	U	0.00889
Trichloroethylene	0.00600	0.03	U	0.0170
Methyl Methacrylate	0.0315	0.13	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0110	0.05	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6041615-01

Sampled: 04/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 912427

Received: 04/16/26 12:39

Comments:

Analyzed: 04/27/26 21:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.390	1.47		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0365	0.17		0.0258
Tetrachloroethylene	0.00690	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0550	0.24		0.0130
m,p-Xylene	0.164	0.71		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0611	0.27		0.0235
1,3,5-Trimethylbenzene	0.00890	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	1.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00700	0.04	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	2.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Description: TMOK
Pressure @ Receipt: 3.5
Comments:

Lab ID: 6041615-02
Canister #: 42180

Sampled: 04/14/26 23:59
Received: 04/16/26 12:39
Analyzed: 04/27/26 22:20

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.496	2.46		0.0237
Chloromethane	0.593	1.23		0.0213
Dichlorotetrafluoroethane	0.0180	0.13		0.0112
Vinyl chloride	6.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00420	0.01	CF, LL, U	0.0215
Bromomethane	0.00980	0.04	U	0.0104
Chloroethane	0.0155	0.04		0.0136
Acetonitrile	226	380.00	CF, E, LK, QB-01	0.184
Acrolein	0.285	0.66	CF, LK	0.0977
Trichlorofluoromethane	0.232	1.31		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.350	1.22	CF, LK	0.0489
Carbon Disulfide	0.0128	0.04	U	0.0473
Trichlorotrifluoroethane	0.0775	0.60		0.0210
trans-1,2-Dichloroethylene	8.00E-4	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0169	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0130	0.05		0.0116
1,1,1-Trichloroethane	9.00E-4	0.00	U	0.00952
Benzene	0.117	0.38		0.0158
Carbon Tetrachloride	0.0744	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00370	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0113	0.05	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6041615-02

Sampled: 04/14/26 23:59

Pressure @ Receipt: 3.5

Canister #: 42180

Received: 04/16/26 12:39

Comments:

Analyzed: 04/27/26 22:20

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.349	1.32		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0299	0.14		0.0258
Tetrachloroethylene	0.00490	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0470	0.21		0.0130
m,p-Xylene	0.144	0.63		0.0418
Bromoform	0.00340	0.04	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0529	0.23		0.0235
1,3,5-Trimethylbenzene	0.00910	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00210	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	2.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042062-01

Sampled: 04/07/26 23:59

Pressure @ Receipt: 4.0

Canister #: 40450

Received: 04/20/26 11:07

Comments: Makeup for 1-8-26

Analyzed: 04/29/26 01:49

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.476	2.36		0.0237
Chloromethane	0.458	0.95		0.0213
Dichlorotetrafluoroethane	0.0179	0.13		0.0112
Vinyl chloride	0.00100	0.00	U	0.0140
1,3-Butadiene	0.0342	0.08	CF, LL	0.0215
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.00850	0.02	U	0.0136
Acetonitrile	476	801.00	E, QB-01, QB-03 CF, LK	0.184
Acrolein	0.138	0.32		0.0977
Trichlorofluoromethane	0.223	1.26		0.00950
Acrylonitrile	0.0148	0.03		0.0116
1,1-Dichloroethene	7.00E-4	0.00	U	0.0158
Dichloromethane	0.747	2.60		0.0489
Carbon Disulfide	0.00840	0.03	U	0.0473
Trichlorotrifluoroethane	0.0850	0.65	LK	0.0210
trans-1,2-Dichloroethylene	0.00870	0.03	U	0.0171
1,1-Dichloroethane	0.00230	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	9.00E-4	0.00	U	0.00905
Chloroform	0.0291	0.14		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0218	0.09		0.0116
1,1,1-Trichloroethane	0.00180	0.01	U	0.00952
Benzene	0.330	1.06		0.0158
Carbon Tetrachloride	0.0726	0.46		0.00750
tert-Amyl Methyl Ether	5.00E-4	0.00	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00230	0.02	U	0.00889
Trichloroethylene	0.0113	0.06	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0206	0.08	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042062-01

Sampled: 04/07/26 23:59

Pressure @ Receipt: 4.0

Canister #: 40450

Received: 04/20/26 11:07

Comments: Makeup for 1-8-26

Analyzed: 04/29/26 01:49

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.684	2.58		0.0380
Dibromochloromethane	0.00130	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0732	0.34		0.0258
Tetrachloroethylene	0.0110	0.07		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0914	0.40		0.0130
m,p-Xylene	0.313	1.36		0.0418
Bromoform	0.00170	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.115	0.50		0.0235
1,3,5-Trimethylbenzene	0.0223	0.11	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	6.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.0123	0.07	U	0.0287
o-Dichlorobenzene	7.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	5.00E-4	0.01	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Description: TMOK	Lab ID: 6042320-01	Sampled: 04/17/26 23:59
Pressure @ Receipt: 3.5	Canister #: 912395	Received: 04/23/26 11:15
Comments: makeup for 1/26/26		Analyzed: 05/01/26 20:17

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.465	2.30		0.0237
Chloromethane	0.518	1.07		0.0213
Dichlorotetrafluoroethane	0.0157	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00400	0.01	CF, LL, U	0.0215
Bromomethane	0.0111	0.04		0.0104
Chloroethane	0.00750	0.02	U	0.0136
Acetonitrile	743	1,250.00	E, QB-03	0.184
Acrolein	0.128	0.29	CF, LK	0.0977
Trichlorofluoromethane	0.216	1.22		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.554	1.93		0.0489
Carbon Disulfide	0.0119	0.04	U	0.0473
Trichlorotrifluoroethane	0.0736	0.57		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0188	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0152	0.06		0.0116
1,1,1-Trichloroethane	ND	ND	U	0.00952
Carbon Tetrachloride	0.0746	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	7.00E-4	0.00	U	0.00889
Trichloroethylene	0.00840	0.05	U	0.0170
Methyl Methacrylate	ND	ND	LK, U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0102	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Description: TMOK	Lab ID: 6042320-01	Sampled: 04/17/26 23:59
Pressure @ Receipt: 3.5	Canister #: 912395	Received: 04/23/26 11:15
Comments: makeup for 1/26/26		Analyzed: 05/01/26 20:17

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Toluene	0.242	0.91		0.0380
Dibromochloromethane	5.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0247	0.12	U	0.0258
Tetrachloroethylene	0.00390	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0360	0.16		0.0130
m,p-Xylene	0.108	0.47		0.0418
Bromoform	0.00340	0.04	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0406	0.18		0.0235
1,3,5-Trimethylbenzene	0.00740	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00200	0.01	U	0.0287
o-Dichlorobenzene	5.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00110	0.01	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6042320-02

Sampled: 04/20/26 23:59

Pressure @ Receipt: 4

Canister #: 33503

Received: 04/23/26 11:15

Comments:

Analyzed: 05/01/26 20:12

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.470	2.33		0.0237
Chloromethane	0.498	1.03		0.0213
Dichlorotetrafluoroethane	0.0175	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0106	0.02	U	0.0215
Bromomethane	0.00830	0.03	U	0.0104
Chloroethane	0.00710	0.02	U	0.0136
Acetonitrile	154	259.00	E	0.184
Acrolein	0.118	0.27	CF, LK	0.0977
Trichlorofluoromethane	0.214	1.20		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.206	0.72		0.0489
Carbon Disulfide	0.00510	0.02	U	0.0473
Trichlorotrifluoroethane	0.0748	0.57		0.0210
trans-1,2-Dichloroethylene	0.00390	0.02	U	0.0171
1,1-Dichloroethane	0.00150	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	5.00E-4	0.00	U	0.00905
Chloroform	0.0180	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0200	0.08		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.183	0.59		0.0158
Carbon Tetrachloride	0.0673	0.42		0.00750
tert-Amyl Methyl Ether	ND	ND	LL, U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00370	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00760	0.03	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Description: TMOK
Pressure @ Receipt: 4
Comments:

Lab ID: 6042320-02
Canister #: 33503

Sampled: 04/20/26 23:59
Received: 04/23/26 11:15
Analyzed: 05/01/26 20:12

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.419	1.58		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0653	0.31		0.0258
Tetrachloroethylene	0.00770	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0410	0.18		0.0130
m,p-Xylene	0.132	0.57		0.0418
Bromoform	0.00120	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0462	0.20		0.0235
1,3,5-Trimethylbenzene	0.00800	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	1.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00220	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	2.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6050126-01

Sampled: 04/26/26 23:59

Pressure @ Receipt: 4.0

Canister #: 911950

Received: 05/01/26 11:28

Comments:

Analyzed: 05/15/26 18:02

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.433	2.15		0.0237
Chloromethane	0.509	1.05		0.0213
Dichlorotetrafluoroethane	0.0145	0.10		0.0112
Vinyl chloride	6.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00390	0.01	CF, LL, U	0.0215
Bromomethane	0.00790	0.03	U	0.0104
Chloroethane	0.00580	0.02	U	0.0136
Acetonitrile	153	257.00	CF, E, LK, QB-01	0.184
Acrolein	0.148	0.34	CF, LK, QB-01	0.0977
Trichlorofluoromethane	0.227	1.28	LK	0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	LJ, U	0.0158
Dichloromethane	0.274	0.95		0.0489
Carbon Disulfide	0.0101	0.03	U	0.0473
Trichlorotrifluoroethane	0.0526	0.40		0.0210
trans-1,2-Dichloroethylene	3.00E-4	0.00	U	0.0171
1,1-Dichloroethane	8.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0154	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0141	0.06		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.0828	0.27		0.0158
Carbon Tetrachloride	0.0726	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	8.00E-4	0.01	U	0.00889
Trichloroethylene	0.00250	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	ND	ND	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6050126-01

Sampled: 04/26/26 23:59

Pressure @ Receipt: 4.0

Canister #: 911950

Received: 05/01/26 11:28

Comments:

Analyzed: 05/15/26 18:02

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.133	0.50		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0181	0.08	U	0.0258
Tetrachloroethylene	0.00440	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0148	0.06		0.0130
m,p-Xylene	0.0433	0.19		0.0418
Bromoform	0.00200	0.02	U	0.0166
Styrene	0.00790	0.03	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0208	0.09	U	0.0235
1,3,5-Trimethylbenzene	0.00320	0.02	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00270	0.02	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	7.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2108 - Summa Canister Prep

Blank (B6D2108-BLK1)

Prepared: 04/14/26 Analyzed: 04/21/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2108 - Summa Canister Prep

Blank (B6D2108-BLK1) Continued

Prepared: 04/14/26 Analyzed: 04/21/26

1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D2708 - Summa Canister Prep

Blank (B6D2708-BLK1)

Prepared: 04/24/26 Analyzed: 04/27/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2708 - Summa Canister Prep

Blank (B6D2708-BLK1) Continued

Prepared: 04/24/26 Analyzed: 04/27/26

Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D2809 - Summa Canister Prep

Blank (B6D2809-BLK1)

Prepared: 04/24/26 Analyzed: 04/28/26

Dichlorodifluoromethane	ND	ppbv				U
-------------------------	----	------	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2809 - Summa Canister Prep

Blank (B6D2809-BLK1) Continued

Prepared: 04/24/26 Analyzed: 04/28/26

Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				LK, U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2809 - Summa Canister Prep

Blank (B6D2809-BLK1) Continued

Prepared: 04/24/26 Analyzed: 04/28/26

n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1)

Prepared: 04/29/26 Analyzed: 05/01/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1) Continued

Prepared: 04/29/26 Analyzed: 05/01/26

Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				LL, U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E1210 - Summa Canister Prep

Blank (B6E1210-BLK1)

Prepared: 04/30/26 Analyzed: 05/01/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1210 - Summa Canister Prep

Blank (B6E1210-BLK1) Continued

Prepared: 04/30/26 Analyzed: 05/01/26

Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				LK, U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1210 - Summa Canister Prep

Blank (B6E1210-BLK1) Continued

Prepared: 04/30/26 Analyzed: 05/01/26

Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E1502 - Summa Canister Prep

Blank (B6E1502-BLK1)

Prepared: 04/29/26 Analyzed: 05/15/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, QB-01, U
Trichlorofluoromethane	ND	ppbv				LK, U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				LJ, U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1502 - Summa Canister Prep

Blank (B6E1502-BLK1) Continued

Prepared: 04/29/26 Analyzed: 05/15/26

Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Duplicate (B6E1502-DUP1)

Source: 6050126-01

Prepared: 04/26/26 Analyzed: 05/15/26

Dichlorodifluoromethane	0.432	ppbv	0.43	0.254	25	
Chloromethane	0.508	ppbv	0.51	0.256	25	
Dichlorotetrafluoroethane	0.0147	ppbv	0.01	1.37	25	
Vinyl chloride	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1502 - Summa Canister Prep

Duplicate (B6E1502-DUP1) Continued

Source: 6050126-01

Prepared: 04/26/26 Analyzed: 05/15/26

1,3-Butadiene	ND	ppbv	ND		25	CF, LL, U
Bromomethane	ND	ppbv	ND		25	U
Chloroethane	ND	ppbv	ND		25	U
Acetonitrile	153	ppbv	153.00	0.0918	25	CF, E, LK, QB-01
Acrolein	0.148	ppbv	0.15	0.0676	25	CF, LK, QB-01
Trichlorofluoromethane	0.228	ppbv	0.23	0.484	25	LK
Acrylonitrile	ND	ppbv	ND		25	U
1,1-Dichloroethene	ND	ppbv	ND		25	LJ, U
Dichloromethane	0.273	ppbv	0.27	0.366	25	
Carbon Disulfide	ND	ppbv	ND		25	U
Trichlorotrifluoroethane	0.0525	ppbv	0.05	0.190	25	
trans-1,2-Dichloroethylene	ND	ppbv	ND		25	U
1,1-Dichloroethane	ND	ppbv	ND		25	U
Methyl tert-Butyl Ether	ND	ppbv	ND		25	U
Chloroprene	ND	ppbv	ND		25	U
cis-1,2-Dichloroethylene	ND	ppbv	ND		25	U
Bromochloromethane	ND	ppbv	ND		25	U
Chloroform	0.0157	ppbv	0.02	1.93	25	
Ethyl tert-Butyl Ether	ND	ppbv	ND		25	U
1,2-Dichloroethane	0.0142	ppbv	0.01	0.707	25	
1,1,1-Trichloroethane	ND	ppbv	ND		25	U
Benzene	0.0826	ppbv	0.08	0.242	25	
Carbon Tetrachloride	0.0725	ppbv	0.07	0.138	25	
tert-Amyl Methyl Ether	ND	ppbv	ND		25	U
1,2-Dichloropropane	ND	ppbv	ND		25	U
Ethyl Acrylate	ND	ppbv	ND		25	U
Bromodichloromethane	ND	ppbv	ND		25	U
Trichloroethylene	ND	ppbv	ND		25	U
Methyl Methacrylate	ND	ppbv	ND		25	U
cis-1,3-Dichloropropene	ND	ppbv	ND		25	U
Methyl Isobutyl Ketone	ND	ppbv	ND		25	U
trans-1,3-Dichloropropene	ND	ppbv	ND		25	U
1,1,2-Trichloroethane	ND	ppbv	ND		25	U
Toluene	0.133	ppbv	0.13	0.451	25	
Dibromochloromethane	ND	ppbv	ND		25	U
1,2-Dibromoethane	ND	ppbv	ND		25	U
n-Octane	ND	ppbv	ND		25	U
Tetrachloroethylene	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1502 - Summa Canister Prep

Duplicate (B6E1502-DUP1) Continued

Source: 6050126-01

Prepared: 04/26/26 Analyzed: 05/15/26

Chlorobenzene	ND	ppbv	ND		25	U
Ethylbenzene	0.0148	ppbv	0.01	0.00	25	
m,p-Xylene	0.0431	ppbv	0.04	0.463	25	
Bromoform	ND	ppbv	ND		25	U
Styrene	ND	ppbv	ND		25	U
1,1,2,2-Tetrachloroethane	ND	ppbv	ND		25	U
o-Xylene	ND	ppbv	ND		25	U
1,3,5-Trimethylbenzene	ND	ppbv	ND		25	U
1,2,4-Trimethylbenzene	ND	ppbv	ND		25	CE, U
m-Dichlorobenzene	ND	ppbv	ND		25	U
p-Dichlorobenzene	ND	ppbv	ND		25	U
o-Dichlorobenzene	ND	ppbv	ND		25	U
1,2,4-Trichlorobenzene	ND	ppbv	ND		25	U
Hexachloro-1,3-butadiene	ND	ppbv	ND		25	U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604068

Calibration Check (2604068-CCV1)

Prepared & Analyzed: 04/21/26

Dichlorodifluoromethane	1.17	ppbv	14.1	30.00
Chloromethane	1.01	ppbv	-1.2	30.00
Dichlorotetrafluoroethane	1.15	ppbv	10.7	30.00
Vinyl chloride	0.891	ppbv	-13.7	30.00
1,3-Butadiene	0.844	ppbv	-18.4	30.00
Bromomethane	1.03	ppbv	-0.3	30.00
Chloroethane	1.08	ppbv	3.9	30.00
Acetonitrile	1.07	ppbv	3.4	30.00
Acrolein	1.14	ppbv	12.0	30.00
Trichlorofluoromethane	1.21	ppbv	20.1	30.00
Acrylonitrile	1.17	ppbv	5.9	30.00
1,1-Dichloroethene	1.22	ppbv	17.2	30.00
Dichloromethane	1.12	ppbv	8.1	30.00
Carbon Disulfide	0.917	ppbv	-15.7	30.00
Trichlorotrifluoroethane	0.907	ppbv	-11.9	30.00
trans-1,2-Dichloroethylene	0.977	ppbv	-6.0	30.00
1,1-Dichloroethane	1.12	ppbv	8.2	30.00
Methyl tert-Butyl Ether	1.09	ppbv	5.1	30.00
Chloroprene	1.07	ppbv	2.9	30.00
cis-1,2-Dichloroethylene	1.11	ppbv	6.7	30.00
Bromochloromethane	1.03	ppbv	0.3	30.00
Chloroform	1.23	ppbv	17.0	30.00
Ethyl tert-Butyl Ether	1.09	ppbv	5.1	30.00
1,2-Dichloroethane	1.21	ppbv	17.0	30.00
1,1,1-Trichloroethane	1.27	ppbv	23.1	30.00
Benzene	1.09	ppbv	5.3	30.00
Carbon Tetrachloride	1.10	ppbv	7.1	30.00
tert-Amyl Methyl Ether	1.06	ppbv	2.5	30.00
1,2-Dichloropropane	1.09	ppbv	4.8	30.00
Ethyl Acrylate	0.983	ppbv	-6.0	30.00
Bromodichloromethane	1.08	ppbv	3.8	30.00
Trichloroethylene	0.938	ppbv	-9.3	30.00
Methyl Methacrylate	0.976	ppbv	-6.3	30.00
cis-1,3-Dichloropropene	1.13	ppbv	8.7	30.00
Methyl Isobutyl Ketone	1.00	ppbv	-3.6	30.00
trans-1,3-Dichloropropene	1.20	ppbv	14.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604068

Calibration Check (2604068-CCV1) Continued

Prepared & Analyzed: 04/21/26

1,1,2-Trichloroethane	0.942	ppbv	-9.9	30.00
Toluene	1.02	ppbv	-2.0	30.00
Dibromochloromethane	1.04	ppbv	0.2	30.00
1,2-Dibromoethane	0.896	ppbv	-14.1	30.00
n-Octane	0.944	ppbv	-9.0	30.00
Tetrachloroethylene	1.08	ppbv	3.9	30.00
Chlorobenzene	0.937	ppbv	-10.9	30.00
Ethylbenzene	0.865	ppbv	-17.3	30.00
m,p-Xylene	1.81	ppbv	-13.5	30.00
Bromoform	1.22	ppbv	17.0	30.00
Styrene	0.951	ppbv	-10.1	30.00
1,1,2,2-Tetrachloroethane	1.27	ppbv	20.4	30.00
o-Xylene	1.00	ppbv	-5.3	30.00
1,3,5-Trimethylbenzene	0.926	ppbv	-12.0	30.00
1,2,4-Trimethylbenzene	0.894	ppbv	-15.2	30.00
m-Dichlorobenzene	0.928	ppbv	-11.0	30.00
p-Dichlorobenzene	0.915	ppbv	-12.5	30.00
o-Dichlorobenzene	0.930	ppbv	-11.1	30.00
1,2,4-Trichlorobenzene	1.10	ppbv	6.4	30.00
Hexachloro-1,3-butadiene	1.18	ppbv	11.5	30.00

Calibration Check (2604068-CCV2)

Prepared: 04/21/26 Analyzed: 04/22/26

Dichlorodifluoromethane	1.19	ppbv	15.8	30.00
Chloromethane	1.02	ppbv	0.09	30.00
Dichlorotetrafluoroethane	1.17	ppbv	13.0	30.00
Vinyl chloride	0.841	ppbv	-18.5	30.00
1,3-Butadiene	0.791	ppbv	-23.5	30.00
Bromomethane	1.04	ppbv	1.0	30.00
Chloroethane	1.13	ppbv	9.0	30.00
Acetonitrile	1.11	ppbv	7.3	30.00
Acrolein	1.13	ppbv	10.7	30.00
Trichlorofluoromethane	1.26	ppbv	24.3	30.00
Acrylonitrile	1.20	ppbv	8.3	30.00
1,1-Dichloroethene	1.20	ppbv	15.4	30.00
Dichloromethane	1.17	ppbv	12.6	30.00
Carbon Disulfide	1.08	ppbv	-0.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604068

Calibration Check (2604068-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

Trichlorotrifluoroethane	0.925	ppbv	-10.2	30.00
trans-1,2-Dichloroethylene	0.955	ppbv	-8.1	30.00
1,1-Dichloroethane	1.17	ppbv	12.8	30.00
Methyl tert-Butyl Ether	1.04	ppbv	0.2	30.00
Chloroprene	1.02	ppbv	-1.8	30.00
cis-1,2-Dichloroethylene	1.10	ppbv	6.6	30.00
Bromochloromethane	1.06	ppbv	3.2	30.00
Chloroform	1.29	ppbv	22.7	30.00
Ethyl tert-Butyl Ether	0.972	ppbv	-6.2	30.00
1,2-Dichloroethane	1.26	ppbv	21.7	30.00
1,1,1-Trichloroethane	1.30	ppbv	26.5	30.00
Benzene	1.11	ppbv	6.8	30.00
Carbon Tetrachloride	1.10	ppbv	7.7	30.00
tert-Amyl Methyl Ether	0.972	ppbv	-6.0	30.00
1,2-Dichloropropane	1.10	ppbv	5.8	30.00
Ethyl Acrylate	0.909	ppbv	-13.1	30.00
Bromodichloromethane	1.13	ppbv	9.2	30.00
Trichloroethylene	0.911	ppbv	-11.9	30.00
Methyl Methacrylate	0.915	ppbv	-12.2	30.00
cis-1,3-Dichloropropene	1.09	ppbv	4.5	30.00
Methyl Isobutyl Ketone	0.967	ppbv	-7.0	30.00
trans-1,3-Dichloropropene	1.18	ppbv	12.0	30.00
1,1,2-Trichloroethane	0.955	ppbv	-8.7	30.00
Toluene	0.979	ppbv	-5.8	30.00
Dibromochloromethane	1.05	ppbv	1.5	30.00
1,2-Dibromoethane	0.881	ppbv	-15.6	30.00
n-Octane	0.865	ppbv	-16.6	30.00
Tetrachloroethylene	1.06	ppbv	1.2	30.00
Chlorobenzene	0.921	ppbv	-12.4	30.00
Ethylbenzene	0.822	ppbv	-21.4	30.00
m,p-Xylene	1.72	ppbv	-17.6	30.00
Bromoform	1.21	ppbv	16.1	30.00
Styrene	0.912	ppbv	-13.8	30.00
1,1,2,2-Tetrachloroethane	1.33	ppbv	26.2	30.00
o-Xylene	0.927	ppbv	-12.3	30.00
1,3,5-Trimethylbenzene	0.832	ppbv	-20.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604068

Calibration Check (2604068-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

1,2,4-Trimethylbenzene	0.803	ppbv	-23.8	30.00
m-Dichlorobenzene	0.911	ppbv	-12.6	30.00
p-Dichlorobenzene	0.890	ppbv	-14.9	30.00
o-Dichlorobenzene	0.908	ppbv	-13.2	30.00
1,2,4-Trichlorobenzene	1.10	ppbv	5.9	30.00
Hexachloro-1,3-butadiene	1.18	ppbv	11.9	30.00

Sequence 2604069

Calibration Check (2604069-CCV1)

Prepared & Analyzed: 04/21/26

Dichlorodifluoromethane	1.16	ppbv	13.1	30.00
Chloromethane	1.08	ppbv	5.5	30.00
Dichlorotetrafluoroethane	1.16	ppbv	12.2	30.00
Vinyl chloride	1.04	ppbv	0.6	30.00
1,3-Butadiene	0.907	ppbv	-12.3	30.00
Bromomethane	1.19	ppbv	15.2	30.00
Chloroethane	1.16	ppbv	11.8	30.00
Acetonitrile	0.806	ppbv	-22.4	30.00
Acrolein	1.11	ppbv	8.6	30.00
Trichlorofluoromethane	1.19	ppbv	18.3	30.00
Acrylonitrile	1.12	ppbv	1.7	30.00
1,1-Dichloroethene	1.23	ppbv	18.5	30.00
Dichloromethane	1.03	ppbv	-0.8	30.00
Carbon Disulfide	1.29	ppbv	18.4	30.00
Trichlorotrifluoroethane	1.28	ppbv	24.7	30.00
trans-1,2-Dichloroethylene	1.13	ppbv	8.8	30.00
1,1-Dichloroethane	1.12	ppbv	8.2	30.00
Methyl tert-Butyl Ether	0.942	ppbv	-9.1	30.00
Chloroprene	0.950	ppbv	-8.5	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.2	30.00
Bromochloromethane	1.17	ppbv	13.8	30.00
Chloroform	1.16	ppbv	10.8	30.00
Ethyl tert-Butyl Ether	0.866	ppbv	-16.5	30.00
1,2-Dichloroethane	1.12	ppbv	8.4	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.4	30.00
Benzene	1.06	ppbv	2.1	30.00
Carbon Tetrachloride	1.14	ppbv	11.5	30.00

QB-01
LK, QB-01

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV1) Continued

Prepared & Analyzed: 04/21/26

tert-Amyl Methyl Ether	0.823	ppbv	-20.4	30.00
1,2-Dichloropropane	0.981	ppbv	-5.5	30.00
Ethyl Acrylate	0.895	ppbv	-14.4	30.00
Bromodichloromethane	1.07	ppbv	3.0	30.00
Trichloroethylene	1.05	ppbv	1.8	30.00
Methyl Methacrylate	0.873	ppbv	-16.2	30.00
cis-1,3-Dichloropropene	0.955	ppbv	-8.5	30.00
Methyl Isobutyl Ketone	0.906	ppbv	-12.9	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-2.2	30.00
1,1,2-Trichloroethane	1.10	ppbv	5.3	30.00
Toluene	0.982	ppbv	-5.6	30.00
Dibromochloromethane	1.08	ppbv	4.1	30.00
1,2-Dibromoethane	1.06	ppbv	1.8	30.00
n-Octane	0.894	ppbv	-13.9	30.00
Tetrachloroethylene	1.10	ppbv	5.0	30.00
Chlorobenzene	1.10	ppbv	4.6	30.00
Ethylbenzene	0.985	ppbv	-5.8	30.00
m,p-Xylene	2.07	ppbv	-0.7	30.00
Bromoform	1.19	ppbv	13.7	30.00
Styrene	1.07	ppbv	1.2	30.00
1,1,2,2-Tetrachloroethane	1.13	ppbv	7.4	30.00
o-Xylene	0.983	ppbv	-6.9	30.00
1,3,5-Trimethylbenzene	0.924	ppbv	-12.2	30.00
1,2,4-Trimethylbenzene	0.912	ppbv	-13.4	30.00
m-Dichlorobenzene	1.07	ppbv	2.7	30.00
p-Dichlorobenzene	1.05	ppbv	0.8	30.00
o-Dichlorobenzene	1.06	ppbv	1.1	30.00
1,2,4-Trichlorobenzene	0.956	ppbv	-7.9	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.6	30.00

Calibration Check (2604069-CCV2)

Prepared: 04/21/26 Analyzed: 04/22/26

Dichlorodifluoromethane	1.18	ppbv	14.4	30.00
Chloromethane	1.08	ppbv	5.6	30.00
Dichlorotetrafluoroethane	1.18	ppbv	13.7	30.00
Vinyl chloride	1.02	ppbv	-1.0	30.00
1,3-Butadiene	0.837	ppbv	-19.0	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

Bromomethane	1.22	ppbv	18.1	30.00	
Chloroethane	1.16	ppbv	12.4	30.00	
Acetonitrile	0.804	ppbv	-22.5	30.00	QB-01
Acrolein	1.07	ppbv	5.1	30.00	LK, QB-01
Trichlorofluoromethane	1.21	ppbv	20.0	30.00	
Acrylonitrile	1.10	ppbv	-0.6	30.00	
1,1-Dichloroethene	1.21	ppbv	16.5	30.00	
Dichloromethane	1.14	ppbv	10.2	30.00	
Carbon Disulfide	1.28	ppbv	17.5	30.00	
Trichlorotrifluoroethane	1.28	ppbv	24.1	30.00	
trans-1,2-Dichloroethylene	1.14	ppbv	10.0	30.00	
1,1-Dichloroethane	1.14	ppbv	10.3	30.00	
Methyl tert-Butyl Ether	0.916	ppbv	-11.5	30.00	
Chloroprene	0.913	ppbv	-12.1	30.00	
cis-1,2-Dichloroethylene	1.07	ppbv	3.6	30.00	
Bromochloromethane	1.22	ppbv	18.2	30.00	
Chloroform	1.18	ppbv	12.6	30.00	
Ethyl tert-Butyl Ether	0.848	ppbv	-18.1	30.00	
1,2-Dichloroethane	1.13	ppbv	9.2	30.00	
1,1,1-Trichloroethane	1.14	ppbv	10.4	30.00	
Benzene	1.06	ppbv	2.3	30.00	
Carbon Tetrachloride	1.17	ppbv	14.0	30.00	
tert-Amyl Methyl Ether	0.810	ppbv	-21.7	30.00	
1,2-Dichloropropane	0.993	ppbv	-4.3	30.00	
Ethyl Acrylate	0.878	ppbv	-16.1	30.00	
Bromodichloromethane	1.07	ppbv	3.4	30.00	
Trichloroethylene	1.07	ppbv	3.3	30.00	
Methyl Methacrylate	0.855	ppbv	-18.0	30.00	
cis-1,3-Dichloropropene	0.968	ppbv	-7.3	30.00	
Methyl Isobutyl Ketone	0.915	ppbv	-12.0	30.00	
trans-1,3-Dichloropropene	1.05	ppbv	0.2	30.00	
1,1,2-Trichloroethane	1.15	ppbv	9.5	30.00	
Toluene	1.00	ppbv	-3.9	30.00	
Dibromochloromethane	1.12	ppbv	8.0	30.00	
1,2-Dibromoethane	1.12	ppbv	6.9	30.00	
n-Octane	0.874	ppbv	-15.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

Tetrachloroethylene	1.15	ppbv	10.4	30.00
Chlorobenzene	1.10	ppbv	4.1	30.00
Ethylbenzene	0.957	ppbv	-8.5	30.00
m,p-Xylene	2.00	ppbv	-4.4	30.00
Bromoform	1.22	ppbv	16.2	30.00
Styrene	1.06	ppbv	0.1	30.00
1,1,2,2-Tetrachloroethane	1.12	ppbv	5.8	30.00
o-Xylene	0.962	ppbv	-8.9	30.00
1,3,5-Trimethylbenzene	0.900	ppbv	-14.4	30.00
1,2,4-Trimethylbenzene	0.903	ppbv	-14.4	30.00
m-Dichlorobenzene	1.08	ppbv	3.8	30.00
p-Dichlorobenzene	1.06	ppbv	1.4	30.00
o-Dichlorobenzene	1.08	ppbv	3.6	30.00
1,2,4-Trichlorobenzene	0.864	ppbv	-16.7	30.00
Hexachloro-1,3-butadiene	0.998	ppbv	-5.4	30.00

Sequence 2604083

Calibration Check (2604083-CCV1)

Prepared & Analyzed: 04/27/26

Dichlorodifluoromethane	1.13	ppbv	9.9	30.00
Chloromethane	1.03	ppbv	0.8	30.00
Dichlorotetrafluoroethane	1.13	ppbv	9.4	30.00
Vinyl chloride	1.03	ppbv	-0.3	30.00
1,3-Butadiene	0.895	ppbv	-13.5	30.00
Bromomethane	1.17	ppbv	13.0	30.00
Chloroethane	1.13	ppbv	9.5	30.00
Acetonitrile	1.05	ppbv	1.0	30.00
Acrolein	1.04	ppbv	2.4	30.00
Trichlorofluoromethane	1.18	ppbv	16.7	30.00
Acrylonitrile	1.08	ppbv	-1.8	30.00
1,1-Dichloroethene	1.16	ppbv	11.2	30.00
Dichloromethane	1.09	ppbv	5.3	30.00
Carbon Disulfide	1.23	ppbv	13.0	30.00
Trichlorotrifluoroethane	1.23	ppbv	19.7	30.00
trans-1,2-Dichloroethylene	1.11	ppbv	6.7	30.00
1,1-Dichloroethane	1.12	ppbv	7.9	30.00
Methyl tert-Butyl Ether	0.932	ppbv	-10.0	30.00

QB-01
LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV1) Continued

Prepared & Analyzed: 04/27/26

Chloroprene	0.898	ppbv	-13.5	30.00
cis-1,2-Dichloroethylene	1.05	ppbv	1.0	30.00
Bromochloromethane	1.19	ppbv	15.5	30.00
Chloroform	1.14	ppbv	9.2	30.00
Ethyl tert-Butyl Ether	0.874	ppbv	-15.6	30.00
1,2-Dichloroethane	1.10	ppbv	6.8	30.00
1,1,1-Trichloroethane	1.13	ppbv	9.5	30.00
Benzene	1.05	ppbv	1.4	30.00
Carbon Tetrachloride	1.15	ppbv	11.8	30.00
tert-Amyl Methyl Ether	0.844	ppbv	-18.3	30.00
1,2-Dichloropropane	1.07	ppbv	2.9	30.00
Ethyl Acrylate	0.855	ppbv	-18.3	30.00
Bromodichloromethane	1.04	ppbv	0.3	30.00
Trichloroethylene	1.05	ppbv	2.0	30.00
Methyl Methacrylate	0.825	ppbv	-20.8	30.00
cis-1,3-Dichloropropene	0.935	ppbv	-10.4	30.00
Methyl Isobutyl Ketone	0.832	ppbv	-20.0	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-2.3	30.00
1,1,2-Trichloroethane	1.09	ppbv	4.4	30.00
Toluene	0.970	ppbv	-6.8	30.00
Dibromochloromethane	1.06	ppbv	2.4	30.00
1,2-Dibromoethane	1.04	ppbv	-0.3	30.00
n-Octane	0.844	ppbv	-18.7	30.00
Tetrachloroethylene	1.13	ppbv	7.8	30.00
Chlorobenzene	1.10	ppbv	5.0	30.00
Ethylbenzene	0.954	ppbv	-8.8	30.00
m,p-Xylene	2.01	ppbv	-3.7	30.00
Bromoform	1.17	ppbv	12.1	30.00
Styrene	1.04	ppbv	-1.9	30.00
1,1,2,2-Tetrachloroethane	1.10	ppbv	4.1	30.00
o-Xylene	0.934	ppbv	-11.5	30.00
1,3,5-Trimethylbenzene	0.865	ppbv	-17.7	30.00
1,2,4-Trimethylbenzene	0.852	ppbv	-19.2	30.00
m-Dichlorobenzene	1.07	ppbv	2.9	30.00
p-Dichlorobenzene	1.05	ppbv	0.3	30.00
o-Dichlorobenzene	1.05	ppbv	0.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV1) Continued

Prepared & Analyzed: 04/27/26

1,2,4-Trichlorobenzene	0.911	ppbv	-12.2	30.00
Hexachloro-1,3-butadiene	1.04	ppbv	-1.2	30.00

Calibration Check (2604083-CCV2)

Prepared: 04/27/26 Analyzed: 04/28/26

Dichlorodifluoromethane	1.11	ppbv	7.9	30.00
Chloromethane	1.03	ppbv	1.5	30.00
Dichlorotetrafluoroethane	1.13	ppbv	9.5	30.00
Vinyl chloride	0.962	ppbv	-6.8	30.00
1,3-Butadiene	0.755	ppbv	-27.0	30.00
Bromomethane	1.16	ppbv	12.6	30.00
Chloroethane	1.14	ppbv	10.1	30.00
Acetonitrile	0.807	ppbv	-22.2	30.00
Acrolein	1.06	ppbv	3.9	30.00
Trichlorofluoromethane	1.14	ppbv	12.6	30.00
Acrylonitrile	1.09	ppbv	-1.1	30.00
1,1-Dichloroethene	1.16	ppbv	11.8	30.00
Dichloromethane	1.15	ppbv	10.8	30.00
Carbon Disulfide	1.24	ppbv	14.0	30.00
Trichlorotrifluoroethane	1.23	ppbv	19.2	30.00
trans-1,2-Dichloroethylene	1.09	ppbv	5.2	30.00
1,1-Dichloroethane	1.12	ppbv	7.7	30.00
Methyl tert-Butyl Ether	0.871	ppbv	-15.9	30.00
Chloroprene	0.834	ppbv	-19.6	30.00
cis-1,2-Dichloroethylene	1.02	ppbv	-1.2	30.00
Bromochloromethane	1.16	ppbv	13.1	30.00
Chloroform	1.13	ppbv	8.1	30.00
Ethyl tert-Butyl Ether	0.796	ppbv	-23.2	30.00
1,2-Dichloroethane	1.08	ppbv	4.1	30.00
1,1,1-Trichloroethane	1.08	ppbv	4.9	30.00
Benzene	1.02	ppbv	-2.1	30.00
Carbon Tetrachloride	1.09	ppbv	6.1	30.00
tert-Amyl Methyl Ether	0.758	ppbv	-26.7	30.00
1,2-Dichloropropane	0.998	ppbv	-3.9	30.00
Ethyl Acrylate	0.835	ppbv	-20.2	30.00
Bromodichloromethane	1.05	ppbv	1.4	30.00
Trichloroethylene	1.04	ppbv	0.4	30.00

QB-01
LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV2) Continued

Prepared: 04/27/26 Analyzed: 04/28/26

Methyl Methacrylate	0.804	ppbv	-22.9	30.00
cis-1,3-Dichloropropene	0.896	ppbv	-14.2	30.00
Methyl Isobutyl Ketone	0.823	ppbv	-20.9	30.00
trans-1,3-Dichloropropene	1.01	ppbv	-3.4	30.00
1,1,2-Trichloroethane	1.12	ppbv	6.9	30.00
Toluene	0.943	ppbv	-9.3	30.00
Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.03	ppbv	-1.3	30.00
n-Octane	0.836	ppbv	-19.5	30.00
Tetrachloroethylene	1.09	ppbv	4.1	30.00
Chlorobenzene	1.11	ppbv	5.1	30.00
Ethylbenzene	0.922	ppbv	-11.9	30.00
m,p-Xylene	1.98	ppbv	-5.1	30.00
Bromoform	1.18	ppbv	12.5	30.00
Styrene	1.03	ppbv	-2.6	30.00
1,1,2,2-Tetrachloroethane	1.15	ppbv	9.1	30.00
o-Xylene	0.892	ppbv	-15.5	30.00
1,3,5-Trimethylbenzene	0.790	ppbv	-24.9	30.00
1,2,4-Trimethylbenzene	0.777	ppbv	-26.3	30.00
m-Dichlorobenzene	1.04	ppbv	-0.5	30.00
p-Dichlorobenzene	1.00	ppbv	-4.1	30.00
o-Dichlorobenzene	1.00	ppbv	-4.2	30.00
1,2,4-Trichlorobenzene	0.830	ppbv	-20.0	30.00
Hexachloro-1,3-butadiene	1.01	ppbv	-4.6	30.00

Sequence 2604090

Calibration Check (2604090-CCV1)

Prepared & Analyzed: 04/28/26

Dichlorodifluoromethane	1.14	ppbv	11.1	30.00
Chloromethane	1.07	ppbv	4.8	30.00
Dichlorotetrafluoroethane	1.17	ppbv	12.9	30.00
Vinyl chloride	1.00	ppbv	-3.1	30.00
1,3-Butadiene	0.812	ppbv	-21.5	30.00
Bromomethane	1.20	ppbv	16.6	30.00
Chloroethane	1.18	ppbv	14.2	30.00
Acetonitrile	1.07	ppbv	3.0	30.00
Acrolein	1.08	ppbv	5.7	30.00

QB-01

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604090

Calibration Check (2604090-CCV1) Continued

Prepared & Analyzed: 04/28/26

Trichlorofluoromethane	1.18	ppbv	16.8	30.00
Acrylonitrile	1.10	ppbv	-0.4	30.00
1,1-Dichloroethene	1.21	ppbv	16.3	30.00
Dichloromethane	1.12	ppbv	8.2	30.00
Carbon Disulfide	1.28	ppbv	17.7	30.00
Trichlorotrifluoroethane	1.27	ppbv	22.9	30.00
trans-1,2-Dichloroethylene	1.10	ppbv	5.9	30.00
1,1-Dichloroethane	1.15	ppbv	11.2	30.00
Methyl tert-Butyl Ether	0.875	ppbv	-15.5	30.00
Chloroprene	0.896	ppbv	-13.7	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.1	30.00
Bromochloromethane	1.20	ppbv	16.6	30.00
Chloroform	1.17	ppbv	11.7	30.00
Ethyl tert-Butyl Ether	0.836	ppbv	-19.3	30.00
1,2-Dichloroethane	1.12	ppbv	7.9	30.00
1,1,1-Trichloroethane	1.12	ppbv	8.3	30.00
Benzene	1.06	ppbv	1.9	30.00
Carbon Tetrachloride	1.13	ppbv	10.2	30.00
tert-Amyl Methyl Ether	0.804	ppbv	-22.2	30.00
1,2-Dichloropropane	1.03	ppbv	-0.5	30.00
Ethyl Acrylate	0.882	ppbv	-15.7	30.00
Bromodichloromethane	1.09	ppbv	4.9	30.00
Trichloroethylene	1.07	ppbv	3.6	30.00
Methyl Methacrylate	0.849	ppbv	-18.5	30.00
cis-1,3-Dichloropropene	0.939	ppbv	-10.0	30.00
Methyl Isobutyl Ketone	0.875	ppbv	-15.9	30.00
trans-1,3-Dichloropropene	1.06	ppbv	0.8	30.00
1,1,2-Trichloroethane	1.15	ppbv	10.0	30.00
Toluene	0.988	ppbv	-5.0	30.00
Dibromochloromethane	1.09	ppbv	5.0	30.00
1,2-Dibromoethane	1.06	ppbv	1.5	30.00
n-Octane	0.882	ppbv	-15.0	30.00
Tetrachloroethylene	1.12	ppbv	7.6	30.00
Chlorobenzene	1.13	ppbv	7.5	30.00
Ethylbenzene	0.959	ppbv	-8.3	30.00
m,p-Xylene	2.06	ppbv	-1.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604090

Calibration Check (2604090-CCV1) Continued

Prepared & Analyzed: 04/28/26

Bromoform	1.20	ppbv	14.3	30.00
Styrene	1.08	ppbv	1.9	30.00
1,1,2,2-Tetrachloroethane	1.17	ppbv	10.7	30.00
o-Xylene	0.934	ppbv	-11.5	30.00
1,3,5-Trimethylbenzene	0.848	ppbv	-19.4	30.00
1,2,4-Trimethylbenzene	0.824	ppbv	-21.8	30.00
m-Dichlorobenzene	1.07	ppbv	2.4	30.00
p-Dichlorobenzene	1.04	ppbv	-0.8	30.00
o-Dichlorobenzene	1.04	ppbv	-1.0	30.00
1,2,4-Trichlorobenzene	0.880	ppbv	-15.2	30.00
Hexachloro-1,3-butadiene	1.03	ppbv	-2.2	30.00

Calibration Check (2604090-CCV2)

Prepared: 04/28/26 Analyzed: 04/29/26

Dichlorodifluoromethane	1.15	ppbv	12.0	30.00
Chloromethane	1.08	ppbv	5.8	30.00
Dichlorotetrafluoroethane	1.19	ppbv	14.5	30.00
Vinyl chloride	0.977	ppbv	-5.3	30.00
1,3-Butadiene	0.747	ppbv	-27.8	30.00
Bromomethane	1.22	ppbv	18.6	30.00
Chloroethane	1.20	ppbv	16.3	30.00
Acetonitrile	1.17	ppbv	12.9	30.00
Acrolein	1.02	ppbv	-0.3	30.00
Trichlorofluoromethane	1.18	ppbv	17.0	30.00
Acrylonitrile	1.03	ppbv	-6.7	30.00
1,1-Dichloroethene	1.28	ppbv	23.5	30.00
Dichloromethane	1.11	ppbv	6.9	30.00
Carbon Disulfide	1.41	ppbv	29.8	30.00
Trichlorotrifluoroethane	1.42	ppbv	37.9	30.00
trans-1,2-Dichloroethylene	1.05	ppbv	0.9	30.00
1,1-Dichloroethane	1.15	ppbv	11.4	30.00
Methyl tert-Butyl Ether	0.797	ppbv	-23.1	30.00
Chloroprene	0.845	ppbv	-18.6	30.00
cis-1,2-Dichloroethylene	1.06	ppbv	2.0	30.00
Bromochloromethane	1.20	ppbv	16.6	30.00
Chloroform	1.17	ppbv	12.1	30.00
Ethyl tert-Butyl Ether	0.800	ppbv	-22.8	30.00

QB-01
LK

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604090

Calibration Check (2604090-CCV2) Continued

Prepared: 04/28/26 Analyzed: 04/29/26

1,2-Dichloroethane	1.11	ppbv	7.3	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.7	30.00
Benzene	1.04	ppbv	0.3	30.00
Carbon Tetrachloride	1.12	ppbv	9.2	30.00
tert-Amyl Methyl Ether	0.777	ppbv	-24.8	30.00
1,2-Dichloropropane	1.15	ppbv	10.5	30.00
Ethyl Acrylate	0.867	ppbv	-17.1	30.00
Bromodichloromethane	1.09	ppbv	5.6	30.00
Trichloroethylene	1.06	ppbv	2.3	30.00
Methyl Methacrylate	0.831	ppbv	-20.3	30.00
cis-1,3-Dichloropropene	0.959	ppbv	-8.2	30.00
Methyl Isobutyl Ketone	0.909	ppbv	-12.6	30.00
trans-1,3-Dichloropropene	1.07	ppbv	1.6	30.00
1,1,2-Trichloroethane	1.18	ppbv	12.8	30.00
Toluene	1.00	ppbv	-3.8	30.00
Dibromochloromethane	1.12	ppbv	7.9	30.00
1,2-Dibromoethane	1.11	ppbv	6.0	30.00
n-Octane	0.885	ppbv	-14.7	30.00
Tetrachloroethylene	1.14	ppbv	9.1	30.00
Chlorobenzene	1.10	ppbv	4.1	30.00
Ethylbenzene	0.934	ppbv	-10.8	30.00
m,p-Xylene	1.96	ppbv	-6.0	30.00
Bromoform	1.19	ppbv	13.4	30.00
Styrene	1.06	ppbv	0.2	30.00
1,1,2,2-Tetrachloroethane	1.14	ppbv	8.5	30.00
o-Xylene	0.931	ppbv	-11.8	30.00
1,3,5-Trimethylbenzene	0.851	ppbv	-19.1	30.00
1,2,4-Trimethylbenzene	0.848	ppbv	-19.6	30.00
m-Dichlorobenzene	1.05	ppbv	0.5	30.00
p-Dichlorobenzene	1.02	ppbv	-2.2	30.00
o-Dichlorobenzene	1.05	ppbv	0.2	30.00
1,2,4-Trichlorobenzene	1.03	ppbv	-0.9	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.03	30.00

Sequence 2605002

Calibration Check (2605002-CCV1)

Prepared & Analyzed: 05/01/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV1) Continued

Prepared & Analyzed: 05/01/26

Dichlorodifluoromethane	1.07	ppbv	3.9	30.00	
Chloromethane	1.04	ppbv	1.9	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	7.6	30.00	
Vinyl chloride	0.944	ppbv	-8.5	30.00	
1,3-Butadiene	0.755	ppbv	-27.0	30.00	
Bromomethane	1.15	ppbv	11.1	30.00	
Chloroethane	1.16	ppbv	11.7	30.00	
Acetonitrile	1.05	ppbv	1.5	30.00	
Acrolein	1.06	ppbv	4.2	30.00	LK
Trichlorofluoromethane	1.09	ppbv	8.3	30.00	
Acrylonitrile	1.07	ppbv	-3.4	30.00	
1,1-Dichloroethene	1.18	ppbv	13.1	30.00	
Dichloromethane	1.09	ppbv	5.5	30.00	
Carbon Disulfide	1.26	ppbv	15.5	30.00	
Trichlorotrifluoroethane	1.21	ppbv	17.3	30.00	
trans-1,2-Dichloroethylene	1.03	ppbv	-0.7	30.00	
1,1-Dichloroethane	1.11	ppbv	7.2	30.00	
Methyl tert-Butyl Ether	0.802	ppbv	-22.6	30.00	
Chloroprene	0.823	ppbv	-20.7	30.00	
cis-1,2-Dichloroethylene	1.02	ppbv	-1.5	30.00	
Bromochloromethane	1.12	ppbv	9.3	30.00	
Chloroform	1.11	ppbv	6.1	30.00	
Ethyl tert-Butyl Ether	0.766	ppbv	-26.1	30.00	
1,2-Dichloroethane	1.05	ppbv	1.2	30.00	
1,1,1-Trichloroethane	1.03	ppbv	-0.3	30.00	
Benzene	0.998	ppbv	-3.8	30.00	
Carbon Tetrachloride	1.03	ppbv	0.3	30.00	
tert-Amyl Methyl Ether	0.735	ppbv	-28.9	30.00	
1,2-Dichloropropane	1.04	ppbv	0.6	30.00	
Ethyl Acrylate	0.865	ppbv	-17.3	30.00	
Bromodichloromethane	1.08	ppbv	4.2	30.00	
Trichloroethylene	1.03	ppbv	-0.07	30.00	
Methyl Methacrylate	0.832	ppbv	-20.2	30.00	
cis-1,3-Dichloropropene	0.902	ppbv	-13.6	30.00	
Methyl Isobutyl Ketone	0.842	ppbv	-19.0	30.00	
trans-1,3-Dichloropropene	1.03	ppbv	-2.1	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV1) Continued

Prepared & Analyzed: 05/01/26

1,1,2-Trichloroethane	1.14	ppbv	9.1	30.00
Toluene	0.952	ppbv	-8.5	30.00
Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.02	ppbv	-2.4	30.00
n-Octane	0.875	ppbv	-15.7	30.00
Tetrachloroethylene	1.07	ppbv	2.7	30.00
Chlorobenzene	1.12	ppbv	6.5	30.00
Ethylbenzene	0.940	ppbv	-10.1	30.00
m,p-Xylene	2.04	ppbv	-2.4	30.00
Bromoform	1.18	ppbv	12.9	30.00
Styrene	1.06	ppbv	0.4	30.00
1,1,2,2-Tetrachloroethane	1.22	ppbv	15.6	30.00
o-Xylene	0.905	ppbv	-14.3	30.00
1,3,5-Trimethylbenzene	0.801	ppbv	-23.8	30.00
1,2,4-Trimethylbenzene	0.776	ppbv	-26.3	30.00
m-Dichlorobenzene	1.04	ppbv	-0.3	30.00
p-Dichlorobenzene	1.01	ppbv	-3.9	30.00
o-Dichlorobenzene	0.996	ppbv	-4.8	30.00
1,2,4-Trichlorobenzene	0.826	ppbv	-20.5	30.00
Hexachloro-1,3-butadiene	0.982	ppbv	-6.8	30.00

Calibration Check (2605002-CCV2)

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorodifluoromethane	1.07	ppbv	3.8	30.00
Chloromethane	1.05	ppbv	3.1	30.00
Dichlorotetrafluoroethane	1.12	ppbv	7.7	30.00
Vinyl chloride	0.945	ppbv	-8.5	30.00
1,3-Butadiene	0.737	ppbv	-28.7	30.00
Bromomethane	1.15	ppbv	11.0	30.00
Chloroethane	1.16	ppbv	12.1	30.00
Acetonitrile	1.10	ppbv	5.6	30.00
Acrolein	1.05	ppbv	3.4	30.00
Trichlorofluoromethane	1.09	ppbv	7.4	30.00
Acrylonitrile	1.06	ppbv	-4.2	30.00
1,1-Dichloroethene	1.17	ppbv	12.9	30.00
Dichloromethane	1.11	ppbv	7.1	30.00
Carbon Disulfide	1.25	ppbv	14.8	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Trichlorotrifluoroethane	1.20	ppbv	16.1	30.00
trans-1,2-Dichloroethylene	1.03	ppbv	-1.2	30.00
1,1-Dichloroethane	1.12	ppbv	7.8	30.00
Methyl tert-Butyl Ether	0.789	ppbv	-23.8	30.00
Chloroprene	0.800	ppbv	-22.9	30.00
cis-1,2-Dichloroethylene	1.02	ppbv	-2.0	30.00
Bromochloromethane	1.12	ppbv	8.5	30.00
Chloroform	1.11	ppbv	5.9	30.00
Ethyl tert-Butyl Ether	0.753	ppbv	-27.3	30.00
1,2-Dichloroethane	1.04	ppbv	0.5	30.00
1,1,1-Trichloroethane	1.02	ppbv	-1.2	30.00
Benzene	0.991	ppbv	-4.5	30.00
Carbon Tetrachloride	1.01	ppbv	-1.3	30.00
tert-Amyl Methyl Ether	0.721	ppbv	-30.2	30.00
1,2-Dichloropropane	1.06	ppbv	2.3	30.00
Ethyl Acrylate	0.869	ppbv	-16.9	30.00
Bromodichloromethane	1.09	ppbv	5.1	30.00
Trichloroethylene	1.03	ppbv	-0.3	30.00
Methyl Methacrylate	0.825	ppbv	-20.8	30.00
cis-1,3-Dichloropropene	0.901	ppbv	-13.7	30.00
Methyl Isobutyl Ketone	0.857	ppbv	-17.6	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-1.6	30.00
1,1,2-Trichloroethane	1.16	ppbv	10.9	30.00
Toluene	0.954	ppbv	-8.3	30.00
Dibromochloromethane	1.06	ppbv	2.1	30.00
1,2-Dibromoethane	1.02	ppbv	-1.8	30.00
n-Octane	0.879	ppbv	-15.3	30.00
Tetrachloroethylene	1.08	ppbv	3.2	30.00
Chlorobenzene	1.12	ppbv	6.1	30.00
Ethylbenzene	0.924	ppbv	-11.7	30.00
m,p-Xylene	2.01	ppbv	-3.6	30.00
Bromoform	1.18	ppbv	12.9	30.00
Styrene	1.05	ppbv	-0.8	30.00
1,1,2,2-Tetrachloroethane	1.23	ppbv	17.0	30.00
o-Xylene	0.884	ppbv	-16.3	30.00
1,3,5-Trimethylbenzene	0.768	ppbv	-27.0	30.00

LL

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

1,2,4-Trimethylbenzene	0.744	ppbv	-29.4	30.00
m-Dichlorobenzene	1.02	ppbv	-2.0	30.00
p-Dichlorobenzene	0.986	ppbv	-5.8	30.00
o-Dichlorobenzene	0.976	ppbv	-6.7	30.00
1,2,4-Trichlorobenzene	0.805	ppbv	-22.4	30.00
Hexachloro-1,3-butadiene	0.972	ppbv	-7.7	30.00

Sequence 2605034

Calibration Check (2605034-CCV1)

Prepared & Analyzed: 05/01/26

Dichlorodifluoromethane	0.961	ppbv	-6.5	30.00
Chloromethane	0.910	ppbv	-10.8	30.00
Dichlorotetrafluoroethane	0.937	ppbv	-9.5	30.00
Vinyl chloride	0.839	ppbv	-18.7	30.00
1,3-Butadiene	0.724	ppbv	-30.0	30.00
Bromomethane	1.03	ppbv	0.2	30.00
Chloroethane	1.04	ppbv	0.5	30.00
Acetonitrile	1.06	ppbv	2.2	30.00
Acrolein	0.996	ppbv	-2.4	30.00
Trichlorofluoromethane	1.01	ppbv	0.02	30.00
Acrylonitrile	1.09	ppbv	-1.7	30.00
1,1-Dichloroethene	1.04	ppbv	0.2	30.00
Dichloromethane	1.00	ppbv	-3.3	30.00
Carbon Disulfide	1.15	ppbv	5.4	30.00
Trichlorotrifluoroethane	1.08	ppbv	4.4	30.00
trans-1,2-Dichloroethylene	0.980	ppbv	-5.8	30.00
1,1-Dichloroethane	1.15	ppbv	11.4	30.00
Methyl tert-Butyl Ether	0.880	ppbv	-15.0	30.00
Chloroprene	0.879	ppbv	-15.3	30.00
cis-1,2-Dichloroethylene	1.09	ppbv	5.2	30.00
Bromochloromethane	1.14	ppbv	11.3	30.00
Chloroform	1.19	ppbv	13.9	30.00
Ethyl tert-Butyl Ether	0.923	ppbv	-10.9	30.00
1,2-Dichloroethane	0.988	ppbv	-4.5	30.00
1,1,1-Trichloroethane	1.08	ppbv	4.9	30.00
Benzene	0.963	ppbv	-7.2	30.00
Carbon Tetrachloride	1.05	ppbv	2.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV1) Continued

Prepared & Analyzed: 05/01/26

tert-Amyl Methyl Ether	0.747	ppbv	-27.7	30.00	LK
1,2-Dichloropropane	1.05	ppbv	0.7	30.00	
Ethyl Acrylate	0.927	ppbv	-11.4	30.00	
Bromodichloromethane	1.20	ppbv	15.9	30.00	
Trichloroethylene	1.08	ppbv	4.9	30.00	
Methyl Methacrylate	0.922	ppbv	-11.5	30.00	
cis-1,3-Dichloropropene	0.888	ppbv	-14.9	30.00	
Methyl Isobutyl Ketone	0.996	ppbv	-4.2	30.00	
trans-1,3-Dichloropropene	0.983	ppbv	-6.4	30.00	
1,1,2-Trichloroethane	1.16	ppbv	11.0	30.00	
Toluene	0.997	ppbv	-4.2	30.00	
Dibromochloromethane	1.17	ppbv	13.0	30.00	
1,2-Dibromoethane	1.07	ppbv	2.9	30.00	
n-Octane	0.924	ppbv	-11.0	30.00	
Tetrachloroethylene	1.09	ppbv	4.0	30.00	
Chlorobenzene	1.15	ppbv	9.2	30.00	
Ethylbenzene	0.953	ppbv	-8.9	30.00	
m,p-Xylene	2.00	ppbv	-4.4	30.00	
Bromoform	1.15	ppbv	10.0	30.00	
Styrene	1.00	ppbv	-5.0	30.00	
1,1,2,2-Tetrachloroethane	1.21	ppbv	15.1	30.00	
o-Xylene	0.907	ppbv	-14.1	30.00	
1,3,5-Trimethylbenzene	0.826	ppbv	-21.4	30.00	
1,2,4-Trimethylbenzene	0.815	ppbv	-22.7	30.00	
m-Dichlorobenzene	1.07	ppbv	2.3	30.00	
p-Dichlorobenzene	1.03	ppbv	-1.8	30.00	
o-Dichlorobenzene	1.05	ppbv	0.6	30.00	
1,2,4-Trichlorobenzene	1.00	ppbv	-3.2	30.00	
Hexachloro-1,3-butadiene	1.23	ppbv	16.9	30.00	

Calibration Check (2605034-CCV2)

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorodifluoromethane	1.07	ppbv	4.3	30.00
Chloromethane	1.03	ppbv	1.0	30.00
Dichlorotetrafluoroethane	1.06	ppbv	2.6	30.00
Vinyl chloride	0.925	ppbv	-10.4	30.00
1,3-Butadiene	0.745	ppbv	-27.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Bromomethane	1.16	ppbv	12.3	30.00
Chloroethane	1.15	ppbv	11.3	30.00
Acetonitrile	1.25	ppbv	20.5	30.00
Acrolein	1.06	ppbv	3.8	30.00
Trichlorofluoromethane	1.08	ppbv	6.6	30.00
Acrylonitrile	1.24	ppbv	12.8	30.00
1,1-Dichloroethene	1.10	ppbv	6.0	30.00
Dichloromethane	1.16	ppbv	11.8	30.00
Carbon Disulfide	1.25	ppbv	15.3	30.00
Trichlorotrifluoroethane	1.18	ppbv	14.7	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.4	30.00
1,1-Dichloroethane	1.10	ppbv	6.5	30.00
Methyl tert-Butyl Ether	0.918	ppbv	-11.4	30.00
Chloroprene	0.801	ppbv	-22.9	30.00
cis-1,2-Dichloroethylene	0.966	ppbv	-6.8	30.00
Bromochloromethane	1.13	ppbv	9.9	30.00
Chloroform	1.21	ppbv	15.4	30.00
Ethyl tert-Butyl Ether	0.766	ppbv	-26.1	30.00
1,2-Dichloroethane	1.11	ppbv	7.3	30.00
1,1,1-Trichloroethane	1.12	ppbv	8.5	30.00
Benzene	1.03	ppbv	-1.1	30.00
Carbon Tetrachloride	1.14	ppbv	10.9	30.00
tert-Amyl Methyl Ether	0.752	ppbv	-27.3	30.00
1,2-Dichloropropane	1.13	ppbv	8.7	30.00
Ethyl Acrylate	0.931	ppbv	-11.0	30.00
Bromodichloromethane	1.25	ppbv	20.6	30.00
Trichloroethylene	1.10	ppbv	6.3	30.00
Methyl Methacrylate	0.918	ppbv	-11.9	30.00
cis-1,3-Dichloropropene	0.872	ppbv	-16.5	30.00
Methyl Isobutyl Ketone	1.01	ppbv	-3.2	30.00
trans-1,3-Dichloropropene	1.00	ppbv	-4.3	30.00
1,1,2-Trichloroethane	1.21	ppbv	15.6	30.00
Toluene	0.990	ppbv	-4.8	30.00
Dibromochloromethane	1.18	ppbv	14.5	30.00
1,2-Dibromoethane	1.07	ppbv	2.7	30.00
n-Octane	0.930	ppbv	-10.4	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605034

Calibration Check (2605034-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

Tetrachloroethylene	1.10	ppbv	5.0	30.00
Chlorobenzene	1.17	ppbv	11.0	30.00
Ethylbenzene	0.936	ppbv	-10.5	30.00
m,p-Xylene	1.98	ppbv	-5.3	30.00
Bromoform	1.18	ppbv	12.5	30.00
Styrene	0.983	ppbv	-7.1	30.00
1,1,2,2-Tetrachloroethane	1.30	ppbv	23.2	30.00
o-Xylene	0.876	ppbv	-17.0	30.00
1,3,5-Trimethylbenzene	0.778	ppbv	-26.1	30.00
1,2,4-Trimethylbenzene	0.768	ppbv	-27.1	30.00
m-Dichlorobenzene	1.07	ppbv	2.8	30.00
p-Dichlorobenzene	1.02	ppbv	-2.3	30.00
o-Dichlorobenzene	1.04	ppbv	-0.6	30.00
1,2,4-Trichlorobenzene	0.961	ppbv	-7.4	30.00
Hexachloro-1,3-butadiene	1.26	ppbv	19.7	30.00

Sequence 2605051

Calibration Check (2605051-CCV1)

Prepared & Analyzed: 05/15/26

Dichlorodifluoromethane	1.02	ppbv	-0.7	30.00
Chloromethane	0.951	ppbv	-6.7	30.00
Dichlorotetrafluoroethane	0.977	ppbv	-5.7	30.00
Vinyl chloride	0.956	ppbv	-7.3	30.00
1,3-Butadiene	0.914	ppbv	-11.7	30.00
Bromomethane	1.01	ppbv	-2.1	30.00
Chloroethane	0.995	ppbv	-3.9	30.00
Acetonitrile	0.917	ppbv	-11.6	30.00
Acrolein	1.01	ppbv	-0.7	30.00
Trichlorofluoromethane	1.13	ppbv	12.3	30.00
Acrylonitrile	1.06	ppbv	-4.2	30.00
1,1-Dichloroethene	1.27	ppbv	22.3	30.00
Dichloromethane	0.937	ppbv	-9.6	30.00
Carbon Disulfide	1.19	ppbv	9.0	30.00
Trichlorotrifluoroethane	1.21	ppbv	17.2	30.00
trans-1,2-Dichloroethylene	0.963	ppbv	-7.4	30.00
1,1-Dichloroethane	0.967	ppbv	-6.6	30.00
Methyl tert-Butyl Ether	1.05	ppbv	1.2	30.00

QB-01
LK, QB-01

LJ

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV1) Continued

Prepared & Analyzed: 05/15/26

Chloroprene	1.03	ppbv	-0.7	30.00
cis-1,2-Dichloroethylene	0.955	ppbv	-7.9	30.00
Bromochloromethane	0.931	ppbv	-9.4	30.00
Chloroform	1.11	ppbv	5.5	30.00
Ethyl tert-Butyl Ether	1.00	ppbv	-3.4	30.00
1,2-Dichloroethane	1.07	ppbv	3.4	30.00
1,1,1-Trichloroethane	1.12	ppbv	9.1	30.00
Benzene	1.03	ppbv	-0.8	30.00
Carbon Tetrachloride	1.12	ppbv	9.8	30.00
tert-Amyl Methyl Ether	1.01	ppbv	-2.7	30.00
1,2-Dichloropropane	1.04	ppbv	0.03	30.00
Ethyl Acrylate	1.12	ppbv	6.7	30.00
Bromodichloromethane	1.17	ppbv	12.8	30.00
Trichloroethylene	1.16	ppbv	12.2	30.00
Methyl Methacrylate	1.15	ppbv	10.8	30.00
cis-1,3-Dichloropropene	1.09	ppbv	4.8	30.00
Methyl Isobutyl Ketone	1.14	ppbv	9.4	30.00
trans-1,3-Dichloropropene	1.12	ppbv	6.2	30.00
1,1,2-Trichloroethane	1.12	ppbv	6.9	30.00
Toluene	1.20	ppbv	15.0	30.00
Dibromochloromethane	1.09	ppbv	5.7	30.00
1,2-Dibromoethane	1.13	ppbv	8.2	30.00
n-Octane	1.18	ppbv	13.5	30.00
Tetrachloroethylene	1.28	ppbv	22.7	30.00
Chlorobenzene	1.17	ppbv	11.5	30.00
Ethylbenzene	1.09	ppbv	3.9	30.00
m,p-Xylene	2.17	ppbv	4.1	30.00
Bromoform	1.30	ppbv	23.9	30.00
Styrene	1.19	ppbv	12.0	30.00
1,1,2,2-Tetrachloroethane	1.21	ppbv	14.5	30.00
o-Xylene	1.21	ppbv	14.4	30.00
1,3,5-Trimethylbenzene	1.19	ppbv	12.7	30.00
1,2,4-Trimethylbenzene	1.18	ppbv	12.3	30.00
m-Dichlorobenzene	1.09	ppbv	4.2	30.00
p-Dichlorobenzene	1.09	ppbv	4.2	30.00
o-Dichlorobenzene	1.09	ppbv	4.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV1) Continued

Prepared & Analyzed: 05/15/26

1,2,4-Trichlorobenzene	1.09	ppbv	5.5	30.00
Hexachloro-1,3-butadiene	1.06	ppbv	0.9	30.00

Calibration Check (2605051-CCV2)

Prepared: 05/15/26 Analyzed: 05/16/26

Dichlorodifluoromethane	1.13	ppbv	10.3	30.00
Chloromethane	1.05	ppbv	2.5	30.00
Dichlorotetrafluoroethane	1.01	ppbv	-2.2	30.00
Vinyl chloride	0.991	ppbv	-4.0	30.00
1,3-Butadiene	0.860	ppbv	-16.8	30.00
Bromomethane	1.00	ppbv	-2.8	30.00
Chloroethane	1.08	ppbv	4.4	30.00
Acetonitrile	0.958	ppbv	-7.7	30.00
Acrolein	1.15	ppbv	12.6	30.00
Trichlorofluoromethane	1.40	ppbv	39.0	30.00
Acrylonitrile	1.08	ppbv	-2.2	30.00
1,1-Dichloroethene	1.18	ppbv	13.3	30.00
Dichloromethane	0.964	ppbv	-7.0	30.00
Carbon Disulfide	1.02	ppbv	-6.4	30.00
Trichlorotrifluoroethane	0.976	ppbv	-5.2	30.00
trans-1,2-Dichloroethylene	0.916	ppbv	-11.9	30.00
1,1-Dichloroethane	1.02	ppbv	-1.9	30.00
Methyl tert-Butyl Ether	1.02	ppbv	-1.3	30.00
Chloroprene	1.15	ppbv	10.5	30.00
cis-1,2-Dichloroethylene	0.995	ppbv	-4.0	30.00
Bromochloromethane	0.929	ppbv	-9.6	30.00
Chloroform	1.26	ppbv	20.2	30.00
Ethyl tert-Butyl Ether	0.966	ppbv	-6.8	30.00
1,2-Dichloroethane	1.28	ppbv	23.8	30.00
1,1,1-Trichloroethane	1.30	ppbv	25.8	30.00
Benzene	1.08	ppbv	3.7	30.00
Carbon Tetrachloride	1.23	ppbv	20.1	30.00
tert-Amyl Methyl Ether	0.937	ppbv	-9.3	30.00
1,2-Dichloropropane	0.995	ppbv	-4.1	30.00
Ethyl Acrylate	1.08	ppbv	3.1	30.00
Bromodichloromethane	1.14	ppbv	10.5	30.00
Trichloroethylene	1.04	ppbv	0.6	30.00

QB-01
LK, QB-01
LK
LJ

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 06/18/26 10:07
SUBMITTED: 04/06/26 to 05/01/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV2) Continued

Prepared: 05/15/26 Analyzed: 05/16/26

Methyl Methacrylate	1.18	ppbv	13.0	30.00
cis-1,3-Dichloropropene	1.06	ppbv	1.5	30.00
Methyl Isobutyl Ketone	1.15	ppbv	10.6	30.00
trans-1,3-Dichloropropene	1.14	ppbv	8.3	30.00
1,1,2-Trichloroethane	1.05	ppbv	0.7	30.00
Toluene	1.19	ppbv	14.9	30.00
Dibromochloromethane	1.07	ppbv	3.3	30.00
1,2-Dibromoethane	1.05	ppbv	0.8	30.00
n-Octane	1.18	ppbv	13.4	30.00
Tetrachloroethylene	1.23	ppbv	17.6	30.00
Chlorobenzene	1.07	ppbv	1.4	30.00
Ethylbenzene	0.994	ppbv	-5.0	30.00
m,p-Xylene	2.02	ppbv	-3.3	30.00
Bromoform	1.30	ppbv	23.9	30.00
Styrene	1.18	ppbv	11.2	30.00
1,1,2,2-Tetrachloroethane	1.23	ppbv	16.8	30.00
o-Xylene	1.20	ppbv	14.0	30.00
1,3,5-Trimethylbenzene	1.15	ppbv	9.4	30.00
1,2,4-Trimethylbenzene	1.15	ppbv	9.1	30.00
m-Dichlorobenzene	0.986	ppbv	-5.4	30.00
p-Dichlorobenzene	0.986	ppbv	-5.7	30.00
o-Dichlorobenzene	0.985	ppbv	-5.8	30.00
1,2,4-Trichlorobenzene	1.05	ppbv	1.1	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.3	30.00



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 06/18/26 10:07

SUBMITTED: 04/06/26 to 05/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
LJ	Identification of analyte is acceptable; reported value is an estimate.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.