



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 09, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/21/26 09:21.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 **FAX:** -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK T5510618	6052203-01	Air	02/25/26 23:59	05/21/26 09:21
TMOK T5510615	6052203-02	Air	03/03/26 23:59	05/21/26 09:21
TMOK T5511324	6052203-03	Air	03/09/26 23:59	05/21/26 09:21
TMOK T5511322	6052203-04	Air	03/15/26 23:59	05/21/26 09:21
TMOK T5511393	6052203-05	Air	03/21/26 23:59	05/21/26 09:21
TMOK T5511391	6052203-06	Air	03/27/26 23:59	05/21/26 09:21
TMOK T5511390 FB	6052203-07	Air	03/28/26 00:00	05/21/26 09:21
TMOK T5511388	6052203-08	Air	04/02/26 23:59	05/21/26 09:21
TMOK T5513364	6052203-09	Air	04/20/26 23:59	05/21/26 09:21
TMOK T5513362	6052203-10	Air	04/26/26 23:59	05/21/26 09:21



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5510618	Lab ID: 6052203-01	Sampled: 02/25/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 18:30
Comments: OK127T31952		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	4.90		0.0548
Arsenic	7440-38-2	0.941	QB-04	0.0245
Beryllium	7440-41-7	0.0248		0.00374
Cadmium	7440-43-9	0.245		0.00365
Chromium	7440-47-3	7.70		2.74
Cobalt	7440-48-4	0.512		0.0120
Lead	7439-92-1	3.76		0.0782
Manganese	7439-96-5	20.9		0.553
Nickel	7440-02-0	2.50		0.225
Selenium	7782-49-2	0.665	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5510615	Lab ID: 6052203-02	Sampled: 03/03/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 19:39
Comments: OK127T31955		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	1.76		0.0548
Arsenic	7440-38-2	0.623	QB-04	0.0245
Beryllium	7440-41-7	0.00803		0.00374
Cadmium	7440-43-9	0.157		0.00365
Chromium	7440-47-3	7.70		2.74
Cobalt	7440-48-4	0.262		0.0120
Lead	7439-92-1	3.22		0.0782
Manganese	7439-96-5	8.21		0.553
Nickel	7440-02-0	1.21		0.225
Selenium	7782-49-2	0.731	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511324	Lab ID: 6052203-03	Sampled: 03/09/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 19:53
Comments: OK127T31957		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.640		0.0548
Arsenic	7440-38-2	0.369	QB-04	0.0245
Beryllium	7440-41-7	0.00323	U	0.00374
Cadmium	7440-43-9	0.0533		0.00365
Chromium	7440-47-3	5.45		2.74
Cobalt	7440-48-4	0.134		0.0120
Lead	7439-92-1	2.03		0.0782
Manganese	7439-96-5	5.46		0.553
Nickel	7440-02-0	0.751		0.225
Selenium	7782-49-2	0.317	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511322	Lab ID: 6052203-04	Sampled: 03/15/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 20:06
Comments: OK127T31959		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.277		0.0548
Arsenic	7440-38-2	0.482	QB-04	0.0245
Beryllium	7440-41-7	0.0266		0.00374
Cadmium	7440-43-9	0.0601		0.00365
Chromium	7440-47-3	6.22		2.74
Cobalt	7440-48-4	0.218		0.0120
Lead	7439-92-1	2.20		0.0782
Manganese	7439-96-5	19.2		0.553
Nickel	7440-02-0	1.04		0.225
Selenium	7782-49-2	0.349	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511393	Lab ID: 6052203-05	Sampled: 03/21/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 20:34
Comments: OK127T31961		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	1.54		0.0548
Arsenic	7440-38-2	0.806	QB-04	0.0245
Beryllium	7440-41-7	0.0231		0.00374
Cadmium	7440-43-9	0.208		0.00365
Chromium	7440-47-3	8.48		2.74
Cobalt	7440-48-4	0.416		0.0120
Lead	7439-92-1	3.50		0.0782
Manganese	7439-96-5	19.8		0.553
Nickel	7440-02-0	2.70		0.225
Selenium	7782-49-2	0.658	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511391	Lab ID: 6052203-06	Sampled: 03/27/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 20:48
Comments: OK127T31963		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.850		0.0548
Arsenic	7440-38-2	0.345	QB-04	0.0245
Beryllium	7440-41-7	0.0299		0.00374
Cadmium	7440-43-9	0.0687		0.00365
Chromium	7440-47-3	7.18		2.74
Cobalt	7440-48-4	0.285		0.0120
Lead	7439-92-1	1.92		0.0782
Manganese	7439-96-5	17.6		0.553
Nickel	7440-02-0	1.20		0.225
Selenium	7782-49-2	0.400	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511390 FB	Lab ID: 6052203-07	Sampled: 03/28/26 00:00
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 21:01
Comments: OK127T31964 FB		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>	
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>	
Antimony	7440-36-0	0.0100	U	0.0548	
Arsenic	7440-38-2	-0.0131	U, QB-04	0.0245	
Beryllium	7440-41-7	-0.00296	U	0.00374	
Cadmium	7440-43-9	8.83E-4	U	0.00365	
Chromium	7440-47-3	4.77	FB-01	2.74	
Cobalt	7440-48-4	0.00667	U	0.0120	
Lead	7439-92-1	0.0215	U	0.0782	
Manganese	7439-96-5	0.234	U	0.553	
Nickel	7440-02-0	0.204	U	0.225	
Selenium	7782-49-2	-0.00770	U, L, QX	0.0383	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5511388	Lab ID: 6052203-08	Sampled: 04/02/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 21:13
Comments: OK127T31966		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.363		0.0548
Arsenic	7440-38-2	0.206	QB-04	0.0245
Beryllium	7440-41-7	0.00566		0.00374
Cadmium	7440-43-9	0.0328		0.00365
Chromium	7440-47-3	5.06		2.74
Cobalt	7440-48-4	0.138		0.0120
Lead	7439-92-1	0.728		0.0782
Manganese	7439-96-5	5.03		0.553
Nickel	7440-02-0	0.733		0.225
Selenium	7782-49-2	0.326	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5513364	Lab ID: 6052203-09	Sampled: 04/20/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 21:26
Comments: OK127T31972		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.751		0.0548
Arsenic	7440-38-2	0.504	QB-04	0.0245
Beryllium	7440-41-7	0.00549		0.00374
Cadmium	7440-43-9	0.0916		0.00365
Chromium	7440-47-3	5.97		2.74
Cobalt	7440-48-4	0.117		0.0120
Lead	7439-92-1	2.77		0.0782
Manganese	7439-96-5	8.08		0.553
Nickel	7440-02-0	0.745		0.225
Selenium	7782-49-2	0.716	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK T5513362	Lab ID: 6052203-10	Sampled: 04/26/26 23:59
Matrix: Air	Sample Volume: 24 m ³	Received: 05/21/26 09:21
	Filter ID:	Analysis Date: 07/01/26 21:39
Comments: OK127T31974		

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.577		0.0548
Arsenic	7440-38-2	0.302	QB-04	0.0245
Beryllium	7440-41-7	0.00102	U	0.00374
Cadmium	7440-43-9	0.0592		0.00365
Chromium	7440-47-3	5.96		2.74
Cobalt	7440-48-4	0.0675		0.0120
Lead	7439-92-1	0.693		0.0782
Manganese	7439-96-5	3.31		0.553
Nickel	7440-02-0	0.661		0.225
Selenium	7782-49-2	0.414	LJ, QX	0.0383



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Inorganics by Compendium Method IO-3.5 - Quality Control										
<i>Batch 2607002 - B6F1101</i>										
Calibration Blank (2607002-CCB1)				Prepared & Analyzed: 07/01/26						
Antimony	-0.804		ng/l							U
Arsenic	-1.11		ng/l							U
Beryllium	-0.00892		ng/l							U
Cadmium	0.405		ng/l							
Chromium	1.74		ng/l							
Cobalt	-0.213		ng/l							U
Lead	10.5		ng/l							
Manganese	1.13		ng/l							
Nickel	-3.53		ng/l							U
Selenium	-0.897		ng/l							LJ, QX, U
Calibration Blank (2607002-CCB2)				Prepared & Analyzed: 07/01/26						
Antimony	-0.735		ng/l							U
Arsenic	-10.8		ng/l							BR, QB-04, U
Beryllium	-0.791		ng/l							U
Cadmium	-0.111		ng/l							U
Chromium	-0.188		ng/l							U
Cobalt	-0.0669		ng/l							U
Lead	6.90		ng/l							
Manganese	0.775		ng/l							
Nickel	-0.495		ng/l							U
Selenium	-7.80		ng/l							LJ, QX, U
Calibration Blank (2607002-CCB3)				Prepared & Analyzed: 07/01/26						
Antimony	-0.783		ng/l							U
Arsenic	-3.34		ng/l							U
Beryllium	-1.00		ng/l							U
Cadmium	-0.0193		ng/l							U
Chromium	1.34		ng/l							
Cobalt	-0.160		ng/l							U
Lead	9.25		ng/l							
Manganese	0.911		ng/l							
Nickel	1.16		ng/l							
Selenium	-6.53		ng/l							LJ, QX, U
Calibration Blank (2607002-CCB4)				Prepared: 07/01/26 Analyzed: 07/02/26						
Antimony	-1.06		ng/l							U
Arsenic	-7.02		ng/l							U
Beryllium	-1.27		ng/l							U
Cadmium	-0.0423		ng/l							U
Chromium	1.18		ng/l							

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2607002 - B6F1101

Calibration Blank (2607002-CCB4) Contin

Prepared: 07/01/26 Analyzed: 07/02/26

Cobalt	-0.186		ng/l							U
Lead	6.85		ng/l							
Manganese	0.105		ng/l							
Nickel	3.45		ng/l							
Selenium	-2.30		ng/l							LJ, QX, U

Calibration Blank (2607002-CCB5)

Prepared: 07/01/26 Analyzed: 07/02/26

Antimony	-1.15		ng/l							U
Arsenic	-4.34		ng/l							U
Beryllium	-1.24		ng/l							U
Cadmium	0.187		ng/l							
Chromium	0.194		ng/l							
Cobalt	-0.200		ng/l							U
Lead	8.54		ng/l							
Manganese	2.27		ng/l							
Nickel	-0.634		ng/l							U
Selenium	-5.52		ng/l							LJ, QX, U

Calibration Check (2607002-CCV1)

Prepared & Analyzed: 07/01/26

Antimony	10100		ng/l	9998.0		101	90-110			
Arsenic	10100		ng/l	10000		101	90-110			
Beryllium	4960		ng/l	5001.5		99.2	90-110			
Cadmium	51700		ng/l	50035		103	90-110			
Chromium	124000		ng/l	120020		104	90-110			
Cobalt	5090		ng/l	5002.0		102	90-110			
Lead	203000		ng/l	200020		101	90-110			
Manganese	104000		ng/l	100040		104	90-110			
Nickel	25500		ng/l	24998		102	90-110			
Selenium	10200		ng/l	10001		102	90-110			LJ, QX

Calibration Check (2607002-CCV2)

Prepared & Analyzed: 07/01/26

Antimony	10000		ng/l	9998.0		100	90-110			
Arsenic	10100		ng/l	10000		101	90-110			
Beryllium	5060		ng/l	5001.5		101	90-110			
Cadmium	51800		ng/l	50035		103	90-110			
Chromium	126000		ng/l	120020		105	90-110			
Cobalt	5180		ng/l	5002.0		104	90-110			
Lead	203000		ng/l	200020		102	90-110			
Manganese	104000		ng/l	100040		104	90-110			
Nickel	25900		ng/l	24998		104	90-110			
Selenium	10100		ng/l	10001		101	90-110			LJ, QX

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2607002 - B6F1101

Calibration Check (2607002-CCV3)

Prepared & Analyzed: 07/01/26

Antimony	10100		ng/l	9998.0		101	90-110			
Arsenic	10200		ng/l	10000		102	90-110			
Beryllium	5120		ng/l	5001.5		102	90-110			
Cadmium	52400		ng/l	50035		105	90-110			
Chromium	127000		ng/l	120020		106	90-110			
Cobalt	5210		ng/l	5002.0		104	90-110			
Lead	204000		ng/l	200020		102	90-110			
Manganese	106000		ng/l	100040		106	90-110			
Nickel	26100		ng/l	24998		104	90-110			
Selenium	10000		ng/l	10001		100	90-110			LJ, QX

Calibration Check (2607002-CCV4)

Prepared: 07/01/26 Analyzed: 07/02/26

Antimony	9730		ng/l	9998.0		97.4	90-110			
Arsenic	9970		ng/l	10000		99.7	90-110			
Beryllium	5080		ng/l	5001.5		102	90-110			
Cadmium	50700		ng/l	50035		101	90-110			
Chromium	122000		ng/l	120020		102	90-110			
Cobalt	5060		ng/l	5002.0		101	90-110			
Lead	199000		ng/l	200020		99.6	90-110			
Manganese	102000		ng/l	100040		102	90-110			
Nickel	25400		ng/l	24998		102	90-110			
Selenium	10000		ng/l	10001		100	90-110			LJ, QX

Calibration Check (2607002-CCV5)

Prepared: 07/01/26 Analyzed: 07/02/26

Antimony	10200		ng/l	9998.0		102	90-110			
Arsenic	10300		ng/l	10000		103	90-110			
Beryllium	5010		ng/l	5001.5		100	90-110			
Cadmium	52900		ng/l	50035		106	90-110			
Chromium	128000		ng/l	120020		106	90-110			
Cobalt	5260		ng/l	5002.0		105	90-110			
Lead	206000		ng/l	200020		103	90-110			
Manganese	106000		ng/l	100040		106	90-110			
Nickel	26300		ng/l	24998		105	90-110			
Selenium	10400		ng/l	10001		104	90-110			LJ, QX

High Cal Check (2607002-HCV1)

Prepared & Analyzed: 07/01/26

Antimony	19400		ng/l	19996		97.1	95-105			
Arsenic	19500		ng/l	20000		97.6	95-105			
Beryllium	9900		ng/l	10003		99.0	95-105			
Cadmium	98300		ng/l	100070		98.2	95-105			
Chromium	237000		ng/l	240050		98.6	95-105			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2607002 - B6F1101

High Cal Check (2607002-HCV1) Continue

Prepared & Analyzed: 07/01/26

Cobalt	9720		ng/l	10004		97.2	95-105			
Lead	392000		ng/l	400040		98.1	95-105			
Manganese	197000		ng/l	200080		98.4	95-105			
Nickel	48600		ng/l	49995		97.2	95-105			
Selenium	19700		ng/l	20002		98.4	95-105			LJ, QX

Initial Cal Blank (2607002-ICB1)

Prepared & Analyzed: 07/01/26

Antimony	-1.05		ng/l							U
Arsenic	2.36		ng/l							
Beryllium	-0.116		ng/l							U
Cadmium	0.265		ng/l							
Chromium	2.68		ng/l							
Cobalt	-0.194		ng/l							U
Lead	11.6		ng/l							
Manganese	1.75		ng/l							
Nickel	-2.08		ng/l							U
Selenium	-3.33		ng/l							LJ, QX, U

Initial Cal Check (2607002-ICV1)

Prepared & Analyzed: 07/01/26

Antimony	9750		ng/l	10000		97.5	90-110			
Arsenic	9760		ng/l	10000		97.6	90-110			
Beryllium	4750		ng/l	5000.0		95.0	90-110			
Cadmium	49200		ng/l	50000		98.3	90-110			
Chromium	116000		ng/l	120000		96.4	90-110			
Cobalt	4960		ng/l	5000.0		99.1	90-110			
Lead	195000		ng/l	200000		97.6	90-110			
Manganese	99800		ng/l	100000		99.8	90-110			
Nickel	25000		ng/l	25000		99.9	90-110			
Selenium	10000		ng/l	10000		100	90-110			LJ, QX

Interference Check A (2607002-IFA1)

Prepared & Analyzed: 07/01/26

Antimony	4.85		ng/l				0-200			
Arsenic	2.53		ng/l				0-200			
Beryllium	1.22		ng/l				0-200			
Cadmium	20.2		ng/l				0-200			
Chromium	24.7		ng/l				0-200			
Cobalt	25.1		ng/l				0-200			
Lead	51.2		ng/l				0-200			
Manganese	38.9		ng/l				0-200			
Nickel	17.7		ng/l				0-200			
Selenium	-0.897		ng/l				0-200			LJ, QX, U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2607002 - B6F1101

Interference Check B (2607002-IFB1)

Prepared & Analyzed: 07/01/26

Antimony	9770		ng/l	9998.0		97.7	80-120			
Arsenic	9670		ng/l	10000		96.7	80-120			
Beryllium	5350		ng/l	5001.5		107	80-120			
Cadmium	49500		ng/l	50035		98.9	80-120			
Chromium	118000		ng/l	120020		98.3	80-120			
Cobalt	4840		ng/l	5002.0		96.8	80-120			
Lead	199000		ng/l	200020		99.6	80-120			
Manganese	97900		ng/l	100040		97.9	80-120			
Nickel	23900		ng/l	24998		95.8	80-120			
Selenium	9760		ng/l	10001		97.6	80-120			LJ, QX

Batch B6F1201 - ICP-MS Extraction

Blank (B6F1201-BLK1)

Prepared: 06/12/26 Analyzed: 07/01/26

Antimony	0.00621	0.0547	ng/m ³ Air							U
Arsenic	-0.00772	0.0245	ng/m ³ Air							U, QB-04
Beryllium	-0.00110	0.00373	ng/m ³ Air							U
Cadmium	-1.79E-4	0.00364	ng/m ³ Air							U
Chromium	0.0861	2.74	ng/m ³ Air							U
Cobalt	-7.17E-4	0.0120	ng/m ³ Air							U
Lead	0.0109	0.0781	ng/m ³ Air							U
Manganese	0.0239	0.552	ng/m ³ Air							U
Nickel	-0.0266	0.225	ng/m ³ Air							U
Selenium	0.00374	0.0382	ng/m ³ Air							LJ, QX, U

LCS (B6F1201-BS1)

Prepared: 06/12/26 Analyzed: 07/01/26

Antimony	10.4	0.0547	ng/m ³ Air	10.397		100	80-120			
Arsenic	20.9	0.0245	ng/m ³ Air	20.799		101	80-120			QB-04
Beryllium	10.5	0.00373	ng/m ³ Air	10.402		101	80-120			
Cadmium	10.6	0.00364	ng/m ³ Air	10.407		102	80-120			
Chromium	106	2.74	ng/m ³ Air	104.01		102	80-120			
Cobalt	10.5	0.0120	ng/m ³ Air	10.403		101	80-120			
Lead	104	0.0781	ng/m ³ Air	104.00		99.6	80-120			
Manganese	64.6	0.552	ng/m ³ Air	62.421		104	80-120			
Nickel	21.3	0.225	ng/m ³ Air	20.797		102	80-120			
Selenium	21.0	0.0382	ng/m ³ Air	20.801		101	80-120			LJ, QX

LCS (B6F1201-BS2)

Prepared: 06/12/26 Analyzed: 07/01/26

Antimony	10.2	0.0547	ng/m ³ Air	10.397		98.1	80-120			
Arsenic	20.4	0.0245	ng/m ³ Air	20.799		98.2	80-120			QB-04
Beryllium	10.5	0.00373	ng/m ³ Air	10.402		101	80-120			
Cadmium	10.3	0.00364	ng/m ³ Air	10.407		99.3	80-120			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch B6F1201 - ICP-MS Extraction

LCS (B6F1201-BS2) Continued

Prepared: 06/12/26 Analyzed: 07/01/26

Chromium	106	2.74	ng/m ³ Air	104.01		102	80-120			
Cobalt	10.3	0.0120	ng/m ³ Air	10.403		98.7	80-120			
Lead	101	0.0781	ng/m ³ Air	104.00		97.3	80-120			
Manganese	64.7	0.552	ng/m ³ Air	62.421		104	80-120			
Nickel	20.9	0.225	ng/m ³ Air	20.797		101	80-120			
Selenium	20.6	0.0382	ng/m ³ Air	20.801		98.8	80-120			LJ, QX

LCS Dup (B6F1201-BSD1)

Prepared: 06/12/26 Analyzed: 07/01/26

Antimony	10.5	0.0547	ng/m ³ Air	10.397		101	80-120	0.892	20	
Arsenic	20.8	0.0245	ng/m ³ Air	20.799		100	80-120	0.528	20	QB-04
Beryllium	10.4	0.00373	ng/m ³ Air	10.402		100	80-120	0.437	20	
Cadmium	10.7	0.00364	ng/m ³ Air	10.407		103	80-120	0.742	20	
Chromium	106	2.74	ng/m ³ Air	104.01		102	80-120	0.332	20	
Cobalt	10.5	0.0120	ng/m ³ Air	10.403		101	80-120	0.0920	20	
Lead	104	0.0781	ng/m ³ Air	104.00		100	80-120	0.548	20	
Manganese	64.7	0.552	ng/m ³ Air	62.421		104	80-120	0.154	20	
Nickel	21.2	0.225	ng/m ³ Air	20.797		102	80-120	0.0546	20	
Selenium	21.0	0.0382	ng/m ³ Air	20.801		101	80-120	0.325	20	LJ, QX

Duplicate (B6F1201-DUP3)

Source: 6052203-04

Prepared: 06/12/26 Analyzed: 07/01/26

Antimony	0.267	0.0548	ng/m ³ Air	0.277				3.69	10	
Arsenic	0.454	0.0245	ng/m ³ Air	0.482				6.11	10	QB-04
Beryllium	0.0289	0.00374	ng/m ³ Air	0.0266				8.56	10	
Cadmium	0.0592	0.00365	ng/m ³ Air	0.0601				1.57	10	
Chromium	6.09	2.74	ng/m ³ Air	6.22				2.10	10	
Cobalt	0.212	0.0120	ng/m ³ Air	0.218				2.96	10	
Lead	2.15	0.0782	ng/m ³ Air	2.20				2.42	10	
Manganese	19.0	0.553	ng/m ³ Air	19.2				1.16	10	
Nickel	1.04	0.225	ng/m ³ Air	1.04				0.348	10	
Selenium	0.379	0.0383	ng/m ³ Air	0.349				8.34	10	LJ, QX



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 **FAX:** -

FILE #: 0475

REPORTED: 07/09/26 14:30

SUBMITTED: 05/21/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
QX	Compound does not meet QC criteria. Results should be considered an estimate.
QB-04	Analyte exceeds continuing calibration blank criteria
LJ	Identification of analyte is acceptable; reported value is an estimate.
FB-01	Analyte exceeds Field Blank criteria.
BR	Sample Value Below Acceptable Range
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard.