



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

August 17, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 06/11/26 11:56 through 07/07/26 09:45.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6061104-03	Air	06/01/26 23:59	06/11/26 11:56
TMOK	6061104-04	Air	06/07/26 23:59	06/11/26 11:56
TMOK	6061711-03	Air	06/13/26 23:59	06/17/26 13:43
TMOK	6062604-02	Air	06/19/26 23:59	06/26/26 10:37
TMOK	6070729-03	Air	06/25/26 23:59	07/07/26 09:45
TMOK FB	6070729-04	Air	06/25/26 23:59	07/07/26 09:45



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Temp Rec'd (°C): -0.9

Lab ID: 6061104-03

Sample Volume: 326.3 m³

Container #: 299

Sampled: 06/01/26 23:59

Received: 06/11/26 11:56

Analyzed: 07/15/26 06:40

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	18.0		1.21		
Acenaphthylene	208-96-8	ND	U	0.00397		
Acenaphthene	83-32-9	3.93		0.128		
Fluorene	86-73-7	4.98		0.0473		
Phenanthrene	85-01-8	9.37	LB	0.129		
Anthracene	120-12-7	0.0561	LB	0.0372		
Fluoranthene	206-44-0	2.54	LB	0.0517		
Pyrene	129-00-0	0.895	LB	0.0567		
Benzo(a)anthracene	56-55-3	0.0211		0.0138		
Chrysene	218-01-9	0.0708		0.0129		
Benzo(b)fluoranthene	205-99-2	0.0368		0.0196		
Benzo(k)fluoranthene	207-08-9	0.0126		0.00827		
Benzo(e)pyrene	192-97-2	0.0276		0.00736		
Benzo(a)pyrene	50-32-8	0.0153		0.00845		
Perylene	198-55-0	ND	U	0.00406		
Dibenz(a,h)anthracene	53-70-3	ND	U	0.00637		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0175		0.00829		
Benzo(g,h,i)perylene	191-24-2	ND	U	0.0294		
Coronene	191-07-1	ND	U	0.0126		
Surrogate: Fluorene-d10		0.820			83.6 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.49			81.2 %	Limit 60-120
Surrogate: Pyrene-d10		0.699			70.0 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.66			85.8 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6061104-04

Sampled: 06/07/26 23:59

Temp Rec'd (°C): -0.9

Sample Volume: 313.267 m³

Received: 06/11/26 11:56

Container #: 280

Analyzed: 07/17/26 23:50

Comments: Arrived without CoC, container 280 sample date 6-7-26, filter present with PUF

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	8.75		1.26		
Acenaphthylene	208-96-8	ND	U	0.00414		
Acenaphthene	83-32-9	1.33		0.133		
Fluorene	86-73-7	1.43		0.0492		
Phenanthrene	85-01-8	3.92	LB, QB-03	0.134		
Anthracene	120-12-7	0.0380	U, LB	0.0388		
Fluoranthene	206-44-0	1.28		0.0538		
Pyrene	129-00-0	0.456		0.0591		
Benzo(a)anthracene	56-55-3	0.0109	U	0.0144		
Chrysene	218-01-9	0.0345		0.0134		
Benzo(b)fluoranthene	205-99-2	0.0217		0.0204		
Benzo(k)fluoranthene	207-08-9	ND	U	0.00862		
Benzo(e)pyrene	192-97-2	0.0140		0.00766		
Benzo(a)pyrene	50-32-8	0.00766	U	0.00880		
Perylene	198-55-0	ND	U	0.00423		
Dibenz(a,h)anthracene	53-70-3	ND	U	0.00664		
Indeno(1,2,3-c,d)pyrene	193-39-5	ND	U	0.00864		
Benzo(g,h,i)perylene	191-24-2	0.0195	U	0.0306		
Coronene	191-07-1	ND	U	0.0131		
Surrogate: Fluorene-d10		0.819			83.6 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.90			90.8 %	Limit 60-120
Surrogate: Pyrene-d10		0.720			72.1 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.78			86.3 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Temp Rec'd (°C): 1.0

Lab ID: 6061711-03

Sample Volume: 310.26 m³

Container #: 146

Sampled: 06/13/26 23:59

Received: 06/17/26 13:43

Analyzed: 07/22/26 23:00

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	16.4		1.28		
Acenaphthylene	208-96-8	ND	U	0.00418		
Acenaphthene	83-32-9	2.10		0.134		
Fluorene	86-73-7	2.46		0.0497		
Phenanthrene	85-01-8	5.88	LB, QB-03	0.135		
Anthracene	120-12-7	0.0719		0.0392		
Fluoranthene	206-44-0	1.76		0.0543		
Pyrene	129-00-0	0.655	LB	0.0597		
Benzo(a)anthracene	56-55-3	0.0261		0.0145		
Chrysene	218-01-9	0.0609		0.0135		
Benzo(b)fluoranthene	205-99-2	0.0284		0.0206		
Benzo(k)fluoranthene	207-08-9	ND	U	0.00870		
Benzo(e)pyrene	192-97-2	ND	U	0.00774		
Benzo(a)pyrene	50-32-8	0.00741	U	0.00889		
Perylene	198-55-0	ND	U	0.00427		
Dibenz(a,h)anthracene	53-70-3	ND	U	0.00670		
Indeno(1,2,3-c,d)pyrene	193-39-5	ND	U	0.00872		
Benzo(g,h,i)perylene	191-24-2	ND	U	0.0309		
Coronene	191-07-1	ND	U	0.0132		
Surrogate: Fluorene-d10		0.827			84.4 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.86			87.0 %	Limit 60-120
Surrogate: Pyrene-d10		0.774			77.4 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.70			84.1 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6062604-02

Sampled: 06/19/26 23:59

Temp Rec'd (°C): 9.6

Sample Volume: 322.1 m³

Received: 06/26/26 10:37

Container #: 621

Analyzed: 07/28/26 04:34

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	13.8	TT	1.23		
Acenaphthylene	208-96-8	ND	TT, U	0.00402		
Acenaphthene	83-32-9	1.25	TT	0.129		
Fluorene	86-73-7	1.67	TT	0.0479		
Phenanthrene	85-01-8	3.24	LB, QB-03, TT	0.130		
Anthracene	120-12-7	0.0447	TT	0.0377		
Fluoranthene	206-44-0	0.812	LB, TT	0.0523		
Pyrene	129-00-0	0.375	LB, TT	0.0575		
Benzo(a)anthracene	56-55-3	0.0199	TT	0.0140		
Chrysene	218-01-9	0.0500	TT	0.0130		
Benzo(b)fluoranthene	205-99-2	0.0407	TT	0.0198		
Benzo(k)fluoranthene	207-08-9	ND	TT, U	0.00838		
Benzo(e)pyrene	192-97-2	0.0292	TT	0.00745		
Benzo(a)pyrene	50-32-8	0.0143	TT	0.00856		
Perylene	198-55-0	ND	TT, U	0.00412		
Dibenz(a,h)anthracene	53-70-3	ND	TT, U	0.00645		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0217	TT	0.00840		
Benzo(g,h,i)perylene	191-24-2	0.0385	TT	0.0298		
Coronene	191-07-1	ND	TT, U	0.0128		
Surrogate: Fluorene-d10		0.771	TT		78.7 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.71	TT		85.6 %	Limit 60-120
Surrogate: Pyrene-d10		0.701	TT		70.2 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.41	TT		77.9 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Temp Rec'd (°C): 5.5

Lab ID: 6070729-03

Sample Volume: 325.4 m³

Container #: 299

Sampled: 06/25/26 23:59

Received: 07/07/26 09:45

Analyzed: 08/05/26 08:28

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	27.8	TT	1.22		
Acenaphthylene	208-96-8	0.182	TT	0.00398		
Acenaphthene	83-32-9	2.29	TT	0.128		
Fluorene	86-73-7	2.75	TT	0.0474		
Phenanthrene	85-01-8	5.93	LB, QB-03, TT	0.129		
Anthracene	120-12-7	0.124	TT	0.0373		
Fluoranthene	206-44-0	1.65	LB, TT	0.0518		
Pyrene	129-00-0	0.720	TT	0.0569		
Benzo(a)anthracene	56-55-3	0.0953	TT	0.0138		
Chrysene	218-01-9	0.132	TT	0.0129		
Benzo(b)fluoranthene	205-99-2	0.148	TT	0.0196		
Benzo(k)fluoranthene	207-08-9	0.0578	TT	0.00830		
Benzo(e)pyrene	192-97-2	0.110	TT	0.00738		
Benzo(a)pyrene	50-32-8	0.113	TT	0.00847		
Perylene	198-55-0	ND	TT, U	0.00407		
Dibenz(a,h)anthracene	53-70-3	ND	TT, U	0.00639		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.113	TT	0.00832		
Benzo(g,h,i)perylene	191-24-2	0.136	TT	0.0295		
Coronene	191-07-1	0.0479	TT	0.0126		
Surrogate: Fluorene-d10		0.754	TT		77.0 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.39	TT		76.2 %	Limit 60-120
Surrogate: Pyrene-d10		0.618	TT		61.9 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.44	TT		79.6 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK FB

Lab ID: 6070729-04

Sampled: 06/25/26 23:59

Temp Rec'd (°C): 5.5

Sample Volume: 325.4 m³

Received: 07/07/26 09:45

Container #: 605

Analyzed: 08/05/26 09:23

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	0.426	TT, U	1.22		
Acenaphthylene	208-96-8	ND	TT, U	0.00398		
Acenaphthene	83-32-9	ND	TT, U	0.128		
Fluorene	86-73-7	0.0175	TT, U	0.0474		
Phenanthrene	85-01-8	0.0575	LB, QB-03, TT, U	0.129		
Anthracene	120-12-7	0.00277	TT, U	0.0373		
Fluoranthene	206-44-0	0.0141	LB, TT, U	0.0518		
Pyrene	129-00-0	0.00860	TT, U	0.0569		
Benzo(a)anthracene	56-55-3	0.00123	TT, U	0.0138		
Chrysene	218-01-9	ND	TT, U	0.0129		
Benzo(b)fluoranthene	205-99-2	ND	TT, U	0.0196		
Benzo(k)fluoranthene	207-08-9	ND	TT, U	0.00830		
Benzo(e)pyrene	192-97-2	ND	TT, U	0.00738		
Benzo(a)pyrene	50-32-8	ND	TT, U	0.00847		
Perylene	198-55-0	ND	TT, U	0.00407		
Dibenz(a,h)anthracene	53-70-3	ND	TT, U	0.00639		
Indeno(1,2,3-c,d)pyrene	193-39-5	ND	TT, U	0.00832		
Benzo(g,h,i)perylene	191-24-2	ND	TT, U	0.0295		
Coronene	191-07-1	ND	TT, U	0.0126		
Surrogate: Fluorene-d10		0.799	TT		81.5 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.71	TT		86.4 %	Limit 60-120
Surrogate: Pyrene-d10		0.692	TT		69.2 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.41	TT		78.7 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606055 - B6E1507

Calibration Blank (2606055-CCB1)

Prepared & Analyzed: 06/17/26

Naphthalene	0.00480	ng/ul		ND					
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00440	ng/ul		ND					
Fluorene	0.00370	ng/ul		ND					
Phenanthrene	0.0115	ng/ul		ND					LB
Anthracene	0.00270	ng/ul		ND					LB
Fluoranthene	0.00880	ng/ul		ND					
Pyrene	0.00480	ng/ul		ND					
Benzo(a)anthracene	0.00140	ng/ul		ND					
Chrysene	0.00220	ng/ul		ND					
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2606055-CCV1)

Prepared & Analyzed: 06/17/26

Naphthalene	7.76	ng/ul	7.5065	ND	103	70-130			
Acenaphthylene	0.517	ng/ul	0.50200	ND	103	70-130			
Acenaphthene	2.59	ng/ul	2.5030	ND	103	70-130			
Fluorene	2.62	ng/ul	2.5022	ND	105	70-130			
Phenanthrene	3.87	ng/ul	3.7538	ND	103	70-130			LB
Anthracene	0.503	ng/ul	0.50125	ND	100	70-130			LB
Fluoranthene	2.57	ng/ul	2.5015	ND	103	70-130			
Pyrene	1.26	ng/ul	1.2520	ND	101	70-130			
Benzo(a)anthracene	0.538	ng/ul	0.50200	ND	107	70-130			
Chrysene	0.515	ng/ul	0.50200	ND	103	70-130			
Benzo(b)fluoranthene	0.527	ng/ul	0.50125	ND	105	70-130			
Benzo(k)fluoranthene	0.541	ng/ul	0.50125	ND	108	70-130			
Benzo(e)pyrene	0.520	ng/ul	0.50125	ND	104	70-130			
Benzo(a)pyrene	0.531	ng/ul	0.50225	ND	106	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606055 - B6E1507

Calibration Check (2606055-CCV1) Contin

Prepared & Analyzed: 06/17/26

Perylene	0.517	ng/ul	0.50125	ND	103	70-130
Dibenz(a,h)anthracene	0.512	ng/ul	0.50275	ND	102	70-130
Indeno(1,2,3-c,d)pyrene	0.523	ng/ul	0.50175	ND	104	70-130
Benzo(g,h,i)perylene	0.515	ng/ul	0.50150	ND	103	70-130
Coronene	0.505	ng/ul	0.50175	ND	101	70-130
Surrogate: Fluorene-d10	0.513	ng/ul	0.49000	ND	105	70-130
Surrogate: Fluoranthene-d10	0.524	ng/ul	0.51000	ND	103	70-130
Surrogate: Pyrene-d10	0.500	ng/ul	0.49950	ND	100	70-130
Surrogate: Benzo(a)pyrene-d12	0.526	ng/ul	0.49850	ND	106	70-130

Batch 2607036 - B6F1703

Calibration Check (2607036-CCV1)

Prepared & Analyzed: 07/09/26

Naphthalene	7.21	ng/ul	7.5065	ND	96.1	70-130
Acenaphthylene	0.500	ng/ul	0.50200	ND	99.7	70-130
Acenaphthene	2.42	ng/ul	2.5030	ND	96.5	70-130
Fluorene	2.50	ng/ul	2.5022	ND	100	70-130
Phenanthrene	3.56	ng/ul	3.7538	ND	94.8	70-130
Anthracene	0.479	ng/ul	0.50125	ND	95.6	70-130
Fluoranthene	2.39	ng/ul	2.5015	ND	95.6	70-130
Pyrene	1.10	ng/ul	1.2520	ND	88.1	70-130
Benzo(a)anthracene	0.497	ng/ul	0.50200	ND	99.1	70-130
Chrysene	0.475	ng/ul	0.50200	ND	94.5	70-130
Benzo(b)fluoranthene	0.498	ng/ul	0.50125	ND	99.3	70-130
Benzo(k)fluoranthene	0.509	ng/ul	0.50125	ND	101	70-130
Benzo(e)pyrene	0.470	ng/ul	0.50125	ND	93.8	70-130
Benzo(a)pyrene	0.489	ng/ul	0.50225	ND	97.4	70-130
Perylene	0.472	ng/ul	0.50125	ND	94.2	70-130
Dibenz(a,h)anthracene	0.425	ng/ul	0.50275	ND	84.5	70-130
Indeno(1,2,3-c,d)pyrene	0.447	ng/ul	0.50175	ND	89.1	70-130
Benzo(g,h,i)perylene	0.446	ng/ul	0.50150	ND	89.0	70-130
Coronene	0.410	ng/ul	0.50175	ND	81.8	70-130
Surrogate: Fluorene-d10	0.532	ng/ul	0.49000	ND	109	70-130
Surrogate: Fluoranthene-d10	0.496	ng/ul	0.51000	ND	97.2	70-130
Surrogate: Pyrene-d10	0.444	ng/ul	0.49950	ND	89.0	70-130
Surrogate: Benzo(a)pyrene-d12	0.483	ng/ul	0.49850	ND	96.9	70-130

LB

Calibration Check (2607036-CCV2)

Prepared: 07/09/26 Analyzed: 07/10/26

Naphthalene	7.23	ng/ul	7.5065	ND	96.3	70-130
Acenaphthylene	0.491	ng/ul	0.50200	ND	97.9	70-130
Acenaphthene	2.43	ng/ul	2.5030	ND	97.2	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607036 - B6F1703

Calibration Check (2607036-CCV2) Contin

Prepared: 07/09/26 Analyzed: 07/10/26

Fluorene	2.52	ng/ul	2.5022	ND	101	70-130			
Phenanthrene	3.57	ng/ul	3.7538	ND	95.1	70-130			
Anthracene	0.479	ng/ul	0.50125	ND	95.6	70-130			LB
Fluoranthene	2.39	ng/ul	2.5015	ND	95.4	70-130			
Pyrene	1.07	ng/ul	1.2520	ND	85.8	70-130			
Benzo(a)anthracene	0.493	ng/ul	0.50200	ND	98.2	70-130			
Chrysene	0.473	ng/ul	0.50200	ND	94.2	70-130			
Benzo(b)fluoranthene	0.496	ng/ul	0.50125	ND	99.0	70-130			
Benzo(k)fluoranthene	0.504	ng/ul	0.50125	ND	101	70-130			
Benzo(e)pyrene	0.471	ng/ul	0.50125	ND	93.9	70-130			
Benzo(a)pyrene	0.485	ng/ul	0.50225	ND	96.6	70-130			
Perylene	0.474	ng/ul	0.50125	ND	94.5	70-130			
Dibenz(a,h)anthracene	0.425	ng/ul	0.50275	ND	84.5	70-130			
Indeno(1,2,3-c,d)pyrene	0.443	ng/ul	0.50175	ND	88.4	70-130			
Benzo(g,h,i)perylene	0.449	ng/ul	0.50150	ND	89.5	70-130			
Coronene	0.403	ng/ul	0.50175	ND	80.2	70-130			
Surrogate: Fluorene-d10	0.530	ng/ul	0.49000	ND	108	70-130			
Surrogate: Fluoranthene-d10	0.493	ng/ul	0.51000	ND	96.6	70-130			
Surrogate: Pyrene-d10	0.430	ng/ul	0.49950	ND	86.0	70-130			
Surrogate: Benzo(a)pyrene-d12	0.481	ng/ul	0.49850	ND	96.4	70-130			

Instrument Blank (2607036-IBL1)

Prepared & Analyzed: 07/09/26

Naphthalene	ND	ng/ul		ND					U
Acenaphthylene	ND	ng/ul		ND					U
Acenaphthene	ND	ng/ul		ND					U
Fluorene	ND	ng/ul		ND					U
Phenanthrene	ND	ng/ul		ND					U
Anthracene	ND	ng/ul		ND					LB, U
Fluoranthene	ND	ng/ul		ND					U
Pyrene	ND	ng/ul		ND					U
Benzo(a)anthracene	ND	ng/ul		ND					U
Chrysene	ND	ng/ul		ND					U
Benzo(b)fluoranthene	ND	ng/ul		ND					U
Benzo(k)fluoranthene	ND	ng/ul		ND					U
Benzo(e)pyrene	ND	ng/ul		ND					U
Benzo(a)pyrene	ND	ng/ul		ND					U
Perylene	ND	ng/ul		ND					U
Dibenz(a,h)anthracene	ND	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607036 - B6F1703

Instrument Blank (2607036-IBL1) Contini

Prepared & Analyzed: 07/09/26

Benzo(g,h,i)perylene	ND	ng/ul		ND					U
Coronene	ND	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Batch 2607048 - B6F1105

Calibration Blank (2607048-CCB1)

Prepared & Analyzed: 07/14/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	7.00E-4	ng/ul		ND					U
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	0.00	ng/ul		ND					U
Pyrene	4.00E-4	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Blank (2607048-CCB2)

Prepared: 07/14/26 Analyzed: 07/15/26

Naphthalene	9.00E-4	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	0.00220	ng/ul		ND					U
Anthracene	0.00	ng/ul		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607048 - B6F1105

Calibration Blank (2607048-CCB2) Contin

Prepared: 07/14/26 Analyzed: 07/15/26

Fluoranthene	0.00110	ng/ul		ND					LB
Pyrene	0.00	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2607048-CCV1)

Prepared & Analyzed: 07/14/26

Naphthalene	7.75	ng/ul	7.5065	ND	103	70-130			
Acenaphthylene	0.524	ng/ul	0.50200	ND	104	70-130			
Acenaphthene	2.61	ng/ul	2.5030	ND	104	70-130			
Fluorene	2.60	ng/ul	2.5022	ND	104	70-130			
Phenanthrene	3.96	ng/ul	3.7538	ND	106	70-130			
Anthracene	0.528	ng/ul	0.50125	ND	105	70-130			
Fluoranthene	2.68	ng/ul	2.5015	ND	107	70-130			LB
Pyrene	1.22	ng/ul	1.2520	ND	97.5	70-130			
Benzo(a)anthracene	0.532	ng/ul	0.50200	ND	106	70-130			
Chrysene	0.531	ng/ul	0.50200	ND	106	70-130			
Benzo(b)fluoranthene	0.534	ng/ul	0.50125	ND	107	70-130			
Benzo(k)fluoranthene	0.533	ng/ul	0.50125	ND	106	70-130			
Benzo(e)pyrene	0.528	ng/ul	0.50125	ND	105	70-130			
Benzo(a)pyrene	0.535	ng/ul	0.50225	ND	106	70-130			
Perylene	0.529	ng/ul	0.50125	ND	106	70-130			
Dibenz(a,h)anthracene	0.554	ng/ul	0.50275	ND	110	70-130			
Indeno(1,2,3-c,d)pyrene	0.544	ng/ul	0.50175	ND	109	70-130			
Benzo(g,h,i)perylene	0.532	ng/ul	0.50150	ND	106	70-130			LB
Coronene	0.538	ng/ul	0.50175	ND	107	70-130			
Surrogate: Fluorene-d10	0.505	ng/ul	0.49000	ND	103	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607048 - B6F1105

Calibration Check (2607048-CCV1) Contin

Prepared & Analyzed: 07/14/26

Surrogate: Fluoranthene-d10	0.542	ng/ul	0.51000	ND	106	70-130
Surrogate: Pyrene-d10	0.481	ng/ul	0.49950	ND	96.2	70-130
Surrogate: Benzo(a)pyrene-d12	0.530	ng/ul	0.49850	ND	106	70-130

Calibration Check (2607048-CCV2)

Prepared: 07/14/26 Analyzed: 07/15/26

Naphthalene	7.81	ng/ul	7.5065	ND	104	70-130	
Acenaphthylene	0.521	ng/ul	0.50200	ND	104	70-130	
Acenaphthene	2.63	ng/ul	2.5030	ND	105	70-130	
Fluorene	2.67	ng/ul	2.5022	ND	107	70-130	
Phenanthrene	3.96	ng/ul	3.7538	ND	105	70-130	
Anthracene	0.550	ng/ul	0.50125	ND	110	70-130	
Fluoranthene	2.81	ng/ul	2.5015	ND	112	70-130	LB
Pyrene	1.18	ng/ul	1.2520	ND	94.5	70-130	
Benzo(a)anthracene	0.539	ng/ul	0.50200	ND	107	70-130	
Chrysene	0.530	ng/ul	0.50200	ND	105	70-130	
Benzo(b)fluoranthene	0.531	ng/ul	0.50125	ND	106	70-130	
Benzo(k)fluoranthene	0.535	ng/ul	0.50125	ND	107	70-130	
Benzo(e)pyrene	0.521	ng/ul	0.50125	ND	104	70-130	
Benzo(a)pyrene	0.532	ng/ul	0.50225	ND	106	70-130	
Perylene	0.532	ng/ul	0.50125	ND	106	70-130	
Dibenz(a,h)anthracene	0.522	ng/ul	0.50275	ND	104	70-130	
Indeno(1,2,3-c,d)pyrene	0.521	ng/ul	0.50175	ND	104	70-130	
Benzo(g,h,i)perylene	0.513	ng/ul	0.50150	ND	102	70-130	LB
Coronene	0.504	ng/ul	0.50175	ND	100	70-130	

Surrogate: Fluorene-d10	0.514	ng/ul	0.49000	ND	105	70-130
Surrogate: Fluoranthene-d10	0.566	ng/ul	0.51000	ND	111	70-130
Surrogate: Pyrene-d10	0.469	ng/ul	0.49950	ND	93.9	70-130
Surrogate: Benzo(a)pyrene-d12	0.527	ng/ul	0.49850	ND	106	70-130

Calibration Check (2607048-CCV3)

Prepared: 07/14/26 Analyzed: 07/15/26

Naphthalene	7.76	ng/ul	7.5065	ND	103	70-130	
Acenaphthylene	0.521	ng/ul	0.50200	ND	104	70-130	
Acenaphthene	2.62	ng/ul	2.5030	ND	105	70-130	
Fluorene	2.64	ng/ul	2.5022	ND	105	70-130	
Phenanthrene	3.92	ng/ul	3.7538	ND	104	70-130	
Anthracene	0.543	ng/ul	0.50125	ND	108	70-130	
Fluoranthene	2.79	ng/ul	2.5015	ND	111	70-130	LB
Pyrene	1.18	ng/ul	1.2520	ND	94.3	70-130	
Benzo(a)anthracene	0.534	ng/ul	0.50200	ND	106	70-130	
Chrysene	0.528	ng/ul	0.50200	ND	105	70-130	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607048 - B6F1105

Calibration Check (2607048-CCV3) Contin

Prepared: 07/14/26 Analyzed: 07/15/26

Benzo(b)fluoranthene	0.530	ng/ul	0.50125	ND	106	70-130			
Benzo(k)fluoranthene	0.534	ng/ul	0.50125	ND	107	70-130			
Benzo(e)pyrene	0.523	ng/ul	0.50125	ND	104	70-130			
Benzo(a)pyrene	0.528	ng/ul	0.50225	ND	105	70-130			
Perylene	0.527	ng/ul	0.50125	ND	105	70-130			
Dibenz(a,h)anthracene	0.512	ng/ul	0.50275	ND	102	70-130			
Indeno(1,2,3-c,d)pyrene	0.508	ng/ul	0.50175	ND	101	70-130			
Benzo(g,h,i)perylene	0.504	ng/ul	0.50150	ND	100	70-130			LB
Coronene	0.477	ng/ul	0.50175	ND	95.0	70-130			
Surrogate: Fluorene-d10	0.509	ng/ul	0.49000	ND	104	70-130			
Surrogate: Fluoranthene-d10	0.560	ng/ul	0.51000	ND	110	70-130			
Surrogate: Pyrene-d10	0.469	ng/ul	0.49950	ND	93.9	70-130			
Surrogate: Benzo(a)pyrene-d12	0.525	ng/ul	0.49850	ND	105	70-130			

Instrument Blank (2607048-IBL1)

Prepared & Analyzed: 07/14/26

Naphthalene	ND	ng/ul		ND					U
Acenaphthylene	ND	ng/ul		ND					U
Acenaphthene	ND	ng/ul		ND					U
Fluorene	ND	ng/ul		ND					U
Phenanthrene	ND	ng/ul		ND					U
Anthracene	ND	ng/ul		ND					U
Fluoranthene	ND	ng/ul		ND					LB, U
Pyrene	ND	ng/ul		ND					U
Benzo(a)anthracene	ND	ng/ul		ND					U
Chrysene	ND	ng/ul		ND					U
Benzo(b)fluoranthene	ND	ng/ul		ND					U
Benzo(k)fluoranthene	ND	ng/ul		ND					U
Benzo(e)pyrene	ND	ng/ul		ND					U
Benzo(a)pyrene	ND	ng/ul		ND					U
Perylene	ND	ng/ul		ND					U
Dibenz(a,h)anthracene	ND	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul		ND					U
Benzo(g,h,i)perylene	ND	ng/ul		ND					LB, U
Coronene	ND	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Batch 2607063 - B6F1703

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607063 - B6F1703

Calibration Blank (2607063-CCB1)

Prepared & Analyzed: 07/17/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	0.00	ng/ul		ND					U
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	0.00	ng/ul		ND					U
Pyrene	0.00	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Blank (2607063-CCB2)

Prepared & Analyzed: 07/17/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	4.00E-4	ng/ul		ND					
Phenanthrene	0.00160	ng/ul		ND					LB
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	9.00E-4	ng/ul		ND					
Pyrene	0.00	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607063 - B6F1703

Calibration Blank (2607063-CCB2) Contin

Prepared & Analyzed: 07/17/26

Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2607063-CCV1)

Prepared & Analyzed: 07/17/26

Naphthalene	7.78	ng/ul	7.5065	ND	104	70-130			
Acenaphthylene	0.525	ng/ul	0.50200	ND	105	70-130			
Acenaphthene	2.61	ng/ul	2.5030	ND	104	70-130			
Fluorene	2.61	ng/ul	2.5022	ND	104	70-130			
Phenanthrene	3.92	ng/ul	3.7538	ND	104	70-130			LB
Anthracene	0.528	ng/ul	0.50125	ND	105	70-130			
Fluoranthene	2.72	ng/ul	2.5015	ND	109	70-130			
Pyrene	1.18	ng/ul	1.2520	ND	94.5	70-130			
Benzo(a)anthracene	0.535	ng/ul	0.50200	ND	107	70-130			
Chrysene	0.529	ng/ul	0.50200	ND	105	70-130			
Benzo(b)fluoranthene	0.534	ng/ul	0.50125	ND	107	70-130			
Benzo(k)fluoranthene	0.534	ng/ul	0.50125	ND	107	70-130			
Benzo(e)pyrene	0.521	ng/ul	0.50125	ND	104	70-130			
Benzo(a)pyrene	0.536	ng/ul	0.50225	ND	107	70-130			
Perylene	0.529	ng/ul	0.50125	ND	105	70-130			
Dibenz(a,h)anthracene	0.565	ng/ul	0.50275	ND	112	70-130			
Indeno(1,2,3-c,d)pyrene	0.555	ng/ul	0.50175	ND	111	70-130			
Benzo(g,h,i)perylene	0.540	ng/ul	0.50150	ND	108	70-130			
Coronene	0.555	ng/ul	0.50175	ND	111	70-130			
Surrogate: Fluorene-d10	0.508	ng/ul	0.49000	ND	104	70-130			
Surrogate: Fluoranthene-d10	0.552	ng/ul	0.51000	ND	108	70-130			
Surrogate: Pyrene-d10	0.473	ng/ul	0.49950	ND	94.8	70-130			
Surrogate: Benzo(a)pyrene-d12	0.530	ng/ul	0.49850	ND	106	70-130			

Calibration Check (2607063-CCV2)

Prepared & Analyzed: 07/17/26

Naphthalene	7.76	ng/ul	7.5065	ND	103	70-130			
Acenaphthylene	0.520	ng/ul	0.50200	ND	104	70-130			
Acenaphthene	2.61	ng/ul	2.5030	ND	104	70-130			
Fluorene	2.65	ng/ul	2.5022	ND	106	70-130			
Phenanthrene	3.89	ng/ul	3.7538	ND	104	70-130			LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607063 - B6F1703

Calibration Check (2607063-CCV2) Contin

Prepared & Analyzed: 07/17/26

Anthracene	0.542	ng/ul	0.50125	ND	108	70-130
Fluoranthene	2.77	ng/ul	2.5015	ND	111	70-130
Pyrene	1.16	ng/ul	1.2520	ND	92.9	70-130
Benzo(a)anthracene	0.539	ng/ul	0.50200	ND	107	70-130
Chrysene	0.528	ng/ul	0.50200	ND	105	70-130
Benzo(b)fluoranthene	0.528	ng/ul	0.50125	ND	105	70-130
Benzo(k)fluoranthene	0.537	ng/ul	0.50125	ND	107	70-130
Benzo(e)pyrene	0.512	ng/ul	0.50125	ND	102	70-130
Benzo(a)pyrene	0.536	ng/ul	0.50225	ND	107	70-130
Perylene	0.526	ng/ul	0.50125	ND	105	70-130
Dibenz(a,h)anthracene	0.557	ng/ul	0.50275	ND	111	70-130
Indeno(1,2,3-c,d)pyrene	0.546	ng/ul	0.50175	ND	109	70-130
Benzo(g,h,i)perylene	0.528	ng/ul	0.50150	ND	105	70-130
Coronene	0.539	ng/ul	0.50175	ND	107	70-130

Surrogate: Fluorene-d10	0.515	ng/ul	0.49000	ND	105	70-130
Surrogate: Fluoranthene-d10	0.558	ng/ul	0.51000	ND	109	70-130
Surrogate: Pyrene-d10	0.461	ng/ul	0.49950	ND	92.4	70-130
Surrogate: Benzo(a)pyrene-d12	0.531	ng/ul	0.49850	ND	107	70-130

Calibration Check (2607063-CCV3)

Prepared: 07/17/26 Analyzed: 07/18/26

Naphthalene	7.71	ng/ul	7.5065	ND	103	70-130
Acenaphthylene	0.515	ng/ul	0.50200	ND	103	70-130
Acenaphthene	2.61	ng/ul	2.5030	ND	104	70-130
Fluorene	2.62	ng/ul	2.5022	ND	105	70-130
Phenanthrene	3.90	ng/ul	3.7538	ND	104	70-130
Anthracene	0.537	ng/ul	0.50125	ND	107	70-130
Fluoranthene	2.72	ng/ul	2.5015	ND	109	70-130
Pyrene	1.18	ng/ul	1.2520	ND	94.3	70-130
Benzo(a)anthracene	0.532	ng/ul	0.50200	ND	106	70-130
Chrysene	0.524	ng/ul	0.50200	ND	104	70-130
Benzo(b)fluoranthene	0.529	ng/ul	0.50125	ND	106	70-130
Benzo(k)fluoranthene	0.536	ng/ul	0.50125	ND	107	70-130
Benzo(e)pyrene	0.518	ng/ul	0.50125	ND	103	70-130
Benzo(a)pyrene	0.532	ng/ul	0.50225	ND	106	70-130
Perylene	0.524	ng/ul	0.50125	ND	104	70-130
Dibenz(a,h)anthracene	0.547	ng/ul	0.50275	ND	109	70-130
Indeno(1,2,3-c,d)pyrene	0.539	ng/ul	0.50175	ND	107	70-130
Benzo(g,h,i)perylene	0.523	ng/ul	0.50150	ND	104	70-130
Coronene	0.523	ng/ul	0.50175	ND	104	70-130

LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607063 - B6F1703

Calibration Check (2607063-CCV3) Contin

Prepared: 07/17/26 Analyzed: 07/18/26

Surrogate: Fluorene-d10	0.512	ng/ul	0.49000	ND	105	70-130
Surrogate: Fluoranthene-d10	0.550	ng/ul	0.51000	ND	108	70-130
Surrogate: Pyrene-d10	0.474	ng/ul	0.49950	ND	94.9	70-130
Surrogate: Benzo(a)pyrene-d12	0.531	ng/ul	0.49850	ND	106	70-130

Instrument Blank (2607063-IBL1)

Prepared & Analyzed: 07/17/26

Naphthalene	ND	ng/ul	ND			U
Acenaphthylene	ND	ng/ul	ND			U
Acenaphthene	ND	ng/ul	ND			U
Fluorene	ND	ng/ul	ND			U
Phenanthrene	ND	ng/ul	ND			LB, U
Anthracene	ND	ng/ul	ND			U
Fluoranthene	ND	ng/ul	ND			U
Pyrene	ND	ng/ul	ND			U
Benzo(a)anthracene	ND	ng/ul	ND			U
Chrysene	ND	ng/ul	ND			U
Benzo(b)fluoranthene	ND	ng/ul	ND			U
Benzo(k)fluoranthene	ND	ng/ul	ND			U
Benzo(e)pyrene	ND	ng/ul	ND			U
Benzo(a)pyrene	ND	ng/ul	ND			U
Perylene	ND	ng/ul	ND			U
Dibenz(a,h)anthracene	ND	ng/ul	ND			U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul	ND			U
Benzo(g,h,i)perylene	ND	ng/ul	ND			U
Coronene	ND	ng/ul	ND			U

Surrogate: Fluorene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Fluoranthene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Pyrene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul	ND		60-120	U

Batch 2607075 - B6F2407

Calibration Blank (2607075-CCB1)

Prepared & Analyzed: 07/22/26

Naphthalene	0.00110	ng/ul	ND			
Acenaphthylene	0.00	ng/ul	ND			U
Acenaphthene	0.00	ng/ul	ND			U
Fluorene	6.00E-4	ng/ul	ND			
Phenanthrene	0.00240	ng/ul	ND			
Anthracene	0.00	ng/ul	ND			U
Fluoranthene	0.00120	ng/ul	ND			
Pyrene	0.00	ng/ul	ND			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607075 - B6F2407

Calibration Blank (2607075-CCB1) Contin

Prepared & Analyzed: 07/22/26

Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2607075-CCV1)

Prepared & Analyzed: 07/22/26

Naphthalene	7.65	ng/ul	7.5065	ND	102	70-130			
Acenaphthylene	0.542	ng/ul	0.50200	ND	108	70-130			
Acenaphthene	2.55	ng/ul	2.5030	ND	102	70-130			
Fluorene	2.52	ng/ul	2.5022	ND	101	70-130			
Phenanthrene	3.90	ng/ul	3.7538	ND	104	70-130			
Anthracene	0.511	ng/ul	0.50125	ND	102	70-130			
Fluoranthene	2.64	ng/ul	2.5015	ND	106	70-130			
Pyrene	1.15	ng/ul	1.2520	ND	92.0	70-130			
Benzo(a)anthracene	0.522	ng/ul	0.50200	ND	104	70-130			
Chrysene	0.518	ng/ul	0.50200	ND	103	70-130			
Benzo(b)fluoranthene	0.520	ng/ul	0.50125	ND	104	70-130			
Benzo(k)fluoranthene	0.528	ng/ul	0.50125	ND	105	70-130			
Benzo(e)pyrene	0.507	ng/ul	0.50125	ND	101	70-130			
Benzo(a)pyrene	0.532	ng/ul	0.50225	ND	106	70-130			
Perylene	0.521	ng/ul	0.50125	ND	104	70-130			
Dibenz(a,h)anthracene	0.584	ng/ul	0.50275	ND	116	70-130			
Indeno(1,2,3-c,d)pyrene	0.574	ng/ul	0.50175	ND	114	70-130			
Benzo(g,h,i)perylene	0.557	ng/ul	0.50150	ND	111	70-130			
Coronene	0.599	ng/ul	0.50175	ND	119	70-130			
Surrogate: Fluorene-d10	0.496	ng/ul	0.49000	ND	101	70-130			
Surrogate: Fluoranthene-d10	0.531	ng/ul	0.51000	ND	104	70-130			
Surrogate: Pyrene-d10	0.459	ng/ul	0.49950	ND	91.9	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607075 - B6F2407

Calibration Check (2607075-CCV1) Contin

Prepared & Analyzed: 07/22/26

Surrogate: Benzo(a)pyrene-d12 0.524 ng/ul 0.49850 ND 105 70-130

Calibration Check (2607075-CCV2)

Prepared & Analyzed: 07/22/26

Naphthalene	7.70	ng/ul	7.5065	ND	103	70-130
Acenaphthylene	0.512	ng/ul	0.50200	ND	102	70-130
Acenaphthene	2.58	ng/ul	2.5030	ND	103	70-130
Fluorene	2.60	ng/ul	2.5022	ND	104	70-130
Phenanthrene	3.89	ng/ul	3.7538	ND	104	70-130
Anthracene	0.526	ng/ul	0.50125	ND	105	70-130
Fluoranthene	2.70	ng/ul	2.5015	ND	108	70-130
Pyrene	1.15	ng/ul	1.2520	ND	92.0	70-130
Benzo(a)anthracene	0.527	ng/ul	0.50200	ND	105	70-130
Chrysene	0.518	ng/ul	0.50200	ND	103	70-130
Benzo(b)fluoranthene	0.514	ng/ul	0.50125	ND	102	70-130
Benzo(k)fluoranthene	0.522	ng/ul	0.50125	ND	104	70-130
Benzo(e)pyrene	0.498	ng/ul	0.50125	ND	99.3	70-130
Benzo(a)pyrene	0.526	ng/ul	0.50225	ND	105	70-130
Perylene	0.514	ng/ul	0.50125	ND	103	70-130
Dibenz(a,h)anthracene	0.559	ng/ul	0.50275	ND	111	70-130
Indeno(1,2,3-c,d)pyrene	0.550	ng/ul	0.50175	ND	110	70-130
Benzo(g,h,i)perylene	0.534	ng/ul	0.50150	ND	106	70-130
Coronene	0.566	ng/ul	0.50175	ND	113	70-130

Surrogate: Fluorene-d10 0.502 ng/ul 0.49000 ND 103 70-130

Surrogate: Fluoranthene-d10 0.545 ng/ul 0.51000 ND 107 70-130

Surrogate: Pyrene-d10 0.459 ng/ul 0.49950 ND 91.9 70-130

Surrogate: Benzo(a)pyrene-d12 0.522 ng/ul 0.49850 ND 105 70-130

Calibration Check (2607075-CCV3)

Prepared: 07/22/26 Analyzed: 07/23/26

Naphthalene	7.62	ng/ul	7.5065	ND	101	70-130
Acenaphthylene	0.511	ng/ul	0.50200	ND	102	70-130
Acenaphthene	2.57	ng/ul	2.5030	ND	103	70-130
Fluorene	2.59	ng/ul	2.5022	ND	103	70-130
Phenanthrene	3.84	ng/ul	3.7538	ND	102	70-130
Anthracene	0.528	ng/ul	0.50125	ND	105	70-130
Fluoranthene	2.70	ng/ul	2.5015	ND	108	70-130
Pyrene	1.17	ng/ul	1.2520	ND	93.7	70-130
Benzo(a)anthracene	0.525	ng/ul	0.50200	ND	105	70-130
Chrysene	0.516	ng/ul	0.50200	ND	103	70-130
Benzo(b)fluoranthene	0.517	ng/ul	0.50125	ND	103	70-130
Benzo(k)fluoranthene	0.525	ng/ul	0.50125	ND	105	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607075 - B6F2407

Calibration Check (2607075-CCV3) Contin

Prepared: 07/22/26 Analyzed: 07/23/26

Benzo(e)pyrene	0.502	ng/ul	0.50125	ND	100	70-130
Benzo(a)pyrene	0.525	ng/ul	0.50225	ND	105	70-130
Perylene	0.516	ng/ul	0.50125	ND	103	70-130
Dibenz(a,h)anthracene	0.553	ng/ul	0.50275	ND	110	70-130
Indeno(1,2,3-c,d)pyrene	0.543	ng/ul	0.50175	ND	108	70-130
Benzo(g,h,i)perylene	0.530	ng/ul	0.50150	ND	106	70-130
Coronene	0.541	ng/ul	0.50175	ND	108	70-130
Surrogate: Fluorene-d10	0.505	ng/ul	0.49000	ND	103	70-130
Surrogate: Fluoranthene-d10	0.543	ng/ul	0.51000	ND	106	70-130
Surrogate: Pyrene-d10	0.467	ng/ul	0.49950	ND	93.6	70-130
Surrogate: Benzo(a)pyrene-d12	0.522	ng/ul	0.49850	ND	105	70-130

Instrument Blank (2607075-IBL1)

Prepared & Analyzed: 07/22/26

Naphthalene	ND	ng/ul	ND			U
Acenaphthylene	ND	ng/ul	ND			U
Acenaphthene	ND	ng/ul	ND			U
Fluorene	ND	ng/ul	ND			U
Phenanthrene	ND	ng/ul	ND			U
Anthracene	ND	ng/ul	ND			U
Fluoranthene	ND	ng/ul	ND			U
Pyrene	ND	ng/ul	ND			U
Benzo(a)anthracene	ND	ng/ul	ND			U
Chrysene	ND	ng/ul	ND			U
Benzo(b)fluoranthene	ND	ng/ul	ND			U
Benzo(k)fluoranthene	ND	ng/ul	ND			U
Benzo(e)pyrene	ND	ng/ul	ND			U
Benzo(a)pyrene	ND	ng/ul	ND			U
Perylene	ND	ng/ul	ND			U
Dibenz(a,h)anthracene	ND	ng/ul	ND			U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul	ND			U
Benzo(g,h,i)perylene	ND	ng/ul	ND			U
Coronene	ND	ng/ul	ND			U
Surrogate: Fluorene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Fluoranthene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Pyrene-d10	0.00	ng/ul	ND		60-120	U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul	ND		60-120	U

Batch 2607099 - B6F3003

Calibration Blank (2607099-CCB1)

Prepared & Analyzed: 07/27/26

Naphthalene	6.00E-4	ng/ul	ND
-------------	---------	-------	----

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607099 - B6F3003

Calibration Blank (2607099-CCB1) Contin

Prepared & Analyzed: 07/27/26

Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	0.00130	ng/ul		ND					
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	6.00E-4	ng/ul		ND					
Pyrene	0.00	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2607099-CCV1)

Prepared & Analyzed: 07/27/26

Naphthalene	7.78	ng/ul	7.5065	ND	104	70-130
Acenaphthylene	0.557	ng/ul	0.50200	ND	111	70-130
Acenaphthene	2.57	ng/ul	2.5030	ND	103	70-130
Fluorene	2.53	ng/ul	2.5022	ND	101	70-130
Phenanthrene	3.92	ng/ul	3.7538	ND	104	70-130
Anthracene	0.510	ng/ul	0.50125	ND	102	70-130
Fluoranthene	2.61	ng/ul	2.5015	ND	104	70-130
Pyrene	1.24	ng/ul	1.2520	ND	99.1	70-130
Benzo(a)anthracene	0.527	ng/ul	0.50200	ND	105	70-130
Chrysene	0.523	ng/ul	0.50200	ND	104	70-130
Benzo(b)fluoranthene	0.524	ng/ul	0.50125	ND	105	70-130
Benzo(k)fluoranthene	0.529	ng/ul	0.50125	ND	105	70-130
Benzo(e)pyrene	0.522	ng/ul	0.50125	ND	104	70-130
Benzo(a)pyrene	0.529	ng/ul	0.50225	ND	105	70-130
Perylene	0.526	ng/ul	0.50125	ND	105	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607099 - B6F3003

Calibration Check (2607099-CCV1) Contin

Prepared & Analyzed: 07/27/26

Dibenz(a,h)anthracene	0.579	ng/ul	0.50275	ND	115	70-130
Indeno(1,2,3-c,d)pyrene	0.576	ng/ul	0.50175	ND	115	70-130
Benzo(g,h,i)perylene	0.569	ng/ul	0.50150	ND	113	70-130
Coronene	0.606	ng/ul	0.50175	ND	121	70-130
Surrogate: Fluorene-d10	0.502	ng/ul	0.49000	ND	102	70-130
Surrogate: Fluoranthene-d10	0.526	ng/ul	0.51000	ND	103	70-130
Surrogate: Pyrene-d10	0.495	ng/ul	0.49950	ND	99.1	70-130
Surrogate: Benzo(a)pyrene-d12	0.524	ng/ul	0.49850	ND	105	70-130

Calibration Check (2607099-CCV2)

Prepared & Analyzed: 07/27/26

Naphthalene	7.76	ng/ul	7.5065	ND	103	70-130
Acenaphthylene	0.518	ng/ul	0.50200	ND	103	70-130
Acenaphthene	2.61	ng/ul	2.5030	ND	104	70-130
Fluorene	2.62	ng/ul	2.5022	ND	105	70-130
Phenanthrene	3.89	ng/ul	3.7538	ND	104	70-130
Anthracene	0.522	ng/ul	0.50125	ND	104	70-130
Fluoranthene	2.67	ng/ul	2.5015	ND	107	70-130
Pyrene	1.22	ng/ul	1.2520	ND	97.7	70-130
Benzo(a)anthracene	0.531	ng/ul	0.50200	ND	106	70-130
Chrysene	0.522	ng/ul	0.50200	ND	104	70-130
Benzo(b)fluoranthene	0.516	ng/ul	0.50125	ND	103	70-130
Benzo(k)fluoranthene	0.523	ng/ul	0.50125	ND	104	70-130
Benzo(e)pyrene	0.513	ng/ul	0.50125	ND	102	70-130
Benzo(a)pyrene	0.527	ng/ul	0.50225	ND	105	70-130
Perylene	0.524	ng/ul	0.50125	ND	104	70-130
Dibenz(a,h)anthracene	0.554	ng/ul	0.50275	ND	110	70-130
Indeno(1,2,3-c,d)pyrene	0.549	ng/ul	0.50175	ND	109	70-130
Benzo(g,h,i)perylene	0.549	ng/ul	0.50150	ND	110	70-130
Coronene	0.580	ng/ul	0.50175	ND	116	70-130
Surrogate: Fluorene-d10	0.509	ng/ul	0.49000	ND	104	70-130
Surrogate: Fluoranthene-d10	0.540	ng/ul	0.51000	ND	106	70-130
Surrogate: Pyrene-d10	0.485	ng/ul	0.49950	ND	97.1	70-130
Surrogate: Benzo(a)pyrene-d12	0.523	ng/ul	0.49850	ND	105	70-130

Calibration Check (2607099-CCV3)

Prepared: 07/27/26 Analyzed: 07/28/26

Naphthalene	7.74	ng/ul	7.5065	ND	103	70-130
Acenaphthylene	0.520	ng/ul	0.50200	ND	104	70-130
Acenaphthene	2.62	ng/ul	2.5030	ND	105	70-130
Fluorene	2.63	ng/ul	2.5022	ND	105	70-130
Phenanthrene	3.91	ng/ul	3.7538	ND	104	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607099 - B6F3003

Calibration Check (2607099-CCV3) Contin

Prepared: 07/27/26 Analyzed: 07/28/26

Anthracene	0.534	ng/ul	0.50125	ND	107	70-130
Fluoranthene	2.75	ng/ul	2.5015	ND	110	70-130
Pyrene	1.16	ng/ul	1.2520	ND	93.0	70-130
Benzo(a)anthracene	0.533	ng/ul	0.50200	ND	106	70-130
Chrysene	0.521	ng/ul	0.50200	ND	104	70-130
Benzo(b)fluoranthene	0.519	ng/ul	0.50125	ND	103	70-130
Benzo(k)fluoranthene	0.526	ng/ul	0.50125	ND	105	70-130
Benzo(e)pyrene	0.506	ng/ul	0.50125	ND	101	70-130
Benzo(a)pyrene	0.529	ng/ul	0.50225	ND	105	70-130
Perylene	0.521	ng/ul	0.50125	ND	104	70-130
Dibenz(a,h)anthracene	0.553	ng/ul	0.50275	ND	110	70-130
Indeno(1,2,3-c,d)pyrene	0.547	ng/ul	0.50175	ND	109	70-130
Benzo(g,h,i)perylene	0.539	ng/ul	0.50150	ND	107	70-130
Coronene	0.554	ng/ul	0.50175	ND	110	70-130
Surrogate: Fluorene-d10	0.512	ng/ul	0.49000	ND	104	70-130
Surrogate: Fluoranthene-d10	0.558	ng/ul	0.51000	ND	109	70-130
Surrogate: Pyrene-d10	0.464	ng/ul	0.49950	ND	93.0	70-130
Surrogate: Benzo(a)pyrene-d12	0.529	ng/ul	0.49850	ND	106	70-130

Instrument Blank (2607099-IBL1)

Prepared & Analyzed: 07/27/26

Naphthalene	ND	ng/ul	ND	U
Acenaphthylene	ND	ng/ul	ND	U
Acenaphthene	ND	ng/ul	ND	U
Fluorene	ND	ng/ul	ND	U
Phenanthrene	ND	ng/ul	ND	U
Anthracene	ND	ng/ul	ND	U
Fluoranthene	ND	ng/ul	ND	U
Pyrene	ND	ng/ul	ND	U
Benzo(a)anthracene	ND	ng/ul	ND	U
Chrysene	ND	ng/ul	ND	U
Benzo(b)fluoranthene	ND	ng/ul	ND	U
Benzo(k)fluoranthene	ND	ng/ul	ND	U
Benzo(e)pyrene	ND	ng/ul	ND	U
Benzo(a)pyrene	ND	ng/ul	ND	U
Perylene	ND	ng/ul	ND	U
Dibenz(a,h)anthracene	ND	ng/ul	ND	U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul	ND	U
Benzo(g,h,i)perylene	ND	ng/ul	ND	U
Coronene	ND	ng/ul	ND	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607099 - B6F3003

Instrument Blank (2607099-IBL1) Contini

Prepared & Analyzed: 07/27/26

Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Batch 2608010 - B6G0803

Calibration Blank (2608010-CCB1)

Prepared: 08/04/26 Analyzed: 08/05/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	7.00E-4	ng/ul		ND					LB
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	0.00	ng/ul		ND					U
Pyrene	0.00	ng/ul		ND					U
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2608010-CCV1)

Prepared & Analyzed: 08/04/26

Naphthalene	7.31	ng/ul	7.5065	ND	97.4	70-130			
Acenaphthylene	0.497	ng/ul	0.50200	ND	98.9	70-130			
Acenaphthene	2.42	ng/ul	2.5030	ND	96.8	70-130			
Fluorene	2.43	ng/ul	2.5022	ND	97.0	70-130			
Phenanthrene	3.67	ng/ul	3.7538	ND	97.9	70-130			LB
Anthracene	0.490	ng/ul	0.50125	ND	97.7	70-130			
Fluoranthene	2.47	ng/ul	2.5015	ND	98.9	70-130			LB
Pyrene	1.14	ng/ul	1.2520	ND	90.7	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2608010 - B6G0803

Calibration Check (2608010-CCV1) Contin

Prepared & Analyzed: 08/04/26

Benzo(a)anthracene	0.484	ng/ul	0.50200	ND	96.5	70-130
Chrysene	0.484	ng/ul	0.50200	ND	96.4	70-130
Benzo(b)fluoranthene	0.481	ng/ul	0.50125	ND	95.9	70-130
Benzo(k)fluoranthene	0.490	ng/ul	0.50125	ND	97.8	70-130
Benzo(e)pyrene	0.478	ng/ul	0.50125	ND	95.4	70-130
Benzo(a)pyrene	0.486	ng/ul	0.50225	ND	96.7	70-130
Perylene	0.482	ng/ul	0.50125	ND	96.1	70-130
Dibenz(a,h)anthracene	0.498	ng/ul	0.50275	ND	99.1	70-130
Indeno(1,2,3-c,d)pyrene	0.496	ng/ul	0.50175	ND	98.9	70-130
Benzo(g,h,i)perylene	0.489	ng/ul	0.50150	ND	97.5	70-130
Coronene	0.498	ng/ul	0.50175	ND	99.2	70-130

Surrogate: Fluorene-d10

0.477 ng/ul 0.49000 ND 97.3 70-130

Surrogate: Fluoranthene-d10

0.500 ng/ul 0.51000 ND 98.0 70-130

Surrogate: Pyrene-d10

0.454 ng/ul 0.49950 ND 90.9 70-130

Surrogate: Benzo(a)pyrene-d12

0.485 ng/ul 0.49850 ND 97.3 70-130

Calibration Check (2608010-CCV2)

Prepared: 08/04/26 Analyzed: 08/05/26

Naphthalene	7.35	ng/ul	7.5065	ND	97.9	70-130
Acenaphthylene	0.478	ng/ul	0.50200	ND	95.3	70-130
Acenaphthene	2.46	ng/ul	2.5030	ND	98.5	70-130
Fluorene	2.49	ng/ul	2.5022	ND	99.6	70-130
Phenanthrene	3.66	ng/ul	3.7538	ND	97.4	70-130
Anthracene	0.503	ng/ul	0.50125	ND	100	70-130
Fluoranthene	2.61	ng/ul	2.5015	ND	104	70-130
Pyrene	1.05	ng/ul	1.2520	ND	84.1	70-130
Benzo(a)anthracene	0.488	ng/ul	0.50200	ND	97.3	70-130
Chrysene	0.483	ng/ul	0.50200	ND	96.3	70-130
Benzo(b)fluoranthene	0.476	ng/ul	0.50125	ND	94.9	70-130
Benzo(k)fluoranthene	0.491	ng/ul	0.50125	ND	97.9	70-130
Benzo(e)pyrene	0.466	ng/ul	0.50125	ND	92.9	70-130
Benzo(a)pyrene	0.485	ng/ul	0.50225	ND	96.6	70-130
Perylene	0.478	ng/ul	0.50125	ND	95.3	70-130
Dibenz(a,h)anthracene	0.493	ng/ul	0.50275	ND	98.1	70-130
Indeno(1,2,3-c,d)pyrene	0.488	ng/ul	0.50175	ND	97.4	70-130
Benzo(g,h,i)perylene	0.479	ng/ul	0.50150	ND	95.5	70-130
Coronene	0.472	ng/ul	0.50175	ND	94.2	70-130

Surrogate: Fluorene-d10

0.480 ng/ul 0.49000 ND 97.9 70-130

Surrogate: Fluoranthene-d10

0.529 ng/ul 0.51000 ND 104 70-130

Surrogate: Pyrene-d10

0.422 ng/ul 0.49950 ND 84.6 70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2608010 - B6G0803

Calibration Check (2608010-CCV2) Contin

Prepared: 08/04/26 Analyzed: 08/05/26

Surrogate: Benzo(a)pyrene-d12 0.489 ng/ul

0.49850 ND 98.1 70-130

Calibration Check (2608010-CCV3)

Prepared: 08/04/26 Analyzed: 08/05/26

Naphthalene 7.31 ng/ul

7.5065 ND 97.4 70-130

Acenaphthylene 0.479 ng/ul

0.50200 ND 95.4 70-130

Acenaphthene 2.46 ng/ul

2.5030 ND 98.4 70-130

Fluorene 2.48 ng/ul

2.5022 ND 99.0 70-130

Phenanthrene 3.66 ng/ul

3.7538 ND 97.4 70-130

Anthracene 0.510 ng/ul

0.50125 ND 102 70-130

Fluoranthene 2.63 ng/ul

2.5015 ND 105 70-130

Pyrene 1.04 ng/ul

1.2520 ND 83.0 70-130

Benzo(a)anthracene 0.487 ng/ul

0.50200 ND 97.0 70-130

Chrysene 0.481 ng/ul

0.50200 ND 95.8 70-130

Benzo(b)fluoranthene 0.473 ng/ul

0.50125 ND 94.4 70-130

Benzo(k)fluoranthene 0.491 ng/ul

0.50125 ND 97.9 70-130

Benzo(e)pyrene 0.466 ng/ul

0.50125 ND 93.0 70-130

Benzo(a)pyrene 0.487 ng/ul

0.50225 ND 96.9 70-130

Perylene 0.478 ng/ul

0.50125 ND 95.4 70-130

Dibenz(a,h)anthracene 0.498 ng/ul

0.50275 ND 99.0 70-130

Indeno(1,2,3-c,d)pyrene 0.494 ng/ul

0.50175 ND 98.4 70-130

Benzo(g,h,i)perylene 0.480 ng/ul

0.50150 ND 95.6 70-130

Coronene 0.470 ng/ul

0.50175 ND 93.7 70-130

Surrogate: Fluorene-d10 0.480 ng/ul

0.49000 ND 97.9 70-130

Surrogate: Fluoranthene-d10 0.535 ng/ul

0.51000 ND 105 70-130

Surrogate: Pyrene-d10 0.417 ng/ul

0.49950 ND 83.4 70-130

Surrogate: Benzo(a)pyrene-d12 0.491 ng/ul

0.49850 ND 98.5 70-130

Instrument Blank (2608010-IBL1)

Prepared & Analyzed: 08/04/26

Naphthalene ND ng/ul

ND

U

Acenaphthylene ND ng/ul

ND

U

Acenaphthene ND ng/ul

ND

U

Fluorene ND ng/ul

ND

U

Phenanthrene ND ng/ul

ND

LB, U

Anthracene ND ng/ul

ND

U

Fluoranthene ND ng/ul

ND

LB, U

Pyrene ND ng/ul

ND

U

Benzo(a)anthracene ND ng/ul

ND

U

Chrysene ND ng/ul

ND

U

Benzo(b)fluoranthene ND ng/ul

ND

U

Benzo(k)fluoranthene ND ng/ul

ND

U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2608010 - B6G0803

Instrument Blank (2608010-IBL1) Contini

Prepared & Analyzed: 08/04/26

Benzo(e)pyrene	ND	ng/ul		ND					U
Benzo(a)pyrene	ND	ng/ul		ND					U
Perylene	ND	ng/ul		ND					U
Dibenz(a,h)anthracene	ND	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	ND	ng/ul		ND					U
Benzo(g,h,i)perylene	ND	ng/ul		ND					U
Coronene	ND	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Batch B6F1105 - TO-13 extraction

Blank (B6F1105-BLK1)

Prepared: 06/11/26 Analyzed: 07/09/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							LB, U
Fluoranthene	ND	ng/m ³ Air							LB, U
Pyrene	ND	ng/m ³ Air							U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U
Surrogate: Fluorene-d10	0.909	ng/ul	0.98000		92.7	60-120			
Surrogate: Fluoranthene-d10	2.82	ng/m ³ Air	3.3333		84.7	60-120			
Surrogate: Pyrene-d10	0.748	ng/ul	0.99900		74.9	60-120			
Surrogate: Benzo(a)pyrene-d12	2.81	ng/m ³ Air	3.3667		83.5	60-120			

LCS (B6F1105-BS1)

Prepared: 06/11/26 Analyzed: 07/09/26

Naphthalene	16.7	ng/m ³ Air	25.000	ND	66.9	60-120			
-------------	------	-----------------------	--------	----	------	--------	--	--	--

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F1105 - TO-13 extraction

LCS (B6F1105-BS1) Continued

Prepared: 06/11/26 Analyzed: 07/09/26

Acenaphthylene	1.08	ng/m ³ Air	1.6667	ND	64.6	60-120			
Acenaphthene	5.83	ng/m ³ Air	8.3333	ND	69.9	60-120			
Fluorene	6.15	ng/m ³ Air	8.3333	ND	73.8	60-120			
Phenanthrene	9.00	ng/m ³ Air	12.600	ND	71.4	60-120			LB
Anthracene	1.05	ng/m ³ Air	1.6633	ND	63.3	60-120			LB
Fluoranthene	6.03	ng/m ³ Air	8.4333	ND	71.5	60-120			LB
Pyrene	2.69	ng/m ³ Air	4.1333	ND	65.0	60-120			LB
Benzo(a)anthracene	1.35	ng/m ³ Air	1.6667	ND	81.1	60-120			
Chrysene	1.23	ng/m ³ Air	1.6667	ND	73.6	60-120			
Benzo(b)fluoranthene	1.33	ng/m ³ Air	1.6667	ND	79.6	60-120			
Benzo(k)fluoranthene	1.32	ng/m ³ Air	1.6433	ND	80.6	60-120			
Benzo(e)pyrene	1.22	ng/m ³ Air	1.6467	ND	73.8	60-120			
Benzo(a)pyrene	1.22	ng/m ³ Air	1.6667	ND	73.4	60-120			
Perylene	1.20	ng/m ³ Air	1.6633	ND	72.3	60-120			
Dibenz(a,h)anthracene	1.14	ng/m ³ Air	1.6667	ND	68.4	60-120			
Indeno(1,2,3-c,d)pyrene	1.22	ng/m ³ Air	1.6667	ND	73.4	60-120			
Benzo(g,h,i)perylene	1.17	ng/m ³ Air	1.6667	ND	70.2	60-120			
Coronene	0.225	ng/m ³ Air	0.33333	ND	67.4	60-120			
Surrogate: Fluorene-d10	0.798	ng/ul	0.98000	ND	81.4	60-120			
Surrogate: Fluoranthene-d10	2.47	ng/m ³ Air	3.3333	ND	74.0	60-120			
Surrogate: Pyrene-d10	0.643	ng/ul	0.99900	ND	64.3	60-120			
Surrogate: Benzo(a)pyrene-d12	2.53	ng/m ³ Air	3.3667	ND	75.0	60-120			

LCS Dup (B6F1105-BS1)

Prepared: 06/11/26 Analyzed: 07/09/26

Naphthalene	16.4	ng/m ³ Air	25.000	ND	65.6	60-120	2.09	20	
Acenaphthylene	1.04	ng/m ³ Air	1.6667	ND	62.6	60-120	3.18	20	
Acenaphthene	5.64	ng/m ³ Air	8.3333	ND	67.7	60-120	3.22	20	
Fluorene	5.90	ng/m ³ Air	8.3333	ND	70.8	60-120	4.22	20	
Phenanthrene	8.95	ng/m ³ Air	12.600	ND	71.0	60-120	0.528	20	LB
Anthracene	1.06	ng/m ³ Air	1.6633	ND	63.9	60-120	1.01	20	LB
Fluoranthene	6.03	ng/m ³ Air	8.4333	ND	71.5	60-120	0.0276	20	LB
Pyrene	2.63	ng/m ³ Air	4.1333	ND	63.7	60-120	2.11	20	LB
Benzo(a)anthracene	1.31	ng/m ³ Air	1.6667	ND	78.8	60-120	2.95	20	
Chrysene	1.20	ng/m ³ Air	1.6667	ND	71.9	60-120	2.42	20	
Benzo(b)fluoranthene	1.30	ng/m ³ Air	1.6667	ND	78.1	60-120	1.85	20	
Benzo(k)fluoranthene	1.30	ng/m ³ Air	1.6433	ND	79.3	60-120	1.65	20	
Benzo(e)pyrene	1.18	ng/m ³ Air	1.6467	ND	71.9	60-120	2.56	20	
Benzo(a)pyrene	1.19	ng/m ³ Air	1.6667	ND	71.6	60-120	2.48	20	
Perylene	1.20	ng/m ³ Air	1.6633	ND	72.0	60-120	0.389	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F1105 - TO-13 extraction

LCS Dup (B6F1105-BSD1) Continued

Prepared: 06/11/26 Analyzed: 07/09/26

Dibenz(a,h)anthracene	1.08	ng/m ³ Air	1.6667	ND	64.9	60-120	5.25	20	
Indeno(1,2,3-c,d)pyrene	1.18	ng/m ³ Air	1.6667	ND	70.9	60-120	3.49	20	
Benzo(g,h,i)perylene	1.12	ng/m ³ Air	1.6667	ND	67.1	60-120	4.54	20	
Coronene	0.206	ng/m ³ Air	0.33333	ND	61.8	60-120	8.67	20	
Surrogate: Fluorene-d10	0.760	ng/ul	0.98000	ND	77.5	60-120			
Surrogate: Fluoranthene-d10	2.42	ng/m ³ Air	3.3333	ND	72.7	60-120			
Surrogate: Pyrene-d10	0.622	ng/ul	0.99900	ND	62.3	60-120			
Surrogate: Benzo(a)pyrene-d12	2.45	ng/m ³ Air	3.3667	ND	72.7	60-120			

Calibration Blank (B6F1105-CCB1)

Prepared: 06/11/26 Analyzed: 07/09/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	0.00190	ng/m ³ Air		ND					LB
Anthracene	0.00340	ng/m ³ Air		ND					LB
Fluoranthene	0.00170	ng/m ³ Air		ND					LB
Pyrene	6.00E-4	ng/m ³ Air		ND					LB
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND		60-120			U

Batch B6F1703 - TO-13 extraction

Blank (B6F1703-BLK1)

Prepared: 06/17/26 Analyzed: 07/09/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F1703 - TO-13 extraction

Blank (B6F1703-BLK1) Continued

Prepared: 06/17/26 Analyzed: 07/09/26

Phenanthrene	ND	ng/m ³ Air							U
Anthracene	ND	ng/m ³ Air							LB, U
Fluoranthene	ND	ng/m ³ Air							U
Pyrene	ND	ng/m ³ Air							U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U

Surrogate: Fluorene-d10 0.851 ng/ul 0.98000 86.8 60-120

Surrogate: Fluoranthene-d10 2.62 ng/m³ Air 3.3333 78.5 60-120

Surrogate: Pyrene-d10 0.698 ng/ul 0.99900 69.9 60-120

Surrogate: Benzo(a)pyrene-d12 2.62 ng/m³ Air 3.3667 77.9 60-120

LCS (B6F1703-BS1)

Prepared: 06/17/26 Analyzed: 07/09/26

Naphthalene	18.0	ng/m ³ Air	25.000	ND	72.0	60-120	
Acenaphthylene	1.14	ng/m ³ Air	1.6667	ND	68.2	60-120	
Acenaphthene	6.21	ng/m ³ Air	8.3333	ND	74.6	60-120	
Fluorene	6.50	ng/m ³ Air	8.3333	ND	78.0	60-120	
Phenanthrene	9.55	ng/m ³ Air	12.600	ND	75.8	60-120	
Anthracene	1.14	ng/m ³ Air	1.6633	ND	68.6	60-120	LB
Fluoranthene	6.39	ng/m ³ Air	8.4333	ND	75.8	60-120	
Pyrene	2.78	ng/m ³ Air	4.1333	ND	67.3	60-120	
Benzo(a)anthracene	1.41	ng/m ³ Air	1.6667	ND	84.3	60-120	
Chrysene	1.28	ng/m ³ Air	1.6667	ND	76.9	60-120	
Benzo(b)fluoranthene	1.40	ng/m ³ Air	1.6667	ND	84.3	60-120	
Benzo(k)fluoranthene	1.41	ng/m ³ Air	1.6433	ND	85.9	60-120	
Benzo(e)pyrene	1.28	ng/m ³ Air	1.6467	ND	77.9	60-120	
Benzo(a)pyrene	1.28	ng/m ³ Air	1.6667	ND	77.1	60-120	
Perylene	1.26	ng/m ³ Air	1.6633	ND	75.6	60-120	
Dibenz(a,h)anthracene	1.17	ng/m ³ Air	1.6667	ND	70.3	60-120	
Indeno(1,2,3-c,d)pyrene	1.27	ng/m ³ Air	1.6667	ND	76.1	60-120	
Benzo(g,h,i)perylene	1.21	ng/m ³ Air	1.6667	ND	72.8	60-120	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F1703 - TO-13 extraction

LCS (B6F1703-BS1) Continued

Prepared: 06/17/26 Analyzed: 07/09/26

Coronene	0.217	ng/m ³ Air	0.33333	ND	65.0	60-120			
Surrogate: Fluorene-d10	0.827	ng/ul	0.98000	ND	84.4	60-120			
Surrogate: Fluoranthene-d10	2.61	ng/m ³ Air	3.3333	ND	78.3	60-120			
Surrogate: Pyrene-d10	0.653	ng/ul	0.99900	ND	65.4	60-120			
Surrogate: Benzo(a)pyrene-d12	2.67	ng/m ³ Air	3.3667	ND	79.4	60-120			

LCS Dup (B6F1703-BS1)

Prepared: 06/17/26 Analyzed: 07/09/26

Naphthalene	16.8	ng/m ³ Air	25.000	ND	67.1	60-120	7.00	20	
Acenaphthylene	1.09	ng/m ³ Air	1.6667	ND	65.3	60-120	4.43	20	
Acenaphthene	5.93	ng/m ³ Air	8.3333	ND	71.1	60-120	4.70	20	
Fluorene	6.18	ng/m ³ Air	8.3333	ND	74.2	60-120	4.96	20	
Phenanthrene	9.14	ng/m ³ Air	12.600	ND	72.5	60-120	4.45	20	
Anthracene	1.08	ng/m ³ Air	1.6633	ND	64.9	60-120	5.49	20	LB
Fluoranthene	6.13	ng/m ³ Air	8.4333	ND	72.7	60-120	4.24	20	
Pyrene	2.70	ng/m ³ Air	4.1333	ND	65.4	60-120	2.72	20	
Benzo(a)anthracene	1.34	ng/m ³ Air	1.6667	ND	80.5	60-120	4.59	20	
Chrysene	1.22	ng/m ³ Air	1.6667	ND	73.2	60-120	4.93	20	
Benzo(b)fluoranthene	1.30	ng/m ³ Air	1.6667	ND	78.2	60-120	7.46	20	
Benzo(k)fluoranthene	1.30	ng/m ³ Air	1.6433	ND	79.3	60-120	8.03	20	
Benzo(e)pyrene	1.21	ng/m ³ Air	1.6467	ND	73.3	60-120	6.13	20	
Benzo(a)pyrene	1.20	ng/m ³ Air	1.6667	ND	72.0	60-120	6.76	20	
Perylene	1.20	ng/m ³ Air	1.6633	ND	72.0	60-120	4.81	20	
Dibenz(a,h)anthracene	1.08	ng/m ³ Air	1.6667	ND	64.5	60-120	8.54	20	
Indeno(1,2,3-c,d)pyrene	1.16	ng/m ³ Air	1.6667	ND	69.9	60-120	8.55	20	
Benzo(g,h,i)perylene	1.12	ng/m ³ Air	1.6667	ND	67.3	60-120	7.79	20	
Coronene	0.201	ng/m ³ Air	0.33333	ND	60.4	60-120	7.34	20	
Surrogate: Fluorene-d10	0.788	ng/ul	0.98000	ND	80.4	60-120			
Surrogate: Fluoranthene-d10	2.51	ng/m ³ Air	3.3333	ND	75.2	60-120			
Surrogate: Pyrene-d10	0.639	ng/ul	0.99900	ND	63.9	60-120			
Surrogate: Benzo(a)pyrene-d12	2.49	ng/m ³ Air	3.3667	ND	74.0	60-120			

Calibration Blank (B6F1703-CCB1)

Prepared: 06/17/26 Analyzed: 07/09/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	0.00	ng/m ³ Air		ND					U
Anthracene	0.00240	ng/m ³ Air		ND					LB
Fluoranthene	0.00	ng/m ³ Air		ND					U
Pyrene	0.00	ng/m ³ Air		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F1703 - TO-13 extraction

Calibration Blank (B6F1703-CCB1) Contin

Prepared: 06/17/26 Analyzed: 07/09/26

Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND		60-120			U

Batch B6F2407 - TO-13 extraction

Blank (B6F2407-BLK1)

Prepared: 06/24/26 Analyzed: 07/22/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							U
Fluoranthene	ND	ng/m ³ Air							U
Pyrene	ND	ng/m ³ Air							LB, U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U
Surrogate: Fluorene-d10	0.749	ng/ul		0.98000	76.4	60-120			
Surrogate: Fluoranthene-d10	2.58	ng/m ³ Air		3.4000	75.9	60-120			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F2407 - TO-13 extraction

Blank (B6F2407-BLK1) Continued

Prepared: 06/24/26 Analyzed: 07/22/26

Surrogate: Pyrene-d10	0.714	ng/ul	0.99900	71.4	60-120
Surrogate: Benzo(a)pyrene-d12	2.60	ng/m ³ Air	3.3233	78.3	60-120

LCS (B6F2407-BS1)

Prepared: 06/24/26 Analyzed: 07/22/26

Naphthalene	18.6	ng/m ³ Air	25.000	ND	74.5	60-120	
Acenaphthylene	1.28	ng/m ³ Air	1.6667	ND	76.8	60-120	
Acenaphthene	6.48	ng/m ³ Air	8.3333	ND	77.8	60-120	
Fluorene	6.60	ng/m ³ Air	8.3333	ND	79.2	60-120	
Phenanthrene	10.2	ng/m ³ Air	12.600	ND	81.0	60-120	LB
Anthracene	1.21	ng/m ³ Air	1.6633	ND	73.0	60-120	
Fluoranthene	7.12	ng/m ³ Air	8.4333	ND	84.4	60-120	
Pyrene	2.88	ng/m ³ Air	4.1333	ND	69.7	60-120	LB
Benzo(a)anthracene	1.46	ng/m ³ Air	1.6667	ND	87.8	60-120	
Chrysene	1.37	ng/m ³ Air	1.6667	ND	82.2	60-120	
Benzo(b)fluoranthene	1.41	ng/m ³ Air	1.6667	ND	84.3	60-120	
Benzo(k)fluoranthene	1.38	ng/m ³ Air	1.6433	ND	84.1	60-120	
Benzo(e)pyrene	1.32	ng/m ³ Air	1.6467	ND	79.9	60-120	
Benzo(a)pyrene	1.35	ng/m ³ Air	1.6667	ND	80.7	60-120	
Perylene	1.34	ng/m ³ Air	1.6633	ND	80.5	60-120	
Dibenz(a,h)anthracene	1.52	ng/m ³ Air	1.6667	ND	90.9	60-120	
Indeno(1,2,3-c,d)pyrene	1.53	ng/m ³ Air	1.6667	ND	91.6	60-120	
Benzo(g,h,i)perylene	1.42	ng/m ³ Air	1.6667	ND	85.0	60-120	
Coronene	0.331	ng/m ³ Air	0.33333	ND	99.2	60-120	

Surrogate: Fluorene-d10	0.781	ng/ul	0.98000	ND	79.7	60-120
Surrogate: Fluoranthene-d10	2.89	ng/m ³ Air	3.4000	ND	85.0	60-120
Surrogate: Pyrene-d10	0.710	ng/ul	0.99900	ND	71.1	60-120
Surrogate: Benzo(a)pyrene-d12	2.81	ng/m ³ Air	3.3233	ND	84.7	60-120

LCS Dup (B6F2407-BSD1)

Prepared: 06/24/26 Analyzed: 07/22/26

Naphthalene	20.3	ng/m ³ Air	25.000	ND	81.0	60-120	8.41	20	
Acenaphthylene	1.36	ng/m ³ Air	1.6667	ND	81.6	60-120	6.06	20	
Acenaphthene	6.94	ng/m ³ Air	8.3333	ND	83.2	60-120	6.74	20	
Fluorene	7.01	ng/m ³ Air	8.3333	ND	84.1	60-120	6.07	20	
Phenanthrene	10.9	ng/m ³ Air	12.600	ND	86.5	60-120	6.51	20	LB
Anthracene	1.26	ng/m ³ Air	1.6633	ND	76.0	60-120	4.06	20	
Fluoranthene	7.34	ng/m ³ Air	8.4333	ND	87.1	60-120	3.14	20	
Pyrene	3.12	ng/m ³ Air	4.1333	ND	75.5	60-120	7.99	20	LB
Benzo(a)anthracene	1.52	ng/m ³ Air	1.6667	ND	91.4	60-120	3.99	20	
Chrysene	1.42	ng/m ³ Air	1.6667	ND	85.3	60-120	3.72	20	
Benzo(b)fluoranthene	1.46	ng/m ³ Air	1.6667	ND	87.5	60-120	3.63	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F2407 - TO-13 extraction

LCS Dup (B6F2407-BSD1) Continued

Prepared: 06/24/26 Analyzed: 07/22/26

Benzo(k)fluoranthene	1.43	ng/m ³ Air	1.6433	ND	87.3	60-120	3.72	20	
Benzo(e)pyrene	1.37	ng/m ³ Air	1.6467	ND	83.3	60-120	4.17	20	
Benzo(a)pyrene	1.40	ng/m ³ Air	1.6667	ND	83.8	60-120	3.77	20	
Perylene	1.40	ng/m ³ Air	1.6633	ND	84.4	60-120	4.69	20	
Dibenz(a,h)anthracene	1.55	ng/m ³ Air	1.6667	ND	93.1	60-120	2.33	20	
Indeno(1,2,3-c,d)pyrene	1.58	ng/m ³ Air	1.6667	ND	94.5	60-120	3.16	20	
Benzo(g,h,i)perylene	1.47	ng/m ³ Air	1.6667	ND	88.2	60-120	3.74	20	
Coronene	0.333	ng/m ³ Air	0.33333	ND	100	60-120	0.803	20	

Surrogate: Fluorene-d10

0.851 ng/ul

0.98000

ND

86.8 60-120

Surrogate: Fluoranthene-d10

3.02 ng/m³ Air

3.4000

ND

88.9 60-120

Surrogate: Pyrene-d10

0.786 ng/ul

0.99900

ND

78.7 60-120

Surrogate: Benzo(a)pyrene-d12

2.91 ng/m³ Air

3.3233

ND

87.7 60-120

Calibration Blank (B6F2407-CCB1)

Prepared: 06/24/26 Analyzed: 07/22/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	0.00120	ng/m ³ Air		ND					LB
Anthracene	0.00	ng/m ³ Air		ND					U
Fluoranthene	0.00	ng/m ³ Air		ND					U
Pyrene	6.00E-4	ng/m ³ Air		ND					LB
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U

Surrogate: Fluorene-d10

0.00 ng/m³ Air

ND

60-120

U

Surrogate: Fluoranthene-d10

0.00 ng/m³ Air

ND

60-120

U

Surrogate: Pyrene-d10

0.00 ng/m³ Air

ND

60-120

U

Surrogate: Benzo(a)pyrene-d12

0.00 ng/m³ Air

ND

60-120

U

Batch B6F3003 - TO-13 extraction

Blank (B6F3003-BLK1)

Prepared: 06/30/26 Analyzed: 07/27/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F3003 - TO-13 extraction

Blank (B6F3003-BLK1) Continued

Prepared: 06/30/26 Analyzed: 07/27/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							U
Fluoranthene	ND	ng/m ³ Air							LB, U
Pyrene	ND	ng/m ³ Air							LB, U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U
Surrogate: Fluorene-d10	0.794	ng/ul	0.98000		81.0	60-120			
Surrogate: Fluoranthene-d10	2.90	ng/m ³ Air	3.4000		85.4	60-120			
Surrogate: Pyrene-d10	0.767	ng/ul	0.99900		76.8	60-120			
Surrogate: Benzo(a)pyrene-d12	2.93	ng/m ³ Air	3.3233		88.3	60-120			

LCS (B6F3003-BS1)

Prepared: 06/30/26 Analyzed: 07/27/26

Naphthalene	20.6	ng/m ³ Air	25.000	ND	82.4	60-120			
Acenaphthylene	1.43	ng/m ³ Air	1.6667	ND	85.9	60-120			
Acenaphthene	7.16	ng/m ³ Air	8.3333	ND	85.9	60-120			
Fluorene	7.14	ng/m ³ Air	8.3333	ND	85.7	60-120			
Phenanthrene	11.4	ng/m ³ Air	12.600	ND	90.4	60-120			LB
Anthracene	1.31	ng/m ³ Air	1.6633	ND	78.6	60-120			
Fluoranthene	7.44	ng/m ³ Air	8.4333	ND	88.2	60-120			LB
Pyrene	3.46	ng/m ³ Air	4.1333	ND	83.8	60-120			LB
Benzo(a)anthracene	1.58	ng/m ³ Air	1.6667	ND	95.1	60-120			
Chrysene	1.49	ng/m ³ Air	1.6667	ND	89.5	60-120			
Benzo(b)fluoranthene	1.52	ng/m ³ Air	1.6667	ND	91.1	60-120			
Benzo(k)fluoranthene	1.48	ng/m ³ Air	1.6433	ND	90.1	60-120			
Benzo(e)pyrene	1.46	ng/m ³ Air	1.6467	ND	88.8	60-120			
Benzo(a)pyrene	1.43	ng/m ³ Air	1.6667	ND	86.1	60-120			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F3003 - TO-13 extraction

LCS (B6F3003-BS1) Continued

Prepared: 06/30/26 Analyzed: 07/27/26

Perylene	1.46	ng/m ³ Air	1.6633	ND	87.8	60-120
Dibenz(a,h)anthracene	1.57	ng/m ³ Air	1.6667	ND	94.4	60-120
Indeno(1,2,3-c,d)pyrene	1.62	ng/m ³ Air	1.6667	ND	96.9	60-120
Benzo(g,h,i)perylene	1.54	ng/m ³ Air	1.6667	ND	92.6	60-120
Coronene	0.349	ng/m ³ Air	0.33333	ND	105	60-120
Surrogate: Fluorene-d10	0.824	ng/ul	0.98000	ND	84.1	60-120
Surrogate: Fluoranthene-d10	3.03	ng/m ³ Air	3.4000	ND	89.0	60-120
Surrogate: Pyrene-d10	0.822	ng/ul	0.99900	ND	82.3	60-120
Surrogate: Benzo(a)pyrene-d12	3.00	ng/m ³ Air	3.3233	ND	90.2	60-120

LCS Dup (B6F3003-BSD1)

Prepared: 06/30/26 Analyzed: 07/27/26

Naphthalene	19.9	ng/m ³ Air	25.000	ND	79.7	60-120	3.31	20	
Acenaphthylene	1.35	ng/m ³ Air	1.6667	ND	80.9	60-120	6.02	20	
Acenaphthene	6.84	ng/m ³ Air	8.3333	ND	82.0	60-120	4.63	20	
Fluorene	6.84	ng/m ³ Air	8.3333	ND	82.1	60-120	4.34	20	
Phenanthrene	10.6	ng/m ³ Air	12.600	ND	83.9	60-120	7.45	20	LB
Anthracene	1.22	ng/m ³ Air	1.6633	ND	73.6	60-120	6.59	20	
Fluoranthene	6.88	ng/m ³ Air	8.4333	ND	81.6	60-120	7.72	20	LB
Pyrene	3.25	ng/m ³ Air	4.1333	ND	78.6	60-120	6.37	20	LB
Benzo(a)anthracene	1.48	ng/m ³ Air	1.6667	ND	88.6	60-120	7.06	20	
Chrysene	1.39	ng/m ³ Air	1.6667	ND	83.1	60-120	7.41	20	
Benzo(b)fluoranthene	1.44	ng/m ³ Air	1.6667	ND	86.2	60-120	5.55	20	
Benzo(k)fluoranthene	1.38	ng/m ³ Air	1.6433	ND	84.3	60-120	6.70	20	
Benzo(e)pyrene	1.38	ng/m ³ Air	1.6467	ND	83.7	60-120	5.89	20	
Benzo(a)pyrene	1.34	ng/m ³ Air	1.6667	ND	80.2	60-120	7.05	20	
Perylene	1.37	ng/m ³ Air	1.6633	ND	82.4	60-120	6.33	20	
Dibenz(a,h)anthracene	1.40	ng/m ³ Air	1.6667	ND	84.3	60-120	11.3	20	
Indeno(1,2,3-c,d)pyrene	1.46	ng/m ³ Air	1.6667	ND	87.7	60-120	9.94	20	
Benzo(g,h,i)perylene	1.40	ng/m ³ Air	1.6667	ND	84.2	60-120	9.44	20	
Coronene	0.305	ng/m ³ Air	0.33333	ND	91.5	60-120	13.5	20	
Surrogate: Fluorene-d10	0.814	ng/ul	0.98000	ND	83.1	60-120			
Surrogate: Fluoranthene-d10	2.79	ng/m ³ Air	3.4000	ND	82.1	60-120			
Surrogate: Pyrene-d10	0.789	ng/ul	0.99900	ND	78.9	60-120			
Surrogate: Benzo(a)pyrene-d12	2.79	ng/m ³ Air	3.3233	ND	83.9	60-120			

Calibration Blank (B6F3003-CCB1)

Prepared: 06/30/26 Analyzed: 07/27/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F3003 - TO-13 extraction

Calibration Blank (B6F3003-CCB1) Contin

Prepared: 06/30/26 Analyzed: 07/27/26

Phenanthrene	0.00120	ng/m ³ Air		ND					LB
Anthracene	0.00	ng/m ³ Air		ND					U
Fluoranthene	7.00E-4	ng/m ³ Air		ND					LB
Pyrene	4.00E-4	ng/m ³ Air		ND					LB
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND		60-120			U

Duplicate (B6F3003-DUP1)

Source: 6062604-02 Prepared: 06/30/26 Analyzed: 07/28/26

Naphthalene	13.7	ng/m ³ Air		13.80		1.10	10		TT
Acenaphthylene	ND	ng/m ³ Air		ND			10		TT, U
Acenaphthene	1.26	ng/m ³ Air		1.25		0.593	10		TT
Fluorene	1.68	ng/m ³ Air		1.67		0.278	10		TT
Phenanthrene	3.21	ng/m ³ Air		3.24		0.664	10		LB, QB-03, TT
Anthracene	0.0450	ng/m ³ Air		0.04		0.692	10		TT
Fluoranthene	0.815	ng/m ³ Air		0.81		0.267	10		LB, TT
Pyrene	0.369	ng/m ³ Air		0.38		1.42	10		LB, TT
Benzo(a)anthracene	0.0202	ng/m ³ Air		0.02		1.55	10		TT
Chrysene	0.0515	ng/m ³ Air		0.05		3.06	10		TT
Benzo(b)fluoranthene	0.0419	ng/m ³ Air		0.04		3.01	10		TT
Benzo(k)fluoranthene	0.0127	ng/m ³ Air		ND			10		TT
Benzo(e)pyrene	0.0298	ng/m ³ Air		0.03		2.11	10		TT
Benzo(a)pyrene	0.0149	ng/m ³ Air		0.01		4.26	10		TT
Perylene	ND	ng/m ³ Air		ND			10		TT, U
Dibenz(a,h)anthracene	ND	ng/m ³ Air		ND			10		TT, U
Indeno(1,2,3-c,d)pyrene	0.0224	ng/m ³ Air		0.02		2.82	10		TT
Benzo(g,h,i)perylene	0.0379	ng/m ³ Air		0.04		1.63	10		TT

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F3003 - TO-13 extraction

Duplicate (B6F3003-DUP1) Continued Source: 6062604-02 Prepared: 06/30/26 Analyzed: 07/28/26

Coronene	ND	ng/m ³ Air	ND				10	TT, U
Surrogate: Fluorene-d10	0.774	ng/ul	0.98000	ND	79.0	60-120		TT
Surrogate: Fluoranthene-d10	2.72	ng/m ³ Air	3.1667	ND	85.8	60-120		TT
Surrogate: Pyrene-d10	0.694	ng/ul	0.99900	ND	69.5	60-120		TT
Surrogate: Benzo(a)pyrene-d12	2.43	ng/m ³ Air	3.0953	ND	78.6	60-120		TT

Batch B6G0803 - TO-13 extraction

Blank (B6G0803-BLK1) Prepared: 07/08/26 Analyzed: 08/04/26

Naphthalene	ND	ng/m ³ Air						LL, S-GC, U
Acenaphthylene	ND	ng/m ³ Air						LL, S-GC, U
Acenaphthene	ND	ng/m ³ Air						LL, S-GC, U
Fluorene	ND	ng/m ³ Air						LL, S-GC, U
Phenanthrene	ND	ng/m ³ Air						LB, LL, S-GC, U
Anthracene	ND	ng/m ³ Air						LL, S-GC, U
Fluoranthene	ND	ng/m ³ Air						S-GC, LB, LL, U
Pyrene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(a)anthracene	ND	ng/m ³ Air						LL, S-GC, U
Chrysene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(b)fluoranthene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(k)fluoranthene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(e)pyrene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(a)pyrene	ND	ng/m ³ Air						LL, S-GC, U
Perylene	ND	ng/m ³ Air						LL, S-GC, U
Dibenz(a,h)anthracene	ND	ng/m ³ Air						LL, S-GC, U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air						LL, S-GC, U
Benzo(g,h,i)perylene	ND	ng/m ³ Air						LL, S-GC, U
Coronene	ND	ng/m ³ Air						LL, S-GC, U
Surrogate: Fluorene-d10	0.549	ng/ul	0.98000		56.0	60-120		LL, S-GC
Surrogate: Fluoranthene-d10	2.36	ng/m ³ Air	3.4000		69.3	60-120		LL, S-GC
Surrogate: Pyrene-d10	0.536	ng/ul	0.99900		53.6	60-120		LL, S-GC
Surrogate: Benzo(a)pyrene-d12	2.25	ng/m ³ Air	3.3233		67.7	60-120		LL, S-GC

LCS (B6G0803-BS1) Prepared: 07/08/26 Analyzed: 08/04/26

Naphthalene	19.4	ng/m ³ Air	25.000	ND	77.7	60-120		
Acenaphthylene	1.29	ng/m ³ Air	1.6667	ND	77.6	60-120		
Acenaphthene	6.75	ng/m ³ Air	8.3333	ND	81.0	60-120		
Fluorene	7.00	ng/m ³ Air	8.3333	ND	84.0	60-120		
Phenanthrene	10.5	ng/m ³ Air	12.600	ND	83.1	60-120		LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6G0803 - TO-13 extraction

LCS (B6G0803-BS1) Continued

Prepared: 07/08/26 Analyzed: 08/04/26

Anthracene	1.27	ng/m ³ Air	1.6633	ND	76.2	60-120			LB
Fluoranthene	7.22	ng/m ³ Air	8.4333	ND	85.6	60-120			
Pyrene	2.88	ng/m ³ Air	4.1333	ND	69.6	60-120			
Benzo(a)anthracene	1.47	ng/m ³ Air	1.6667	ND	88.1	60-120			
Chrysene	1.37	ng/m ³ Air	1.6667	ND	82.5	60-120			
Benzo(b)fluoranthene	1.41	ng/m ³ Air	1.6667	ND	84.4	60-120			
Benzo(k)fluoranthene	1.38	ng/m ³ Air	1.6433	ND	84.0	60-120			
Benzo(e)pyrene	1.31	ng/m ³ Air	1.6467	ND	79.7	60-120			
Benzo(a)pyrene	1.33	ng/m ³ Air	1.6667	ND	80.0	60-120			
Perylene	1.34	ng/m ³ Air	1.6633	ND	80.7	60-120			
Dibenz(a,h)anthracene	1.38	ng/m ³ Air	1.6667	ND	83.0	60-120			
Indeno(1,2,3-c,d)pyrene	1.41	ng/m ³ Air	1.6667	ND	84.8	60-120			
Benzo(g,h,i)perylene	1.32	ng/m ³ Air	1.6667	ND	79.4	60-120			
Coronene	0.287	ng/m ³ Air	0.33333	ND	86.1	60-120			
Surrogate: Fluorene-d10	0.795	ng/ul	0.98000	ND	81.1	60-120			
Surrogate: Fluoranthene-d10	2.94	ng/m ³ Air	3.4000	ND	86.5	60-120			
Surrogate: Pyrene-d10	0.690	ng/ul	0.99900	ND	69.0	60-120			
Surrogate: Benzo(a)pyrene-d12	2.80	ng/m ³ Air	3.3233	ND	84.3	60-120			

LCS Dup (B6G0803-BSD1)

Prepared: 07/08/26 Analyzed: 08/04/26

Naphthalene	19.6	ng/m ³ Air	25.000	ND	78.5	60-120	1.03	20	
Acenaphthylene	1.31	ng/m ³ Air	1.6667	ND	78.7	60-120	1.46	20	
Acenaphthene	6.87	ng/m ³ Air	8.3333	ND	82.4	60-120	1.78	20	
Fluorene	7.11	ng/m ³ Air	8.3333	ND	85.3	60-120	1.54	20	
Phenanthrene	10.6	ng/m ³ Air	12.600	ND	84.5	60-120	1.72	20	LB
Anthracene	1.28	ng/m ³ Air	1.6633	ND	77.1	60-120	1.18	20	
Fluoranthene	7.52	ng/m ³ Air	8.4333	ND	89.2	60-120	4.14	20	LB
Pyrene	2.99	ng/m ³ Air	4.1333	ND	72.4	60-120	3.98	20	
Benzo(a)anthracene	1.50	ng/m ³ Air	1.6667	ND	90.0	60-120	2.11	20	
Chrysene	1.40	ng/m ³ Air	1.6667	ND	84.2	60-120	2.06	20	
Benzo(b)fluoranthene	1.44	ng/m ³ Air	1.6667	ND	86.5	60-120	2.44	20	
Benzo(k)fluoranthene	1.42	ng/m ³ Air	1.6433	ND	86.6	60-120	3.02	20	
Benzo(e)pyrene	1.34	ng/m ³ Air	1.6467	ND	81.7	60-120	2.46	20	
Benzo(a)pyrene	1.36	ng/m ³ Air	1.6667	ND	81.5	60-120	1.88	20	
Perylene	1.36	ng/m ³ Air	1.6633	ND	82.0	60-120	1.60	20	
Dibenz(a,h)anthracene	1.39	ng/m ³ Air	1.6667	ND	83.3	60-120	0.409	20	
Indeno(1,2,3-c,d)pyrene	1.42	ng/m ³ Air	1.6667	ND	85.3	60-120	0.658	20	
Benzo(g,h,i)perylene	1.34	ng/m ³ Air	1.6667	ND	80.4	60-120	1.25	20	
Coronene	0.280	ng/m ³ Air	0.33333	ND	83.9	60-120	2.59	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6G0803 - TO-13 extraction

LCS Dup (B6G0803-BSD1) Continued

Prepared: 07/08/26 Analyzed: 08/04/26

Surrogate: Fluorene-d10	0.803	ng/ul	0.98000	ND	81.9	60-120
Surrogate: Fluoranthene-d10	3.06	ng/m ³ Air	3.4000	ND	90.0	60-120
Surrogate: Pyrene-d10	0.708	ng/ul	0.99900	ND	70.9	60-120
Surrogate: Benzo(a)pyrene-d12	2.83	ng/m ³ Air	3.3233	ND	85.2	60-120

Calibration Blank (B6G0803-CCB1)

Prepared: 07/08/26 Analyzed: 08/04/26

Naphthalene	0.00	ng/m ³ Air		ND			U
Acenaphthylene	0.00	ng/m ³ Air		ND			U
Acenaphthene	0.00	ng/m ³ Air		ND			U
Fluorene	0.00	ng/m ³ Air		ND			U
Phenanthrene	5.00E-4	ng/m ³ Air		ND			LB
Anthracene	0.00	ng/m ³ Air		ND			U
Fluoranthene	0.00	ng/m ³ Air		ND			U
Pyrene	0.00	ng/m ³ Air		ND			U
Benzo(a)anthracene	0.00	ng/m ³ Air		ND			U
Chrysene	0.00	ng/m ³ Air		ND			U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND			U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND			U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND			U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND			U
Perylene	0.00	ng/m ³ Air		ND			U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND			U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND			U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND			U
Coronene	0.00	ng/m ³ Air		ND			U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air	ND		60-120		U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air	ND		60-120		U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air	ND		60-120		U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air	ND		60-120		U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606055

Calibration Check (2606055-CCV1)

Prepared & Analyzed: 06/17/26

Naphthalene	7.76	ng/ul	3.4	30.00	
Acenaphthylene	0.517	ng/ul	2.9	30.00	
Acenaphthene	2.59	ng/ul	3.3	30.00	
Fluorene	2.62	ng/ul	4.7	30.00	
Phenanthrene	3.87	ng/ul	3.1	30.00	LB
Anthracene	0.503	ng/ul	0.3	30.00	LB
Fluoranthene	2.57	ng/ul	2.6	30.00	
Pyrene	1.26	ng/ul	0.7	30.00	
Benzo(a)anthracene	0.538	ng/ul	7.2	30.00	
Chrysene	0.515	ng/ul	2.6	30.00	
Benzo(b)fluoranthene	0.527	ng/ul	5.2	30.00	
Benzo(k)fluoranthene	0.541	ng/ul	8.0	30.00	
Benzo(e)pyrene	0.520	ng/ul	3.7	30.00	
Benzo(a)pyrene	0.531	ng/ul	5.8	30.00	
Perylene	0.517	ng/ul	3.1	30.00	
Dibenz(a,h)anthracene	0.512	ng/ul	1.9	30.00	
Indeno(1,2,3-c,d)pyrene	0.523	ng/ul	4.3	30.00	
Benzo(g,h,i)perylene	0.515	ng/ul	2.8	30.00	
Coronene	0.505	ng/ul	0.6	30.00	
Surrogate: Fluorene-d10	0.513	ng/ul	4.7	30.00	
Surrogate: Fluoranthene-d10	0.524	ng/ul	2.8	30.00	
Surrogate: Pyrene-d10	0.500	ng/ul	0.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.526	ng/ul	5.5	30.00	

Sequence 2607036

Calibration Check (2607036-CCV1)

Prepared & Analyzed: 07/09/26

Naphthalene	7.21	ng/ul	-3.9	30.00	
Acenaphthylene	0.500	ng/ul	-0.3	30.00	
Acenaphthene	2.42	ng/ul	-3.5	30.00	
Fluorene	2.50	ng/ul	0.04	30.00	
Phenanthrene	3.56	ng/ul	-5.2	30.00	
Anthracene	0.479	ng/ul	-4.4	30.00	LB
Fluoranthene	2.39	ng/ul	-4.4	30.00	
Pyrene	1.10	ng/ul	-11.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607036

Calibration Check (2607036-CCV1) Continued

Prepared & Analyzed: 07/09/26

Benzo(a)anthracene	0.497	ng/ul	-0.9	30.00
Chrysene	0.475	ng/ul	-5.5	30.00
Benzo(b)fluoranthene	0.498	ng/ul	-0.7	30.00
Benzo(k)fluoranthene	0.509	ng/ul	1.5	30.00
Benzo(e)pyrene	0.470	ng/ul	-6.2	30.00
Benzo(a)pyrene	0.489	ng/ul	-2.6	30.00
Perylene	0.472	ng/ul	-5.8	30.00
Dibenz(a,h)anthracene	0.425	ng/ul	-15.5	30.00
Indeno(1,2,3-c,d)pyrene	0.447	ng/ul	-10.9	30.00
Benzo(g,h,i)perylene	0.446	ng/ul	-11.0	30.00
Coronene	0.410	ng/ul	-18.3	30.00

Surrogate: Fluorene-d10

0.532 ng/ul 8.6 30.00

Surrogate: Fluoranthene-d10

0.496 ng/ul -2.8 30.00

Surrogate: Pyrene-d10

0.444 ng/ul -11.0 30.00

Surrogate: Benzo(a)pyrene-d12

0.483 ng/ul -3.1 30.00

Calibration Check (2607036-CCV2)

Prepared: 07/09/26 Analyzed: 07/10/26

Naphthalene	7.23	ng/ul	-3.7	30.00
Acenaphthylene	0.491	ng/ul	-2.1	30.00
Acenaphthene	2.43	ng/ul	-2.8	30.00
Fluorene	2.52	ng/ul	0.6	30.00
Phenanthrene	3.57	ng/ul	-4.9	30.00
Anthracene	0.479	ng/ul	-4.4	30.00
Fluoranthene	2.39	ng/ul	-4.6	30.00
Pyrene	1.07	ng/ul	-14.2	30.00
Benzo(a)anthracene	0.493	ng/ul	-1.8	30.00
Chrysene	0.473	ng/ul	-5.8	30.00
Benzo(b)fluoranthene	0.496	ng/ul	-1.0	30.00
Benzo(k)fluoranthene	0.504	ng/ul	0.6	30.00
Benzo(e)pyrene	0.471	ng/ul	-6.1	30.00
Benzo(a)pyrene	0.485	ng/ul	-3.3	30.00
Perylene	0.474	ng/ul	-5.5	30.00
Dibenz(a,h)anthracene	0.425	ng/ul	-15.5	30.00
Indeno(1,2,3-c,d)pyrene	0.443	ng/ul	-11.6	30.00
Benzo(g,h,i)perylene	0.449	ng/ul	-10.5	30.00
Coronene	0.403	ng/ul	-19.8	30.00

LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607036

Calibration Check (2607036-CCV2) Continued

Prepared: 07/09/26 Analyzed: 07/10/26

Surrogate: Fluorene-d10	0.530	ng/ul	8.2	30.00
Surrogate: Fluoranthene-d10	0.493	ng/ul	-3.4	30.00
Surrogate: Pyrene-d10	0.430	ng/ul	-14.0	30.00
Surrogate: Benzo(a)pyrene-d12	0.481	ng/ul	-3.6	30.00

Sequence 2607048

Calibration Check (2607048-CCV1)

Prepared & Analyzed: 07/14/26

Naphthalene	7.75	ng/ul	3.3	30.00	
Acenaphthylene	0.524	ng/ul	4.3	30.00	
Acenaphthene	2.61	ng/ul	4.4	30.00	
Fluorene	2.60	ng/ul	3.7	30.00	
Phenanthrene	3.96	ng/ul	5.6	30.00	
Anthracene	0.528	ng/ul	5.4	30.00	
Fluoranthene	2.68	ng/ul	7.3	30.00	LB
Pyrene	1.22	ng/ul	-2.5	30.00	
Benzo(a)anthracene	0.532	ng/ul	6.0	30.00	
Chrysene	0.531	ng/ul	5.7	30.00	
Benzo(b)fluoranthene	0.534	ng/ul	6.6	30.00	
Benzo(k)fluoranthene	0.533	ng/ul	6.4	30.00	
Benzo(e)pyrene	0.528	ng/ul	5.3	30.00	
Benzo(a)pyrene	0.535	ng/ul	6.5	30.00	
Perylene	0.529	ng/ul	5.6	30.00	
Dibenz(a,h)anthracene	0.554	ng/ul	10.2	30.00	
Indeno(1,2,3-c,d)pyrene	0.544	ng/ul	8.5	30.00	
Benzo(g,h,i)perylene	0.532	ng/ul	6.1	30.00	LB
Coronene	0.538	ng/ul	7.2	30.00	

Surrogate: Fluorene-d10	0.505	ng/ul	3.1	30.00
Surrogate: Fluoranthene-d10	0.542	ng/ul	6.4	30.00
Surrogate: Pyrene-d10	0.481	ng/ul	-3.8	30.00
Surrogate: Benzo(a)pyrene-d12	0.530	ng/ul	6.2	30.00

Calibration Check (2607048-CCV2)

Prepared: 07/14/26 Analyzed: 07/15/26

Naphthalene	7.81	ng/ul	4.0	30.00
Acenaphthylene	0.521	ng/ul	3.8	30.00
Acenaphthene	2.63	ng/ul	5.2	30.00
Fluorene	2.67	ng/ul	6.6	30.00
Phenanthrene	3.96	ng/ul	5.4	30.00
Anthracene	0.550	ng/ul	9.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607048

Calibration Check (2607048-CCV2) Continued

Prepared: 07/14/26 Analyzed: 07/15/26

Fluoranthene	2.81	ng/ul	12.3	30.00	LB
Pyrene	1.18	ng/ul	-5.5	30.00	
Benzo(a)anthracene	0.539	ng/ul	7.4	30.00	
Chrysene	0.530	ng/ul	5.5	30.00	
Benzo(b)fluoranthene	0.531	ng/ul	6.0	30.00	
Benzo(k)fluoranthene	0.535	ng/ul	6.7	30.00	
Benzo(e)pyrene	0.521	ng/ul	4.0	30.00	
Benzo(a)pyrene	0.532	ng/ul	6.0	30.00	
Perylene	0.532	ng/ul	6.0	30.00	
Dibenz(a,h)anthracene	0.522	ng/ul	3.9	30.00	
Indeno(1,2,3-c,d)pyrene	0.521	ng/ul	3.8	30.00	
Benzo(g,h,i)perylene	0.513	ng/ul	2.3	30.00	LB
Coronene	0.504	ng/ul	0.3	30.00	
Surrogate: Fluorene-d10	0.514	ng/ul	4.9	30.00	
Surrogate: Fluoranthene-d10	0.566	ng/ul	11.0	30.00	
Surrogate: Pyrene-d10	0.469	ng/ul	-6.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.527	ng/ul	5.7	30.00	

Calibration Check (2607048-CCV3)

Prepared: 07/14/26 Analyzed: 07/15/26

Naphthalene	7.76	ng/ul	3.4	30.00	
Acenaphthylene	0.521	ng/ul	3.7	30.00	
Acenaphthene	2.62	ng/ul	4.8	30.00	
Fluorene	2.64	ng/ul	5.3	30.00	
Phenanthrene	3.92	ng/ul	4.4	30.00	
Anthracene	0.543	ng/ul	8.3	30.00	
Fluoranthene	2.79	ng/ul	11.4	30.00	LB
Pyrene	1.18	ng/ul	-5.7	30.00	
Benzo(a)anthracene	0.534	ng/ul	6.5	30.00	
Chrysene	0.528	ng/ul	5.1	30.00	
Benzo(b)fluoranthene	0.530	ng/ul	5.7	30.00	
Benzo(k)fluoranthene	0.534	ng/ul	6.6	30.00	
Benzo(e)pyrene	0.523	ng/ul	4.4	30.00	
Benzo(a)pyrene	0.528	ng/ul	5.1	30.00	
Perylene	0.527	ng/ul	5.1	30.00	
Dibenz(a,h)anthracene	0.512	ng/ul	1.7	30.00	
Indeno(1,2,3-c,d)pyrene	0.508	ng/ul	1.2	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607048

Calibration Check (2607048-CCV3) Continued

Prepared: 07/14/26 Analyzed: 07/15/26

Benzo(g,h,i)perylene	0.504	ng/ul	0.4	30.00	LB
Coronene	0.477	ng/ul	-5.0	30.00	
Surrogate: Fluorene-d10	0.509	ng/ul	3.8	30.00	
Surrogate: Fluoranthene-d10	0.560	ng/ul	9.9	30.00	
Surrogate: Pyrene-d10	0.469	ng/ul	-6.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.525	ng/ul	5.3	30.00	

Sequence 2607063

Calibration Check (2607063-CCV1)

Prepared & Analyzed: 07/17/26

Naphthalene	7.78	ng/ul	3.7	30.00	LB
Acenaphthylene	0.525	ng/ul	4.7	30.00	
Acenaphthene	2.61	ng/ul	4.2	30.00	
Fluorene	2.61	ng/ul	4.4	30.00	
Phenanthrene	3.92	ng/ul	4.5	30.00	
Anthracene	0.528	ng/ul	5.3	30.00	
Fluoranthene	2.72	ng/ul	8.7	30.00	
Pyrene	1.18	ng/ul	-5.5	30.00	
Benzo(a)anthracene	0.535	ng/ul	6.5	30.00	
Chrysene	0.529	ng/ul	5.5	30.00	
Benzo(b)fluoranthene	0.534	ng/ul	6.6	30.00	
Benzo(k)fluoranthene	0.534	ng/ul	6.5	30.00	
Benzo(e)pyrene	0.521	ng/ul	4.0	30.00	
Benzo(a)pyrene	0.536	ng/ul	6.7	30.00	
Perylene	0.529	ng/ul	5.5	30.00	
Dibenz(a,h)anthracene	0.565	ng/ul	12.4	30.00	
Indeno(1,2,3-c,d)pyrene	0.555	ng/ul	10.5	30.00	
Benzo(g,h,i)perylene	0.540	ng/ul	7.6	30.00	
Coronene	0.555	ng/ul	10.7	30.00	
Surrogate: Fluorene-d10	0.508	ng/ul	3.8	30.00	
Surrogate: Fluoranthene-d10	0.552	ng/ul	8.2	30.00	
Surrogate: Pyrene-d10	0.473	ng/ul	-5.2	30.00	
Surrogate: Benzo(a)pyrene-d12	0.530	ng/ul	6.3	30.00	

Calibration Check (2607063-CCV2)

Prepared & Analyzed: 07/17/26

Naphthalene	7.76	ng/ul	3.4	30.00	
Acenaphthylene	0.520	ng/ul	3.5	30.00	
Acenaphthene	2.61	ng/ul	4.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607063

Calibration Check (2607063-CCV2) Continued

Prepared & Analyzed: 07/17/26

Fluorene	2.65	ng/ul	6.1	30.00	LB
Phenanthrene	3.89	ng/ul	3.7	30.00	
Anthracene	0.542	ng/ul	8.1	30.00	
Fluoranthene	2.77	ng/ul	10.6	30.00	
Pyrene	1.16	ng/ul	-7.2	30.00	
Benzo(a)anthracene	0.539	ng/ul	7.3	30.00	
Chrysene	0.528	ng/ul	5.2	30.00	
Benzo(b)fluoranthene	0.528	ng/ul	5.2	30.00	
Benzo(k)fluoranthene	0.537	ng/ul	7.1	30.00	
Benzo(e)pyrene	0.512	ng/ul	2.2	30.00	
Benzo(a)pyrene	0.536	ng/ul	6.7	30.00	
Perylene	0.526	ng/ul	4.8	30.00	
Dibenz(a,h)anthracene	0.557	ng/ul	10.8	30.00	
Indeno(1,2,3-c,d)pyrene	0.546	ng/ul	8.9	30.00	
Benzo(g,h,i)perylene	0.528	ng/ul	5.4	30.00	
Coronene	0.539	ng/ul	7.4	30.00	
Surrogate: Fluorene-d10	0.515	ng/ul	5.2	30.00	
Surrogate: Fluoranthene-d10	0.558	ng/ul	9.3	30.00	
Surrogate: Pyrene-d10	0.461	ng/ul	-7.6	30.00	
Surrogate: Benzo(a)pyrene-d12	0.531	ng/ul	6.6	30.00	

Calibration Check (2607063-CCV3)

Prepared: 07/17/26 Analyzed: 07/18/26

Naphthalene	7.71	ng/ul	2.7	30.00	LB
Acenaphthylene	0.515	ng/ul	2.6	30.00	
Acenaphthene	2.61	ng/ul	4.4	30.00	
Fluorene	2.62	ng/ul	4.8	30.00	
Phenanthrene	3.90	ng/ul	3.9	30.00	
Anthracene	0.537	ng/ul	7.1	30.00	
Fluoranthene	2.72	ng/ul	8.8	30.00	
Pyrene	1.18	ng/ul	-5.7	30.00	
Benzo(a)anthracene	0.532	ng/ul	6.0	30.00	
Chrysene	0.524	ng/ul	4.5	30.00	
Benzo(b)fluoranthene	0.529	ng/ul	5.5	30.00	
Benzo(k)fluoranthene	0.536	ng/ul	6.9	30.00	
Benzo(e)pyrene	0.518	ng/ul	3.3	30.00	
Benzo(a)pyrene	0.532	ng/ul	5.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607063

Calibration Check (2607063-CCV3) Continued

Prepared: 07/17/26 Analyzed: 07/18/26

Perylene	0.524	ng/ul	4.5	30.00
Dibenz(a,h)anthracene	0.547	ng/ul	8.8	30.00
Indeno(1,2,3-c,d)pyrene	0.539	ng/ul	7.4	30.00
Benzo(g,h,i)perylene	0.523	ng/ul	4.3	30.00
Coronene	0.523	ng/ul	4.2	30.00
Surrogate: Fluorene-d10	0.512	ng/ul	4.5	30.00
Surrogate: Fluoranthene-d10	0.550	ng/ul	7.9	30.00
Surrogate: Pyrene-d10	0.474	ng/ul	-5.1	30.00
Surrogate: Benzo(a)pyrene-d12	0.531	ng/ul	6.5	30.00

Sequence 2607075

Calibration Check (2607075-CCV1)

Prepared & Analyzed: 07/22/26

Naphthalene	7.65	ng/ul	1.9	30.00
Acenaphthylene	0.542	ng/ul	8.0	30.00
Acenaphthene	2.55	ng/ul	1.9	30.00
Fluorene	2.52	ng/ul	0.9	30.00
Phenanthrene	3.90	ng/ul	3.9	30.00
Anthracene	0.511	ng/ul	2.0	30.00
Fluoranthene	2.64	ng/ul	5.6	30.00
Pyrene	1.15	ng/ul	-8.0	30.00
Benzo(a)anthracene	0.522	ng/ul	4.1	30.00
Chrysene	0.518	ng/ul	3.2	30.00
Benzo(b)fluoranthene	0.520	ng/ul	3.8	30.00
Benzo(k)fluoranthene	0.528	ng/ul	5.3	30.00
Benzo(e)pyrene	0.507	ng/ul	1.1	30.00
Benzo(a)pyrene	0.532	ng/ul	5.8	30.00
Perylene	0.521	ng/ul	4.0	30.00
Dibenz(a,h)anthracene	0.584	ng/ul	16.2	30.00
Indeno(1,2,3-c,d)pyrene	0.574	ng/ul	14.4	30.00
Benzo(g,h,i)perylene	0.557	ng/ul	11.1	30.00
Coronene	0.599	ng/ul	19.3	30.00
Surrogate: Fluorene-d10	0.496	ng/ul	1.2	30.00
Surrogate: Fluoranthene-d10	0.531	ng/ul	4.1	30.00
Surrogate: Pyrene-d10	0.459	ng/ul	-8.1	30.00
Surrogate: Benzo(a)pyrene-d12	0.524	ng/ul	5.0	30.00

Calibration Check (2607075-CCV2)

Prepared & Analyzed: 07/22/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607075

Calibration Check (2607075-CCV2) Continued

Prepared & Analyzed: 07/22/26

Naphthalene	7.70	ng/ul	2.6	30.00
Acenaphthylene	0.512	ng/ul	2.1	30.00
Acenaphthene	2.58	ng/ul	2.9	30.00
Fluorene	2.60	ng/ul	3.8	30.00
Phenanthrene	3.89	ng/ul	3.5	30.00
Anthracene	0.526	ng/ul	4.9	30.00
Fluoranthene	2.70	ng/ul	8.1	30.00
Pyrene	1.15	ng/ul	-8.0	30.00
Benzo(a)anthracene	0.527	ng/ul	5.0	30.00
Chrysene	0.518	ng/ul	3.1	30.00
Benzo(b)fluoranthene	0.514	ng/ul	2.5	30.00
Benzo(k)fluoranthene	0.522	ng/ul	4.2	30.00
Benzo(e)pyrene	0.498	ng/ul	-0.7	30.00
Benzo(a)pyrene	0.526	ng/ul	4.6	30.00
Perylene	0.514	ng/ul	2.6	30.00
Dibenz(a,h)anthracene	0.559	ng/ul	11.2	30.00
Indeno(1,2,3-c,d)pyrene	0.550	ng/ul	9.6	30.00
Benzo(g,h,i)perylene	0.534	ng/ul	6.5	30.00
Coronene	0.566	ng/ul	12.7	30.00
Surrogate: Fluorene-d10	0.502	ng/ul	2.5	30.00
Surrogate: Fluoranthene-d10	0.545	ng/ul	6.8	30.00
Surrogate: Pyrene-d10	0.459	ng/ul	-8.1	30.00
Surrogate: Benzo(a)pyrene-d12	0.522	ng/ul	4.7	30.00

Calibration Check (2607075-CCV3)

Prepared: 07/22/26 Analyzed: 07/23/26

Naphthalene	7.62	ng/ul	1.4	30.00
Acenaphthylene	0.511	ng/ul	1.7	30.00
Acenaphthene	2.57	ng/ul	2.6	30.00
Fluorene	2.59	ng/ul	3.4	30.00
Phenanthrene	3.84	ng/ul	2.4	30.00
Anthracene	0.528	ng/ul	5.2	30.00
Fluoranthene	2.70	ng/ul	7.8	30.00
Pyrene	1.17	ng/ul	-6.3	30.00
Benzo(a)anthracene	0.525	ng/ul	4.6	30.00
Chrysene	0.516	ng/ul	2.9	30.00
Benzo(b)fluoranthene	0.517	ng/ul	3.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607075

Calibration Check (2607075-CCV3) Continued

Prepared: 07/22/26 Analyzed: 07/23/26

Benzo(k)fluoranthene	0.525	ng/ul	4.7	30.00
Benzo(e)pyrene	0.502	ng/ul	0.06	30.00
Benzo(a)pyrene	0.525	ng/ul	4.5	30.00
Perylene	0.516	ng/ul	2.9	30.00
Dibenz(a,h)anthracene	0.553	ng/ul	10.0	30.00
Indeno(1,2,3-c,d)pyrene	0.543	ng/ul	8.3	30.00
Benzo(g,h,i)perylene	0.530	ng/ul	5.7	30.00
Coronene	0.541	ng/ul	7.8	30.00

Surrogate: Fluorene-d10	0.505	ng/ul	3.1	30.00
Surrogate: Fluoranthene-d10	0.543	ng/ul	6.5	30.00
Surrogate: Pyrene-d10	0.467	ng/ul	-6.4	30.00
Surrogate: Benzo(a)pyrene-d12	0.522	ng/ul	4.7	30.00

Sequence 2607099

Calibration Check (2607099-CCV1)

Prepared & Analyzed: 07/27/26

Naphthalene	7.78	ng/ul	3.7	30.00
Acenaphthylene	0.557	ng/ul	10.9	30.00
Acenaphthene	2.57	ng/ul	2.6	30.00
Fluorene	2.53	ng/ul	1.0	30.00
Phenanthrene	3.92	ng/ul	4.4	30.00
Anthracene	0.510	ng/ul	1.8	30.00
Fluoranthene	2.61	ng/ul	4.5	30.00
Pyrene	1.24	ng/ul	-0.9	30.00
Benzo(a)anthracene	0.527	ng/ul	5.0	30.00
Chrysene	0.523	ng/ul	4.1	30.00
Benzo(b)fluoranthene	0.524	ng/ul	4.6	30.00
Benzo(k)fluoranthene	0.529	ng/ul	5.5	30.00
Benzo(e)pyrene	0.522	ng/ul	4.1	30.00
Benzo(a)pyrene	0.529	ng/ul	5.3	30.00
Perylene	0.526	ng/ul	4.9	30.00
Dibenz(a,h)anthracene	0.579	ng/ul	15.2	30.00
Indeno(1,2,3-c,d)pyrene	0.576	ng/ul	14.8	30.00
Benzo(g,h,i)perylene	0.569	ng/ul	13.5	30.00
Coronene	0.606	ng/ul	20.9	30.00

Surrogate: Fluorene-d10	0.502	ng/ul	2.4	30.00
Surrogate: Fluoranthene-d10	0.526	ng/ul	3.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607099

Calibration Check (2607099-CCV1) Continued

Prepared & Analyzed: 07/27/26

Surrogate: Pyrene-d10	0.495	ng/ul	-0.9	30.00
Surrogate: Benzo(a)pyrene-d12	0.524	ng/ul	5.1	30.00

Calibration Check (2607099-CCV2)

Prepared & Analyzed: 07/27/26

Naphthalene	7.76	ng/ul	3.4	30.00
Acenaphthylene	0.518	ng/ul	3.2	30.00
Acenaphthene	2.61	ng/ul	4.3	30.00
Fluorene	2.62	ng/ul	4.7	30.00
Phenanthrene	3.89	ng/ul	3.5	30.00
Anthracene	0.522	ng/ul	4.1	30.00
Fluoranthene	2.67	ng/ul	6.8	30.00
Pyrene	1.22	ng/ul	-2.3	30.00
Benzo(a)anthracene	0.531	ng/ul	5.8	30.00
Chrysene	0.522	ng/ul	4.1	30.00
Benzo(b)fluoranthene	0.516	ng/ul	2.9	30.00
Benzo(k)fluoranthene	0.523	ng/ul	4.2	30.00
Benzo(e)pyrene	0.513	ng/ul	2.4	30.00
Benzo(a)pyrene	0.527	ng/ul	4.9	30.00
Perylene	0.524	ng/ul	4.5	30.00
Dibenz(a,h)anthracene	0.554	ng/ul	10.2	30.00
Indeno(1,2,3-c,d)pyrene	0.549	ng/ul	9.5	30.00
Benzo(g,h,i)perylene	0.549	ng/ul	9.5	30.00
Coronene	0.580	ng/ul	15.7	30.00

Surrogate: Fluorene-d10	0.509	ng/ul	3.9	30.00
Surrogate: Fluoranthene-d10	0.540	ng/ul	5.8	30.00
Surrogate: Pyrene-d10	0.485	ng/ul	-2.9	30.00
Surrogate: Benzo(a)pyrene-d12	0.523	ng/ul	5.0	30.00

Calibration Check (2607099-CCV3)

Prepared: 07/27/26 Analyzed: 07/28/26

Naphthalene	7.74	ng/ul	3.1	30.00
Acenaphthylene	0.520	ng/ul	3.5	30.00
Acenaphthene	2.62	ng/ul	4.9	30.00
Fluorene	2.63	ng/ul	5.0	30.00
Phenanthrene	3.91	ng/ul	4.0	30.00
Anthracene	0.534	ng/ul	6.5	30.00
Fluoranthene	2.75	ng/ul	10.1	30.00
Pyrene	1.16	ng/ul	-7.0	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607099

Calibration Check (2607099-CCV3) Continued

Prepared: 07/27/26 Analyzed: 07/28/26

Benzo(a)anthracene	0.533	ng/ul	6.3	30.00
Chrysene	0.521	ng/ul	3.7	30.00
Benzo(b)fluoranthene	0.519	ng/ul	3.5	30.00
Benzo(k)fluoranthene	0.526	ng/ul	4.9	30.00
Benzo(e)pyrene	0.506	ng/ul	1.0	30.00
Benzo(a)pyrene	0.529	ng/ul	5.4	30.00
Perylene	0.521	ng/ul	3.9	30.00
Dibenz(a,h)anthracene	0.553	ng/ul	10.0	30.00
Indeno(1,2,3-c,d)pyrene	0.547	ng/ul	9.1	30.00
Benzo(g,h,i)perylene	0.539	ng/ul	7.4	30.00
Coronene	0.554	ng/ul	10.4	30.00

Surrogate: Fluorene-d10	0.512	ng/ul	4.5	30.00
Surrogate: Fluoranthene-d10	0.558	ng/ul	9.5	30.00
Surrogate: Pyrene-d10	0.464	ng/ul	-7.1	30.00
Surrogate: Benzo(a)pyrene-d12	0.529	ng/ul	6.2	30.00

Sequence 2608010

Calibration Check (2608010-CCV1)

Prepared & Analyzed: 08/04/26

Naphthalene	7.31	ng/ul	-2.6	30.00	
Acenaphthylene	0.497	ng/ul	-1.1	30.00	
Acenaphthene	2.42	ng/ul	-3.2	30.00	
Fluorene	2.43	ng/ul	-3.0	30.00	
Phenanthrene	3.67	ng/ul	-2.1	30.00	LB
Anthracene	0.490	ng/ul	-2.4	30.00	
Fluoranthene	2.47	ng/ul	-1.1	30.00	LB
Pyrene	1.14	ng/ul	-9.3	30.00	
Benzo(a)anthracene	0.484	ng/ul	-3.5	30.00	
Chrysene	0.484	ng/ul	-3.6	30.00	
Benzo(b)fluoranthene	0.481	ng/ul	-4.1	30.00	
Benzo(k)fluoranthene	0.490	ng/ul	-2.2	30.00	
Benzo(e)pyrene	0.478	ng/ul	-4.6	30.00	
Benzo(a)pyrene	0.486	ng/ul	-3.3	30.00	
Perylene	0.482	ng/ul	-3.9	30.00	
Dibenz(a,h)anthracene	0.498	ng/ul	-0.9	30.00	
Indeno(1,2,3-c,d)pyrene	0.496	ng/ul	-1.1	30.00	
Benzo(g,h,i)perylene	0.489	ng/ul	-2.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2608010

Calibration Check (2608010-CCV1) Continued

Prepared & Analyzed: 08/04/26

Coronene	0.498	ng/ul	-0.8	30.00	
Surrogate: Fluorene-d10	0.477	ng/ul	-2.7	30.00	
Surrogate: Fluoranthene-d10	0.500	ng/ul	-2.0	30.00	
Surrogate: Pyrene-d10	0.454	ng/ul	-9.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.485	ng/ul	-2.7	30.00	

Calibration Check (2608010-CCV2)

Prepared: 08/04/26 Analyzed: 08/05/26

Naphthalene	7.35	ng/ul	-2.1	30.00	
Acenaphthylene	0.478	ng/ul	-4.7	30.00	
Acenaphthene	2.46	ng/ul	-1.5	30.00	
Fluorene	2.49	ng/ul	-0.4	30.00	
Phenanthrene	3.66	ng/ul	-2.6	30.00	LB
Anthracene	0.503	ng/ul	0.4	30.00	
Fluoranthene	2.61	ng/ul	4.4	30.00	LB
Pyrene	1.05	ng/ul	-15.9	30.00	
Benzo(a)anthracene	0.488	ng/ul	-2.7	30.00	
Chrysene	0.483	ng/ul	-3.7	30.00	
Benzo(b)fluoranthene	0.476	ng/ul	-5.1	30.00	
Benzo(k)fluoranthene	0.491	ng/ul	-2.1	30.00	
Benzo(e)pyrene	0.466	ng/ul	-7.1	30.00	
Benzo(a)pyrene	0.485	ng/ul	-3.4	30.00	
Perylene	0.478	ng/ul	-4.7	30.00	
Dibenz(a,h)anthracene	0.493	ng/ul	-1.9	30.00	
Indeno(1,2,3-c,d)pyrene	0.488	ng/ul	-2.6	30.00	
Benzo(g,h,i)perylene	0.479	ng/ul	-4.5	30.00	
Coronene	0.472	ng/ul	-5.9	30.00	
Surrogate: Fluorene-d10	0.480	ng/ul	-2.2	30.00	
Surrogate: Fluoranthene-d10	0.529	ng/ul	3.7	30.00	
Surrogate: Pyrene-d10	0.422	ng/ul	-15.4	30.00	
Surrogate: Benzo(a)pyrene-d12	0.489	ng/ul	-1.9	30.00	

Calibration Check (2608010-CCV3)

Prepared: 08/04/26 Analyzed: 08/05/26

Naphthalene	7.31	ng/ul	-2.6	30.00	
Acenaphthylene	0.479	ng/ul	-4.6	30.00	
Acenaphthene	2.46	ng/ul	-1.6	30.00	
Fluorene	2.48	ng/ul	-1.0	30.00	
Phenanthrene	3.66	ng/ul	-2.6	30.00	LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2608010

Calibration Check (2608010-CCV3) Continued

Prepared: 08/04/26 Analyzed: 08/05/26

Anthracene	0.510	ng/ul	1.8	30.00	LB
Fluoranthene	2.63	ng/ul	5.3	30.00	
Pyrene	1.04	ng/ul	-17.0	30.00	
Benzo(a)anthracene	0.487	ng/ul	-3.0	30.00	
Chrysene	0.481	ng/ul	-4.2	30.00	
Benzo(b)fluoranthene	0.473	ng/ul	-5.6	30.00	
Benzo(k)fluoranthene	0.491	ng/ul	-2.1	30.00	
Benzo(e)pyrene	0.466	ng/ul	-7.0	30.00	
Benzo(a)pyrene	0.487	ng/ul	-3.0	30.00	
Perylene	0.478	ng/ul	-4.6	30.00	
Dibenz(a,h)anthracene	0.498	ng/ul	-1.0	30.00	
Indeno(1,2,3-c,d)pyrene	0.494	ng/ul	-1.6	30.00	
Benzo(g,h,i)perylene	0.480	ng/ul	-4.4	30.00	
Coronene	0.470	ng/ul	-6.3	30.00	
Surrogate: Fluorene-d10	0.480	ng/ul	-2.1	30.00	
Surrogate: Fluoranthene-d10	0.535	ng/ul	5.0	30.00	
Surrogate: Pyrene-d10	0.417	ng/ul	-16.6	30.00	
Surrogate: Benzo(a)pyrene-d12	0.491	ng/ul	-1.6	30.00	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/17/26 16:07

SUBMITTED: 06/11/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
S-GC	Surrogate recovery outside of control limits. The data was accepted based on valid recovery of the remaining surrogate.
QB-03	Analyte exceeds sampling media lot blank criteria.
LL	Analyte identified; Reported value may be biased low.
LB	Lab blank value above acceptable limit.
INV-LB	Lot Blank for this sample is invalid.
FX	Filter Integrity Issue
FB-01	Analyte exceeds Field Blank criteria.
D	This result obtained by dilution.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Matrix Spikes were not performed as required by Compendium Method TO-13A. Matrix spikes are not required by ASTM D2609. This deviation from the method has been approved by the US EPA under the contract for UATMP, NATTS, CSATAM, PAMS, and NMOC Support (US EPA Contract No. 68-D-09-048).

Note: This test is accredited under the 2016 TNI Standard; however acenaphthylene is not accredited.