



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

August 12, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 06/11/26 11:56 through 07/07/26 09:45.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6061711-01	Air	06/01/26 23:59	06/17/26 13:43
TMOK	6062316-01	Air	06/13/26 23:59	06/23/26 11:42
TMOK D1	6062426-02	Air	06/19/26 23:59	06/24/26 13:06
TMOK D2	6062426-03	Air	06/19/26 23:59	06/24/26 13:06
TMOK	6070729-05	Air	06/25/26 23:59	07/07/26 09:45



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6061711-01

Sampled: 06/01/26 23:59

Pressure @ Receipt: 3.0

Canister #: 41551

Received: 06/17/26 13:43

Comments:

Analyzed: 06/18/26 17:46

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.460	2.28		0.0237
Chloromethane	0.590	1.22		0.0213
Dichlorotetrafluoroethane	0.0161	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00730	0.02	CF, LL, U	0.0215
Bromomethane	0.00840	0.03	U	0.0104
Chloroethane	0.00990	0.03	U	0.0136
Acetonitrile	184	310.00	E, LJ, LK	0.184
Acrolein	0.300	0.69	CF, LK	0.0977
Trichlorofluoromethane	0.200	1.13		0.00950
Acrylonitrile	0.0190	0.04		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	2.92	10.20		0.0489
Carbon Disulfide	0.0964	0.30		0.0473
Trichlorotrifluoroethane	0.0645	0.50		0.0210
trans-1,2-Dichloroethylene	0.00750	0.03	U	0.0171
1,1-Dichloroethane	0.00110	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0232	0.11		0.0134
Ethyl tert-Butyl Ether	0.00250	0.01	U	0.00893
1,2-Dichloroethane	0.0155	0.06		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.165	0.53		0.0158
Carbon Tetrachloride	0.0680	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	LK, U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00350	0.02	U	0.0170
Methyl Methacrylate	ND	ND	LK, U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0327	0.13	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6061711-01

Sampled: 06/01/26 23:59

Pressure @ Receipt: 3.0

Canister #: 41551

Received: 06/17/26 13:43

Comments:

Analyzed: 06/18/26 17:46

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	1.92	7.25		0.0380
Dibromochloromethane	5.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.109	0.51		0.0258
Tetrachloroethylene	0.00960	0.07		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.133	0.58		0.0130
m,p-Xylene	0.343	1.49		0.0418
Bromoform	0.00100	0.01	U	0.0166
Styrene	0.0643	0.27		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U, LK	0.0190
o-Xylene	0.0730	0.32		0.0235
1,3,5-Trimethylbenzene	0.00670	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	5.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00430	0.03	U	0.0287
o-Dichlorobenzene	7.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00190	0.01	U	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6062316-01

Sampled: 06/13/26 23:59

Pressure @ Receipt: 4.0

Canister #: 18830

Received: 06/23/26 11:42

Comments:

Analyzed: 06/29/26 23:16

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.447	2.22		0.0237
Chloromethane	0.526	1.09		0.0213
Dichlorotetrafluoroethane	0.0161	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00730	0.02	U	0.0215
Bromomethane	0.00800	0.03	U	0.0104
Chloroethane	0.00540	0.01	U	0.0136
Acetonitrile	285	479.00	E	0.184
Acrolein	0.292	0.67	CF, LK	0.0977
Trichlorofluoromethane	0.228	1.28		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.260	0.91		0.0489
Carbon Disulfide	0.0170	0.05	U	0.0473
Trichlorotrifluoroethane	0.0594	0.46		0.0210
trans-1,2-Dichloroethylene	0.00170	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	1.00E-4	0.00	U	0.00905
Chloroform	0.0181	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0100	0.04	U	0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.135	0.43		0.0158
Carbon Tetrachloride	0.0681	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00380	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0144	0.06	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6062316-01

Sampled: 06/13/26 23:59

Pressure @ Receipt: 4.0

Canister #: 18830

Received: 06/23/26 11:42

Comments:

Analyzed: 06/29/26 23:16

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.556	2.10		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0556	0.26		0.0258
Tetrachloroethylene	0.00660	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0557	0.24		0.0130
m,p-Xylene	0.180	0.78		0.0418
Bromoform	0.00180	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0706	0.31		0.0235
1,3,5-Trimethylbenzene	0.0171	0.08	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00250	0.02	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1

Lab ID: 6062426-02

Sampled: 06/19/26 23:59

Pressure @ Receipt: 4.50

Canister #: 18884

Received: 06/24/26 13:06

Comments: Duplicate Makeup from 3-15-26

Analyzed: 06/30/26 21:11

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.459	2.27		0.0237
Chloromethane	0.478	0.99		0.0213
Dichlorotetrafluoroethane	0.0166	0.12		0.0112
Vinyl chloride	4.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0126	0.03	U	0.0215
Bromomethane	0.00850	0.03	U	0.0104
Chloroethane	0.00630	0.02	U	0.0136
Acetonitrile	233	392.00	D-F, E	0.184
Acrolein	0.187	0.43		0.0977
Trichlorofluoromethane	0.211	1.19		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.183	0.64	CF, LK	0.0489
Carbon Disulfide	0.0654	0.20	QB-03	0.0473
Trichlorotrifluoroethane	0.0622	0.48		0.0210
trans-1,2-Dichloroethylene	0.00310	0.01	U	0.0171
1,1-Dichloroethane	0.00100	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	1.00E-4	0.00	U	0.00905
Chloroform	0.0194	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0143	0.06		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.102	0.33	D-F	0.0158
Carbon Tetrachloride	0.0712	0.45		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00120	0.01	U	0.00889
Trichloroethylene	0.00500	0.03	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0120	0.05	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1

Lab ID: 6062426-02

Sampled: 06/19/26 23:59

Pressure @ Receipt: 4.50

Canister #: 18884

Received: 06/24/26 13:06

Comments: Duplicate Makeup from 3-15-26

Analyzed: 06/30/26 21:11

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.201	0.76	D-F	0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0144	0.07	U	0.0258
Tetrachloroethylene	0.00630	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0275	0.12		0.0130
m,p-Xylene	0.0778	0.34		0.0418
Bromoform	7.00E-4	0.01	U	0.0166
Styrene	0.0130	0.06	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0304	0.13		0.0235
1,3,5-Trimethylbenzene	0.00650	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00170	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	3.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2

Lab ID: 6062426-03

Sampled: 06/19/26 23:59

Pressure @ Receipt: 4.00

Canister #: 33512

Received: 06/24/26 13:06

Comments: Duplicate Makeup from 3-15-26

Analyzed: 06/30/26 23:13

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.465	2.30		0.0237
Chloromethane	0.500	1.03		0.0213
Dichlorotetrafluoroethane	0.0170	0.12		0.0112
Vinyl chloride	5.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00820	0.02	CF, LL, U	0.0215
Bromomethane	0.00850	0.03	U	0.0104
Chloroethane	0.00690	0.02	U	0.0136
Acetonitrile	46.4	78.10	D-F, E	0.184
Acrolein	0.221	0.51	CF, LK	0.0977
Trichlorofluoromethane	0.220	1.24		0.00950
Acrylonitrile	0.0185	0.04		0.0116
1,1-Dichloroethene	1.00E-4	0.00	U	0.0158
Dichloromethane	0.205	0.71		0.0489
Carbon Disulfide	0.0144	0.04	U	0.0473
Trichlorotrifluoroethane	0.0622	0.48		0.0210
trans-1,2-Dichloroethylene	0.00640	0.03	U	0.0171
1,1-Dichloroethane	0.00110	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0307	0.15		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0154	0.06		0.0116
1,1,1-Trichloroethane	0.00130	0.01	U	0.00952
Benzene	0.156	0.50	D-F	0.0158
Carbon Tetrachloride	0.0701	0.44		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00100	0.01	U	0.00889
Trichloroethylene	0.00410	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0242	0.10	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2

Lab ID: 6062426-03

Sampled: 06/19/26 23:59

Pressure @ Receipt: 4.00

Canister #: 33512

Received: 06/24/26 13:06

Comments: Duplicate Makeup from 3-15-26

Analyzed: 06/30/26 23:13

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.636	2.40	D-F	0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0249	0.12	U	0.0258
Tetrachloroethylene	0.0171	0.12		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0499	0.22		0.0130
m,p-Xylene	0.141	0.61		0.0418
Bromoform	6.00E-4	0.01	U	0.0166
Styrene	0.0402	0.17		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0540	0.24		0.0235
1,3,5-Trimethylbenzene	0.00960	0.05	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00240	0.01	U	0.0287
o-Dichlorobenzene	ND	ND	U	0.0364
1,2,4-Trichlorobenzene	9.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6070729-05

Sampled: 06/25/26 23:59

Pressure @ Receipt: 4.5

Canister #: 40468

Received: 07/07/26 09:45

Comments:

Analyzed: 07/14/26 00:16

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.425	2.11		0.0237
Chloromethane	0.606	1.25		0.0213
Dichlorotetrafluoroethane	0.0165	0.12		0.0112
Vinyl chloride	8.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0118	0.03	CF, LL, U	0.0215
Bromomethane	0.0104	0.04		0.0104
Chloroethane	0.00850	0.02	U	0.0136
Acetonitrile	218	367.00	E	0.184
Acrolein	0.225	0.52	CF, LK	0.0977
Trichlorofluoromethane	0.171	0.96		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.189	0.66		0.0489
Carbon Disulfide	0.0117	0.04	U	0.0473
Trichlorotrifluoroethane	0.0544	0.42		0.0210
trans-1,2-Dichloroethylene	0.00600	0.02	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	5.00E-4	0.00	U	0.00905
Chloroform	0.0172	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.00730	0.03	U	0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.181	0.58		0.0158
Carbon Tetrachloride	0.0621	0.39		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00130	0.01	U	0.00889
Trichloroethylene	0.00460	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0203	0.08	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6070729-05

Sampled: 06/25/26 23:59

Pressure @ Receipt: 4.5

Canister #: 40468

Received: 07/07/26 09:45

Comments:

Analyzed: 07/14/26 00:16

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.382	1.44		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0391	0.18		0.0258
Tetrachloroethylene	0.0102	0.07		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0580	0.25		0.0130
m,p-Xylene	0.159	0.69		0.0418
Bromoform	0.00160	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0613	0.27		0.0235
1,3,5-Trimethylbenzene	0.00940	0.05	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	1.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00280	0.02	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F1802 - Summa Canister Prep

Blank (B6F1802-BLK1)

Prepared: 06/05/26 Analyzed: 06/18/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				LK, U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				LK, U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				LK, U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F1802 - Summa Canister Prep

Blank (B6F1802-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/18/26

1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				LK, U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6F2902 - Summa Canister Prep

Blank (B6F2902-BLK1)

Prepared: 06/25/26 Analyzed: 06/29/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F2902 - Summa Canister Prep

Blank (B6F2902-BLK1) Continued

Prepared: 06/25/26 Analyzed: 06/29/26

Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6F3001 - Summa Canister Prep

Blank (B6F3001-BLK1)

Prepared: 06/25/26 Analyzed: 06/30/26

Dichlorodifluoromethane	ND	ppbv				U
-------------------------	----	------	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F3001 - Summa Canister Prep

Blank (B6F3001-BLK1) Continued

Prepared: 06/25/26 Analyzed: 06/30/26

Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F3001 - Summa Canister Prep

Blank (B6F3001-BLK1) Continued

Prepared: 06/25/26 Analyzed: 06/30/26

n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Duplicate (B6F3001-DUP1)

Source: 6062426-02

Prepared: 06/19/26 Analyzed: 06/30/26

Dichlorodifluoromethane	0.454	ppbv	0.46	1.16	25	
Chloromethane	0.472	ppbv	0.48	1.35	25	
Dichlorotetrafluoroethane	0.0164	ppbv	0.02	1.21	25	
Vinyl chloride	ND	ppbv	ND		25	U
1,3-Butadiene	ND	ppbv	ND		25	U
Bromomethane	ND	ppbv	ND		25	U
Chloroethane	ND	ppbv	ND		25	U
Acetonitrile	227	ppbv	233.00	2.45	25	E
Acrolein	0.185	ppbv	0.19	0.914	25	
Trichlorofluoromethane	0.205	ppbv	0.21	2.69	25	
Acrylonitrile	ND	ppbv	ND		25	CE, U
1,1-Dichloroethene	ND	ppbv	ND		25	U
Dichloromethane	0.180	ppbv	0.18	1.88	25	CF, LK
Carbon Disulfide	0.0636	ppbv	0.07	2.79	25	QB-03
Trichlorotrifluoroethane	0.0609	ppbv	0.06	2.11	25	
trans-1,2-Dichloroethylene	ND	ppbv	ND		25	U
1,1-Dichloroethane	ND	ppbv	ND		25	U
Methyl tert-Butyl Ether	ND	ppbv	ND		25	U
Chloroprene	ND	ppbv	ND		25	U
cis-1,2-Dichloroethylene	ND	ppbv	ND		25	U
Bromochloromethane	ND	ppbv	ND		25	U
Chloroform	0.0189	ppbv	0.02	2.61	25	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F3001 - Summa Canister Prep

Duplicate (B6F3001-DUP1) Continued Source: 6062426-02 Prepared: 06/19/26 Analyzed: 06/30/26

Ethyl tert-Butyl Ether	ND	ppbv	ND		25	U
1,2-Dichloroethane	0.0139	ppbv	0.01	2.84	25	
1,1,1-Trichloroethane	ND	ppbv	ND		25	U
Benzene	0.0990	ppbv	0.10	2.89	25	
Carbon Tetrachloride	0.0691	ppbv	0.07	2.99	25	
tert-Amyl Methyl Ether	ND	ppbv	ND		25	U
1,2-Dichloropropane	ND	ppbv	ND		25	U
Ethyl Acrylate	ND	ppbv	ND		25	U
Bromodichloromethane	ND	ppbv	ND		25	U
Trichloroethylene	ND	ppbv	ND		25	U
Methyl Methacrylate	ND	ppbv	ND		25	U
cis-1,3-Dichloropropene	ND	ppbv	ND		25	U
Methyl Isobutyl Ketone	ND	ppbv	ND		25	U
trans-1,3-Dichloropropene	ND	ppbv	ND		25	U
1,1,2-Trichloroethane	ND	ppbv	ND		25	U
Toluene	0.211	ppbv	0.20	4.86	25	
Dibromochloromethane	ND	ppbv	ND		25	U
1,2-Dibromoethane	ND	ppbv	ND		25	U
n-Octane	ND	ppbv	ND		25	U
Tetrachloroethylene	ND	ppbv	ND		25	U
Chlorobenzene	ND	ppbv	ND		25	U
Ethylbenzene	0.0270	ppbv	0.03	1.83	25	
m,p-Xylene	0.0761	ppbv	0.08	2.21	25	
Bromoform	ND	ppbv	ND		25	U
Styrene	ND	ppbv	ND		25	U
1,1,2,2-Tetrachloroethane	ND	ppbv	ND		25	U
o-Xylene	0.0295	ppbv	0.03	3.01	25	
1,3,5-Trimethylbenzene	ND	ppbv	ND		25	U
1,2,4-Trimethylbenzene	ND	ppbv	ND		25	CE, U
m-Dichlorobenzene	ND	ppbv	ND		25	U
p-Dichlorobenzene	ND	ppbv	ND		25	U
o-Dichlorobenzene	ND	ppbv	ND		25	U
1,2,4-Trichlorobenzene	ND	ppbv	ND		25	U
Hexachloro-1,3-butadiene	ND	ppbv	ND		25	U

Duplicate (B6F3001-DUP2) Source: 6062426-03 Prepared: 06/19/26 Analyzed: 07/01/26

Dichlorodifluoromethane	0.470	ppbv	0.47	1.03	25	
Chloromethane	0.508	ppbv	0.50	1.65	25	
Dichlorotetrafluoroethane	0.0167	ppbv	0.02	1.78	25	
Vinyl chloride	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F3001 - Summa Canister Prep

Duplicate (B6F3001-DUP2) Continued

Source: 6062426-03

Prepared: 06/19/26 Analyzed: 07/01/26

1,3-Butadiene	ND	ppbv	ND		25	CF, LL, U
Bromomethane	ND	ppbv	ND		25	U
Chloroethane	ND	ppbv	ND		25	U
Acetonitrile	46.5	ppbv	46.40	0.169	25	E
Acrolein	0.225	ppbv	0.22	1.48	25	CF, LK
Trichlorofluoromethane	0.222	ppbv	0.22	0.543	25	
Acrylonitrile	0.0195	ppbv	0.02	5.26	25	
1,1-Dichloroethene	ND	ppbv	ND		25	U
Dichloromethane	0.210	ppbv	0.21	2.41	25	
Carbon Disulfide	ND	ppbv	ND		25	U
Trichlorotrifluoroethane	0.0624	ppbv	0.06	0.321	25	
trans-1,2-Dichloroethylene	ND	ppbv	ND		25	U
1,1-Dichloroethane	ND	ppbv	ND		25	U
Methyl tert-Butyl Ether	ND	ppbv	ND		25	U
Chloroprene	ND	ppbv	ND		25	U
cis-1,2-Dichloroethylene	ND	ppbv	ND		25	U
Bromochloromethane	ND	ppbv	ND		25	U
Chloroform	0.0307	ppbv	0.03	0.00	25	
Ethyl tert-Butyl Ether	ND	ppbv	ND		25	U
1,2-Dichloroethane	0.0154	ppbv	0.02	0.00	25	
1,1,1-Trichloroethane	ND	ppbv	ND		25	U
Benzene	0.156	ppbv	0.16	0.00	25	
Carbon Tetrachloride	0.0703	ppbv	0.07	0.285	25	
tert-Amyl Methyl Ether	ND	ppbv	ND		25	U
1,2-Dichloropropane	ND	ppbv	ND		25	U
Ethyl Acrylate	ND	ppbv	ND		25	U
Bromodichloromethane	ND	ppbv	ND		25	U
Trichloroethylene	ND	ppbv	ND		25	U
Methyl Methacrylate	ND	ppbv	ND		25	U
cis-1,3-Dichloropropene	ND	ppbv	ND		25	U
Methyl Isobutyl Ketone	ND	ppbv	ND		25	U
trans-1,3-Dichloropropene	ND	ppbv	ND		25	U
1,1,2-Trichloroethane	ND	ppbv	ND		25	U
Toluene	0.639	ppbv	0.64	0.455	25	
Dibromochloromethane	ND	ppbv	ND		25	U
1,2-Dibromoethane	ND	ppbv	ND		25	U
n-Octane	ND	ppbv	ND		25	U
Tetrachloroethylene	0.0171	ppbv	0.02	0.00	25	
Chlorobenzene	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F3001 - Summa Canister Prep

Duplicate (B6F3001-DUP2) Continued Source: 6062426-03 Prepared: 06/19/26 Analyzed: 07/01/26

Ethylbenzene	0.0500	ppbv	0.05	0.200	25	
m,p-Xylene	0.142	ppbv	0.14	0.212	25	
Bromoform	ND	ppbv	ND		25	U
Styrene	0.0403	ppbv	0.04	0.248	25	
1,1,2,2-Tetrachloroethane	ND	ppbv	ND		25	U
o-Xylene	0.0539	ppbv	0.05	0.185	25	
1,3,5-Trimethylbenzene	ND	ppbv	ND		25	U
1,2,4-Trimethylbenzene	ND	ppbv	ND		25	CE, U
m-Dichlorobenzene	ND	ppbv	ND		25	U
p-Dichlorobenzene	ND	ppbv	ND		25	U
o-Dichlorobenzene	ND	ppbv	ND		25	U
1,2,4-Trichlorobenzene	ND	ppbv	ND		25	U
Hexachloro-1,3-butadiene	ND	ppbv	ND		25	U

Batch B6G1302 - Summa Canister Prep

Blank (B6G1302-BLK1) Prepared & Analyzed: 07/13/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6G1302 - Summa Canister Prep

Blank (B6G1302-BLK1) Continued

Prepared & Analyzed: 07/13/26

1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606071

Calibration Check (2606071-CCV1)

Prepared & Analyzed: 06/18/26

Dichlorodifluoromethane	1.05	ppbv	2.5	30.00
Chloromethane	1.17	ppbv	15.0	30.00
Dichlorotetrafluoroethane	1.08	ppbv	4.3	30.00
Vinyl chloride	1.09	ppbv	5.6	30.00
1,3-Butadiene	1.08	ppbv	4.1	30.00
Bromomethane	1.03	ppbv	-0.02	30.00
Chloroethane	1.17	ppbv	12.5	30.00
Acetonitrile	1.31	ppbv	25.9	30.00
Acrolein	1.14	ppbv	11.3	30.00
Trichlorofluoromethane	1.05	ppbv	3.5	30.00
Acrylonitrile	1.28	ppbv	16.4	30.00
1,1-Dichloroethene	1.17	ppbv	12.6	30.00
Dichloromethane	1.23	ppbv	19.2	30.00
Carbon Disulfide	1.15	ppbv	6.1	30.00
Trichlorotrifluoroethane	1.05	ppbv	1.8	30.00
trans-1,2-Dichloroethylene	1.06	ppbv	1.5	30.00
1,1-Dichloroethane	1.19	ppbv	15.0	30.00
Methyl tert-Butyl Ether	1.08	ppbv	4.3	30.00
Chloroprene	1.15	ppbv	11.2	30.00
cis-1,2-Dichloroethylene	1.18	ppbv	14.2	30.00
Bromochloromethane	1.02	ppbv	-0.8	30.00
Chloroform	1.15	ppbv	10.0	30.00
Ethyl tert-Butyl Ether	1.14	ppbv	10.4	30.00
1,2-Dichloroethane	1.18	ppbv	14.5	30.00
1,1,1-Trichloroethane	1.10	ppbv	6.4	30.00
Benzene	1.13	ppbv	8.6	30.00
Carbon Tetrachloride	1.05	ppbv	2.5	30.00
tert-Amyl Methyl Ether	1.09	ppbv	5.5	30.00
1,2-Dichloropropane	1.30	ppbv	24.8	30.00
Ethyl Acrylate	1.29	ppbv	23.0	30.00
Bromodichloromethane	1.21	ppbv	17.2	30.00
Trichloroethylene	1.09	ppbv	5.3	30.00
Methyl Methacrylate	1.31	ppbv	26.0	30.00
cis-1,3-Dichloropropene	1.26	ppbv	21.1	30.00
Methyl Isobutyl Ketone	1.26	ppbv	21.3	30.00
trans-1,3-Dichloropropene	1.23	ppbv	17.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606071

Calibration Check (2606071-CCV1) Continued

Prepared & Analyzed: 06/18/26

1,1,2-Trichloroethane	1.16	ppbv	11.0	30.00
Toluene	1.21	ppbv	16.0	30.00
Dibromochloromethane	1.12	ppbv	8.1	30.00
1,2-Dibromoethane	1.14	ppbv	8.8	30.00
n-Octane	1.29	ppbv	24.0	30.00
Tetrachloroethylene	1.07	ppbv	2.1	30.00
Chlorobenzene	1.17	ppbv	11.0	30.00
Ethylbenzene	1.15	ppbv	10.2	30.00
m,p-Xylene	2.44	ppbv	16.9	30.00
Bromoform	1.13	ppbv	8.5	30.00
Styrene	1.19	ppbv	12.1	30.00
1,1,2,2-Tetrachloroethane	1.29	ppbv	22.8	30.00
o-Xylene	1.16	ppbv	10.2	30.00
1,3,5-Trimethylbenzene	1.07	ppbv	2.0	30.00
1,2,4-Trimethylbenzene	1.08	ppbv	2.3	30.00
m-Dichlorobenzene	1.10	ppbv	5.8	30.00
p-Dichlorobenzene	1.09	ppbv	4.1	30.00
o-Dichlorobenzene	1.10	ppbv	5.0	30.00
1,2,4-Trichlorobenzene	0.967	ppbv	-6.8	30.00
Hexachloro-1,3-butadiene	1.01	ppbv	-4.0	30.00

Calibration Check (2606071-CCV2)

Prepared: 06/18/26 Analyzed: 06/19/26

Dichlorodifluoromethane	1.10	ppbv	7.3	30.00
Chloromethane	1.23	ppbv	21.0	30.00
Dichlorotetrafluoroethane	1.09	ppbv	5.5	30.00
Vinyl chloride	1.09	ppbv	5.9	30.00
1,3-Butadiene	1.00	ppbv	-2.9	30.00
Bromomethane	1.08	ppbv	4.6	30.00
Chloroethane	1.21	ppbv	16.3	30.00
Acetonitrile	1.43	ppbv	37.6	30.00
Acrolein	1.06	ppbv	4.4	30.00
Trichlorofluoromethane	1.07	ppbv	5.8	30.00
Acrylonitrile	1.27	ppbv	14.8	30.00
1,1-Dichloroethene	1.18	ppbv	13.0	30.00
Dichloromethane	1.28	ppbv	23.6	30.00
Carbon Disulfide	1.20	ppbv	10.1	30.00

LJ, LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606071

Calibration Check (2606071-CCV2) Continued

Prepared: 06/18/26 Analyzed: 06/19/26

Trichlorotrifluoroethane	1.06	ppbv	2.6	30.00	
trans-1,2-Dichloroethylene	1.06	ppbv	2.1	30.00	
1,1-Dichloroethane	1.21	ppbv	17.0	30.00	
Methyl tert-Butyl Ether	1.05	ppbv	1.8	30.00	
Chloroprene	1.13	ppbv	8.5	30.00	
cis-1,2-Dichloroethylene	1.19	ppbv	15.3	30.00	
Bromochloromethane	1.01	ppbv	-1.6	30.00	
Chloroform	1.16	ppbv	10.8	30.00	
Ethyl tert-Butyl Ether	1.10	ppbv	6.6	30.00	
1,2-Dichloroethane	1.19	ppbv	15.3	30.00	
1,1,1-Trichloroethane	1.09	ppbv	6.1	30.00	
Benzene	1.12	ppbv	8.0	30.00	
Carbon Tetrachloride	1.05	ppbv	2.4	30.00	
tert-Amyl Methyl Ether	1.05	ppbv	1.9	30.00	
1,2-Dichloropropane	1.35	ppbv	30.1	30.00	LK
Ethyl Acrylate	1.32	ppbv	25.9	30.00	
Bromodichloromethane	1.28	ppbv	23.3	30.00	
Trichloroethylene	1.10	ppbv	6.2	30.00	
Methyl Methacrylate	1.36	ppbv	31.0	30.00	LK
cis-1,3-Dichloropropene	1.30	ppbv	24.3	30.00	
Methyl Isobutyl Ketone	1.33	ppbv	28.1	30.00	
trans-1,3-Dichloropropene	1.29	ppbv	22.8	30.00	
1,1,2-Trichloroethane	1.24	ppbv	18.5	30.00	
Toluene	1.24	ppbv	19.2	30.00	
Dibromochloromethane	1.11	ppbv	7.2	30.00	
1,2-Dibromoethane	1.12	ppbv	7.3	30.00	
n-Octane	1.35	ppbv	29.8	30.00	
Tetrachloroethylene	1.00	ppbv	-4.0	30.00	
Chlorobenzene	1.20	ppbv	14.2	30.00	
Ethylbenzene	1.15	ppbv	9.7	30.00	
m,p-Xylene	2.48	ppbv	18.7	30.00	
Bromoform	1.14	ppbv	9.0	30.00	
Styrene	1.19	ppbv	12.4	30.00	
1,1,2,2-Tetrachloroethane	1.41	ppbv	33.7	30.00	LK
o-Xylene	1.15	ppbv	8.7	30.00	
1,3,5-Trimethylbenzene	1.03	ppbv	-2.2	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606071

Calibration Check (2606071-CCV2) Continued

Prepared: 06/18/26 Analyzed: 06/19/26

1,2,4-Trimethylbenzene	1.04	ppbv	-1.2	30.00
m-Dichlorobenzene	1.09	ppbv	4.4	30.00
p-Dichlorobenzene	1.06	ppbv	1.8	30.00
o-Dichlorobenzene	1.07	ppbv	2.7	30.00
1,2,4-Trichlorobenzene	0.899	ppbv	-13.4	30.00
Hexachloro-1,3-butadiene	1.00	ppbv	-4.7	30.00

Sequence 2606111

Calibration Check (2606111-CCV1)

Prepared & Analyzed: 06/29/26

Dichlorodifluoromethane	1.07	ppbv	4.2	30.00
Chloromethane	1.04	ppbv	1.8	30.00
Dichlorotetrafluoroethane	1.05	ppbv	1.5	30.00
Vinyl chloride	1.04	ppbv	1.1	30.00
1,3-Butadiene	1.04	ppbv	0.2	30.00
Bromomethane	1.08	ppbv	4.6	30.00
Chloroethane	1.09	ppbv	5.2	30.00
Acetonitrile	1.12	ppbv	8.0	30.00
Acrolein	1.05	ppbv	2.7	30.00
Trichlorofluoromethane	1.10	ppbv	8.6	30.00
Acrylonitrile	1.14	ppbv	3.5	30.00
1,1-Dichloroethene	1.03	ppbv	-0.9	30.00
Dichloromethane	1.10	ppbv	6.5	30.00
Carbon Disulfide	1.00	ppbv	-8.0	30.00
Trichlorotrifluoroethane	0.990	ppbv	-3.9	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.8	30.00
1,1-Dichloroethane	1.09	ppbv	4.8	30.00
Methyl tert-Butyl Ether	1.12	ppbv	8.1	30.00
Chloroprene	1.05	ppbv	1.4	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	4.3	30.00
Bromochloromethane	1.08	ppbv	4.9	30.00
Chloroform	1.12	ppbv	7.0	30.00
Ethyl tert-Butyl Ether	1.09	ppbv	5.4	30.00
1,2-Dichloroethane	1.11	ppbv	7.5	30.00
1,1,1-Trichloroethane	1.12	ppbv	8.6	30.00
Benzene	1.07	ppbv	2.9	30.00
Carbon Tetrachloride	1.11	ppbv	8.6	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606111

Calibration Check (2606111-CCV1) Continued

Prepared & Analyzed: 06/29/26

tert-Amyl Methyl Ether	1.08	ppbv	4.8	30.00
1,2-Dichloropropane	1.17	ppbv	13.1	30.00
Ethyl Acrylate	1.06	ppbv	1.4	30.00
Bromodichloromethane	1.11	ppbv	7.1	30.00
Trichloroethylene	1.08	ppbv	4.2	30.00
Methyl Methacrylate	1.08	ppbv	3.6	30.00
cis-1,3-Dichloropropene	1.05	ppbv	1.0	30.00
Methyl Isobutyl Ketone	1.02	ppbv	-2.3	30.00
trans-1,3-Dichloropropene	1.13	ppbv	8.0	30.00
1,1,2-Trichloroethane	1.09	ppbv	4.4	30.00
Toluene	1.09	ppbv	4.9	30.00
Dibromochloromethane	1.10	ppbv	6.4	30.00
1,2-Dibromoethane	1.07	ppbv	2.7	30.00
n-Octane	1.12	ppbv	7.7	30.00
Tetrachloroethylene	1.11	ppbv	6.4	30.00
Chlorobenzene	1.11	ppbv	5.1	30.00
Ethylbenzene	1.09	ppbv	4.4	30.00
m,p-Xylene	2.34	ppbv	12.2	30.00
Bromoform	1.16	ppbv	10.6	30.00
Styrene	1.21	ppbv	13.9	30.00
1,1,2,2-Tetrachloroethane	1.14	ppbv	8.4	30.00
o-Xylene	1.13	ppbv	7.3	30.00
1,3,5-Trimethylbenzene	1.07	ppbv	1.9	30.00
1,2,4-Trimethylbenzene	1.09	ppbv	3.6	30.00
m-Dichlorobenzene	1.13	ppbv	8.6	30.00
p-Dichlorobenzene	1.14	ppbv	9.0	30.00
o-Dichlorobenzene	1.12	ppbv	7.1	30.00
1,2,4-Trichlorobenzene	1.03	ppbv	-0.9	30.00
Hexachloro-1,3-butadiene	1.09	ppbv	3.4	30.00

Calibration Check (2606111-CCV2)

Prepared: 06/29/26 Analyzed: 06/30/26

Dichlorodifluoromethane	1.08	ppbv	4.7	30.00
Chloromethane	1.07	ppbv	4.6	30.00
Dichlorotetrafluoroethane	1.07	ppbv	3.4	30.00
Vinyl chloride	1.01	ppbv	-2.1	30.00
1,3-Butadiene	0.942	ppbv	-8.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606111

Calibration Check (2606111-CCV2) Continued

Prepared: 06/29/26 Analyzed: 06/30/26

Bromomethane	1.07	ppbv	3.6	30.00
Chloroethane	1.12	ppbv	8.5	30.00
Acetonitrile	1.16	ppbv	11.3	30.00
Acrolein	1.10	ppbv	8.0	30.00
Trichlorofluoromethane	1.07	ppbv	5.6	30.00
Acrylonitrile	1.13	ppbv	2.3	30.00
1,1-Dichloroethene	1.06	ppbv	1.5	30.00
Dichloromethane	1.12	ppbv	8.5	30.00
Carbon Disulfide	1.01	ppbv	-7.0	30.00
Trichlorotrifluoroethane	0.982	ppbv	-4.6	30.00
trans-1,2-Dichloroethylene	1.02	ppbv	-1.5	30.00
1,1-Dichloroethane	1.09	ppbv	5.5	30.00
Methyl tert-Butyl Ether	1.03	ppbv	-0.1	30.00
Chloroprene	1.00	ppbv	-3.5	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	3.8	30.00
Bromochloromethane	1.05	ppbv	2.3	30.00
Chloroform	1.11	ppbv	5.5	30.00
Ethyl tert-Butyl Ether	1.02	ppbv	-1.7	30.00
1,2-Dichloroethane	1.10	ppbv	6.1	30.00
1,1,1-Trichloroethane	1.07	ppbv	3.6	30.00
Benzene	1.05	ppbv	0.7	30.00
Carbon Tetrachloride	1.05	ppbv	2.5	30.00
tert-Amyl Methyl Ether	1.01	ppbv	-2.2	30.00
1,2-Dichloropropane	1.22	ppbv	17.1	30.00
Ethyl Acrylate	1.07	ppbv	2.8	30.00
Bromodichloromethane	1.13	ppbv	8.9	30.00
Trichloroethylene	1.05	ppbv	1.3	30.00
Methyl Methacrylate	1.07	ppbv	3.0	30.00
cis-1,3-Dichloropropene	1.04	ppbv	-0.5	30.00
Methyl Isobutyl Ketone	1.02	ppbv	-1.7	30.00
trans-1,3-Dichloropropene	1.14	ppbv	9.0	30.00
1,1,2-Trichloroethane	1.11	ppbv	6.3	30.00
Toluene	1.08	ppbv	4.3	30.00
Dibromochloromethane	1.09	ppbv	5.3	30.00
1,2-Dibromoethane	1.06	ppbv	1.3	30.00
n-Octane	1.16	ppbv	11.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606111

Calibration Check (2606111-CCV2) Continued

Prepared: 06/29/26 Analyzed: 06/30/26

Tetrachloroethylene	1.05	ppbv	0.4	30.00
Chlorobenzene	1.10	ppbv	4.3	30.00
Ethylbenzene	1.07	ppbv	2.5	30.00
m,p-Xylene	2.35	ppbv	12.4	30.00
Bromoform	1.11	ppbv	6.5	30.00
Styrene	1.20	ppbv	13.1	30.00
1,1,2,2-Tetrachloroethane	1.20	ppbv	13.7	30.00
o-Xylene	1.11	ppbv	5.2	30.00
1,3,5-Trimethylbenzene	1.02	ppbv	-3.0	30.00
1,2,4-Trimethylbenzene	1.03	ppbv	-1.9	30.00
m-Dichlorobenzene	1.09	ppbv	4.9	30.00
p-Dichlorobenzene	1.09	ppbv	4.5	30.00
o-Dichlorobenzene	1.07	ppbv	2.2	30.00
1,2,4-Trichlorobenzene	0.944	ppbv	-9.1	30.00
Hexachloro-1,3-butadiene	1.04	ppbv	-1.6	30.00

Sequence 2606120

Calibration Check (2606120-CCV1)

Prepared & Analyzed: 06/30/26

Dichlorodifluoromethane	1.06	ppbv	3.4	30.00
Chloromethane	1.06	ppbv	3.5	30.00
Dichlorotetrafluoroethane	1.06	ppbv	2.1	30.00
Vinyl chloride	1.00	ppbv	-2.7	30.00
1,3-Butadiene	0.968	ppbv	-6.4	30.00
Bromomethane	1.07	ppbv	3.6	30.00
Chloroethane	1.14	ppbv	9.9	30.00
Acetonitrile	1.13	ppbv	9.2	30.00
Acrolein	1.12	ppbv	9.6	30.00
Trichlorofluoromethane	1.08	ppbv	7.3	30.00
Acrylonitrile	1.14	ppbv	2.9	30.00
1,1-Dichloroethene	1.07	ppbv	3.0	30.00
Dichloromethane	1.12	ppbv	7.7	30.00
Carbon Disulfide	1.03	ppbv	-5.3	30.00
Trichlorotrifluoroethane	0.994	ppbv	-3.5	30.00
trans-1,2-Dichloroethylene	1.03	ppbv	-0.6	30.00
1,1-Dichloroethane	1.08	ppbv	4.5	30.00
Methyl tert-Butyl Ether	1.06	ppbv	2.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606120

Calibration Check (2606120-CCV1) Continued

Prepared & Analyzed: 06/30/26

Chloroprene	1.03	ppbv	-0.5	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	3.8	30.00
Bromochloromethane	1.05	ppbv	2.3	30.00
Chloroform	1.10	ppbv	5.4	30.00
Ethyl tert-Butyl Ether	1.04	ppbv	0.8	30.00
1,2-Dichloroethane	1.10	ppbv	6.4	30.00
1,1,1-Trichloroethane	1.07	ppbv	4.2	30.00
Benzene	1.05	ppbv	1.5	30.00
Carbon Tetrachloride	1.06	ppbv	3.3	30.00
tert-Amyl Methyl Ether	1.03	ppbv	-0.3	30.00
1,2-Dichloropropane	1.09	ppbv	5.1	30.00
Ethyl Acrylate	1.08	ppbv	3.4	30.00
Bromodichloromethane	1.13	ppbv	9.2	30.00
Trichloroethylene	1.06	ppbv	2.6	30.00
Methyl Methacrylate	1.09	ppbv	4.5	30.00
cis-1,3-Dichloropropene	1.06	ppbv	1.6	30.00
Methyl Isobutyl Ketone	1.03	ppbv	-1.0	30.00
trans-1,3-Dichloropropene	1.15	ppbv	9.8	30.00
1,1,2-Trichloroethane	1.11	ppbv	6.0	30.00
Toluene	1.10	ppbv	6.2	30.00
Dibromochloromethane	1.10	ppbv	6.4	30.00
1,2-Dibromoethane	1.07	ppbv	2.3	30.00
n-Octane	1.16	ppbv	12.2	30.00
Tetrachloroethylene	1.07	ppbv	2.3	30.00
Chlorobenzene	1.10	ppbv	4.3	30.00
Ethylbenzene	1.09	ppbv	4.3	30.00
m,p-Xylene	2.37	ppbv	13.4	30.00
Bromoform	1.13	ppbv	7.6	30.00
Styrene	1.21	ppbv	14.0	30.00
1,1,2,2-Tetrachloroethane	1.19	ppbv	12.6	30.00
o-Xylene	1.13	ppbv	7.3	30.00
1,3,5-Trimethylbenzene	1.05	ppbv	-0.4	30.00
1,2,4-Trimethylbenzene	1.06	ppbv	1.0	30.00
m-Dichlorobenzene	1.11	ppbv	6.4	30.00
p-Dichlorobenzene	1.11	ppbv	6.4	30.00
o-Dichlorobenzene	1.09	ppbv	4.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606120

Calibration Check (2606120-CCV1) Continued

Prepared & Analyzed: 06/30/26

1,2,4-Trichlorobenzene	0.963	ppbv	-7.2	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.8	30.00

Calibration Check (2606120-CCV2)

Prepared: 06/30/26 Analyzed: 07/01/26

Dichlorodifluoromethane	1.07	ppbv	3.9	30.00
Chloromethane	1.05	ppbv	3.4	30.00
Dichlorotetrafluoroethane	1.06	ppbv	2.2	30.00
Vinyl chloride	0.987	ppbv	-4.4	30.00
1,3-Butadiene	0.916	ppbv	-11.4	30.00
Bromomethane	1.06	ppbv	2.9	30.00
Chloroethane	1.12	ppbv	8.4	30.00
Acetonitrile	1.16	ppbv	11.8	30.00
Acrolein	1.12	ppbv	9.6	30.00
Trichlorofluoromethane	1.07	ppbv	5.8	30.00
Acrylonitrile	1.12	ppbv	1.5	30.00
1,1-Dichloroethene	1.07	ppbv	2.6	30.00
Dichloromethane	1.18	ppbv	13.8	30.00
Carbon Disulfide	1.01	ppbv	-6.8	30.00
Trichlorotrifluoroethane	0.988	ppbv	-4.1	30.00
trans-1,2-Dichloroethylene	1.02	ppbv	-2.3	30.00
1,1-Dichloroethane	1.08	ppbv	4.7	30.00
Methyl tert-Butyl Ether	1.02	ppbv	-1.1	30.00
Chloroprene	0.992	ppbv	-4.5	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.0	30.00
Bromochloromethane	1.04	ppbv	1.4	30.00
Chloroform	1.10	ppbv	5.0	30.00
Ethyl tert-Butyl Ether	1.00	ppbv	-3.4	30.00
1,2-Dichloroethane	1.10	ppbv	6.1	30.00
1,1,1-Trichloroethane	1.06	ppbv	2.7	30.00
Benzene	1.04	ppbv	-0.3	30.00
Carbon Tetrachloride	1.04	ppbv	1.8	30.00
tert-Amyl Methyl Ether	0.992	ppbv	-4.1	30.00
1,2-Dichloropropane	1.09	ppbv	5.3	30.00
Ethyl Acrylate	1.06	ppbv	1.7	30.00
Bromodichloromethane	1.13	ppbv	9.1	30.00
Trichloroethylene	1.04	ppbv	0.7	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606120

Calibration Check (2606120-CCV2) Continued

Prepared: 06/30/26 Analyzed: 07/01/26

Methyl Methacrylate	1.06	ppbv	1.9	30.00
cis-1,3-Dichloropropene	1.03	ppbv	-1.4	30.00
Methyl Isobutyl Ketone	1.01	ppbv	-3.1	30.00
trans-1,3-Dichloropropene	1.13	ppbv	7.3	30.00
1,1,2-Trichloroethane	1.10	ppbv	5.4	30.00
Toluene	1.08	ppbv	3.6	30.00
Dibromochloromethane	1.08	ppbv	4.1	30.00
1,2-Dibromoethane	1.05	ppbv	0.2	30.00
n-Octane	1.14	ppbv	10.2	30.00
Tetrachloroethylene	1.03	ppbv	-1.0	30.00
Chlorobenzene	1.09	ppbv	3.6	30.00
Ethylbenzene	1.07	ppbv	2.1	30.00
m,p-Xylene	2.33	ppbv	11.8	30.00
Bromoform	1.11	ppbv	5.7	30.00
Styrene	1.18	ppbv	11.3	30.00
1,1,2,2-Tetrachloroethane	1.20	ppbv	13.9	30.00
o-Xylene	1.10	ppbv	4.3	30.00
1,3,5-Trimethylbenzene	1.00	ppbv	-4.9	30.00
1,2,4-Trimethylbenzene	1.02	ppbv	-3.6	30.00
m-Dichlorobenzene	1.09	ppbv	4.2	30.00
p-Dichlorobenzene	1.08	ppbv	3.5	30.00
o-Dichlorobenzene	1.06	ppbv	1.4	30.00
1,2,4-Trichlorobenzene	0.926	ppbv	-10.8	30.00
Hexachloro-1,3-butadiene	1.02	ppbv	-2.8	30.00

Sequence 2607044

Calibration Check (2607044-CCV1)

Prepared & Analyzed: 07/13/26

Dichlorodifluoromethane	1.23	ppbv	19.6	30.00
Chloromethane	1.19	ppbv	16.5	30.00
Dichlorotetrafluoroethane	1.19	ppbv	14.8	30.00
Vinyl chloride	1.13	ppbv	9.5	30.00
1,3-Butadiene	0.978	ppbv	-5.4	30.00
Bromomethane	1.23	ppbv	19.4	30.00
Chloroethane	1.23	ppbv	19.2	30.00
Acetonitrile	1.24	ppbv	19.8	30.00
Acrolein	1.05	ppbv	3.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2607044

Calibration Check (2607044-CCV1) Continued

Prepared & Analyzed: 07/13/26

Trichlorofluoromethane	0.950	ppbv	-6.0	30.00
Acrylonitrile	1.21	ppbv	9.5	30.00
1,1-Dichloroethene	0.927	ppbv	-10.8	30.00
Dichloromethane	1.13	ppbv	9.2	30.00
Carbon Disulfide	0.932	ppbv	-14.4	30.00
Trichlorotrifluoroethane	0.898	ppbv	-12.8	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.7	30.00
1,1-Dichloroethane	1.08	ppbv	3.9	30.00
Methyl tert-Butyl Ether	1.10	ppbv	5.8	30.00
Chloroprene	0.878	ppbv	-15.4	30.00
cis-1,2-Dichloroethylene	1.05	ppbv	1.6	30.00
Bromochloromethane	1.06	ppbv	3.3	30.00
Chloroform	1.07	ppbv	1.7	30.00
Ethyl tert-Butyl Ether	1.07	ppbv	3.3	30.00
1,2-Dichloroethane	0.942	ppbv	-8.9	30.00
1,1,1-Trichloroethane	1.02	ppbv	-1.3	30.00
Benzene	1.07	ppbv	3.3	30.00
Carbon Tetrachloride	0.979	ppbv	-4.4	30.00
tert-Amyl Methyl Ether	1.09	ppbv	5.0	30.00
1,2-Dichloropropane	1.28	ppbv	23.0	30.00
Ethyl Acrylate	1.11	ppbv	5.9	30.00
Bromodichloromethane	1.08	ppbv	4.7	30.00
Trichloroethylene	1.08	ppbv	4.7	30.00
Methyl Methacrylate	1.07	ppbv	2.6	30.00
cis-1,3-Dichloropropene	1.08	ppbv	3.4	30.00
Methyl Isobutyl Ketone	1.22	ppbv	17.3	30.00
trans-1,3-Dichloropropene	1.12	ppbv	6.8	30.00
1,1,2-Trichloroethane	1.17	ppbv	12.0	30.00
Toluene	1.13	ppbv	8.6	30.00
Dibromochloromethane	1.11	ppbv	7.7	30.00
1,2-Dibromoethane	1.14	ppbv	8.8	30.00
n-Octane	1.23	ppbv	18.7	30.00
Tetrachloroethylene	1.17	ppbv	11.8	30.00
Chlorobenzene	1.12	ppbv	6.9	30.00
Ethylbenzene	1.08	ppbv	2.8	30.00
m,p-Xylene	2.21	ppbv	5.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2607044

Calibration Check (2607044-CCV1) Continued

Prepared & Analyzed: 07/13/26

Bromoform	1.18	ppbv	12.7	30.00
Styrene	1.11	ppbv	5.2	30.00
1,1,2,2-Tetrachloroethane	1.24	ppbv	17.5	30.00
o-Xylene	1.09	ppbv	2.8	30.00
1,3,5-Trimethylbenzene	0.926	ppbv	-12.0	30.00
1,2,4-Trimethylbenzene	1.02	ppbv	-2.8	30.00
m-Dichlorobenzene	1.15	ppbv	10.2	30.00
p-Dichlorobenzene	1.15	ppbv	10.4	30.00
o-Dichlorobenzene	1.14	ppbv	8.6	30.00
1,2,4-Trichlorobenzene	1.11	ppbv	7.1	30.00
Hexachloro-1,3-butadiene	1.19	ppbv	13.2	30.00

Calibration Check (2607044-CCV2)

Prepared: 07/13/26 Analyzed: 07/14/26

Dichlorodifluoromethane	0.990	ppbv	-3.7	30.00
Chloromethane	1.11	ppbv	9.3	30.00
Dichlorotetrafluoroethane	1.06	ppbv	2.5	30.00
Vinyl chloride	1.05	ppbv	1.5	30.00
1,3-Butadiene	0.854	ppbv	-17.4	30.00
Bromomethane	1.01	ppbv	-2.2	30.00
Chloroethane	1.07	ppbv	3.0	30.00
Acetonitrile	1.29	ppbv	24.6	30.00
Acrolein	1.05	ppbv	3.2	30.00
Trichlorofluoromethane	0.905	ppbv	-10.4	30.00
Acrylonitrile	1.19	ppbv	8.2	30.00
1,1-Dichloroethene	0.908	ppbv	-12.7	30.00
Dichloromethane	1.16	ppbv	12.3	30.00
Carbon Disulfide	0.908	ppbv	-16.6	30.00
Trichlorotrifluoroethane	0.875	ppbv	-15.0	30.00
trans-1,2-Dichloroethylene	1.04	ppbv	-0.2	30.00
1,1-Dichloroethane	1.10	ppbv	6.5	30.00
Methyl tert-Butyl Ether	1.02	ppbv	-1.1	30.00
Chloroprene	0.839	ppbv	-19.2	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.6	30.00
Bromochloromethane	1.05	ppbv	2.6	30.00
Chloroform	1.08	ppbv	2.9	30.00
Ethyl tert-Butyl Ether	1.01	ppbv	-2.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2607044

Calibration Check (2607044-CCV2) Continued

Prepared: 07/13/26 Analyzed: 07/14/26

1,2-Dichloroethane	1.02	ppbv	-1.4	30.00
1,1,1-Trichloroethane	1.02	ppbv	-0.7	30.00
Benzene	1.07	ppbv	3.0	30.00
Carbon Tetrachloride	0.989	ppbv	-3.4	30.00
tert-Amyl Methyl Ether	1.06	ppbv	2.1	30.00
1,2-Dichloropropane	1.24	ppbv	19.5	30.00
Ethyl Acrylate	1.16	ppbv	10.7	30.00
Bromodichloromethane	1.25	ppbv	21.0	30.00
Trichloroethylene	1.09	ppbv	5.4	30.00
Methyl Methacrylate	1.13	ppbv	8.9	30.00
cis-1,3-Dichloropropene	1.16	ppbv	10.8	30.00
Methyl Isobutyl Ketone	1.32	ppbv	26.8	30.00
trans-1,3-Dichloropropene	1.22	ppbv	16.0	30.00
1,1,2-Trichloroethane	1.26	ppbv	20.2	30.00
Toluene	1.19	ppbv	14.8	30.00
Dibromochloromethane	1.21	ppbv	16.6	30.00
1,2-Dibromoethane	1.10	ppbv	5.6	30.00
n-Octane	1.26	ppbv	21.1	30.00
Tetrachloroethylene	1.19	ppbv	14.1	30.00
Chlorobenzene	1.10	ppbv	4.9	30.00
Ethylbenzene	1.09	ppbv	4.4	30.00
m,p-Xylene	2.25	ppbv	7.6	30.00
Bromoform	1.10	ppbv	5.6	30.00
Styrene	1.15	ppbv	8.6	30.00
1,1,2,2-Tetrachloroethane	1.28	ppbv	21.7	30.00
o-Xylene	1.18	ppbv	11.4	30.00
1,3,5-Trimethylbenzene	0.831	ppbv	-21.0	30.00
1,2,4-Trimethylbenzene	0.791	ppbv	-25.0	30.00
m-Dichlorobenzene	1.00	ppbv	-3.9	30.00
p-Dichlorobenzene	1.00	ppbv	-4.4	30.00
o-Dichlorobenzene	1.00	ppbv	-4.3	30.00
1,2,4-Trichlorobenzene	0.986	ppbv	-5.0	30.00
Hexachloro-1,3-butadiene	1.15	ppbv	9.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 08/12/26 12:46

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
LJ	Identification of analyte is acceptable; reported value is an estimate.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
D-F	Duplicate exceeds DQO criteria.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.