



Eastern Research Group  
601 Keystone Park Drive  
Suite 700  
Morrisville, NC 27560

May 14, 2026

Mr. Bryan Sims  
Oklahoma Dept. of Environmental Quality  
707 N. Robinson, PO Box 1677  
Oklahoma City, OK 73101-1677  
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 03/18/26 11:17 through 04/01/26 13:04.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower  
Program Manager  
[randy.bower@erg.com](mailto:randy.bower@erg.com)

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify [randy.bower@erg.com](mailto:randy.bower@erg.com) and delete the report without retaining any copies.



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

**ATTN:** Mr. Bryan Sims

**PHONE:** (405) 702-0100

**FILE #:** 0475

**REPORTED:** 05/14/26 10:06

**SUBMITTED:** 03/18/26 to 04/01/26

**AQS SITE CODE:** 40 143 1127

**SITE CODE:** TMOK

### ANALYTICAL REPORT FOR SAMPLES

| <u>SampleName</u> | <u>LabNumber</u> | <u>Matrix</u> | <u>Sampled</u> | <u>Received</u> |
|-------------------|------------------|---------------|----------------|-----------------|
| TMOK D1           | 6031817-04       | Air           | 03/15/26 23:59 | 03/18/26 11:17  |
| TMOK D2           | 6031817-05       | Air           | 03/15/26 23:59 | 03/18/26 11:17  |
| TMOK              | 6032352-03       | Air           | 03/17/26 23:59 | 03/23/26 11:07  |
| TMOK              | 6032352-04       | Air           | 03/18/26 23:59 | 03/23/26 11:07  |
| TMOK              | 6040116-03       | Air           | 03/21/26 23:59 | 04/01/26 13:04  |
| TMOK              | 6040116-04       | Air           | 03/27/26 23:59 | 04/01/26 13:04  |
| TMOK FB           | 6040116-05       | Air           | 03/27/26 23:59 | 04/01/26 13:04  |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

|                             |                                   |                                 |
|-----------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK D1 | <b>Lab ID:</b> 6031817-04         | <b>Sampled:</b> 03/15/26 23:59  |
| <b>Temp Rec'd (°C):</b> 2.7 | <b>Sample Volume:</b> 1146.1 L    | <b>Received:</b> 03/18/26 11:17 |
|                             | <b>Cartridge Pouch #:</b> 26G0952 | <b>Analyzed:</b> 04/06/26 16:53 |

**Comments:**

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 2.09                        | QB-01       | 0.0552                      |
| Acetaldehyde    | 1.22                        |             | 0.0229                      |
| Acetone         | 2.18                        |             | 0.0846                      |
| Propionaldehyde | 0.370                       |             | 0.181                       |
| Crotonaldehyde  | 0.0194                      |             | 0.0166                      |
| Butyraldehyde   | 0.151                       |             | 0.00977                     |
| 2-Butanone      | 1.41                        |             | 0.0368                      |
| Benzaldehyde    | 0.0818                      |             | 0.0378                      |
| Valeraldehyde   | 0.130                       |             | 0.0106                      |
| Hexaldehyde     | 0.0802                      |             | 0.0106                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

|                             |                                   |                                 |
|-----------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK D2 | <b>Lab ID:</b> 6031817-05         | <b>Sampled:</b> 03/15/26 23:59  |
| <b>Temp Rec'd (°C):</b> 2.7 | <b>Sample Volume:</b> 1145.4 L    | <b>Received:</b> 03/18/26 11:17 |
|                             | <b>Cartridge Pouch #:</b> 26G0953 | <b>Analyzed:</b> 04/06/26 17:26 |

Comments:

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 2.06                        | QB-01       | 0.0553                      |
| Acetaldehyde    | 1.20                        |             | 0.0229                      |
| Acetone         | 2.15                        |             | 0.0847                      |
| Propionaldehyde | 0.369                       |             | 0.181                       |
| Crotonaldehyde  | 0.0247                      |             | 0.0166                      |
| Butyraldehyde   | 0.147                       |             | 0.00978                     |
| 2-Butanone      | 1.39                        |             | 0.0368                      |
| Benzaldehyde    | 0.0810                      |             | 0.0378                      |
| Valeraldehyde   | 0.145                       |             | 0.0106                      |
| Hexaldehyde     | 0.0894                      |             | 0.0107                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

---

|                                                      |                                   |                                 |
|------------------------------------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK                             | <b>Lab ID:</b> 6032352-03         | <b>Sampled:</b> 03/17/26 23:59  |
| <b>Temp Rec'd (°C):</b> 21.1                         | <b>Sample Volume:</b> 1156.3 L    | <b>Received:</b> 03/23/26 11:07 |
|                                                      | <b>Cartridge Pouch #:</b> 26G0947 | <b>Analyzed:</b> 03/26/26 19:28 |
| <b>Comments:</b> Received Monday, Makeup for 2/19/26 |                                   |                                 |

---

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 2.34                        | TT          | 0.0547                      |
| Acetaldehyde    | 2.03                        | TT          | 0.0227                      |
| Acetone         | 3.78                        | TT          | 0.0839                      |
| Propionaldehyde | 0.372                       | TT          | 0.179                       |
| Crotonaldehyde  | ND                          | TT, U       | 0.0164                      |
| Butyraldehyde   | 0.132                       | TT          | 0.00969                     |
| 2-Butanone      | 0.882                       | TT          | 0.0365                      |
| Benzaldehyde    | 0.0450                      | TT          | 0.0374                      |
| Valeraldehyde   | 0.110                       | TT          | 0.0105                      |
| Hexaldehyde     | 0.0549                      | TT          | 0.0106                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

---

|                                                     |                                   |                                 |
|-----------------------------------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK                            | <b>Lab ID:</b> 6032352-04         | <b>Sampled:</b> 03/18/26 23:59  |
| <b>Temp Rec'd (°C):</b> 21.1                        | <b>Sample Volume:</b> 1155 L      | <b>Received:</b> 03/23/26 11:07 |
|                                                     | <b>Cartridge Pouch #:</b> 26M0024 | <b>Analyzed:</b> 04/06/26 12:49 |
| <b>Comments:</b> Received Monday, Makeup for 2/7/26 |                                   |                                 |

---

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 3.98                        | TT          | 0.0548                      |
| Acetaldehyde    | 2.65                        | TT          | 0.0227                      |
| Acetone         | 6.45                        | TT          | 0.0840                      |
| Propionaldehyde | 0.739                       | QB-01, TT   | 0.179                       |
| Crotonaldehyde  | 0.0496                      | TT          | 0.0165                      |
| Butyraldehyde   | 0.238                       | TT          | 0.00970                     |
| 2-Butanone      | 2.09                        | TT          | 0.0365                      |
| Benzaldehyde    | 0.109                       | TT          | 0.0375                      |
| Valeraldehyde   | 0.195                       | TT          | 0.0105                      |
| Hexaldehyde     | 0.116                       | TT          | 0.0106                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

|                             |                                   |                                 |
|-----------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK    | <b>Lab ID:</b> 6040116-03         | <b>Sampled:</b> 03/21/26 23:59  |
| <b>Temp Rec'd (°C):</b> 8.6 | <b>Sample Volume:</b> 1156.3 L    | <b>Received:</b> 04/01/26 13:04 |
|                             | <b>Cartridge Pouch #:</b> 26G0954 | <b>Analyzed:</b> 04/02/26 18:45 |

**Comments:**

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 5.87                        | QB-01       | 0.0547                      |
| Acetaldehyde    | 4.44                        |             | 0.0227                      |
| Acetone         | 3.80                        |             | 0.0839                      |
| Propionaldehyde | 1.15                        |             | 0.179                       |
| Crotonaldehyde  | 0.0687                      |             | 0.0164                      |
| Butyraldehyde   | 0.510                       |             | 0.00969                     |
| 2-Butanone      | 2.36                        |             | 0.0365                      |
| Benzaldehyde    | 0.188                       |             | 0.0374                      |
| Valeraldehyde   | 0.411                       |             | 0.0105                      |
| Hexaldehyde     | 0.173                       |             | 0.0106                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

|                             |                                   |                                 |
|-----------------------------|-----------------------------------|---------------------------------|
| <b>Description:</b> TMOK    | <b>Lab ID:</b> 6040116-04         | <b>Sampled:</b> 03/27/26 23:59  |
| <b>Temp Rec'd (°C):</b> 0.5 | <b>Sample Volume:</b> 1156.4 L    | <b>Received:</b> 04/01/26 13:04 |
|                             | <b>Cartridge Pouch #:</b> 26G0955 | <b>Analyzed:</b> 04/07/26 16:19 |

**Comments:**

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 2.15                        |             | 0.0547                      |
| Acetaldehyde    | 1.23                        |             | 0.0227                      |
| Acetone         | 2.93                        |             | 0.0839                      |
| Propionaldehyde | 0.447                       | QB-01       | 0.179                       |
| Crotonaldehyde  | ND                          | U           | 0.0164                      |
| Butyraldehyde   | 0.174                       |             | 0.00969                     |
| 2-Butanone      | 0.457                       |             | 0.0365                      |
| Benzaldehyde    | 0.0768                      |             | 0.0374                      |
| Valeraldehyde   | 0.107                       |             | 0.0105                      |
| Hexaldehyde     | 0.102                       |             | 0.0105                      |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK FB

Lab ID: 6040116-05

Sampled: 03/27/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1156.4 L

Received: 04/01/26 13:04

Cartridge Pouch #: 26G0956

Analyzed: 04/07/26 16:35

Comments:

### Air Toxics by EPA Compendium Method TO-11A

| <u>Analyte</u>  | <u>Results</u>              | <u>Flag</u> | <u>MDL</u>                  |
|-----------------|-----------------------------|-------------|-----------------------------|
|                 | <u>ug/m<sup>3</sup> Air</u> |             | <u>ug/m<sup>3</sup> Air</u> |
| Formaldehyde    | 0.0369                      | U           | 0.0547                      |
| Acetaldehyde    | 0.0318                      |             | 0.0227                      |
| Acetone         | 0.0587                      | U           | 0.0839                      |
| Propionaldehyde | 0.133                       | QB-01, U    | 0.179                       |
| Crotonaldehyde  | ND                          | U           | 0.0164                      |
| Butyraldehyde   | ND                          | U           | 0.00969                     |
| 2-Butanone      | ND                          | U           | 0.0365                      |
| Benzaldehyde    | ND                          | U           | 0.0374                      |
| Valeraldehyde   | ND                          | U           | 0.0105                      |
| Hexaldehyde     | ND                          | U           | 0.0105                      |



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603095 - B6C2509

### Calibration Blank (2603095-CCB1)

Prepared & Analyzed: 03/26/26

|                 |      |       |  |  |  |  |  |  |   |
|-----------------|------|-------|--|--|--|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Blank (2603095-CCB2)

Prepared & Analyzed: 03/26/26

|                 |      |       |  |  |  |  |  |  |   |
|-----------------|------|-------|--|--|--|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Check (2603095-CCV1)

Prepared & Analyzed: 03/26/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.146 | ug/ml | 0.15000 | 97.0 | 84.9-115.1 |
| Acetaldehyde    | 0.149 | ug/ml | 0.15000 | 99.1 | 84.9-115.1 |
| Acetone         | 0.147 | ug/ml | 0.15000 | 98.1 | 84.9-115.1 |
| Propionaldehyde | 0.145 | ug/ml | 0.15000 | 97.0 | 84.9-115.1 |
| Crotonaldehyde  | 0.155 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Butyraldehyde   | 0.147 | ug/ml | 0.15000 | 98.1 | 84.9-115.1 |
| 2-Butanone      | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Benzaldehyde    | 0.155 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Valeraldehyde   | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |

### Calibration Check (2603095-CCV2)

Prepared & Analyzed: 03/26/26

|              |       |       |         |      |            |
|--------------|-------|-------|---------|------|------------|
| Formaldehyde | 0.146 | ug/ml | 0.15000 | 97.4 | 84.9-115.1 |
| Acetaldehyde | 0.149 | ug/ml | 0.15000 | 99.1 | 84.9-115.1 |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2603095 - B6C2509

### Calibration Check (2603095-CCV2) Contin

Prepared & Analyzed: 03/26/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Acetone         | 0.149 | ug/ml | 0.15000 | 99.1 | 84.9-115.1 |
| Propionaldehyde | 0.147 | ug/ml | 0.15000 | 98.2 | 84.9-115.1 |
| Crotonaldehyde  | 0.149 | ug/ml | 0.15000 | 99.3 | 84.9-115.1 |
| Butyraldehyde   | 0.144 | ug/ml | 0.15000 | 96.2 | 84.9-115.1 |
| 2-Butanone      | 0.147 | ug/ml | 0.15000 | 98.3 | 84.9-115.1 |
| Benzaldehyde    | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Valeraldehyde   | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Hexaldehyde     | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |

### Calibration Check (2603095-CCV3)

Prepared & Analyzed: 03/26/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.144 | ug/ml | 0.15000 | 95.8 | 84.9-115.1 |
| Acetaldehyde    | 0.150 | ug/ml | 0.15000 | 99.7 | 84.9-115.1 |
| Acetone         | 0.148 | ug/ml | 0.15000 | 98.5 | 84.9-115.1 |
| Propionaldehyde | 0.147 | ug/ml | 0.15000 | 97.9 | 84.9-115.1 |
| Crotonaldehyde  | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Butyraldehyde   | 0.146 | ug/ml | 0.15000 | 97.3 | 84.9-115.1 |
| 2-Butanone      | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Benzaldehyde    | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Valeraldehyde   | 0.148 | ug/ml | 0.15000 | 98.8 | 84.9-115.1 |
| Hexaldehyde     | 0.147 | ug/ml | 0.15000 | 97.9 | 84.9-115.1 |

Batch 2604008 - B6D0204

### Calibration Blank (2604008-CCB1)

Prepared & Analyzed: 04/02/26

|                 |      |       |  |  |  |   |
|-----------------|------|-------|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  | U |

### Calibration Blank (2604008-CCB2)

Prepared: 04/02/26 Analyzed: 04/03/26

|                 |      |       |  |  |  |   |
|-----------------|------|-------|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  | U |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604008 - B6D0204

### Calibration Blank (2604008-CCB2) Contin

Prepared: 04/02/26 Analyzed: 04/03/26

|               |      |       |  |  |  |  |  |  |   |
|---------------|------|-------|--|--|--|--|--|--|---|
| Butyraldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Check (2604008-CCV1)

Prepared & Analyzed: 04/02/26

|                 |       |       |         |     |            |  |  |  |  |
|-----------------|-------|-------|---------|-----|------------|--|--|--|--|
| Formaldehyde    | 0.151 | ug/ml | 0.15000 | 100 | 84.9-115.1 |  |  |  |  |
| Acetaldehyde    | 0.151 | ug/ml | 0.15000 | 100 | 84.9-115.1 |  |  |  |  |
| Acetone         | 0.152 | ug/ml | 0.15000 | 102 | 84.9-115.1 |  |  |  |  |
| Propionaldehyde | 0.159 | ug/ml | 0.15000 | 106 | 84.9-115.1 |  |  |  |  |
| Crotonaldehyde  | 0.155 | ug/ml | 0.15000 | 103 | 84.9-115.1 |  |  |  |  |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 | 100 | 84.9-115.1 |  |  |  |  |
| 2-Butanone      | 0.151 | ug/ml | 0.15000 | 101 | 84.9-115.1 |  |  |  |  |
| Benzaldehyde    | 0.163 | ug/ml | 0.15000 | 108 | 84.9-115.1 |  |  |  |  |
| Valeraldehyde   | 0.158 | ug/ml | 0.15000 | 106 | 84.9-115.1 |  |  |  |  |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 | 100 | 84.9-115.1 |  |  |  |  |

### Calibration Check (2604008-CCV2)

Prepared & Analyzed: 04/02/26

|                 |       |       |         |      |            |  |  |  |  |
|-----------------|-------|-------|---------|------|------------|--|--|--|--|
| Formaldehyde    | 0.150 | ug/ml | 0.15000 | 99.9 | 84.9-115.1 |  |  |  |  |
| Acetaldehyde    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |  |  |  |  |
| Acetone         | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |  |  |  |  |
| Propionaldehyde | 0.156 | ug/ml | 0.15000 | 104  | 84.9-115.1 |  |  |  |  |
| Crotonaldehyde  | 0.157 | ug/ml | 0.15000 | 105  | 84.9-115.1 |  |  |  |  |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |  |  |  |  |
| 2-Butanone      | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |  |  |  |  |
| Benzaldehyde    | 0.161 | ug/ml | 0.15000 | 107  | 84.9-115.1 |  |  |  |  |
| Valeraldehyde   | 0.157 | ug/ml | 0.15000 | 105  | 84.9-115.1 |  |  |  |  |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |  |  |  |  |

### Calibration Check (2604008-CCV3)

Prepared & Analyzed: 04/02/26

|                 |       |       |         |      |            |  |  |  |  |
|-----------------|-------|-------|---------|------|------------|--|--|--|--|
| Formaldehyde    | 0.150 | ug/ml | 0.15000 | 99.8 | 84.9-115.1 |  |  |  |  |
| Acetaldehyde    | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |  |  |  |  |
| Acetone         | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |  |  |  |  |
| Propionaldehyde | 0.157 | ug/ml | 0.15000 | 104  | 84.9-115.1 |  |  |  |  |
| Crotonaldehyde  | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |  |  |  |  |
| Butyraldehyde   | 0.148 | ug/ml | 0.15000 | 99.0 | 84.9-115.1 |  |  |  |  |
| 2-Butanone      | 0.150 | ug/ml | 0.15000 | 99.8 | 84.9-115.1 |  |  |  |  |
| Benzaldehyde    | 0.161 | ug/ml | 0.15000 | 107  | 84.9-115.1 |  |  |  |  |
| Valeraldehyde   | 0.159 | ug/ml | 0.15000 | 106  | 84.9-115.1 |  |  |  |  |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604008 - B6D0204

### Calibration Check (2604008-CCV3) Contin

Prepared & Analyzed: 04/02/26

|             |       |       |         |     |            |
|-------------|-------|-------|---------|-----|------------|
| Hexaldehyde | 0.151 | ug/ml | 0.15000 | 101 | 84.9-115.1 |
|-------------|-------|-------|---------|-----|------------|

### Calibration Check (2604008-CCV4)

Prepared: 04/02/26 Analyzed: 04/03/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Acetaldehyde    | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Acetone         | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Propionaldehyde | 0.158 | ug/ml | 0.15000 | 105  | 84.9-115.1 |
| Crotonaldehyde  | 0.154 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Butyraldehyde   | 0.148 | ug/ml | 0.15000 | 98.9 | 84.9-115.1 |
| 2-Butanone      | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Benzaldehyde    | 0.159 | ug/ml | 0.15000 | 106  | 84.9-115.1 |
| Valeraldehyde   | 0.160 | ug/ml | 0.15000 | 107  | 84.9-115.1 |
| Hexaldehyde     | 0.149 | ug/ml | 0.15000 | 99.1 | 84.9-115.1 |

### Calibration Check (2604008-CCV5)

Prepared: 04/02/26 Analyzed: 04/03/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.148 | ug/ml | 0.15000 | 99.0 | 84.9-115.1 |
| Acetaldehyde    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Acetone         | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Propionaldehyde | 0.157 | ug/ml | 0.15000 | 105  | 84.9-115.1 |
| Crotonaldehyde  | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 | 99.9 | 84.9-115.1 |
| 2-Butanone      | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Benzaldehyde    | 0.158 | ug/ml | 0.15000 | 106  | 84.9-115.1 |
| Valeraldehyde   | 0.157 | ug/ml | 0.15000 | 105  | 84.9-115.1 |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 | 99.8 | 84.9-115.1 |

Batch 2604019 - B6C2702

### Calibration Blank (2604019-CCB1)

Prepared & Analyzed: 04/06/26

|                 |      |       |  |  |   |
|-----------------|------|-------|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  | U |

### Calibration Blank (2604019-CCB2)

Prepared & Analyzed: 04/06/26

|              |      |       |  |  |   |
|--------------|------|-------|--|--|---|
| Formaldehyde | 0.00 | ug/ml |  |  | U |
|--------------|------|-------|--|--|---|

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604019 - B6C2702

### Calibration Blank (2604019-CCB2) Contin

Prepared & Analyzed: 04/06/26

|                 |      |       |  |  |  |  |  |  |   |
|-----------------|------|-------|--|--|--|--|--|--|---|
| Acetaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Check (2604019-CCV1)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.148 | ug/ml | 0.15000 | 98.7 | 84.9-115.1 |
| Acetaldehyde    | 0.154 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Acetone         | 0.154 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Propionaldehyde | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Crotonaldehyde  | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| 2-Butanone      | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Benzaldehyde    | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Valeraldehyde   | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Hexaldehyde     | 0.147 | ug/ml | 0.15000 | 98.0 | 84.9-115.1 |

### Calibration Check (2604019-CCV2)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.149 | ug/ml | 0.15000 | 99.4 | 84.9-115.1 |
| Acetaldehyde    | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Acetone         | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Propionaldehyde | 0.145 | ug/ml | 0.15000 | 96.6 | 84.9-115.1 |
| Crotonaldehyde  | 0.150 | ug/ml | 0.15000 | 99.9 | 84.9-115.1 |
| Butyraldehyde   | 0.147 | ug/ml | 0.15000 | 97.7 | 84.9-115.1 |
| 2-Butanone      | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Benzaldehyde    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Valeraldehyde   | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Hexaldehyde     | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |

### Calibration Check (2604019-CCV3)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.147 | ug/ml | 0.15000 | 98.2 | 84.9-115.1 |
| Acetaldehyde    | 0.151 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Acetone         | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Propionaldehyde | 0.149 | ug/ml | 0.15000 | 99.5 | 84.9-115.1 |
| Crotonaldehyde  | 0.149 | ug/ml | 0.15000 | 99.6 | 84.9-115.1 |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604019 - B6C2702

### Calibration Check (2604019-CCV3) Contin

Prepared & Analyzed: 04/06/26

|               |       |       |         |      |            |
|---------------|-------|-------|---------|------|------------|
| Butyraldehyde | 0.148 | ug/ml | 0.15000 | 98.7 | 84.9-115.1 |
| 2-Butanone    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Benzaldehyde  | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Valeraldehyde | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Hexaldehyde   | 0.148 | ug/ml | 0.15000 | 99.0 | 84.9-115.1 |

### Calibration Check (2604019-CCV4)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.148 | ug/ml | 0.15000 | 98.7 | 84.9-115.1 |
| Acetaldehyde    | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Acetone         | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Propionaldehyde | 0.149 | ug/ml | 0.15000 | 99.5 | 84.9-115.1 |
| Crotonaldehyde  | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Butyraldehyde   | 0.148 | ug/ml | 0.15000 | 98.5 | 84.9-115.1 |
| 2-Butanone      | 0.150 | ug/ml | 0.15000 | 99.9 | 84.9-115.1 |
| Benzaldehyde    | 0.156 | ug/ml | 0.15000 | 104  | 84.9-115.1 |
| Valeraldehyde   | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 | 99.8 | 84.9-115.1 |

Batch 2604020 - B6C3010

### Calibration Blank (2604020-CCB1)

Prepared & Analyzed: 04/06/26

|                 |      |       |  |  |   |
|-----------------|------|-------|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  | U |

### Calibration Blank (2604020-CCB2)

Prepared & Analyzed: 04/06/26

|                 |      |       |  |  |   |
|-----------------|------|-------|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  | U |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604020 - B6C3010

### Calibration Blank (2604020-CCB2) Contin

Prepared & Analyzed: 04/06/26

|               |      |       |  |  |  |  |  |  |   |
|---------------|------|-------|--|--|--|--|--|--|---|
| Valeraldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Check (2604020-CCV1)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |  |      |            |  |  |  |
|-----------------|-------|-------|---------|--|------|------------|--|--|--|
| Formaldehyde    | 0.151 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Acetaldehyde    | 0.151 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Acetone         | 0.151 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Propionaldehyde | 0.158 | ug/ml | 0.15000 |  | 105  | 84.9-115.1 |  |  |  |
| Crotonaldehyde  | 0.156 | ug/ml | 0.15000 |  | 104  | 84.9-115.1 |  |  |  |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 |  | 99.9 | 84.9-115.1 |  |  |  |
| 2-Butanone      | 0.151 | ug/ml | 0.15000 |  | 100  | 84.9-115.1 |  |  |  |
| Benzaldehyde    | 0.162 | ug/ml | 0.15000 |  | 108  | 84.9-115.1 |  |  |  |
| Valeraldehyde   | 0.157 | ug/ml | 0.15000 |  | 104  | 84.9-115.1 |  |  |  |
| Hexaldehyde     | 0.151 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |

### Calibration Check (2604020-CCV2)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |  |      |            |  |  |  |
|-----------------|-------|-------|---------|--|------|------------|--|--|--|
| Formaldehyde    | 0.151 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Acetaldehyde    | 0.152 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Acetone         | 0.153 | ug/ml | 0.15000 |  | 102  | 84.9-115.1 |  |  |  |
| Propionaldehyde | 0.156 | ug/ml | 0.15000 |  | 104  | 84.9-115.1 |  |  |  |
| Crotonaldehyde  | 0.158 | ug/ml | 0.15000 |  | 105  | 84.9-115.1 |  |  |  |
| Butyraldehyde   | 0.152 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| 2-Butanone      | 0.152 | ug/ml | 0.15000 |  | 102  | 84.9-115.1 |  |  |  |
| Benzaldehyde    | 0.158 | ug/ml | 0.15000 |  | 106  | 84.9-115.1 |  |  |  |
| Valeraldehyde   | 0.157 | ug/ml | 0.15000 |  | 105  | 84.9-115.1 |  |  |  |
| Hexaldehyde     | 0.150 | ug/ml | 0.15000 |  | 99.7 | 84.9-115.1 |  |  |  |

### Calibration Check (2604020-CCV3)

Prepared & Analyzed: 04/06/26

|                 |       |       |         |  |      |            |  |  |  |
|-----------------|-------|-------|---------|--|------|------------|--|--|--|
| Formaldehyde    | 0.152 | ug/ml | 0.15000 |  | 101  | 84.9-115.1 |  |  |  |
| Acetaldehyde    | 0.153 | ug/ml | 0.15000 |  | 102  | 84.9-115.1 |  |  |  |
| Acetone         | 0.153 | ug/ml | 0.15000 |  | 102  | 84.9-115.1 |  |  |  |
| Propionaldehyde | 0.154 | ug/ml | 0.15000 |  | 103  | 84.9-115.1 |  |  |  |
| Crotonaldehyde  | 0.155 | ug/ml | 0.15000 |  | 103  | 84.9-115.1 |  |  |  |
| Butyraldehyde   | 0.150 | ug/ml | 0.15000 |  | 99.9 | 84.9-115.1 |  |  |  |
| 2-Butanone      | 0.149 | ug/ml | 0.15000 |  | 99.5 | 84.9-115.1 |  |  |  |
| Benzaldehyde    | 0.157 | ug/ml | 0.15000 |  | 105  | 84.9-115.1 |  |  |  |
| Valeraldehyde   | 0.159 | ug/ml | 0.15000 |  | 106  | 84.9-115.1 |  |  |  |
| Hexaldehyde     | 0.149 | ug/ml | 0.15000 |  | 99.2 | 84.9-115.1 |  |  |  |

Batch 2604026 - B6D0608

### Calibration Blank (2604026-CCB1)

Prepared & Analyzed: 04/07/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604026 - B6D0608

### Calibration Blank (2604026-CCB1) Contin

Prepared & Analyzed: 04/07/26

|                 |      |       |  |  |  |  |  |  |   |
|-----------------|------|-------|--|--|--|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Blank (2604026-CCB2)

Prepared & Analyzed: 04/07/26

|                 |      |       |  |  |  |  |  |  |   |
|-----------------|------|-------|--|--|--|--|--|--|---|
| Formaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Acetone         | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Propionaldehyde | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Crotonaldehyde  | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Butyraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| 2-Butanone      | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Benzaldehyde    | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Valeraldehyde   | 0.00 | ug/ml |  |  |  |  |  |  | U |
| Hexaldehyde     | 0.00 | ug/ml |  |  |  |  |  |  | U |

### Calibration Check (2604026-CCV1)

Prepared & Analyzed: 04/07/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.147 | ug/ml | 0.15000 | 98.0 | 84.9-115.1 |
| Acetaldehyde    | 0.149 | ug/ml | 0.15000 | 99.6 | 84.9-115.1 |
| Acetone         | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Propionaldehyde | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Crotonaldehyde  | 0.152 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Butyraldehyde   | 0.147 | ug/ml | 0.15000 | 98.1 | 84.9-115.1 |
| 2-Butanone      | 0.149 | ug/ml | 0.15000 | 99.0 | 84.9-115.1 |
| Benzaldehyde    | 0.158 | ug/ml | 0.15000 | 105  | 84.9-115.1 |
| Valeraldehyde   | 0.155 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Hexaldehyde     | 0.147 | ug/ml | 0.15000 | 98.3 | 84.9-115.1 |

### Calibration Check (2604026-CCV2)

Prepared & Analyzed: 04/07/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.149 | ug/ml | 0.15000 | 99.3 | 84.9-115.1 |
| Acetaldehyde    | 0.149 | ug/ml | 0.15000 | 99.1 | 84.9-115.1 |
| Acetone         | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Propionaldehyde | 0.153 | ug/ml | 0.15000 | 102  | 84.9-115.1 |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604026 - B6D0608

### Calibration Check (2604026-CCV2) Contin

Prepared & Analyzed: 04/07/26

|                |       |       |         |      |            |
|----------------|-------|-------|---------|------|------------|
| Crotonaldehyde | 0.152 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Butyraldehyde  | 0.147 | ug/ml | 0.15000 | 98.3 | 84.9-115.1 |
| 2-Butanone     | 0.148 | ug/ml | 0.15000 | 98.6 | 84.9-115.1 |
| Benzaldehyde   | 0.157 | ug/ml | 0.15000 | 105  | 84.9-115.1 |
| Valeraldehyde  | 0.151 | ug/ml | 0.15000 | 101  | 84.9-115.1 |
| Hexaldehyde    | 0.148 | ug/ml | 0.15000 | 98.8 | 84.9-115.1 |

### Calibration Check (2604026-CCV3)

Prepared & Analyzed: 04/07/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.147 | ug/ml | 0.15000 | 97.9 | 84.9-115.1 |
| Acetaldehyde    | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Acetone         | 0.150 | ug/ml | 0.15000 | 100  | 84.9-115.1 |
| Propionaldehyde | 0.155 | ug/ml | 0.15000 | 103  | 84.9-115.1 |
| Crotonaldehyde  | 0.154 | ug/ml | 0.15000 | 102  | 84.9-115.1 |
| Butyraldehyde   | 0.147 | ug/ml | 0.15000 | 97.9 | 84.9-115.1 |
| 2-Butanone      | 0.147 | ug/ml | 0.15000 | 97.9 | 84.9-115.1 |
| Benzaldehyde    | 0.162 | ug/ml | 0.15000 | 108  | 84.9-115.1 |
| Valeraldehyde   | 0.156 | ug/ml | 0.15000 | 104  | 84.9-115.1 |
| Hexaldehyde     | 0.148 | ug/ml | 0.15000 | 98.8 | 84.9-115.1 |

Batch B6C2509 - TO-11A extraction

### Blank (B6C2509-BLK1)

Prepared: 03/25/26 Analyzed: 03/26/26

|                 |        |                       |  |  |   |
|-----------------|--------|-----------------------|--|--|---|
| Formaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Acetaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Acetone         | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Propionaldehyde | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Crotonaldehyde  | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Butyraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  | U |
| 2-Butanone      | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Benzaldehyde    | 0.0514 | ug/m <sup>3</sup> Air |  |  |   |
| Valeraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  | U |
| Hexaldehyde     | ND     | ug/m <sup>3</sup> Air |  |  | U |

### LCS (B6C2509-BS1)

Prepared: 03/25/26 Analyzed: 03/26/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.285 | ug/ml | 0.30000 | 94.9 | 79.9-120.1 |
| Acetaldehyde    | 0.294 | ug/ml | 0.30000 | 97.9 | 79.9-120.1 |
| Acetone         | 0.304 | ug/ml | 0.30000 | 101  | 69.9-130.1 |
| Propionaldehyde | 0.335 | ug/ml | 0.30000 | 112  | 69.9-130.1 |
| Crotonaldehyde  | 0.291 | ug/ml | 0.30000 | 96.9 | 69.9-130.1 |
| Butyraldehyde   | 0.265 | ug/ml | 0.30000 | 88.2 | 69.9-130.1 |
| 2-Butanone      | 0.244 | ug/ml | 0.30000 | 81.3 | 69.9-130.1 |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C2509 - TO-11A extraction

### LCS (B6C2509-BS1) Continued

Prepared: 03/25/26 Analyzed: 03/26/26

|               |       |       |         |      |            |
|---------------|-------|-------|---------|------|------------|
| Benzaldehyde  | 0.315 | ug/ml | 0.30000 | 105  | 69.9-130.1 |
| Valeraldehyde | 0.266 | ug/ml | 0.30000 | 88.7 | 69.9-130.1 |
| Hexaldehyde   | 0.256 | ug/ml | 0.30000 | 85.4 | 69.9-130.1 |

### LCS Dup (B6C2509-BSD1)

Prepared: 03/25/26 Analyzed: 03/26/26

|                 |       |       |         |      |            |       |    |
|-----------------|-------|-------|---------|------|------------|-------|----|
| Formaldehyde    | 0.287 | ug/ml | 0.30000 | 95.7 | 79.9-120.1 | 0.838 | 20 |
| Acetaldehyde    | 0.293 | ug/ml | 0.30000 | 97.6 | 79.9-120.1 | 0.324 | 20 |
| Acetone         | 0.300 | ug/ml | 0.30000 | 100  | 69.9-130.1 | 1.20  | 20 |
| Propionaldehyde | 0.330 | ug/ml | 0.30000 | 110  | 69.9-130.1 | 1.71  | 20 |
| Crotonaldehyde  | 0.290 | ug/ml | 0.30000 | 96.5 | 69.9-130.1 | 0.355 | 20 |
| Butyraldehyde   | 0.262 | ug/ml | 0.30000 | 87.2 | 69.9-130.1 | 1.18  | 20 |
| 2-Butanone      | 0.242 | ug/ml | 0.30000 | 80.7 | 69.9-130.1 | 0.660 | 20 |
| Benzaldehyde    | 0.314 | ug/ml | 0.30000 | 105  | 69.9-130.1 | 0.369 | 20 |
| Valeraldehyde   | 0.261 | ug/ml | 0.30000 | 86.9 | 69.9-130.1 | 1.94  | 20 |
| Hexaldehyde     | 0.255 | ug/ml | 0.30000 | 85.1 | 69.9-130.1 | 0.349 | 20 |

Batch B6C2702 - TO-11A extraction

### Blank (B6C2702-BLK1)

Prepared: 03/27/26 Analyzed: 04/06/26

|                 |        |                       |  |  |  |  |          |
|-----------------|--------|-----------------------|--|--|--|--|----------|
| Formaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Acetaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Acetone         | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Propionaldehyde | ND     | ug/m <sup>3</sup> Air |  |  |  |  | QB-01, U |
| Crotonaldehyde  | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Butyraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| 2-Butanone      | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Benzaldehyde    | 0.0622 | ug/m <sup>3</sup> Air |  |  |  |  |          |
| Valeraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |
| Hexaldehyde     | ND     | ug/m <sup>3</sup> Air |  |  |  |  | U        |

### LCS (B6C2702-BS1)

Prepared: 03/27/26 Analyzed: 04/06/26

|                 |       |       |         |      |            |
|-----------------|-------|-------|---------|------|------------|
| Formaldehyde    | 0.289 | ug/ml | 0.30000 | 96.2 | 79.9-120.1 |
| Acetaldehyde    | 0.295 | ug/ml | 0.30000 | 98.4 | 79.9-120.1 |
| Acetone         | 0.311 | ug/ml | 0.30000 | 104  | 69.9-130.1 |
| Propionaldehyde | 0.347 | ug/ml | 0.30000 | 116  | 69.9-130.1 |
| Crotonaldehyde  | 0.278 | ug/ml | 0.30000 | 92.6 | 69.9-130.1 |
| Butyraldehyde   | 0.260 | ug/ml | 0.30000 | 86.8 | 69.9-130.1 |
| 2-Butanone      | 0.241 | ug/ml | 0.30000 | 80.2 | 69.9-130.1 |
| Benzaldehyde    | 0.310 | ug/ml | 0.30000 | 103  | 69.9-130.1 |
| Valeraldehyde   | 0.262 | ug/ml | 0.30000 | 87.3 | 69.9-130.1 |
| Hexaldehyde     | 0.253 | ug/ml | 0.30000 | 84.4 | 69.9-130.1 |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C2702 - TO-11A extraction

### LCS Dup (B6C2702-BSD1)

Prepared: 03/27/26 Analyzed: 04/06/26

|                 |       |       |         |  |      |            |        |    |  |
|-----------------|-------|-------|---------|--|------|------------|--------|----|--|
| Formaldehyde    | 0.288 | ug/ml | 0.30000 |  | 96.1 | 79.9-120.1 | 0.183  | 20 |  |
| Acetaldehyde    | 0.295 | ug/ml | 0.30000 |  | 98.3 | 79.9-120.1 | 0.0712 | 20 |  |
| Acetone         | 0.308 | ug/ml | 0.30000 |  | 103  | 69.9-130.1 | 0.763  | 20 |  |
| Propionaldehyde | 0.344 | ug/ml | 0.30000 |  | 115  | 69.9-130.1 | 0.714  | 20 |  |
| Crotonaldehyde  | 0.277 | ug/ml | 0.30000 |  | 92.4 | 69.9-130.1 | 0.286  | 20 |  |
| Butyraldehyde   | 0.257 | ug/ml | 0.30000 |  | 85.8 | 69.9-130.1 | 1.17   | 20 |  |
| 2-Butanone      | 0.241 | ug/ml | 0.30000 |  | 80.2 | 69.9-130.1 | 0.0241 | 20 |  |
| Benzaldehyde    | 0.314 | ug/ml | 0.30000 |  | 105  | 69.9-130.1 | 1.35   | 20 |  |
| Valeraldehyde   | 0.254 | ug/ml | 0.30000 |  | 84.7 | 69.9-130.1 | 3.04   | 20 |  |
| Hexaldehyde     | 0.252 | ug/ml | 0.30000 |  | 84.1 | 69.9-130.1 | 0.378  | 20 |  |

### Duplicate (B6C2702-DUP7)

Source: 6031817-04 Prepared: 03/27/26 Analyzed: 04/06/26

|                 |        |                       |  |        |  |  |        |      |       |
|-----------------|--------|-----------------------|--|--------|--|--|--------|------|-------|
| Formaldehyde    | 2.09   | ug/m <sup>3</sup> Air |  | 2.09   |  |  | 0.201  | 10.1 |       |
| Acetaldehyde    | 1.23   | ug/m <sup>3</sup> Air |  | 1.22   |  |  | 0.531  | 10.1 |       |
| Acetone         | 2.17   | ug/m <sup>3</sup> Air |  | 2.18   |  |  | 0.454  | 10.1 |       |
| Propionaldehyde | 0.386  | ug/m <sup>3</sup> Air |  | 0.370  |  |  | 4.28   | 10.1 | QB-01 |
| Crotonaldehyde  | 0.0210 | ug/m <sup>3</sup> Air |  | 0.0194 |  |  | 8.09   | 10.1 |       |
| Butyraldehyde   | 0.151  | ug/m <sup>3</sup> Air |  | 0.151  |  |  | 0.0318 | 10.1 |       |
| 2-Butanone      | 1.43   | ug/m <sup>3</sup> Air |  | 1.41   |  |  | 1.17   | 10.1 |       |
| Benzaldehyde    | 0.0826 | ug/m <sup>3</sup> Air |  | 0.0818 |  |  | 0.894  | 10.1 |       |
| Valeraldehyde   | 0.138  | ug/m <sup>3</sup> Air |  | 0.130  |  |  | 5.54   | 10.1 |       |
| Hexaldehyde     | 0.0922 | ug/m <sup>3</sup> Air |  | 0.0802 |  |  | 14.0   | 10.1 |       |

### Duplicate (B6C2702-DUP8)

Source: 6031817-05 Prepared: 03/27/26 Analyzed: 04/06/26

|                 |        |                       |  |        |  |  |        |      |       |
|-----------------|--------|-----------------------|--|--------|--|--|--------|------|-------|
| Formaldehyde    | 2.06   | ug/m <sup>3</sup> Air |  | 2.06   |  |  | 0.0328 | 10.1 |       |
| Acetaldehyde    | 1.20   | ug/m <sup>3</sup> Air |  | 1.20   |  |  | 0.347  | 10.1 |       |
| Acetone         | 2.15   | ug/m <sup>3</sup> Air |  | 2.15   |  |  | 0.125  | 10.1 |       |
| Propionaldehyde | 0.388  | ug/m <sup>3</sup> Air |  | 0.369  |  |  | 5.08   | 10.1 | QB-01 |
| Crotonaldehyde  | 0.0172 | ug/m <sup>3</sup> Air |  | 0.0247 |  |  | 35.8   | 10.1 |       |
| Butyraldehyde   | 0.144  | ug/m <sup>3</sup> Air |  | 0.147  |  |  | 2.05   | 10.1 |       |
| 2-Butanone      | 1.40   | ug/m <sup>3</sup> Air |  | 1.39   |  |  | 0.485  | 10.1 |       |
| Benzaldehyde    | 0.0755 | ug/m <sup>3</sup> Air |  | 0.0810 |  |  | 6.97   | 10.1 |       |
| Valeraldehyde   | 0.134  | ug/m <sup>3</sup> Air |  | 0.145  |  |  | 7.95   | 10.1 |       |
| Hexaldehyde     | 0.0852 | ug/m <sup>3</sup> Air |  | 0.0894 |  |  | 4.83   | 10.1 |       |

Batch B6C3010 - TO-11A extraction

### Blank (B6C3010-BLK1)

Prepared: 03/30/26 Analyzed: 04/06/26

|              |        |                       |  |  |  |  |  |  |   |
|--------------|--------|-----------------------|--|--|--|--|--|--|---|
| Formaldehyde | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |
| Acetaldehyde | 0.0334 | ug/m <sup>3</sup> Air |  |  |  |  |  |  |   |
| Acetone      | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6C3010 - TO-11A extraction

### Blank (B6C3010-BLK1) Continued

Prepared: 03/30/26 Analyzed: 04/06/26

|                 |        |                       |  |  |  |  |  |  |          |
|-----------------|--------|-----------------------|--|--|--|--|--|--|----------|
| Propionaldehyde | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | QB-01, U |
| Crotonaldehyde  | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Butyraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| 2-Butanone      | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Benzaldehyde    | 0.0600 | ug/m <sup>3</sup> Air |  |  |  |  |  |  |          |
| Valeraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Hexaldehyde     | 0.0227 | ug/m <sup>3</sup> Air |  |  |  |  |  |  |          |

### LCS (B6C3010-BS1)

Prepared: 03/30/26 Analyzed: 04/06/26

|                 |       |       |         |      |            |  |  |  |
|-----------------|-------|-------|---------|------|------------|--|--|--|
| Formaldehyde    | 0.300 | ug/ml | 0.30000 | 99.9 | 79.9-120.1 |  |  |  |
| Acetaldehyde    | 0.296 | ug/ml | 0.30000 | 98.7 | 79.9-120.1 |  |  |  |
| Acetone         | 0.308 | ug/ml | 0.30000 | 103  | 69.9-130.1 |  |  |  |
| Propionaldehyde | 0.369 | ug/ml | 0.30000 | 123  | 69.9-130.1 |  |  |  |
| Crotonaldehyde  | 0.290 | ug/ml | 0.30000 | 96.6 | 69.9-130.1 |  |  |  |
| Butyraldehyde   | 0.265 | ug/ml | 0.30000 | 88.4 | 69.9-130.1 |  |  |  |
| 2-Butanone      | 0.242 | ug/ml | 0.30000 | 80.7 | 69.9-130.1 |  |  |  |
| Benzaldehyde    | 0.321 | ug/ml | 0.30000 | 107  | 69.9-130.1 |  |  |  |
| Valeraldehyde   | 0.265 | ug/ml | 0.30000 | 88.4 | 69.9-130.1 |  |  |  |
| Hexaldehyde     | 0.256 | ug/ml | 0.30000 | 85.5 | 69.9-130.1 |  |  |  |

### LCS Dup (B6C3010-BSD1)

Prepared: 03/30/26 Analyzed: 04/06/26

|                 |       |       |         |      |            |       |    |  |
|-----------------|-------|-------|---------|------|------------|-------|----|--|
| Formaldehyde    | 0.303 | ug/ml | 0.30000 | 101  | 79.9-120.1 | 1.08  | 20 |  |
| Acetaldehyde    | 0.298 | ug/ml | 0.30000 | 99.5 | 79.9-120.1 | 0.787 | 20 |  |
| Acetone         | 0.310 | ug/ml | 0.30000 | 103  | 69.9-130.1 | 0.634 | 20 |  |
| Propionaldehyde | 0.367 | ug/ml | 0.30000 | 122  | 69.9-130.1 | 0.395 | 20 |  |
| Crotonaldehyde  | 0.293 | ug/ml | 0.30000 | 97.5 | 69.9-130.1 | 0.939 | 20 |  |
| Butyraldehyde   | 0.267 | ug/ml | 0.30000 | 89.0 | 69.9-130.1 | 0.665 | 20 |  |
| 2-Butanone      | 0.243 | ug/ml | 0.30000 | 80.9 | 69.9-130.1 | 0.280 | 20 |  |
| Benzaldehyde    | 0.328 | ug/ml | 0.30000 | 109  | 69.9-130.1 | 2.16  | 20 |  |
| Valeraldehyde   | 0.275 | ug/ml | 0.30000 | 91.6 | 69.9-130.1 | 3.54  | 20 |  |
| Hexaldehyde     | 0.259 | ug/ml | 0.30000 | 86.3 | 69.9-130.1 | 0.987 | 20 |  |

Batch B6D0203 - TO-11A extraction

### Blank (B6D0203-BLK1)

Prepared & Analyzed: 04/02/26

|                 |    |                       |  |  |  |  |  |          |
|-----------------|----|-----------------------|--|--|--|--|--|----------|
| Formaldehyde    | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | U        |
| Acetaldehyde    | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | U        |
| Acetone         | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | U        |
| Propionaldehyde | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | QB-01, U |
| Crotonaldehyde  | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | U        |
| Butyraldehyde   | ND | ug/m <sup>3</sup> Air |  |  |  |  |  | U        |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D0203 - TO-11A extraction

### Blank (B6D0203-BLK1) Continued

Prepared & Analyzed: 04/02/26

|               |        |                       |  |  |  |  |  |  |   |
|---------------|--------|-----------------------|--|--|--|--|--|--|---|
| 2-Butanone    | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |
| Benzaldehyde  | 0.0505 | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |
| Valeraldehyde | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |
| Hexaldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U |

### LCS (B6D0203-BS1)

Prepared & Analyzed: 04/02/26

|                 |       |       |         |  |      |            |  |  |  |
|-----------------|-------|-------|---------|--|------|------------|--|--|--|
| Formaldehyde    | 0.298 | ug/ml | 0.30000 |  | 99.4 | 79.9-120.1 |  |  |  |
| Acetaldehyde    | 0.298 | ug/ml | 0.30000 |  | 99.4 | 79.9-120.1 |  |  |  |
| Acetone         | 0.304 | ug/ml | 0.30000 |  | 101  | 69.9-130.1 |  |  |  |
| Propionaldehyde | 0.361 | ug/ml | 0.30000 |  | 120  | 69.9-130.1 |  |  |  |
| Crotonaldehyde  | 0.297 | ug/ml | 0.30000 |  | 98.9 | 69.9-130.1 |  |  |  |
| Butyraldehyde   | 0.266 | ug/ml | 0.30000 |  | 88.7 | 69.9-130.1 |  |  |  |
| 2-Butanone      | 0.238 | ug/ml | 0.30000 |  | 79.4 | 69.9-130.1 |  |  |  |
| Benzaldehyde    | 0.321 | ug/ml | 0.30000 |  | 107  | 69.9-130.1 |  |  |  |
| Valeraldehyde   | 0.271 | ug/ml | 0.30000 |  | 90.3 | 69.9-130.1 |  |  |  |
| Hexaldehyde     | 0.255 | ug/ml | 0.30000 |  | 85.0 | 69.9-130.1 |  |  |  |

### LCS Dup (B6D0203-BSD1)

Prepared & Analyzed: 04/02/26

|                 |       |       |         |  |      |            |       |    |  |
|-----------------|-------|-------|---------|--|------|------------|-------|----|--|
| Formaldehyde    | 0.308 | ug/ml | 0.30000 |  | 103  | 79.9-120.1 | 3.28  | 20 |  |
| Acetaldehyde    | 0.301 | ug/ml | 0.30000 |  | 100  | 79.9-120.1 | 0.958 | 20 |  |
| Acetone         | 0.307 | ug/ml | 0.30000 |  | 102  | 69.9-130.1 | 0.896 | 20 |  |
| Propionaldehyde | 0.363 | ug/ml | 0.30000 |  | 121  | 69.9-130.1 | 0.399 | 20 |  |
| Crotonaldehyde  | 0.300 | ug/ml | 0.30000 |  | 99.8 | 69.9-130.1 | 0.940 | 20 |  |
| Butyraldehyde   | 0.269 | ug/ml | 0.30000 |  | 89.7 | 69.9-130.1 | 1.05  | 20 |  |
| 2-Butanone      | 0.244 | ug/ml | 0.30000 |  | 81.2 | 69.9-130.1 | 2.23  | 20 |  |
| Benzaldehyde    | 0.324 | ug/ml | 0.30000 |  | 108  | 69.9-130.1 | 0.836 | 20 |  |
| Valeraldehyde   | 0.273 | ug/ml | 0.30000 |  | 91.1 | 69.9-130.1 | 0.844 | 20 |  |
| Hexaldehyde     | 0.261 | ug/ml | 0.30000 |  | 87.1 | 69.9-130.1 | 2.41  | 20 |  |

### Duplicate (B6D0203-DUP1)

Source: 6040116-03 Prepared & Analyzed: 04/02/26

|                 |        |                       |  |        |  |  |       |      |       |
|-----------------|--------|-----------------------|--|--------|--|--|-------|------|-------|
| Formaldehyde    | 5.91   | ug/m <sup>3</sup> Air |  | 5.87   |  |  | 0.528 | 10.1 |       |
| Acetaldehyde    | 4.48   | ug/m <sup>3</sup> Air |  | 4.44   |  |  | 0.827 | 10.1 |       |
| Acetone         | 3.83   | ug/m <sup>3</sup> Air |  | 3.80   |  |  | 0.687 | 10.1 |       |
| Propionaldehyde | 1.16   | ug/m <sup>3</sup> Air |  | 1.15   |  |  | 0.709 | 10.1 | QB-01 |
| Crotonaldehyde  | 0.0696 | ug/m <sup>3</sup> Air |  | 0.0687 |  |  | 1.24  | 10.1 |       |
| Butyraldehyde   | 0.496  | ug/m <sup>3</sup> Air |  | 0.510  |  |  | 2.89  | 10.1 |       |
| 2-Butanone      | 2.37   | ug/m <sup>3</sup> Air |  | 2.36   |  |  | 0.344 | 10.1 |       |
| Benzaldehyde    | 0.218  | ug/m <sup>3</sup> Air |  | 0.188  |  |  | 14.7  | 10.1 |       |
| Valeraldehyde   | 0.337  | ug/m <sup>3</sup> Air |  | 0.411  |  |  | 19.6  | 10.1 |       |
| Hexaldehyde     | 0.183  | ug/m <sup>3</sup> Air |  | 0.173  |  |  | 6.13  | 10.1 |       |

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



# CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 10:06

SUBMITTED: 03/18/26 to 04/01/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

| Analyte | Result | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

## Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D0608 - TO-11A extraction

### Blank (B6D0608-BLK1)

Prepared: 04/06/26 Analyzed: 04/07/26

|                 |        |                       |  |  |  |  |  |  |          |
|-----------------|--------|-----------------------|--|--|--|--|--|--|----------|
| Formaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Acetaldehyde    | 0.0264 | ug/m <sup>3</sup> Air |  |  |  |  |  |  |          |
| Acetone         | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Propionaldehyde | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | QB-01, U |
| Crotonaldehyde  | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Butyraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| 2-Butanone      | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Benzaldehyde    | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Valeraldehyde   | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |
| Hexaldehyde     | ND     | ug/m <sup>3</sup> Air |  |  |  |  |  |  | U        |

### LCS (B6D0608-BS1)

Prepared: 04/06/26 Analyzed: 04/07/26

|                 |       |       |         |      |            |  |  |  |  |
|-----------------|-------|-------|---------|------|------------|--|--|--|--|
| Formaldehyde    | 0.297 | ug/ml | 0.30000 | 98.8 | 79.9-120.1 |  |  |  |  |
| Acetaldehyde    | 0.292 | ug/ml | 0.30000 | 97.2 | 79.9-120.1 |  |  |  |  |
| Acetone         | 0.298 | ug/ml | 0.30000 | 99.5 | 69.9-130.1 |  |  |  |  |
| Propionaldehyde | 0.347 | ug/ml | 0.30000 | 116  | 69.9-130.1 |  |  |  |  |
| Crotonaldehyde  | 0.291 | ug/ml | 0.30000 | 97.0 | 69.9-130.1 |  |  |  |  |
| Butyraldehyde   | 0.263 | ug/ml | 0.30000 | 87.6 | 69.9-130.1 |  |  |  |  |
| 2-Butanone      | 0.236 | ug/ml | 0.30000 | 78.7 | 69.9-130.1 |  |  |  |  |
| Benzaldehyde    | 0.317 | ug/ml | 0.30000 | 106  | 69.9-130.1 |  |  |  |  |
| Valeraldehyde   | 0.265 | ug/ml | 0.30000 | 88.3 | 69.9-130.1 |  |  |  |  |
| Hexaldehyde     | 0.251 | ug/ml | 0.30000 | 83.8 | 69.9-130.1 |  |  |  |  |

### LCS Dup (B6D0608-BSD1)

Prepared: 04/06/26 Analyzed: 04/07/26

|                 |       |       |         |      |            |       |    |  |  |
|-----------------|-------|-------|---------|------|------------|-------|----|--|--|
| Formaldehyde    | 0.290 | ug/ml | 0.30000 | 96.7 | 79.9-120.1 | 2.20  | 20 |  |  |
| Acetaldehyde    | 0.289 | ug/ml | 0.30000 | 96.4 | 79.9-120.1 | 0.816 | 20 |  |  |
| Acetone         | 0.297 | ug/ml | 0.30000 | 98.9 | 69.9-130.1 | 0.601 | 20 |  |  |
| Propionaldehyde | 0.338 | ug/ml | 0.30000 | 113  | 69.9-130.1 | 2.73  | 20 |  |  |
| Crotonaldehyde  | 0.286 | ug/ml | 0.30000 | 95.3 | 69.9-130.1 | 1.70  | 20 |  |  |
| Butyraldehyde   | 0.255 | ug/ml | 0.30000 | 84.9 | 69.9-130.1 | 3.11  | 20 |  |  |
| 2-Butanone      | 0.234 | ug/ml | 0.30000 | 77.9 | 69.9-130.1 | 1.05  | 20 |  |  |
| Benzaldehyde    | 0.310 | ug/ml | 0.30000 | 103  | 69.9-130.1 | 2.22  | 20 |  |  |
| Valeraldehyde   | 0.261 | ug/ml | 0.30000 | 87.1 | 69.9-130.1 | 1.42  | 20 |  |  |
| Hexaldehyde     | 0.249 | ug/ml | 0.30000 | 83.0 | 69.9-130.1 | 1.02  | 20 |  |  |



## CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

**ATTN:** Mr. Bryan Sims

**PHONE:** (405) 702-0100

**FILE #:** 0475

**REPORTED:** 05/14/26 10:06

**SUBMITTED:** 03/18/26 to 04/01/26

**AQS SITE CODE:** 40 143 1127

**SITE CODE:** TMOK

### Notes and Definitions

|       |                                                                                   |
|-------|-----------------------------------------------------------------------------------|
| U     | Under Detection Limit                                                             |
| TT    | Transport temperature is out of specs.                                            |
| S-Rec | Sample received by laboratory outside of initial extraction hold time - 14 days - |
| QB-01 | Analyte exceeds method blank criteria                                             |
| D-F   | Duplicate exceeds DQO criteria.                                                   |
| CE    | Not reportable due to a co-eluting compound.                                      |
| ND    | Analyte NOT DETECTED                                                              |
| NR    | Not Reported                                                                      |
| MDL   | Method Detection Limit                                                            |
| RPD   | Relative Percent Difference                                                       |

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.