



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

May 14, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 03/18/26 11:17 through 04/06/26 10:13.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6031817-01	Air	03/13/26 23:59	03/18/26 11:17
TMOK	6032352-01	Air	03/17/26 23:59	03/23/26 11:07
TMOK	6032352-02	Air	03/18/26 23:59	03/23/26 11:07
TMOK	6040116-01	Air	03/21/26 23:59	04/01/26 13:04
TMOK	6040116-02	Air	03/27/26 23:59	04/01/26 13:04
TMOK	6040617-01	Air	03/31/26 23:59	04/06/26 10:13



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6031817-01

Sampled: 03/13/26 23:59

Pressure @ Receipt: 4.5

Canister #: 43398

Received: 03/18/26 11:17

Comments: Makeup for 3-9-26

Analyzed: 03/19/26 21:44

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.499	2.47		0.0237
Chloromethane	0.496	1.03		0.0213
Dichlorotetrafluoroethane	0.0171	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00360	0.01	U, CF, LL	0.0215
Bromomethane	0.00750	0.03	U	0.0104
Chloroethane	0.00710	0.02	U	0.0136
Acetonitrile	180	303.00	E, QB-03	0.184
Acrolein	0.132	0.30	CF, LK	0.0977
Trichlorofluoromethane	0.226	1.27		0.00950
Acrylonitrile	0.0143	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.359	1.25		0.0489
Carbon Disulfide	0.00520	0.02	U	0.0473
Trichlorotrifluoroethane	0.0701	0.54		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	9.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	0.00460	0.02	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0163	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0208	0.08		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.120	0.38		0.0158
Carbon Tetrachloride	0.0734	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00450	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0141	0.06	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6031817-01

Sampled: 03/13/26 23:59

Pressure @ Receipt: 4.5

Canister #: 43398

Received: 03/18/26 11:17

Comments: Makeup for 3-9-26

Analyzed: 03/19/26 21:44

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.221	0.84		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0341	0.16		0.0258
Tetrachloroethylene	0.00420	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0198	0.09		0.0130
m,p-Xylene	0.0526	0.23		0.0418
Bromoform	0.00220	0.02	U	0.0166
Styrene	0.0146	0.06	U	0.0273
1,1,2,2-Tetrachloroethane	6.00E-4	0.00	U	0.0190
o-Xylene	0.0213	0.09	U	0.0235
1,3,5-Trimethylbenzene	0.00430	0.02	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	5.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00250	0.02	U	0.0287
o-Dichlorobenzene	7.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00300	0.02	U, QB-01	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6032352-01

Sampled: 03/17/26 23:59

Pressure @ Receipt: 3.5

Canister #: 35153

Received: 03/23/26 11:07

Comments: Makeup for 2-19-26

Analyzed: 03/26/26 20:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.540	2.68		0.0237
Chloromethane	0.486	1.01		0.0213
Dichlorotetrafluoroethane	0.0189	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0216	0.05		0.0215
Bromomethane	0.00830	0.03	U	0.0104
Chloroethane	0.00620	0.02	U	0.0136
Acetonitrile	204	343.00	E	0.184
Acrolein	0.0912	0.21	U	0.0977
Trichlorofluoromethane	0.242	1.36		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.148	0.52		0.0489
Carbon Disulfide	0.0371	0.12	LK, U	0.0473
Trichlorotrifluoroethane	0.0806	0.62		0.0210
trans-1,2-Dichloroethylene	0.00160	0.01	U	0.0171
1,1-Dichloroethane	0.00120	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0205	0.10		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0200	0.08		0.0116
1,1,1-Trichloroethane	0.00130	0.01	U	0.00952
Benzene	0.270	0.86		0.0158
Carbon Tetrachloride	0.0722	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00610	0.03	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0105	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6032352-01

Sampled: 03/17/26 23:59

Pressure @ Receipt: 3.5

Canister #: 35153

Received: 03/23/26 11:07

Comments: Makeup for 2-19-26

Analyzed: 03/26/26 20:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.428	1.62		0.0380
Dibromochloromethane	ND	ND	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0533	0.25		0.0258
Tetrachloroethylene	0.00690	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0538	0.23		0.0130
m,p-Xylene	0.183	0.80		0.0418
Bromoform	0.00280	0.03	U	0.0166
Styrene	0.0309	0.13		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0696	0.30		0.0235
1,3,5-Trimethylbenzene	0.0170	0.08	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00470	0.03	U	0.0287
o-Dichlorobenzene	7.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00140	0.01	U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6032352-02

Sampled: 03/18/26 23:59

Pressure @ Receipt: 3.0

Canister #: 43346

Received: 03/23/26 11:07

Comments: Makeup for 2-7-26

Analyzed: 03/25/26 17:36

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.512	2.54		0.0237
Chloromethane	0.491	1.02		0.0213
Dichlorotetrafluoroethane	0.0183	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0228	0.05	CF, LL	0.0215
Bromomethane	0.00850	0.03	U	0.0104
Chloroethane	0.00840	0.02	U	0.0136
Acetonitrile	143	241.00		0.184
Acrolein	0.131	0.30	CF, LK	0.0977
Trichlorofluoromethane	0.232	1.31		0.00950
Acrylonitrile	0.0117	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.159	0.55		0.0489
Carbon Disulfide	ND	ND	LK, U	0.0473
Trichlorotrifluoroethane	0.0714	0.55		0.0210
trans-1,2-Dichloroethylene	0.00190	0.01	U	0.0171
1,1-Dichloroethane	0.00140	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	6.00E-4	0.00	U	0.00905
Chloroform	0.0226	0.11		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0184	0.07		0.0116
1,1,1-Trichloroethane	0.00160	0.01	U	0.00952
Benzene	0.272	0.87		0.0158
Carbon Tetrachloride	0.0693	0.44		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00570	0.03	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0163	0.07	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6032352-02

Sampled: 03/18/26 23:59

Pressure @ Receipt: 3.0

Canister #: 43346

Received: 03/23/26 11:07

Comments: Makeup for 2-7-26

Analyzed: 03/25/26 17:36

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.899	3.39		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.124	0.58		0.0258
Tetrachloroethylene	0.00830	0.06	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.122	0.53		0.0130
m,p-Xylene	0.442	1.92		0.0418
Bromoform	0.00180	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.167	0.73		0.0235
1,3,5-Trimethylbenzene	0.0360	0.18		0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	0.00170	0.01	U	0.0277
p-Dichlorobenzene	0.00710	0.04	U	0.0287
o-Dichlorobenzene	0.00190	0.01	U	0.0364
1,2,4-Trichlorobenzene	0.00640	0.05	QB-01, U	0.0792
Hexachloro-1,3-butadiene	0.00130	0.01	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040116-01

Sampled: 03/21/26 23:59

Pressure @ Receipt: 4.50

Canister #: 912392

Received: 04/01/26 13:04

Comments:

Analyzed: 04/15/26 19:50

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.525	2.60		0.0237
Chloromethane	0.508	1.05		0.0213
Dichlorotetrafluoroethane	0.0180	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0166	0.04	CF, LL, U	0.0215
Bromomethane	0.0101	0.04	U	0.0104
Chloroethane	0.00690	0.02	U	0.0136
Acetonitrile	160	269.00	E, QB-03	0.184
Acrolein	0.260	0.60	CF, LK	0.0977
Trichlorofluoromethane	0.224	1.26		0.00950
Acrylonitrile	0.0239	0.05		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.228	0.79	QB-01	0.0489
Carbon Disulfide	0.00780	0.02	U	0.0473
Trichlorotrifluoroethane	0.0744	0.57		0.0210
trans-1,2-Dichloroethylene	0.00820	0.03	U	0.0171
1,1-Dichloroethane	7.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0221	0.11		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0148	0.06		0.0116
1,1,1-Trichloroethane	0.00160	0.01	U	0.00952
Benzene	0.466	1.49		0.0158
Carbon Tetrachloride	0.0749	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00320	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0230	0.09	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040116-01

Sampled: 03/21/26 23:59

Pressure @ Receipt: 4.50

Canister #: 912392

Received: 04/01/26 13:04

Comments:

Analyzed: 04/15/26 19:50

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	2.18	8.23		0.0380
Dibromochloromethane	ND	ND	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.452	2.12		0.0258
Tetrachloroethylene	0.0152	0.10		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0978	0.43		0.0130
m,p-Xylene	0.343	1.49		0.0418
Bromoform	0.00200	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.153	0.67		0.0235
1,3,5-Trimethylbenzene	0.0299	0.15	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	5.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00590	0.04	U	0.0287
o-Dichlorobenzene	8.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00240	0.02	QB-01, U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040116-02

Sampled: 03/27/26 23:59

Pressure @ Receipt: 4.50

Canister #: 912410

Received: 04/01/26 13:04

Comments:

Analyzed: 04/16/26 00:04

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.520	2.58		0.0237
Chloromethane	0.469	0.97		0.0213
Dichlorotetrafluoroethane	0.0179	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00590	0.01	U	0.0215
Bromomethane	0.00730	0.03	U	0.0104
Chloroethane	0.00520	0.01	U	0.0136
Acetonitrile	155	261.00	E, QB-03	0.184
Acrolein	0.103	0.24	CF, LK	0.0977
Trichlorofluoromethane	0.219	1.23		0.00950
Acrylonitrile	0.0114	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.163	0.57	QB-01	0.0489
Carbon Disulfide	0.00340	0.01	U	0.0473
Trichlorotrifluoroethane	0.0736	0.57		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	0.00100	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0164	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0184	0.07		0.0116
1,1,1-Trichloroethane	0.00120	0.01	U	0.00952
Benzene	0.115	0.37		0.0158
Carbon Tetrachloride	0.0733	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00260	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0107	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040116-02

Sampled: 03/27/26 23:59

Pressure @ Receipt: 4.50

Canister #: 912410

Received: 04/01/26 13:04

Comments:

Analyzed: 04/16/26 00:04

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.128	0.48		0.0380
Dibromochloromethane	2.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0202	0.09	U	0.0258
Tetrachloroethylene	0.00520	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0159	0.07		0.0130
m,p-Xylene	0.0506	0.22		0.0418
Bromoform	0.00220	0.02	U	0.0166
Styrene	0.0106	0.05	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0241	0.11		0.0235
1,3,5-Trimethylbenzene	0.00350	0.02	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	7.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00220	0.01	U	0.0287
o-Dichlorobenzene	ND	ND	U	0.0364
1,2,4-Trichlorobenzene	0.00140	0.01	QB-01, U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040617-01

Sampled: 03/31/26 23:59

Pressure @ Receipt: 3.5

Canister #: 43324

Received: 04/06/26 10:13

Comments: Make up for 1-2-26

Analyzed: 04/21/26 14:53

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.514	2.55		0.0237
Chloromethane	0.491	1.02		0.0213
Dichlorotetrafluoroethane	0.0185	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00380	0.01	CF, LL, U	0.0215
Bromomethane	0.00910	0.04	U	0.0104
Chloroethane	0.00920	0.02	U	0.0136
Acetonitrile	271	456.00	E	0.184
Acrolein	0.202	0.46	CF, LK	0.0977
Trichlorofluoromethane	0.238	1.34		0.00950
Acrylonitrile	0.0145	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.618	2.15		0.0489
Carbon Disulfide	0.00710	0.02	U	0.0473
Trichlorotrifluoroethane	0.0800	0.61		0.0210
trans-1,2-Dichloroethylene	6.00E-4	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0159	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0135	0.05		0.0116
1,1,1-Trichloroethane	0.00120	0.01	U	0.00952
Benzene	0.104	0.33		0.0158
Carbon Tetrachloride	0.0776	0.49		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	5.00E-4	0.00	U	0.00889
Trichloroethylene	0.00480	0.03	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0102	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6040617-01

Sampled: 03/31/26 23:59

Pressure @ Receipt: 3.5

Canister #: 43324

Received: 04/06/26 10:13

Comments: Make up for 1-2-26

Analyzed: 04/21/26 14:53

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.224	0.85		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0259	0.12		0.0258
Tetrachloroethylene	0.00470	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0349	0.15		0.0130
m,p-Xylene	0.104	0.45		0.0418
Bromoform	0.00140	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0395	0.17		0.0235
1,3,5-Trimethylbenzene	0.00680	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00210	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	4.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C1905 - Summa Canister Prep

Blank (B6C1905-BLK1)

Prepared: 03/04/26 Analyzed: 03/19/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C1905 - Summa Canister Prep

Blank (B6C1905-BLK1) Continued

Prepared: 03/04/26 Analyzed: 03/19/26

1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6C2508 - Summa Canister Prep

Blank (B6C2508-BLK1)

Prepared: 03/18/26 Analyzed: 03/25/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				LK, U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2508 - Summa Canister Prep

Blank (B6C2508-BLK1) Continued

Prepared: 03/18/26 Analyzed: 03/25/26

Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6C2602 - Summa Canister Prep

Blank (B6C2602-BLK1)

Prepared: 03/18/26 Analyzed: 03/26/26

Dichlorodifluoromethane	ND	ppbv				U
-------------------------	----	------	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2602 - Summa Canister Prep

Blank (B6C2602-BLK1) Continued

Prepared: 03/18/26 Analyzed: 03/26/26

Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				LK, U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2602 - Summa Canister Prep

Blank (B6C2602-BLK1) Continued

Prepared: 03/18/26 Analyzed: 03/26/26

n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D1504 - Summa Canister Prep

Blank (B6D1504-BLK1)

Prepared: 04/07/26 Analyzed: 04/15/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				QB-01, U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D1504 - Summa Canister Prep

Blank (B6D1504-BLK1) Continued

Prepared: 04/07/26 Analyzed: 04/15/26

Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D2107 - Summa Canister Prep

Blank (B6D2107-BLK1)

Prepared: 04/17/26 Analyzed: 04/21/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2107 - Summa Canister Prep

Blank (B6D2107-BLK1) Continued

Prepared: 04/17/26 Analyzed: 04/21/26

Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 05/14/26 15:18
SUBMITTED: 03/18/26 to 04/06/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2107 - Summa Canister Prep

Blank (B6D2107-BLK1) Continued

Prepared: 04/17/26 Analyzed: 04/21/26

Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603065

Calibration Check (2603065-CCV1)

Prepared & Analyzed: 03/19/26

Dichlorodifluoromethane	1.03	ppbv	0.6	30.00
Chloromethane	1.00	ppbv	-1.5	30.00
Dichlorotetrafluoroethane	1.02	ppbv	-1.6	30.00
Vinyl chloride	0.998	ppbv	-3.3	30.00
1,3-Butadiene	0.962	ppbv	-6.9	30.00
Bromomethane	0.995	ppbv	-3.6	30.00
Chloroethane	1.04	ppbv	-0.07	30.00
Acetonitrile	1.10	ppbv	5.6	30.00
Acrolein	1.09	ppbv	6.8	30.00
Trichlorofluoromethane	1.08	ppbv	6.5	30.00
Acrylonitrile	1.17	ppbv	6.2	30.00
1,1-Dichloroethene	1.11	ppbv	7.0	30.00
Dichloromethane	1.10	ppbv	6.5	30.00
Carbon Disulfide	1.09	ppbv	0.4	30.00
Trichlorotrifluoroethane	1.06	ppbv	3.1	30.00
trans-1,2-Dichloroethylene	1.08	ppbv	3.5	30.00
1,1-Dichloroethane	1.11	ppbv	7.1	30.00
Methyl tert-Butyl Ether	1.14	ppbv	10.1	30.00
Chloroprene	1.16	ppbv	11.9	30.00
cis-1,2-Dichloroethylene	1.14	ppbv	9.8	30.00
Bromochloromethane	1.06	ppbv	3.1	30.00
Chloroform	1.12	ppbv	6.7	30.00
Ethyl tert-Butyl Ether	1.16	ppbv	12.2	30.00
1,2-Dichloroethane	1.14	ppbv	10.3	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.9	30.00
Benzene	1.10	ppbv	6.4	30.00
Carbon Tetrachloride	1.09	ppbv	6.3	30.00
tert-Amyl Methyl Ether	1.14	ppbv	9.9	30.00
1,2-Dichloropropane	1.11	ppbv	7.1	30.00
Ethyl Acrylate	1.14	ppbv	9.2	30.00
Bromodichloromethane	1.11	ppbv	7.2	30.00
Trichloroethylene	1.06	ppbv	3.0	30.00
Methyl Methacrylate	1.13	ppbv	8.9	30.00
cis-1,3-Dichloropropene	1.14	ppbv	9.1	30.00
Methyl Isobutyl Ketone	1.14	ppbv	10.0	30.00
trans-1,3-Dichloropropene	1.15	ppbv	9.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603065

Calibration Check (2603065-CCV1) Continued

Prepared & Analyzed: 03/19/26

1,1,2-Trichloroethane	1.09	ppbv	4.1	30.00	
Toluene	1.12	ppbv	8.1	30.00	
Dibromochloromethane	1.08	ppbv	4.8	30.00	
1,2-Dibromoethane	1.10	ppbv	5.0	30.00	
n-Octane	1.16	ppbv	11.6	30.00	
Tetrachloroethylene	1.08	ppbv	3.8	30.00	
Chlorobenzene	1.10	ppbv	4.5	30.00	
Ethylbenzene	1.13	ppbv	7.9	30.00	
m,p-Xylene	2.31	ppbv	10.7	30.00	
Bromoform	1.10	ppbv	5.0	30.00	
Styrene	1.10	ppbv	4.2	30.00	
1,1,2,2-Tetrachloroethane	1.11	ppbv	5.5	30.00	
o-Xylene	1.15	ppbv	9.1	30.00	
1,3,5-Trimethylbenzene	1.05	ppbv	0.01	30.00	
1,2,4-Trimethylbenzene	1.09	ppbv	3.6	30.00	
m-Dichlorobenzene	1.12	ppbv	7.4	30.00	
p-Dichlorobenzene	1.12	ppbv	7.5	30.00	
o-Dichlorobenzene	1.11	ppbv	6.4	30.00	
1,2,4-Trichlorobenzene	1.03	ppbv	-0.3	30.00	QB-01
Hexachloro-1,3-butadiene	1.01	ppbv	-4.4	30.00	

Calibration Check (2603065-CCV2)

Prepared: 03/19/26 Analyzed: 03/20/26

Dichlorodifluoromethane	1.14	ppbv	10.4	30.00	
Chloromethane	1.12	ppbv	10.0	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	6.8	30.00	
Vinyl chloride	1.06	ppbv	3.0	30.00	
1,3-Butadiene	0.954	ppbv	-7.7	30.00	
Bromomethane	1.08	ppbv	4.3	30.00	
Chloroethane	1.14	ppbv	9.8	30.00	
Acetonitrile	1.16	ppbv	11.6	30.00	
Acrolein	1.12	ppbv	10.2	30.00	
Trichlorofluoromethane	1.15	ppbv	13.7	30.00	
Acrylonitrile	1.15	ppbv	4.2	30.00	
1,1-Dichloroethene	1.18	ppbv	13.0	30.00	
Dichloromethane	1.16	ppbv	11.9	30.00	
Carbon Disulfide	1.18	ppbv	8.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603065

Calibration Check (2603065-CCV2) Continued

Prepared: 03/19/26 Analyzed: 03/20/26

Trichlorotrifluoroethane	1.15	ppbv	11.8	30.00
trans-1,2-Dichloroethylene	1.05	ppbv	0.8	30.00
1,1-Dichloroethane	1.14	ppbv	10.5	30.00
Methyl tert-Butyl Ether	1.02	ppbv	-1.4	30.00
Chloroprene	1.00	ppbv	-3.2	30.00
cis-1,2-Dichloroethylene	1.11	ppbv	7.2	30.00
Bromochloromethane	1.04	ppbv	1.2	30.00
Chloroform	1.14	ppbv	9.0	30.00
Ethyl tert-Butyl Ether	1.00	ppbv	-3.4	30.00
1,2-Dichloroethane	1.18	ppbv	14.6	30.00
1,1,1-Trichloroethane	1.13	ppbv	9.3	30.00
Benzene	1.03	ppbv	-0.6	30.00
Carbon Tetrachloride	1.10	ppbv	7.3	30.00
tert-Amyl Methyl Ether	0.962	ppbv	-7.0	30.00
1,2-Dichloropropane	1.16	ppbv	12.1	30.00
Ethyl Acrylate	1.09	ppbv	4.3	30.00
Bromodichloromethane	1.20	ppbv	15.7	30.00
Trichloroethylene	1.05	ppbv	1.9	30.00
Methyl Methacrylate	1.11	ppbv	6.2	30.00
cis-1,3-Dichloropropene	1.08	ppbv	3.6	30.00
Methyl Isobutyl Ketone	1.15	ppbv	11.0	30.00
trans-1,3-Dichloropropene	1.17	ppbv	11.4	30.00
1,1,2-Trichloroethane	1.15	ppbv	10.0	30.00
Toluene	1.07	ppbv	3.2	30.00
Dibromochloromethane	1.11	ppbv	7.5	30.00
1,2-Dibromoethane	1.11	ppbv	6.4	30.00
n-Octane	1.13	ppbv	9.3	30.00
Tetrachloroethylene	1.06	ppbv	1.7	30.00
Chlorobenzene	1.11	ppbv	5.4	30.00
Ethylbenzene	1.08	ppbv	2.8	30.00
m,p-Xylene	2.24	ppbv	7.2	30.00
Bromoform	1.10	ppbv	5.4	30.00
Styrene	1.05	ppbv	-1.2	30.00
1,1,2,2-Tetrachloroethane	1.18	ppbv	11.8	30.00
o-Xylene	1.06	ppbv	0.6	30.00
1,3,5-Trimethylbenzene	0.942	ppbv	-10.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603065

Calibration Check (2603065-CCV2) Continued

Prepared: 03/19/26 Analyzed: 03/20/26

1,2,4-Trimethylbenzene	0.982	ppbv	-6.9	30.00	QB-01
m-Dichlorobenzene	1.08	ppbv	3.9	30.00	
p-Dichlorobenzene	1.07	ppbv	2.3	30.00	
o-Dichlorobenzene	1.07	ppbv	2.5	30.00	
1,2,4-Trichlorobenzene	1.00	ppbv	-3.5	30.00	
Hexachloro-1,3-butadiene	0.990	ppbv	-6.0	30.00	

Sequence 2603086

Calibration Check (2603086-CCV1)

Prepared & Analyzed: 03/25/26

Dichlorodifluoromethane	1.14	ppbv	11.0	30.00
Chloromethane	1.12	ppbv	9.5	30.00
Dichlorotetrafluoroethane	1.14	ppbv	10.4	30.00
Vinyl chloride	1.11	ppbv	7.9	30.00
1,3-Butadiene	1.14	ppbv	9.8	30.00
Bromomethane	1.13	ppbv	9.5	30.00
Chloroethane	1.12	ppbv	8.3	30.00
Acetonitrile	1.07	ppbv	3.2	30.00
Acrolein	1.04	ppbv	1.9	30.00
Trichlorofluoromethane	1.12	ppbv	11.0	30.00
Acrylonitrile	1.16	ppbv	4.8	30.00
1,1-Dichloroethene	1.14	ppbv	9.8	30.00
Dichloromethane	1.13	ppbv	9.4	30.00
Carbon Disulfide	1.38	ppbv	26.4	30.00
Trichlorotrifluoroethane	1.13	ppbv	10.0	30.00
trans-1,2-Dichloroethylene	1.09	ppbv	4.4	30.00
1,1-Dichloroethane	1.12	ppbv	7.8	30.00
Methyl tert-Butyl Ether	1.11	ppbv	6.9	30.00
Chloroprene	1.10	ppbv	6.1	30.00
cis-1,2-Dichloroethylene	1.12	ppbv	7.7	30.00
Bromochloromethane	1.09	ppbv	5.6	30.00
Chloroform	1.13	ppbv	7.7	30.00
Ethyl tert-Butyl Ether	1.10	ppbv	6.4	30.00
1,2-Dichloroethane	1.12	ppbv	8.2	30.00
1,1,1-Trichloroethane	1.10	ppbv	7.1	30.00
Benzene	1.09	ppbv	4.9	30.00
Carbon Tetrachloride	1.09	ppbv	6.6	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603086

Calibration Check (2603086-CCV1) Continued

Prepared & Analyzed: 03/25/26

tert-Amyl Methyl Ether	1.10	ppbv	6.6	30.00	
1,2-Dichloropropane	1.10	ppbv	5.7	30.00	
Ethyl Acrylate	1.14	ppbv	8.6	30.00	
Bromodichloromethane	1.18	ppbv	13.5	30.00	
Trichloroethylene	1.14	ppbv	9.9	30.00	
Methyl Methacrylate	1.14	ppbv	9.2	30.00	
cis-1,3-Dichloropropene	1.14	ppbv	9.3	30.00	
Methyl Isobutyl Ketone	1.14	ppbv	10.0	30.00	
trans-1,3-Dichloropropene	1.16	ppbv	10.5	30.00	
1,1,2-Trichloroethane	1.15	ppbv	9.9	30.00	
Toluene	1.13	ppbv	8.4	30.00	
Dibromochloromethane	1.13	ppbv	9.7	30.00	
1,2-Dibromoethane	1.09	ppbv	4.6	30.00	
n-Octane	1.13	ppbv	9.0	30.00	
Tetrachloroethylene	1.11	ppbv	6.3	30.00	
Chlorobenzene	1.10	ppbv	4.3	30.00	
Ethylbenzene	1.16	ppbv	10.5	30.00	
m,p-Xylene	2.33	ppbv	11.6	30.00	
Bromoform	1.13	ppbv	8.4	30.00	
Styrene	1.10	ppbv	4.1	30.00	
1,1,2,2-Tetrachloroethane	1.18	ppbv	11.6	30.00	
o-Xylene	1.16	ppbv	10.3	30.00	
1,3,5-Trimethylbenzene	1.11	ppbv	5.2	30.00	
1,2,4-Trimethylbenzene	1.07	ppbv	1.7	30.00	
m-Dichlorobenzene	1.15	ppbv	10.0	30.00	
p-Dichlorobenzene	1.15	ppbv	9.8	30.00	
o-Dichlorobenzene	1.15	ppbv	10.0	30.00	
1,2,4-Trichlorobenzene	1.11	ppbv	6.6	30.00	QB-01
Hexachloro-1,3-butadiene	1.08	ppbv	2.4	30.00	

Calibration Check (2603086-CCV2)

Prepared: 03/25/26 Analyzed: 03/26/26

Dichlorodifluoromethane	1.23	ppbv	19.5	30.00	
Chloromethane	1.13	ppbv	11.3	30.00	
Dichlorotetrafluoroethane	1.19	ppbv	15.0	30.00	
Vinyl chloride	1.11	ppbv	7.6	30.00	
1,3-Butadiene	1.12	ppbv	8.2	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603086

Calibration Check (2603086-CCV2) Continued

Prepared: 03/25/26 Analyzed: 03/26/26

Bromomethane	1.21	ppbv	17.1	30.00	
Chloroethane	1.16	ppbv	11.9	30.00	
Acetonitrile	1.06	ppbv	2.3	30.00	
Acrolein	0.960	ppbv	-5.8	30.00	
Trichlorofluoromethane	1.22	ppbv	20.6	30.00	
Acrylonitrile	1.05	ppbv	-4.9	30.00	
1,1-Dichloroethene	1.17	ppbv	12.8	30.00	
Dichloromethane	1.16	ppbv	12.0	30.00	
Carbon Disulfide	1.42	ppbv	30.9	30.00	LK
Trichlorotrifluoroethane	1.25	ppbv	21.6	30.00	
trans-1,2-Dichloroethylene	1.08	ppbv	3.4	30.00	
1,1-Dichloroethane	1.11	ppbv	6.9	30.00	
Methyl tert-Butyl Ether	1.06	ppbv	2.6	30.00	
Chloroprene	1.02	ppbv	-1.5	30.00	
cis-1,2-Dichloroethylene	1.07	ppbv	3.8	30.00	
Bromochloromethane	1.03	ppbv	0.4	30.00	
Chloroform	1.17	ppbv	11.2	30.00	
Ethyl tert-Butyl Ether	1.01	ppbv	-2.7	30.00	
1,2-Dichloroethane	1.16	ppbv	12.0	30.00	
1,1,1-Trichloroethane	1.18	ppbv	14.9	30.00	
Benzene	1.07	ppbv	2.7	30.00	
Carbon Tetrachloride	1.09	ppbv	6.9	30.00	
tert-Amyl Methyl Ether	1.04	ppbv	0.9	30.00	
1,2-Dichloropropane	1.23	ppbv	18.2	30.00	
Ethyl Acrylate	1.14	ppbv	9.3	30.00	
Bromodichloromethane	1.29	ppbv	24.8	30.00	
Trichloroethylene	1.17	ppbv	13.2	30.00	
Methyl Methacrylate	1.14	ppbv	9.1	30.00	
cis-1,3-Dichloropropene	1.16	ppbv	10.7	30.00	
Methyl Isobutyl Ketone	1.18	ppbv	13.0	30.00	
trans-1,3-Dichloropropene	1.23	ppbv	17.2	30.00	
1,1,2-Trichloroethane	1.27	ppbv	21.1	30.00	
Toluene	1.18	ppbv	13.3	30.00	
Dibromochloromethane	1.14	ppbv	10.4	30.00	
1,2-Dibromoethane	1.20	ppbv	14.8	30.00	
n-Octane	1.10	ppbv	6.1	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603086

Calibration Check (2603086-CCV2) Continued

Prepared: 03/25/26 Analyzed: 03/26/26

Tetrachloroethylene	1.18	ppbv	13.2	30.00	
Chlorobenzene	1.13	ppbv	7.0	30.00	
Ethylbenzene	1.15	ppbv	10.2	30.00	
m,p-Xylene	2.37	ppbv	13.5	30.00	
Bromoform	1.15	ppbv	10.3	30.00	
Styrene	1.13	ppbv	6.8	30.00	
1,1,2,2-Tetrachloroethane	1.27	ppbv	20.7	30.00	
o-Xylene	1.17	ppbv	10.5	30.00	
1,3,5-Trimethylbenzene	1.13	ppbv	7.3	30.00	
1,2,4-Trimethylbenzene	1.09	ppbv	3.4	30.00	
m-Dichlorobenzene	1.18	ppbv	12.9	30.00	
p-Dichlorobenzene	1.17	ppbv	11.6	30.00	
o-Dichlorobenzene	1.18	ppbv	13.0	30.00	
1,2,4-Trichlorobenzene	1.08	ppbv	4.1	30.00	QB-01
Hexachloro-1,3-butadiene	1.25	ppbv	18.2	30.00	

Sequence 2603093

Calibration Check (2603093-CCV1)

Prepared & Analyzed: 03/26/26

Dichlorodifluoromethane	1.19	ppbv	15.9	30.00	
Chloromethane	1.12	ppbv	9.8	30.00	
Dichlorotetrafluoroethane	1.17	ppbv	12.6	30.00	
Vinyl chloride	1.09	ppbv	5.8	30.00	
1,3-Butadiene	1.10	ppbv	6.0	30.00	
Bromomethane	1.18	ppbv	14.2	30.00	
Chloroethane	1.15	ppbv	10.8	30.00	
Acetonitrile	1.05	ppbv	1.6	30.00	
Acrolein	0.996	ppbv	-2.4	30.00	
Trichlorofluoromethane	1.20	ppbv	18.8	30.00	
Acrylonitrile	1.09	ppbv	-1.0	30.00	
1,1-Dichloroethene	1.13	ppbv	8.8	30.00	
Dichloromethane	1.14	ppbv	10.5	30.00	
Carbon Disulfide	1.41	ppbv	29.5	30.00	
Trichlorotrifluoroethane	1.22	ppbv	18.8	30.00	
trans-1,2-Dichloroethylene	1.07	ppbv	3.2	30.00	
1,1-Dichloroethane	1.11	ppbv	7.1	30.00	
Methyl tert-Butyl Ether	1.07	ppbv	2.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603093

Calibration Check (2603093-CCV1) Continued

Prepared & Analyzed: 03/26/26

Chloroprene	1.00	ppbv	-3.4	30.00
cis-1,2-Dichloroethylene	1.06	ppbv	2.6	30.00
Bromochloromethane	1.11	ppbv	8.3	30.00
Chloroform	1.12	ppbv	6.6	30.00
Ethyl tert-Butyl Ether	1.04	ppbv	0.0	30.00
1,2-Dichloroethane	1.09	ppbv	5.6	30.00
1,1,1-Trichloroethane	1.14	ppbv	10.4	30.00
Benzene	1.03	ppbv	-0.5	30.00
Carbon Tetrachloride	1.19	ppbv	16.3	30.00
tert-Amyl Methyl Ether	1.06	ppbv	2.3	30.00
1,2-Dichloropropane	1.16	ppbv	11.8	30.00
Ethyl Acrylate	1.07	ppbv	2.5	30.00
Bromodichloromethane	1.13	ppbv	9.5	30.00
Trichloroethylene	1.11	ppbv	7.1	30.00
Methyl Methacrylate	1.04	ppbv	-0.01	30.00
cis-1,3-Dichloropropene	1.09	ppbv	4.3	30.00
Methyl Isobutyl Ketone	1.07	ppbv	3.1	30.00
trans-1,3-Dichloropropene	1.12	ppbv	7.1	30.00
1,1,2-Trichloroethane	1.15	ppbv	9.6	30.00
Toluene	1.05	ppbv	0.9	30.00
Dibromochloromethane	1.05	ppbv	1.5	30.00
1,2-Dibromoethane	1.09	ppbv	4.1	30.00
n-Octane	0.984	ppbv	-5.2	30.00
Tetrachloroethylene	1.06	ppbv	1.5	30.00
Chlorobenzene	1.11	ppbv	5.9	30.00
Ethylbenzene	1.14	ppbv	9.4	30.00
m,p-Xylene	2.35	ppbv	12.6	30.00
Bromoform	1.17	ppbv	12.2	30.00
Styrene	1.11	ppbv	5.2	30.00
1,1,2,2-Tetrachloroethane	1.26	ppbv	19.2	30.00
o-Xylene	1.15	ppbv	8.6	30.00
1,3,5-Trimethylbenzene	1.12	ppbv	6.1	30.00
1,2,4-Trimethylbenzene	1.08	ppbv	2.4	30.00
m-Dichlorobenzene	1.16	ppbv	11.3	30.00
p-Dichlorobenzene	1.15	ppbv	10.1	30.00
o-Dichlorobenzene	1.16	ppbv	11.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603093

Calibration Check (2603093-CCV1) Continued

Prepared & Analyzed: 03/26/26

1,2,4-Trichlorobenzene	1.06	ppbv	2.6	30.00
Hexachloro-1,3-butadiene	1.24	ppbv	17.8	30.00

Calibration Check (2603093-CCV2)

Prepared: 03/26/26 Analyzed: 03/27/26

Dichlorodifluoromethane	1.25	ppbv	22.0	30.00
Chloromethane	1.16	ppbv	13.8	30.00
Dichlorotetrafluoroethane	1.21	ppbv	16.9	30.00
Vinyl chloride	1.11	ppbv	7.4	30.00
1,3-Butadiene	1.06	ppbv	2.7	30.00
Bromomethane	1.23	ppbv	19.1	30.00
Chloroethane	1.18	ppbv	14.0	30.00
Acetonitrile	1.14	ppbv	10.3	30.00
Acrolein	1.05	ppbv	3.3	30.00
Trichlorofluoromethane	1.24	ppbv	22.5	30.00
Acrylonitrile	1.16	ppbv	4.8	30.00
1,1-Dichloroethene	1.19	ppbv	14.8	30.00
Dichloromethane	1.19	ppbv	15.0	30.00
Carbon Disulfide	1.44	ppbv	32.6	30.00
Trichlorotrifluoroethane	1.28	ppbv	24.6	30.00
trans-1,2-Dichloroethylene	1.06	ppbv	1.7	30.00
1,1-Dichloroethane	1.15	ppbv	11.3	30.00
Methyl tert-Butyl Ether	1.03	ppbv	-0.8	30.00
Chloroprene	0.968	ppbv	-6.8	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	4.6	30.00
Bromochloromethane	1.14	ppbv	11.3	30.00
Chloroform	1.17	ppbv	12.1	30.00
Ethyl tert-Butyl Ether	0.997	ppbv	-3.8	30.00
1,2-Dichloroethane	1.17	ppbv	12.8	30.00
1,1,1-Trichloroethane	1.21	ppbv	17.4	30.00
Benzene	1.07	ppbv	2.9	30.00
Carbon Tetrachloride	1.24	ppbv	21.4	30.00
tert-Amyl Methyl Ether	1.02	ppbv	-1.7	30.00
1,2-Dichloropropane	1.15	ppbv	10.5	30.00
Ethyl Acrylate	1.02	ppbv	-2.7	30.00
Bromodichloromethane	1.23	ppbv	18.5	30.00
Trichloroethylene	1.15	ppbv	11.4	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603093

Calibration Check (2603093-CCV2) Continued

Prepared: 03/26/26 Analyzed: 03/27/26

Methyl Methacrylate	0.989	ppbv	-5.1	30.00
cis-1,3-Dichloropropene	1.01	ppbv	-3.4	30.00
Methyl Isobutyl Ketone	1.05	ppbv	1.4	30.00
trans-1,3-Dichloropropene	1.10	ppbv	5.0	30.00
1,1,2-Trichloroethane	1.17	ppbv	11.6	30.00
Toluene	1.02	ppbv	-2.3	30.00
Dibromochloromethane	1.04	ppbv	0.5	30.00
1,2-Dibromoethane	1.08	ppbv	3.5	30.00
n-Octane	0.945	ppbv	-8.9	30.00
Tetrachloroethylene	1.07	ppbv	3.0	30.00
Chlorobenzene	1.14	ppbv	8.8	30.00
Ethylbenzene	1.09	ppbv	4.3	30.00
m,p-Xylene	2.27	ppbv	9.0	30.00
Bromoform	1.20	ppbv	15.0	30.00
Styrene	1.10	ppbv	3.7	30.00
1,1,2,2-Tetrachloroethane	1.35	ppbv	28.4	30.00
o-Xylene	1.09	ppbv	3.6	30.00
1,3,5-Trimethylbenzene	1.04	ppbv	-1.2	30.00
1,2,4-Trimethylbenzene	1.00	ppbv	-4.7	30.00
m-Dichlorobenzene	1.15	ppbv	10.4	30.00
p-Dichlorobenzene	1.13	ppbv	7.8	30.00
o-Dichlorobenzene	1.15	ppbv	10.2	30.00
1,2,4-Trichlorobenzene	1.07	ppbv	2.9	30.00
Hexachloro-1,3-butadiene	1.28	ppbv	21.4	30.00

Sequence 2604048

Calibration Check (2604048-CCV1)

Prepared & Analyzed: 04/15/26

Dichlorodifluoromethane	1.24	ppbv	20.8	30.00
Chloromethane	1.12	ppbv	9.5	30.00
Dichlorotetrafluoroethane	1.10	ppbv	6.6	30.00
Vinyl chloride	1.08	ppbv	4.8	30.00
1,3-Butadiene	1.08	ppbv	4.5	30.00
Bromomethane	1.14	ppbv	10.7	30.00
Chloroethane	1.11	ppbv	7.0	30.00
Acetonitrile	1.14	ppbv	9.7	30.00
Acrolein	1.14	ppbv	11.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604048

Calibration Check (2604048-CCV1) Continued

Prepared & Analyzed: 04/15/26

Trichlorofluoromethane	1.17	ppbv	15.7	30.00	QB-01
Acrylonitrile	1.20	ppbv	8.6	30.00	
1,1-Dichloroethene	1.16	ppbv	11.9	30.00	
Dichloromethane	1.16	ppbv	12.1	30.00	
Carbon Disulfide	1.10	ppbv	1.3	30.00	
Trichlorotrifluoroethane	1.18	ppbv	14.6	30.00	
trans-1,2-Dichloroethylene	1.10	ppbv	5.7	30.00	
1,1-Dichloroethane	1.12	ppbv	8.0	30.00	
Methyl tert-Butyl Ether	1.15	ppbv	10.8	30.00	
Chloroprene	1.14	ppbv	10.0	30.00	
cis-1,2-Dichloroethylene	1.14	ppbv	10.2	30.00	
Bromochloromethane	1.09	ppbv	5.9	30.00	
Chloroform	1.19	ppbv	13.8	30.00	
Ethyl tert-Butyl Ether	1.19	ppbv	14.7	30.00	
1,2-Dichloroethane	1.19	ppbv	15.3	30.00	
1,1,1-Trichloroethane	1.21	ppbv	17.7	30.00	
Benzene	1.13	ppbv	8.9	30.00	
Carbon Tetrachloride	1.20	ppbv	17.5	30.00	
tert-Amyl Methyl Ether	1.20	ppbv	15.6	30.00	
1,2-Dichloropropane	1.16	ppbv	11.4	30.00	
Ethyl Acrylate	1.15	ppbv	9.7	30.00	
Bromodichloromethane	1.15	ppbv	11.2	30.00	
Trichloroethylene	1.13	ppbv	9.0	30.00	
Methyl Methacrylate	1.16	ppbv	11.4	30.00	
cis-1,3-Dichloropropene	1.23	ppbv	17.8	30.00	
Methyl Isobutyl Ketone	1.15	ppbv	10.2	30.00	
trans-1,3-Dichloropropene	1.23	ppbv	16.8	30.00	
1,1,2-Trichloroethane	1.05	ppbv	0.4	30.00	
Toluene	1.14	ppbv	9.5	30.00	
Dibromochloromethane	1.13	ppbv	9.7	30.00	
1,2-Dibromoethane	1.07	ppbv	2.9	30.00	
n-Octane	1.17	ppbv	13.1	30.00	
Tetrachloroethylene	1.24	ppbv	18.3	30.00	
Chlorobenzene	1.14	ppbv	8.3	30.00	
Ethylbenzene	1.06	ppbv	1.7	30.00	
m,p-Xylene	2.20	ppbv	5.2	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604048

Calibration Check (2604048-CCV1) Continued

Prepared & Analyzed: 04/15/26

Bromoform	1.30	ppbv	24.3	30.00	
Styrene	1.14	ppbv	7.5	30.00	
1,1,2,2-Tetrachloroethane	1.28	ppbv	21.5	30.00	
o-Xylene	1.20	ppbv	13.2	30.00	
1,3,5-Trimethylbenzene	1.19	ppbv	13.4	30.00	
1,2,4-Trimethylbenzene	1.17	ppbv	10.8	30.00	
m-Dichlorobenzene	1.09	ppbv	4.9	30.00	
p-Dichlorobenzene	1.10	ppbv	4.7	30.00	
o-Dichlorobenzene	1.10	ppbv	4.8	30.00	
1,2,4-Trichlorobenzene	1.19	ppbv	15.0	30.00	QB-01
Hexachloro-1,3-butadiene	1.23	ppbv	17.1	30.00	

Calibration Check (2604048-CCV2)

Prepared: 04/15/26 Analyzed: 04/16/26

Dichlorodifluoromethane	1.19	ppbv	15.9	30.00	
Chloromethane	1.03	ppbv	1.1	30.00	
Dichlorotetrafluoroethane	1.16	ppbv	12.1	30.00	
Vinyl chloride	0.994	ppbv	-3.7	30.00	
1,3-Butadiene	0.981	ppbv	-5.2	30.00	
Bromomethane	1.07	ppbv	3.9	30.00	
Chloroethane	1.04	ppbv	0.7	30.00	
Acetonitrile	1.08	ppbv	3.9	30.00	
Acrolein	1.03	ppbv	1.0	30.00	
Trichlorofluoromethane	1.11	ppbv	9.9	30.00	
Acrylonitrile	1.14	ppbv	3.2	30.00	
1,1-Dichloroethene	1.08	ppbv	3.7	30.00	
Dichloromethane	1.10	ppbv	6.0	30.00	QB-01
Carbon Disulfide	1.03	ppbv	-5.0	30.00	
Trichlorotrifluoroethane	1.14	ppbv	11.0	30.00	
trans-1,2-Dichloroethylene	1.01	ppbv	-2.7	30.00	
1,1-Dichloroethane	1.12	ppbv	8.5	30.00	
Methyl tert-Butyl Ether	1.11	ppbv	7.0	30.00	
Chloroprene	1.11	ppbv	6.6	30.00	
cis-1,2-Dichloroethylene	1.09	ppbv	5.4	30.00	
Bromochloromethane	1.01	ppbv	-1.7	30.00	
Chloroform	1.16	ppbv	10.6	30.00	
Ethyl tert-Butyl Ether	1.13	ppbv	8.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604048

Calibration Check (2604048-CCV2) Continued

Prepared: 04/15/26 Analyzed: 04/16/26

1,2-Dichloroethane	1.15	ppbv	11.2	30.00
1,1,1-Trichloroethane	1.20	ppbv	16.4	30.00
Benzene	1.08	ppbv	3.9	30.00
Carbon Tetrachloride	1.10	ppbv	7.6	30.00
tert-Amyl Methyl Ether	1.12	ppbv	8.7	30.00
1,2-Dichloropropane	1.10	ppbv	5.6	30.00
Ethyl Acrylate	1.08	ppbv	2.8	30.00
Bromodichloromethane	1.05	ppbv	1.5	30.00
Trichloroethylene	1.03	ppbv	-0.3	30.00
Methyl Methacrylate	1.08	ppbv	4.1	30.00
cis-1,3-Dichloropropene	1.15	ppbv	10.5	30.00
Methyl Isobutyl Ketone	1.08	ppbv	3.4	30.00
trans-1,3-Dichloropropene	1.19	ppbv	13.3	30.00
1,1,2-Trichloroethane	0.974	ppbv	-6.9	30.00
Toluene	1.08	ppbv	3.6	30.00
Dibromochloromethane	1.03	ppbv	-0.6	30.00
1,2-Dibromoethane	0.970	ppbv	-7.0	30.00
n-Octane	1.08	ppbv	4.3	30.00
Tetrachloroethylene	1.15	ppbv	10.4	30.00
Chlorobenzene	1.03	ppbv	-1.7	30.00
Ethylbenzene	0.970	ppbv	-7.3	30.00
m,p-Xylene	2.01	ppbv	-3.6	30.00
Bromoform	1.21	ppbv	15.8	30.00
Styrene	1.06	ppbv	-0.2	30.00
1,1,2,2-Tetrachloroethane	1.22	ppbv	15.8	30.00
o-Xylene	1.12	ppbv	6.0	30.00
1,3,5-Trimethylbenzene	1.12	ppbv	6.0	30.00
1,2,4-Trimethylbenzene	1.09	ppbv	3.4	30.00
m-Dichlorobenzene	1.00	ppbv	-3.7	30.00
p-Dichlorobenzene	1.00	ppbv	-4.0	30.00
o-Dichlorobenzene	1.01	ppbv	-3.6	30.00
1,2,4-Trichlorobenzene	1.12	ppbv	7.8	30.00
Hexachloro-1,3-butadiene	1.21	ppbv	14.8	30.00

QB-01

Sequence 2604069

Calibration Check (2604069-CCV1)

Prepared & Analyzed: 04/21/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV1) Continued

Prepared & Analyzed: 04/21/26

Dichlorodifluoromethane	1.16	ppbv	13.1	30.00
Chloromethane	1.08	ppbv	5.5	30.00
Dichlorotetrafluoroethane	1.16	ppbv	12.2	30.00
Vinyl chloride	1.04	ppbv	0.6	30.00
1,3-Butadiene	0.907	ppbv	-12.3	30.00
Bromomethane	1.19	ppbv	15.2	30.00
Chloroethane	1.16	ppbv	11.8	30.00
Acetonitrile	0.806	ppbv	-22.4	30.00
Acrolein	1.11	ppbv	8.6	30.00
Trichlorofluoromethane	1.19	ppbv	18.3	30.00
Acrylonitrile	1.12	ppbv	1.7	30.00
1,1-Dichloroethene	1.23	ppbv	18.5	30.00
Dichloromethane	1.03	ppbv	-0.8	30.00
Carbon Disulfide	1.29	ppbv	18.4	30.00
Trichlorotrifluoroethane	1.28	ppbv	24.7	30.00
trans-1,2-Dichloroethylene	1.13	ppbv	8.8	30.00
1,1-Dichloroethane	1.12	ppbv	8.2	30.00
Methyl tert-Butyl Ether	0.942	ppbv	-9.1	30.00
Chloroprene	0.950	ppbv	-8.5	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.2	30.00
Bromochloromethane	1.17	ppbv	13.8	30.00
Chloroform	1.16	ppbv	10.8	30.00
Ethyl tert-Butyl Ether	0.866	ppbv	-16.5	30.00
1,2-Dichloroethane	1.12	ppbv	8.4	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.4	30.00
Benzene	1.06	ppbv	2.1	30.00
Carbon Tetrachloride	1.14	ppbv	11.5	30.00
tert-Amyl Methyl Ether	0.823	ppbv	-20.4	30.00
1,2-Dichloropropane	0.981	ppbv	-5.5	30.00
Ethyl Acrylate	0.895	ppbv	-14.4	30.00
Bromodichloromethane	1.07	ppbv	3.0	30.00
Trichloroethylene	1.05	ppbv	1.8	30.00
Methyl Methacrylate	0.873	ppbv	-16.2	30.00
cis-1,3-Dichloropropene	0.955	ppbv	-8.5	30.00
Methyl Isobutyl Ketone	0.906	ppbv	-12.9	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-2.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV1) Continued

Prepared & Analyzed: 04/21/26

1,1,2-Trichloroethane	1.10	ppbv	5.3	30.00
Toluene	0.982	ppbv	-5.6	30.00
Dibromochloromethane	1.08	ppbv	4.1	30.00
1,2-Dibromoethane	1.06	ppbv	1.8	30.00
n-Octane	0.894	ppbv	-13.9	30.00
Tetrachloroethylene	1.10	ppbv	5.0	30.00
Chlorobenzene	1.10	ppbv	4.6	30.00
Ethylbenzene	0.985	ppbv	-5.8	30.00
m,p-Xylene	2.07	ppbv	-0.7	30.00
Bromoform	1.19	ppbv	13.7	30.00
Styrene	1.07	ppbv	1.2	30.00
1,1,2,2-Tetrachloroethane	1.13	ppbv	7.4	30.00
o-Xylene	0.983	ppbv	-6.9	30.00
1,3,5-Trimethylbenzene	0.924	ppbv	-12.2	30.00
1,2,4-Trimethylbenzene	0.912	ppbv	-13.4	30.00
m-Dichlorobenzene	1.07	ppbv	2.7	30.00
p-Dichlorobenzene	1.05	ppbv	0.8	30.00
o-Dichlorobenzene	1.06	ppbv	1.1	30.00
1,2,4-Trichlorobenzene	0.956	ppbv	-7.9	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.6	30.00

Calibration Check (2604069-CCV2)

Prepared: 04/21/26 Analyzed: 04/22/26

Dichlorodifluoromethane	1.18	ppbv	14.4	30.00
Chloromethane	1.08	ppbv	5.6	30.00
Dichlorotetrafluoroethane	1.18	ppbv	13.7	30.00
Vinyl chloride	1.02	ppbv	-1.0	30.00
1,3-Butadiene	0.837	ppbv	-19.0	30.00
Bromomethane	1.22	ppbv	18.1	30.00
Chloroethane	1.16	ppbv	12.4	30.00
Acetonitrile	0.804	ppbv	-22.5	30.00
Acrolein	1.07	ppbv	5.1	30.00
Trichlorofluoromethane	1.21	ppbv	20.0	30.00
Acrylonitrile	1.10	ppbv	-0.6	30.00
1,1-Dichloroethene	1.21	ppbv	16.5	30.00
Dichloromethane	1.14	ppbv	10.2	30.00
Carbon Disulfide	1.28	ppbv	17.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604069

Calibration Check (2604069-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

Trichlorotrifluoroethane	1.28	ppbv	24.1	30.00
trans-1,2-Dichloroethylene	1.14	ppbv	10.0	30.00
1,1-Dichloroethane	1.14	ppbv	10.3	30.00
Methyl tert-Butyl Ether	0.916	ppbv	-11.5	30.00
Chloroprene	0.913	ppbv	-12.1	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.6	30.00
Bromochloromethane	1.22	ppbv	18.2	30.00
Chloroform	1.18	ppbv	12.6	30.00
Ethyl tert-Butyl Ether	0.848	ppbv	-18.1	30.00
1,2-Dichloroethane	1.13	ppbv	9.2	30.00
1,1,1-Trichloroethane	1.14	ppbv	10.4	30.00
Benzene	1.06	ppbv	2.3	30.00
Carbon Tetrachloride	1.17	ppbv	14.0	30.00
tert-Amyl Methyl Ether	0.810	ppbv	-21.7	30.00
1,2-Dichloropropane	0.993	ppbv	-4.3	30.00
Ethyl Acrylate	0.878	ppbv	-16.1	30.00
Bromodichloromethane	1.07	ppbv	3.4	30.00
Trichloroethylene	1.07	ppbv	3.3	30.00
Methyl Methacrylate	0.855	ppbv	-18.0	30.00
cis-1,3-Dichloropropene	0.968	ppbv	-7.3	30.00
Methyl Isobutyl Ketone	0.915	ppbv	-12.0	30.00
trans-1,3-Dichloropropene	1.05	ppbv	0.2	30.00
1,1,2-Trichloroethane	1.15	ppbv	9.5	30.00
Toluene	1.00	ppbv	-3.9	30.00
Dibromochloromethane	1.12	ppbv	8.0	30.00
1,2-Dibromoethane	1.12	ppbv	6.9	30.00
n-Octane	0.874	ppbv	-15.8	30.00
Tetrachloroethylene	1.15	ppbv	10.4	30.00
Chlorobenzene	1.10	ppbv	4.1	30.00
Ethylbenzene	0.957	ppbv	-8.5	30.00
m,p-Xylene	2.00	ppbv	-4.4	30.00
Bromoform	1.22	ppbv	16.2	30.00
Styrene	1.06	ppbv	0.1	30.00
1,1,2,2-Tetrachloroethane	1.12	ppbv	5.8	30.00
o-Xylene	0.962	ppbv	-8.9	30.00
1,3,5-Trimethylbenzene	0.900	ppbv	-14.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 05/14/26 15:18
SUBMITTED: 03/18/26 to 04/06/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control
Sequence 2604069

Calibration Check (2604069-CCV2) Continued

Prepared: 04/21/26 Analyzed: 04/22/26

1,2,4-Trimethylbenzene	0.903	ppbv	-14.4	30.00
m-Dichlorobenzene	1.08	ppbv	3.8	30.00
p-Dichlorobenzene	1.06	ppbv	1.4	30.00
o-Dichlorobenzene	1.08	ppbv	3.6	30.00
1,2,4-Trichlorobenzene	0.864	ppbv	-16.7	30.00
Hexachloro-1,3-butadiene	0.998	ppbv	-5.4	30.00



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 05/14/26 15:18

SUBMITTED: 03/18/26 to 04/06/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.