



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 08, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/07/26 10:54 through 06/03/26 13:31.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6050702-02	Air	05/02/26 23:59	05/07/26 10:54
TMOK	6051320-02	Air	05/08/26 23:59	05/13/26 12:55
TMOK D1	6052714-01	Air	05/14/26 23:59	05/27/26 12:59
TMOK D2	6052714-02	Air	05/14/26 23:59	05/27/26 12:59
TMOK	6052714-03	Air	05/20/26 23:59	05/27/26 12:59
TMOK	6060308-01	Air	05/26/26 23:59	06/03/26 13:31
TMOK FB	6060308-02	Air	05/26/26 23:59	06/03/26 13:31



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6050702-02	Sampled: 05/02/26 23:59
Temp Rec'd (°C): 3.4	Sample Volume: 1156.1 L	Received: 05/07/26 10:54
	Cartridge Pouch #: 26G0963	Analyzed: 05/11/26 16:37

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.84		0.0548
Acetaldehyde	1.33		0.0227
Acetone	4.01		0.0839
Propionaldehyde	0.476		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.149		0.00969
2-Butanone	0.535		0.0365
Benzaldehyde	0.107		0.0375
Valeraldehyde	0.0687		0.0105
Hexaldehyde	0.142		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6051320-02	Sampled: 05/08/26 23:59
Temp Rec'd (°C): 4.9	Sample Volume: 1156.1 L	Received: 05/13/26 12:55
	Cartridge Pouch #: 26G0964	Analyzed: 05/19/26 19:52

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	4.11		0.0548
Acetaldehyde	1.81		0.0227
Acetone	1.86		0.0839
Propionaldehyde	0.400		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.192		0.00969
2-Butanone	0.555		0.0365
Benzaldehyde	0.138		0.0375
Valeraldehyde	0.234		0.0105
Hexaldehyde	0.228		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1	Lab ID: 6052714-01	Sampled: 05/14/26 23:59
Temp Rec'd (°C): 3.5	Sample Volume: 1144.5 L	Received: 05/27/26 12:59
	Cartridge Pouch #: 26G0965	Analyzed: 05/30/26 09:22

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	5.19		0.0553
Acetaldehyde	1.67		0.0229
Acetone	2.21		0.0848
Propionaldehyde	0.320		0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.148		0.00979
2-Butanone	0.319		0.0369
Benzaldehyde	0.138		0.0378
Valeraldehyde	0.199		0.0106
Hexaldehyde	0.209		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2	Lab ID: 6052714-02	Sampled: 05/14/26 23:59
Temp Rec'd (°C): 3.5	Sample Volume: 1144.2 L	Received: 05/27/26 12:59
	Cartridge Pouch #: 26G0966	Analyzed: 05/30/26 09:55

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	5.26		0.0553
Acetaldehyde	1.68		0.0229
Acetone	2.15		0.0848
Propionaldehyde	0.266		0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.155		0.00979
2-Butanone	0.316		0.0369
Benzaldehyde	0.131		0.0378
Valeraldehyde	0.200		0.0106
Hexaldehyde	0.221		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6052714-03	Sampled: 05/20/26 23:59
Temp Rec'd (°C): 3.5	Sample Volume: 1156.6 L	Received: 05/27/26 12:59
	Cartridge Pouch #: 26G0967	Analyzed: 06/02/26 16:49

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.962		0.0547
Acetaldehyde	1.24		0.0227
Acetone	0.888		0.0839
Propionaldehyde	0.117	U	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.0663		0.00968
2-Butanone	0.0808		0.0365
Benzaldehyde	0.0750		0.0374
Valeraldehyde	0.117		0.0105
Hexaldehyde	0.0869		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK	Lab ID: 6060308-01	Sampled: 05/26/26 23:59
Temp Rec'd (°C): 4.8	Sample Volume: 1155.8 L	Received: 06/03/26 13:31
	Cartridge Pouch #: 26G0968	Analyzed: 06/04/26 22:17

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.54		0.0548
Acetaldehyde	1.45		0.0227
Acetone	1.19		0.0839
Propionaldehyde	0.260		0.179
Crotonaldehyde	1.34		0.0164
Butyraldehyde	0.168		0.00969
2-Butanone	0.285		0.0365
Benzaldehyde	0.152		0.0375
Valeraldehyde	0.264		0.0105
Hexaldehyde	0.208		0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK FB

Lab ID: 6060308-02

Sampled: 05/26/26 23:59

Temp Rec'd (°C): 4.8

Sample Volume: 1155.8 L

Received: 06/03/26 13:31

Cartridge Pouch #: 26G0969

Analyzed: 06/04/26 22:33

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0334	U	0.0548
Acetaldehyde	0.0215	U	0.0227
Acetone	0.0420	U	0.0839
Propionaldehyde	0.0324	U	0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	ND	U	0.00969
2-Butanone	ND	U	0.0365
Benzaldehyde	0.0394		0.0375
Valeraldehyde	ND	U	0.0105
Hexaldehyde	ND	U	0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Blank (2605031-CCB1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2605031-CCB2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605031-CCV1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.143	ug/ml	0.15000		95.4	84.9-115.1			
Acetaldehyde	0.146	ug/ml	0.15000		97.3	84.9-115.1			
Acetone	0.147	ug/ml	0.15000		97.7	84.9-115.1			
Propionaldehyde	0.149	ug/ml	0.15000		99.4	84.9-115.1			
Crotonaldehyde	0.149	ug/ml	0.15000		99.6	84.9-115.1			
Butyraldehyde	0.140	ug/ml	0.15000		93.4	84.9-115.1			
2-Butanone	0.145	ug/ml	0.15000		96.7	84.9-115.1			
Benzaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1			
Valeraldehyde	0.146	ug/ml	0.15000		97.1	84.9-115.1			
Hexaldehyde	0.145	ug/ml	0.15000		96.5	84.9-115.1			

Calibration Check (2605031-CCV2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000		96.1	84.9-115.1			
Acetaldehyde	0.146	ug/ml	0.15000		97.5	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Check (2605031-CCV2) Contin

Prepared & Analyzed: 05/11/26

Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1

Calibration Check (2605031-CCV3)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.6	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	95.7	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch 2605066 - B6E1503

Calibration Blank (2605066-CCB1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605066-CCB2)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605066 - B6E1503

Calibration Blank (2605066-CCB2) Contin

Prepared & Analyzed: 05/19/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605066-CCV1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.0	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Hexaldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1

Calibration Check (2605066-CCV2)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1

Calibration Check (2605066-CCV3)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.5	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.5	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605066 - B6E1503

Calibration Check (2605066-CCV3) Contin

Prepared & Analyzed: 05/19/26

Hexaldehyde	0.142	ug/ml	0.15000	94.5	84.9-115.1
-------------	-------	-------	---------	------	------------

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605069-CCB2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605069-CCV1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.9	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
2-Butanone	0.154	ug/ml	0.15000	102	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1

Calibration Check (2605069-CCV2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Check (2605069-CCV2) Contin

Prepared & Analyzed: 05/20/26

Acetaldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Calibration Check (2605069-CCV3)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch 2605092 - B6E2903

Calibration Blank (2605092-CCB1)

Prepared & Analyzed: 05/29/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2605092-CCB2)

Prepared: 05/29/26 Analyzed: 05/30/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605092 - B6E2903

Calibration Blank (2605092-CCB2) Contin

Prepared: 05/29/26 Analyzed: 05/30/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605092-CCV1)

Prepared & Analyzed: 05/29/26

Formaldehyde	0.145	ug/ml	0.15000	97.0	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.2	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.8	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1

Calibration Check (2605092-CCV2)

Prepared & Analyzed: 05/29/26

Formaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.155	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1

Calibration Check (2605092-CCV3)

Prepared & Analyzed: 05/29/26

Formaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.4	84.9-115.1
Propionaldehyde	0.145	ug/ml	0.15000	97.0	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605092 - B6E2903

Calibration Check (2605092-CCV3) Contin

Prepared & Analyzed: 05/29/26

Valeraldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1

Calibration Check (2605092-CCV4)

Prepared: 05/29/26 Analyzed: 05/30/26

Formaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.8	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.7	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1

Calibration Check (2605092-CCV5)

Prepared: 05/29/26 Analyzed: 05/30/26

Formaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.9	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.2	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1

Calibration Check (2605092-CCV6)

Prepared: 05/29/26 Analyzed: 05/30/26

Formaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.1	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.6	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	95.0	84.9-115.1

Batch 2606005 - B6F0104

Calibration Blank (2606005-CCB1)

Prepared & Analyzed: 06/02/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606005 - B6F0104

Calibration Blank (2606005-CCB1) Contin

Prepared & Analyzed: 06/02/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2606005-CCB2)

Prepared & Analyzed: 06/02/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2606005-CCV1)

Prepared & Analyzed: 06/02/26

Formaldehyde	0.147	ug/ml	0.15000		97.7	84.9-115.1			
Acetaldehyde	0.148	ug/ml	0.15000		98.6	84.9-115.1			
Acetone	0.148	ug/ml	0.15000		98.5	84.9-115.1			
Propionaldehyde	0.146	ug/ml	0.15000		97.6	84.9-115.1			
Crotonaldehyde	0.151	ug/ml	0.15000		101	84.9-115.1			
Butyraldehyde	0.143	ug/ml	0.15000		95.0	84.9-115.1			
2-Butanone	0.145	ug/ml	0.15000		96.9	84.9-115.1			
Benzaldehyde	0.154	ug/ml	0.15000		102	84.9-115.1			
Valeraldehyde	0.151	ug/ml	0.15000		101	84.9-115.1			
Hexaldehyde	0.143	ug/ml	0.15000		95.0	84.9-115.1			

Calibration Check (2606005-CCV2)

Prepared & Analyzed: 06/02/26

Formaldehyde	0.146	ug/ml	0.15000		97.6	84.9-115.1			
Acetaldehyde	0.148	ug/ml	0.15000		98.4	84.9-115.1			
Acetone	0.149	ug/ml	0.15000		99.4	84.9-115.1			
Propionaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606005 - B6F0104

Calibration Check (2606005-CCV2) Contin

Prepared & Analyzed: 06/02/26

Crotonaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	97.9	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1

Calibration Check (2606005-CCV3)

Prepared & Analyzed: 06/02/26

Formaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.9	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1

Batch 2606020 - B6F0406

Calibration Blank (2606020-CCB1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2606020-CCB2)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606020 - B6F0406

Calibration Blank (2606020-CCB2) Contin

Prepared & Analyzed: 06/04/26

Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2606020-CCV1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.149	ug/ml	0.15000		99.2	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		99.8	84.9-115.1			
Acetone	0.151	ug/ml	0.15000		101	84.9-115.1			
Propionaldehyde	0.151	ug/ml	0.15000		100	84.9-115.1			
Crotonaldehyde	0.152	ug/ml	0.15000		102	84.9-115.1			
Butyraldehyde	0.147	ug/ml	0.15000		98.3	84.9-115.1			
2-Butanone	0.153	ug/ml	0.15000		102	84.9-115.1			
Benzaldehyde	0.158	ug/ml	0.15000		105	84.9-115.1			
Valeraldehyde	0.151	ug/ml	0.15000		101	84.9-115.1			
Hexaldehyde	0.147	ug/ml	0.15000		97.7	84.9-115.1			

Calibration Check (2606020-CCV2)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.149	ug/ml	0.15000		99.0	84.9-115.1			
Acetaldehyde	0.152	ug/ml	0.15000		102	84.9-115.1			
Acetone	0.152	ug/ml	0.15000		101	84.9-115.1			
Propionaldehyde	0.149	ug/ml	0.15000		99.5	84.9-115.1			
Crotonaldehyde	0.156	ug/ml	0.15000		104	84.9-115.1			
Butyraldehyde	0.148	ug/ml	0.15000		98.5	84.9-115.1			
2-Butanone	0.152	ug/ml	0.15000		101	84.9-115.1			
Benzaldehyde	0.155	ug/ml	0.15000		104	84.9-115.1			
Valeraldehyde	0.154	ug/ml	0.15000		103	84.9-115.1			
Hexaldehyde	0.144	ug/ml	0.15000		95.8	84.9-115.1			

Calibration Check (2606020-CCV3)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.147	ug/ml	0.15000		97.9	84.9-115.1			
Acetaldehyde	0.152	ug/ml	0.15000		101	84.9-115.1			
Acetone	0.149	ug/ml	0.15000		99.4	84.9-115.1			
Propionaldehyde	0.148	ug/ml	0.15000		98.7	84.9-115.1			
Crotonaldehyde	0.156	ug/ml	0.15000		104	84.9-115.1			
Butyraldehyde	0.149	ug/ml	0.15000		99.5	84.9-115.1			
2-Butanone	0.149	ug/ml	0.15000		99.0	84.9-115.1			
Benzaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1			
Valeraldehyde	0.152	ug/ml	0.15000		102	84.9-115.1			
Hexaldehyde	0.147	ug/ml	0.15000		97.9	84.9-115.1			

Batch B6E1108 - TO-11A extraction

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1108 - TO-11A extraction

Blank (B6E1108-BLK1)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0325	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							CE, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0448	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E1108-BS1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.288	ug/ml	0.30000	95.9	79.9-120.1				
Acetaldehyde	0.293	ug/ml	0.30000	97.6	79.9-120.1				
Acetone	0.294	ug/ml	0.30000	98.0	69.9-130.1				
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1				
Crotonaldehyde	0.276	ug/ml	0.30000	92.0	69.9-130.1				
Butyraldehyde	0.263	ug/ml	0.30000	87.6	69.9-130.1				
2-Butanone	0.238	ug/ml	0.30000	79.4	69.9-130.1				
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1				
Valeraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1				
Hexaldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1				

LCS Dup (B6E1108-BSD1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.290	ug/ml	0.30000	96.6	79.9-120.1	0.660	20		
Acetaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1	0.305	20		
Acetone	0.293	ug/ml	0.30000	97.5	69.9-130.1	0.439	20		
Propionaldehyde	0.348	ug/ml	0.30000	116	69.9-130.1	0.877	20		
Crotonaldehyde	0.271	ug/ml	0.30000	90.4	69.9-130.1	1.74	20		
Butyraldehyde	0.259	ug/ml	0.30000	86.2	69.9-130.1	1.60	20		
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.778	20		
Benzaldehyde	0.315	ug/ml	0.30000	105	69.9-130.1	0.358	20		
Valeraldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	2.43	20		
Hexaldehyde	0.249	ug/ml	0.30000	83.1	69.9-130.1	2.06	20		

Batch B6E1503 - TO-11A extraction

Blank (B6E1503-BLK1)

Prepared: 05/15/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1503 - TO-11A extraction

Blank (B6E1503-BLK1) Continued

Prepared: 05/15/26 Analyzed: 05/22/26

Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E1503-BS1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.325	ug/ml	0.30000	108	79.9-120.1		
Acetaldehyde	0.285	ug/ml	0.30000	95.0	79.9-120.1		
Acetone	0.295	ug/ml	0.30000	98.5	69.9-130.1		
Propionaldehyde	0.266	ug/ml	0.30000	88.6	69.9-130.1		
Crotonaldehyde	0.278	ug/ml	0.30000	92.6	69.9-130.1		
Butyraldehyde	0.250	ug/ml	0.30000	83.3	69.9-130.1		
2-Butanone	0.241	ug/ml	0.30000	80.4	69.9-130.1		
Benzaldehyde	0.308	ug/ml	0.30000	103	69.9-130.1		
Valeraldehyde	0.249	ug/ml	0.30000	83.0	69.9-130.1		
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1		

LCS Dup (B6E1503-BSD1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.332	ug/ml	0.30000	111	79.9-120.1	2.31	20
Acetaldehyde	0.289	ug/ml	0.30000	96.4	79.9-120.1	1.46	20
Acetone	0.301	ug/ml	0.30000	100	69.9-130.1	2.03	20
Propionaldehyde	0.267	ug/ml	0.30000	89.1	69.9-130.1	0.612	20
Crotonaldehyde	0.283	ug/ml	0.30000	94.4	69.9-130.1	1.90	20
Butyraldehyde	0.255	ug/ml	0.30000	85.1	69.9-130.1	2.14	20
2-Butanone	0.243	ug/ml	0.30000	81.0	69.9-130.1	0.708	20
Benzaldehyde	0.311	ug/ml	0.30000	104	69.9-130.1	1.04	20
Valeraldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	2.11	20
Hexaldehyde	0.243	ug/ml	0.30000	81.0	69.9-130.1	1.12	20

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1)

Prepared: 05/20/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air						U
Acetaldehyde	ND	ug/m ³ Air						U
Acetone	ND	ug/m ³ Air						U
Propionaldehyde	ND	ug/m ³ Air						U
Crotonaldehyde	ND	ug/m ³ Air						U
Butyraldehyde	ND	ug/m ³ Air						U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1) Continued

Prepared: 05/20/26 Analyzed: 05/22/26

2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E2006-BS1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.298	ug/ml	0.30000		99.2	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.296	ug/ml	0.30000		98.8	69.9-130.1			
Propionaldehyde	0.269	ug/ml	0.30000		89.6	69.9-130.1			
Crotonaldehyde	0.291	ug/ml	0.30000		97.1	69.9-130.1			
Butyraldehyde	0.257	ug/ml	0.30000		85.5	69.9-130.1			
2-Butanone	0.240	ug/ml	0.30000		79.9	69.9-130.1			
Benzaldehyde	0.310	ug/ml	0.30000		103	69.9-130.1			
Valeraldehyde	0.255	ug/ml	0.30000		85.0	69.9-130.1			
Hexaldehyde	0.244	ug/ml	0.30000		81.4	69.9-130.1			

LCS Dup (B6E2006-BSD1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.297	ug/ml	0.30000		98.9	79.9-120.1	0.349	20	
Acetaldehyde	0.290	ug/ml	0.30000		96.7	79.9-120.1	0.499	20	
Acetone	0.294	ug/ml	0.30000		98.1	69.9-130.1	0.653	20	
Propionaldehyde	0.267	ug/ml	0.30000		88.9	69.9-130.1	0.876	20	
Crotonaldehyde	0.281	ug/ml	0.30000		93.7	69.9-130.1	3.50	20	
Butyraldehyde	0.253	ug/ml	0.30000		84.3	69.9-130.1	1.36	20	
2-Butanone	0.240	ug/ml	0.30000		80.1	69.9-130.1	0.186	20	
Benzaldehyde	0.307	ug/ml	0.30000		102	69.9-130.1	0.986	20	
Valeraldehyde	0.254	ug/ml	0.30000		84.8	69.9-130.1	0.252	20	
Hexaldehyde	0.246	ug/ml	0.30000		82.0	69.9-130.1	0.678	20	

Batch B6E2802 - TO-11A extraction

Blank (B6E2802-BLK1)

Prepared: 05/28/26 Analyzed: 05/29/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0450	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2802 - TO-11A extraction

Blank (B6E2802-BLK1) Continued

Prepared: 05/28/26 Analyzed: 05/29/26

Hexaldehyde	ND	ug/m ³ Air							U
-------------	----	-----------------------	--	--	--	--	--	--	---

LCS (B6E2802-BS1)

Prepared: 05/28/26 Analyzed: 05/29/26

Formaldehyde	0.260	ug/ml	0.30000		86.8	79.9-120.1			
Acetaldehyde	0.271	ug/ml	0.30000		90.3	79.9-120.1			
Acetone	0.275	ug/ml	0.30000		91.8	69.9-130.1			
Propionaldehyde	0.250	ug/ml	0.30000		83.3	69.9-130.1			
Crotonaldehyde	0.259	ug/ml	0.30000		86.3	69.9-130.1			
Butyraldehyde	0.234	ug/ml	0.30000		77.9	69.9-130.1			
2-Butanone	0.223	ug/ml	0.30000		74.2	69.9-130.1			
Benzaldehyde	0.284	ug/ml	0.30000		94.7	69.9-130.1			
Valeraldehyde	0.236	ug/ml	0.30000		78.5	69.9-130.1			
Hexaldehyde	0.227	ug/ml	0.30000		75.7	69.9-130.1			

LCS Dup (B6E2802-BSD1)

Prepared: 05/28/26 Analyzed: 05/29/26

Formaldehyde	0.273	ug/ml	0.30000		91.2	79.9-120.1	4.90	20	
Acetaldehyde	0.283	ug/ml	0.30000		94.4	79.9-120.1	4.45	20	
Acetone	0.290	ug/ml	0.30000		96.7	69.9-130.1	5.17	20	
Propionaldehyde	0.264	ug/ml	0.30000		88.1	69.9-130.1	5.54	20	
Crotonaldehyde	0.269	ug/ml	0.30000		89.8	69.9-130.1	3.97	20	
Butyraldehyde	0.247	ug/ml	0.30000		82.4	69.9-130.1	5.56	20	
2-Butanone	0.231	ug/ml	0.30000		76.9	69.9-130.1	3.48	20	
Benzaldehyde	0.298	ug/ml	0.30000		99.4	69.9-130.1	4.83	20	
Valeraldehyde	0.245	ug/ml	0.30000		81.5	69.9-130.1	3.72	20	
Hexaldehyde	0.241	ug/ml	0.30000		80.3	69.9-130.1	5.95	20	

Duplicate (B6E2802-DUPH)

Source: 6052714-01 Prepared: 05/28/26 Analyzed: 05/30/26

Formaldehyde	5.20	ug/m ³ Air		5.19			0.161	10.1	
Acetaldehyde	1.66	ug/m ³ Air		1.67			0.229	10.1	
Acetone	2.20	ug/m ³ Air		2.21			0.456	10.1	
Propionaldehyde	0.319	ug/m ³ Air		0.320			0.406	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.152	ug/m ³ Air		0.148			2.52	10.1	
2-Butanone	0.312	ug/m ³ Air		0.319			1.95	10.1	
Benzaldehyde	0.144	ug/m ³ Air		0.138			3.99	10.1	
Valeraldehyde	0.195	ug/m ³ Air		0.199			2.22	10.1	
Hexaldehyde	0.206	ug/m ³ Air		0.209			1.23	10.1	

Duplicate (B6E2802-DUPI)

Source: 6052714-02 Prepared: 05/28/26 Analyzed: 05/30/26

Formaldehyde	5.26	ug/m ³ Air		5.26			0.101	10.1	
Acetaldehyde	1.68	ug/m ³ Air		1.68			0.0469	10.1	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2802 - TO-11A extraction

Duplicate (B6E2802-DUPI) Continued Source: 6052714-02 Prepared: 05/28/26 Analyzed: 05/30/26

Acetone	2.20	ug/m ³ Air		2.15			2.65	10.1	
Propionaldehyde	0.315	ug/m ³ Air		0.266			16.8	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.139	ug/m ³ Air		0.155			10.8	10.1	
2-Butanone	0.311	ug/m ³ Air		0.316			1.59	10.1	
Benzaldehyde	0.135	ug/m ³ Air		0.131			2.60	10.1	
Valeraldehyde	0.173	ug/m ³ Air		0.200			14.8	10.1	
Hexaldehyde	0.182	ug/m ³ Air		0.221			19.1	10.1	

Batch B6F0104 - TO-11A extraction

Blank (B6F0104-BLK1)

Prepared: 06/01/26 Analyzed: 06/02/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0321	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0448	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6F0104-BS1)

Prepared: 06/01/26 Analyzed: 06/02/26

Formaldehyde	0.283	ug/ml	0.30000		94.2	79.9-120.1			
Acetaldehyde	0.285	ug/ml	0.30000		94.9	79.9-120.1			
Acetone	0.295	ug/ml	0.30000		98.3	69.9-130.1			
Propionaldehyde	0.279	ug/ml	0.30000		93.0	69.9-130.1			
Crotonaldehyde	0.282	ug/ml	0.30000		93.9	69.9-130.1			
Butyraldehyde	0.248	ug/ml	0.30000		82.5	69.9-130.1			
2-Butanone	0.237	ug/ml	0.30000		79.1	69.9-130.1			
Benzaldehyde	0.308	ug/ml	0.30000		103	69.9-130.1			
Valeraldehyde	0.247	ug/ml	0.30000		82.3	69.9-130.1			
Hexaldehyde	0.243	ug/ml	0.30000		80.9	69.9-130.1			

LCS Dup (B6F0104-BSD1)

Prepared: 06/01/26 Analyzed: 06/02/26

Formaldehyde	0.285	ug/ml	0.30000		95.1	79.9-120.1	0.937	20	
Acetaldehyde	0.286	ug/ml	0.30000		95.3	79.9-120.1	0.407	20	
Acetone	0.294	ug/ml	0.30000		98.1	69.9-130.1	0.266	20	
Propionaldehyde	0.277	ug/ml	0.30000		92.2	69.9-130.1	0.859	20	
Crotonaldehyde	0.283	ug/ml	0.30000		94.2	69.9-130.1	0.293	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6F0104 - TO-11A extraction

LCS Dup (B6F0104-BSD1) Continued

Prepared: 06/01/26 Analyzed: 06/02/26

Butyraldehyde	0.252	ug/ml	0.30000		83.8	69.9-130.1	1.61	20	
2-Butanone	0.236	ug/ml	0.30000		78.8	69.9-130.1	0.474	20	
Benzaldehyde	0.304	ug/ml	0.30000		101	69.9-130.1	1.40	20	
Valeraldehyde	0.250	ug/ml	0.30000		83.2	69.9-130.1	1.16	20	
Hexaldehyde	0.244	ug/ml	0.30000		81.2	69.9-130.1	0.313	20	

Batch B6F0406 - TO-11A extraction

Blank (B6F0406-BLK1)

Prepared & Analyzed: 06/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0621	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6F0406-BS1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.284	ug/ml	0.30000		94.5	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.4	79.9-120.1			
Acetone	0.296	ug/ml	0.30000		98.6	69.9-130.1			
Propionaldehyde	0.280	ug/ml	0.30000		93.4	69.9-130.1			
Crotonaldehyde	0.287	ug/ml	0.30000		95.7	69.9-130.1			
Butyraldehyde	0.253	ug/ml	0.30000		84.2	69.9-130.1			
2-Butanone	0.240	ug/ml	0.30000		80.1	69.9-130.1			
Benzaldehyde	0.303	ug/ml	0.30000		101	69.9-130.1			
Valeraldehyde	0.254	ug/ml	0.30000		84.8	69.9-130.1			
Hexaldehyde	0.247	ug/ml	0.30000		82.3	69.9-130.1			

LCS Dup (B6F0406-BSD1)

Prepared & Analyzed: 06/04/26

Formaldehyde	0.283	ug/ml	0.30000		94.4	79.9-120.1	0.0994	20	
Acetaldehyde	0.290	ug/ml	0.30000		96.7	79.9-120.1	0.299	20	
Acetone	0.295	ug/ml	0.30000		98.2	69.9-130.1	0.424	20	
Propionaldehyde	0.267	ug/ml	0.30000		89.1	69.9-130.1	4.70	20	
Crotonaldehyde	0.279	ug/ml	0.30000		93.0	69.9-130.1	2.78	20	
Butyraldehyde	0.251	ug/ml	0.30000		83.7	69.9-130.1	0.669	20	
2-Butanone	0.238	ug/ml	0.30000		79.4	69.9-130.1	0.914	20	
Benzaldehyde	0.305	ug/ml	0.30000		102	69.9-130.1	0.418	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6F0406 - TO-11A extraction

LCS Dup (B6F0406-BSD1) Continued

Prepared & Analyzed: 06/04/26

Valeraldehyde	0.253	ug/ml	0.30000	84.3	69.9-130.1	0.601	20
Hexaldehyde	0.243	ug/ml	0.30000	81.1	69.9-130.1	1.39	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/08/26 09:29

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
O-05	This sample was extracted outside of the EPA recommended holding time.
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
D-F	Duplicate exceeds DQO criteria.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.