



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 16, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/07/26 10:54 through 06/03/26 13:31.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6050702-03	Air	05/02/26 23:59	05/07/26 10:54
TMOK	6051320-03	Air	05/08/26 23:59	05/13/26 12:55
TMOK	6052714-05	Air	05/20/26 23:59	05/27/26 12:59
TMOK	6060308-03	Air	05/26/26 23:59	06/03/26 13:31
TMOK FB	6060308-04	Air	05/26/26 23:59	06/03/26 13:31



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK
Temp Rec'd (°C): 3.4

Lab ID: 6050702-03
Sample Volume: 303.3 m³
Container #: 443

Sampled: 05/02/26 23:59
Received: 05/07/26 10:54
Analyzed: 06/04/26 18:29

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	35.8		1.31		
Acenaphthylene	208-96-8	0.164		0.00427		
Acenaphthene	83-32-9	2.22		0.137		
Fluorene	86-73-7	2.50		0.0508		
Phenanthrene	85-01-8	5.51	LB, QB-03	0.138		
Anthracene	120-12-7	0.206		0.0401		
Fluoranthene	206-44-0	1.78	LB	0.0556		
Pyrene	129-00-0	2.07	LB	0.0610		
Benzo(a)anthracene	56-55-3	0.239		0.0148		
Chrysene	218-01-9	0.597		0.0138		
Benzo(b)fluoranthene	205-99-2	0.155		0.0211		
Benzo(k)fluoranthene	207-08-9	0.0488		0.00890		
Benzo(e)pyrene	192-97-2	0.172		0.00791		
Benzo(a)pyrene	50-32-8	0.0547		0.00909		
Perylene	198-55-0	ND	U	0.00437		
Dibenz(a,h)anthracene	53-70-3	ND	U	0.00685		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0926		0.00892		
Benzo(g,h,i)perylene	191-24-2	0.217		0.0317		
Coronene	191-07-1	0.116		0.0136		
Surrogate: Fluorene-d10		0.915			93.4 %	Limit 60-120
Surrogate: Fluoranthene-d10		3.13			94.8 %	Limit 60-120
Surrogate: Pyrene-d10		0.841			84.2 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		3.08			92.4 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK
Temp Rec'd (°C): 4.9

Lab ID: 6051320-03
Sample Volume: 325 m³
Container #: 54

Sampled: 05/08/26 23:59
Received: 05/13/26 12:55
Analyzed: 06/17/26 07:15

Comments:

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	21.3	LB, TT	1.22		
Acenaphthylene	208-96-8	ND	U, TT	0.00399		
Acenaphthene	83-32-9	1.88	TT	0.128		
Fluorene	86-73-7	2.38	TT	0.0474		
Phenanthrene	85-01-8	5.49	LB, QB-03, TT	0.129		
Anthracene	120-12-7	0.0489	TT	0.0374		
Fluoranthene	206-44-0	1.19	LB, TT	0.0519		
Pyrene	129-00-0	0.568	LB, TT	0.0570		
Benzo(a)anthracene	56-55-3	0.0462	TT	0.0138		
Chrysene	218-01-9	0.125	TT	0.0129		
Benzo(b)fluoranthene	205-99-2	0.0523	TT	0.0197		
Benzo(k)fluoranthene	207-08-9	0.0175	TT	0.00831		
Benzo(e)pyrene	192-97-2	0.0477	TT	0.00738		
Benzo(a)pyrene	50-32-8	0.0225	TT	0.00848		
Perylene	198-55-0	ND	U, TT	0.00408		
Dibenz(a,h)anthracene	53-70-3	ND	U, TT	0.00640		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0277	TT	0.00833		
Benzo(g,h,i)perylene	191-24-2	0.0511	TT	0.0295		
Coronene	191-07-1	ND	U, TT	0.0126		
Surrogate: Fluorene-d10		0.820	TT		83.7 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.68	TT		87.1 %	Limit 60-120
Surrogate: Pyrene-d10		0.723	TT		72.4 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.61	TT		84.0 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK
Temp Rec'd (°C): 3.5

Lab ID: 6052714-05
Sample Volume: 311.6 m³
Container #: 554

Sampled: 05/20/26 23:59
Received: 05/27/26 12:59
Analyzed: 06/30/26 23:24

Comments: CoC arrived separate from Sample

Air Toxics by EPA Compendium Method TO-13

Analyte	CAS Number	Results	Flag	MDL	% Rec.	% Rec Limits
		ng/m ³ Air		ng/m ³ Air		
Naphthalene	91-20-3	8.53		1.27		
Acenaphthylene	208-96-8	0.118		0.00416		
Acenaphthene	83-32-9	0.448		0.134		
Fluorene	86-73-7	0.617		0.0495		
Phenanthrene	85-01-8	1.43	LB, QB-03	0.135		
Anthracene	120-12-7	0.0716		0.0390		
Fluoranthene	206-44-0	0.431	LB	0.0541		
Pyrene	129-00-0	0.280		0.0594		
Benzo(a)anthracene	56-55-3	0.0234		0.0144		
Chrysene	218-01-9	0.0411		0.0135		
Benzo(b)fluoranthene	205-99-2	0.0414		0.0205		
Benzo(k)fluoranthene	207-08-9	0.0144		0.00866		
Benzo(e)pyrene	192-97-2	0.0334		0.00770		
Benzo(a)pyrene	50-32-8	0.0209		0.00885		
Perylene	198-55-0	ND	U	0.00426		
Dibenz(a,h)anthracene	53-70-3	ND	U	0.00667		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0250		0.00868		
Benzo(g,h,i)perylene	191-24-2	0.0475		0.0308		
Coronene	191-07-1	0.0209		0.0132		
Surrogate: Fluorene-d10		0.829			84.6 %	Limit 60-120
Surrogate: Fluoranthene-d10		3.02			94.0 %	Limit 60-120
Surrogate: Pyrene-d10		0.767			76.8 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.77			85.5 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK
Temp Rec'd (°C): 4.8

Lab ID: 6060308-03
Sample Volume: 334.3 m³
Container #: 196

Sampled: 05/26/26 23:59
Received: 06/03/26 13:31
Analyzed: 07/03/26 03:57

Comments:

Air Toxics by EPA Compendium Method TO-13

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>	<u>% Rec.</u>	<u>% Rec Limits</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>		
Naphthalene	91-20-3	29.7	TT	1.18		
Acenaphthylene	208-96-8	ND	U, TT	0.00388		
Acenaphthene	83-32-9	3.37	QB-03, TT	0.125		
Fluorene	86-73-7	3.43	TT	0.0461		
Phenanthrene	85-01-8	6.71	LB, QB-03, TT	0.126		
Anthracene	120-12-7	0.0826	LB, QB-03, TT	0.0363		
Fluoranthene	206-44-0	1.33	LB, QB-03, TT	0.0504		
Pyrene	129-00-0	0.538	LB, QB-03, TT	0.0554		
Benzo(a)anthracene	56-55-3	0.0365	QB-03, TT	0.0135		
Chrysene	218-01-9	0.0751	QB-03, TT	0.0126		
Benzo(b)fluoranthene	205-99-2	0.0900	QB-03, TT	0.0191		
Benzo(k)fluoranthene	207-08-9	0.0215	QB-03, TT	0.00808		
Benzo(e)pyrene	192-97-2	ND	U, TT	0.00718		
Benzo(a)pyrene	50-32-8	0.0344	QB-03, TT	0.00825		
Perylene	198-55-0	ND	U, TT	0.00397		
Dibenz(a,h)anthracene	53-70-3	ND	U, TT	0.00622		
Indeno(1,2,3-c,d)pyrene	193-39-5	0.0646	QB-03, TT	0.00809		
Benzo(g,h,i)perylene	191-24-2	0.0912	QB-03, TT	0.0287		
Coronene	191-07-1	0.0553	TT	0.0123		
Surrogate: Fluorene-d10		0.950	TT		96.9 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.43	TT		81.2 %	Limit 60-120
Surrogate: Pyrene-d10		0.672	TT		67.2 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.68	TT		88.9 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK FB
Temp Rec'd (°C): 4.8

Lab ID: 6060308-04
Sample Volume: 334.3 m³
Container #: 510

Sampled: 05/26/26 23:59
Received: 06/03/26 13:31
Analyzed: 07/03/26 04:50

Comments:

Air Toxics by EPA Compendium Method TO-13

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>	<u>% Rec.</u>	<u>% Rec Limits</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>		
Naphthalene	91-20-3	0.311	U, TT	1.18		
Acenaphthylene	208-96-8	ND	U, TT	0.00388		
Acenaphthene	83-32-9	ND	U, TT	0.125		
Fluorene	86-73-7	0.0120	U, TT	0.0461		
Phenanthrene	85-01-8	0.0344	U, LB, QB-03, TT	0.126		
Anthracene	120-12-7	ND	U, TT	0.0363		
Fluoranthene	206-44-0	0.0126	U, LB, QB-03, TT	0.0504		
Pyrene	129-00-0	0.00927	U, LB, QB-03, TT	0.0554		
Benzo(a)anthracene	56-55-3	ND	U, TT	0.0135		
Chrysene	218-01-9	ND	U, TT	0.0126		
Benzo(b)fluoranthene	205-99-2	ND	U, TT	0.0191		
Benzo(k)fluoranthene	207-08-9	ND	U, TT	0.00808		
Benzo(e)pyrene	192-97-2	ND	U, TT	0.00718		
Benzo(a)pyrene	50-32-8	ND	U, TT	0.00825		
Perylene	198-55-0	ND	U, TT	0.00397		
Dibenz(a,h)anthracene	53-70-3	ND	U, TT	0.00622		
Indeno(1,2,3-c,d)pyrene	193-39-5	ND	U, TT	0.00809		
Benzo(g,h,i)perylene	191-24-2	ND	U, TT	0.0287		
Coronene	191-07-1	ND	U, TT	0.0123		
Surrogate: Fluorene-d10		0.940	TT		95.9 %	Limit 60-120
Surrogate: Fluoranthene-d10		2.56	TT		85.7 %	Limit 60-120
Surrogate: Pyrene-d10		0.672	TT		67.3 %	Limit 60-120
Surrogate: Benzo(a)pyrene-d12		2.70	TT		89.2 %	Limit 60-120



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606004 - B6E1305

Calibration Blank (2606004-CCB1)

Prepared & Analyzed: 06/04/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	0.00	ng/ul		ND					U
Phenanthrene	6.00E-4	ng/ul		ND					
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	3.00E-4	ng/ul		ND					LB
Pyrene	3.00E-4	ng/ul		ND					LB
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2606004-CCV1)

Prepared & Analyzed: 06/04/26

Naphthalene	7.56	ng/ul	7.5065	ND	101	70-130			
Acenaphthylene	0.530	ng/ul	0.50200	ND	106	70-130			
Acenaphthene	2.51	ng/ul	2.5030	ND	100	70-130			
Fluorene	2.50	ng/ul	2.5022	ND	99.7	70-130			
Phenanthrene	3.87	ng/ul	3.7538	ND	103	70-130			
Anthracene	0.494	ng/ul	0.50125	ND	98.5	70-130			
Fluoranthene	2.39	ng/ul	2.5015	ND	95.7	70-130			LB
Pyrene	1.32	ng/ul	1.2520	ND	106	70-130			LB
Benzo(a)anthracene	0.520	ng/ul	0.50200	ND	104	70-130			
Chrysene	0.516	ng/ul	0.50200	ND	103	70-130			
Benzo(b)fluoranthene	0.514	ng/ul	0.50125	ND	103	70-130			
Benzo(k)fluoranthene	0.515	ng/ul	0.50125	ND	103	70-130			
Benzo(e)pyrene	0.514	ng/ul	0.50125	ND	103	70-130			
Benzo(a)pyrene	0.516	ng/ul	0.50225	ND	103	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606004 - B6E1305

Calibration Check (2606004-CCV1) Contin

Prepared & Analyzed: 06/04/26

Perylene	0.515	ng/ul	0.50125	ND	103	70-130
Dibenz(a,h)anthracene	0.533	ng/ul	0.50275	ND	106	70-130
Indeno(1,2,3-c,d)pyrene	0.535	ng/ul	0.50175	ND	107	70-130
Benzo(g,h,i)perylene	0.540	ng/ul	0.50150	ND	108	70-130
Coronene	0.544	ng/ul	0.50175	ND	108	70-130
Surrogate: Fluorene-d10	0.492	ng/ul	0.49000	ND	101	70-130
Surrogate: Fluoranthene-d10	0.473	ng/ul	0.50000	ND	94.6	70-130
Surrogate: Pyrene-d10	0.525	ng/ul	0.49950	ND	105	70-130
Surrogate: Benzo(a)pyrene-d12	0.520	ng/ul	0.50500	ND	103	70-130

Calibration Check (2606004-CCV2)

Prepared & Analyzed: 06/04/26

Naphthalene	7.58	ng/ul	7.5065	ND	101	70-130
Acenaphthylene	0.501	ng/ul	0.50200	ND	99.8	70-130
Acenaphthene	2.56	ng/ul	2.5030	ND	102	70-130
Fluorene	2.58	ng/ul	2.5022	ND	103	70-130
Phenanthrene	3.85	ng/ul	3.7538	ND	103	70-130
Anthracene	0.514	ng/ul	0.50125	ND	103	70-130
Fluoranthene	2.49	ng/ul	2.5015	ND	99.5	70-130
Pyrene	1.40	ng/ul	1.2520	ND	112	70-130
Benzo(a)anthracene	0.524	ng/ul	0.50200	ND	104	70-130
Chrysene	0.515	ng/ul	0.50200	ND	103	70-130
Benzo(b)fluoranthene	0.516	ng/ul	0.50125	ND	103	70-130
Benzo(k)fluoranthene	0.517	ng/ul	0.50125	ND	103	70-130
Benzo(e)pyrene	0.517	ng/ul	0.50125	ND	103	70-130
Benzo(a)pyrene	0.514	ng/ul	0.50225	ND	102	70-130
Perylene	0.514	ng/ul	0.50125	ND	103	70-130
Dibenz(a,h)anthracene	0.498	ng/ul	0.50275	ND	99.0	70-130
Indeno(1,2,3-c,d)pyrene	0.500	ng/ul	0.50175	ND	99.6	70-130
Benzo(g,h,i)perylene	0.510	ng/ul	0.50150	ND	102	70-130
Coronene	0.502	ng/ul	0.50175	ND	100	70-130
Surrogate: Fluorene-d10	0.497	ng/ul	0.49000	ND	101	70-130
Surrogate: Fluoranthene-d10	0.496	ng/ul	0.50000	ND	99.2	70-130
Surrogate: Pyrene-d10	0.554	ng/ul	0.49950	ND	111	70-130
Surrogate: Benzo(a)pyrene-d12	0.518	ng/ul	0.50500	ND	103	70-130

Calibration Check (2606004-CCV3)

Prepared: 06/04/26 Analyzed: 06/05/26

Naphthalene	7.55	ng/ul	7.5065	ND	101	70-130
Acenaphthylene	0.503	ng/ul	0.50200	ND	100	70-130
Acenaphthene	2.56	ng/ul	2.5030	ND	102	70-130
Fluorene	2.58	ng/ul	2.5022	ND	103	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606004 - B6E1305

Calibration Check (2606004-CCV3) Contin

Prepared: 06/04/26 Analyzed: 06/05/26

Phenanthrene	3.84	ng/ul	3.7538	ND	102	70-130			
Anthracene	0.525	ng/ul	0.50125	ND	105	70-130			
Fluoranthene	2.61	ng/ul	2.5015	ND	104	70-130			LB
Pyrene	1.24	ng/ul	1.2520	ND	99.1	70-130			LB
Benzo(a)anthracene	0.523	ng/ul	0.50200	ND	104	70-130			
Chrysene	0.514	ng/ul	0.50200	ND	102	70-130			
Benzo(b)fluoranthene	0.511	ng/ul	0.50125	ND	102	70-130			
Benzo(k)fluoranthene	0.519	ng/ul	0.50125	ND	104	70-130			
Benzo(e)pyrene	0.507	ng/ul	0.50125	ND	101	70-130			
Benzo(a)pyrene	0.515	ng/ul	0.50225	ND	103	70-130			
Perylene	0.514	ng/ul	0.50125	ND	102	70-130			
Dibenz(a,h)anthracene	0.514	ng/ul	0.50275	ND	102	70-130			
Indeno(1,2,3-c,d)pyrene	0.510	ng/ul	0.50175	ND	102	70-130			
Benzo(g,h,i)perylene	0.511	ng/ul	0.50150	ND	102	70-130			
Coronene	0.493	ng/ul	0.50175	ND	98.3	70-130			

Surrogate: Fluorene-d10 0.499 ng/ul 0.49000 ND 102 70-130

Surrogate: Fluoranthene-d10 0.520 ng/ul 0.50000 ND 104 70-130

Surrogate: Pyrene-d10 0.497 ng/ul 0.49950 ND 99.5 70-130

Surrogate: Benzo(a)pyrene-d12 0.524 ng/ul 0.50500 ND 104 70-130

Batch 2606056 - B6E1507

Calibration Blank (2606056-CCB1)

Prepared: 06/16/26 Analyzed: 06/17/26

Naphthalene	0.00330	ng/ul		ND					
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00140	ng/ul		ND					
Fluorene	0.00120	ng/ul		ND					
Phenanthrene	0.00300	ng/ul		ND					
Anthracene	0.00	ng/ul		ND					U
Fluoranthene	0.00190	ng/ul		ND					
Pyrene	9.00E-4	ng/ul		ND					LB
Benzo(a)anthracene	2.00E-4	ng/ul		ND					
Chrysene	3.00E-4	ng/ul		ND					
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606056 - B6E1507

Calibration Blank (2606056-CCB1) Contin

Prepared: 06/16/26 Analyzed: 06/17/26

Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2606056-CCV1)

Prepared & Analyzed: 06/16/26

Naphthalene	6.23	ng/ul	7.5065	ND	83.0	70-130			
Acenaphthylene	0.418	ng/ul	0.50200	ND	83.3	70-130			
Acenaphthene	2.07	ng/ul	2.5030	ND	82.8	70-130			
Fluorene	2.08	ng/ul	2.5022	ND	82.9	70-130			
Phenanthrene	3.12	ng/ul	3.7538	ND	83.2	70-130			
Anthracene	0.410	ng/ul	0.50125	ND	81.7	70-130			
Fluoranthene	2.06	ng/ul	2.5015	ND	82.3	70-130			
Pyrene	1.01	ng/ul	1.2520	ND	81.0	70-130			LB
Benzo(a)anthracene	0.416	ng/ul	0.50200	ND	82.8	70-130			
Chrysene	0.413	ng/ul	0.50200	ND	82.3	70-130			
Benzo(b)fluoranthene	0.416	ng/ul	0.50125	ND	83.1	70-130			
Benzo(k)fluoranthene	0.415	ng/ul	0.50125	ND	82.8	70-130			
Benzo(e)pyrene	0.415	ng/ul	0.50125	ND	82.8	70-130			
Benzo(a)pyrene	0.416	ng/ul	0.50225	ND	82.9	70-130			
Perylene	0.413	ng/ul	0.50125	ND	82.4	70-130			
Dibenz(a,h)anthracene	0.424	ng/ul	0.50275	ND	84.4	70-130			
Indeno(1,2,3-c,d)pyrene	0.424	ng/ul	0.50175	ND	84.5	70-130			
Benzo(g,h,i)perylene	0.424	ng/ul	0.50150	ND	84.6	70-130			
Coronene	0.441	ng/ul	0.50175	ND	87.9	70-130			
Surrogate: Fluorene-d10	0.483	ng/ul	0.49000	ND	98.5	70-130			
Surrogate: Fluoranthene-d10	0.496	ng/ul	0.51000	ND	97.3	70-130			
Surrogate: Pyrene-d10	0.481	ng/ul	0.49950	ND	96.3	70-130			
Surrogate: Benzo(a)pyrene-d12	0.500	ng/ul	0.49850	ND	100	70-130			

Calibration Check (2606056-CCV2)

Prepared & Analyzed: 06/16/26

Naphthalene	6.25	ng/ul	7.5065	ND	83.3	70-130			
Acenaphthylene	0.412	ng/ul	0.50200	ND	82.1	70-130			
Acenaphthene	2.09	ng/ul	2.5030	ND	83.6	70-130			
Fluorene	2.11	ng/ul	2.5022	ND	84.4	70-130			
Phenanthrene	3.10	ng/ul	3.7538	ND	82.7	70-130			
Anthracene	0.418	ng/ul	0.50125	ND	83.3	70-130			
Fluoranthene	2.07	ng/ul	2.5015	ND	82.7	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606056 - B6E1507

Calibration Check (2606056-CCV2) Contin

Prepared & Analyzed: 06/16/26

Pyrene	1.01	ng/ul	1.2520	ND	80.3	70-130			LB
Benzo(a)anthracene	0.420	ng/ul	0.50200	ND	83.7	70-130			
Chrysene	0.412	ng/ul	0.50200	ND	82.0	70-130			
Benzo(b)fluoranthene	0.415	ng/ul	0.50125	ND	82.8	70-130			
Benzo(k)fluoranthene	0.417	ng/ul	0.50125	ND	83.2	70-130			
Benzo(e)pyrene	0.408	ng/ul	0.50125	ND	81.3	70-130			
Benzo(a)pyrene	0.413	ng/ul	0.50225	ND	82.3	70-130			
Perylene	0.410	ng/ul	0.50125	ND	81.7	70-130			
Dibenz(a,h)anthracene	0.414	ng/ul	0.50275	ND	82.4	70-130			
Indeno(1,2,3-c,d)pyrene	0.412	ng/ul	0.50175	ND	82.2	70-130			
Benzo(g,h,i)perylene	0.412	ng/ul	0.50150	ND	82.2	70-130			
Coronene	0.421	ng/ul	0.50175	ND	83.9	70-130			

Surrogate: Fluorene-d10

0.487

ng/ul

0.49000

ND

99.4

70-130

Surrogate: Fluoranthene-d10

0.501

ng/ul

0.51000

ND

98.3

70-130

Surrogate: Pyrene-d10

0.481

ng/ul

0.49950

ND

96.3

70-130

Surrogate: Benzo(a)pyrene-d12

0.501

ng/ul

0.49850

ND

101

70-130

Calibration Check (2606056-CCV3)

Prepared: 06/16/26 Analyzed: 06/17/26

Naphthalene	6.23	ng/ul	7.5065	ND	82.9	70-130			
Acenaphthylene	0.412	ng/ul	0.50200	ND	82.0	70-130			
Acenaphthene	2.09	ng/ul	2.5030	ND	83.4	70-130			
Fluorene	2.10	ng/ul	2.5022	ND	83.8	70-130			
Phenanthrene	3.11	ng/ul	3.7538	ND	82.8	70-130			
Anthracene	0.436	ng/ul	0.50125	ND	86.9	70-130			
Fluoranthene	2.14	ng/ul	2.5015	ND	85.7	70-130			
Pyrene	0.969	ng/ul	1.2520	ND	77.4	70-130			LB
Benzo(a)anthracene	0.419	ng/ul	0.50200	ND	83.5	70-130			
Chrysene	0.410	ng/ul	0.50200	ND	81.7	70-130			
Benzo(b)fluoranthene	0.410	ng/ul	0.50125	ND	81.7	70-130			
Benzo(k)fluoranthene	0.416	ng/ul	0.50125	ND	82.9	70-130			
Benzo(e)pyrene	0.406	ng/ul	0.50125	ND	80.9	70-130			
Benzo(a)pyrene	0.413	ng/ul	0.50225	ND	82.3	70-130			
Perylene	0.411	ng/ul	0.50125	ND	81.9	70-130			
Dibenz(a,h)anthracene	0.416	ng/ul	0.50275	ND	82.8	70-130			
Indeno(1,2,3-c,d)pyrene	0.413	ng/ul	0.50175	ND	82.3	70-130			
Benzo(g,h,i)perylene	0.409	ng/ul	0.50150	ND	81.5	70-130			
Coronene	0.407	ng/ul	0.50175	ND	81.1	70-130			

Surrogate: Fluorene-d10

0.487

ng/ul

0.49000

ND

99.3

70-130

Surrogate: Fluoranthene-d10

0.523

ng/ul

0.51000

ND

103

70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606056 - B6E1507

Calibration Check (2606056-CCV3) Contin

Prepared: 06/16/26 Analyzed: 06/17/26

Surrogate: Pyrene-d10	0.464	ng/ul	0.49950	ND	92.8	70-130
Surrogate: Benzo(a)pyrene-d12	0.506	ng/ul	0.49850	ND	101	70-130

Calibration Check (2606056-CCV4)

Prepared: 06/16/26 Analyzed: 06/17/26

Naphthalene	6.24	ng/ul	7.5065	ND	83.2	70-130
Acenaphthylene	0.411	ng/ul	0.50200	ND	81.8	70-130
Acenaphthene	2.09	ng/ul	2.5030	ND	83.5	70-130
Fluorene	2.08	ng/ul	2.5022	ND	83.1	70-130
Phenanthrene	3.09	ng/ul	3.7538	ND	82.2	70-130
Anthracene	0.426	ng/ul	0.50125	ND	85.1	70-130
Fluoranthene	2.12	ng/ul	2.5015	ND	84.8	70-130
Pyrene	0.999	ng/ul	1.2520	ND	79.8	70-130
Benzo(a)anthracene	0.415	ng/ul	0.50200	ND	82.7	70-130
Chrysene	0.412	ng/ul	0.50200	ND	82.1	70-130
Benzo(b)fluoranthene	0.407	ng/ul	0.50125	ND	81.1	70-130
Benzo(k)fluoranthene	0.411	ng/ul	0.50125	ND	82.0	70-130
Benzo(e)pyrene	0.408	ng/ul	0.50125	ND	81.4	70-130
Benzo(a)pyrene	0.411	ng/ul	0.50225	ND	81.8	70-130
Perylene	0.409	ng/ul	0.50125	ND	81.6	70-130
Dibenz(a,h)anthracene	0.410	ng/ul	0.50275	ND	81.6	70-130
Indeno(1,2,3-c,d)pyrene	0.409	ng/ul	0.50175	ND	81.6	70-130
Benzo(g,h,i)perylene	0.408	ng/ul	0.50150	ND	81.3	70-130
Coronene	0.402	ng/ul	0.50175	ND	80.0	70-130
Surrogate: Fluorene-d10	0.483	ng/ul	0.49000	ND	98.6	70-130
Surrogate: Fluoranthene-d10	0.513	ng/ul	0.51000	ND	101	70-130
Surrogate: Pyrene-d10	0.479	ng/ul	0.49950	ND	95.9	70-130
Surrogate: Benzo(a)pyrene-d12	0.499	ng/ul	0.49850	ND	100	70-130

LB

Batch 2606119 - B6F0203

Calibration Blank (2606119-CCB1)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	6.00E-4	ng/ul	ND	
Acenaphthylene	0.00	ng/ul	ND	U
Acenaphthene	0.00	ng/ul	ND	U
Fluorene	0.00	ng/ul	ND	U
Phenanthrene	6.00E-4	ng/ul	ND	
Anthracene	0.00	ng/ul	ND	U
Fluoranthene	0.00	ng/ul	ND	U
Pyrene	0.00	ng/ul	ND	U
Benzo(a)anthracene	0.00	ng/ul	ND	U
Chrysene	0.00	ng/ul	ND	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606119 - B6F0203

Calibration Blank (2606119-CCB1) Contin

Prepared: 06/30/26 Analyzed: 07/01/26

Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Blank (2606119-CCB2)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	0.00290	ng/ul		ND					
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00120	ng/ul		ND					
Fluorene	0.00100	ng/ul		ND					
Phenanthrene	0.00210	ng/ul		ND					
Anthracene	4.00E-4	ng/ul		ND					
Fluoranthene	0.00210	ng/ul		ND					LB
Pyrene	0.00100	ng/ul		ND					
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	4.00E-4	ng/ul		ND					
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U
Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2606119-CCV1)

Prepared & Analyzed: 06/30/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606119 - B6F0203

Calibration Check (2606119-CCV1) Contin

Prepared & Analyzed: 06/30/26

Naphthalene	7.33	ng/ul	7.5065	ND	97.6	70-130			
Acenaphthylene	0.513	ng/ul	0.50200	ND	102	70-130			
Acenaphthene	2.44	ng/ul	2.5030	ND	97.4	70-130			
Fluorene	2.44	ng/ul	2.5022	ND	97.6	70-130			
Phenanthrene	3.67	ng/ul	3.7538	ND	97.7	70-130			
Anthracene	0.484	ng/ul	0.50125	ND	96.6	70-130			
Fluoranthene	2.51	ng/ul	2.5015	ND	100	70-130			LB
Pyrene	1.12	ng/ul	1.2520	ND	89.8	70-130			
Benzo(a)anthracene	0.485	ng/ul	0.50200	ND	96.7	70-130			
Chrysene	0.484	ng/ul	0.50200	ND	96.3	70-130			
Benzo(b)fluoranthene	0.483	ng/ul	0.50125	ND	96.4	70-130			
Benzo(k)fluoranthene	0.483	ng/ul	0.50125	ND	96.4	70-130			
Benzo(e)pyrene	0.479	ng/ul	0.50125	ND	95.6	70-130			
Benzo(a)pyrene	0.486	ng/ul	0.50225	ND	96.8	70-130			
Perylene	0.482	ng/ul	0.50125	ND	96.3	70-130			
Dibenz(a,h)anthracene	0.517	ng/ul	0.50275	ND	103	70-130			
Indeno(1,2,3-c,d)pyrene	0.509	ng/ul	0.50175	ND	102	70-130			
Benzo(g,h,i)perylene	0.497	ng/ul	0.50150	ND	99.1	70-130			
Coronene	0.519	ng/ul	0.50175	ND	103	70-130			
Surrogate: Fluorene-d10	0.481	ng/ul	0.49000	ND	98.2	70-130			
Surrogate: Fluoranthene-d10	0.509	ng/ul	0.51000	ND	99.9	70-130			
Surrogate: Pyrene-d10	0.452	ng/ul	0.49950	ND	90.5	70-130			
Surrogate: Benzo(a)pyrene-d12	0.487	ng/ul	0.49850	ND	97.6	70-130			

Calibration Check (2606119-CCV2)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.32	ng/ul	7.5065	ND	97.6	70-130			
Acenaphthylene	0.492	ng/ul	0.50200	ND	98.0	70-130			
Acenaphthene	2.47	ng/ul	2.5030	ND	98.6	70-130			
Fluorene	2.49	ng/ul	2.5022	ND	99.4	70-130			
Phenanthrene	3.64	ng/ul	3.7538	ND	96.9	70-130			
Anthracene	0.501	ng/ul	0.50125	ND	100	70-130			
Fluoranthene	2.58	ng/ul	2.5015	ND	103	70-130			LB
Pyrene	1.05	ng/ul	1.2520	ND	84.1	70-130			
Benzo(a)anthracene	0.494	ng/ul	0.50200	ND	98.4	70-130			
Chrysene	0.484	ng/ul	0.50200	ND	96.4	70-130			
Benzo(b)fluoranthene	0.485	ng/ul	0.50125	ND	96.7	70-130			
Benzo(k)fluoranthene	0.487	ng/ul	0.50125	ND	97.2	70-130			
Benzo(e)pyrene	0.471	ng/ul	0.50125	ND	93.9	70-130			
Benzo(a)pyrene	0.489	ng/ul	0.50225	ND	97.4	70-130			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606119 - B6F0203

Calibration Check (2606119-CCV2) Contin

Prepared: 06/30/26 Analyzed: 07/01/26

Perylene	0.483	ng/ul	0.50125	ND	96.3	70-130
Dibenz(a,h)anthracene	0.498	ng/ul	0.50275	ND	99.0	70-130
Indeno(1,2,3-c,d)pyrene	0.490	ng/ul	0.50175	ND	97.6	70-130
Benzo(g,h,i)perylene	0.478	ng/ul	0.50150	ND	95.3	70-130
Coronene	0.482	ng/ul	0.50175	ND	96.1	70-130

Surrogate: Fluorene-d10	0.484	ng/ul	0.49000	ND	98.9	70-130
Surrogate: Fluoranthene-d10	0.525	ng/ul	0.51000	ND	103	70-130
Surrogate: Pyrene-d10	0.420	ng/ul	0.49950	ND	84.1	70-130
Surrogate: Benzo(a)pyrene-d12	0.489	ng/ul	0.49850	ND	98.1	70-130

Calibration Check (2606119-CCV3)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.34	ng/ul	7.5065	ND	97.7	70-130
Acenaphthylene	0.491	ng/ul	0.50200	ND	97.8	70-130
Acenaphthene	2.47	ng/ul	2.5030	ND	98.6	70-130
Fluorene	2.47	ng/ul	2.5022	ND	98.8	70-130
Phenanthrene	3.66	ng/ul	3.7538	ND	97.5	70-130
Anthracene	0.507	ng/ul	0.50125	ND	101	70-130
Fluoranthene	2.56	ng/ul	2.5015	ND	102	70-130
Pyrene	1.06	ng/ul	1.2520	ND	84.9	70-130
Benzo(a)anthracene	0.488	ng/ul	0.50200	ND	97.2	70-130
Chrysene	0.480	ng/ul	0.50200	ND	95.5	70-130
Benzo(b)fluoranthene	0.482	ng/ul	0.50125	ND	96.1	70-130
Benzo(k)fluoranthene	0.486	ng/ul	0.50125	ND	96.9	70-130
Benzo(e)pyrene	0.472	ng/ul	0.50125	ND	94.3	70-130
Benzo(a)pyrene	0.487	ng/ul	0.50225	ND	96.9	70-130
Perylene	0.481	ng/ul	0.50125	ND	95.9	70-130
Dibenz(a,h)anthracene	0.493	ng/ul	0.50275	ND	98.1	70-130
Indeno(1,2,3-c,d)pyrene	0.487	ng/ul	0.50175	ND	97.1	70-130
Benzo(g,h,i)perylene	0.480	ng/ul	0.50150	ND	95.7	70-130
Coronene	0.476	ng/ul	0.50175	ND	94.8	70-130

Surrogate: Fluorene-d10	0.483	ng/ul	0.49000	ND	98.5	70-130
Surrogate: Fluoranthene-d10	0.520	ng/ul	0.51000	ND	102	70-130
Surrogate: Pyrene-d10	0.425	ng/ul	0.49950	ND	85.0	70-130
Surrogate: Benzo(a)pyrene-d12	0.488	ng/ul	0.49850	ND	97.9	70-130

Calibration Check (2606119-CCV4)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.31	ng/ul	7.5065	ND	97.4	70-130
Acenaphthylene	0.490	ng/ul	0.50200	ND	97.5	70-130
Acenaphthene	2.46	ng/ul	2.5030	ND	98.3	70-130
Fluorene	2.46	ng/ul	2.5022	ND	98.5	70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2606119 - B6F0203

Calibration Check (2606119-CCV4) Contin

Prepared: 06/30/26 Analyzed: 07/01/26

Phenanthrene	3.64	ng/ul	3.7538	ND	96.9	70-130			
Anthracene	0.500	ng/ul	0.50125	ND	99.7	70-130			
Fluoranthene	2.54	ng/ul	2.5015	ND	101	70-130			LB
Pyrene	1.07	ng/ul	1.2520	ND	85.2	70-130			
Benzo(a)anthracene	0.489	ng/ul	0.50200	ND	97.4	70-130			
Chrysene	0.482	ng/ul	0.50200	ND	96.1	70-130			
Benzo(b)fluoranthene	0.481	ng/ul	0.50125	ND	95.9	70-130			
Benzo(k)fluoranthene	0.485	ng/ul	0.50125	ND	96.8	70-130			
Benzo(e)pyrene	0.474	ng/ul	0.50125	ND	94.6	70-130			
Benzo(a)pyrene	0.484	ng/ul	0.50225	ND	96.3	70-130			
Perylene	0.482	ng/ul	0.50125	ND	96.3	70-130			
Dibenz(a,h)anthracene	0.495	ng/ul	0.50275	ND	98.4	70-130			
Indeno(1,2,3-c,d)pyrene	0.490	ng/ul	0.50175	ND	97.7	70-130			
Benzo(g,h,i)perylene	0.481	ng/ul	0.50150	ND	95.9	70-130			
Coronene	0.476	ng/ul	0.50175	ND	94.9	70-130			
Surrogate: Fluorene-d10	0.481	ng/ul	0.49000	ND	98.2	70-130			
Surrogate: Fluoranthene-d10	0.518	ng/ul	0.51000	ND	102	70-130			
Surrogate: Pyrene-d10	0.429	ng/ul	0.49950	ND	85.9	70-130			
Surrogate: Benzo(a)pyrene-d12	0.488	ng/ul	0.49850	ND	97.9	70-130			

Batch 2607007 - B6F0404

Calibration Blank (2607007-CCB1)

Prepared: 07/02/26 Analyzed: 07/03/26

Naphthalene	0.00	ng/ul		ND					U
Acenaphthylene	0.00	ng/ul		ND					U
Acenaphthene	0.00	ng/ul		ND					U
Fluorene	9.00E-4	ng/ul		ND					
Phenanthrene	0.00340	ng/ul		ND					LB
Anthracene	0.00140	ng/ul		ND					LB
Fluoranthene	0.00200	ng/ul		ND					LB
Pyrene	0.00100	ng/ul		ND					LB
Benzo(a)anthracene	0.00	ng/ul		ND					U
Chrysene	0.00	ng/ul		ND					U
Benzo(b)fluoranthene	0.00	ng/ul		ND					U
Benzo(k)fluoranthene	0.00	ng/ul		ND					U
Benzo(e)pyrene	0.00	ng/ul		ND					U
Benzo(a)pyrene	0.00	ng/ul		ND					U
Perylene	0.00	ng/ul		ND					U
Dibenz(a,h)anthracene	0.00	ng/ul		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/ul		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607007 - B6F0404

Calibration Blank (2607007-CCB1) Contin

Prepared: 07/02/26 Analyzed: 07/03/26

Benzo(g,h,i)perylene	0.00	ng/ul		ND					U
Coronene	0.00	ng/ul		ND					U
Surrogate: Fluorene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/ul		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/ul		ND		60-120			U

Calibration Check (2607007-CCV1)

Prepared & Analyzed: 07/02/26

Naphthalene	7.31	ng/ul	7.5065	ND	97.4	70-130			
Acenaphthylene	0.496	ng/ul	0.50200	ND	98.7	70-130			
Acenaphthene	2.45	ng/ul	2.5030	ND	97.7	70-130			
Fluorene	2.60	ng/ul	2.5022	ND	104	70-130			
Phenanthrene	3.66	ng/ul	3.7538	ND	97.5	70-130			LB
Anthracene	0.495	ng/ul	0.50125	ND	98.7	70-130			LB
Fluoranthene	2.49	ng/ul	2.5015	ND	99.4	70-130			LB
Pyrene	1.01	ng/ul	1.2520	ND	80.6	70-130			LB
Benzo(a)anthracene	0.522	ng/ul	0.50200	ND	104	70-130			
Chrysene	0.488	ng/ul	0.50200	ND	97.1	70-130			
Benzo(b)fluoranthene	0.506	ng/ul	0.50125	ND	101	70-130			
Benzo(k)fluoranthene	0.537	ng/ul	0.50125	ND	107	70-130			
Benzo(e)pyrene	0.465	ng/ul	0.50125	ND	92.8	70-130			
Benzo(a)pyrene	0.523	ng/ul	0.50225	ND	104	70-130			
Perylene	0.485	ng/ul	0.50125	ND	96.7	70-130			
Dibenz(a,h)anthracene	0.550	ng/ul	0.50275	ND	109	70-130			
Indeno(1,2,3-c,d)pyrene	0.545	ng/ul	0.50175	ND	109	70-130			
Benzo(g,h,i)perylene	0.506	ng/ul	0.50150	ND	101	70-130			
Coronene	0.520	ng/ul	0.50175	ND	104	70-130			
Surrogate: Fluorene-d10	0.558	ng/ul	0.49000	ND	114	70-130			
Surrogate: Fluoranthene-d10	0.516	ng/ul	0.51000	ND	101	70-130			
Surrogate: Pyrene-d10	0.408	ng/ul	0.49950	ND	81.7	70-130			
Surrogate: Benzo(a)pyrene-d12	0.517	ng/ul	0.49850	ND	104	70-130			

Calibration Check (2607007-CCV2)

Prepared & Analyzed: 07/02/26

Naphthalene	7.37	ng/ul	7.5065	ND	98.2	70-130			
Acenaphthylene	0.496	ng/ul	0.50200	ND	98.8	70-130			
Acenaphthene	2.47	ng/ul	2.5030	ND	98.7	70-130			
Fluorene	2.58	ng/ul	2.5022	ND	103	70-130			
Phenanthrene	3.63	ng/ul	3.7538	ND	96.7	70-130			LB
Anthracene	0.495	ng/ul	0.50125	ND	98.8	70-130			LB
Fluoranthene	2.45	ng/ul	2.5015	ND	98.1	70-130			LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607007 - B6F0404

Calibration Check (2607007-CCV2) Contin

Prepared & Analyzed: 07/02/26

Pyrene	0.988	ng/ul	1.2520	ND	79.0	70-130			LB
Benzo(a)anthracene	0.516	ng/ul	0.50200	ND	103	70-130			
Chrysene	0.483	ng/ul	0.50200	ND	96.2	70-130			
Benzo(b)fluoranthene	0.498	ng/ul	0.50125	ND	99.4	70-130			
Benzo(k)fluoranthene	0.530	ng/ul	0.50125	ND	106	70-130			
Benzo(e)pyrene	0.457	ng/ul	0.50125	ND	91.3	70-130			
Benzo(a)pyrene	0.524	ng/ul	0.50225	ND	104	70-130			
Perylene	0.482	ng/ul	0.50125	ND	96.1	70-130			
Dibenz(a,h)anthracene	0.554	ng/ul	0.50275	ND	110	70-130			
Indeno(1,2,3-c,d)pyrene	0.551	ng/ul	0.50175	ND	110	70-130			
Benzo(g,h,i)perylene	0.508	ng/ul	0.50150	ND	101	70-130			
Coronene	0.535	ng/ul	0.50175	ND	107	70-130			

Surrogate: Fluorene-d10

0.563

ng/ul

0.49000

ND

115

70-130

Surrogate: Fluoranthene-d10

0.516

ng/ul

0.51000

ND

101

70-130

Surrogate: Pyrene-d10

0.401

ng/ul

0.49950

ND

80.4

70-130

Surrogate: Benzo(a)pyrene-d12

0.517

ng/ul

0.49850

ND

104

70-130

Calibration Check (2607007-CCV3)

Prepared: 07/02/26 Analyzed: 07/03/26

Naphthalene	7.37	ng/ul	7.5065	ND	98.2	70-130			
Acenaphthylene	0.492	ng/ul	0.50200	ND	98.0	70-130			
Acenaphthene	2.44	ng/ul	2.5030	ND	97.4	70-130			
Fluorene	2.58	ng/ul	2.5022	ND	103	70-130			
Phenanthrene	3.61	ng/ul	3.7538	ND	96.3	70-130			LB
Anthracene	0.488	ng/ul	0.50125	ND	97.3	70-130			LB
Fluoranthene	2.45	ng/ul	2.5015	ND	97.9	70-130			LB
Pyrene	0.987	ng/ul	1.2520	ND	78.8	70-130			LB
Benzo(a)anthracene	0.516	ng/ul	0.50200	ND	103	70-130			
Chrysene	0.485	ng/ul	0.50200	ND	96.7	70-130			
Benzo(b)fluoranthene	0.506	ng/ul	0.50125	ND	101	70-130			
Benzo(k)fluoranthene	0.542	ng/ul	0.50125	ND	108	70-130			
Benzo(e)pyrene	0.461	ng/ul	0.50125	ND	92.0	70-130			
Benzo(a)pyrene	0.520	ng/ul	0.50225	ND	103	70-130			
Perylene	0.487	ng/ul	0.50125	ND	97.2	70-130			
Dibenz(a,h)anthracene	0.547	ng/ul	0.50275	ND	109	70-130			
Indeno(1,2,3-c,d)pyrene	0.542	ng/ul	0.50175	ND	108	70-130			
Benzo(g,h,i)perylene	0.506	ng/ul	0.50150	ND	101	70-130			
Coronene	0.528	ng/ul	0.50175	ND	105	70-130			

Surrogate: Fluorene-d10

0.555

ng/ul

0.49000

ND

113

70-130

Surrogate: Fluoranthene-d10

0.508

ng/ul

0.51000

ND

99.6

70-130

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch 2607007 - B6F0404

Calibration Check (2607007-CCV3) Contin

Prepared: 07/02/26 Analyzed: 07/03/26

Surrogate: Pyrene-d10	0.395	ng/ul	0.49950	ND	79.1	70-130
Surrogate: Benzo(a)pyrene-d12	0.516	ng/ul	0.49850	ND	104	70-130

Batch B6E1305 - TO-13 extraction

Blank (B6E1305-BLK1)

Prepared: 05/13/26 Analyzed: 06/04/26

Naphthalene	ND	ng/m ³ Air					U
Acenaphthylene	ND	ng/m ³ Air					U
Acenaphthene	ND	ng/m ³ Air					U
Fluorene	ND	ng/m ³ Air					U
Phenanthrene	ND	ng/m ³ Air					LB, U
Anthracene	ND	ng/m ³ Air					U
Fluoranthene	ND	ng/m ³ Air					LB, U
Pyrene	ND	ng/m ³ Air					LB, U
Benzo(a)anthracene	ND	ng/m ³ Air					U
Chrysene	ND	ng/m ³ Air					U
Benzo(b)fluoranthene	ND	ng/m ³ Air					U
Benzo(k)fluoranthene	ND	ng/m ³ Air					U
Benzo(e)pyrene	ND	ng/m ³ Air					U
Benzo(a)pyrene	ND	ng/m ³ Air					U
Perylene	ND	ng/m ³ Air					U
Dibenz(a,h)anthracene	ND	ng/m ³ Air					U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air					U
Benzo(g,h,i)perylene	ND	ng/m ³ Air					U
Coronene	ND	ng/m ³ Air					U

Surrogate: Fluorene-d10	0.781	ng/ul	0.98000		79.7	60-120
Surrogate: Fluoranthene-d10	2.60	ng/m ³ Air	3.3333		77.9	60-120
Surrogate: Pyrene-d10	0.863	ng/ul	0.99900		86.4	60-120
Surrogate: Benzo(a)pyrene-d12	2.87	ng/m ³ Air	3.3667		85.2	60-120

LCS (B6E1305-BS1)

Prepared: 05/13/26 Analyzed: 06/04/26

Naphthalene	17.2	ng/m ³ Air	25.000	ND	68.8	60-120	
Acenaphthylene	1.19	ng/m ³ Air	1.6667	ND	71.5	60-120	
Acenaphthene	5.95	ng/m ³ Air	8.3333	ND	71.4	60-120	
Fluorene	6.08	ng/m ³ Air	8.3333	ND	73.0	60-120	
Phenanthrene	9.51	ng/m ³ Air	12.600	ND	75.4	60-120	LB
Anthracene	1.10	ng/m ³ Air	1.6633	ND	66.2	60-120	
Fluoranthene	5.99	ng/m ³ Air	8.4333	ND	71.0	60-120	LB
Pyrene	3.11	ng/m ³ Air	4.1333	ND	75.2	60-120	LB
Benzo(a)anthracene	1.33	ng/m ³ Air	1.6667	ND	79.7	60-120	
Chrysene	1.25	ng/m ³ Air	1.6667	ND	75.1	60-120	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6E1305 - TO-13 extraction

LCS (B6E1305-BS1) Continued

Prepared: 05/13/26 Analyzed: 06/04/26

Benzo(b)fluoranthene	1.28	ng/m ³ Air	1.6667	ND	76.8	60-120
Benzo(k)fluoranthene	1.23	ng/m ³ Air	1.6433	ND	75.1	60-120
Benzo(e)pyrene	1.24	ng/m ³ Air	1.6467	ND	75.2	60-120
Benzo(a)pyrene	1.20	ng/m ³ Air	1.6667	ND	71.7	60-120
Perylene	1.23	ng/m ³ Air	1.6633	ND	73.7	60-120
Dibenz(a,h)anthracene	1.25	ng/m ³ Air	1.6667	ND	75.1	60-120
Indeno(1,2,3-c,d)pyrene	1.29	ng/m ³ Air	1.6667	ND	77.6	60-120
Benzo(g,h,i)perylene	1.25	ng/m ³ Air	1.6667	ND	75.1	60-120
Coronene	0.277	ng/m ³ Air	0.33333	ND	83.2	60-120

Surrogate: Fluorene-d10	0.867	ng/ul	0.98000	ND	88.5	60-120
Surrogate: Fluoranthene-d10	2.86	ng/m ³ Air	3.3333	ND	85.9	60-120
Surrogate: Pyrene-d10	0.899	ng/ul	0.99900	ND	90.0	60-120
Surrogate: Benzo(a)pyrene-d12	3.00	ng/m ³ Air	3.3667	ND	89.1	60-120

LCS Dup (B6E1305-BSD1)

Prepared: 05/13/26 Analyzed: 06/04/26

Naphthalene	17.4	ng/m ³ Air	25.000	ND	69.5	60-120	0.945	20	
Acenaphthylene	1.16	ng/m ³ Air	1.6667	ND	69.6	60-120	2.64	20	
Acenaphthene	5.90	ng/m ³ Air	8.3333	ND	70.8	60-120	0.894	20	
Fluorene	6.03	ng/m ³ Air	8.3333	ND	72.3	60-120	0.853	20	
Phenanthrene	9.27	ng/m ³ Air	12.600	ND	73.6	60-120	2.50	20	LB
Anthracene	1.08	ng/m ³ Air	1.6633	ND	65.0	60-120	1.74	20	
Fluoranthene	5.98	ng/m ³ Air	8.4333	ND	70.9	60-120	0.156	20	LB
Pyrene	2.90	ng/m ³ Air	4.1333	ND	70.3	60-120	6.85	20	LB
Benzo(a)anthracene	1.29	ng/m ³ Air	1.6667	ND	77.6	60-120	2.70	20	
Chrysene	1.22	ng/m ³ Air	1.6667	ND	73.2	60-120	2.51	20	
Benzo(b)fluoranthene	1.25	ng/m ³ Air	1.6667	ND	75.1	60-120	2.26	20	
Benzo(k)fluoranthene	1.22	ng/m ³ Air	1.6433	ND	74.0	60-120	1.50	20	
Benzo(e)pyrene	1.20	ng/m ³ Air	1.6467	ND	72.8	60-120	3.34	20	
Benzo(a)pyrene	1.17	ng/m ³ Air	1.6667	ND	70.3	60-120	2.03	20	
Perylene	1.20	ng/m ³ Air	1.6633	ND	71.9	60-120	2.50	20	
Dibenz(a,h)anthracene	1.24	ng/m ³ Air	1.6667	ND	74.1	60-120	1.34	20	
Indeno(1,2,3-c,d)pyrene	1.27	ng/m ³ Air	1.6667	ND	76.0	60-120	2.06	20	
Benzo(g,h,i)perylene	1.22	ng/m ³ Air	1.6667	ND	73.1	60-120	2.70	20	
Coronene	0.266	ng/m ³ Air	0.33333	ND	79.7	60-120	4.30	20	

Surrogate: Fluorene-d10	0.834	ng/ul	0.98000	ND	85.1	60-120
Surrogate: Fluoranthene-d10	2.77	ng/m ³ Air	3.3333	ND	83.0	60-120
Surrogate: Pyrene-d10	0.809	ng/ul	0.99900	ND	80.9	60-120
Surrogate: Benzo(a)pyrene-d12	2.86	ng/m ³ Air	3.3667	ND	84.8	60-120

Calibration Blank (B6E1305-CCB1)

Prepared: 05/13/26 Analyzed: 06/04/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6E1305 - TO-13 extraction

Calibration Blank (B6E1305-CCB1) Contin

Prepared: 05/13/26 Analyzed: 06/04/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	4.00E-4	ng/m ³ Air		ND					LB
Anthracene	0.00	ng/m ³ Air		ND					U
Fluoranthene	3.00E-4	ng/m ³ Air		ND					LB
Pyrene	4.00E-4	ng/m ³ Air		ND					LB
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND		60-120			U

Batch B6E1403 - TO-13 extraction

Blank (B6E1403-BLK1)

Prepared: 05/14/26 Analyzed: 06/16/26

Naphthalene	ND	ng/m ³ Air							LB, U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							U
Fluoranthene	ND	ng/m ³ Air							LB, U
Pyrene	ND	ng/m ³ Air							LB, U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6E1403 - TO-13 extraction

Blank (B6E1403-BLK1) Continued

Prepared: 05/14/26 Analyzed: 06/16/26

Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U
Surrogate: Fluorene-d10	0.750	ng/ul	0.98000		76.6	60-120			
Surrogate: Fluoranthene-d10	2.68	ng/m ³ Air	3.3333		80.3	60-120			
Surrogate: Pyrene-d10	0.746	ng/ul	0.99900		74.6	60-120			
Surrogate: Benzo(a)pyrene-d12	2.77	ng/m ³ Air	3.3667		82.2	60-120			

LCS Dup (B6E1403-BSD1)

Prepared: 05/14/26 Analyzed: 06/16/26

Naphthalene	16.5	ng/m ³ Air	25.000	ND	65.9	60-120	39.1	20	LB
Acenaphthylene	1.11	ng/m ³ Air	1.6667	ND	66.4	60-120	33.5	20	
Acenaphthene	5.76	ng/m ³ Air	8.3333	ND	69.1	60-120	36.9	20	
Fluorene	5.92	ng/m ³ Air	8.3333	ND	71.1	60-120	37.6	20	
Phenanthrene	9.12	ng/m ³ Air	12.600	ND	72.4	60-120	32.5	20	LB
Anthracene	1.07	ng/m ³ Air	1.6633	ND	64.5	60-120	33.7	20	
Fluoranthene	6.13	ng/m ³ Air	8.4333	ND	72.7	60-120	31.6	20	LB
Pyrene	2.86	ng/m ³ Air	4.1333	ND	69.1	60-120	31.5	20	LB
Benzo(a)anthracene	1.31	ng/m ³ Air	1.6667	ND	78.6	60-120	31.3	20	
Chrysene	1.23	ng/m ³ Air	1.6667	ND	74.0	60-120	30.7	20	
Benzo(b)fluoranthene	1.28	ng/m ³ Air	1.6667	ND	76.7	60-120	29.3	20	
Benzo(k)fluoranthene	1.25	ng/m ³ Air	1.6433	ND	76.2	60-120	30.2	20	
Benzo(e)pyrene	1.23	ng/m ³ Air	1.6467	ND	74.6	60-120	26.1	20	
Benzo(a)pyrene	1.19	ng/m ³ Air	1.6667	ND	71.6	60-120	26.7	20	
Perylene	1.23	ng/m ³ Air	1.6633	ND	73.7	60-120	25.4	20	
Dibenz(a,h)anthracene	1.24	ng/m ³ Air	1.6667	ND	74.6	60-120	26.7	20	
Indeno(1,2,3-c,d)pyrene	1.28	ng/m ³ Air	1.6667	ND	76.8	60-120	25.4	20	
Benzo(g,h,i)perylene	1.24	ng/m ³ Air	1.6667	ND	74.5	60-120	23.7	20	
Coronene	0.278	ng/m ³ Air	0.33333	ND	83.4	60-120	23.3	20	
Surrogate: Fluorene-d10	0.791	ng/ul	0.98000	ND	80.8	60-120			
Surrogate: Fluoranthene-d10	2.82	ng/m ³ Air	3.3333	ND	84.5	60-120			
Surrogate: Pyrene-d10	0.776	ng/ul	0.99900	ND	77.7	60-120			
Surrogate: Benzo(a)pyrene-d12	2.89	ng/m ³ Air	3.3667	ND	85.9	60-120			

Calibration Blank (B6E1403-CCB1)

Prepared: 05/14/26 Analyzed: 06/16/26

Naphthalene	3.00E-4	ng/m ³ Air		ND					LB
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6E1403 - TO-13 extraction

Calibration Blank (B6E1403-CCB1) Contin

Prepared: 05/14/26 Analyzed: 06/16/26

Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	5.00E-4	ng/m ³ Air		ND					LB
Anthracene	0.00	ng/m ³ Air		ND					U
Fluoranthene	5.00E-4	ng/m ³ Air		ND					LB
Pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND		60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND		60-120			U

Batch B6F0203 - TO-13 extraction

Blank (B6F0203-BLK1)

Prepared: 06/02/26 Analyzed: 06/30/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							U
Fluoranthene	ND	ng/m ³ Air							LB, U
Pyrene	ND	ng/m ³ Air							U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F0203 - TO-13 extraction

Blank (B6F0203-BLK1) Continued

Prepared: 06/02/26 Analyzed: 06/30/26

Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U
Surrogate: Fluorene-d10	0.827	ng/ul	0.98000		84.4	60-120			
Surrogate: Fluoranthene-d10	2.92	ng/m ³ Air	3.3333		87.7	60-120			
Surrogate: Pyrene-d10	0.814	ng/ul	0.99900		81.5	60-120			
Surrogate: Benzo(a)pyrene-d12	2.81	ng/m ³ Air	3.3667		83.4	60-120			

LCS (B6F0203-BS1)

Prepared: 06/02/26 Analyzed: 06/30/26

Naphthalene	20.0	ng/m ³ Air	25.000	ND	79.9	60-120			
Acenaphthylene	1.38	ng/m ³ Air	1.6667	ND	83.0	60-120			
Acenaphthene	6.96	ng/m ³ Air	8.3333	ND	83.6	60-120			
Fluorene	7.05	ng/m ³ Air	8.3333	ND	84.6	60-120			
Phenanthrene	10.8	ng/m ³ Air	12.600	ND	85.8	60-120			LB
Anthracene	1.28	ng/m ³ Air	1.6633	ND	77.0	60-120			
Fluoranthene	7.40	ng/m ³ Air	8.4333	ND	87.7	60-120			LB
Pyrene	3.09	ng/m ³ Air	4.1333	ND	74.8	60-120			
Benzo(a)anthracene	1.48	ng/m ³ Air	1.6667	ND	88.8	60-120			
Chrysene	1.39	ng/m ³ Air	1.6667	ND	83.5	60-120			
Benzo(b)fluoranthene	1.43	ng/m ³ Air	1.6667	ND	86.1	60-120			
Benzo(k)fluoranthene	1.40	ng/m ³ Air	1.6433	ND	85.1	60-120			
Benzo(e)pyrene	1.36	ng/m ³ Air	1.6467	ND	82.8	60-120			
Benzo(a)pyrene	1.34	ng/m ³ Air	1.6667	ND	80.2	60-120			
Perylene	1.37	ng/m ³ Air	1.6633	ND	82.1	60-120			
Dibenz(a,h)anthracene	1.42	ng/m ³ Air	1.6667	ND	85.3	60-120			
Indeno(1,2,3-c,d)pyrene	1.45	ng/m ³ Air	1.6667	ND	87.2	60-120			
Benzo(g,h,i)perylene	1.37	ng/m ³ Air	1.6667	ND	82.1	60-120			
Coronene	0.299	ng/m ³ Air	0.33333	ND	89.6	60-120			
Surrogate: Fluorene-d10	0.821	ng/ul	0.98000	ND	83.8	60-120			
Surrogate: Fluoranthene-d10	2.99	ng/m ³ Air	3.3333	ND	89.8	60-120			
Surrogate: Pyrene-d10	0.746	ng/ul	0.99900	ND	74.7	60-120			
Surrogate: Benzo(a)pyrene-d12	2.82	ng/m ³ Air	3.3667	ND	83.7	60-120			

LCS Dup (B6F0203-BS1)

Prepared: 06/02/26 Analyzed: 06/30/26

Naphthalene	19.3	ng/m ³ Air	25.000	ND	77.0	60-120	3.73	20	
Acenaphthylene	1.34	ng/m ³ Air	1.6667	ND	80.3	60-120	3.38	20	
Acenaphthene	6.77	ng/m ³ Air	8.3333	ND	81.2	60-120	2.85	20	
Fluorene	6.93	ng/m ³ Air	8.3333	ND	83.2	60-120	1.70	20	
Phenanthrene	10.6	ng/m ³ Air	12.600	ND	84.2	60-120	1.81	20	LB
Anthracene	1.26	ng/m ³ Air	1.6633	ND	76.0	60-120	1.31	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F0203 - TO-13 extraction

LCS Dup (B6F0203-BSD1) Continued

Prepared: 06/02/26 Analyzed: 06/30/26

Fluoranthene	7.28	ng/m ³ Air	8.4333	ND	86.3	60-120	1.64	20	LB
Pyrene	3.02	ng/m ³ Air	4.1333	ND	73.0	60-120	2.49	20	
Benzo(a)anthracene	1.46	ng/m ³ Air	1.6667	ND	87.5	60-120	1.57	20	
Chrysene	1.37	ng/m ³ Air	1.6667	ND	82.4	60-120	1.37	20	
Benzo(b)fluoranthene	1.41	ng/m ³ Air	1.6667	ND	84.4	60-120	1.95	20	
Benzo(k)fluoranthene	1.38	ng/m ³ Air	1.6433	ND	83.7	60-120	1.66	20	
Benzo(e)pyrene	1.34	ng/m ³ Air	1.6467	ND	81.5	60-120	1.55	20	
Benzo(a)pyrene	1.32	ng/m ³ Air	1.6667	ND	79.5	60-120	0.827	20	
Perylene	1.35	ng/m ³ Air	1.6633	ND	81.0	60-120	1.40	20	
Dibenz(a,h)anthracene	1.41	ng/m ³ Air	1.6667	ND	84.6	60-120	0.894	20	
Indeno(1,2,3-c,d)pyrene	1.44	ng/m ³ Air	1.6667	ND	86.2	60-120	1.08	20	
Benzo(g,h,i)perylene	1.35	ng/m ³ Air	1.6667	ND	81.1	60-120	1.23	20	
Coronene	0.298	ng/m ³ Air	0.33333	ND	89.5	60-120	0.112	20	
Surrogate: Fluorene-d10	0.800	ng/ul	0.98000	ND	81.7	60-120			
Surrogate: Fluoranthene-d10	3.01	ng/m ³ Air	3.3333	ND	90.4	60-120			
Surrogate: Pyrene-d10	0.729	ng/ul	0.99900	ND	73.0	60-120			
Surrogate: Benzo(a)pyrene-d12	2.84	ng/m ³ Air	3.3667	ND	84.3	60-120			

Calibration Blank (B6F0203-CCB1)

Prepared: 06/02/26 Analyzed: 06/30/26

Naphthalene	0.00	ng/m ³ Air		ND					U
Acenaphthylene	0.00	ng/m ³ Air		ND					U
Acenaphthene	0.00	ng/m ³ Air		ND					U
Fluorene	0.00	ng/m ³ Air		ND					U
Phenanthrene	4.00E-4	ng/m ³ Air		ND					LB
Anthracene	0.00	ng/m ³ Air		ND					U
Fluoranthene	3.00E-4	ng/m ³ Air		ND					LB
Pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)anthracene	0.00	ng/m ³ Air		ND					U
Chrysene	0.00	ng/m ³ Air		ND					U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND					U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND					U
Perylene	0.00	ng/m ³ Air		ND					U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND					U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND					U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND					U
Coronene	0.00	ng/m ³ Air		ND					U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND		60-120			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F0203 - TO-13 extraction

Calibration Blank (B6F0203-CCB1) Contin

Prepared: 06/02/26 Analyzed: 06/30/26

Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air	ND			60-120			U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air	ND			60-120			U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air	ND			60-120			U

Batch B6F0404 - TO-13 extraction

Blank (B6F0404-BLK1)

Prepared: 06/04/26 Analyzed: 07/02/26

Naphthalene	ND	ng/m ³ Air							U
Acenaphthylene	ND	ng/m ³ Air							U
Acenaphthene	ND	ng/m ³ Air							U
Fluorene	ND	ng/m ³ Air							U
Phenanthrene	ND	ng/m ³ Air							LB, U
Anthracene	ND	ng/m ³ Air							LB, U
Fluoranthene	ND	ng/m ³ Air							LB, U
Pyrene	ND	ng/m ³ Air							LB, U
Benzo(a)anthracene	ND	ng/m ³ Air							U
Chrysene	ND	ng/m ³ Air							U
Benzo(b)fluoranthene	ND	ng/m ³ Air							U
Benzo(k)fluoranthene	ND	ng/m ³ Air							U
Benzo(e)pyrene	ND	ng/m ³ Air							U
Benzo(a)pyrene	ND	ng/m ³ Air							U
Perylene	ND	ng/m ³ Air							U
Dibenz(a,h)anthracene	ND	ng/m ³ Air							U
Indeno(1,2,3-c,d)pyrene	ND	ng/m ³ Air							U
Benzo(g,h,i)perylene	ND	ng/m ³ Air							U
Coronene	ND	ng/m ³ Air							U

Surrogate: Fluorene-d10	0.892	ng/ul	0.98000		91.1	60-120			
Surrogate: Fluoranthene-d10	2.81	ng/m ³ Air	3.3333		84.4	60-120			
Surrogate: Pyrene-d10	0.669	ng/ul	0.99900		66.9	60-120			
Surrogate: Benzo(a)pyrene-d12	3.04	ng/m ³ Air	3.3667		90.3	60-120			

LCS (B6F0404-BS1)

Prepared: 06/04/26 Analyzed: 07/02/26

Naphthalene	18.3	ng/m ³ Air	25.000	ND	73.2	60-120			
Acenaphthylene	1.14	ng/m ³ Air	1.6667	ND	68.6	60-120			
Acenaphthene	6.37	ng/m ³ Air	8.3333	ND	76.4	60-120			
Fluorene	6.75	ng/m ³ Air	8.3333	ND	81.0	60-120			
Phenanthrene	9.82	ng/m ³ Air	12.600	ND	77.9	60-120			LB
Anthracene	1.17	ng/m ³ Air	1.6633	ND	70.4	60-120			LB
Fluoranthene	6.50	ng/m ³ Air	8.4333	ND	77.0	60-120			LB
Pyrene	2.68	ng/m ³ Air	4.1333	ND	64.9	60-120			LB
Benzo(a)anthracene	1.46	ng/m ³ Air	1.6667	ND	87.6	60-120			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Batch B6F0404 - TO-13 extraction

LCS (B6F0404-BS1) Continued

Prepared: 06/04/26 Analyzed: 07/02/26

Chrysene	1.32	ng/m ³ Air	1.6667	ND	79.0	60-120
Benzo(b)fluoranthene	1.41	ng/m ³ Air	1.6667	ND	84.4	60-120
Benzo(k)fluoranthene	1.45	ng/m ³ Air	1.6433	ND	88.1	60-120
Benzo(e)pyrene	1.25	ng/m ³ Air	1.6467	ND	75.8	60-120
Benzo(a)pyrene	1.36	ng/m ³ Air	1.6667	ND	81.8	60-120
Perylene	1.30	ng/m ³ Air	1.6633	ND	78.0	60-120
Dibenz(a,h)anthracene	1.45	ng/m ³ Air	1.6667	ND	87.1	60-120
Indeno(1,2,3-c,d)pyrene	1.51	ng/m ³ Air	1.6667	ND	90.5	60-120
Benzo(g,h,i)perylene	1.36	ng/m ³ Air	1.6667	ND	81.3	60-120
Coronene	0.287	ng/m ³ Air	0.33333	ND	86.0	60-120

Surrogate: Fluorene-d10	0.915	ng/ul	0.98000	ND	93.3	60-120
Surrogate: Fluoranthene-d10	2.71	ng/m ³ Air	3.3333	ND	81.4	60-120
Surrogate: Pyrene-d10	0.682	ng/ul	0.99900	ND	68.3	60-120
Surrogate: Benzo(a)pyrene-d12	2.89	ng/m ³ Air	3.3667	ND	85.9	60-120

Calibration Blank (B6F0404-CCB1)

Prepared: 06/04/26 Analyzed: 07/02/26

Naphthalene	0.00	ng/m ³ Air		ND		U
Acenaphthylene	0.00	ng/m ³ Air		ND		U
Acenaphthene	0.00	ng/m ³ Air		ND		U
Fluorene	0.00	ng/m ³ Air		ND		U
Phenanthrene	0.00190	ng/m ³ Air		ND		LB
Anthracene	0.00	ng/m ³ Air		ND		U
Fluoranthene	0.00190	ng/m ³ Air		ND		LB
Pyrene	0.00270	ng/m ³ Air		ND		LB
Benzo(a)anthracene	0.00	ng/m ³ Air		ND		U
Chrysene	0.00	ng/m ³ Air		ND		U
Benzo(b)fluoranthene	0.00	ng/m ³ Air		ND		U
Benzo(k)fluoranthene	0.00	ng/m ³ Air		ND		U
Benzo(e)pyrene	0.00	ng/m ³ Air		ND		U
Benzo(a)pyrene	0.00	ng/m ³ Air		ND		U
Perylene	0.00	ng/m ³ Air		ND		U
Dibenz(a,h)anthracene	0.00	ng/m ³ Air		ND		U
Indeno(1,2,3-c,d)pyrene	0.00	ng/m ³ Air		ND		U
Benzo(g,h,i)perylene	0.00	ng/m ³ Air		ND		U
Coronene	0.00	ng/m ³ Air		ND		U
Surrogate: Fluorene-d10	0.00	ng/m ³ Air		ND	60-120	U
Surrogate: Fluoranthene-d10	0.00	ng/m ³ Air		ND	60-120	U
Surrogate: Pyrene-d10	0.00	ng/m ³ Air		ND	60-120	U
Surrogate: Benzo(a)pyrene-d12	0.00	ng/m ³ Air		ND	60-120	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606004

Calibration Check (2606004-CCV1)

Prepared & Analyzed: 06/04/26

Naphthalene	7.56	ng/ul	0.7	30.00	
Acenaphthylene	0.530	ng/ul	5.7	30.00	
Acenaphthene	2.51	ng/ul	0.2	30.00	
Fluorene	2.50	ng/ul	-0.3	30.00	
Phenanthrene	3.87	ng/ul	3.0	30.00	
Anthracene	0.494	ng/ul	-1.5	30.00	
Fluoranthene	2.39	ng/ul	-4.3	30.00	LB
Pyrene	1.32	ng/ul	5.6	30.00	LB
Benzo(a)anthracene	0.520	ng/ul	3.7	30.00	
Chrysene	0.516	ng/ul	2.8	30.00	
Benzo(b)fluoranthene	0.514	ng/ul	2.5	30.00	
Benzo(k)fluoranthene	0.515	ng/ul	2.8	30.00	
Benzo(e)pyrene	0.514	ng/ul	2.6	30.00	
Benzo(a)pyrene	0.516	ng/ul	2.7	30.00	
Perylene	0.515	ng/ul	2.7	30.00	
Dibenz(a,h)anthracene	0.533	ng/ul	6.0	30.00	
Indeno(1,2,3-c,d)pyrene	0.535	ng/ul	6.7	30.00	
Benzo(g,h,i)perylene	0.540	ng/ul	7.7	30.00	
Coronene	0.544	ng/ul	8.4	30.00	

Surrogate: Fluorene-d10

0.492

ng/ul

0.5

30.00

Surrogate: Fluoranthene-d10

0.473

ng/ul

-5.4

30.00

Surrogate: Pyrene-d10

0.525

ng/ul

5.1

30.00

Surrogate: Benzo(a)pyrene-d12

0.520

ng/ul

2.9

30.00

Calibration Check (2606004-CCV2)

Prepared & Analyzed: 06/04/26

Naphthalene	7.58	ng/ul	1.0	30.00	
Acenaphthylene	0.501	ng/ul	-0.2	30.00	
Acenaphthene	2.56	ng/ul	2.3	30.00	
Fluorene	2.58	ng/ul	3.1	30.00	
Phenanthrene	3.85	ng/ul	2.6	30.00	
Anthracene	0.514	ng/ul	2.6	30.00	
Fluoranthene	2.49	ng/ul	-0.5	30.00	LB
Pyrene	1.40	ng/ul	11.7	30.00	LB
Benzo(a)anthracene	0.524	ng/ul	4.3	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606004

Calibration Check (2606004-CCV2) Continued

Prepared & Analyzed: 06/04/26

Chrysene	0.515	ng/ul	2.6	30.00
Benzo(b)fluoranthene	0.516	ng/ul	2.9	30.00
Benzo(k)fluoranthene	0.517	ng/ul	3.1	30.00
Benzo(e)pyrene	0.517	ng/ul	3.1	30.00
Benzo(a)pyrene	0.514	ng/ul	2.2	30.00
Perylene	0.514	ng/ul	2.6	30.00
Dibenz(a,h)anthracene	0.498	ng/ul	-1.1	30.00
Indeno(1,2,3-c,d)pyrene	0.500	ng/ul	-0.4	30.00
Benzo(g,h,i)perylene	0.510	ng/ul	1.7	30.00
Coronene	0.502	ng/ul	0.07	30.00

Surrogate: Fluorene-d10	0.497	ng/ul	1.3	30.00
Surrogate: Fluoranthene-d10	0.496	ng/ul	-0.8	30.00
Surrogate: Pyrene-d10	0.554	ng/ul	10.9	30.00
Surrogate: Benzo(a)pyrene-d12	0.518	ng/ul	2.6	30.00

Calibration Check (2606004-CCV3)

Prepared: 06/04/26 Analyzed: 06/05/26

Naphthalene	7.55	ng/ul	0.5	30.00
Acenaphthylene	0.503	ng/ul	0.2	30.00
Acenaphthene	2.56	ng/ul	2.3	30.00
Fluorene	2.58	ng/ul	3.2	30.00
Phenanthrene	3.84	ng/ul	2.3	30.00
Anthracene	0.525	ng/ul	4.7	30.00
Fluoranthene	2.61	ng/ul	4.3	30.00
Pyrene	1.24	ng/ul	-0.9	30.00
Benzo(a)anthracene	0.523	ng/ul	4.2	30.00
Chrysene	0.514	ng/ul	2.4	30.00
Benzo(b)fluoranthene	0.511	ng/ul	2.0	30.00
Benzo(k)fluoranthene	0.519	ng/ul	3.6	30.00
Benzo(e)pyrene	0.507	ng/ul	1.2	30.00
Benzo(a)pyrene	0.515	ng/ul	2.6	30.00
Perylene	0.514	ng/ul	2.4	30.00
Dibenz(a,h)anthracene	0.514	ng/ul	2.2	30.00
Indeno(1,2,3-c,d)pyrene	0.510	ng/ul	1.7	30.00
Benzo(g,h,i)perylene	0.511	ng/ul	1.8	30.00
Coronene	0.493	ng/ul	-1.7	30.00

Surrogate: Fluorene-d10	0.499	ng/ul	1.9	30.00
-------------------------	-------	-------	-----	-------

LB

LB

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606004

Calibration Check (2606004-CCV3) Continued

Prepared: 06/04/26 Analyzed: 06/05/26

Surrogate: Fluoranthene-d10	0.520	ng/ul	4.0	30.00
Surrogate: Pyrene-d10	0.497	ng/ul	-0.5	30.00
Surrogate: Benzo(a)pyrene-d12	0.524	ng/ul	3.7	30.00

Sequence 2606056

Calibration Check (2606056-CCV1)

Prepared & Analyzed: 06/16/26

Naphthalene	6.23	ng/ul	-17.0	30.00
Acenaphthylene	0.418	ng/ul	-16.7	30.00
Acenaphthene	2.07	ng/ul	-17.2	30.00
Fluorene	2.08	ng/ul	-17.1	30.00
Phenanthrene	3.12	ng/ul	-16.8	30.00
Anthracene	0.410	ng/ul	-18.3	30.00
Fluoranthene	2.06	ng/ul	-17.7	30.00
Pyrene	1.01	ng/ul	-19.0	30.00
Benzo(a)anthracene	0.416	ng/ul	-17.2	30.00
Chrysene	0.413	ng/ul	-17.7	30.00
Benzo(b)fluoranthene	0.416	ng/ul	-17.0	30.00
Benzo(k)fluoranthene	0.415	ng/ul	-17.2	30.00
Benzo(e)pyrene	0.415	ng/ul	-17.3	30.00
Benzo(a)pyrene	0.416	ng/ul	-17.1	30.00
Perylene	0.413	ng/ul	-17.6	30.00
Dibenz(a,h)anthracene	0.424	ng/ul	-15.7	30.00
Indeno(1,2,3-c,d)pyrene	0.424	ng/ul	-15.5	30.00
Benzo(g,h,i)perylene	0.424	ng/ul	-15.4	30.00
Coronene	0.441	ng/ul	-12.1	30.00

LB

Surrogate: Fluorene-d10	0.483	ng/ul	-1.5	30.00
Surrogate: Fluoranthene-d10	0.496	ng/ul	-2.7	30.00
Surrogate: Pyrene-d10	0.481	ng/ul	-3.7	30.00
Surrogate: Benzo(a)pyrene-d12	0.500	ng/ul	0.2	30.00

Calibration Check (2606056-CCV2)

Prepared & Analyzed: 06/16/26

Naphthalene	6.25	ng/ul	-16.7	30.00
Acenaphthylene	0.412	ng/ul	-17.9	30.00
Acenaphthene	2.09	ng/ul	-16.4	30.00
Fluorene	2.11	ng/ul	-15.6	30.00
Phenanthrene	3.10	ng/ul	-17.3	30.00
Anthracene	0.418	ng/ul	-16.7	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606056

Calibration Check (2606056-CCV2) Continued

Prepared & Analyzed: 06/16/26

Fluoranthene	2.07	ng/ul	-17.3	30.00	LB
Pyrene	1.01	ng/ul	-19.7	30.00	
Benzo(a)anthracene	0.420	ng/ul	-16.3	30.00	
Chrysene	0.412	ng/ul	-18.0	30.00	
Benzo(b)fluoranthene	0.415	ng/ul	-17.2	30.00	
Benzo(k)fluoranthene	0.417	ng/ul	-16.8	30.00	
Benzo(e)pyrene	0.408	ng/ul	-18.7	30.00	
Benzo(a)pyrene	0.413	ng/ul	-17.7	30.00	
Perylene	0.410	ng/ul	-18.3	30.00	
Dibenz(a,h)anthracene	0.414	ng/ul	-17.6	30.00	
Indeno(1,2,3-c,d)pyrene	0.412	ng/ul	-17.8	30.00	
Benzo(g,h,i)perylene	0.412	ng/ul	-17.8	30.00	
Coronene	0.421	ng/ul	-16.1	30.00	
Surrogate: Fluorene-d10	0.487	ng/ul	-0.6	30.00	
Surrogate: Fluoranthene-d10	0.501	ng/ul	-1.7	30.00	
Surrogate: Pyrene-d10	0.481	ng/ul	-3.7	30.00	
Surrogate: Benzo(a)pyrene-d12	0.501	ng/ul	0.5	30.00	

Calibration Check (2606056-CCV3)

Prepared: 06/16/26 Analyzed: 06/17/26

Naphthalene	6.23	ng/ul	-17.1	30.00	LB
Acenaphthylene	0.412	ng/ul	-18.0	30.00	
Acenaphthene	2.09	ng/ul	-16.6	30.00	
Fluorene	2.10	ng/ul	-16.2	30.00	
Phenanthrene	3.11	ng/ul	-17.2	30.00	
Anthracene	0.436	ng/ul	-13.1	30.00	
Fluoranthene	2.14	ng/ul	-14.3	30.00	
Pyrene	0.969	ng/ul	-22.6	30.00	
Benzo(a)anthracene	0.419	ng/ul	-16.5	30.00	
Chrysene	0.410	ng/ul	-18.3	30.00	
Benzo(b)fluoranthene	0.410	ng/ul	-18.3	30.00	
Benzo(k)fluoranthene	0.416	ng/ul	-17.1	30.00	
Benzo(e)pyrene	0.406	ng/ul	-19.1	30.00	
Benzo(a)pyrene	0.413	ng/ul	-17.7	30.00	
Perylene	0.411	ng/ul	-18.1	30.00	
Dibenz(a,h)anthracene	0.416	ng/ul	-17.2	30.00	
Indeno(1,2,3-c,d)pyrene	0.413	ng/ul	-17.7	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606056

Calibration Check (2606056-CCV3) Continued

Prepared: 06/16/26 Analyzed: 06/17/26

Benzo(g,h,i)perylene	0.409	ng/ul	-18.5	30.00	
Coronene	0.407	ng/ul	-18.9	30.00	
Surrogate: Fluorene-d10	0.487	ng/ul	-0.7	30.00	
Surrogate: Fluoranthene-d10	0.523	ng/ul	2.6	30.00	
Surrogate: Pyrene-d10	0.464	ng/ul	-7.2	30.00	
Surrogate: Benzo(a)pyrene-d12	0.506	ng/ul	1.5	30.00	

Calibration Check (2606056-CCV4)

Prepared: 06/16/26 Analyzed: 06/17/26

Naphthalene	6.24	ng/ul	-16.8	30.00	
Acenaphthylene	0.411	ng/ul	-18.2	30.00	
Acenaphthene	2.09	ng/ul	-16.5	30.00	
Fluorene	2.08	ng/ul	-16.9	30.00	
Phenanthrene	3.09	ng/ul	-17.8	30.00	
Anthracene	0.426	ng/ul	-14.9	30.00	
Fluoranthene	2.12	ng/ul	-15.2	30.00	
Pyrene	0.999	ng/ul	-20.2	30.00	LB
Benzo(a)anthracene	0.415	ng/ul	-17.3	30.00	
Chrysene	0.412	ng/ul	-17.9	30.00	
Benzo(b)fluoranthene	0.407	ng/ul	-18.9	30.00	
Benzo(k)fluoranthene	0.411	ng/ul	-18.0	30.00	
Benzo(e)pyrene	0.408	ng/ul	-18.6	30.00	
Benzo(a)pyrene	0.411	ng/ul	-18.2	30.00	
Perylene	0.409	ng/ul	-18.4	30.00	
Dibenz(a,h)anthracene	0.410	ng/ul	-18.5	30.00	
Indeno(1,2,3-c,d)pyrene	0.409	ng/ul	-18.4	30.00	
Benzo(g,h,i)perylene	0.408	ng/ul	-18.7	30.00	
Coronene	0.402	ng/ul	-20.0	30.00	
Surrogate: Fluorene-d10	0.483	ng/ul	-1.4	30.00	
Surrogate: Fluoranthene-d10	0.513	ng/ul	0.6	30.00	
Surrogate: Pyrene-d10	0.479	ng/ul	-4.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.499	ng/ul	0.2	30.00	

Sequence 2606119

Calibration Check (2606119-CCV1)

Prepared & Analyzed: 06/30/26

Naphthalene	7.33	ng/ul	-2.4	30.00	
Acenaphthylene	0.513	ng/ul	2.2	30.00	
Acenaphthene	2.44	ng/ul	-2.6	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606119

Calibration Check (2606119-CCV1) Continued

Prepared & Analyzed: 06/30/26

Fluorene	2.44	ng/ul	-2.4	30.00	LB
Phenanthrene	3.67	ng/ul	-2.3	30.00	
Anthracene	0.484	ng/ul	-3.4	30.00	
Fluoranthene	2.51	ng/ul	0.3	30.00	
Pyrene	1.12	ng/ul	-10.2	30.00	
Benzo(a)anthracene	0.485	ng/ul	-3.3	30.00	
Chrysene	0.484	ng/ul	-3.7	30.00	
Benzo(b)fluoranthene	0.483	ng/ul	-3.6	30.00	
Benzo(k)fluoranthene	0.483	ng/ul	-3.6	30.00	
Benzo(e)pyrene	0.479	ng/ul	-4.4	30.00	
Benzo(a)pyrene	0.486	ng/ul	-3.2	30.00	
Perylene	0.482	ng/ul	-3.7	30.00	
Dibenz(a,h)anthracene	0.517	ng/ul	2.9	30.00	
Indeno(1,2,3-c,d)pyrene	0.509	ng/ul	1.5	30.00	
Benzo(g,h,i)perylene	0.497	ng/ul	-0.9	30.00	
Coronene	0.519	ng/ul	3.5	30.00	
Surrogate: Fluorene-d10	0.481	ng/ul	-1.8	30.00	
Surrogate: Fluoranthene-d10	0.509	ng/ul	-0.1	30.00	
Surrogate: Pyrene-d10	0.452	ng/ul	-9.5	30.00	
Surrogate: Benzo(a)pyrene-d12	0.487	ng/ul	-2.4	30.00	

Calibration Check (2606119-CCV2)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.32	ng/ul	-2.4	30.00	LB
Acenaphthylene	0.492	ng/ul	-2.0	30.00	
Acenaphthene	2.47	ng/ul	-1.4	30.00	
Fluorene	2.49	ng/ul	-0.6	30.00	
Phenanthrene	3.64	ng/ul	-3.1	30.00	
Anthracene	0.501	ng/ul	0.03	30.00	
Fluoranthene	2.58	ng/ul	3.3	30.00	
Pyrene	1.05	ng/ul	-15.9	30.00	
Benzo(a)anthracene	0.494	ng/ul	-1.6	30.00	
Chrysene	0.484	ng/ul	-3.6	30.00	
Benzo(b)fluoranthene	0.485	ng/ul	-3.3	30.00	
Benzo(k)fluoranthene	0.487	ng/ul	-2.8	30.00	
Benzo(e)pyrene	0.471	ng/ul	-6.1	30.00	
Benzo(a)pyrene	0.489	ng/ul	-2.6	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606119

Calibration Check (2606119-CCV2) Continued

Prepared: 06/30/26 Analyzed: 07/01/26

Perylene	0.483	ng/ul	-3.7	30.00
Dibenz(a,h)anthracene	0.498	ng/ul	-1.0	30.00
Indeno(1,2,3-c,d)pyrene	0.490	ng/ul	-2.4	30.00
Benzo(g,h,i)perylene	0.478	ng/ul	-4.7	30.00
Coronene	0.482	ng/ul	-3.9	30.00

Surrogate: Fluorene-d10	0.484	ng/ul	-1.1	30.00
Surrogate: Fluoranthene-d10	0.525	ng/ul	2.9	30.00
Surrogate: Pyrene-d10	0.420	ng/ul	-15.9	30.00
Surrogate: Benzo(a)pyrene-d12	0.489	ng/ul	-1.9	30.00

Calibration Check (2606119-CCV3)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.34	ng/ul	-2.3	30.00
Acenaphthylene	0.491	ng/ul	-2.2	30.00
Acenaphthene	2.47	ng/ul	-1.4	30.00
Fluorene	2.47	ng/ul	-1.2	30.00
Phenanthrene	3.66	ng/ul	-2.5	30.00
Anthracene	0.507	ng/ul	1.2	30.00
Fluoranthene	2.56	ng/ul	2.5	30.00
Pyrene	1.06	ng/ul	-15.1	30.00
Benzo(a)anthracene	0.488	ng/ul	-2.8	30.00
Chrysene	0.480	ng/ul	-4.5	30.00
Benzo(b)fluoranthene	0.482	ng/ul	-3.9	30.00
Benzo(k)fluoranthene	0.486	ng/ul	-3.1	30.00
Benzo(e)pyrene	0.472	ng/ul	-5.7	30.00
Benzo(a)pyrene	0.487	ng/ul	-3.1	30.00
Perylene	0.481	ng/ul	-4.1	30.00
Dibenz(a,h)anthracene	0.493	ng/ul	-1.9	30.00
Indeno(1,2,3-c,d)pyrene	0.487	ng/ul	-2.9	30.00
Benzo(g,h,i)perylene	0.480	ng/ul	-4.3	30.00
Coronene	0.476	ng/ul	-5.2	30.00

Surrogate: Fluorene-d10	0.483	ng/ul	-1.5	30.00
Surrogate: Fluoranthene-d10	0.520	ng/ul	2.0	30.00
Surrogate: Pyrene-d10	0.425	ng/ul	-15.0	30.00
Surrogate: Benzo(a)pyrene-d12	0.488	ng/ul	-2.2	30.00

Calibration Check (2606119-CCV4)

Prepared: 06/30/26 Analyzed: 07/01/26

Naphthalene	7.31	ng/ul	-2.6	30.00
-------------	------	-------	------	-------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2606119

Calibration Check (2606119-CCV4) Continued

Prepared: 06/30/26 Analyzed: 07/01/26

Acenaphthylene	0.490	ng/ul	-2.4	30.00	
Acenaphthene	2.46	ng/ul	-1.7	30.00	
Fluorene	2.46	ng/ul	-1.5	30.00	
Phenanthrene	3.64	ng/ul	-3.1	30.00	
Anthracene	0.500	ng/ul	-0.3	30.00	
Fluoranthene	2.54	ng/ul	1.5	30.00	LB
Pyrene	1.07	ng/ul	-14.8	30.00	
Benzo(a)anthracene	0.489	ng/ul	-2.6	30.00	
Chrysene	0.482	ng/ul	-3.9	30.00	
Benzo(b)fluoranthene	0.481	ng/ul	-4.1	30.00	
Benzo(k)fluoranthene	0.485	ng/ul	-3.2	30.00	
Benzo(e)pyrene	0.474	ng/ul	-5.4	30.00	
Benzo(a)pyrene	0.484	ng/ul	-3.7	30.00	
Perylene	0.482	ng/ul	-3.7	30.00	
Dibenz(a,h)anthracene	0.495	ng/ul	-1.6	30.00	
Indeno(1,2,3-c,d)pyrene	0.490	ng/ul	-2.3	30.00	
Benzo(g,h,i)perylene	0.481	ng/ul	-4.1	30.00	
Coronene	0.476	ng/ul	-5.1	30.00	
Surrogate: Fluorene-d10	0.481	ng/ul	-1.8	30.00	
Surrogate: Fluoranthene-d10	0.518	ng/ul	1.6	30.00	
Surrogate: Pyrene-d10	0.429	ng/ul	-14.1	30.00	
Surrogate: Benzo(a)pyrene-d12	0.488	ng/ul	-2.1	30.00	

Sequence 2607007

Calibration Check (2607007-CCV1)

Prepared & Analyzed: 07/02/26

Naphthalene	7.31	ng/ul	-2.6	30.00	
Acenaphthylene	0.496	ng/ul	-1.3	30.00	
Acenaphthene	2.45	ng/ul	-2.3	30.00	
Fluorene	2.60	ng/ul	3.7	30.00	
Phenanthrene	3.66	ng/ul	-2.5	30.00	LB
Anthracene	0.495	ng/ul	-1.3	30.00	LB
Fluoranthene	2.49	ng/ul	-0.6	30.00	LB
Pyrene	1.01	ng/ul	-19.4	30.00	LB
Benzo(a)anthracene	0.522	ng/ul	3.9	30.00	
Chrysene	0.488	ng/ul	-2.9	30.00	
Benzo(b)fluoranthene	0.506	ng/ul	1.0	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607007

Calibration Check (2607007-CCV1) Continued

Prepared & Analyzed: 07/02/26

Benzo(k)fluoranthene	0.537	ng/ul	7.2	30.00
Benzo(e)pyrene	0.465	ng/ul	-7.2	30.00
Benzo(a)pyrene	0.523	ng/ul	4.2	30.00
Perylene	0.485	ng/ul	-3.3	30.00
Dibenz(a,h)anthracene	0.550	ng/ul	9.4	30.00
Indeno(1,2,3-c,d)pyrene	0.545	ng/ul	8.6	30.00
Benzo(g,h,i)perylene	0.506	ng/ul	1.0	30.00
Coronene	0.520	ng/ul	3.6	30.00

Surrogate: Fluorene-d10	0.558	ng/ul	13.8	30.00
Surrogate: Fluoranthene-d10	0.516	ng/ul	1.1	30.00
Surrogate: Pyrene-d10	0.408	ng/ul	-18.3	30.00
Surrogate: Benzo(a)pyrene-d12	0.517	ng/ul	3.7	30.00

Calibration Check (2607007-CCV2)

Prepared & Analyzed: 07/02/26

Naphthalene	7.37	ng/ul	-1.8	30.00	
Acenaphthylene	0.496	ng/ul	-1.2	30.00	
Acenaphthene	2.47	ng/ul	-1.3	30.00	
Fluorene	2.58	ng/ul	3.3	30.00	
Phenanthrene	3.63	ng/ul	-3.3	30.00	LB
Anthracene	0.495	ng/ul	-1.2	30.00	LB
Fluoranthene	2.45	ng/ul	-1.9	30.00	LB
Pyrene	0.988	ng/ul	-21.0	30.00	LB
Benzo(a)anthracene	0.516	ng/ul	2.9	30.00	
Chrysene	0.483	ng/ul	-3.8	30.00	
Benzo(b)fluoranthene	0.498	ng/ul	-0.6	30.00	
Benzo(k)fluoranthene	0.530	ng/ul	5.7	30.00	
Benzo(e)pyrene	0.457	ng/ul	-8.7	30.00	
Benzo(a)pyrene	0.524	ng/ul	4.3	30.00	
Perylene	0.482	ng/ul	-3.9	30.00	
Dibenz(a,h)anthracene	0.554	ng/ul	10.2	30.00	
Indeno(1,2,3-c,d)pyrene	0.551	ng/ul	9.7	30.00	
Benzo(g,h,i)perylene	0.508	ng/ul	1.4	30.00	
Coronene	0.535	ng/ul	6.6	30.00	

Surrogate: Fluorene-d10	0.563	ng/ul	14.8	30.00
Surrogate: Fluoranthene-d10	0.516	ng/ul	1.1	30.00
Surrogate: Pyrene-d10	0.401	ng/ul	-19.6	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-13 - Quality Control

Sequence 2607007

Calibration Check (2607007-CCV2) Continued

Prepared & Analyzed: 07/02/26

Surrogate: Benzo(a)pyrene-d12	0.517	ng/ul	3.8	30.00
-------------------------------	-------	-------	-----	-------

Calibration Check (2607007-CCV3)

Prepared: 07/02/26 Analyzed: 07/03/26

Naphthalene	7.37	ng/ul	-1.8	30.00	
Acenaphthylene	0.492	ng/ul	-2.0	30.00	
Acenaphthene	2.44	ng/ul	-2.6	30.00	
Fluorene	2.58	ng/ul	3.1	30.00	
Phenanthrene	3.61	ng/ul	-3.7	30.00	LB
Anthracene	0.488	ng/ul	-2.7	30.00	LB
Fluoranthene	2.45	ng/ul	-2.1	30.00	LB
Pyrene	0.987	ng/ul	-21.2	30.00	LB
Benzo(a)anthracene	0.516	ng/ul	2.8	30.00	
Chrysene	0.485	ng/ul	-3.3	30.00	
Benzo(b)fluoranthene	0.506	ng/ul	0.9	30.00	
Benzo(k)fluoranthene	0.542	ng/ul	8.0	30.00	
Benzo(e)pyrene	0.461	ng/ul	-7.9	30.00	
Benzo(a)pyrene	0.520	ng/ul	3.5	30.00	
Perylene	0.487	ng/ul	-2.8	30.00	
Dibenz(a,h)anthracene	0.547	ng/ul	8.7	30.00	
Indeno(1,2,3-c,d)pyrene	0.542	ng/ul	8.1	30.00	
Benzo(g,h,i)perylene	0.506	ng/ul	0.9	30.00	
Coronene	0.528	ng/ul	5.3	30.00	

Surrogate: Fluorene-d10	0.555	ng/ul	13.3	30.00
-------------------------	-------	-------	------	-------

Surrogate: Fluoranthene-d10	0.508	ng/ul	-0.3	30.00
-----------------------------	-------	-------	------	-------

Surrogate: Pyrene-d10	0.395	ng/ul	-20.9	30.00
-----------------------	-------	-------	-------	-------

Surrogate: Benzo(a)pyrene-d12	0.516	ng/ul	3.5	30.00
-------------------------------	-------	-------	-----	-------



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/16/26 09:24

SUBMITTED: 05/07/26 to 06/03/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
S-GC	Surrogate recovery outside of control limits. The data was accepted based on valid recovery of the remaining surrogate.
QB-03	Analyte exceeds sampling media lot blank criteria.
O-05	This sample was extracted outside of the EPA recommended holding time.
LL	Analyte identified; Reported value may be biased low.
LB	Lab blank value above acceptable limit.
FX	Filter Integrity Issue
FB-01	Analyte exceeds Field Blank criteria.
D-F	Duplicate exceeds DQO criteria.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Matrix Spikes were not performed as required by Compendium Method TO-13A. Matrix spikes are not required by ASTM D2609. This deviation from the method has been approved by the US EPA under the contract for UATMP, NATTS, CSATAM, PAMS, and NMOC Support (US EPA Contract No. 68-D-09-048).

Note: This test is accredited under the 2016 TNI Standard; however acenaphthylene is not accredited.