



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 15, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TMOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/07/26 10:54 through 06/16/26 11:28.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TMOK	6050702-01	Air	05/02/26 23:59	05/07/26 10:54
TMOK	6051320-01	Air	05/08/26 23:59	05/13/26 12:55
TMOK D1	6060179-01	Air	05/14/26 23:59	06/01/26 12:02
TMOK D2	6060179-02	Air	05/14/26 23:59	06/01/26 12:02
TMOK	6060179-03	Air	05/20/26 23:59	06/01/26 12:02
TMOK	6060179-04	Air	05/24/26 23:59	06/01/26 12:02
TMOK	6060179-05	Air	05/26/26 23:59	06/01/26 12:02
TMOK	6061659-01	Air	05/30/26 23:59	06/16/26 11:28



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6050702-01

Sampled: 05/02/26 23:59

Pressure @ Receipt: 4.0

Canister #: 18864

Received: 05/07/26 10:54

Comments:

Analyzed: 05/19/26 18:15

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.462	2.29		0.0237
Chloromethane	0.425	0.88		0.0213
Dichlorotetrafluoroethane	0.0168	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0293	0.07		0.0215
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.00660	0.02	U	0.0136
Acetonitrile	199	335.00	E, QB-01, QB-03	0.184
Acrolein	0.136	0.31		0.0977
Trichlorofluoromethane	0.207	1.17		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	5.00E-4	0.00	U	0.0158
Dichloromethane	0.177	0.62		0.0489
Carbon Disulfide	0.00470	0.01	U	0.0473
Trichlorotrifluoroethane	0.0623	0.48		0.0210
trans-1,2-Dichloroethylene	0.00130	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	6.00E-4	0.00	U	0.00905
Chloroform	0.0190	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0178	0.07		0.0116
1,1,1-Trichloroethane	0.00130	0.01	U	0.00952
Benzene	0.195	0.62		0.0158
Carbon Tetrachloride	0.0644	0.41		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00140	0.01	U	0.00889
Trichloroethylene	0.00370	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00920	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6050702-01

Sampled: 05/02/26 23:59

Pressure @ Receipt: 4.0

Canister #: 18864

Received: 05/07/26 10:54

Comments:

Analyzed: 05/19/26 18:15

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.304	1.15	CF, LK	0.0380
Dibromochloromethane	8.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0332	0.16		0.0258
Tetrachloroethylene	0.00650	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0423	0.18		0.0130
m,p-Xylene	0.132	0.57		0.0418
Bromoform	0.00120	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0513	0.22		0.0235
1,3,5-Trimethylbenzene	0.00990	0.05	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	8.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00950	0.06	U	0.0287
o-Dichlorobenzene	0.00110	0.01	U	0.0364
1,2,4-Trichlorobenzene	0.00190	0.01	QB-01, U	0.0792
Hexachloro-1,3-butadiene	6.00E-4	0.01	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6051320-01

Sampled: 05/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43301

Received: 05/13/26 12:55

Comments:

Analyzed: 05/21/26 20:18

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.458	2.27		0.0237
Chloromethane	0.431	0.89		0.0213
Dichlorotetrafluoroethane	0.0171	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0103	0.02	U	0.0215
Bromomethane	0.00900	0.04	U	0.0104
Chloroethane	0.0143	0.04		0.0136
Acetonitrile	175	294.00	E, QB-03	0.184
Acrolein	0.168	0.39	CF, LK	0.0977
Trichlorofluoromethane	0.234	1.32		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.236	0.82		0.0489
Carbon Disulfide	0.00780	0.02	LK, U	0.0473
Trichlorotrifluoroethane	0.0837	0.64	LK	0.0210
trans-1,2-Dichloroethylene	0.00590	0.02	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	4.00E-4	0.00	U	0.00905
Chloroform	0.0193	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0184	0.07		0.0116
1,1,1-Trichloroethane	0.00140	0.01	U	0.00952
Benzene	0.194	0.62		0.0158
Carbon Tetrachloride	0.0681	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00110	0.01	LK, U	0.00889
Trichloroethylene	0.00420	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00990	0.04	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6051320-01

Sampled: 05/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43301

Received: 05/13/26 12:55

Comments:

Analyzed: 05/21/26 20:18

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.510	1.93		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0745	0.35		0.0258
Tetrachloroethylene	0.00780	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0557	0.24		0.0130
m,p-Xylene	0.179	0.78		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0661	0.29		0.0235
1,3,5-Trimethylbenzene	0.0129	0.06	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00310	0.02	U	0.0287
o-Dichlorobenzene	5.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	9.00E-4	0.01	QB-01, U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1

Lab ID: 6060179-01

Sampled: 05/14/26 23:59

Pressure @ Receipt: 3.5

Canister #: 912427

Received: 06/01/26 12:02

Comments:

Analyzed: 06/08/26 17:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.505	2.50		0.0237
Chloromethane	0.546	1.13		0.0213
Dichlorotetrafluoroethane	0.0185	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0113	0.03	CF, LL, U	0.0215
Bromomethane	0.00880	0.03	U	0.0104
Chloroethane	0.00610	0.02	U	0.0136
Acetonitrile	209	352.00	D-F, E	0.184
Acrolein	0.112	0.26		0.0977
Trichlorofluoromethane	0.230	1.29	LK	0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.253	0.88	D-F	0.0489
Carbon Disulfide	0.00980	0.03	U	0.0473
Trichlorotrifluoroethane	0.0758	0.58		0.0210
trans-1,2-Dichloroethylene	0.00200	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0209	0.10		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0207	0.08		0.0116
1,1,1-Trichloroethane	0.00140	0.01	U	0.00952
Benzene	0.148	0.47		0.0158
Carbon Tetrachloride	0.0760	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	8.00E-4	0.01	U	0.00889
Trichloroethylene	0.00260	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0147	0.06	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D1

Lab ID: 6060179-01

Sampled: 05/14/26 23:59

Pressure @ Receipt: 3.5

Canister #: 912427

Received: 06/01/26 12:02

Comments:

Analyzed: 06/08/26 17:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.259	0.98		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0300	0.14		0.0258
Tetrachloroethylene	0.00560	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0318	0.14		0.0130
m,p-Xylene	0.0985	0.43		0.0418
Bromoform	9.00E-4	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0383	0.17		0.0235
1,3,5-Trimethylbenzene	0.00600	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00640	0.04	U	0.0287
o-Dichlorobenzene	6.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00210	0.02	QB-01, U	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2

Lab ID: 6060179-02

Sampled: 05/14/26 23:59

Pressure @ Receipt: 4.5

Canister #: 911694

Received: 06/01/26 12:02

Comments:

Analyzed: 06/08/26 19:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.473	2.34		0.0237
Chloromethane	0.517	1.07		0.0213
Dichlorotetrafluoroethane	0.0170	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0120	0.03	CF, LL, U	0.0215
Bromomethane	0.00820	0.03	U	0.0104
Chloroethane	ND	ND	U	0.0136
Acetonitrile	540	908.00	D-F, E	0.184
Acrolein	0.121	0.28		0.0977
Trichlorofluoromethane	0.222	1.25	LK	0.00950
Acrylonitrile	0.0120	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.700	2.44	D-F	0.0489
Carbon Disulfide	0.00440	0.01	U	0.0473
Trichlorotrifluoroethane	0.0732	0.56		0.0210
trans-1,2-Dichloroethylene	0.00190	0.01	U	0.0171
1,1-Dichloroethane	0.00150	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0209	0.10		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0208	0.08		0.0116
1,1,1-Trichloroethane	0.00130	0.01	U	0.00952
Benzene	0.141	0.45		0.0158
Carbon Tetrachloride	0.0711	0.45		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00260	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0181	0.07	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK D2

Lab ID: 6060179-02

Sampled: 05/14/26 23:59

Pressure @ Receipt: 4.5

Canister #: 911694

Received: 06/01/26 12:02

Comments:

Analyzed: 06/08/26 19:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.251	0.95		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0280	0.13		0.0258
Tetrachloroethylene	0.00520	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0323	0.14		0.0130
m,p-Xylene	0.102	0.44		0.0418
Bromoform	8.00E-4	0.01	U	0.0166
Styrene	0.0176	0.08	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0390	0.17		0.0235
1,3,5-Trimethylbenzene	0.00700	0.03	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00630	0.04	U	0.0287
o-Dichlorobenzene	4.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00150	0.01	QB-01, U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-03

Sampled: 05/20/26 23:59

Pressure @ Receipt: 4.5

Canister #: 43376

Received: 06/01/26 12:02

Comments:

Analyzed: 06/09/26 22:21

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.529	2.62		0.0237
Chloromethane	0.560	1.16		0.0213
Dichlorotetrafluoroethane	0.0190	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00460	0.01	CF, LL, U	0.0215
Bromomethane	0.00820	0.03	U	0.0104
Chloroethane	0.00850	0.02	U	0.0136
Acetonitrile	161	271.00	E	0.184
Acrolein	0.118	0.27	CF, LK	0.0977
Trichlorofluoromethane	0.232	1.31		0.00950
Acrylonitrile	ND	ND	CE, U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.198	0.69		0.0489
Carbon Disulfide	0.00440	0.01	U	0.0473
Trichlorotrifluoroethane	0.0742	0.57		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	0.00100	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0162	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0177	0.07		0.0116
1,1,1-Trichloroethane	ND	ND	U	0.00952
Benzene	0.0804	0.26		0.0158
Carbon Tetrachloride	0.0732	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00230	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00730	0.03	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-03

Sampled: 05/20/26 23:59

Pressure @ Receipt: 4.5

Canister #: 43376

Received: 06/01/26 12:02

Comments:

Analyzed: 06/09/26 22:21

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.110	0.42	QB-03	0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0128	0.06	U	0.0258
Tetrachloroethylene	0.00350	0.02	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0154	0.07		0.0130
m,p-Xylene	0.0418	0.18		0.0418
Bromoform	0.00100	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	LK, U	0.0190
o-Xylene	0.0169	0.07	U	0.0235
1,3,5-Trimethylbenzene	0.00190	0.01	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	7.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00180	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00100	0.01	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-04

Sampled: 05/24/26 23:59

Pressure @ Receipt: 4.0

Canister #: 912404

Received: 06/01/26 12:02

Comments: Makeup for 2-1-26

Analyzed: 06/09/26 23:23

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.513	2.54		0.0237
Chloromethane	0.583	1.21		0.0213
Dichlorotetrafluoroethane	0.0183	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0218	0.05	CF, LL	0.0215
Bromomethane	0.00930	0.04	U	0.0104
Chloroethane	0.00680	0.02	U	0.0136
Acetonitrile	127	214.00	E	0.184
Acrolein	0.130	0.30	CF, LK	0.0977
Trichlorofluoromethane	0.226	1.27		0.00950
Acrylonitrile	0.0136	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.785	2.73		0.0489
Carbon Disulfide	0.0100	0.03	U	0.0473
Trichlorotrifluoroethane	0.0748	0.57		0.0210
trans-1,2-Dichloroethylene	0.00170	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0254	0.12		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0159	0.06		0.0116
1,1,1-Trichloroethane	0.00120	0.01	U	0.00952
Benzene	0.226	0.72		0.0158
Carbon Tetrachloride	0.0726	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00310	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0112	0.05	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-04

Sampled: 05/24/26 23:59

Pressure @ Receipt: 4.0

Canister #: 912404

Received: 06/01/26 12:02

Comments: Makeup for 2-1-26

Analyzed: 06/09/26 23:23

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.516	1.95		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0473	0.22		0.0258
Tetrachloroethylene	0.00410	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0761	0.33		0.0130
m,p-Xylene	0.228	0.99		0.0418
Bromoform	8.00E-4	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	LK, U	0.0190
o-Xylene	0.0834	0.36		0.0235
1,3,5-Trimethylbenzene	0.0126	0.06	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.0111	0.07	U	0.0287
o-Dichlorobenzene	ND	ND	U	0.0364
1,2,4-Trichlorobenzene	7.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-05

Sampled: 05/26/26 23:59

Pressure @ Receipt: 3.5

Canister #: 43343

Received: 06/01/26 12:02

Comments:

Analyzed: 06/10/26 00:25

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.506	2.51		0.0237
Chloromethane	0.576	1.19		0.0213
Dichlorotetrafluoroethane	0.0183	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0222	0.05	CF, LL	0.0215
Bromomethane	0.00890	0.03	U	0.0104
Chloroethane	0.0114	0.03	U	0.0136
Acetonitrile	267	449.00	E	0.184
Acrolein	0.135	0.31		0.0977
Trichlorofluoromethane	0.226	1.27		0.00950
Acrylonitrile	0.0116	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	1.44	5.01	CF, LK	0.0489
Carbon Disulfide	0.0113	0.04	U	0.0473
Trichlorotrifluoroethane	0.0730	0.56		0.0210
trans-1,2-Dichloroethylene	0.0118	0.05	U	0.0171
1,1-Dichloroethane	0.00110	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0273	0.13		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0155	0.06		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.228	0.73		0.0158
Carbon Tetrachloride	0.0725	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00140	0.01	U	0.00889
Trichloroethylene	0.00500	0.03	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0261	0.11	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6060179-05

Sampled: 05/26/26 23:59

Pressure @ Receipt: 3.5

Canister #: 43343

Received: 06/01/26 12:02

Comments:

Analyzed: 06/10/26 00:25

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.492	1.86		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0472	0.22		0.0258
Tetrachloroethylene	0.00980	0.07		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0655	0.29		0.0130
m,p-Xylene	0.214	0.93		0.0418
Bromoform	9.00E-4	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	LK, U	0.0190
o-Xylene	0.0804	0.35		0.0235
1,3,5-Trimethylbenzene	0.0161	0.08	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	ND	ND	U	0.0277
p-Dichlorobenzene	0.0101	0.06	U	0.0287
o-Dichlorobenzene	ND	ND	U	0.0364
1,2,4-Trichlorobenzene	6.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6061659-01

Sampled: 05/30/26 23:59

Pressure @ Receipt: 3.5

Canister #: 19651

Received: 06/16/26 11:28

Comments: Make up for 1-14-26

Analyzed: 06/18/26 01:24

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.478	2.37		0.0237
Chloromethane	0.613	1.27		0.0213
Dichlorotetrafluoroethane	0.0174	0.12		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0115	0.03	U	0.0215
Bromomethane	0.00920	0.04	U	0.0104
Chloroethane	0.00860	0.02	U	0.0136
Acetonitrile	352	592.00	E, LJ	0.184
Acrolein	0.221	0.51	CF, LK	0.0977
Trichlorofluoromethane	0.205	1.15		0.00950
Acrylonitrile	0.0150	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	1.79	6.23	QB-03	0.0489
Carbon Disulfide	0.0590	0.18	QB-03	0.0473
Trichlorotrifluoroethane	0.0675	0.52		0.0210
trans-1,2-Dichloroethylene	0.00310	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0186	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0141	0.06		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.126	0.40		0.0158
Carbon Tetrachloride	0.0689	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00250	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0182	0.07	U	0.0578
trans-1,3-Dichloropropene	ND	ND	U	0.00822

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Description: TMOK

Lab ID: 6061659-01

Sampled: 05/30/26 23:59

Pressure @ Receipt: 3.5

Canister #: 19651

Received: 06/16/26 11:28

Comments: Make up for 1-14-26

Analyzed: 06/18/26 01:24

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.394	1.49		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	ND	ND	CE, U	0.0258
Tetrachloroethylene	0.00510	0.03	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0447	0.19		0.0130
m,p-Xylene	0.140	0.61		0.0418
Bromoform	0.00100	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0434	0.19		0.0235
1,3,5-Trimethylbenzene	0.00750	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00260	0.02	U	0.0287
o-Dichlorobenzene	ND	ND	U	0.0364
1,2,4-Trichlorobenzene	7.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1908 - Summa Canister Prep

Blank (B6E1908-BLK1)

Prepared: 05/06/26 Analyzed: 05/19/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1908 - Summa Canister Prep

Blank (B6E1908-BLK1) Continued

Prepared: 05/06/26 Analyzed: 05/19/26

1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E2102 - Summa Canister Prep

Blank (B6E2102-BLK1)

Prepared: 05/06/26 Analyzed: 05/21/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				LK, U
Trichlorotrifluoroethane	ND	ppbv				LK, U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 07/15/26 12:50
SUBMITTED: 05/07/26 to 06/16/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E2102 - Summa Canister Prep

Blank (B6E2102-BLK1) Continued

Prepared: 05/06/26 Analyzed: 05/21/26

Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				LK, U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6F0802 - Summa Canister Prep

Blank (B6F0802-BLK1)

Prepared: 06/05/26 Analyzed: 06/08/26

Dichlorodifluoromethane	ND	ppbv				U
-------------------------	----	------	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0802 - Summa Canister Prep

Blank (B6F0802-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/08/26

Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				LK, U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0802 - Summa Canister Prep

Blank (B6F0802-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/08/26

n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Duplicate (B6F0802-DUP1)

Source: 6060179-01

Prepared: 05/14/26 Analyzed: 06/08/26

Dichlorodifluoromethane	0.528	ppbv	0.51	4.42	25	
Chloromethane	0.572	ppbv	0.55	4.63	25	
Dichlorotetrafluoroethane	0.0193	ppbv	0.02	4.23	25	
Vinyl chloride	ND	ppbv	ND		25	U
1,3-Butadiene	ND	ppbv	ND		25	CF, LL, U
Bromomethane	ND	ppbv	ND		25	U
Chloroethane	ND	ppbv	ND		25	U
Acetonitrile	212	ppbv	209.00	1.31	25	E
Acrolein	0.111	ppbv	0.11	0.898	25	
Trichlorofluoromethane	0.239	ppbv	0.23	3.66	25	LK
Acrylonitrile	ND	ppbv	ND		25	U
1,1-Dichloroethene	ND	ppbv	ND		25	U
Dichloromethane	0.252	ppbv	0.25	0.475	25	
Carbon Disulfide	ND	ppbv	ND		25	U
Trichlorotrifluoroethane	0.0770	ppbv	0.08	1.57	25	
trans-1,2-Dichloroethylene	ND	ppbv	ND		25	U
1,1-Dichloroethane	ND	ppbv	ND		25	U
Methyl tert-Butyl Ether	ND	ppbv	ND		25	U
Chloroprene	ND	ppbv	ND		25	U
cis-1,2-Dichloroethylene	ND	ppbv	ND		25	U
Bromochloromethane	ND	ppbv	ND		25	U
Chloroform	0.0210	ppbv	0.02	0.477	25	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0802 - Summa Canister Prep

Duplicate (B6F0802-DUP1) Continued

Source: 6060179-01

Prepared: 05/14/26 Analyzed: 06/08/26

Ethyl tert-Butyl Ether	ND	ppbv	ND		25	U
1,2-Dichloroethane	0.0207	ppbv	0.02	0.00	25	
1,1,1-Trichloroethane	ND	ppbv	ND		25	U
Benzene	0.148	ppbv	0.15	0.00	25	
Carbon Tetrachloride	0.0761	ppbv	0.08	0.131	25	
tert-Amyl Methyl Ether	ND	ppbv	ND		25	U
1,2-Dichloropropane	ND	ppbv	ND		25	U
Ethyl Acrylate	ND	ppbv	ND		25	U
Bromodichloromethane	ND	ppbv	ND		25	U
Trichloroethylene	ND	ppbv	ND		25	U
Methyl Methacrylate	ND	ppbv	ND		25	U
cis-1,3-Dichloropropene	ND	ppbv	ND		25	U
Methyl Isobutyl Ketone	ND	ppbv	ND		25	U
trans-1,3-Dichloropropene	ND	ppbv	ND		25	U
1,1,2-Trichloroethane	ND	ppbv	ND		25	U
Toluene	0.258	ppbv	0.26	0.271	25	
Dibromochloromethane	ND	ppbv	ND		25	U
1,2-Dibromoethane	ND	ppbv	ND		25	U
n-Octane	0.0277	ppbv	0.03	7.97	25	
Tetrachloroethylene	ND	ppbv	ND		25	U
Chlorobenzene	ND	ppbv	ND		25	U
Ethylbenzene	0.0317	ppbv	0.03	0.315	25	
m,p-Xylene	0.0972	ppbv	0.10	1.33	25	
Bromoform	ND	ppbv	ND		25	U
Styrene	ND	ppbv	ND		25	CE, U
1,1,2,2-Tetrachloroethane	ND	ppbv	ND		25	U
o-Xylene	0.0377	ppbv	0.04	1.58	25	
1,3,5-Trimethylbenzene	ND	ppbv	ND		25	U
1,2,4-Trimethylbenzene	ND	ppbv	ND		25	CE, U
m-Dichlorobenzene	ND	ppbv	ND		25	U
p-Dichlorobenzene	ND	ppbv	ND		25	U
o-Dichlorobenzene	ND	ppbv	ND		25	U
1,2,4-Trichlorobenzene	ND	ppbv	ND		25	QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv	ND		25	U

Duplicate (B6F0802-DUP2)

Source: 6060179-02

Prepared: 05/14/26 Analyzed: 06/08/26

Dichlorodifluoromethane	0.489	ppbv	0.47	3.22	25	
Chloromethane	0.530	ppbv	0.52	2.47	25	
Dichlorotetrafluoroethane	0.0171	ppbv	0.02	0.587	25	
Vinyl chloride	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0802 - Summa Canister Prep

Duplicate (B6F0802-DUP2) Continued

Source: 6060179-02

Prepared: 05/14/26 Analyzed: 06/08/26

1,3-Butadiene	ND	ppbv	ND		25	CF, LL, U
Bromomethane	ND	ppbv	ND		25	U
Chloroethane	ND	ppbv	ND		25	U
Acetonitrile	535	ppbv	540.00	0.968	25	E
Acrolein	0.120	ppbv	0.12	0.914	25	
Trichlorofluoromethane	0.221	ppbv	0.22	0.587	25	LK
Acrylonitrile	0.0118	ppbv	0.01	1.68	25	
1,1-Dichloroethene	ND	ppbv	ND		25	U
Dichloromethane	0.696	ppbv	0.70	0.530	25	
Carbon Disulfide	ND	ppbv	ND		25	U
Trichlorotrifluoroethane	0.0730	ppbv	0.07	0.274	25	
trans-1,2-Dichloroethylene	ND	ppbv	ND		25	U
1,1-Dichloroethane	ND	ppbv	ND		25	U
Methyl tert-Butyl Ether	ND	ppbv	ND		25	U
Chloroprene	ND	ppbv	ND		25	U
cis-1,2-Dichloroethylene	ND	ppbv	ND		25	U
Bromochloromethane	ND	ppbv	ND		25	U
Chloroform	0.0203	ppbv	0.02	2.91	25	
Ethyl tert-Butyl Ether	ND	ppbv	ND		25	U
1,2-Dichloroethane	0.0205	ppbv	0.02	1.45	25	
1,1,1-Trichloroethane	ND	ppbv	ND		25	U
Benzene	0.140	ppbv	0.14	0.568	25	
Carbon Tetrachloride	0.0709	ppbv	0.07	0.282	25	
tert-Amyl Methyl Ether	ND	ppbv	ND		25	U
1,2-Dichloropropane	ND	ppbv	ND		25	U
Ethyl Acrylate	ND	ppbv	ND		25	U
Bromodichloromethane	ND	ppbv	ND		25	U
Trichloroethylene	ND	ppbv	ND		25	U
Methyl Methacrylate	ND	ppbv	ND		25	U
cis-1,3-Dichloropropene	ND	ppbv	ND		25	U
Methyl Isobutyl Ketone	ND	ppbv	ND		25	U
trans-1,3-Dichloropropene	ND	ppbv	ND		25	U
1,1,2-Trichloroethane	ND	ppbv	ND		25	U
Toluene	0.251	ppbv	0.25	0.119	25	
Dibromochloromethane	ND	ppbv	ND		25	U
1,2-Dibromoethane	ND	ppbv	ND		25	U
n-Octane	0.0280	ppbv	0.03	0.00	25	
Tetrachloroethylene	ND	ppbv	ND		25	U
Chlorobenzene	ND	ppbv	ND		25	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0802 - Summa Canister Prep

Duplicate (B6F0802-DUP2) Continued Source: 6060179-02 Prepared: 05/14/26 Analyzed: 06/08/26

Ethylbenzene	0.0321	ppbv	0.03	0.621	25	
m,p-Xylene	0.101	ppbv	0.10	0.987	25	
Bromoform	ND	ppbv	ND		25	U
Styrene	ND	ppbv	ND		25	U
1,1,2,2-Tetrachloroethane	ND	ppbv	ND		25	U
o-Xylene	0.0384	ppbv	0.04	1.55	25	
1,3,5-Trimethylbenzene	ND	ppbv	ND		25	U
1,2,4-Trimethylbenzene	ND	ppbv	ND		25	CE, U
m-Dichlorobenzene	ND	ppbv	ND		25	U
p-Dichlorobenzene	ND	ppbv	ND		25	U
o-Dichlorobenzene	ND	ppbv	ND		25	U
1,2,4-Trichlorobenzene	ND	ppbv	ND		25	QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv	ND		25	U

Batch B6F0903 - Summa Canister Prep

Blank (B6F0903-BLK1) Prepared: 06/05/26 Analyzed: 06/09/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0903 - Summa Canister Prep

Blank (B6F0903-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/09/26

1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				LK, U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6F1702 - Summa Canister Prep

Blank (B6F1702-BLK1)

Prepared: 06/05/26 Analyzed: 06/17/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F1702 - Summa Canister Prep

Blank (B6F1702-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/17/26

Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				LJ, U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 0475
REPORTED: 07/15/26 12:50
SUBMITTED: 05/07/26 to 06/16/26
AQS SITE CODE: 40 143 1127
SITE CODE: TMOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F1702 - Summa Canister Prep

Blank (B6F1702-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/17/26

m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV1)

Prepared & Analyzed: 05/19/26

Dichlorodifluoromethane	1.07	ppbv	4.1	30.00	QB-01
Chloromethane	1.06	ppbv	3.7	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	7.1	30.00	
Vinyl chloride	0.998	ppbv	-3.2	30.00	
1,3-Butadiene	0.858	ppbv	-17.0	30.00	
Bromomethane	1.12	ppbv	8.8	30.00	
Chloroethane	1.16	ppbv	11.7	30.00	
Acetonitrile	1.12	ppbv	7.6	30.00	
Acrolein	1.06	ppbv	3.8	30.00	
Trichlorofluoromethane	1.08	ppbv	7.1	30.00	
Acrylonitrile	1.13	ppbv	2.2	30.00	
1,1-Dichloroethene	1.11	ppbv	6.9	30.00	
Dichloromethane	1.14	ppbv	9.7	30.00	
Carbon Disulfide	1.38	ppbv	26.6	30.00	
Trichlorotrifluoroethane	1.21	ppbv	17.9	30.00	
trans-1,2-Dichloroethylene	1.07	ppbv	2.5	30.00	
1,1-Dichloroethane	1.14	ppbv	10.2	30.00	
Methyl tert-Butyl Ether	0.981	ppbv	-5.3	30.00	
Chloroprene	0.927	ppbv	-10.7	30.00	
cis-1,2-Dichloroethylene	0.965	ppbv	-6.8	30.00	
Bromochloromethane	1.00	ppbv	-2.7	30.00	
Chloroform	1.12	ppbv	6.8	30.00	
Ethyl tert-Butyl Ether	0.980	ppbv	-5.5	30.00	
1,2-Dichloroethane	1.07	ppbv	3.5	30.00	
1,1,1-Trichloroethane	1.06	ppbv	2.5	30.00	
Benzene	1.03	ppbv	-0.8	30.00	
Carbon Tetrachloride	1.03	ppbv	0.2	30.00	
tert-Amyl Methyl Ether	0.917	ppbv	-11.3	30.00	
1,2-Dichloropropane	1.23	ppbv	18.9	30.00	
Ethyl Acrylate	1.01	ppbv	-3.6	30.00	
Bromodichloromethane	1.22	ppbv	18.2	30.00	
Trichloroethylene	1.15	ppbv	11.1	30.00	
Methyl Methacrylate	1.03	ppbv	-1.2	30.00	
cis-1,3-Dichloropropene	1.07	ppbv	2.3	30.00	
Methyl Isobutyl Ketone	1.14	ppbv	9.1	30.00	
trans-1,3-Dichloropropene	1.16	ppbv	10.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV1) Continued

Prepared & Analyzed: 05/19/26

1,1,2-Trichloroethane	1.21	ppbv	16.0	30.00	
Toluene	1.08	ppbv	3.4	30.00	
Dibromochloromethane	1.21	ppbv	17.5	30.00	
1,2-Dibromoethane	1.12	ppbv	7.0	30.00	
n-Octane	1.04	ppbv	0.2	30.00	
Tetrachloroethylene	1.14	ppbv	8.9	30.00	
Chlorobenzene	1.11	ppbv	5.5	30.00	
Ethylbenzene	1.00	ppbv	-4.2	30.00	
m,p-Xylene	2.08	ppbv	-0.4	30.00	
Bromoform	1.15	ppbv	9.5	30.00	
Styrene	1.01	ppbv	-4.6	30.00	
1,1,2,2-Tetrachloroethane	1.24	ppbv	17.6	30.00	
o-Xylene	0.967	ppbv	-8.4	30.00	
1,3,5-Trimethylbenzene	0.871	ppbv	-17.2	30.00	
1,2,4-Trimethylbenzene	0.850	ppbv	-19.3	30.00	
m-Dichlorobenzene	1.08	ppbv	3.2	30.00	
p-Dichlorobenzene	1.05	ppbv	0.2	30.00	
o-Dichlorobenzene	1.05	ppbv	0.7	30.00	
1,2,4-Trichlorobenzene	0.958	ppbv	-7.7	30.00	QB-01
Hexachloro-1,3-butadiene	1.12	ppbv	6.6	30.00	

Calibration Check (2605061-CCV2)

Prepared: 05/19/26 Analyzed: 05/20/26

Dichlorodifluoromethane	1.12	ppbv	8.6	30.00	
Chloromethane	1.08	ppbv	5.6	30.00	
Dichlorotetrafluoroethane	1.14	ppbv	10.5	30.00	
Vinyl chloride	0.978	ppbv	-5.2	30.00	
1,3-Butadiene	0.798	ppbv	-22.8	30.00	
Bromomethane	1.17	ppbv	13.8	30.00	
Chloroethane	1.19	ppbv	14.8	30.00	
Acetonitrile	1.13	ppbv	9.0	30.00	QB-01
Acrolein	1.08	ppbv	6.1	30.00	
Trichlorofluoromethane	1.11	ppbv	9.8	30.00	
Acrylonitrile	1.14	ppbv	3.4	30.00	
1,1-Dichloroethene	1.14	ppbv	9.5	30.00	
Dichloromethane	1.15	ppbv	10.8	30.00	
Carbon Disulfide	1.22	ppbv	12.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV2) Continued

Prepared: 05/19/26 Analyzed: 05/20/26

Trichlorotrifluoroethane	1.08	ppbv	4.4	30.00
trans-1,2-Dichloroethylene	1.06	ppbv	2.4	30.00
1,1-Dichloroethane	1.05	ppbv	1.0	30.00
Methyl tert-Butyl Ether	0.963	ppbv	-7.0	30.00
Chloroprene	0.779	ppbv	-24.9	30.00
cis-1,2-Dichloroethylene	0.903	ppbv	-12.9	30.00
Bromochloromethane	0.999	ppbv	-2.9	30.00
Chloroform	1.13	ppbv	7.9	30.00
Ethyl tert-Butyl Ether	0.749	ppbv	-27.7	30.00
1,2-Dichloroethane	1.08	ppbv	4.2	30.00
1,1,1-Trichloroethane	1.06	ppbv	3.2	30.00
Benzene	1.00	ppbv	-3.7	30.00
Carbon Tetrachloride	1.02	ppbv	0.02	30.00
tert-Amyl Methyl Ether	0.864	ppbv	-16.4	30.00
1,2-Dichloropropane	1.25	ppbv	20.1	30.00
Ethyl Acrylate	0.980	ppbv	-6.3	30.00
Bromodichloromethane	1.27	ppbv	22.3	30.00
Trichloroethylene	1.16	ppbv	12.0	30.00
Methyl Methacrylate	0.997	ppbv	-4.3	30.00
cis-1,3-Dichloropropene	1.04	ppbv	-0.8	30.00
Methyl Isobutyl Ketone	1.12	ppbv	7.4	30.00
trans-1,3-Dichloropropene	1.15	ppbv	10.0	30.00
1,1,2-Trichloroethane	1.24	ppbv	18.4	30.00
Toluene	1.05	ppbv	0.7	30.00
Dibromochloromethane	1.23	ppbv	19.4	30.00
1,2-Dibromoethane	1.11	ppbv	6.8	30.00
n-Octane	1.02	ppbv	-2.2	30.00
Tetrachloroethylene	1.16	ppbv	11.4	30.00
Chlorobenzene	1.12	ppbv	6.1	30.00
Ethylbenzene	0.966	ppbv	-7.6	30.00
m,p-Xylene	2.05	ppbv	-2.0	30.00
Bromoform	1.18	ppbv	12.6	30.00
Styrene	0.994	ppbv	-6.0	30.00
1,1,2,2-Tetrachloroethane	1.30	ppbv	22.9	30.00
o-Xylene	0.934	ppbv	-11.5	30.00
1,3,5-Trimethylbenzene	0.829	ppbv	-21.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV2) Continued

Prepared: 05/19/26 Analyzed: 05/20/26

1,2,4-Trimethylbenzene	0.810	ppbv	-23.2	30.00	QB-01
m-Dichlorobenzene	1.09	ppbv	4.6	30.00	
p-Dichlorobenzene	1.05	ppbv	0.7	30.00	
o-Dichlorobenzene	1.06	ppbv	1.7	30.00	
1,2,4-Trichlorobenzene	0.962	ppbv	-7.3	30.00	
Hexachloro-1,3-butadiene	1.13	ppbv	7.6	30.00	

Sequence 2605070

Calibration Check (2605070-CCV1)

Prepared & Analyzed: 05/21/26

Dichlorodifluoromethane	1.14	ppbv	10.8	30.00
Chloromethane	1.12	ppbv	9.4	30.00
Dichlorotetrafluoroethane	1.17	ppbv	12.5	30.00
Vinyl chloride	0.979	ppbv	-5.2	30.00
1,3-Butadiene	0.799	ppbv	-22.7	30.00
Bromomethane	1.17	ppbv	13.4	30.00
Chloroethane	1.23	ppbv	18.4	30.00
Acetonitrile	1.11	ppbv	7.0	30.00
Acrolein	1.14	ppbv	11.5	30.00
Trichlorofluoromethane	1.23	ppbv	21.3	30.00
Acrylonitrile	1.12	ppbv	1.8	30.00
1,1-Dichloroethene	1.23	ppbv	18.7	30.00
Dichloromethane	1.19	ppbv	14.5	30.00
Carbon Disulfide	1.26	ppbv	15.6	30.00
Trichlorotrifluoroethane	1.09	ppbv	5.7	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.6	30.00
1,1-Dichloroethane	1.17	ppbv	13.1	30.00
Methyl tert-Butyl Ether	0.978	ppbv	-5.6	30.00
Chloroprene	0.891	ppbv	-14.2	30.00
cis-1,2-Dichloroethylene	0.945	ppbv	-8.8	30.00
Bromochloromethane	1.01	ppbv	-1.9	30.00
Chloroform	1.20	ppbv	14.1	30.00
Ethyl tert-Butyl Ether	0.943	ppbv	-9.0	30.00
1,2-Dichloroethane	1.14	ppbv	10.0	30.00
1,1,1-Trichloroethane	1.15	ppbv	12.0	30.00
Benzene	1.02	ppbv	-2.2	30.00
Carbon Tetrachloride	1.10	ppbv	7.3	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605070

Calibration Check (2605070-CCV1) Continued

Prepared & Analyzed: 05/21/26

tert-Amyl Methyl Ether	0.882	ppbv	-14.7	30.00	
1,2-Dichloropropane	1.23	ppbv	18.9	30.00	
Ethyl Acrylate	1.00	ppbv	-4.4	30.00	
Bromodichloromethane	1.34	ppbv	29.4	30.00	
Trichloroethylene	1.16	ppbv	12.5	30.00	
Methyl Methacrylate	0.997	ppbv	-4.3	30.00	
cis-1,3-Dichloropropene	1.05	ppbv	0.1	30.00	
Methyl Isobutyl Ketone	1.11	ppbv	7.0	30.00	
trans-1,3-Dichloropropene	1.19	ppbv	13.7	30.00	
1,1,2-Trichloroethane	1.27	ppbv	21.9	30.00	
Toluene	1.06	ppbv	2.2	30.00	
Dibromochloromethane	1.26	ppbv	22.3	30.00	
1,2-Dibromoethane	1.15	ppbv	10.1	30.00	
n-Octane	1.01	ppbv	-2.9	30.00	
Tetrachloroethylene	1.20	ppbv	15.3	30.00	
Chlorobenzene	1.14	ppbv	8.6	30.00	
Ethylbenzene	0.988	ppbv	-5.6	30.00	
m,p-Xylene	2.09	ppbv	-0.06	30.00	
Bromoform	1.25	ppbv	19.8	30.00	
Styrene	0.983	ppbv	-7.1	30.00	
1,1,2,2-Tetrachloroethane	1.31	ppbv	24.1	30.00	
o-Xylene	0.932	ppbv	-11.8	30.00	
1,3,5-Trimethylbenzene	0.816	ppbv	-22.5	30.00	
1,2,4-Trimethylbenzene	0.794	ppbv	-24.6	30.00	
m-Dichlorobenzene	1.09	ppbv	4.7	30.00	
p-Dichlorobenzene	1.05	ppbv	-0.02	30.00	
o-Dichlorobenzene	1.05	ppbv	0.7	30.00	
1,2,4-Trichlorobenzene	0.943	ppbv	-9.2	30.00	QB-01
Hexachloro-1,3-butadiene	1.12	ppbv	6.5	30.00	

Calibration Check (2605070-CCV2)

Prepared: 05/21/26 Analyzed: 05/22/26

Dichlorodifluoromethane	1.16	ppbv	12.7	30.00	
Chloromethane	1.10	ppbv	8.0	30.00	
Dichlorotetrafluoroethane	1.13	ppbv	9.3	30.00	
Vinyl chloride	0.955	ppbv	-7.4	30.00	
1,3-Butadiene	0.791	ppbv	-23.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605070

Calibration Check (2605070-CCV2) Continued

Prepared: 05/21/26 Analyzed: 05/22/26

Bromomethane	1.09	ppbv	6.0	30.00	
Chloroethane	1.18	ppbv	14.3	30.00	
Acetonitrile	1.12	ppbv	7.7	30.00	
Acrolein	1.16	ppbv	13.7	30.00	
Trichlorofluoromethane	1.27	ppbv	25.6	30.00	
Acrylonitrile	1.12	ppbv	1.6	30.00	
1,1-Dichloroethene	1.27	ppbv	22.2	30.00	
Dichloromethane	1.18	ppbv	13.5	30.00	
Carbon Disulfide	1.56	ppbv	43.8	30.00	LK
Trichlorotrifluoroethane	1.40	ppbv	35.7	30.00	LK
trans-1,2-Dichloroethylene	1.04	ppbv	0.05	30.00	
1,1-Dichloroethane	1.17	ppbv	13.4	30.00	
Methyl tert-Butyl Ether	0.985	ppbv	-4.9	30.00	
Chloroprene	0.892	ppbv	-14.0	30.00	
cis-1,2-Dichloroethylene	0.872	ppbv	-15.8	30.00	
Bromochloromethane	0.948	ppbv	-7.7	30.00	
Chloroform	1.18	ppbv	12.4	30.00	
Ethyl tert-Butyl Ether	0.740	ppbv	-28.6	30.00	
1,2-Dichloroethane	1.13	ppbv	9.4	30.00	
1,1,1-Trichloroethane	1.16	ppbv	12.2	30.00	
Benzene	0.961	ppbv	-7.4	30.00	
Carbon Tetrachloride	1.09	ppbv	6.8	30.00	
tert-Amyl Methyl Ether	0.836	ppbv	-19.1	30.00	
1,2-Dichloropropane	1.21	ppbv	17.0	30.00	
Ethyl Acrylate	0.985	ppbv	-5.8	30.00	
Bromodichloromethane	1.37	ppbv	32.2	30.00	LK
Trichloroethylene	1.13	ppbv	8.8	30.00	
Methyl Methacrylate	0.979	ppbv	-6.0	30.00	
cis-1,3-Dichloropropene	1.02	ppbv	-2.7	30.00	
Methyl Isobutyl Ketone	1.11	ppbv	7.0	30.00	
trans-1,3-Dichloropropene	1.16	ppbv	10.7	30.00	
1,1,2-Trichloroethane	1.28	ppbv	21.9	30.00	
Toluene	1.05	ppbv	0.6	30.00	
Dibromochloromethane	1.24	ppbv	20.4	30.00	
1,2-Dibromoethane	1.14	ppbv	9.3	30.00	
n-Octane	0.975	ppbv	-6.1	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605070

Calibration Check (2605070-CCV2) Continued

Prepared: 05/21/26 Analyzed: 05/22/26

Tetrachloroethylene	1.21	ppbv	15.8	30.00	
Chlorobenzene	1.14	ppbv	8.3	30.00	
Ethylbenzene	0.973	ppbv	-7.0	30.00	
m,p-Xylene	2.05	ppbv	-1.6	30.00	
Bromoform	1.28	ppbv	22.7	30.00	
Styrene	0.967	ppbv	-8.6	30.00	
1,1,2,2-Tetrachloroethane	1.31	ppbv	24.5	30.00	
o-Xylene	0.923	ppbv	-12.6	30.00	
1,3,5-Trimethylbenzene	0.803	ppbv	-23.7	30.00	
1,2,4-Trimethylbenzene	0.784	ppbv	-25.6	30.00	
m-Dichlorobenzene	1.07	ppbv	2.7	30.00	
p-Dichlorobenzene	1.03	ppbv	-2.0	30.00	
o-Dichlorobenzene	1.04	ppbv	-0.8	30.00	
1,2,4-Trichlorobenzene	0.986	ppbv	-5.0	30.00	QB-01
Hexachloro-1,3-butadiene	1.16	ppbv	10.2	30.00	

Sequence 2606026

Calibration Check (2606026-CCV1)

Prepared & Analyzed: 06/08/26

Dichlorodifluoromethane	1.17	ppbv	13.9	30.00	
Chloromethane	1.18	ppbv	16.1	30.00	
Dichlorotetrafluoroethane	1.18	ppbv	13.9	30.00	
Vinyl chloride	1.16	ppbv	12.3	30.00	
1,3-Butadiene	1.02	ppbv	-1.2	30.00	
Bromomethane	1.16	ppbv	12.7	30.00	
Chloroethane	1.21	ppbv	17.2	30.00	
Acetonitrile	1.09	ppbv	5.5	30.00	
Acrolein	1.01	ppbv	-1.3	30.00	
Trichlorofluoromethane	1.32	ppbv	31.0	30.00	LK
Acrylonitrile	1.20	ppbv	9.0	30.00	
1,1-Dichloroethene	1.14	ppbv	9.2	30.00	
Dichloromethane	1.19	ppbv	14.6	30.00	
Carbon Disulfide	1.24	ppbv	13.6	30.00	
Trichlorotrifluoroethane	1.24	ppbv	20.4	30.00	
trans-1,2-Dichloroethylene	1.13	ppbv	8.9	30.00	
1,1-Dichloroethane	1.18	ppbv	13.6	30.00	
Methyl tert-Butyl Ether	1.18	ppbv	14.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606026

Calibration Check (2606026-CCV1) Continued

Prepared & Analyzed: 06/08/26

Chloroprene	1.08	ppbv	4.0	30.00
cis-1,2-Dichloroethylene	1.12	ppbv	8.1	30.00
Bromochloromethane	1.17	ppbv	13.9	30.00
Chloroform	1.24	ppbv	18.3	30.00
Ethyl tert-Butyl Ether	1.16	ppbv	12.4	30.00
1,2-Dichloroethane	1.12	ppbv	7.8	30.00
1,1,1-Trichloroethane	1.24	ppbv	20.7	30.00
Benzene	1.15	ppbv	11.0	30.00
Carbon Tetrachloride	1.21	ppbv	17.9	30.00
tert-Amyl Methyl Ether	1.21	ppbv	16.7	30.00
1,2-Dichloropropane	1.15	ppbv	11.2	30.00
Ethyl Acrylate	1.11	ppbv	6.5	30.00
Bromodichloromethane	1.15	ppbv	10.6	30.00
Trichloroethylene	1.06	ppbv	2.6	30.00
Methyl Methacrylate	1.07	ppbv	2.4	30.00
cis-1,3-Dichloropropene	1.14	ppbv	9.5	30.00
Methyl Isobutyl Ketone	1.05	ppbv	1.3	30.00
trans-1,3-Dichloropropene	1.14	ppbv	9.0	30.00
1,1,2-Trichloroethane	1.08	ppbv	3.1	30.00
Toluene	1.07	ppbv	3.3	30.00
Dibromochloromethane	1.25	ppbv	20.9	30.00
1,2-Dibromoethane	1.11	ppbv	6.2	30.00
n-Octane	0.966	ppbv	-6.9	30.00
Tetrachloroethylene	1.08	ppbv	3.5	30.00
Chlorobenzene	1.09	ppbv	3.4	30.00
Ethylbenzene	1.03	ppbv	-1.6	30.00
m,p-Xylene	2.08	ppbv	-0.3	30.00
Bromoform	1.30	ppbv	24.2	30.00
Styrene	1.08	ppbv	2.3	30.00
1,1,2,2-Tetrachloroethane	1.26	ppbv	19.5	30.00
o-Xylene	1.06	ppbv	0.009	30.00
1,3,5-Trimethylbenzene	1.01	ppbv	-4.4	30.00
1,2,4-Trimethylbenzene	0.981	ppbv	-7.0	30.00
m-Dichlorobenzene	1.09	ppbv	4.3	30.00
p-Dichlorobenzene	1.08	ppbv	2.8	30.00
o-Dichlorobenzene	1.09	ppbv	4.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606026

Calibration Check (2606026-CCV1) Continued

Prepared & Analyzed: 06/08/26

1,2,4-Trichlorobenzene	1.14	ppbv	9.9	30.00	QB-01
Hexachloro-1,3-butadiene	1.15	ppbv	9.6	30.00	

Calibration Check (2606026-CCV2)

Prepared: 06/08/26 Analyzed: 06/09/26

Dichlorodifluoromethane	1.22	ppbv	18.5	30.00	
Chloromethane	1.24	ppbv	21.6	30.00	
Dichlorotetrafluoroethane	1.25	ppbv	20.4	30.00	
Vinyl chloride	1.15	ppbv	11.0	30.00	
1,3-Butadiene	0.955	ppbv	-7.6	30.00	
Bromomethane	1.20	ppbv	16.6	30.00	
Chloroethane	1.25	ppbv	21.0	30.00	
Acetonitrile	1.17	ppbv	12.6	30.00	
Acrolein	0.987	ppbv	-3.2	30.00	
Trichlorofluoromethane	1.20	ppbv	18.5	30.00	
Acrylonitrile	1.25	ppbv	13.1	30.00	
1,1-Dichloroethene	1.14	ppbv	9.8	30.00	
Dichloromethane	1.25	ppbv	20.9	30.00	
Carbon Disulfide	1.28	ppbv	17.7	30.00	
Trichlorotrifluoroethane	1.24	ppbv	20.7	30.00	
trans-1,2-Dichloroethylene	1.15	ppbv	11.0	30.00	
1,1-Dichloroethane	1.22	ppbv	17.7	30.00	
Methyl tert-Butyl Ether	1.10	ppbv	5.9	30.00	
Chloroprene	1.02	ppbv	-1.6	30.00	
cis-1,2-Dichloroethylene	1.15	ppbv	11.4	30.00	
Bromochloromethane	1.19	ppbv	15.8	30.00	
Chloroform	1.24	ppbv	18.4	30.00	
Ethyl tert-Butyl Ether	1.10	ppbv	5.9	30.00	
1,2-Dichloroethane	1.16	ppbv	12.2	30.00	
1,1,1-Trichloroethane	1.22	ppbv	18.1	30.00	
Benzene	1.16	ppbv	11.7	30.00	
Carbon Tetrachloride	1.17	ppbv	13.9	30.00	
tert-Amyl Methyl Ether	1.13	ppbv	9.1	30.00	
1,2-Dichloropropane	1.22	ppbv	17.2	30.00	
Ethyl Acrylate	1.10	ppbv	4.9	30.00	
Bromodichloromethane	1.20	ppbv	15.4	30.00	
Trichloroethylene	1.08	ppbv	4.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606026

Calibration Check (2606026-CCV2) Continued

Prepared: 06/08/26 Analyzed: 06/09/26

Methyl Methacrylate	1.06	ppbv	1.8	30.00	
cis-1,3-Dichloropropene	1.16	ppbv	10.9	30.00	
Methyl Isobutyl Ketone	1.10	ppbv	5.7	30.00	
trans-1,3-Dichloropropene	1.12	ppbv	7.1	30.00	
1,1,2-Trichloroethane	1.18	ppbv	13.1	30.00	
Toluene	1.02	ppbv	-1.5	30.00	
Dibromochloromethane	1.22	ppbv	17.7	30.00	
1,2-Dibromoethane	1.10	ppbv	5.4	30.00	
n-Octane	0.935	ppbv	-9.9	30.00	
Tetrachloroethylene	1.02	ppbv	-2.0	30.00	
Chlorobenzene	1.14	ppbv	8.6	30.00	
Ethylbenzene	1.03	ppbv	-1.7	30.00	
m,p-Xylene	2.14	ppbv	2.4	30.00	
Bromoform	1.30	ppbv	24.2	30.00	
Styrene	1.09	ppbv	2.7	30.00	
1,1,2,2-Tetrachloroethane	1.35	ppbv	28.1	30.00	
o-Xylene	1.01	ppbv	-4.4	30.00	
1,3,5-Trimethylbenzene	0.934	ppbv	-11.3	30.00	
1,2,4-Trimethylbenzene	0.915	ppbv	-13.2	30.00	
m-Dichlorobenzene	1.12	ppbv	7.2	30.00	
p-Dichlorobenzene	1.09	ppbv	4.4	30.00	
o-Dichlorobenzene	1.11	ppbv	6.5	30.00	
1,2,4-Trichlorobenzene	1.11	ppbv	7.3	30.00	QB-01
Hexachloro-1,3-butadiene	1.16	ppbv	10.2	30.00	

Sequence 2606030

Calibration Check (2606030-CCV1)

Prepared & Analyzed: 06/09/26

Dichlorodifluoromethane	1.18	ppbv	14.8	30.00	
Chloromethane	1.21	ppbv	18.3	30.00	
Dichlorotetrafluoroethane	1.22	ppbv	17.3	30.00	
Vinyl chloride	1.13	ppbv	9.4	30.00	
1,3-Butadiene	0.962	ppbv	-7.0	30.00	
Bromomethane	1.19	ppbv	15.2	30.00	
Chloroethane	1.26	ppbv	21.2	30.00	
Acetonitrile	1.18	ppbv	13.3	30.00	
Acrolein	1.01	ppbv	-0.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606030

Calibration Check (2606030-CCV1) Continued

Prepared & Analyzed: 06/09/26

Trichlorofluoromethane	1.29	ppbv	27.5	30.00
Acrylonitrile	1.26	ppbv	14.3	30.00
1,1-Dichloroethene	1.16	ppbv	11.1	30.00
Dichloromethane	1.25	ppbv	20.4	30.00
Carbon Disulfide	1.28	ppbv	17.7	30.00
Trichlorotrifluoroethane	1.24	ppbv	20.2	30.00
trans-1,2-Dichloroethylene	1.16	ppbv	11.1	30.00
1,1-Dichloroethane	1.22	ppbv	17.3	30.00
Methyl tert-Butyl Ether	1.13	ppbv	9.2	30.00
Chloroprene	1.06	ppbv	1.9	30.00
cis-1,2-Dichloroethylene	1.16	ppbv	11.7	30.00
Bromochloromethane	1.18	ppbv	15.2	30.00
Chloroform	1.24	ppbv	18.2	30.00
Ethyl tert-Butyl Ether	1.13	ppbv	9.1	30.00
1,2-Dichloroethane	1.16	ppbv	12.3	30.00
1,1,1-Trichloroethane	1.22	ppbv	18.6	30.00
Benzene	1.16	ppbv	12.2	30.00
Carbon Tetrachloride	1.17	ppbv	13.9	30.00
tert-Amyl Methyl Ether	1.15	ppbv	11.3	30.00
1,2-Dichloropropane	1.22	ppbv	17.3	30.00
Ethyl Acrylate	1.12	ppbv	6.7	30.00
Bromodichloromethane	1.19	ppbv	14.7	30.00
Trichloroethylene	1.08	ppbv	4.1	30.00
Methyl Methacrylate	1.08	ppbv	3.4	30.00
cis-1,3-Dichloropropene	1.15	ppbv	10.6	30.00
Methyl Isobutyl Ketone	1.09	ppbv	5.0	30.00
trans-1,3-Dichloropropene	1.15	ppbv	9.4	30.00
1,1,2-Trichloroethane	1.11	ppbv	6.0	30.00
Toluene	1.06	ppbv	1.7	30.00
Dibromochloromethane	1.24	ppbv	19.8	30.00
1,2-Dibromoethane	1.10	ppbv	5.7	30.00
n-Octane	0.973	ppbv	-6.3	30.00
Tetrachloroethylene	1.02	ppbv	-2.7	30.00
Chlorobenzene	1.15	ppbv	9.4	30.00
Ethylbenzene	1.04	ppbv	-0.7	30.00
m,p-Xylene	2.16	ppbv	3.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606030

Calibration Check (2606030-CCV1) Continued

Prepared & Analyzed: 06/09/26

Bromoform	1.33	ppbv	26.8	30.00
Styrene	1.09	ppbv	3.1	30.00
1,1,2,2-Tetrachloroethane	1.35	ppbv	28.4	30.00
o-Xylene	1.02	ppbv	-3.8	30.00
1,3,5-Trimethylbenzene	0.926	ppbv	-11.9	30.00
1,2,4-Trimethylbenzene	0.901	ppbv	-14.5	30.00
m-Dichlorobenzene	1.11	ppbv	6.3	30.00
p-Dichlorobenzene	1.08	ppbv	3.5	30.00
o-Dichlorobenzene	1.10	ppbv	5.0	30.00
1,2,4-Trichlorobenzene	1.10	ppbv	5.8	30.00
Hexachloro-1,3-butadiene	1.15	ppbv	9.5	30.00

Calibration Check (2606030-CCV2)

Prepared: 06/09/26 Analyzed: 06/10/26

Dichlorodifluoromethane	1.26	ppbv	22.6	30.00
Chloromethane	1.32	ppbv	29.5	30.00
Dichlorotetrafluoroethane	1.29	ppbv	24.6	30.00
Vinyl chloride	1.19	ppbv	15.8	30.00
1,3-Butadiene	0.981	ppbv	-5.1	30.00
Bromomethane	1.25	ppbv	20.7	30.00
Chloroethane	1.30	ppbv	25.2	30.00
Acetonitrile	1.27	ppbv	22.6	30.00
Acrolein	0.974	ppbv	-4.5	30.00
Trichlorofluoromethane	1.21	ppbv	20.1	30.00
Acrylonitrile	1.23	ppbv	11.2	30.00
1,1-Dichloroethene	1.17	ppbv	12.8	30.00
Dichloromethane	1.29	ppbv	24.9	30.00
Carbon Disulfide	1.33	ppbv	22.0	30.00
Trichlorotrifluoroethane	1.25	ppbv	21.5	30.00
trans-1,2-Dichloroethylene	1.17	ppbv	12.2	30.00
1,1-Dichloroethane	1.24	ppbv	19.4	30.00
Methyl tert-Butyl Ether	1.12	ppbv	8.1	30.00
Chloroprene	1.05	ppbv	0.7	30.00
cis-1,2-Dichloroethylene	1.18	ppbv	13.6	30.00
Bromochloromethane	1.18	ppbv	15.2	30.00
Chloroform	1.25	ppbv	18.8	30.00
Ethyl tert-Butyl Ether	1.12	ppbv	7.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606030

Calibration Check (2606030-CCV2) Continued

Prepared: 06/09/26 Analyzed: 06/10/26

1,2-Dichloroethane	1.17	ppbv	13.0	30.00	
1,1,1-Trichloroethane	1.21	ppbv	17.7	30.00	
Benzene	1.16	ppbv	12.2	30.00	
Carbon Tetrachloride	1.15	ppbv	12.8	30.00	
tert-Amyl Methyl Ether	1.13	ppbv	9.5	30.00	
1,2-Dichloropropane	1.28	ppbv	23.7	30.00	
Ethyl Acrylate	1.15	ppbv	10.0	30.00	
Bromodichloromethane	1.31	ppbv	26.1	30.00	
Trichloroethylene	1.10	ppbv	6.3	30.00	
Methyl Methacrylate	1.15	ppbv	10.7	30.00	
cis-1,3-Dichloropropene	1.06	ppbv	2.0	30.00	
Methyl Isobutyl Ketone	1.08	ppbv	3.4	30.00	
trans-1,3-Dichloropropene	1.20	ppbv	13.9	30.00	
1,1,2-Trichloroethane	1.26	ppbv	20.1	30.00	
Toluene	1.05	ppbv	0.7	30.00	
Dibromochloromethane	1.26	ppbv	21.7	30.00	
1,2-Dibromoethane	1.15	ppbv	10.0	30.00	
n-Octane	0.987	ppbv	-4.9	30.00	
Tetrachloroethylene	1.04	ppbv	-0.6	30.00	
Chlorobenzene	1.17	ppbv	11.1	30.00	
Ethylbenzene	1.04	ppbv	-0.4	30.00	
m,p-Xylene	2.18	ppbv	4.3	30.00	
Bromoform	1.30	ppbv	24.3	30.00	
Styrene	1.10	ppbv	4.0	30.00	
1,1,2,2-Tetrachloroethane	1.41	ppbv	34.1	30.00	LK
o-Xylene	1.02	ppbv	-3.8	30.00	
1,3,5-Trimethylbenzene	0.924	ppbv	-12.1	30.00	
1,2,4-Trimethylbenzene	0.912	ppbv	-13.4	30.00	
m-Dichlorobenzene	1.12	ppbv	7.9	30.00	
p-Dichlorobenzene	1.09	ppbv	4.5	30.00	
o-Dichlorobenzene	1.11	ppbv	6.5	30.00	
1,2,4-Trichlorobenzene	1.12	ppbv	7.9	30.00	
Hexachloro-1,3-butadiene	1.15	ppbv	9.0	30.00	

Sequence 2606069

Calibration Check (2606069-CCV1)

Prepared & Analyzed: 06/17/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606069

Calibration Check (2606069-CCV1) Continued

Prepared & Analyzed: 06/17/26

Dichlorodifluoromethane	1.08	ppbv	5.5	30.00	LJ
Chloromethane	1.15	ppbv	13.1	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	7.5	30.00	
Vinyl chloride	1.08	ppbv	5.0	30.00	
1,3-Butadiene	1.01	ppbv	-2.1	30.00	
Bromomethane	1.10	ppbv	6.4	30.00	
Chloroethane	1.17	ppbv	13.3	30.00	
Acetonitrile	1.25	ppbv	20.1	30.00	
Acrolein	1.09	ppbv	6.5	30.00	
Trichlorofluoromethane	1.06	ppbv	5.0	30.00	
Acrylonitrile	1.26	ppbv	14.5	30.00	
1,1-Dichloroethene	1.15	ppbv	10.9	30.00	
Dichloromethane	1.20	ppbv	16.0	30.00	
Carbon Disulfide	1.19	ppbv	9.3	30.00	
Trichlorotrifluoroethane	1.12	ppbv	8.5	30.00	
trans-1,2-Dichloroethylene	1.10	ppbv	6.2	30.00	
1,1-Dichloroethane	1.17	ppbv	13.2	30.00	
Methyl tert-Butyl Ether	1.09	ppbv	4.8	30.00	
Chloroprene	1.13	ppbv	8.6	30.00	
cis-1,2-Dichloroethylene	1.16	ppbv	11.9	30.00	
Bromochloromethane	1.10	ppbv	7.0	30.00	
Chloroform	1.16	ppbv	11.0	30.00	
Ethyl tert-Butyl Ether	1.11	ppbv	7.6	30.00	
1,2-Dichloroethane	1.15	ppbv	10.9	30.00	
1,1,1-Trichloroethane	1.12	ppbv	8.3	30.00	
Benzene	1.14	ppbv	10.2	30.00	
Carbon Tetrachloride	1.09	ppbv	6.6	30.00	
tert-Amyl Methyl Ether	1.09	ppbv	5.1	30.00	
1,2-Dichloropropane	1.20	ppbv	15.1	30.00	
Ethyl Acrylate	1.18	ppbv	12.4	30.00	
Bromodichloromethane	1.15	ppbv	11.4	30.00	
Trichloroethylene	1.11	ppbv	7.2	30.00	
Methyl Methacrylate	1.17	ppbv	12.1	30.00	
cis-1,3-Dichloropropene	1.18	ppbv	12.6	30.00	
Methyl Isobutyl Ketone	1.14	ppbv	9.5	30.00	
trans-1,3-Dichloropropene	1.12	ppbv	6.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606069

Calibration Check (2606069-CCV1) Continued

Prepared & Analyzed: 06/17/26

1,1,2-Trichloroethane	1.14	ppbv	8.8	30.00
Toluene	1.09	ppbv	5.2	30.00
Dibromochloromethane	1.09	ppbv	5.0	30.00
1,2-Dibromoethane	1.07	ppbv	2.6	30.00
n-Octane	1.11	ppbv	6.5	30.00
Tetrachloroethylene	1.05	ppbv	0.7	30.00
Chlorobenzene	1.16	ppbv	10.7	30.00
Ethylbenzene	1.09	ppbv	3.8	30.00
m,p-Xylene	2.35	ppbv	12.7	30.00
Bromoform	1.17	ppbv	11.7	30.00
Styrene	1.18	ppbv	12.0	30.00
1,1,2,2-Tetrachloroethane	1.28	ppbv	21.8	30.00
o-Xylene	1.12	ppbv	6.2	30.00
1,3,5-Trimethylbenzene	1.04	ppbv	-1.6	30.00
1,2,4-Trimethylbenzene	1.04	ppbv	-1.4	30.00
m-Dichlorobenzene	1.13	ppbv	8.1	30.00
p-Dichlorobenzene	1.11	ppbv	6.0	30.00
o-Dichlorobenzene	1.12	ppbv	6.7	30.00
1,2,4-Trichlorobenzene	1.01	ppbv	-2.9	30.00
Hexachloro-1,3-butadiene	1.11	ppbv	4.8	30.00

Calibration Check (2606069-CCV2)

Prepared: 06/17/26 Analyzed: 06/18/26

Dichlorodifluoromethane	1.13	ppbv	10.0	30.00
Chloromethane	1.22	ppbv	19.7	30.00
Dichlorotetrafluoroethane	1.16	ppbv	12.1	30.00
Vinyl chloride	1.11	ppbv	7.4	30.00
1,3-Butadiene	0.995	ppbv	-3.7	30.00
Bromomethane	1.12	ppbv	8.4	30.00
Chloroethane	1.21	ppbv	16.4	30.00
Acetonitrile	1.28	ppbv	23.7	30.00
Acrolein	1.09	ppbv	6.8	30.00
Trichlorofluoromethane	1.08	ppbv	6.6	30.00
Acrylonitrile	1.27	ppbv	14.8	30.00
1,1-Dichloroethene	1.15	ppbv	10.6	30.00
Dichloromethane	1.21	ppbv	16.9	30.00
Carbon Disulfide	1.20	ppbv	10.6	30.00

LJ

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606069

Calibration Check (2606069-CCV2) Continued

Prepared: 06/17/26 Analyzed: 06/18/26

Trichlorotrifluoroethane	1.10	ppbv	6.9	30.00
trans-1,2-Dichloroethylene	1.08	ppbv	4.2	30.00
1,1-Dichloroethane	1.18	ppbv	13.9	30.00
Methyl tert-Butyl Ether	1.06	ppbv	2.3	30.00
Chloroprene	1.09	ppbv	5.4	30.00
cis-1,2-Dichloroethylene	1.16	ppbv	11.6	30.00
Bromochloromethane	1.08	ppbv	5.2	30.00
Chloroform	1.16	ppbv	10.4	30.00
Ethyl tert-Butyl Ether	1.09	ppbv	5.7	30.00
1,2-Dichloroethane	1.15	ppbv	11.3	30.00
1,1,1-Trichloroethane	1.10	ppbv	7.1	30.00
Benzene	1.14	ppbv	9.4	30.00
Carbon Tetrachloride	1.07	ppbv	4.5	30.00
tert-Amyl Methyl Ether	1.06	ppbv	2.8	30.00
1,2-Dichloropropane	1.27	ppbv	22.2	30.00
Ethyl Acrylate	1.18	ppbv	12.6	30.00
Bromodichloromethane	1.21	ppbv	17.2	30.00
Trichloroethylene	1.09	ppbv	5.0	30.00
Methyl Methacrylate	1.13	ppbv	8.1	30.00
cis-1,3-Dichloropropene	0.964	ppbv	-7.7	30.00
Methyl Isobutyl Ketone	1.05	ppbv	1.1	30.00
trans-1,3-Dichloropropene	1.13	ppbv	7.7	30.00
1,1,2-Trichloroethane	1.18	ppbv	13.0	30.00
Toluene	1.00	ppbv	-3.6	30.00
Dibromochloromethane	1.08	ppbv	4.7	30.00
1,2-Dibromoethane	1.10	ppbv	5.2	30.00
n-Octane	1.10	ppbv	5.8	30.00
Tetrachloroethylene	1.02	ppbv	-2.4	30.00
Chlorobenzene	1.17	ppbv	11.0	30.00
Ethylbenzene	1.12	ppbv	6.7	30.00
m,p-Xylene	2.40	ppbv	15.2	30.00
Bromoform	1.15	ppbv	9.8	30.00
Styrene	1.17	ppbv	11.0	30.00
1,1,2,2-Tetrachloroethane	1.33	ppbv	26.2	30.00
o-Xylene	1.12	ppbv	5.9	30.00
1,3,5-Trimethylbenzene	1.04	ppbv	-1.3	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606069

Calibration Check (2606069-CCV2) Continued

Prepared: 06/17/26 Analyzed: 06/18/26

1,2,4-Trimethylbenzene	1.05	ppbv	-0.4	30.00
m-Dichlorobenzene	1.13	ppbv	8.2	30.00
p-Dichlorobenzene	1.11	ppbv	5.9	30.00
o-Dichlorobenzene	1.12	ppbv	6.6	30.00
1,2,4-Trichlorobenzene	0.988	ppbv	-4.8	30.00
Hexachloro-1,3-butadiene	1.06	ppbv	0.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 0475

REPORTED: 07/15/26 12:50

SUBMITTED: 05/07/26 to 06/16/26

AQS SITE CODE: 40 143 1127

SITE CODE: TMOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
LJ	Identification of analyte is acceptable; reported value is an estimate.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
D-F	Duplicate exceeds DQO criteria.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.