



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 13, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/07/26 11:10 through 05/19/26 11:46.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6041608-01	Air	04/08/26 23:59	04/16/26 12:03
TOOK FB	6041608-02	Air	04/08/26 23:59	04/16/26 12:03
TOOK	6042227-01	Air	04/14/26 23:59	04/22/26 14:00
TOOK	6042227-02	Air	04/20/26 23:59	04/22/26 14:00
TOOK D2	6043006-02	Air	04/28/26 23:59	04/30/26 11:05
TOOK	6051326-01	Air	05/08/26 23:59	05/13/26 13:42
TOOK	6051949-01	Air	04/26/26 23:59	05/19/26 11:46
TOOK FB	6051949-02	Air	04/26/26 23:59	05/19/26 11:46
TOOK D2	6051949-04	Air	05/14/26 23:59	05/19/26 11:46



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041608-01

Sampled: 04/08/26 23:59

Temp Rec'd (°C): 2.7

Sample Volume: 1161.2 L

Received: 04/16/26 12:03

Cartridge Pouch #: 26G1037

Analyzed: 04/20/26 21:20

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.47		0.0545
Acetaldehyde	2.24		0.0226
Acetone	4.98		0.0835
Propionaldehyde	0.697	QB-01	0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.263		0.00965
2-Butanone	0.809		0.0363
Benzaldehyde	0.0981		0.0373
Valeraldehyde	0.0753		0.0104
Hexaldehyde	0.0923		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK FB	Lab ID: 6041608-02	Sampled: 04/08/26 23:59
Temp Rec'd (°C): 2.7	Sample Volume: 1161.2 L	Received: 04/16/26 12:03
	Cartridge Pouch #: 26M0031	Analyzed: 04/20/26 21:36
Comments: FB Makeup from 3-27-26		

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0358	U	0.0545
Acetaldehyde	0.0277		0.0226
Acetone	0.0684	U	0.0835
Propionaldehyde	0.0717	QB-01, U	0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	ND	U	0.00965
2-Butanone	ND	U	0.0363
Benzaldehyde	0.0375		0.0373
Valeraldehyde	ND	U	0.0104
Hexaldehyde	ND	U	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6042227-01

Sampled: 04/14/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1160.1 L

Received: 04/22/26 14:00

Cartridge Pouch #: 26G1038

Analyzed: 04/23/26 19:58

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.99		0.0546
Acetaldehyde	1.24		0.0226
Acetone	0.791		0.0836
Propionaldehyde	0.381		0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.241		0.00965
2-Butanone	0.170		0.0364
Benzaldehyde	0.0872		0.0373
Valeraldehyde	0.106		0.0104
Hexaldehyde	0.0828		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6042227-02

Sampled: 04/20/26 23:59

Temp Rec'd (°C): -0.5

Sample Volume: 1161.7 L

Received: 04/22/26 14:00

Cartridge Pouch #: 26G1039

Analyzed: 05/04/26 14:54

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.45		0.0545
Acetaldehyde	1.76		0.0226
Acetone	4.38		0.0835
Propionaldehyde	0.695	QB-01	0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.227		0.00964
2-Butanone	0.661		0.0363
Benzaldehyde	0.0712		0.0373
Valeraldehyde	0.0636		0.0104
Hexaldehyde	0.0720		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK D2	Lab ID: 6043006-02	Sampled: 04/28/26 23:59
Temp Rec'd (°C): 8.6	Sample Volume: 1159.9 L	Received: 04/30/26 11:05
	Cartridge Pouch #: 26M0035	Analyzed: 05/07/26 18:52

Comments: Makeup from 2/13/26

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.43		0.0546
Acetaldehyde	1.01		0.0226
Acetone	1.55		0.0836
Propionaldehyde	0.325		0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.136		0.00966
2-Butanone	0.542		0.0364
Benzaldehyde	0.0851		0.0373
Valeraldehyde	0.0570		0.0104
Hexaldehyde	0.0737		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6051326-01

Sampled: 05/08/26 23:59

Temp Rec'd (°C): 4.9

Sample Volume: 1162.1 L

Received: 05/13/26 13:42

Cartridge Pouch #: 26G1043

Analyzed: 05/19/26 20:57

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.94		0.0545
Acetaldehyde	2.17		0.0225
Acetone	2.82		0.0835
Propionaldehyde	0.604		0.178
Crotonaldehyde	ND	U	0.0163
Butyraldehyde	0.281		0.00964
2-Butanone	0.495		0.0363
Benzaldehyde	0.148		0.0373
Valeraldehyde	0.292		0.0104
Hexaldehyde	0.203		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK	Lab ID: 6051949-01	Sampled: 04/26/26 23:59
Temp Rec'd (°C): 4.5	Sample Volume: 1165.3 L	Received: 05/19/26 11:46
	Cartridge Pouch #: 26G1040	Analyzed: 05/20/26 19:33

Comments: Recd past extraction HT

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.66	DNPH, S-Rec	0.0543
Acetaldehyde	1.13	DNPH, S-Rec	0.0225
Acetone	0.592	DNPH, S-Rec	0.0832
Propionaldehyde	0.253	DNPH, S-Rec	0.178
Crotonaldehyde	ND	DNPH, S-Rec, U	0.0163
Butyraldehyde	0.174	DNPH, S-Rec	0.00961
2-Butanone	0.499	DNPH, S-Rec	0.0362
Benzaldehyde	0.109	DNPH, S-Rec	0.0372
Valeraldehyde	0.130	DNPH, S-Rec	0.0104
Hexaldehyde	0.140	DNPH, S-Rec	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK FB	Lab ID: 6051949-02	Sampled: 04/26/26 23:59
Temp Rec'd (°C): 4.5	Sample Volume: 1165.3 L	Received: 05/19/26 11:46
	Cartridge Pouch #: 26G1041	Analyzed: 05/20/26 19:50

Comments: Recd past extraction HT

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0369	S-Rec, U	0.0543
Acetaldehyde	0.0178	S-Rec, U	0.0225
Acetone	0.0359	S-Rec, U	0.0832
Propionaldehyde	0.0534	S-Rec, U	0.178
Crotonaldehyde	ND	S-Rec, U	0.0163
Butyraldehyde	ND	S-Rec, U	0.00961
2-Butanone	ND	S-Rec, U	0.0362
Benzaldehyde	0.0408	S-Rec	0.0372
Valeraldehyde	ND	S-Rec, U	0.0104
Hexaldehyde	ND	S-Rec, U	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK D2	Lab ID: 6051949-04	Sampled: 05/14/26 23:59
Temp Rec'd (°C): 4.5	Sample Volume: 1157.8 L	Received: 05/19/26 11:46
	Cartridge Pouch #: 26G1045	Analyzed: 05/22/26 20:26

Comments: Temperature tolerance

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	4.47		0.0547
Acetaldehyde	2.14		0.0226
Acetone	2.11		0.0838
Propionaldehyde	0.564		0.179
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	0.184		0.00967
2-Butanone	0.258		0.0364
Benzaldehyde	0.0892		0.0374
Valeraldehyde	0.183		0.0105
Hexaldehyde	0.126		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604065 - B6D1701

Calibration Blank (2604065-CCB1)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2604065-CCB2)

Prepared: 04/20/26 Analyzed: 04/21/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604065-CCV1)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.146	ug/ml	0.15000		97.7	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			
Acetone	0.149	ug/ml	0.15000		99.6	84.9-115.1			
Propionaldehyde	0.149	ug/ml	0.15000		99.3	84.9-115.1			
Crotonaldehyde	0.149	ug/ml	0.15000		99.1	84.9-115.1			
Butyraldehyde	0.147	ug/ml	0.15000		98.1	84.9-115.1			
2-Butanone	0.151	ug/ml	0.15000		101	84.9-115.1			
Benzaldehyde	0.156	ug/ml	0.15000		104	84.9-115.1			
Valeraldehyde	0.152	ug/ml	0.15000		101	84.9-115.1			
Hexaldehyde	0.147	ug/ml	0.15000		98.1	84.9-115.1			

Calibration Check (2604065-CCV2)

Prepared & Analyzed: 04/20/26

Formaldehyde	0.146	ug/ml	0.15000		97.1	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		99.7	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604065 - B6D1701

Calibration Check (2604065-CCV2) Contin

Prepared & Analyzed: 04/20/26

Acetone	0.149	ug/ml	0.15000	99.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.2	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1

Calibration Check (2604065-CCV3)

Prepared: 04/20/26 Analyzed: 04/21/26

Formaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Acetaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Acetone	0.150	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.3	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1

Batch 2604076 - B6D2305

Calibration Blank (2604076-CCB1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2604076-CCB2)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604076 - B6D2305

Calibration Blank (2604076-CCB2) Contin

Prepared & Analyzed: 04/23/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604076-CCV1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1

Calibration Check (2604076-CCV2)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.8	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1

Calibration Check (2604076-CCV3)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.4	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604076 - B6D2305

Calibration Check (2604076-CCV3) Contin

Prepared & Analyzed: 04/23/26

Hexaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
-------------	-------	-------	---------	------	------------

Batch 2605011 - B6E0103

Calibration Blank (2605011-CCB1)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605011-CCB2)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605011-CCV1)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.0	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1
Crotonaldehyde	0.148	ug/ml	0.15000	98.4	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Benzaldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1

Calibration Check (2605011-CCV2)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605011 - B6E0103

Calibration Check (2605011-CCV2) Contin

Prepared & Analyzed: 05/04/26

Acetaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Butyraldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.4	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Hexaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1

Calibration Check (2605011-CCV3)

Prepared & Analyzed: 05/04/26

Formaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.6	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Crotonaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
2-Butanone	0.140	ug/ml	0.15000	93.2	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1

Batch 2605024 - B6E0702

Calibration Blank (2605024-CCB1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605024-CCB2)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605024 - B6E0702

Calibration Blank (2605024-CCB2) Contin

Prepared & Analyzed: 05/07/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605024-CCV1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.4	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1

Calibration Check (2605024-CCV2)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2605024-CCV3)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	94.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.1	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605024 - B6E0702

Calibration Check (2605024-CCV3) Contin

Prepared & Analyzed: 05/07/26

Valeraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1

Batch 2605066 - B6E1503

Calibration Blank (2605066-CCB1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605066-CCB2)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605066-CCV1)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.0	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Hexaldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1

Calibration Check (2605066-CCV2)

Prepared & Analyzed: 05/19/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605066 - B6E1503

Calibration Check (2605066-CCV2) Contin

Prepared & Analyzed: 05/19/26

Formaldehyde	0.146	ug/ml	0.15000	97.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1

Calibration Check (2605066-CCV3)

Prepared & Analyzed: 05/19/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.5	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.5	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.5	84.9-115.1

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2605069-CCB2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Blank (2605069-CCB2) Contin

Prepared & Analyzed: 05/20/26

Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605069-CCV1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.9	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
2-Butanone	0.154	ug/ml	0.15000	102	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.9	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1

Calibration Check (2605069-CCV2)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.3	84.9-115.1
Acetone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Calibration Check (2605069-CCV3)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.4	84.9-115.1
Propionaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Crotonaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Butyraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605069 - B6E2006

Calibration Check (2605069-CCV3) Contin

Prepared & Analyzed: 05/20/26

Benzaldehyde	0.158	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch 2605076 - B6E2205

Calibration Blank (2605076-CCB1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605076-CCB2)

Prepared: 05/22/26 Analyzed: 05/23/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605076-CCV1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.4	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.8	84.9-115.1
Crotonaldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.8	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.3	84.9-115.1
Benzaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Valeraldehyde	0.142	ug/ml	0.15000	94.8	84.9-115.1
Hexaldehyde	0.140	ug/ml	0.15000	93.4	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605076 - B6E2205

Calibration Check (2605076-CCV2)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.0	84.9-115.1
Acetone	0.144	ug/ml	0.15000	96.1	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.139	ug/ml	0.15000	92.7	84.9-115.1
2-Butanone	0.143	ug/ml	0.15000	95.0	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1

Calibration Check (2605076-CCV3)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.142	ug/ml	0.15000	95.0	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Propionaldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.9	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	96.0	84.9-115.1
Benzaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Hexaldehyde	0.139	ug/ml	0.15000	92.8	84.9-115.1

Calibration Check (2605076-CCV4)

Prepared: 05/22/26 Analyzed: 05/23/26

Formaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Propionaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	94.1	84.9-115.1
2-Butanone	0.142	ug/ml	0.15000	94.9	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Valeraldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1
Hexaldehyde	0.141	ug/ml	0.15000	94.0	84.9-115.1

Batch B6D1701 - TO-11A extraction

Blank (B6D1701-BLK1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	ND	ug/m ³ Air		U
Acetaldehyde	0.0326	ug/m ³ Air		
Acetone	ND	ug/m ³ Air		U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1701 - TO-11A extraction

Blank (B6D1701-BLK1) Continued

Prepared: 04/17/26 Analyzed: 04/20/26

Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0685	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6D1701-BLK3)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D1701-BS1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	0.298	ug/ml	0.30000		99.2	79.9-120.1			
Acetaldehyde	0.308	ug/ml	0.30000		103	79.9-120.1			
Acetone	0.315	ug/ml	0.30000		105	69.9-130.1			
Propionaldehyde	0.343	ug/ml	0.30000		114	69.9-130.1			
Crotonaldehyde	0.294	ug/ml	0.30000		97.9	69.9-130.1			
Butyraldehyde	0.273	ug/ml	0.30000		91.0	69.9-130.1			
2-Butanone	0.250	ug/ml	0.30000		83.3	69.9-130.1			
Benzaldehyde	0.331	ug/ml	0.30000		110	69.9-130.1			
Valeraldehyde	0.271	ug/ml	0.30000		90.3	69.9-130.1			
Hexaldehyde	0.265	ug/ml	0.30000		88.2	69.9-130.1			

LCS Dup (B6D1701-BSD1)

Prepared: 04/17/26 Analyzed: 04/20/26

Formaldehyde	0.286	ug/ml	0.30000		95.2	79.9-120.1	4.09	20	
Acetaldehyde	0.294	ug/ml	0.30000		98.1	79.9-120.1	4.44	20	
Acetone	0.299	ug/ml	0.30000		99.8	69.9-130.1	4.98	20	
Propionaldehyde	0.328	ug/ml	0.30000		109	69.9-130.1	4.29	20	
Crotonaldehyde	0.282	ug/ml	0.30000		93.9	69.9-130.1	4.18	20	
Butyraldehyde	0.259	ug/ml	0.30000		86.2	69.9-130.1	5.34	20	
2-Butanone	0.239	ug/ml	0.30000		79.8	69.9-130.1	4.31	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1701 - TO-11A extraction

LCS Dup (B6D1701-BSD1) Continued

Prepared: 04/17/26 Analyzed: 04/20/26

Benzaldehyde	0.320	ug/ml	0.30000	107	69.9-130.1	3.30	20
Valeraldehyde	0.265	ug/ml	0.30000	88.3	69.9-130.1	2.23	20
Hexaldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	3.84	20

Batch B6D2305 - TO-11A extraction

Blank (B6D2305-BLK1)

Prepared & Analyzed: 04/23/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0281	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	0.0441	ug/m ³ Air					
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6D2305-BLK3)

Prepared & Analyzed: 04/23/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6D2305-BS1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.283	ug/ml	0.30000	94.3	79.9-120.1
Acetaldehyde	0.294	ug/ml	0.30000	98.1	79.9-120.1
Acetone	0.297	ug/ml	0.30000	98.9	69.9-130.1
Propionaldehyde	0.329	ug/ml	0.30000	110	69.9-130.1
Crotonaldehyde	0.289	ug/ml	0.30000	96.3	69.9-130.1
Butyraldehyde	0.262	ug/ml	0.30000	87.2	69.9-130.1
2-Butanone	0.240	ug/ml	0.30000	79.9	69.9-130.1
Benzaldehyde	0.317	ug/ml	0.30000	106	69.9-130.1
Valeraldehyde	0.263	ug/ml	0.30000	87.7	69.9-130.1
Hexaldehyde	0.256	ug/ml	0.30000	85.4	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D2305 - TO-11A extraction

LCS Dup (B6D2305-BS1)

Prepared & Analyzed: 04/23/26

Formaldehyde	0.284	ug/ml	0.30000		94.6	79.9-120.1	0.236	20	
Acetaldehyde	0.296	ug/ml	0.30000		98.8	79.9-120.1	0.747	20	
Acetone	0.298	ug/ml	0.30000		99.2	69.9-130.1	0.274	20	
Propionaldehyde	0.330	ug/ml	0.30000		110	69.9-130.1	0.258	20	
Crotonaldehyde	0.288	ug/ml	0.30000		96.1	69.9-130.1	0.144	20	
Butyraldehyde	0.261	ug/ml	0.30000		87.0	69.9-130.1	0.274	20	
2-Butanone	0.237	ug/ml	0.30000		79.0	69.9-130.1	1.25	20	
Benzaldehyde	0.319	ug/ml	0.30000		106	69.9-130.1	0.707	20	
Valeraldehyde	0.260	ug/ml	0.30000		86.6	69.9-130.1	1.32	20	
Hexaldehyde	0.255	ug/ml	0.30000		84.9	69.9-130.1	0.592	20	

Batch B6E0103 - TO-11A extraction

Blank (B6E0103-BLK1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0369	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E0103-BLK3)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E0103-BS1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	0.290	ug/ml	0.30000		96.6	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.298	ug/ml	0.30000		99.3	69.9-130.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0103 - TO-11A extraction

LCS (B6E0103-BS1) Continued

Prepared: 05/01/26 Analyzed: 05/04/26

Propionaldehyde	0.370	ug/ml	0.30000	123	69.9-130.1
Crotonaldehyde	0.286	ug/ml	0.30000	95.4	69.9-130.1
Butyraldehyde	0.259	ug/ml	0.30000	86.4	69.9-130.1
2-Butanone	0.237	ug/ml	0.30000	79.2	69.9-130.1
Benzaldehyde	0.312	ug/ml	0.30000	104	69.9-130.1
Valeraldehyde	0.261	ug/ml	0.30000	86.9	69.9-130.1
Hexaldehyde	0.249	ug/ml	0.30000	83.0	69.9-130.1

LCS Dup (B6E0103-BSD1)

Prepared: 05/01/26 Analyzed: 05/04/26

Formaldehyde	0.288	ug/ml	0.30000	96.0	79.9-120.1	0.640	20
Acetaldehyde	0.287	ug/ml	0.30000	95.7	79.9-120.1	0.479	20
Acetone	0.293	ug/ml	0.30000	97.5	69.9-130.1	1.79	20
Propionaldehyde	0.376	ug/ml	0.30000	125	69.9-130.1	1.60	20
Crotonaldehyde	0.280	ug/ml	0.30000	93.3	69.9-130.1	2.16	20
Butyraldehyde	0.256	ug/ml	0.30000	85.2	69.9-130.1	1.45	20
2-Butanone	0.236	ug/ml	0.30000	78.5	69.9-130.1	0.807	20
Benzaldehyde	0.312	ug/ml	0.30000	104	69.9-130.1	0.119	20
Valeraldehyde	0.258	ug/ml	0.30000	86.1	69.9-130.1	0.976	20
Hexaldehyde	0.250	ug/ml	0.30000	83.3	69.9-130.1	0.401	20

Batch B6E0702 - TO-11A extraction

Blank (B6E0702-BLK1)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0303	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	0.0517	ug/m ³ Air					
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6E0702-BLK3)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Air Toxics by EPA Compendium Method TO-11A - Quality Control									
<i>Batch B6E0702 - TO-11A extraction</i>									
Blank (B6E0702-BLK3) Continued					Prepared & Analyzed: 05/07/26				
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U
LCS (B6E0702-BS1)					Prepared & Analyzed: 05/07/26				
Formaldehyde	0.288	ug/ml	0.30000		96.0	79.9-120.1			
Acetaldehyde	0.294	ug/ml	0.30000		98.1	79.9-120.1			
Acetone	0.294	ug/ml	0.30000		98.1	69.9-130.1			
Propionaldehyde	0.335	ug/ml	0.30000		112	69.9-130.1			
Crotonaldehyde	0.277	ug/ml	0.30000		92.2	69.9-130.1			
Butyraldehyde	0.260	ug/ml	0.30000		86.8	69.9-130.1			
2-Butanone	0.242	ug/ml	0.30000		80.7	69.9-130.1			
Benzaldehyde	0.310	ug/ml	0.30000		103	69.9-130.1			
Valeraldehyde	0.259	ug/ml	0.30000		86.4	69.9-130.1			
Hexaldehyde	0.251	ug/ml	0.30000		83.6	69.9-130.1			
LCS Dup (B6E0702-BSD1)					Prepared & Analyzed: 05/07/26				
Formaldehyde	0.288	ug/ml	0.30000		95.9	79.9-120.1	0.125	20	
Acetaldehyde	0.294	ug/ml	0.30000		97.9	79.9-120.1	0.145	20	
Acetone	0.293	ug/ml	0.30000		97.6	69.9-130.1	0.495	20	
Propionaldehyde	0.334	ug/ml	0.30000		111	69.9-130.1	0.412	20	
Crotonaldehyde	0.277	ug/ml	0.30000		92.4	69.9-130.1	0.162	20	
Butyraldehyde	0.261	ug/ml	0.30000		87.0	69.9-130.1	0.315	20	
2-Butanone	0.244	ug/ml	0.30000		81.4	69.9-130.1	0.906	20	
Benzaldehyde	0.314	ug/ml	0.30000		105	69.9-130.1	1.30	20	
Valeraldehyde	0.259	ug/ml	0.30000		86.5	69.9-130.1	0.0858	20	
Hexaldehyde	0.251	ug/ml	0.30000		83.6	69.9-130.1	0.102	20	
Duplicate (B6E0702-DUP1)					Source: 6043006-02 Prepared & Analyzed: 05/07/26				
Formaldehyde	1.43	ug/m ³ Air		1.43			0.124	10.1	
Acetaldehyde	1.00	ug/m ³ Air		1.01			0.255	10.1	
Acetone	1.56	ug/m ³ Air		1.55			0.148	10.1	
Propionaldehyde	0.306	ug/m ³ Air		0.325			5.93	10.1	
Crotonaldehyde	ND	ug/m ³ Air		ND				10.1	U
Butyraldehyde	0.140	ug/m ³ Air		0.136			3.05	10.1	
2-Butanone	0.550	ug/m ³ Air		0.542			1.39	10.1	
Benzaldehyde	0.0802	ug/m ³ Air		0.0851			5.85	10.1	
Valeraldehyde	0.0629	ug/m ³ Air		0.0570			9.83	10.1	
Hexaldehyde	0.0726	ug/m ³ Air		0.0737			1.59	10.1	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1503 - TO-11A extraction

Blank (B6E1503-BLK1)

Prepared: 05/15/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E1503-BLK3)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E1503-BS1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.325	ug/ml	0.30000	108	79.9-120.1
Acetaldehyde	0.285	ug/ml	0.30000	95.0	79.9-120.1
Acetone	0.295	ug/ml	0.30000	98.5	69.9-130.1
Propionaldehyde	0.266	ug/ml	0.30000	88.6	69.9-130.1
Crotonaldehyde	0.278	ug/ml	0.30000	92.6	69.9-130.1
Butyraldehyde	0.250	ug/ml	0.30000	83.3	69.9-130.1
2-Butanone	0.241	ug/ml	0.30000	80.4	69.9-130.1
Benzaldehyde	0.308	ug/ml	0.30000	103	69.9-130.1
Valeraldehyde	0.249	ug/ml	0.30000	83.0	69.9-130.1
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1

LCS Dup (B6E1503-BSD1)

Prepared: 05/15/26 Analyzed: 05/19/26

Formaldehyde	0.332	ug/ml	0.30000	111	79.9-120.1	2.31	20
Acetaldehyde	0.289	ug/ml	0.30000	96.4	79.9-120.1	1.46	20
Acetone	0.301	ug/ml	0.30000	100	69.9-130.1	2.03	20
Propionaldehyde	0.267	ug/ml	0.30000	89.1	69.9-130.1	0.612	20

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1503 - TO-11A extraction

LCS Dup (B6E1503-BSD1) Continued

Prepared: 05/15/26 Analyzed: 05/19/26

Crotonaldehyde	0.283	ug/ml	0.30000		94.4	69.9-130.1	1.90	20	
Butyraldehyde	0.255	ug/ml	0.30000		85.1	69.9-130.1	2.14	20	
2-Butanone	0.243	ug/ml	0.30000		81.0	69.9-130.1	0.708	20	
Benzaldehyde	0.311	ug/ml	0.30000		104	69.9-130.1	1.04	20	
Valeraldehyde	0.254	ug/ml	0.30000		84.8	69.9-130.1	2.11	20	
Hexaldehyde	0.243	ug/ml	0.30000		81.0	69.9-130.1	1.12	20	

Batch B6E2006 - TO-11A extraction

Blank (B6E2006-BLK1)

Prepared: 05/20/26 Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E2006-BLK3)

Prepared & Analyzed: 05/20/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E2006-BS1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.298	ug/ml	0.30000		99.2	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.296	ug/ml	0.30000		98.8	69.9-130.1			
Propionaldehyde	0.269	ug/ml	0.30000		89.6	69.9-130.1			
Crotonaldehyde	0.291	ug/ml	0.30000		97.1	69.9-130.1			
Butyraldehyde	0.257	ug/ml	0.30000		85.5	69.9-130.1			
2-Butanone	0.240	ug/ml	0.30000		79.9	69.9-130.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2006 - TO-11A extraction

LCS (B6E2006-BS1) Continued

Prepared & Analyzed: 05/20/26

Benzaldehyde	0.310	ug/ml	0.30000	103	69.9-130.1
Valeraldehyde	0.255	ug/ml	0.30000	85.0	69.9-130.1
Hexaldehyde	0.244	ug/ml	0.30000	81.4	69.9-130.1

LCS Dup (B6E2006-BSD1)

Prepared & Analyzed: 05/20/26

Formaldehyde	0.297	ug/ml	0.30000	98.9	79.9-120.1	0.349	20
Acetaldehyde	0.290	ug/ml	0.30000	96.7	79.9-120.1	0.499	20
Acetone	0.294	ug/ml	0.30000	98.1	69.9-130.1	0.653	20
Propionaldehyde	0.267	ug/ml	0.30000	88.9	69.9-130.1	0.876	20
Crotonaldehyde	0.281	ug/ml	0.30000	93.7	69.9-130.1	3.50	20
Butyraldehyde	0.253	ug/ml	0.30000	84.3	69.9-130.1	1.36	20
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.186	20
Benzaldehyde	0.307	ug/ml	0.30000	102	69.9-130.1	0.986	20
Valeraldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	0.252	20
Hexaldehyde	0.246	ug/ml	0.30000	82.0	69.9-130.1	0.678	20

Batch B6E2205 - TO-11A extraction

Blank (B6E2205-BLK1)

Prepared & Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0284	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6E2205-BLK3)

Prepared & Analyzed: 05/22/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E2205 - TO-11A extraction

LCS (B6E2205-BS1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.274	ug/ml	0.30000		91.5	79.9-120.1
Acetaldehyde	0.288	ug/ml	0.30000		96.1	79.9-120.1
Acetone	0.289	ug/ml	0.30000		96.2	69.9-130.1
Propionaldehyde	0.302	ug/ml	0.30000		101	69.9-130.1
Crotonaldehyde	0.281	ug/ml	0.30000		93.5	69.9-130.1
Butyraldehyde	0.245	ug/ml	0.30000		81.8	69.9-130.1
2-Butanone	0.230	ug/ml	0.30000		76.5	69.9-130.1
Benzaldehyde	0.298	ug/ml	0.30000		99.3	69.9-130.1
Valeraldehyde	0.248	ug/ml	0.30000		82.7	69.9-130.1
Hexaldehyde	0.242	ug/ml	0.30000		80.8	69.9-130.1

LCS Dup (B6E2205-BSD1)

Prepared & Analyzed: 05/22/26

Formaldehyde	0.274	ug/ml	0.30000		91.2	79.9-120.1	0.300	20
Acetaldehyde	0.285	ug/ml	0.30000		95.0	79.9-120.1	1.19	20
Acetone	0.286	ug/ml	0.30000		95.3	69.9-130.1	0.879	20
Propionaldehyde	0.298	ug/ml	0.30000		99.4	69.9-130.1	1.45	20
Crotonaldehyde	0.277	ug/ml	0.30000		92.4	69.9-130.1	1.27	20
Butyraldehyde	0.246	ug/ml	0.30000		81.9	69.9-130.1	0.192	20
2-Butanone	0.228	ug/ml	0.30000		75.9	69.9-130.1	0.879	20
Benzaldehyde	0.296	ug/ml	0.30000		98.7	69.9-130.1	0.681	20
Valeraldehyde	0.245	ug/ml	0.30000		81.6	69.9-130.1	1.42	20
Hexaldehyde	0.243	ug/ml	0.30000		81.0	69.9-130.1	0.229	20

Duplicate (B6E2205-DUP9)

Source: 6051949-04 Prepared & Analyzed: 05/22/26

Formaldehyde	4.47	ug/m ³ Air		4.47		0.0124	10.1
Acetaldehyde	2.14	ug/m ³ Air		2.14		0.145	10.1
Acetone	2.10	ug/m ³ Air		2.11		0.444	10.1
Propionaldehyde	0.571	ug/m ³ Air		0.564		1.14	10.1
Crotonaldehyde	ND	ug/m ³ Air		ND			10.1
Butyraldehyde	0.188	ug/m ³ Air		0.184		1.86	10.1
2-Butanone	0.261	ug/m ³ Air		0.258		1.00	10.1
Benzaldehyde	0.0909	ug/m ³ Air		0.0892		1.91	10.1
Valeraldehyde	0.191	ug/m ³ Air		0.183		4.24	10.1
Hexaldehyde	0.107	ug/m ³ Air		0.126		15.8	10.1

U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:26

SUBMITTED: 04/16/26 to 05/19/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
QB-01	Analyte exceeds method blank criteria
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
D-F	Duplicate exceeds DQO criteria.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.