



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 15, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/15/26 12:17 through 05/01/26 11:31.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6041537-01	Air	04/01/26 23:59	04/15/26 12:17
TOOK	6041537-02	Air	04/02/26 23:59	04/15/26 12:17
TOOK	6041537-03	Air	04/08/26 23:59	04/15/26 12:17
TOOK	6042063-01	Air	04/14/26 23:59	04/20/26 11:11
TOOK	6042412-01	Air	04/20/26 23:59	04/24/26 11:36
TOOK	6050127-01	Air	04/26/26 23:59	05/01/26 11:31



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK	Lab ID: 6041537-01	Sampled: 04/01/26 23:59
Pressure @ Receipt: 3.5	Canister #: 41629	Received: 04/15/26 12:17
Comments: Makeup for 2-7-26		Analyzed: 04/24/26 13:36

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.508	2.52		0.0237
Chloromethane	0.545	1.13		0.0213
Dichlorotetrafluoroethane	0.0176	0.12		0.0112
Vinyl chloride	7.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00820	0.02	CF, LL, U	0.0215
Ethylene oxide	0.109	0.20	CF, LK	0.0276
Bromomethane	0.00920	0.04	U	0.0104
Chloroethane	0.0186	0.05		0.0136
Acetonitrile	17.2	28.90	E	0.184
Acrolein	0.219	0.50	CF, LK	0.0977
Trichlorofluoromethane	0.236	1.33		0.00950
Acrylonitrile	0.0137	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.0982	0.34		0.0489
Carbon Disulfide	0.0108	0.03	U	0.0473
Trichlorotrifluoroethane	0.0800	0.61		0.0210
trans-1,2-Dichloroethylene	4.00E-4	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	0.0187	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0116	0.05		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.205	0.66		0.0158
Carbon Tetrachloride	0.0779	0.49		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U, CE	0.00889
Trichloroethylene	0.00220	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0190	0.08	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041537-01

Sampled: 04/01/26 23:59

Pressure @ Receipt: 3.5

Canister #: 41629

Received: 04/15/26 12:17

Comments: Makeup for 2-7-26

Analyzed: 04/24/26 13:36

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	0.0184	0.10		0.00749
Toluene	0.361	1.36		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0613	0.29		0.0258
Tetrachloroethylene	0.0164	0.11		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0428	0.19		0.0130
m,p-Xylene	0.129	0.56		0.0418
Bromoform	0.00260	0.03	U	0.0166
Styrene	0.0169	0.07	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0440	0.19		0.0235
1,3,5-Trimethylbenzene	0.00960	0.05	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00190	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041537-02

Sampled: 04/02/26 23:59

Pressure @ Receipt: 4.5

Canister #: 40446

Received: 04/15/26 12:17

Comments:

Analyzed: 04/24/26 14:37

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.513	2.54		0.0237
Chloromethane	0.550	1.14		0.0213
Dichlorotetrafluoroethane	0.0183	0.13		0.0112
Vinyl chloride	6.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00580	0.01	CF, LL, U	0.0215
Ethylene oxide	0.0682	0.12		0.0276
Bromomethane	0.00930	0.04	U	0.0104
Chloroethane	0.00790	0.02	U	0.0136
Acetonitrile	72.3	122.00	E, QB-03	0.184
Acrolein	0.210	0.48	CF, LK	0.0977
Trichlorofluoromethane	0.236	1.33		0.00950
Acrylonitrile	0.0102	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.120	0.42	QB-03	0.0489
Carbon Disulfide	0.0118	0.04	U	0.0473
Trichlorotrifluoroethane	0.0799	0.61		0.0210
trans-1,2-Dichloroethylene	0.00140	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0168	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0108	0.04	U	0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.190	0.61		0.0158
Carbon Tetrachloride	0.0775	0.49		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	CE, U	0.00889
Trichloroethylene	0.00220	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	ND	ND	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041537-02

Sampled: 04/02/26 23:59

Pressure @ Receipt: 4.5

Canister #: 40446

Received: 04/15/26 12:17

Comments:

Analyzed: 04/24/26 14:37

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.446	1.68		0.0380
Dibromochloromethane	5.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0837	0.39		0.0258
Tetrachloroethylene	0.0133	0.09		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0588	0.26		0.0130
m,p-Xylene	0.191	0.83		0.0418
Bromoform	0.00160	0.02	U	0.0166
Styrene	0.0329	0.14		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0685	0.30		0.0235
1,3,5-Trimethylbenzene	0.0129	0.06	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00170	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	4.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041537-03

Sampled: 04/08/26 23:59

Pressure @ Receipt: 4.5

Canister #: 912408

Received: 04/15/26 12:17

Comments:

Analyzed: 04/27/26 14:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.492	2.44		0.0237
Chloromethane	0.497	1.03		0.0213
Dichlorotetrafluoroethane	0.0175	0.12		0.0112
Vinyl chloride	0.00110	0.00	U	0.0140
1,3-Butadiene	0.0140	0.03	CF, LL, U	0.0215
Ethylene oxide	0.0839	0.15		0.0276
Bromomethane	0.00880	0.03	U	0.0104
Chloroethane	0.0104	0.03	U	0.0136
Acetonitrile	80.4	135.00	E, QB-01	0.184
Acrolein	0.185	0.43	CF, LK	0.0977
Trichlorofluoromethane	0.233	1.31		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.235	0.82		0.0489
Carbon Disulfide	0.0117	0.04	U	0.0473
Trichlorotrifluoroethane	0.0775	0.60		0.0210
trans-1,2-Dichloroethylene	0.00530	0.02	U	0.0171
1,1-Dichloroethane	0.00200	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	4.00E-4	0.00	U	0.00905
Chloroform	0.0268	0.13		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0305	0.12		0.0116
1,1,1-Trichloroethane	0.00150	0.01	U	0.00952
Benzene	0.397	1.27		0.0158
Carbon Tetrachloride	0.0761	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00310	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0127	0.05	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6041537-03

Sampled: 04/08/26 23:59

Pressure @ Receipt: 4.5

Canister #: 912408

Received: 04/15/26 12:17

Comments:

Analyzed: 04/27/26 14:14

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.876	3.31		0.0380
Dibromochloromethane	7.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0712	0.33		0.0258
Tetrachloroethylene	0.0185	0.13		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.109	0.47		0.0130
m,p-Xylene	0.366	1.59		0.0418
Bromoform	9.00E-4	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.131	0.57		0.0235
1,3,5-Trimethylbenzene	0.0228	0.11	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00260	0.02	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	3.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK
Pressure @ Receipt: 4.0
Comments:

Lab ID: 6042063-01
Canister #: 35124

Sampled: 04/14/26 23:59
Received: 04/20/26 11:11
Analyzed: 04/30/26 23:02

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.423	2.10		0.0237
Chloromethane	0.456	0.94		0.0213
Dichlorotetrafluoroethane	0.0142	0.10		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00560	0.01	U	0.0215
Ethylene oxide	0.0534	0.10		0.0276
Bromomethane	0.0121	0.05		0.0104
Chloroethane	0.00830	0.02	U	0.0136
Acetonitrile	73.7	124.00	E	0.184
Acrolein	0.226	0.52	CF, LK	0.0977
Trichlorofluoromethane	0.212	1.19		0.00950
Acrylonitrile	ND	ND	U	0.0116
1,1-Dichloroethene	0.00110	0.00	U	0.0158
Dichloromethane	0.123	0.43		0.0489
Carbon Disulfide	0.0160	0.05	U	0.0473
Trichlorotrifluoroethane	0.0689	0.53		0.0210
trans-1,2-Dichloroethylene	0.00210	0.01	U	0.0171
1,1-Dichloroethane	0.00180	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	0.00120	0.01	U	0.00905
Chloroform	0.0183	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0135	0.05		0.0116
1,1,1-Trichloroethane	0.00190	0.01	U	0.00952
Benzene	0.137	0.44		0.0158
Carbon Tetrachloride	0.0700	0.44		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00170	0.01	U	0.00889
Trichloroethylene	0.00320	0.02	U	0.0170
Methyl Methacrylate	ND	ND	LK, U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0140	0.06	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6042063-01

Sampled: 04/14/26 23:59

Pressure @ Receipt: 4.0

Canister #: 35124

Received: 04/20/26 11:11

Comments:

Analyzed: 04/30/26 23:02

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.303	1.14		0.0380
Dibromochloromethane	0.00150	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0370	0.17		0.0258
Tetrachloroethylene	0.0155	0.11		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0334	0.15		0.0130
m,p-Xylene	0.0984	0.43		0.0418
Bromoform	0.00360	0.04	U	0.0166
Styrene	0.0170	0.07	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0356	0.16		0.0235
1,3,5-Trimethylbenzene	0.00790	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	0.00120	0.01	U	0.0277
p-Dichlorobenzene	0.00340	0.02	U	0.0287
o-Dichlorobenzene	0.00120	0.01	U	0.0364
1,2,4-Trichlorobenzene	0.00270	0.02	QB-01, U	0.0792
Hexachloro-1,3-butadiene	8.00E-4	0.01	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK
Pressure @ Receipt: 3.5
Comments:

Lab ID: 6042412-01
Canister #: 35123

Sampled: 04/20/26 23:59
Received: 04/24/26 11:36
Analyzed: 05/02/26 03:18

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.471	2.33		0.0237
Chloromethane	0.528	1.09		0.0213
Dichlorotetrafluoroethane	0.0181	0.13		0.0112
Vinyl chloride	7.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0152	0.03	U	0.0215
Ethylene oxide	0.0417	0.08		0.0276
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.00870	0.02	U	0.0136
Acetonitrile	51.5	86.60	E	0.184
Acrolein	0.166	0.38	CF, LK	0.0977
Trichlorofluoromethane	0.216	1.22		0.00950
Acrylonitrile	0.0117	0.03		0.0116
1,1-Dichloroethene	3.00E-4	0.00	U	0.0158
Dichloromethane	0.144	0.50	CF, LK	0.0489
Carbon Disulfide	0.00870	0.03	U	0.0473
Trichlorotrifluoroethane	0.0755	0.58		0.0210
trans-1,2-Dichloroethylene	0.00180	0.01	U	0.0171
1,1-Dichloroethane	0.00160	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	7.00E-4	0.00	U	0.00905
Chloroform	0.0219	0.11		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0202	0.08		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.296	0.95		0.0158
Carbon Tetrachloride	0.0685	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	LL, U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00170	0.01	U	0.00889
Trichloroethylene	0.00260	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00980	0.04	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6042412-01

Sampled: 04/20/26 23:59

Pressure @ Receipt: 3.5

Canister #: 35123

Received: 04/24/26 11:36

Comments:

Analyzed: 05/02/26 03:18

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.464	1.75		0.0380
Dibromochloromethane	0.00110	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0549	0.26		0.0258
Tetrachloroethylene	0.0132	0.09		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0558	0.24		0.0130
m,p-Xylene	0.171	0.74		0.0418
Bromoform	0.00120	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0588	0.26		0.0235
1,3,5-Trimethylbenzene	0.0110	0.05	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00370	0.02	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK
Pressure @ Receipt: 4.0
Comments:

Lab ID: 6050127-01
Canister #: 40460

Sampled: 04/26/26 23:59
Received: 05/01/26 11:31
Analyzed: 05/15/26 20:03

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
Dichlorodifluoromethane	0.436	2.16		0.0237
Chloromethane	0.527	1.09		0.0213
Dichlorotetrafluoroethane	0.0146	0.10		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00390	0.01	CF, LL, U	0.0215
Ethylene oxide	0.0806	0.15	CF, LK	0.0276
Bromomethane	0.00780	0.03	U	0.0104
Chloroethane	0.00790	0.02	U	0.0136
Acetonitrile	44.6	75.00	E, QB-01	0.184
Acrolein	0.203	0.47	CF, LK, QB-01	0.0977
Trichlorofluoromethane	0.230	1.29	LK	0.00950
Acrylonitrile	0.0138	0.03		0.0116
1,1-Dichloroethene	ND	ND	LJ, U	0.0158
Dichloromethane	0.125	0.44	CF, LK	0.0489
Carbon Disulfide	0.0114	0.04	U	0.0473
Trichlorotrifluoroethane	0.0531	0.41		0.0210
trans-1,2-Dichloroethylene	2.00E-4	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	ND	ND	U	0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0137	0.06		0.0116
1,1,1-Trichloroethane	9.00E-4	0.00	U	0.00952
Benzene	0.253	0.81		0.0158
Carbon Tetrachloride	0.0725	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00240	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0163	0.07	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK
Pressure @ Receipt: 4.0
Comments:

Lab ID: 6050127-01
Canister #: 40460

Sampled: 04/26/26 23:59
Received: 05/01/26 11:31
Analyzed: 05/15/26 20:03

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	1.19	4.49	CF, LK	0.0380
Dibromochloromethane	ND	ND	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.920	4.31		0.0258
Tetrachloroethylene	0.0136	0.09		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.144	0.63		0.0130
m,p-Xylene	0.487	2.12		0.0418
Bromoform	0.00170	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.186	0.81		0.0235
1,3,5-Trimethylbenzene	0.0195	0.10	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00250	0.02	U	0.0287
o-Dichlorobenzene	4.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	9.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2402 - Summa Canister Prep

Blank (B6D2402-BLK1)

Prepared & Analyzed: 04/24/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2402 - Summa Canister Prep

Blank (B6D2402-BLK1) Continued

Prepared & Analyzed: 04/24/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D2708 - Summa Canister Prep

Blank (B6D2708-BLK1)

Prepared: 04/24/26 Analyzed: 04/27/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D2708 - Summa Canister Prep

Blank (B6D2708-BLK1) Continued

Prepared: 04/24/26 Analyzed: 04/27/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D3003 - Summa Canister Prep

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D3003 - Summa Canister Prep

Blank (B6D3003-BLK1)

Prepared: 04/27/26 Analyzed: 04/30/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				LK, U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D3003 - Summa Canister Prep

Blank (B6D3003-BLK1) Continued

Prepared: 04/27/26 Analyzed: 04/30/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1)

Prepared: 04/29/26 Analyzed: 05/01/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				LK, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E0102 - Summa Canister Prep

Blank (B6E0102-BLK1) Continued

Prepared: 04/29/26 Analyzed: 05/01/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				LL, U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E1502 - Summa Canister Prep

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1502 - Summa Canister Prep

Blank (B6E1502-BLK1)

Prepared: 04/29/26 Analyzed: 05/15/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				LK, QB-01, U
Trichlorofluoromethane	ND	ppbv				LK, U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				LJ, U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1502 - Summa Canister Prep

Blank (B6E1502-BLK1) Continued

Prepared: 04/29/26 Analyzed: 05/15/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604077

Calibration Check (2604077-CCV1)

Prepared & Analyzed: 04/24/26

Dichlorodifluoromethane	1.16	ppbv	13.2	30.00
Chloromethane	1.07	ppbv	4.7	30.00
Dichlorotetrafluoroethane	1.18	ppbv	13.5	30.00
Vinyl chloride	1.02	ppbv	-1.5	30.00
1,3-Butadiene	0.841	ppbv	-18.7	30.00
Ethylene oxide	0.973	ppbv	1.8	30.00
Bromomethane	1.21	ppbv	17.0	30.00
Chloroethane	1.19	ppbv	14.6	30.00
Acetonitrile	1.08	ppbv	4.2	30.00
Acrolein	1.09	ppbv	7.0	30.00
Trichlorofluoromethane	1.20	ppbv	18.9	30.00
Acrylonitrile	1.10	ppbv	-0.3	30.00
1,1-Dichloroethene	1.21	ppbv	16.1	30.00
Dichloromethane	1.11	ppbv	7.3	30.00
Carbon Disulfide	1.27	ppbv	16.6	30.00
Trichlorotrifluoroethane	1.27	ppbv	23.0	30.00
trans-1,2-Dichloroethylene	1.12	ppbv	7.6	30.00
1,1-Dichloroethane	1.15	ppbv	10.7	30.00
Methyl tert-Butyl Ether	0.899	ppbv	-13.2	30.00
Chloroprene	0.919	ppbv	-11.5	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	4.0	30.00
Bromochloromethane	1.21	ppbv	17.3	30.00
Chloroform	1.17	ppbv	12.1	30.00
Ethyl tert-Butyl Ether	0.857	ppbv	-17.2	30.00
1,2-Dichloroethane	1.13	ppbv	9.3	30.00
1,1,1-Trichloroethane	1.13	ppbv	10.0	30.00
Benzene	1.07	ppbv	3.3	30.00
Carbon Tetrachloride	1.15	ppbv	12.7	30.00
tert-Amyl Methyl Ether	0.814	ppbv	-21.3	30.00
1,2-Dichloropropane	1.01	ppbv	-2.9	30.00
Ethyl Acrylate	0.874	ppbv	-16.5	30.00
Bromodichloromethane	1.07	ppbv	3.4	30.00
Trichloroethylene	1.07	ppbv	3.6	30.00
Methyl Methacrylate	0.855	ppbv	-18.0	30.00
cis-1,3-Dichloropropene	0.950	ppbv	-9.0	30.00
Methyl Isobutyl Ketone	0.857	ppbv	-17.6	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604077

Calibration Check (2604077-CCV1) Continued

Prepared & Analyzed: 04/24/26

trans-1,3-Dichloropropene	1.05	ppbv	-0.2	30.00
1,1,2-Trichloroethane	1.13	ppbv	7.7	30.00
Toluene	0.989	ppbv	-4.9	30.00
Dibromochloromethane	1.08	ppbv	4.5	30.00
1,2-Dibromoethane	1.07	ppbv	2.2	30.00
n-Octane	0.877	ppbv	-15.5	30.00
Tetrachloroethylene	1.13	ppbv	8.1	30.00
Chlorobenzene	1.12	ppbv	6.9	30.00
Ethylbenzene	0.972	ppbv	-7.0	30.00
m,p-Xylene	2.07	ppbv	-0.7	30.00
Bromoform	1.20	ppbv	15.0	30.00
Styrene	1.07	ppbv	1.6	30.00
1,1,2,2-Tetrachloroethane	1.15	ppbv	8.8	30.00
o-Xylene	0.956	ppbv	-9.4	30.00
1,3,5-Trimethylbenzene	0.880	ppbv	-16.3	30.00
1,2,4-Trimethylbenzene	0.864	ppbv	-18.0	30.00
m-Dichlorobenzene	1.08	ppbv	3.7	30.00
p-Dichlorobenzene	1.06	ppbv	0.9	30.00
o-Dichlorobenzene	1.06	ppbv	1.1	30.00
1,2,4-Trichlorobenzene	0.913	ppbv	-12.0	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.6	30.00

Calibration Check (2604077-CCV2)

Prepared: 04/24/26 Analyzed: 04/25/26

Dichlorodifluoromethane	1.14	ppbv	10.8	30.00
Chloromethane	1.05	ppbv	3.3	30.00
Dichlorotetrafluoroethane	1.16	ppbv	12.1	30.00
Vinyl chloride	0.984	ppbv	-4.7	30.00
1,3-Butadiene	0.790	ppbv	-23.6	30.00
Ethylene oxide	0.948	ppbv	-0.8	30.00
Bromomethane	1.18	ppbv	14.8	30.00
Chloroethane	1.15	ppbv	11.5	30.00
Acetonitrile	1.08	ppbv	4.4	30.00
Acrolein	1.08	ppbv	5.5	30.00
Trichlorofluoromethane	1.17	ppbv	16.3	30.00
Acrylonitrile	1.08	ppbv	-2.4	30.00
1,1-Dichloroethene	1.19	ppbv	14.3	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604077

Calibration Check (2604077-CCV2) Continued

Prepared: 04/24/26 Analyzed: 04/25/26

Dichloromethane	1.09	ppbv	5.6	30.00
Carbon Disulfide	1.25	ppbv	15.3	30.00
Trichlorotrifluoroethane	1.25	ppbv	21.7	30.00
trans-1,2-Dichloroethylene	1.09	ppbv	4.4	30.00
1,1-Dichloroethane	1.13	ppbv	8.7	30.00
Methyl tert-Butyl Ether	0.856	ppbv	-17.4	30.00
Chloroprene	0.871	ppbv	-16.1	30.00
cis-1,2-Dichloroethylene	1.05	ppbv	1.2	30.00
Bromochloromethane	1.18	ppbv	15.2	30.00
Chloroform	1.16	ppbv	10.3	30.00
Ethyl tert-Butyl Ether	0.817	ppbv	-21.1	30.00
1,2-Dichloroethane	1.11	ppbv	7.1	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.3	30.00
Benzene	1.04	ppbv	0.4	30.00
Carbon Tetrachloride	1.12	ppbv	9.6	30.00
tert-Amyl Methyl Ether	0.770	ppbv	-25.5	30.00
1,2-Dichloropropane	1.00	ppbv	-3.6	30.00
Ethyl Acrylate	0.849	ppbv	-18.8	30.00
Bromodichloromethane	1.06	ppbv	2.2	30.00
Trichloroethylene	1.05	ppbv	1.7	30.00
Methyl Methacrylate	0.822	ppbv	-21.1	30.00
cis-1,3-Dichloropropene	0.920	ppbv	-11.9	30.00
Methyl Isobutyl Ketone	0.832	ppbv	-20.0	30.00
trans-1,3-Dichloropropene	1.02	ppbv	-2.4	30.00
1,1,2-Trichloroethane	1.12	ppbv	7.1	30.00
Toluene	0.964	ppbv	-7.3	30.00
Dibromochloromethane	1.06	ppbv	2.9	30.00
1,2-Dibromoethane	1.05	ppbv	0.5	30.00
n-Octane	0.851	ppbv	-18.0	30.00
Tetrachloroethylene	1.10	ppbv	5.7	30.00
Chlorobenzene	1.10	ppbv	5.0	30.00
Ethylbenzene	0.943	ppbv	-9.9	30.00
m,p-Xylene	2.01	ppbv	-3.6	30.00
Bromoform	1.18	ppbv	13.3	30.00
Styrene	1.05	ppbv	-1.0	30.00
1,1,2,2-Tetrachloroethane	1.14	ppbv	8.0	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604077

Calibration Check (2604077-CCV2) Continued

Prepared: 04/24/26 Analyzed: 04/25/26

o-Xylene	0.920	ppbv	-12.9	30.00
1,3,5-Trimethylbenzene	0.828	ppbv	-21.3	30.00
1,2,4-Trimethylbenzene	0.812	ppbv	-22.9	30.00
m-Dichlorobenzene	1.05	ppbv	0.6	30.00
p-Dichlorobenzene	1.02	ppbv	-2.6	30.00
o-Dichlorobenzene	1.02	ppbv	-2.6	30.00
1,2,4-Trichlorobenzene	0.864	ppbv	-16.8	30.00
Hexachloro-1,3-butadiene	1.01	ppbv	-4.1	30.00

Sequence 2604083

Calibration Check (2604083-CCV1)

Prepared & Analyzed: 04/27/26

Dichlorodifluoromethane	1.13	ppbv	9.9	30.00
Chloromethane	1.03	ppbv	0.8	30.00
Dichlorotetrafluoroethane	1.13	ppbv	9.4	30.00
Vinyl chloride	1.03	ppbv	-0.3	30.00
1,3-Butadiene	0.895	ppbv	-13.5	30.00
Ethylene oxide	0.932	ppbv	-2.5	30.00
Bromomethane	1.17	ppbv	13.0	30.00
Chloroethane	1.13	ppbv	9.5	30.00
Acetonitrile	1.05	ppbv	1.0	30.00
Acrolein	1.04	ppbv	2.4	30.00
Trichlorofluoromethane	1.18	ppbv	16.7	30.00
Acrylonitrile	1.08	ppbv	-1.8	30.00
1,1-Dichloroethene	1.16	ppbv	11.2	30.00
Dichloromethane	1.09	ppbv	5.3	30.00
Carbon Disulfide	1.23	ppbv	13.0	30.00
Trichlorotrifluoroethane	1.23	ppbv	19.7	30.00
trans-1,2-Dichloroethylene	1.11	ppbv	6.7	30.00
1,1-Dichloroethane	1.12	ppbv	7.9	30.00
Methyl tert-Butyl Ether	0.932	ppbv	-10.0	30.00
Chloroprene	0.898	ppbv	-13.5	30.00
cis-1,2-Dichloroethylene	1.05	ppbv	1.0	30.00
Bromochloromethane	1.19	ppbv	15.5	30.00
Chloroform	1.14	ppbv	9.2	30.00
Ethyl tert-Butyl Ether	0.874	ppbv	-15.6	30.00
1,2-Dichloroethane	1.10	ppbv	6.8	30.00

QB-01
LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV1) Continued

Prepared & Analyzed: 04/27/26

1,1,1-Trichloroethane	1.13	ppbv	9.5	30.00
Benzene	1.05	ppbv	1.4	30.00
Carbon Tetrachloride	1.15	ppbv	11.8	30.00
tert-Amyl Methyl Ether	0.844	ppbv	-18.3	30.00
1,2-Dichloropropane	1.07	ppbv	2.9	30.00
Ethyl Acrylate	0.855	ppbv	-18.3	30.00
Bromodichloromethane	1.04	ppbv	0.3	30.00
Trichloroethylene	1.05	ppbv	2.0	30.00
Methyl Methacrylate	0.825	ppbv	-20.8	30.00
cis-1,3-Dichloropropene	0.935	ppbv	-10.4	30.00
Methyl Isobutyl Ketone	0.832	ppbv	-20.0	30.00
trans-1,3-Dichloropropene	1.03	ppbv	-2.3	30.00
1,1,2-Trichloroethane	1.09	ppbv	4.4	30.00
Toluene	0.970	ppbv	-6.8	30.00
Dibromochloromethane	1.06	ppbv	2.4	30.00
1,2-Dibromoethane	1.04	ppbv	-0.3	30.00
n-Octane	0.844	ppbv	-18.7	30.00
Tetrachloroethylene	1.13	ppbv	7.8	30.00
Chlorobenzene	1.10	ppbv	5.0	30.00
Ethylbenzene	0.954	ppbv	-8.8	30.00
m,p-Xylene	2.01	ppbv	-3.7	30.00
Bromoform	1.17	ppbv	12.1	30.00
Styrene	1.04	ppbv	-1.9	30.00
1,1,2,2-Tetrachloroethane	1.10	ppbv	4.1	30.00
o-Xylene	0.934	ppbv	-11.5	30.00
1,3,5-Trimethylbenzene	0.865	ppbv	-17.7	30.00
1,2,4-Trimethylbenzene	0.852	ppbv	-19.2	30.00
m-Dichlorobenzene	1.07	ppbv	2.9	30.00
p-Dichlorobenzene	1.05	ppbv	0.3	30.00
o-Dichlorobenzene	1.05	ppbv	0.5	30.00
1,2,4-Trichlorobenzene	0.911	ppbv	-12.2	30.00
Hexachloro-1,3-butadiene	1.04	ppbv	-1.2	30.00

Calibration Check (2604083-CCV2)

Prepared: 04/27/26 Analyzed: 04/28/26

Dichlorodifluoromethane	1.11	ppbv	7.9	30.00
Chloromethane	1.03	ppbv	1.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV2) Continued

Prepared: 04/27/26 Analyzed: 04/28/26

Dichlorotetrafluoroethane	1.13	ppbv	9.5	30.00	
Vinyl chloride	0.962	ppbv	-6.8	30.00	
1,3-Butadiene	0.755	ppbv	-27.0	30.00	
Ethylene oxide	0.924	ppbv	-3.3	30.00	
Bromomethane	1.16	ppbv	12.6	30.00	
Chloroethane	1.14	ppbv	10.1	30.00	
Acetonitrile	0.807	ppbv	-22.2	30.00	QB-01
Acrolein	1.06	ppbv	3.9	30.00	LK
Trichlorofluoromethane	1.14	ppbv	12.6	30.00	
Acrylonitrile	1.09	ppbv	-1.1	30.00	
1,1-Dichloroethene	1.16	ppbv	11.8	30.00	
Dichloromethane	1.15	ppbv	10.8	30.00	
Carbon Disulfide	1.24	ppbv	14.0	30.00	
Trichlorotrifluoroethane	1.23	ppbv	19.2	30.00	
trans-1,2-Dichloroethylene	1.09	ppbv	5.2	30.00	
1,1-Dichloroethane	1.12	ppbv	7.7	30.00	
Methyl tert-Butyl Ether	0.871	ppbv	-15.9	30.00	
Chloroprene	0.834	ppbv	-19.6	30.00	
cis-1,2-Dichloroethylene	1.02	ppbv	-1.2	30.00	
Bromochloromethane	1.16	ppbv	13.1	30.00	
Chloroform	1.13	ppbv	8.1	30.00	
Ethyl tert-Butyl Ether	0.796	ppbv	-23.2	30.00	
1,2-Dichloroethane	1.08	ppbv	4.1	30.00	
1,1,1-Trichloroethane	1.08	ppbv	4.9	30.00	
Benzene	1.02	ppbv	-2.1	30.00	
Carbon Tetrachloride	1.09	ppbv	6.1	30.00	
tert-Amyl Methyl Ether	0.758	ppbv	-26.7	30.00	
1,2-Dichloropropane	0.998	ppbv	-3.9	30.00	
Ethyl Acrylate	0.835	ppbv	-20.2	30.00	
Bromodichloromethane	1.05	ppbv	1.4	30.00	
Trichloroethylene	1.04	ppbv	0.4	30.00	
Methyl Methacrylate	0.804	ppbv	-22.9	30.00	
cis-1,3-Dichloropropene	0.896	ppbv	-14.2	30.00	
Methyl Isobutyl Ketone	0.823	ppbv	-20.9	30.00	
trans-1,3-Dichloropropene	1.01	ppbv	-3.4	30.00	
1,1,2-Trichloroethane	1.12	ppbv	6.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604083

Calibration Check (2604083-CCV2) Continued

Prepared: 04/27/26 Analyzed: 04/28/26

Toluene	0.943	ppbv	-9.3	30.00
Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.03	ppbv	-1.3	30.00
n-Octane	0.836	ppbv	-19.5	30.00
Tetrachloroethylene	1.09	ppbv	4.1	30.00
Chlorobenzene	1.11	ppbv	5.1	30.00
Ethylbenzene	0.922	ppbv	-11.9	30.00
m,p-Xylene	1.98	ppbv	-5.1	30.00
Bromoform	1.18	ppbv	12.5	30.00
Styrene	1.03	ppbv	-2.6	30.00
1,1,2,2-Tetrachloroethane	1.15	ppbv	9.1	30.00
o-Xylene	0.892	ppbv	-15.5	30.00
1,3,5-Trimethylbenzene	0.790	ppbv	-24.9	30.00
1,2,4-Trimethylbenzene	0.777	ppbv	-26.3	30.00
m-Dichlorobenzene	1.04	ppbv	-0.5	30.00
p-Dichlorobenzene	1.00	ppbv	-4.1	30.00
o-Dichlorobenzene	1.00	ppbv	-4.2	30.00
1,2,4-Trichlorobenzene	0.830	ppbv	-20.0	30.00
Hexachloro-1,3-butadiene	1.01	ppbv	-4.6	30.00

Sequence 2604097

Calibration Check (2604097-CCV1)

Prepared & Analyzed: 04/30/26

Dichlorodifluoromethane	0.984	ppbv	-4.2	30.00
Chloromethane	0.916	ppbv	-10.2	30.00
Dichlorotetrafluoroethane	0.961	ppbv	-7.2	30.00
Vinyl chloride	0.948	ppbv	-8.1	30.00
1,3-Butadiene	0.949	ppbv	-8.2	30.00
Ethylene oxide	0.954	ppbv	-0.2	30.00
Bromomethane	0.995	ppbv	-3.6	30.00
Chloroethane	1.01	ppbv	-2.5	30.00
Acetonitrile	0.970	ppbv	-6.5	30.00
Acrolein	1.02	ppbv	-0.2	30.00
Trichlorofluoromethane	1.01	ppbv	0.09	30.00
Acrylonitrile	1.08	ppbv	-2.1	30.00
1,1-Dichloroethene	1.02	ppbv	-2.2	30.00
Dichloromethane	0.966	ppbv	-6.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604097

Calibration Check (2604097-CCV1) Continued

Prepared & Analyzed: 04/30/26

Carbon Disulfide	1.01	ppbv	-7.5	30.00
Trichlorotrifluoroethane	1.01	ppbv	-1.7	30.00
trans-1,2-Dichloroethylene	0.978	ppbv	-6.0	30.00
1,1-Dichloroethane	0.987	ppbv	-4.7	30.00
Methyl tert-Butyl Ether	0.995	ppbv	-3.9	30.00
Chloroprene	0.967	ppbv	-6.8	30.00
cis-1,2-Dichloroethylene	0.961	ppbv	-7.2	30.00
Bromochloromethane	0.970	ppbv	-5.6	30.00
Chloroform	0.981	ppbv	-6.4	30.00
Ethyl tert-Butyl Ether	0.971	ppbv	-6.3	30.00
1,2-Dichloroethane	1.01	ppbv	-2.6	30.00
1,1,1-Trichloroethane	0.986	ppbv	-4.3	30.00
Benzene	0.928	ppbv	-10.6	30.00
Carbon Tetrachloride	0.998	ppbv	-2.6	30.00
tert-Amyl Methyl Ether	0.921	ppbv	-10.9	30.00
1,2-Dichloropropane	1.07	ppbv	3.1	30.00
Ethyl Acrylate	1.03	ppbv	-1.1	30.00
Bromodichloromethane	1.05	ppbv	1.3	30.00
Trichloroethylene	1.05	ppbv	1.9	30.00
Methyl Methacrylate	0.992	ppbv	-4.8	30.00
cis-1,3-Dichloropropene	1.03	ppbv	-1.0	30.00
Methyl Isobutyl Ketone	0.986	ppbv	-5.2	30.00
trans-1,3-Dichloropropene	1.07	ppbv	1.5	30.00
1,1,2-Trichloroethane	1.02	ppbv	-2.6	30.00
Toluene	1.01	ppbv	-3.0	30.00
Dibromochloromethane	1.07	ppbv	3.3	30.00
1,2-Dibromoethane	1.02	ppbv	-2.3	30.00
n-Octane	0.990	ppbv	-4.6	30.00
Tetrachloroethylene	1.04	ppbv	-0.5	30.00
Chlorobenzene	1.05	ppbv	0.2	30.00
Ethylbenzene	0.996	ppbv	-4.8	30.00
m,p-Xylene	2.05	ppbv	-1.6	30.00
Bromoform	1.06	ppbv	1.4	30.00
Styrene	1.03	ppbv	-2.2	30.00
1,1,2,2-Tetrachloroethane	1.03	ppbv	-2.1	30.00
o-Xylene	1.02	ppbv	-3.4	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604097

Calibration Check (2604097-CCV1) Continued

Prepared & Analyzed: 04/30/26

1,3,5-Trimethylbenzene	0.999	ppbv	-5.0	30.00	QB-01
1,2,4-Trimethylbenzene	0.988	ppbv	-6.3	30.00	
m-Dichlorobenzene	1.04	ppbv	0.09	30.00	
p-Dichlorobenzene	1.04	ppbv	-1.0	30.00	
o-Dichlorobenzene	1.04	ppbv	-0.7	30.00	
1,2,4-Trichlorobenzene	0.994	ppbv	-4.2	30.00	
Hexachloro-1,3-butadiene	1.06	ppbv	0.1	30.00	

Calibration Check (2604097-CCV2)

Prepared: 04/30/26 Analyzed: 05/01/26

Dichlorodifluoromethane	1.07	ppbv	3.9	30.00
Chloromethane	1.02	ppbv	0.2	30.00
Dichlorotetrafluoroethane	1.05	ppbv	1.0	30.00
Vinyl chloride	0.943	ppbv	-8.6	30.00
1,3-Butadiene	0.781	ppbv	-24.5	30.00
Ethylene oxide	1.07	ppbv	12.1	30.00
Bromomethane	1.19	ppbv	15.3	30.00
Chloroethane	1.18	ppbv	13.5	30.00
Acetonitrile	1.20	ppbv	15.4	30.00
Acrolein	1.09	ppbv	6.8	30.00
Trichlorofluoromethane	1.14	ppbv	13.3	30.00
Acrylonitrile	1.13	ppbv	2.7	30.00
1,1-Dichloroethene	1.19	ppbv	14.0	30.00
Dichloromethane	1.14	ppbv	9.7	30.00
Carbon Disulfide	1.41	ppbv	29.6	30.00
Trichlorotrifluoroethane	1.16	ppbv	12.8	30.00
trans-1,2-Dichloroethylene	0.998	ppbv	-4.0	30.00
1,1-Dichloroethane	1.07	ppbv	3.5	30.00
Methyl tert-Butyl Ether	0.852	ppbv	-17.8	30.00
Chloroprene	0.857	ppbv	-17.4	30.00
cis-1,2-Dichloroethylene	0.964	ppbv	-6.9	30.00
Bromochloromethane	1.15	ppbv	11.5	30.00
Chloroform	1.22	ppbv	16.2	30.00
Ethyl tert-Butyl Ether	0.800	ppbv	-22.8	30.00
1,2-Dichloroethane	1.10	ppbv	6.6	30.00
1,1,1-Trichloroethane	1.15	ppbv	11.9	30.00
Benzene	1.05	ppbv	1.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604097

Calibration Check (2604097-CCV2) Continued

Prepared: 04/30/26 Analyzed: 05/01/26

Carbon Tetrachloride	1.17	ppbv	14.3	30.00	
tert-Amyl Methyl Ether	0.797	ppbv	-22.9	30.00	
1,2-Dichloropropane	1.04	ppbv	0.7	30.00	
Ethyl Acrylate	0.916	ppbv	-12.4	30.00	
Bromodichloromethane	1.21	ppbv	16.8	30.00	
Trichloroethylene	1.09	ppbv	5.1	30.00	
Methyl Methacrylate	0.907	ppbv	-12.9	30.00	LK
cis-1,3-Dichloropropene	0.880	ppbv	-15.7	30.00	
Methyl Isobutyl Ketone	0.973	ppbv	-6.5	30.00	
trans-1,3-Dichloropropene	0.990	ppbv	-5.7	30.00	
1,1,2-Trichloroethane	1.18	ppbv	12.5	30.00	
Toluene	1.00	ppbv	-3.7	30.00	
Dibromochloromethane	1.18	ppbv	14.1	30.00	
1,2-Dibromoethane	1.08	ppbv	3.4	30.00	
n-Octane	0.921	ppbv	-11.3	30.00	
Tetrachloroethylene	1.09	ppbv	4.7	30.00	
Chlorobenzene	1.14	ppbv	8.5	30.00	
Ethylbenzene	0.934	ppbv	-10.7	30.00	
m,p-Xylene	1.96	ppbv	-6.4	30.00	
Bromoform	1.14	ppbv	9.4	30.00	
Styrene	0.988	ppbv	-6.6	30.00	
1,1,2,2-Tetrachloroethane	1.20	ppbv	14.0	30.00	
o-Xylene	0.886	ppbv	-16.1	30.00	
1,3,5-Trimethylbenzene	0.794	ppbv	-24.5	30.00	
1,2,4-Trimethylbenzene	0.776	ppbv	-26.4	30.00	
m-Dichlorobenzene	1.03	ppbv	-1.2	30.00	
p-Dichlorobenzene	0.990	ppbv	-5.4	30.00	
o-Dichlorobenzene	1.00	ppbv	-4.2	30.00	
1,2,4-Trichlorobenzene	0.928	ppbv	-10.6	30.00	QB-01
Hexachloro-1,3-butadiene	1.14	ppbv	8.4	30.00	

Sequence 2605002

Calibration Check (2605002-CCV1)

Prepared & Analyzed: 05/01/26

Dichlorodifluoromethane	1.07	ppbv	3.9	30.00	
Chloromethane	1.04	ppbv	1.9	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	7.6	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV1) Continued

Prepared & Analyzed: 05/01/26

Vinyl chloride	0.944	ppbv	-8.5	30.00	LK
1,3-Butadiene	0.755	ppbv	-27.0	30.00	
Ethylene oxide	0.934	ppbv	-2.3	30.00	
Bromomethane	1.15	ppbv	11.1	30.00	
Chloroethane	1.16	ppbv	11.7	30.00	
Acetonitrile	1.05	ppbv	1.5	30.00	
Acrolein	1.06	ppbv	4.2	30.00	
Trichlorofluoromethane	1.09	ppbv	8.3	30.00	
Acrylonitrile	1.07	ppbv	-3.4	30.00	
1,1-Dichloroethene	1.18	ppbv	13.1	30.00	
Dichloromethane	1.09	ppbv	5.5	30.00	
Carbon Disulfide	1.26	ppbv	15.5	30.00	
Trichlorotrifluoroethane	1.21	ppbv	17.3	30.00	
trans-1,2-Dichloroethylene	1.03	ppbv	-0.7	30.00	
1,1-Dichloroethane	1.11	ppbv	7.2	30.00	
Methyl tert-Butyl Ether	0.802	ppbv	-22.6	30.00	
Chloroprene	0.823	ppbv	-20.7	30.00	
cis-1,2-Dichloroethylene	1.02	ppbv	-1.5	30.00	
Bromochloromethane	1.12	ppbv	9.3	30.00	
Chloroform	1.11	ppbv	6.1	30.00	
Ethyl tert-Butyl Ether	0.766	ppbv	-26.1	30.00	
1,2-Dichloroethane	1.05	ppbv	1.2	30.00	
1,1,1-Trichloroethane	1.03	ppbv	-0.3	30.00	
Benzene	0.998	ppbv	-3.8	30.00	
Carbon Tetrachloride	1.03	ppbv	0.3	30.00	
tert-Amyl Methyl Ether	0.735	ppbv	-28.9	30.00	
1,2-Dichloropropane	1.04	ppbv	0.6	30.00	
Ethyl Acrylate	0.865	ppbv	-17.3	30.00	
Bromodichloromethane	1.08	ppbv	4.2	30.00	
Trichloroethylene	1.03	ppbv	-0.07	30.00	
Methyl Methacrylate	0.832	ppbv	-20.2	30.00	
cis-1,3-Dichloropropene	0.902	ppbv	-13.6	30.00	
Methyl Isobutyl Ketone	0.842	ppbv	-19.0	30.00	
trans-1,3-Dichloropropene	1.03	ppbv	-2.1	30.00	
1,1,2-Trichloroethane	1.14	ppbv	9.1	30.00	
Toluene	0.952	ppbv	-8.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 06/15/26 09:51
SUBMITTED: 04/15/26 to 05/01/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV1) Continued

Prepared & Analyzed: 05/01/26

Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.02	ppbv	-2.4	30.00
n-Octane	0.875	ppbv	-15.7	30.00
Tetrachloroethylene	1.07	ppbv	2.7	30.00
Chlorobenzene	1.12	ppbv	6.5	30.00
Ethylbenzene	0.940	ppbv	-10.1	30.00
m,p-Xylene	2.04	ppbv	-2.4	30.00
Bromoform	1.18	ppbv	12.9	30.00
Styrene	1.06	ppbv	0.4	30.00
1,1,2,2-Tetrachloroethane	1.22	ppbv	15.6	30.00
o-Xylene	0.905	ppbv	-14.3	30.00
1,3,5-Trimethylbenzene	0.801	ppbv	-23.8	30.00
1,2,4-Trimethylbenzene	0.776	ppbv	-26.3	30.00
m-Dichlorobenzene	1.04	ppbv	-0.3	30.00
p-Dichlorobenzene	1.01	ppbv	-3.9	30.00
o-Dichlorobenzene	0.996	ppbv	-4.8	30.00
1,2,4-Trichlorobenzene	0.826	ppbv	-20.5	30.00
Hexachloro-1,3-butadiene	0.982	ppbv	-6.8	30.00

Calibration Check (2605002-CCV2)

Prepared: 05/01/26 Analyzed: 05/02/26

Dichlorodifluoromethane	1.07	ppbv	3.8	30.00
Chloromethane	1.05	ppbv	3.1	30.00
Dichlorotetrafluoroethane	1.12	ppbv	7.7	30.00
Vinyl chloride	0.945	ppbv	-8.5	30.00
1,3-Butadiene	0.737	ppbv	-28.7	30.00
Ethylene oxide	0.934	ppbv	-2.3	30.00
Bromomethane	1.15	ppbv	11.0	30.00
Chloroethane	1.16	ppbv	12.1	30.00
Acetonitrile	1.10	ppbv	5.6	30.00
Acrolein	1.05	ppbv	3.4	30.00
Trichlorofluoromethane	1.09	ppbv	7.4	30.00
Acrylonitrile	1.06	ppbv	-4.2	30.00
1,1-Dichloroethene	1.17	ppbv	12.9	30.00
Dichloromethane	1.11	ppbv	7.1	30.00
Carbon Disulfide	1.25	ppbv	14.8	30.00
Trichlorotrifluoroethane	1.20	ppbv	16.1	30.00

LK

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

trans-1,2-Dichloroethylene	1.03	ppbv	-1.2	30.00	
1,1-Dichloroethane	1.12	ppbv	7.8	30.00	
Methyl tert-Butyl Ether	0.789	ppbv	-23.8	30.00	
Chloroprene	0.800	ppbv	-22.9	30.00	
cis-1,2-Dichloroethylene	1.02	ppbv	-2.0	30.00	
Bromochloromethane	1.12	ppbv	8.5	30.00	
Chloroform	1.11	ppbv	5.9	30.00	
Ethyl tert-Butyl Ether	0.753	ppbv	-27.3	30.00	
1,2-Dichloroethane	1.04	ppbv	0.5	30.00	
1,1,1-Trichloroethane	1.02	ppbv	-1.2	30.00	
Benzene	0.991	ppbv	-4.5	30.00	
Carbon Tetrachloride	1.01	ppbv	-1.3	30.00	
tert-Amyl Methyl Ether	0.721	ppbv	-30.2	30.00	LL
1,2-Dichloropropane	1.06	ppbv	2.3	30.00	
Ethyl Acrylate	0.869	ppbv	-16.9	30.00	
Bromodichloromethane	1.09	ppbv	5.1	30.00	
Trichloroethylene	1.03	ppbv	-0.3	30.00	
Methyl Methacrylate	0.825	ppbv	-20.8	30.00	
cis-1,3-Dichloropropene	0.901	ppbv	-13.7	30.00	
Methyl Isobutyl Ketone	0.857	ppbv	-17.6	30.00	
trans-1,3-Dichloropropene	1.03	ppbv	-1.6	30.00	
1,1,2-Trichloroethane	1.16	ppbv	10.9	30.00	
Toluene	0.954	ppbv	-8.3	30.00	
Dibromochloromethane	1.06	ppbv	2.1	30.00	
1,2-Dibromoethane	1.02	ppbv	-1.8	30.00	
n-Octane	0.879	ppbv	-15.3	30.00	
Tetrachloroethylene	1.08	ppbv	3.2	30.00	
Chlorobenzene	1.12	ppbv	6.1	30.00	
Ethylbenzene	0.924	ppbv	-11.7	30.00	
m,p-Xylene	2.01	ppbv	-3.6	30.00	
Bromoform	1.18	ppbv	12.9	30.00	
Styrene	1.05	ppbv	-0.8	30.00	
1,1,2,2-Tetrachloroethane	1.23	ppbv	17.0	30.00	
o-Xylene	0.884	ppbv	-16.3	30.00	
1,3,5-Trimethylbenzene	0.768	ppbv	-27.0	30.00	
1,2,4-Trimethylbenzene	0.744	ppbv	-29.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605002

Calibration Check (2605002-CCV2) Continued

Prepared: 05/01/26 Analyzed: 05/02/26

m-Dichlorobenzene	1.02	ppbv	-2.0	30.00
p-Dichlorobenzene	0.986	ppbv	-5.8	30.00
o-Dichlorobenzene	0.976	ppbv	-6.7	30.00
1,2,4-Trichlorobenzene	0.805	ppbv	-22.4	30.00
Hexachloro-1,3-butadiene	0.972	ppbv	-7.7	30.00

Sequence 2605051

Calibration Check (2605051-CCV1)

Prepared & Analyzed: 05/15/26

Dichlorodifluoromethane	1.02	ppbv	-0.7	30.00
Chloromethane	0.951	ppbv	-6.7	30.00
Dichlorotetrafluoroethane	0.977	ppbv	-5.7	30.00
Vinyl chloride	0.956	ppbv	-7.3	30.00
1,3-Butadiene	0.914	ppbv	-11.7	30.00
Ethylene oxide	0.951	ppbv	-0.5	30.00
Bromomethane	1.01	ppbv	-2.1	30.00
Chloroethane	0.995	ppbv	-3.9	30.00
Acetonitrile	0.917	ppbv	-11.6	30.00
Acrolein	1.01	ppbv	-0.7	30.00
Trichlorofluoromethane	1.13	ppbv	12.3	30.00
Acrylonitrile	1.06	ppbv	-4.2	30.00
1,1-Dichloroethene	1.27	ppbv	22.3	30.00
Dichloromethane	0.937	ppbv	-9.6	30.00
Carbon Disulfide	1.19	ppbv	9.0	30.00
Trichlorotrifluoroethane	1.21	ppbv	17.2	30.00
trans-1,2-Dichloroethylene	0.963	ppbv	-7.4	30.00
1,1-Dichloroethane	0.967	ppbv	-6.6	30.00
Methyl tert-Butyl Ether	1.05	ppbv	1.2	30.00
Chloroprene	1.03	ppbv	-0.7	30.00
cis-1,2-Dichloroethylene	0.955	ppbv	-7.9	30.00
Bromochloromethane	0.931	ppbv	-9.4	30.00
Chloroform	1.11	ppbv	5.5	30.00
Ethyl tert-Butyl Ether	1.00	ppbv	-3.4	30.00
1,2-Dichloroethane	1.07	ppbv	3.4	30.00
1,1,1-Trichloroethane	1.12	ppbv	9.1	30.00
Benzene	1.03	ppbv	-0.8	30.00
Carbon Tetrachloride	1.12	ppbv	9.8	30.00

QB-01
LK, QB-01

LJ



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV1) Continued

Prepared & Analyzed: 05/15/26

tert-Amyl Methyl Ether	1.01	ppbv	-2.7	30.00
1,2-Dichloropropane	1.04	ppbv	0.03	30.00
Ethyl Acrylate	1.12	ppbv	6.7	30.00
Bromodichloromethane	1.17	ppbv	12.8	30.00
Trichloroethylene	1.16	ppbv	12.2	30.00
Methyl Methacrylate	1.15	ppbv	10.8	30.00
cis-1,3-Dichloropropene	1.09	ppbv	4.8	30.00
Methyl Isobutyl Ketone	1.14	ppbv	9.4	30.00
trans-1,3-Dichloropropene	1.12	ppbv	6.2	30.00
1,1,2-Trichloroethane	1.12	ppbv	6.9	30.00
Toluene	1.20	ppbv	15.0	30.00
Dibromochloromethane	1.09	ppbv	5.7	30.00
1,2-Dibromoethane	1.13	ppbv	8.2	30.00
n-Octane	1.18	ppbv	13.5	30.00
Tetrachloroethylene	1.28	ppbv	22.7	30.00
Chlorobenzene	1.17	ppbv	11.5	30.00
Ethylbenzene	1.09	ppbv	3.9	30.00
m,p-Xylene	2.17	ppbv	4.1	30.00
Bromoform	1.30	ppbv	23.9	30.00
Styrene	1.19	ppbv	12.0	30.00
1,1,1,2-Tetrachloroethane	1.21	ppbv	14.5	30.00
o-Xylene	1.21	ppbv	14.4	30.00
1,3,5-Trimethylbenzene	1.19	ppbv	12.7	30.00
1,2,4-Trimethylbenzene	1.18	ppbv	12.3	30.00
m-Dichlorobenzene	1.09	ppbv	4.2	30.00
p-Dichlorobenzene	1.09	ppbv	4.2	30.00
o-Dichlorobenzene	1.09	ppbv	4.2	30.00
1,2,4-Trichlorobenzene	1.09	ppbv	5.5	30.00
Hexachloro-1,3-butadiene	1.06	ppbv	0.9	30.00

Calibration Check (2605051-CCV2)

Prepared: 05/15/26 Analyzed: 05/16/26

Dichlorodifluoromethane	1.13	ppbv	10.3	30.00
Chloromethane	1.05	ppbv	2.5	30.00
Dichlorotetrafluoroethane	1.01	ppbv	-2.2	30.00
Vinyl chloride	0.991	ppbv	-4.0	30.00
1,3-Butadiene	0.860	ppbv	-16.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV2) Continued

Prepared: 05/15/26 Analyzed: 05/16/26

Ethylene oxide	1.16	ppbv	21.8	30.00	
Bromomethane	1.00	ppbv	-2.8	30.00	
Chloroethane	1.08	ppbv	4.4	30.00	
Acetonitrile	0.958	ppbv	-7.7	30.00	QB-01
Acrolein	1.15	ppbv	12.6	30.00	LK, QB-01
Trichlorofluoromethane	1.40	ppbv	39.0	30.00	LK
Acrylonitrile	1.08	ppbv	-2.2	30.00	
1,1-Dichloroethene	1.18	ppbv	13.3	30.00	LJ
Dichloromethane	0.964	ppbv	-7.0	30.00	
Carbon Disulfide	1.02	ppbv	-6.4	30.00	
Trichlorotrifluoroethane	0.976	ppbv	-5.2	30.00	
trans-1,2-Dichloroethylene	0.916	ppbv	-11.9	30.00	
1,1-Dichloroethane	1.02	ppbv	-1.9	30.00	
Methyl tert-Butyl Ether	1.02	ppbv	-1.3	30.00	
Chloroprene	1.15	ppbv	10.5	30.00	
cis-1,2-Dichloroethylene	0.995	ppbv	-4.0	30.00	
Bromochloromethane	0.929	ppbv	-9.6	30.00	
Chloroform	1.26	ppbv	20.2	30.00	
Ethyl tert-Butyl Ether	0.966	ppbv	-6.8	30.00	
1,2-Dichloroethane	1.28	ppbv	23.8	30.00	
1,1,1-Trichloroethane	1.30	ppbv	25.8	30.00	
Benzene	1.08	ppbv	3.7	30.00	
Carbon Tetrachloride	1.23	ppbv	20.1	30.00	
tert-Amyl Methyl Ether	0.937	ppbv	-9.3	30.00	
1,2-Dichloropropane	0.995	ppbv	-4.1	30.00	
Ethyl Acrylate	1.08	ppbv	3.1	30.00	
Bromodichloromethane	1.14	ppbv	10.5	30.00	
Trichloroethylene	1.04	ppbv	0.6	30.00	
Methyl Methacrylate	1.18	ppbv	13.0	30.00	
cis-1,3-Dichloropropene	1.06	ppbv	1.5	30.00	
Methyl Isobutyl Ketone	1.15	ppbv	10.6	30.00	
trans-1,3-Dichloropropene	1.14	ppbv	8.3	30.00	
1,1,2-Trichloroethane	1.05	ppbv	0.7	30.00	
Toluene	1.19	ppbv	14.9	30.00	
Dibromochloromethane	1.07	ppbv	3.3	30.00	
1,2-Dibromoethane	1.05	ppbv	0.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605051

Calibration Check (2605051-CCV2) Continued

Prepared: 05/15/26 Analyzed: 05/16/26

n-Octane	1.18	ppbv	13.4	30.00
Tetrachloroethylene	1.23	ppbv	17.6	30.00
Chlorobenzene	1.07	ppbv	1.4	30.00
Ethylbenzene	0.994	ppbv	-5.0	30.00
m,p-Xylene	2.02	ppbv	-3.3	30.00
Bromoform	1.30	ppbv	23.9	30.00
Styrene	1.18	ppbv	11.2	30.00
1,1,2,2-Tetrachloroethane	1.23	ppbv	16.8	30.00
o-Xylene	1.20	ppbv	14.0	30.00
1,3,5-Trimethylbenzene	1.15	ppbv	9.4	30.00
1,2,4-Trimethylbenzene	1.15	ppbv	9.1	30.00
m-Dichlorobenzene	0.986	ppbv	-5.4	30.00
p-Dichlorobenzene	0.986	ppbv	-5.7	30.00
o-Dichlorobenzene	0.985	ppbv	-5.8	30.00
1,2,4-Trichlorobenzene	1.05	ppbv	1.1	30.00
Hexachloro-1,3-butadiene	1.05	ppbv	-0.3	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/15/26 09:51

SUBMITTED: 04/15/26 to 05/01/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
LJ	Identification of analyte is acceptable; reported value is an estimate.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.