



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 09, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/01/26 08:51 through 05/27/26 14:43.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 **FAX:** -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK Q9551577	6040205-01	Air	12/03/25 23:59	04/01/26 08:51
TOOK Q9551578	6040205-02	Air	12/09/25 23:59	04/01/26 08:51
TOOK Q9551580	6040205-03	Air	12/15/25 23:59	04/01/26 08:51
TOOK Q9551579	6040205-04	Air	12/21/25 23:59	04/01/26 08:51
TOOK Q9551581	6040205-05	Air	12/27/25 23:59	04/01/26 08:51
TOOK Q9551583 TB	6040205-06	Air	12/27/25 00:00	04/01/26 08:51
TOOK Q9551584 FB	6040205-07	Air	12/27/25 00:00	04/01/26 08:51
TOOK Q9551582	6040205-08	Air	01/02/26 23:59	04/01/26 08:51
TOOK Q9551585	6040205-09	Air	01/08/26 23:59	04/01/26 08:51
TOOK Q9551586	6040205-10	Air	01/14/26 23:59	04/01/26 08:51
TOOK Q9551587	6040205-11	Air	01/21/26 23:59	04/01/26 08:51
TOOK Q9551588	6040205-12	Air	01/26/26 23:59	04/01/26 08:51
TOOK Q9551589 TB	6040205-13	Air	01/26/26 00:00	04/01/26 08:51
TOOK Q9551590 FB	6040205-14	Air	01/26/26 00:00	04/01/26 08:51
TOOK Q9551591	6050623-01	Air	02/01/26 23:59	05/04/26 09:30
TOOK Q9551595	6050623-02	Air	02/25/26 23:59	05/04/26 09:30
TOOK Q9551597	6050623-03	Air	03/15/26 23:59	05/04/26 09:30
TOOK Q9551598	6050623-04	Air	03/21/26 23:59	05/04/26 09:30
TOOK Q9551599	6050623-05	Air	03/27/26 23:59	05/04/26 09:30
TOOK Q9551600	6052727-01	Air	04/02/26 23:59	05/27/26 14:43
TOOK Q9551601	6052727-02	Air	04/08/26 23:59	05/27/26 14:43

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 **FAX:** -

TOOK Q9551602	6052727-03	Air	04/14/26 23:59	05/27/26 14:43
TOOK Q9551603	6052727-04	Air	04/20/26 23:59	05/27/26 14:43
TOOK Q9551605 TB	6052727-06	Air	04/26/26 00:00	05/27/26 14:43
TOOK Q9551604 FB	6052727-07	Air	04/26/26 00:00	05/27/26 14:43

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551577

Lab ID: 6040205-01

Sampled: 12/03/25 23:59

Matrix: Air

Sample Volume: 2246.23 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 02:07

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.597	SL	0.0240
Arsenic	7440-38-2	0.765		0.00614
Beryllium	7440-41-7	0.0154		0.00127
Cadmium	7440-43-9	0.169		0.00355
Chromium	7440-47-3	2.67		1.68
Cobalt	7440-48-4	0.669		0.0337
Lead	7439-92-1	2.96		0.0891
Manganese	7439-96-5	21.0		0.532
Nickel	7440-02-0	1.86		0.531
Selenium	7782-49-2	0.491	LJ, QX	0.00739



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551578

Lab ID: 6040205-02

Sampled: 12/09/25 23:59

Matrix: Air

Sample Volume: 2204.66 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 02:24

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.456	SL	0.0244
Arsenic	7440-38-2	0.795		0.00625
Beryllium	7440-41-7	0.0153		0.00130
Cadmium	7440-43-9	0.168		0.00362
Chromium	7440-47-3	3.11		1.71
Cobalt	7440-48-4	0.330		0.0343
Lead	7439-92-1	3.61		0.0908
Manganese	7439-96-5	24.6		0.542
Nickel	7440-02-0	2.37		0.541
Selenium	7782-49-2	1.09	LJ, QX	0.00753



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551580

Lab ID: 6040205-03

Sampled: 12/15/25 23:59

Matrix: Air

Sample Volume: 2260.79 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 02:40

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.312	SL	0.0238
Arsenic	7440-38-2	0.605		0.00610
Beryllium	7440-41-7	0.0196		0.00127
Cadmium	7440-43-9	0.133		0.00353
Chromium	7440-47-3	6.17		1.67
Cobalt	7440-48-4	0.410		0.0335
Lead	7439-92-1	2.92		0.0885
Manganese	7439-96-5	29.3		0.529
Nickel	7440-02-0	6.53		0.528
Selenium	7782-49-2	0.744	LJ, QX	0.00734



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551579

Lab ID: 6040205-04

Sampled: 12/21/25 23:59

Matrix: Air

Sample Volume: 2221.96 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 02:56

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>	
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>	
Antimony	7440-36-0	0.286	SL	0.0242	
Arsenic	7440-38-2	0.520		0.00620	
Beryllium	7440-41-7	0.0134		0.00129	
Cadmium	7440-43-9	0.159		0.00359	
Chromium	7440-47-3	2.10		1.70	
Cobalt	7440-48-4	0.459		0.0340	
Lead	7439-92-1	2.10		0.0901	
Manganese	7439-96-5	14.1		0.538	
Nickel	7440-02-0	0.868		0.537	
Selenium	7782-49-2	0.311	LJ, QX	0.00747	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551581

Lab ID: 6040205-05

Sampled: 12/27/25 23:59

Matrix: Air

Sample Volume: 2089.43 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 19:38

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.961	SL	0.0258
Arsenic	7440-38-2	0.861		0.00660
Beryllium	7440-41-7	0.0117		0.00137
Cadmium	7440-43-9	0.214		0.00382
Chromium	7440-47-3	2.34		1.81
Cobalt	7440-48-4	0.253		0.0362
Lead	7439-92-1	4.10		0.0958
Manganese	7439-96-5	17.8		0.572
Nickel	7440-02-0	1.09		0.571
Selenium	7782-49-2	0.857		0.00794



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551583 TB	Lab ID: 6040205-06	Sampled: 12/27/25 00:00
Matrix: Air	Sample Volume: 2089.43 m ³	Received: 04/01/26 08:51
	Filter ID:	Analysis Date: 04/24/26 19:57

Comments: Trip Blank

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.0138	U, SL	0.0258
Arsenic	7440-38-2	0.00480	U	0.00660
Beryllium	7440-41-7	8.57E-4	U	0.00137
Cadmium	7440-43-9	0.00146	U	0.00382
Chromium	7440-47-3	1.28	U	1.81
Cobalt	7440-48-4	0.0196	U	0.0362
Lead	7439-92-1	0.0618	U	0.0958
Manganese	7439-96-5	0.225	U	0.572
Nickel	7440-02-0	0.239	U	0.571
Selenium	7782-49-2	0.00285	U	0.00794



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551584 FB	Lab ID: 6040205-07	Sampled: 12/27/25 00:00
Matrix: Air	Sample Volume: 2089.43 m ³	Received: 04/01/26 08:51
	Filter ID:	Analysis Date: 04/24/26 20:16

Comments: Field Blank

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.00735	U, SL	0.0258
Arsenic	7440-38-2	0.00456	U	0.00660
Beryllium	7440-41-7	7.76E-4	U	0.00137
Cadmium	7440-43-9	0.00159	U	0.00382
Chromium	7440-47-3	1.40	U	1.81
Cobalt	7440-48-4	0.0231	U	0.0362
Lead	7439-92-1	0.0570	U	0.0958
Manganese	7439-96-5	0.213	U	0.572
Nickel	7440-02-0	0.359	U	0.571
Selenium	7782-49-2	0.00380	U	0.00794



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551582

Lab ID: 6040205-08

Sampled: 01/02/26 23:59

Matrix: Air

Sample Volume: 2181.67 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 20:29

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.787	SL	0.0247
Arsenic	7440-38-2	1.11		0.00632
Beryllium	7440-41-7	0.0175		0.00131
Cadmium	7440-43-9	0.236		0.00365
Chromium	7440-47-3	2.79		1.73
Cobalt	7440-48-4	1.16		0.0347
Lead	7439-92-1	4.15		0.0917
Manganese	7439-96-5	25.5		0.548
Nickel	7440-02-0	1.59		0.547
Selenium	7782-49-2	0.766		0.00761



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551585

Lab ID: 6040205-09

Sampled: 01/08/26 23:59

Matrix: Air

Sample Volume: 2093.6 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 20:49

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.306	SL	0.0257
Arsenic	7440-38-2	0.441		0.00658
Beryllium	7440-41-7	0.00971		0.00137
Cadmium	7440-43-9	0.0866		0.00381
Chromium	7440-47-3	2.05		1.80
Cobalt	7440-48-4	0.237		0.0361
Lead	7439-92-1	2.01		0.0956
Manganese	7439-96-5	12.8		0.571
Nickel	7440-02-0	1.00		0.570
Selenium	7782-49-2	0.609		0.00793



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551586

Lab ID: 6040205-10

Sampled: 01/14/26 23:59

Matrix: Air

Sample Volume: 2231.14 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 21:07

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.221	SL	0.0241
Arsenic	7440-38-2	0.220		0.00618
Beryllium	7440-41-7	0.0137		0.00128
Cadmium	7440-43-9	0.0640		0.00357
Chromium	7440-47-3	2.07		1.69
Cobalt	7440-48-4	0.184		0.0339
Lead	7439-92-1	1.34		0.0897
Manganese	7439-96-5	15.4		0.536
Nickel	7440-02-0	0.932		0.535
Selenium	7782-49-2	0.169		0.00744



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551587

Lab ID: 6040205-11

Sampled: 01/21/26 23:59

Matrix: Air

Sample Volume: 2316.15 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 21:44

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	2.14	SL	0.0233
Arsenic	7440-38-2	0.805		0.00595
Beryllium	7440-41-7	0.0236		0.00124
Cadmium	7440-43-9	0.183		0.00344
Chromium	7440-47-3	3.64		1.63
Cobalt	7440-48-4	0.506		0.0327
Lead	7439-92-1	3.66		0.0864
Manganese	7439-96-5	36.2		0.516
Nickel	7440-02-0	1.95		0.515
Selenium	7782-49-2	0.426		0.00717



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551588

Lab ID: 6040205-12

Sampled: 01/26/26 23:59

Matrix: Air

Sample Volume: 2583.11 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 22:03

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.349	SL	0.0208
Arsenic	7440-38-2	0.875		0.00533
Beryllium	7440-41-7	0.00507		0.00111
Cadmium	7440-43-9	0.319		0.00309
Chromium	7440-47-3	1.72		1.46
Cobalt	7440-48-4	0.155		0.0293
Lead	7439-92-1	3.51		0.0775
Manganese	7439-96-5	8.84		0.463
Nickel	7440-02-0	0.610		0.462
Selenium	7782-49-2	0.372		0.00642



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551589 TB

Lab ID: 6040205-13

Sampled: 01/26/26 00:00

Matrix: Air

Sample Volume: 2583.11 m³

Received: 04/01/26 08:51

Filter ID:

Analysis Date: 04/24/26 22:21

Comments: Trip Blank

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.00671	U, SL	0.0208
Arsenic	7440-38-2	0.00350	U	0.00533
Beryllium	7440-41-7	6.13E-4	U	0.00111
Cadmium	7440-43-9	0.00139	U	0.00309
Chromium	7440-47-3	1.00	U	1.46
Cobalt	7440-48-4	0.0208	U	0.0293
Lead	7439-92-1	0.0487	U	0.0775
Manganese	7439-96-5	0.168	U	0.463
Nickel	7440-02-0	0.195	U	0.462
Selenium	7782-49-2	0.00313	U	0.00642



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551590 FB	Lab ID: 6040205-14	Sampled: 01/26/26 00:00
Matrix: Air	Sample Volume: 2583.11 m ³	Received: 04/01/26 08:51
	Filter ID:	Analysis Date: 04/24/26 23:26

Comments: Field Blank

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.00572	U, SL	0.0208
Arsenic	7440-38-2	0.00313	U	0.00533
Beryllium	7440-41-7	5.07E-4	U	0.00111
Cadmium	7440-43-9	0.00109	U	0.00309
Chromium	7440-47-3	1.06	U	1.46
Cobalt	7440-48-4	0.0154	U	0.0293
Lead	7439-92-1	0.0480	U	0.0775
Manganese	7439-96-5	0.185	U	0.463
Nickel	7440-02-0	0.206	U	0.462
Selenium	7782-49-2	0.00197	U	0.00642



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551591

Lab ID: 6050623-01

Sampled: 02/01/26 23:59

Matrix: Air

Sample Volume: 2231.11 m³

Received: 05/04/26 09:30

Filter ID:

Analysis Date: 06/30/26 00:47

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.362	SL	0.0241
Arsenic	7440-38-2	0.418		0.00618
Beryllium	7440-41-7	0.00671		0.00128
Cadmium	7440-43-9	0.116		0.00357
Chromium	7440-47-3	1.72		1.69
Cobalt	7440-48-4	0.129		0.0339
Lead	7439-92-1	2.44		0.0897
Manganese	7439-96-5	9.31		0.536
Nickel	7440-02-0	0.613		0.535
Selenium	7782-49-2	0.842		0.00744



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551595

Lab ID: 6050623-02

Sampled: 02/25/26 23:59

Matrix: Air

Sample Volume: 2069.09 m³

Received: 05/04/26 09:30

Filter ID:

Analysis Date: 06/30/26 01:04

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.589	SL	0.0260
Arsenic	7440-38-2	0.760		0.00666
Beryllium	7440-41-7	0.0328		0.00138
Cadmium	7440-43-9	0.166		0.00385
Chromium	7440-47-3	2.64		1.82
Cobalt	7440-48-4	0.439		0.0366
Lead	7439-92-1	5.38		0.0967
Manganese	7439-96-5	38.2		0.578
Nickel	7440-02-0	1.54		0.576
Selenium	7782-49-2	2.00		0.00802



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551597

Lab ID: 6050623-03

Sampled: 03/15/26 23:59

Matrix: Air

Sample Volume: 2150.53 m³

Received: 05/04/26 09:30

Filter ID:

Analysis Date: 06/30/26 01:21

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.219	SL	0.0250
Arsenic	7440-38-2	0.863		0.00641
Beryllium	7440-41-7	0.0471		0.00133
Cadmium	7440-43-9	0.506		0.00371
Chromium	7440-47-3	2.46		1.76
Cobalt	7440-48-4	0.496		0.0352
Lead	7439-92-1	5.43		0.0931
Manganese	7439-96-5	40.0		0.556
Nickel	7440-02-0	1.60		0.555
Selenium	7782-49-2	0.551		0.00772



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551598

Lab ID: 6050623-04

Sampled: 03/21/26 23:59

Matrix: Air

Sample Volume: 1961.8 m³

Received: 05/04/26 09:30

Filter ID:

Analysis Date: 06/30/26 01:58

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.774	SL	0.0275
Arsenic	7440-38-2	1.13		0.00702
Beryllium	7440-41-7	0.0344		0.00146
Cadmium	7440-43-9	0.290		0.00406
Chromium	7440-47-3	3.46		1.92
Cobalt	7440-48-4	0.505		0.0386
Lead	7439-92-1	5.75		0.102
Manganese	7439-96-5	41.7		0.610
Nickel	7440-02-0	1.87		0.608
Selenium	7782-49-2	2.10		0.00846



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551599

Lab ID: 6050623-05

Sampled: 03/27/26 23:59

Matrix: Air

Sample Volume: 2063.2 m³

Received: 05/04/26 09:30

Filter ID:

Analysis Date: 06/30/26 02:17

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.466	SL	0.0261
Arsenic	7440-38-2	0.652		0.00668
Beryllium	7440-41-7	0.0303		0.00139
Cadmium	7440-43-9	0.254		0.00386
Chromium	7440-47-3	2.66		1.83
Cobalt	7440-48-4	0.506		0.0367
Lead	7439-92-1	3.79		0.0970
Manganese	7439-96-5	37.0		0.580
Nickel	7440-02-0	1.83		0.578
Selenium	7782-49-2	0.668		0.00804



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551600

Lab ID: 6052727-01

Sampled: 04/02/26 23:59

Matrix: Air

Sample Volume: 1913.44 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 06/30/26 22:55

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>	
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>	
Antimony	7440-36-0	0.258	SL	0.0281	
Arsenic	7440-38-2	0.443		0.00720	
Beryllium	7440-41-7	0.0132		0.00150	
Cadmium	7440-43-9	0.105		0.00417	
Chromium	7440-47-3	1.89	U	1.97	
Cobalt	7440-48-4	0.155		0.0395	
Lead	7439-92-1	1.63		0.105	
Manganese	7439-96-5	14.1		0.625	
Nickel	7440-02-0	1.08		0.623	
Selenium	7782-49-2	0.737	LJ, QX	0.00867	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551601

Lab ID: 6052727-02

Sampled: 04/08/26 23:59

Matrix: Air

Sample Volume: 2082.77 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 06/30/26 23:09

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.590	SL	0.0259
Arsenic	7440-38-2	0.859		0.00662
Beryllium	7440-41-7	0.0188		0.00137
Cadmium	7440-43-9	0.811		0.00383
Chromium	7440-47-3	2.58		1.81
Cobalt	7440-48-4	0.323		0.0363
Lead	7439-92-1	4.57		0.0961
Manganese	7439-96-5	25.2		0.574
Nickel	7440-02-0	1.65		0.573
Selenium	7782-49-2	2.13	LJ, QX	0.00797



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551602

Lab ID: 6052727-03

Sampled: 04/14/26 23:59

Matrix: Air

Sample Volume: 1964.33 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 06/30/26 23:25

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.228	SL	0.0274
Arsenic	7440-38-2	0.655		0.00702
Beryllium	7440-41-7	0.0301		0.00146
Cadmium	7440-43-9	0.124		0.00406
Chromium	7440-47-3	2.48		1.92
Cobalt	7440-48-4	0.278		0.0385
Lead	7439-92-1	1.70		0.102
Manganese	7439-96-5	19.7		0.609
Nickel	7440-02-0	1.70		0.607
Selenium	7782-49-2	0.773	LJ, QX	0.00845



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551603

Lab ID: 6052727-04

Sampled: 04/20/26 23:59

Matrix: Air

Sample Volume: 2069.19 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 06/30/26 23:39

Comments:

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>		<u>MDL</u>	
		<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>	
Antimony	7440-36-0	0.383	SL	0.0260	
Arsenic	7440-38-2	0.743		0.00666	
Beryllium	7440-41-7	0.0187		0.00138	
Cadmium	7440-43-9	0.155		0.00385	
Chromium	7440-47-3	2.22		1.82	
Cobalt	7440-48-4	0.256		0.0366	
Lead	7439-92-1	2.85		0.0967	
Manganese	7439-96-5	21.2		0.578	
Nickel	7440-02-0	1.39		0.576	
Selenium	7782-49-2	0.432	LJ, QX	0.00802	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551605 TB

Lab ID: 6052727-06

Sampled: 04/26/26 00:00

Matrix: Air

Sample Volume: 2069.19 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 06/30/26 23:54

Comments: Trip Blank

Inorganics by Compendium Method IO-3.5

Results

MDL

<u>Analyte</u>	<u>CAS Number</u>	<u>ng/m³ Air</u>	<u>Flag</u>	<u>ng/m³ Air</u>
Antimony	7440-36-0	0.00738	U, SL	0.0260
Arsenic	7440-38-2	0.00469	U	0.00666
Beryllium	7440-41-7	9.66E-4	U	0.00138
Cadmium	7440-43-9	0.00148	U	0.00385
Chromium	7440-47-3	1.27	U	1.82
Cobalt	7440-48-4	0.0163	U	0.0366
Lead	7439-92-1	0.0545	U	0.0967
Manganese	7439-96-5	0.152	U	0.578
Nickel	7440-02-0	0.207	U	0.576
Selenium	7782-49-2	0.00116	U, LJ, QX	0.00802



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK Q9551604 FB

Lab ID: 6052727-07

Sampled: 04/26/26 00:00

Matrix: Air

Sample Volume: 2069.19 m³

Received: 05/27/26 14:43

Filter ID:

Analysis Date: 07/01/26 00:07

Comments: Field Blank

Inorganics by Compendium Method IO-3.5

<u>Analyte</u>	<u>CAS Number</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
		<u>ng/m³ Air</u>		<u>ng/m³ Air</u>
Antimony	7440-36-0	0.00740	U, SL	0.0260
Arsenic	7440-38-2	0.00351	U	0.00666
Beryllium	7440-41-7	7.30E-4	U	0.00138
Cadmium	7440-43-9	0.00178	U	0.00385
Chromium	7440-47-3	1.27	U	1.82
Cobalt	7440-48-4	0.0173	U	0.0366
Lead	7439-92-1	0.0536	U	0.0967
Manganese	7439-96-5	0.123	U	0.578
Nickel	7440-02-0	0.198	U	0.576
Selenium	7782-49-2	-1.89E-6	U, LJ, QX	0.00802



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604071 - B6D1502

Calibration Blank (2604071-CCB1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	0.404		ng/l							
Arsenic	1.86		ng/l							
Beryllium	-1.33		ng/l							U
Cadmium	0.220		ng/l							
Chromium	3.69		ng/l							
Cobalt	0.551		ng/l							
Lead	2.20		ng/l							
Manganese	6.03		ng/l							
Nickel	0.630		ng/l							
Selenium	16.3		ng/l							LJ, QX

Calibration Blank (2604071-CCB2)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	0.00929		ng/l							
Arsenic	6.10		ng/l							
Beryllium	-1.47		ng/l							U
Cadmium	0.0873		ng/l							
Chromium	3.67		ng/l							
Cobalt	0.526		ng/l							
Lead	1.47		ng/l							
Manganese	4.42		ng/l							
Nickel	-0.111		ng/l							U
Selenium	10.4		ng/l							LJ, QX

Calibration Blank (2604071-CCB3)

Prepared: 04/22/26 Analyzed: 04/24/26

Antimony	0.0149		ng/l							
Arsenic	6.02		ng/l							
Beryllium	-1.77		ng/l							U
Cadmium	0.177		ng/l							
Chromium	4.48		ng/l							
Cobalt	0.838		ng/l							
Lead	2.40		ng/l							
Manganese	6.44		ng/l							
Nickel	1.52		ng/l							
Selenium	13.5		ng/l							LJ, QX

Calibration Blank (2604071-CCB4)

Prepared: 04/22/26 Analyzed: 04/24/26

Antimony	0.688		ng/l							
Arsenic	4.92		ng/l							
Beryllium	-1.68		ng/l							U
Cadmium	0.435		ng/l							
Chromium	7.91		ng/l							

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604071 - B6D1502

Calibration Blank (2604071-CCB4) Contin

Prepared: 04/22/26 Analyzed: 04/24/26

Cobalt	1.27		ng/l							
Lead	4.72		ng/l							
Manganese	12.6		ng/l							
Nickel	5.18		ng/l							
Selenium	16.4		ng/l							LJ, QX

Calibration Check (2604071-CCV1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	18700		ng/l	19996		93.4	90-110			
Arsenic	19100		ng/l	20000		95.3	90-110			
Beryllium	5150		ng/l	5001.5		103	90-110			
Cadmium	18900		ng/l	20014		94.6	90-110			
Chromium	229000		ng/l	240050		95.4	90-110			
Cobalt	48200		ng/l	50020		96.3	90-110			
Lead	190000		ng/l	200020		95.0	90-110			
Manganese	482000		ng/l	500200		96.4	90-110			
Nickel	117000		ng/l	119990		97.1	90-110			
Selenium	19000		ng/l	20002		95.0	90-110			LJ, QX

Calibration Check (2604071-CCV2)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	19200		ng/l	19996		96.0	90-110			
Arsenic	19700		ng/l	20000		98.7	90-110			
Beryllium	4770		ng/l	5001.5		95.4	90-110			
Cadmium	19600		ng/l	20014		98.1	90-110			
Chromium	241000		ng/l	240050		101	90-110			
Cobalt	49700		ng/l	50020		99.3	90-110			
Lead	196000		ng/l	200020		97.8	90-110			
Manganese	503000		ng/l	500200		101	90-110			
Nickel	121000		ng/l	119990		101	90-110			
Selenium	19700		ng/l	20002		98.7	90-110			LJ, QX

Calibration Check (2604071-CCV3)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	19100		ng/l	19996		95.3	90-110			
Arsenic	19500		ng/l	20000		97.7	90-110			
Beryllium	4970		ng/l	5001.5		99.3	90-110			
Cadmium	19600		ng/l	20014		98.0	90-110			
Chromium	245000		ng/l	240050		102	90-110			
Cobalt	49000		ng/l	50020		97.9	90-110			
Lead	195000		ng/l	200020		97.3	90-110			
Manganese	497000		ng/l	500200		99.4	90-110			
Nickel	119000		ng/l	119990		99.2	90-110			
Selenium	19100		ng/l	20002		95.7	90-110			LJ, QX

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604071 - B6D1502

Calibration Check (2604071-CCV4)

Prepared: 04/22/26 Analyzed: 04/24/26

Antimony	19300		ng/l	19996		96.5	90-110			
Arsenic	19700		ng/l	20000		98.4	90-110			
Beryllium	4980		ng/l	5001.5		99.6	90-110			
Cadmium	19800		ng/l	20014		99.0	90-110			
Chromium	249000		ng/l	240050		104	90-110			
Cobalt	49100		ng/l	50020		98.1	90-110			
Lead	197000		ng/l	200020		98.4	90-110			
Manganese	504000		ng/l	500200		101	90-110			
Nickel	119000		ng/l	119990		99.5	90-110			
Selenium	19300		ng/l	20002		96.4	90-110			LJ, QX

High Cal Check (2604071-HCV1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	39100		ng/l	39992		97.8	95-105			
Arsenic	39900		ng/l	40000		99.8	95-105			
Beryllium	10400		ng/l	10003		104	95-105			
Cadmium	39200		ng/l	40028		98.0	95-105			
Chromium	479000		ng/l	480100		99.7	95-105			
Cobalt	99100		ng/l	100040		99.0	95-105			
Lead	397000		ng/l	400040		99.3	95-105			
Manganese	995000		ng/l	1.0004E6		99.4	95-105			
Nickel	237000		ng/l	239980		98.6	95-105			
Selenium	39600		ng/l	40004		99.0	95-105			LJ, QX

Initial Cal Blank (2604071-ICB1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	0.337		ng/l							
Arsenic	0.862		ng/l							
Beryllium	-0.878		ng/l							U
Cadmium	0.206		ng/l							
Chromium	4.76		ng/l							
Cobalt	0.628		ng/l							
Lead	3.73		ng/l							
Manganese	8.34		ng/l							
Nickel	-0.499		ng/l							U
Selenium	9.32		ng/l							LJ, QX

Initial Cal Check (2604071-ICV1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	18700		ng/l	20000		93.4	90-110			
Arsenic	19300		ng/l	20000		96.4	90-110			
Beryllium	5260		ng/l	5000.0		105	90-110			
Cadmium	19300		ng/l	20000		96.3	90-110			
Chromium	236000		ng/l	240000		98.5	90-110			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604071 - B6D1502

Initial Cal Check (2604071-ICV1) Continu

Prepared: 04/22/26 Analyzed: 04/23/26

Cobalt	48200		ng/l	50000		96.4	90-110			
Lead	194000		ng/l	200000		97.0	90-110			
Manganese	477000		ng/l	500000		95.5	90-110			
Nickel	116000		ng/l	120000		97.0	90-110			
Selenium	19700		ng/l	20000		98.5	90-110			LJ, QX

Interference Check A (2604071-IFA1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	0.00		ng/l				80-120			U
Arsenic	0.00		ng/l				80-120			U
Beryllium	0.00		ng/l				80-120			U
Cadmium	0.00		ng/l				80-120			U
Chromium	0.00		ng/l				80-120			U
Cobalt	0.00		ng/l				80-120			U
Lead	0.00		ng/l				80-120			U
Manganese	0.00		ng/l				80-120			U
Nickel	0.00		ng/l				80-120			U
Selenium	0.00		ng/l				80-120			LJ, QX, U

Interference Check B (2604071-IFB1)

Prepared: 04/22/26 Analyzed: 04/23/26

Antimony	20100		ng/l	19996		101	80-120			
Arsenic	20700		ng/l	20000		103	80-120			
Beryllium	4970		ng/l	5001.5		99.4	80-120			
Cadmium	19500		ng/l	20014		97.6	80-120			
Chromium	235000		ng/l	240050		98.1	80-120			
Cobalt	49900		ng/l	50020		99.7	80-120			
Lead	205000		ng/l	200020		103	80-120			
Manganese	517000		ng/l	500200		103	80-120			
Nickel	117000		ng/l	119990		97.4	80-120			
Selenium	19100		ng/l	20002		95.4	80-120			LJ, QX

Batch 2604078 - B6D1601

Calibration Blank (2604078-CCB1)

Prepared & Analyzed: 04/24/26

Antimony	0.132		ng/l							
Arsenic	3.93		ng/l							
Beryllium	0.0424		ng/l							
Cadmium	0.296		ng/l							
Chromium	0.434		ng/l							
Cobalt	0.246		ng/l							
Lead	2.03		ng/l							
Manganese	5.16		ng/l							
Nickel	-21.1		ng/l							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604078 - B6D1601

Calibration Blank (2604078-CCB1) Contin

Prepared & Analyzed: 04/24/26

Selenium	6.38	ng/l
----------	------	------

Calibration Blank (2604078-CCB2)

Prepared & Analyzed: 04/24/26

Antimony	0.484	ng/l	
Arsenic	-0.549	ng/l	U
Beryllium	-0.765	ng/l	U
Cadmium	0.295	ng/l	
Chromium	1.43	ng/l	
Cobalt	0.412	ng/l	
Lead	2.57	ng/l	
Manganese	4.68	ng/l	
Nickel	-20.5	ng/l	U
Selenium	3.99	ng/l	

Calibration Blank (2604078-CCB3)

Prepared & Analyzed: 04/24/26

Antimony	0.756	ng/l	
Arsenic	5.04	ng/l	
Beryllium	-1.01	ng/l	U
Cadmium	0.215	ng/l	
Chromium	0.483	ng/l	
Cobalt	0.0108	ng/l	
Lead	1.69	ng/l	
Manganese	2.79	ng/l	
Nickel	-21.4	ng/l	U
Selenium	2.60	ng/l	

Calibration Blank (2604078-CCB4)

Prepared: 04/24/26 Analyzed: 04/25/26

Antimony	0.148	ng/l	
Arsenic	3.67	ng/l	
Beryllium	-1.70	ng/l	U
Cadmium	0.0894	ng/l	
Chromium	0.862	ng/l	
Cobalt	0.244	ng/l	
Lead	1.63	ng/l	
Manganese	4.17	ng/l	
Nickel	-20.3	ng/l	U
Selenium	4.74	ng/l	

Calibration Check (2604078-CCV1)

Prepared & Analyzed: 04/24/26

Antimony	19400	ng/l	19996	97.2	90-110
Arsenic	19500	ng/l	20000	97.4	90-110

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604078 - B6D1601

Calibration Check (2604078-CCV1) Contin

Prepared & Analyzed: 04/24/26

Beryllium	4740	ng/l	5001.5	94.8	90-110
Cadmium	19600	ng/l	20014	97.8	90-110
Chromium	242000	ng/l	240050	101	90-110
Cobalt	49600	ng/l	50020	99.1	90-110
Lead	194000	ng/l	200020	97.2	90-110
Manganese	489000	ng/l	500200	97.8	90-110
Nickel	119000	ng/l	119990	98.9	90-110
Selenium	19400	ng/l	20002	97.2	90-110

Calibration Check (2604078-CCV2)

Prepared & Analyzed: 04/24/26

Antimony	20400	ng/l	19996	102	90-110
Arsenic	20200	ng/l	20000	101	90-110
Beryllium	4800	ng/l	5001.5	96.0	90-110
Cadmium	20400	ng/l	20014	102	90-110
Chromium	254000	ng/l	240050	106	90-110
Cobalt	50800	ng/l	50020	102	90-110
Lead	203000	ng/l	200020	101	90-110
Manganese	513000	ng/l	500200	103	90-110
Nickel	122000	ng/l	119990	101	90-110
Selenium	20500	ng/l	20002	102	90-110

Calibration Check (2604078-CCV3)

Prepared & Analyzed: 04/24/26

Antimony	20400	ng/l	19996	102	90-110
Arsenic	20300	ng/l	20000	102	90-110
Beryllium	4610	ng/l	5001.5	92.2	90-110
Cadmium	20500	ng/l	20014	103	90-110
Chromium	254000	ng/l	240050	106	90-110
Cobalt	50900	ng/l	50020	102	90-110
Lead	203000	ng/l	200020	102	90-110
Manganese	526000	ng/l	500200	105	90-110
Nickel	123000	ng/l	119990	102	90-110
Selenium	20100	ng/l	20002	100	90-110

Calibration Check (2604078-CCV4)

Prepared: 04/24/26 Analyzed: 04/25/26

Antimony	20300	ng/l	19996	101	90-110
Arsenic	20300	ng/l	20000	102	90-110
Beryllium	4730	ng/l	5001.5	94.5	90-110
Cadmium	20700	ng/l	20014	103	90-110
Chromium	254000	ng/l	240050	106	90-110
Cobalt	51100	ng/l	50020	102	90-110
Lead	203000	ng/l	200020	102	90-110

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604078 - B6D1601

Calibration Check (2604078-CCV4) Contin

Prepared: 04/24/26 Analyzed: 04/25/26

Manganese	532000		ng/l	500200		106	90-110			
Nickel	123000		ng/l	119990		103	90-110			
Selenium	20300		ng/l	20002		102	90-110			

High Cal Check (2604078-HCV1)

Prepared & Analyzed: 04/24/26

Antimony	40500		ng/l	39992		101	95-105			
Arsenic	40600		ng/l	40000		101	95-105			
Beryllium	9780		ng/l	10003		97.7	95-105			
Cadmium	40500		ng/l	40028		101	95-105			
Chromium	483000		ng/l	480100		101	95-105			
Cobalt	101000		ng/l	100040		101	95-105			
Lead	405000		ng/l	400040		101	95-105			
Manganese	1.02E6		ng/l	1.0004E6		102	95-105			
Nickel	240000		ng/l	239980		99.8	95-105			
Selenium	40400		ng/l	40004		101	95-105			

Initial Cal Blank (2604078-ICB1)

Prepared & Analyzed: 04/24/26

Antimony	0.201		ng/l							
Arsenic	-2.00		ng/l							U
Beryllium	-0.364		ng/l							U
Cadmium	0.207		ng/l							
Chromium	3.25		ng/l							
Cobalt	0.216		ng/l							
Lead	1.74		ng/l							
Manganese	5.62		ng/l							
Nickel	-17.2		ng/l							U
Selenium	3.58		ng/l							

Initial Cal Check (2604078-ICV1)

Prepared & Analyzed: 04/24/26

Antimony	19900		ng/l	20000		99.5	90-110			
Arsenic	20100		ng/l	20000		100	90-110			
Beryllium	4800		ng/l	5000.0		95.9	90-110			
Cadmium	20400		ng/l	20000		102	90-110			
Chromium	251000		ng/l	240000		105	90-110			
Cobalt	50900		ng/l	50000		102	90-110			
Lead	203000		ng/l	200000		101	90-110			
Manganese	502000		ng/l	500000		100	90-110			
Nickel	122000		ng/l	120000		102	90-110			
Selenium	20500		ng/l	20000		102	90-110			

Interference Check A (2604078-IFA1)

Prepared & Analyzed: 04/24/26

Antimony	0.00		ng/l				80-120			U
----------	------	--	------	--	--	--	--------	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2604078 - B6D1601

Interference Check A (2604078-IFA1) Co

Prepared & Analyzed: 04/24/26

Arsenic	0.00		ng/l				80-120			U
Beryllium	0.00		ng/l				80-120			U
Cadmium	0.00		ng/l				80-120			U
Chromium	0.00		ng/l				80-120			U
Cobalt	0.00		ng/l				80-120			U
Lead	0.00		ng/l				80-120			U
Manganese	0.00		ng/l				80-120			U
Nickel	0.00		ng/l				80-120			U
Selenium	0.00		ng/l				80-120			U

Interference Check B (2604078-IFB1)

Prepared & Analyzed: 04/24/26

Antimony	21100		ng/l	19996		106	80-120			
Arsenic	21100		ng/l	20000		105	80-120			
Beryllium	4670		ng/l	5001.5		93.3	80-120			
Cadmium	20300		ng/l	20014		102	80-120			
Chromium	242000		ng/l	240050		101	80-120			
Cobalt	51600		ng/l	50020		103	80-120			
Lead	213000		ng/l	200020		107	80-120			
Manganese	534000		ng/l	500200		107	80-120			
Nickel	120000		ng/l	119990		100	80-120			
Selenium	19800		ng/l	20002		98.8	80-120			

Batch 2606112 - B6F0901

Calibration Blank (2606112-CCB1)

Prepared & Analyzed: 06/29/26

Antimony	2.07		ng/l							
Arsenic	7.81		ng/l							
Beryllium	-0.576		ng/l							U
Cadmium	0.00416		ng/l							
Chromium	3.15		ng/l							
Cobalt	0.387		ng/l							
Lead	2.05		ng/l							
Manganese	3.59		ng/l							
Nickel	-3.20		ng/l							U
Selenium	2.66		ng/l							

Calibration Blank (2606112-CCB2)

Prepared & Analyzed: 06/29/26

Antimony	1.08		ng/l							
Arsenic	6.86		ng/l							
Beryllium	-0.566		ng/l							U
Cadmium	0.285		ng/l							
Chromium	5.87		ng/l							

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606112 - B6F0901

Calibration Blank (2606112-CCB2) Contin

Prepared & Analyzed: 06/29/26

Cobalt	0.436		ng/l							
Lead	4.33		ng/l							
Manganese	8.89		ng/l							
Nickel	-4.44		ng/l							U
Selenium	2.55		ng/l							

Calibration Blank (2606112-CCB3)

Prepared & Analyzed: 06/29/26

Antimony	0.945		ng/l							
Arsenic	5.94		ng/l							
Beryllium	-1.49		ng/l							U
Cadmium	0.174		ng/l							
Chromium	6.43		ng/l							
Cobalt	0.653		ng/l							
Lead	4.51		ng/l							
Manganese	12.0		ng/l							
Nickel	-2.78		ng/l							U
Selenium	-8.25		ng/l							U

Calibration Blank (2606112-CCB4)

Prepared: 06/29/26 Analyzed: 06/30/26

Antimony	0.593		ng/l							
Arsenic	8.46		ng/l							
Beryllium	-1.13		ng/l							U
Cadmium	0.207		ng/l							
Chromium	7.43		ng/l							
Cobalt	0.622		ng/l							
Lead	4.70		ng/l							
Manganese	14.5		ng/l							
Nickel	-3.08		ng/l							U
Selenium	3.54		ng/l							

Calibration Check (2606112-CCV1)

Prepared & Analyzed: 06/29/26

Antimony	20500	ng/l	19996	102	90-110
Arsenic	19900	ng/l	20000	99.5	90-110
Beryllium	5080	ng/l	5001.5	102	90-110
Cadmium	20700	ng/l	20014	103	90-110
Chromium	252000	ng/l	240050	105	90-110
Cobalt	51100	ng/l	50020	102	90-110
Lead	202000	ng/l	200020	101	90-110
Manganese	513000	ng/l	500200	103	90-110
Nickel	123000	ng/l	119990	102	90-110
Selenium	19900	ng/l	20002	99.4	90-110

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606112 - B6F0901

Calibration Check (2606112-CCV2)

Prepared & Analyzed: 06/29/26

Antimony	20800		ng/l	19996		104	90-110			
Arsenic	19900		ng/l	20000		99.7	90-110			
Beryllium	5190		ng/l	5001.5		104	90-110			
Cadmium	20400		ng/l	20014		102	90-110			
Chromium	241000		ng/l	240050		100	90-110			
Cobalt	48700		ng/l	50020		97.5	90-110			
Lead	201000		ng/l	200020		100	90-110			
Manganese	498000		ng/l	500200		99.5	90-110			
Nickel	117000		ng/l	119990		97.6	90-110			
Selenium	19900		ng/l	20002		99.4	90-110			

Calibration Check (2606112-CCV3)

Prepared & Analyzed: 06/29/26

Antimony	20900		ng/l	19996		105	90-110			
Arsenic	20000		ng/l	20000		99.9	90-110			
Beryllium	5040		ng/l	5001.5		101	90-110			
Cadmium	20800		ng/l	20014		104	90-110			
Chromium	248000		ng/l	240050		103	90-110			
Cobalt	49500		ng/l	50020		99.0	90-110			
Lead	202000		ng/l	200020		101	90-110			
Manganese	507000		ng/l	500200		101	90-110			
Nickel	121000		ng/l	119990		100	90-110			
Selenium	20000		ng/l	20002		99.9	90-110			

Calibration Check (2606112-CCV4)

Prepared: 06/29/26 Analyzed: 06/30/26

Antimony	20600		ng/l	19996		103	90-110			
Arsenic	20100		ng/l	20000		100	90-110			
Beryllium	5020		ng/l	5001.5		100	90-110			
Cadmium	20500		ng/l	20014		102	90-110			
Chromium	244000		ng/l	240050		102	90-110			
Cobalt	48600		ng/l	50020		97.2	90-110			
Lead	201000		ng/l	200020		101	90-110			
Manganese	500000		ng/l	500200		99.9	90-110			
Nickel	119000		ng/l	119990		99.0	90-110			
Selenium	19600		ng/l	20002		97.8	90-110			

High Cal Check (2606112-HCV1)

Prepared & Analyzed: 06/29/26

Antimony	40100		ng/l	39992		100	95-105			
Arsenic	40300		ng/l	40000		101	95-105			
Beryllium	9860		ng/l	10003		98.6	95-105			
Cadmium	40000		ng/l	40028		100	95-105			
Chromium	483000		ng/l	480100		101	95-105			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606112 - B6F0901

High Cal Check (2606112-HCV1) Continue

Prepared & Analyzed: 06/29/26

Cobalt	99300		ng/l	100040		99.3	95-105			
Lead	404000		ng/l	400040		101	95-105			
Manganese	1.01E6		ng/l	1.0004E6		101	95-105			
Nickel	239000		ng/l	239980		99.7	95-105			
Selenium	40100		ng/l	40004		100	95-105			

Initial Cal Blank (2606112-ICB1)

Prepared & Analyzed: 06/29/26

Antimony	2.34		ng/l							
Arsenic	-2.22		ng/l							U
Beryllium	-0.515		ng/l							U
Cadmium	0.120		ng/l							
Chromium	5.40		ng/l							
Cobalt	0.299		ng/l							
Lead	3.21		ng/l							
Manganese	4.38		ng/l							
Nickel	-9.22		ng/l							U
Selenium	-5.59		ng/l							U

Initial Cal Check (2606112-ICV1)

Prepared & Analyzed: 06/29/26

Antimony	19500		ng/l	20000		97.6	90-110			
Arsenic	19800		ng/l	20000		99.1	90-110			
Beryllium	4820		ng/l	5000.0		96.5	90-110			
Cadmium	19900		ng/l	20000		99.7	90-110			
Chromium	240000		ng/l	240000		100	90-110			
Cobalt	50100		ng/l	50000		100	90-110			
Lead	199000		ng/l	200000		99.7	90-110			
Manganese	511000		ng/l	500000		102	90-110			
Nickel	120000		ng/l	120000		100	90-110			
Selenium	20600		ng/l	20000		103	90-110			

Interference Check A (2606112-IFA1)

Prepared & Analyzed: 06/29/26

Antimony	0.00		ng/l				80-120			U
Arsenic	0.00		ng/l				80-120			U
Beryllium	0.00		ng/l				80-120			U
Cadmium	0.00		ng/l				80-120			U
Chromium	0.00		ng/l				80-120			U
Cobalt	0.00		ng/l				80-120			U
Lead	0.00		ng/l				80-120			U
Manganese	0.00		ng/l				80-120			U
Nickel	0.00		ng/l				80-120			U
Selenium	0.00		ng/l				80-120			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606112 - B6F0901

Interference Check B (2606112-IFB1)

Prepared & Analyzed: 06/29/26

Antimony	20900		ng/l	19996		105	80-120			
Arsenic	20700		ng/l	20000		103	80-120			
Beryllium	5000		ng/l	5001.5		100	80-120			
Cadmium	20400		ng/l	20014		102	80-120			
Chromium	242000		ng/l	240050		101	80-120			
Cobalt	50600		ng/l	50020		101	80-120			
Lead	211000		ng/l	200020		105	80-120			
Manganese	509000		ng/l	500200		102	80-120			
Nickel	119000		ng/l	119990		99.4	80-120			
Selenium	19200		ng/l	20002		96.1	80-120			

Batch 2606118 - B6F2501

Calibration Blank (2606118-CCB1)

Prepared & Analyzed: 06/30/26

Antimony	2.30		ng/l							
Arsenic	7.46		ng/l							
Beryllium	-0.268		ng/l							U
Cadmium	-0.0188		ng/l							U
Chromium	3.74		ng/l							
Cobalt	0.305		ng/l							
Lead	2.64		ng/l							
Manganese	8.37		ng/l							
Nickel	-0.401		ng/l							U
Selenium	-0.193		ng/l							LJ, QX, U

Calibration Blank (2606118-CCB2)

Prepared & Analyzed: 06/30/26

Antimony	0.733		ng/l							
Arsenic	5.12		ng/l							
Beryllium	-0.344		ng/l							U
Cadmium	0.0214		ng/l							
Chromium	5.09		ng/l							
Cobalt	0.0595		ng/l							
Lead	2.75		ng/l							
Manganese	10.2		ng/l							
Nickel	2.54		ng/l							
Selenium	4.11		ng/l							LJ, QX

Calibration Blank (2606118-CCB3)

Prepared & Analyzed: 06/30/26

Antimony	1.20		ng/l							
Arsenic	5.91		ng/l							
Beryllium	-0.827		ng/l							U
Cadmium	-0.0671		ng/l							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606118 - B6F2501

Calibration Blank (2606118-CCB3) Contin

Prepared & Analyzed: 06/30/26

Chromium	4.24		ng/l							
Cobalt	0.00955		ng/l							
Lead	2.43		ng/l							
Manganese	7.30		ng/l							
Nickel	-1.71		ng/l							U
Selenium	-0.400		ng/l							LJ, QX, U

Calibration Blank (2606118-CCB4)

Prepared: 06/30/26 Analyzed: 07/01/26

Antimony	0.748		ng/l							
Arsenic	8.76		ng/l							
Beryllium	-1.22		ng/l							U
Cadmium	0.110		ng/l							
Chromium	5.53		ng/l							
Cobalt	0.290		ng/l							
Lead	3.23		ng/l							
Manganese	9.07		ng/l							
Nickel	1.21		ng/l							
Selenium	4.58		ng/l							LJ, QX

Calibration Check (2606118-CCV1)

Prepared & Analyzed: 06/30/26

Antimony	20000		ng/l	19996		100	90-110			
Arsenic	20100		ng/l	20000		100	90-110			
Beryllium	5030		ng/l	5001.5		101	90-110			
Cadmium	20100		ng/l	20014		100	90-110			
Chromium	247000		ng/l	240050		103	90-110			
Cobalt	50800		ng/l	50020		102	90-110			
Lead	200000		ng/l	200020		99.9	90-110			
Manganese	506000		ng/l	500200		101	90-110			
Nickel	122000		ng/l	119990		102	90-110			
Selenium	20000		ng/l	20002		100	90-110			LJ, QX

Calibration Check (2606118-CCV2)

Prepared & Analyzed: 06/30/26

Antimony	20000		ng/l	19996		100	90-110			
Arsenic	20000		ng/l	20000		100	90-110			
Beryllium	5010		ng/l	5001.5		100	90-110			
Cadmium	20200		ng/l	20014		101	90-110			
Chromium	249000		ng/l	240050		104	90-110			
Cobalt	50200		ng/l	50020		100	90-110			
Lead	200000		ng/l	200020		100	90-110			
Manganese	514000		ng/l	500200		103	90-110			
Nickel	122000		ng/l	119990		102	90-110			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606118 - B6F2501

Calibration Check (2606118-CCV2) Contin

Prepared & Analyzed: 06/30/26

Selenium	19800		ng/l	20002		99.2	90-110			LJ, QX
----------	-------	--	------	-------	--	------	--------	--	--	--------

Calibration Check (2606118-CCV3)

Prepared & Analyzed: 06/30/26

Antimony	20200		ng/l	19996		101	90-110			
Arsenic	20200		ng/l	20000		101	90-110			
Beryllium	4910		ng/l	5001.5		98.2	90-110			
Cadmium	20400		ng/l	20014		102	90-110			
Chromium	250000		ng/l	240050		104	90-110			
Cobalt	50400		ng/l	50020		101	90-110			
Lead	201000		ng/l	200020		101	90-110			
Manganese	522000		ng/l	500200		104	90-110			
Nickel	123000		ng/l	119990		103	90-110			
Selenium	20100		ng/l	20002		100	90-110			LJ, QX

Calibration Check (2606118-CCV4)

Prepared: 06/30/26 Analyzed: 07/01/26

Antimony	20100		ng/l	19996		101	90-110			
Arsenic	20000		ng/l	20000		100	90-110			
Beryllium	4850		ng/l	5001.5		96.9	90-110			
Cadmium	20400		ng/l	20014		102	90-110			
Chromium	249000		ng/l	240050		104	90-110			
Cobalt	49600		ng/l	50020		99.1	90-110			
Lead	202000		ng/l	200020		101	90-110			
Manganese	513000		ng/l	500200		103	90-110			
Nickel	121000		ng/l	119990		101	90-110			
Selenium	20100		ng/l	20002		100	90-110			LJ, QX

High Cal Check (2606118-HCV1)

Prepared & Analyzed: 06/30/26

Antimony	38900		ng/l	39992		97.3	95-105			
Arsenic	39300		ng/l	40000		98.1	95-105			
Beryllium	9860		ng/l	10003		98.6	95-105			
Cadmium	38500		ng/l	40028		96.1	95-105			
Chromium	464000		ng/l	480100		96.7	95-105			
Cobalt	97600		ng/l	100040		97.6	95-105			
Lead	391000		ng/l	400040		97.7	95-105			
Manganese	968000		ng/l	1.0004E6		96.8	95-105			
Nickel	233000		ng/l	239980		97.0	95-105			
Selenium	38800		ng/l	40004		96.9	95-105			LJ, QX

Initial Cal Blank (2606118-ICB1)

Prepared & Analyzed: 06/30/26

Antimony	4.62		ng/l							
Arsenic	4.88		ng/l							
Beryllium	-0.0422		ng/l							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606118 - B6F2501

Initial Cal Blank (2606118-ICB1) Continuu

Prepared & Analyzed: 06/30/26

Cadmium	0.0842		ng/l							
Chromium	8.89		ng/l							
Cobalt	0.489		ng/l							
Lead	8.93		ng/l							
Manganese	13.5		ng/l							
Nickel	-1.41		ng/l							U
Selenium	-12.7		ng/l							LJ, QX, U

Initial Cal Check (2606118-ICV1)

Prepared & Analyzed: 06/30/26

Antimony	19700		ng/l	20000		98.5	90-110			
Arsenic	19700		ng/l	20000		98.4	90-110			
Beryllium	4600		ng/l	5000.0		92.1	90-110			
Cadmium	19900		ng/l	20000		99.4	90-110			
Chromium	241000		ng/l	240000		100	90-110			
Cobalt	50200		ng/l	50000		100	90-110			
Lead	198000		ng/l	200000		99.2	90-110			
Manganese	512000		ng/l	500000		102	90-110			
Nickel	121000		ng/l	120000		100	90-110			
Selenium	20100		ng/l	20000		101	90-110			LJ, QX

Interference Check A (2606118-IFA1)

Prepared & Analyzed: 06/30/26

Antimony	0.00		ng/l				80-120			U
Arsenic	0.00		ng/l				80-120			U
Beryllium	0.00		ng/l				80-120			U
Cadmium	0.00		ng/l				80-120			U
Chromium	0.00		ng/l				80-120			U
Cobalt	0.00		ng/l				80-120			U
Lead	0.00		ng/l				80-120			U
Manganese	0.00		ng/l				80-120			U
Nickel	0.00		ng/l				80-120			U
Selenium	0.00		ng/l				80-120			LJ, QX, U

Interference Check B (2606118-IFB1)

Prepared & Analyzed: 06/30/26

Antimony	20700		ng/l	19996		103	80-120			
Arsenic	20800		ng/l	20000		104	80-120			
Beryllium	4930		ng/l	5001.5		98.5	80-120			
Cadmium	20000		ng/l	20014		99.7	80-120			
Chromium	241000		ng/l	240050		101	80-120			
Cobalt	50800		ng/l	50020		102	80-120			
Lead	209000		ng/l	200020		104	80-120			
Manganese	508000		ng/l	500200		101	80-120			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch 2606118 - B6F2501

Interference Check B (2606118-IFB1) Co

Prepared & Analyzed: 06/30/26

Nickel	119000		ng/l	119990		99.3	80-120			
Selenium	19500		ng/l	20002		97.5	80-120			LJ, QX

Batch B6D1502 - ICP-MS Extraction

Blank (B6D1502-BLK1)

Prepared: 04/15/26 Analyzed: 04/23/26

Antimony	5.53E-4	0.0331	ng/m ³ Air							U, SL
Arsenic	0.00152	0.00847	ng/m ³ Air							U
Beryllium	-3.56E-4	0.00176	ng/m ³ Air							U
Cadmium	1.62E-4	0.00490	ng/m ³ Air							U
Chromium	0.0281	2.32	ng/m ³ Air							U
Cobalt	6.04E-4	0.465	ng/m ³ Air							U
Lead	0.00178	0.123	ng/m ³ Air							U
Manganese	0.0234	0.735	ng/m ³ Air							U
Nickel	0.00480	0.733	ng/m ³ Air							U
Selenium	0.00300	0.0102	ng/m ³ Air							LJ, QX, U

LCS (B6D1502-BS1)

Prepared: 04/15/26 Analyzed: 04/23/26

Antimony	0.919	0.0331	ng/m ³ Air	1.3826		66.5	80-120			SL
Arsenic	2.72	0.00847	ng/m ³ Air	2.7658		98.3	80-120			
Beryllium	1.37	0.00176	ng/m ³ Air	1.3833		99.0	80-120			
Cadmium	1.38	0.00490	ng/m ³ Air	1.3839		99.8	80-120			
Chromium	15.9	2.32	ng/m ³ Air	13.832		115	80-120			
Cobalt	1.38	0.465	ng/m ³ Air	1.3835		99.7	80-120			
Lead	13.8	0.123	ng/m ³ Air	13.831		99.9	80-120			
Manganese	9.00	0.735	ng/m ³ Air	8.3008		108	80-120			
Nickel	3.20	0.733	ng/m ³ Air	2.7656		116	80-120			
Selenium	2.76	0.0102	ng/m ³ Air	2.7661		99.6	80-120			LJ, QX

LCS (B6D1502-BS2)

Prepared: 04/15/26 Analyzed: 04/23/26

Antimony	1.33	0.0331	ng/m ³ Air	1.3826		95.9	80-120			SL
Arsenic	2.71	0.00847	ng/m ³ Air	2.7658		97.9	80-120			
Beryllium	1.36	0.00176	ng/m ³ Air	1.3833		98.0	80-120			
Cadmium	1.37	0.00490	ng/m ³ Air	1.3839		98.9	80-120			
Chromium	14.6	2.32	ng/m ³ Air	13.832		105	80-120			
Cobalt	1.36	0.465	ng/m ³ Air	1.3835		98.1	80-120			
Lead	13.6	0.123	ng/m ³ Air	13.831		98.1	80-120			
Manganese	8.55	0.735	ng/m ³ Air	8.3008		103	80-120			
Nickel	2.82	0.733	ng/m ³ Air	2.7656		102	80-120			
Selenium	2.69	0.0102	ng/m ³ Air	2.7661		97.1	80-120			LJ, QX

Batch B6D1601 - ICP-MS Extraction

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch B6D1601 - ICP-MS Extraction

Blank (B6D1601-BLK1)

Prepared: 04/16/26 Analyzed: 04/24/26

Antimony	0.00211	0.0331	ng/m ³ Air							SL, U
Arsenic	1.53E-4	0.00847	ng/m ³ Air							U
Beryllium	-1.51E-4	0.00176	ng/m ³ Air							U
Cadmium	1.02E-4	0.00490	ng/m ³ Air							U
Chromium	0.0175	2.32	ng/m ³ Air							U
Cobalt	2.78E-4	0.465	ng/m ³ Air							U
Lead	0.00263	0.123	ng/m ³ Air							U
Manganese	0.00913	0.735	ng/m ³ Air							U
Nickel	-0.00135	0.733	ng/m ³ Air							U
Selenium	0.00293	0.0102	ng/m ³ Air							U

LCS (B6D1601-BS1)

Prepared: 04/16/26 Analyzed: 04/24/26

Antimony	0.981	0.0331	ng/m ³ Air	1.3826		70.9	80-120			SL
Arsenic	2.73	0.00847	ng/m ³ Air	2.7658		98.8	80-120			
Beryllium	1.29	0.00176	ng/m ³ Air	1.3833		93.3	80-120			
Cadmium	1.39	0.00490	ng/m ³ Air	1.3839		101	80-120			
Chromium	15.3	2.32	ng/m ³ Air	13.832		111	80-120			
Cobalt	1.39	0.465	ng/m ³ Air	1.3835		100	80-120			
Lead	13.9	0.123	ng/m ³ Air	13.831		100	80-120			
Manganese	9.25	0.735	ng/m ³ Air	8.3008		111	80-120			
Nickel	3.02	0.733	ng/m ³ Air	2.7656		109	80-120			
Selenium	2.78	0.0102	ng/m ³ Air	2.7661		101	80-120			

LCS (B6D1601-BS2)

Prepared: 04/16/26 Analyzed: 04/24/26

Antimony	1.38	0.0331	ng/m ³ Air	1.3826		99.6	80-120			SL
Arsenic	2.74	0.00847	ng/m ³ Air	2.7658		99.0	80-120			
Beryllium	1.28	0.00176	ng/m ³ Air	1.3833		92.8	80-120			
Cadmium	1.40	0.00490	ng/m ³ Air	1.3839		101	80-120			
Chromium	14.7	2.32	ng/m ³ Air	13.832		107	80-120			
Cobalt	1.39	0.465	ng/m ³ Air	1.3835		101	80-120			
Lead	14.0	0.123	ng/m ³ Air	13.831		101	80-120			
Manganese	8.80	0.735	ng/m ³ Air	8.3008		106	80-120			
Nickel	2.86	0.733	ng/m ³ Air	2.7656		103	80-120			
Selenium	2.73	0.0102	ng/m ³ Air	2.7661		98.8	80-120			

Duplicate (B6D1601-DUP3)

Source: 6040205-10

Prepared: 04/16/26 Analyzed: 04/24/26

Antimony	0.219	0.0241	ng/m ³ Air	0.221		0.535	10	SL
Arsenic	0.215	0.00618	ng/m ³ Air	0.220		2.06	10	
Beryllium	0.0138	0.00128	ng/m ³ Air	0.0137		0.574	10	
Cadmium	0.0632	0.00357	ng/m ³ Air	0.0640		1.26	10	
Chromium	2.05	1.69	ng/m ³ Air	2.07		1.07	10	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch B6D1601 - ICP-MS Extraction

Duplicate (B6D1601-DUP3) Continued Source: 6040205-10 Prepared: 04/16/26 Analyzed: 04/24/26

Cobalt	0.182	0.339	ng/m ³ Air	0.184	0.903	10
Lead	1.32	0.0897	ng/m ³ Air	1.34	0.984	10
Manganese	15.2	0.536	ng/m ³ Air	15.4	1.15	10
Nickel	0.917	0.535	ng/m ³ Air	0.932	1.63	10
Selenium	0.176	0.00744	ng/m ³ Air	0.169	4.05	10

Batch B6F0901 - ICP-MS Extraction

Blank (B6F0901-BLK1) Prepared: 06/09/26 Analyzed: 06/29/26

Antimony	7.44E-4	0.0331	ng/m ³ Air			SL, U
Arsenic	0.00257	0.00847	ng/m ³ Air			U
Beryllium	-3.70E-5	0.00176	ng/m ³ Air			U
Cadmium	7.87E-5	0.00490	ng/m ³ Air			U
Chromium	0.140	2.32	ng/m ³ Air			U
Cobalt	5.53E-4	0.465	ng/m ³ Air			U
Lead	0.00172	0.123	ng/m ³ Air			U
Manganese	0.0266	0.735	ng/m ³ Air			U
Nickel	-0.00165	0.733	ng/m ³ Air			U
Selenium	7.93E-4	0.0102	ng/m ³ Air			U

LCS (B6F0901-BS1) Prepared: 06/09/26 Analyzed: 06/29/26

Antimony	0.988	0.0331	ng/m ³ Air	1.3826	71.5	80-120	SL
Arsenic	2.75	0.00847	ng/m ³ Air	2.7658	99.3	80-120	
Beryllium	1.51	0.00176	ng/m ³ Air	1.3833	109	80-120	
Cadmium	1.45	0.00490	ng/m ³ Air	1.3839	105	80-120	
Chromium	14.9	2.32	ng/m ³ Air	13.832	108	80-120	
Cobalt	1.40	0.465	ng/m ³ Air	1.3835	101	80-120	
Lead	13.9	0.123	ng/m ³ Air	13.831	101	80-120	
Manganese	8.87	0.735	ng/m ³ Air	8.3008	107	80-120	
Nickel	3.03	0.733	ng/m ³ Air	2.7656	110	80-120	
Selenium	2.75	0.0102	ng/m ³ Air	2.7661	99.3	80-120	

LCS (B6F0901-BS2) Prepared: 06/09/26 Analyzed: 06/29/26

Antimony	1.43	0.0331	ng/m ³ Air	1.3826	103	80-120	SL
Arsenic	2.73	0.00847	ng/m ³ Air	2.7658	98.7	80-120	
Beryllium	1.39	0.00176	ng/m ³ Air	1.3833	101	80-120	
Cadmium	1.43	0.00490	ng/m ³ Air	1.3839	104	80-120	
Chromium	14.2	2.32	ng/m ³ Air	13.832	102	80-120	
Cobalt	1.38	0.465	ng/m ³ Air	1.3835	99.6	80-120	
Lead	13.6	0.123	ng/m ³ Air	13.831	98.2	80-120	
Manganese	8.48	0.735	ng/m ³ Air	8.3008	102	80-120	
Nickel	2.80	0.733	ng/m ³ Air	2.7656	101	80-120	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch B6F0901 - ICP-MS Extraction

LCS (B6F0901-BS2) Continued

Prepared: 06/09/26 Analyzed: 06/29/26

Selenium	2.72	0.0102	ng/m ³ Air	2.7661	98.3	80-120			
Duplicate (B6F0901-DUP3)				Source: 6050623-03		Prepared: 06/09/26 Analyzed: 06/30/26			
Antimony	0.218	0.0250	ng/m ³ Air	0.219		0.425	10	SL	
Arsenic	0.869	0.00641	ng/m ³ Air	0.863		0.660	10		
Beryllium	0.0484	0.00133	ng/m ³ Air	0.0471		2.80	10		
Cadmium	0.512	0.00371	ng/m ³ Air	0.506		1.10	10		
Chromium	2.45	1.76	ng/m ³ Air	2.46		0.372	10		
Cobalt	0.498	0.352	ng/m ³ Air	0.496		0.320	10		
Lead	5.43	0.0931	ng/m ³ Air	5.43		0.0630	10		
Manganese	40.1	0.556	ng/m ³ Air	40.0		0.351	10		
Nickel	1.61	0.555	ng/m ³ Air	1.60		0.787	10		
Selenium	0.550	0.00772	ng/m ³ Air	0.551		0.208	10		

Batch B6F2501 - ICP-MS Extraction

Blank (B6F2501-BLK1)

Prepared: 06/25/26 Analyzed: 06/30/26

Antimony	4.09E-4	0.0331	ng/m ³ Air					U, SL	
Arsenic	0.00300	0.00847	ng/m ³ Air					U	
Beryllium	-6.72E-5	0.00176	ng/m ³ Air					U	
Cadmium	3.82E-5	0.00490	ng/m ³ Air					U	
Chromium	0.0224	2.32	ng/m ³ Air					U	
Cobalt	1.32E-4	0.465	ng/m ³ Air					U	
Lead	0.00100	0.123	ng/m ³ Air					U	
Manganese	1.12E-4	0.735	ng/m ³ Air					U	
Nickel	-0.00654	0.733	ng/m ³ Air					U	
Selenium	0.00135	0.0102	ng/m ³ Air					U, LJ, QX	

LCS (B6F2501-BS1)

Prepared: 06/25/26 Analyzed: 06/30/26

Antimony	0.963	0.0331	ng/m ³ Air	1.3826	69.6	80-120		SL	
Arsenic	2.78	0.00847	ng/m ³ Air	2.7658	100	80-120			
Beryllium	1.30	0.00176	ng/m ³ Air	1.3833	93.9	80-120			
Cadmium	1.43	0.00490	ng/m ³ Air	1.3839	104	80-120			
Chromium	15.1	2.32	ng/m ³ Air	13.832	109	80-120			
Cobalt	1.41	0.465	ng/m ³ Air	1.3835	102	80-120			
Lead	14.0	0.123	ng/m ³ Air	13.831	101	80-120			
Manganese	8.82	0.735	ng/m ³ Air	8.3008	106	80-120			
Nickel	3.08	0.733	ng/m ³ Air	2.7656	111	80-120			
Selenium	2.77	0.0102	ng/m ³ Air	2.7661	100	80-120		LJ, QX	

LCS (B6F2501-BS2)

Prepared: 06/25/26 Analyzed: 06/30/26

Antimony	1.39	0.0331	ng/m ³ Air	1.3826	100	80-120		SL	
----------	------	--------	-----------------------	--------	-----	--------	--	----	--

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 FAX: -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	PQL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Inorganics by Compendium Method IO-3.5 - Quality Control

Batch B6F2501 - ICP-MS Extraction

LCS (B6F2501-BS2) Continued

Prepared: 06/25/26 Analyzed: 06/30/26

Arsenic	2.75	0.00847	ng/m ³ Air	2.7658		99.3	80-120			
Beryllium	1.37	0.00176	ng/m ³ Air	1.3833		99.2	80-120			
Cadmium	1.39	0.00490	ng/m ³ Air	1.3839		100	80-120			
Chromium	14.1	2.32	ng/m ³ Air	13.832		102	80-120			
Cobalt	1.38	0.465	ng/m ³ Air	1.3835		100	80-120			
Lead	13.6	0.123	ng/m ³ Air	13.831		98.2	80-120			
Manganese	8.49	0.735	ng/m ³ Air	8.3008		102	80-120			
Nickel	2.81	0.733	ng/m ³ Air	2.7656		102	80-120			
Selenium	2.79	0.0102	ng/m ³ Air	2.7661		101	80-120			LJ, QX



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100 **FAX:** -

FILE #: 3508.00

REPORTED: 07/09/26 14:25

SUBMITTED: 04/01/26 to 05/27/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
SL	The spike recovery was outside acceptance limits. Reported value may be biased low.
QX	Compound does not meet QC criteria. Results should be considered an estimate.
LJ	Identification of analyte is acceptable; reported value is an estimate.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard.