



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

April 14, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 02/26/26 12:45 through 03/20/26 15:48.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6022628-03	Air	02/01/26 23:59	02/26/26 12:45
TOOK	6032029-01	Air	02/27/26 23:59	03/20/26 15:48



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6022628-03

Sampled: 02/01/26 23:59

Pressure @ Receipt: 3.0

Canister #: 35154

Received: 02/26/26 12:45

Comments:

Analyzed: 03/02/26 15:23

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.430	2.13		0.0237
Chloromethane	0.430	0.89		0.0213
Dichlorotetrafluoroethane	0.0145	0.10		0.0112
Vinyl chloride	8.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0109	0.02	U, CF, LL	0.0215
Ethylene oxide	0.0260	0.05	U, CF, LK	0.0276
Bromomethane	0.00630	0.02	U	0.0104
Chloroethane	ND	ND	U	0.0136
Acetonitrile	10.5	17.70	CF, E, LK	0.184
Acrolein	0.0937	0.22	U, CF, LK, QB-01	0.0977
Trichlorofluoromethane	0.189	1.06		0.00950
Acrylonitrile	0.00770	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.109	0.38		0.0489
Carbon Disulfide	0.0277	0.09	U	0.0473
Trichlorotrifluoroethane	0.0801	0.62		0.0210
trans-1,2-Dichloroethylene	8.00E-4	0.00	U	0.0171
1,1-Dichloroethane	0.00110	0.00	U	0.00655
Methyl tert-Butyl Ether	0.00240	0.01	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0186	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0170	0.07		0.0116
1,1,1-Trichloroethane	8.00E-4	0.00	U	0.00952
Benzene	0.240	0.77		0.0158
Carbon Tetrachloride	0.0682	0.43		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00170	0.01	U	0.00889
Trichloroethylene	0.00430	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00760	0.03	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6022628-03

Sampled: 02/01/26 23:59

Pressure @ Receipt: 3.0

Canister #: 35154

Received: 02/26/26 12:45

Comments:

Analyzed: 03/02/26 15:23

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.300	1.13		0.0380
Dibromochloromethane	9.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0855	0.40		0.0258
Tetrachloroethylene	0.00690	0.05	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0367	0.16		0.0130
m,p-Xylene	0.0998	0.43		0.0418
Bromoform	0.00190	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0400	0.17		0.0235
1,3,5-Trimethylbenzene	0.00780	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00130	0.01	U	0.0287
o-Dichlorobenzene	4.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	9.00E-4	0.01	U, QB-01	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6032029-01

Sampled: 02/27/26 23:59

Pressure @ Receipt: 3.5

Canister #: 41661

Received: 03/20/26 15:48

Comments:

Analyzed: 03/24/26 01:09

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.423	2.10		0.0237
Chloromethane	0.484	1.00		0.0213
Dichlorotetrafluoroethane	0.0162	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0642	0.14		0.0215
Ethylene oxide	0.0506	0.09		0.0276
Bromomethane	0.00640	0.02	U	0.0104
Chloroethane	0.0106	0.03	U	0.0136
Acetonitrile	18.2	30.60	E, QB-01	0.184
Acrolein	0.219	0.50	CF, LK, QB-01	0.0977
Trichlorofluoromethane	0.232	1.31		0.00950
Acrylonitrile	0.0109	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.204	0.71		0.0489
Carbon Disulfide	0.0187	0.06	U	0.0473
Trichlorotrifluoroethane	0.0738	0.57		0.0210
trans-1,2-Dichloroethylene	0.0175	0.07		0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	ND	ND	U	0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0181	0.07		0.0116
1,1,1-Trichloroethane	0.00130	0.01	U	0.00952
Benzene	0.882	2.82		0.0158
Carbon Tetrachloride	0.0718	0.45		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00550	0.03	U	0.0170
Methyl Methacrylate	ND	ND	QB-03, U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0411	0.17	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6032029-01

Sampled: 02/27/26 23:59

Pressure @ Receipt: 3.5

Canister #: 41661

Received: 03/20/26 15:48

Comments:

Analyzed: 03/24/26 01:09

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	2.32	8.76		0.0380
Dibromochloromethane	0.00170	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.534	2.50		0.0258
Tetrachloroethylene	0.0225	0.15		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.235	1.02		0.0130
m,p-Xylene	0.944	4.11		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.366	1.59		0.0235
1,3,5-Trimethylbenzene	0.107	0.53		0.0316
1,2,4-Trimethylbenzene	0.396	1.95		0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00960	0.06	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	1.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6B2705 - Summa Canister Prep

Blank (B6B2705-BLK1)

Prepared: 02/11/26 Analyzed: 03/02/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				QB-01, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6B2705 - Summa Canister Prep

Blank (B6B2705-BLK1) Continued

Prepared: 02/11/26 Analyzed: 03/02/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6C2308 - Summa Canister Prep

Blank (B6C2308-BLK1)

Prepared: 03/04/26 Analyzed: 03/23/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				QB-01, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2308 - Summa Canister Prep

Blank (B6C2308-BLK1) Continued

Prepared: 03/04/26 Analyzed: 03/23/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2602077

Calibration Check (2602077-CCV1)

Prepared: 02/27/26 Analyzed: 03/02/26

Dichlorodifluoromethane	0.989	ppbv	-3.8	30.00	QB-01
Chloromethane	0.939	ppbv	-7.9	30.00	
Dichlorotetrafluoroethane	0.949	ppbv	-8.4	30.00	
Vinyl chloride	0.940	ppbv	-8.9	30.00	
1,3-Butadiene	0.961	ppbv	-7.0	30.00	
Ethylene oxide	0.891	ppbv	-6.8	30.00	
Bromomethane	0.979	ppbv	-5.1	30.00	
Chloroethane	0.936	ppbv	-9.7	30.00	
Acetonitrile	0.979	ppbv	-5.7	30.00	
Acrolein	0.934	ppbv	-8.4	30.00	
Trichlorofluoromethane	0.969	ppbv	-4.1	30.00	
Acrylonitrile	1.06	ppbv	-3.6	30.00	
1,1-Dichloroethene	1.08	ppbv	3.7	30.00	
Dichloromethane	0.967	ppbv	-6.7	30.00	
Carbon Disulfide	1.34	ppbv	22.8	30.00	
Trichlorotrifluoroethane	0.994	ppbv	-3.5	30.00	
trans-1,2-Dichloroethylene	0.985	ppbv	-5.3	30.00	
1,1-Dichloroethane	0.978	ppbv	-5.6	30.00	
Methyl tert-Butyl Ether	1.05	ppbv	1.5	30.00	
Chloroprene	1.03	ppbv	-0.8	30.00	
cis-1,2-Dichloroethylene	1.00	ppbv	-3.4	30.00	
Bromochloromethane	0.927	ppbv	-9.8	30.00	
Chloroform	1.02	ppbv	-2.2	30.00	
Ethyl tert-Butyl Ether	1.05	ppbv	1.0	30.00	
1,2-Dichloroethane	1.02	ppbv	-1.6	30.00	
1,1,1-Trichloroethane	1.03	ppbv	-0.3	30.00	
Benzene	1.02	ppbv	-2.0	30.00	
Carbon Tetrachloride	1.05	ppbv	2.3	30.00	
tert-Amyl Methyl Ether	1.06	ppbv	2.8	30.00	
1,2-Dichloropropane	1.09	ppbv	5.1	30.00	
Ethyl Acrylate	1.12	ppbv	6.7	30.00	
Bromodichloromethane	1.13	ppbv	8.8	30.00	
Trichloroethylene	1.11	ppbv	7.0	30.00	
Methyl Methacrylate	1.11	ppbv	6.9	30.00	
cis-1,3-Dichloropropene	1.09	ppbv	4.2	30.00	
Methyl Isobutyl Ketone	1.16	ppbv	11.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2602077

Calibration Check (2602077-CCV1) Continued

Prepared: 02/27/26 Analyzed: 03/02/26

trans-1,3-Dichloropropene	1.13	ppbv	7.3	30.00	
1,1,2-Trichloroethane	1.08	ppbv	2.9	30.00	
Toluene	1.12	ppbv	7.5	30.00	
Dibromochloromethane	1.04	ppbv	0.4	30.00	
1,2-Dibromoethane	1.08	ppbv	3.0	30.00	
n-Octane	1.18	ppbv	13.8	30.00	
Tetrachloroethylene	1.09	ppbv	4.2	30.00	
Chlorobenzene	1.13	ppbv	7.7	30.00	
Ethylbenzene	1.08	ppbv	3.2	30.00	
m,p-Xylene	2.22	ppbv	6.3	30.00	
Bromoform	1.05	ppbv	0.7	30.00	
Styrene	1.13	ppbv	6.7	30.00	
1,1,2,2-Tetrachloroethane	1.12	ppbv	6.1	30.00	
o-Xylene	1.13	ppbv	6.6	30.00	
1,3,5-Trimethylbenzene	1.08	ppbv	2.4	30.00	
1,2,4-Trimethylbenzene	1.08	ppbv	2.4	30.00	
m-Dichlorobenzene	0.993	ppbv	-4.7	30.00	
p-Dichlorobenzene	0.988	ppbv	-5.6	30.00	
o-Dichlorobenzene	0.988	ppbv	-5.5	30.00	
1,2,4-Trichlorobenzene	0.906	ppbv	-12.7	30.00	QB-01
Hexachloro-1,3-butadiene	0.951	ppbv	-9.8	30.00	

Calibration Check (2602077-CCV2)

Prepared: 02/27/26 Analyzed: 03/03/26

Dichlorodifluoromethane	0.969	ppbv	-5.7	30.00	
Chloromethane	0.959	ppbv	-6.0	30.00	
Dichlorotetrafluoroethane	0.934	ppbv	-9.8	30.00	
Vinyl chloride	0.903	ppbv	-12.5	30.00	
1,3-Butadiene	0.818	ppbv	-20.9	30.00	
Ethylene oxide	0.928	ppbv	-2.9	30.00	
Bromomethane	0.975	ppbv	-5.6	30.00	
Chloroethane	0.943	ppbv	-9.0	30.00	
Acetonitrile	1.06	ppbv	2.6	30.00	
Acrolein	0.920	ppbv	-9.8	30.00	QB-01
Trichlorofluoromethane	0.965	ppbv	-4.4	30.00	
Acrylonitrile	1.07	ppbv	-2.7	30.00	
1,1-Dichloroethene	1.07	ppbv	2.6	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2602077

Calibration Check (2602077-CCV2) Continued

Prepared: 02/27/26 Analyzed: 03/03/26

Dichloromethane	0.982	ppbv	-5.2	30.00
Carbon Disulfide	1.09	ppbv	0.2	30.00
Trichlorotrifluoroethane	0.985	ppbv	-4.3	30.00
trans-1,2-Dichloroethylene	0.932	ppbv	-10.4	30.00
1,1-Dichloroethane	0.994	ppbv	-4.0	30.00
Methyl tert-Butyl Ether	0.976	ppbv	-5.8	30.00
Chloroprene	0.915	ppbv	-11.9	30.00
cis-1,2-Dichloroethylene	0.952	ppbv	-8.1	30.00
Bromochloromethane	0.928	ppbv	-9.7	30.00
Chloroform	1.06	ppbv	1.1	30.00
Ethyl tert-Butyl Ether	0.932	ppbv	-10.1	30.00
1,2-Dichloroethane	1.02	ppbv	-1.3	30.00
1,1,1-Trichloroethane	1.05	ppbv	2.2	30.00
Benzene	0.994	ppbv	-4.2	30.00
Carbon Tetrachloride	1.05	ppbv	3.0	30.00
tert-Amyl Methyl Ether	0.936	ppbv	-9.5	30.00
1,2-Dichloropropane	1.10	ppbv	5.7	30.00
Ethyl Acrylate	1.05	ppbv	0.3	30.00
Bromodichloromethane	1.13	ppbv	9.5	30.00
Trichloroethylene	1.07	ppbv	3.7	30.00
Methyl Methacrylate	1.07	ppbv	2.3	30.00
cis-1,3-Dichloropropene	1.01	ppbv	-3.3	30.00
Methyl Isobutyl Ketone	1.15	ppbv	10.3	30.00
trans-1,3-Dichloropropene	1.07	ppbv	1.5	30.00
1,1,2-Trichloroethane	1.07	ppbv	2.3	30.00
Toluene	1.09	ppbv	4.7	30.00
Dibromochloromethane	1.05	ppbv	1.6	30.00
1,2-Dibromoethane	1.05	ppbv	0.8	30.00
n-Octane	1.14	ppbv	10.2	30.00
Tetrachloroethylene	1.10	ppbv	5.6	30.00
Chlorobenzene	1.10	ppbv	4.7	30.00
Ethylbenzene	0.966	ppbv	-7.6	30.00
m,p-Xylene	2.03	ppbv	-2.8	30.00
Bromoform	1.08	ppbv	3.5	30.00
Styrene	1.09	ppbv	3.5	30.00
1,1,2,2-Tetrachloroethane	1.17	ppbv	11.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2602077

Calibration Check (2602077-CCV2) Continued

Prepared: 02/27/26 Analyzed: 03/03/26

o-Xylene	1.04	ppbv	-1.6	30.00	
1,3,5-Trimethylbenzene	0.966	ppbv	-8.2	30.00	
1,2,4-Trimethylbenzene	0.973	ppbv	-7.7	30.00	
m-Dichlorobenzene	0.948	ppbv	-9.0	30.00	
p-Dichlorobenzene	0.931	ppbv	-11.0	30.00	
o-Dichlorobenzene	0.937	ppbv	-10.4	30.00	
1,2,4-Trichlorobenzene	0.903	ppbv	-13.0	30.00	QB-01
Hexachloro-1,3-butadiene	0.967	ppbv	-8.3	30.00	

Sequence 2603073

Calibration Check (2603073-CCV1)

Prepared & Analyzed: 03/23/26

Dichlorodifluoromethane	0.960	ppbv	-6.6	30.00	
Chloromethane	0.902	ppbv	-11.6	30.00	
Dichlorotetrafluoroethane	0.944	ppbv	-8.9	30.00	
Vinyl chloride	0.941	ppbv	-8.9	30.00	
1,3-Butadiene	0.847	ppbv	-18.1	30.00	
Ethylene oxide	1.09	ppbv	13.8	30.00	
Bromomethane	0.912	ppbv	-11.6	30.00	
Chloroethane	0.936	ppbv	-9.7	30.00	
Acetonitrile	0.880	ppbv	-15.2	30.00	QB-01
Acrolein	1.08	ppbv	5.9	30.00	QB-01
Trichlorofluoromethane	1.31	ppbv	29.4	30.00	
Acrylonitrile	1.13	ppbv	2.0	30.00	
1,1-Dichloroethene	1.05	ppbv	1.0	30.00	
Dichloromethane	0.975	ppbv	-5.9	30.00	
Carbon Disulfide	0.969	ppbv	-10.9	30.00	
Trichlorotrifluoroethane	0.948	ppbv	-7.9	30.00	
trans-1,2-Dichloroethylene	1.02	ppbv	-2.3	30.00	
1,1-Dichloroethane	1.10	ppbv	5.9	30.00	
Methyl tert-Butyl Ether	1.03	ppbv	-0.3	30.00	
Chloroprene	1.04	ppbv	-0.06	30.00	
cis-1,2-Dichloroethylene	1.07	ppbv	2.9	30.00	
Bromochloromethane	1.03	ppbv	0.1	30.00	
Chloroform	1.15	ppbv	9.9	30.00	
Ethyl tert-Butyl Ether	1.03	ppbv	-0.5	30.00	
1,2-Dichloroethane	1.18	ppbv	13.9	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV1) Continued

Prepared & Analyzed: 03/23/26

1,1,1-Trichloroethane	1.16	ppbv	12.6	30.00
Benzene	1.07	ppbv	2.6	30.00
Carbon Tetrachloride	1.12	ppbv	9.2	30.00
tert-Amyl Methyl Ether	1.04	ppbv	0.1	30.00
1,2-Dichloropropane	1.10	ppbv	5.7	30.00
Ethyl Acrylate	1.07	ppbv	2.2	30.00
Bromodichloromethane	1.13	ppbv	9.0	30.00
Trichloroethylene	1.04	ppbv	0.6	30.00
Methyl Methacrylate	1.10	ppbv	5.9	30.00
cis-1,3-Dichloropropene	1.07	ppbv	2.3	30.00
Methyl Isobutyl Ketone	1.17	ppbv	12.5	30.00
trans-1,3-Dichloropropene	1.10	ppbv	4.7	30.00
1,1,2-Trichloroethane	1.09	ppbv	4.1	30.00
Toluene	1.10	ppbv	5.6	30.00
Dibromochloromethane	1.09	ppbv	5.2	30.00
1,2-Dibromoethane	1.09	ppbv	4.8	30.00
n-Octane	1.11	ppbv	6.5	30.00
Tetrachloroethylene	1.10	ppbv	4.9	30.00
Chlorobenzene	1.07	ppbv	1.7	30.00
Ethylbenzene	1.06	ppbv	1.7	30.00
m,p-Xylene	2.21	ppbv	6.0	30.00
Bromoform	1.07	ppbv	2.6	30.00
Styrene	1.15	ppbv	8.9	30.00
1,1,2,2-Tetrachloroethane	1.06	ppbv	0.9	30.00
o-Xylene	1.18	ppbv	11.7	30.00
1,3,5-Trimethylbenzene	1.11	ppbv	5.7	30.00
1,2,4-Trimethylbenzene	1.11	ppbv	5.3	30.00
m-Dichlorobenzene	1.09	ppbv	4.2	30.00
p-Dichlorobenzene	1.08	ppbv	3.5	30.00
o-Dichlorobenzene	1.10	ppbv	4.9	30.00
1,2,4-Trichlorobenzene	0.865	ppbv	-16.6	30.00
Hexachloro-1,3-butadiene	0.970	ppbv	-8.0	30.00

Calibration Check (2603073-CCV2)

Prepared: 03/23/26 Analyzed: 03/24/26

Dichlorodifluoromethane	0.891	ppbv	-13.3	30.00
Chloromethane	1.08	ppbv	6.0	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV2) Continued

Prepared: 03/23/26 Analyzed: 03/24/26

Dichlorotetrafluoroethane	1.03	ppbv	-0.4	30.00	
Vinyl chloride	0.920	ppbv	-10.9	30.00	
1,3-Butadiene	0.762	ppbv	-26.3	30.00	
Ethylene oxide	1.15	ppbv	20.6	30.00	
Bromomethane	1.03	ppbv	-0.3	30.00	
Chloroethane	1.07	ppbv	3.5	30.00	
Acetonitrile	0.972	ppbv	-6.4	30.00	QB-01
Acrolein	1.12	ppbv	9.9	30.00	QB-01
Trichlorofluoromethane	1.27	ppbv	26.0	30.00	
Acrylonitrile	1.16	ppbv	5.2	30.00	
1,1-Dichloroethene	1.09	ppbv	4.8	30.00	
Dichloromethane	1.03	ppbv	-0.9	30.00	
Carbon Disulfide	1.01	ppbv	-7.4	30.00	
Trichlorotrifluoroethane	1.11	ppbv	7.7	30.00	
trans-1,2-Dichloroethylene	0.988	ppbv	-5.0	30.00	
1,1-Dichloroethane	1.12	ppbv	8.6	30.00	
Methyl tert-Butyl Ether	0.997	ppbv	-3.8	30.00	
Chloroprene	0.998	ppbv	-3.9	30.00	
cis-1,2-Dichloroethylene	1.06	ppbv	2.4	30.00	
Bromochloromethane	1.04	ppbv	1.5	30.00	
Chloroform	1.19	ppbv	13.2	30.00	
Ethyl tert-Butyl Ether	0.992	ppbv	-4.2	30.00	
1,2-Dichloroethane	1.19	ppbv	14.9	30.00	
1,1,1-Trichloroethane	1.18	ppbv	14.3	30.00	
Benzene	1.08	ppbv	3.8	30.00	
Carbon Tetrachloride	1.11	ppbv	8.7	30.00	
tert-Amyl Methyl Ether	1.00	ppbv	-3.2	30.00	
1,2-Dichloropropane	1.12	ppbv	8.0	30.00	
Ethyl Acrylate	1.06	ppbv	1.5	30.00	
Bromodichloromethane	1.13	ppbv	8.7	30.00	
Trichloroethylene	1.02	ppbv	-1.0	30.00	
Methyl Methacrylate	1.09	ppbv	5.0	30.00	
cis-1,3-Dichloropropene	1.05	ppbv	0.7	30.00	
Methyl Isobutyl Ketone	1.18	ppbv	13.6	30.00	
trans-1,3-Dichloropropene	1.08	ppbv	3.2	30.00	
1,1,2-Trichloroethane	1.08	ppbv	3.7	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV2) Continued

Prepared: 03/23/26 Analyzed: 03/24/26

Toluene	1.11	ppbv	6.7	30.00
Dibromochloromethane	1.08	ppbv	4.4	30.00
1,2-Dibromoethane	1.08	ppbv	3.3	30.00
n-Octane	1.13	ppbv	8.5	30.00
Tetrachloroethylene	1.09	ppbv	4.4	30.00
Chlorobenzene	1.06	ppbv	0.4	30.00
Ethylbenzene	1.03	ppbv	-1.6	30.00
m,p-Xylene	2.15	ppbv	2.9	30.00
Bromoform	1.06	ppbv	1.3	30.00
Styrene	1.15	ppbv	8.7	30.00
1,1,2,2-Tetrachloroethane	1.08	ppbv	2.2	30.00
o-Xylene	1.17	ppbv	10.5	30.00
1,3,5-Trimethylbenzene	1.09	ppbv	3.7	30.00
1,2,4-Trimethylbenzene	1.09	ppbv	3.3	30.00
m-Dichlorobenzene	1.06	ppbv	1.4	30.00
p-Dichlorobenzene	1.04	ppbv	-0.3	30.00
o-Dichlorobenzene	1.05	ppbv	0.8	30.00
1,2,4-Trichlorobenzene	0.836	ppbv	-19.4	30.00
Hexachloro-1,3-butadiene	0.947	ppbv	-10.1	30.00



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 04/14/26 14:07

SUBMITTED: 02/26/26 to 03/20/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.