



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

August 14, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 06/11/26 13:04 through 07/07/26 09:51.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6061710-02	Air	06/13/26 23:59	06/17/26 13:28
TOOK	6062603-01	Air	06/19/26 23:59	06/26/26 10:35
TOOK	6070730-01	Air	06/25/26 23:59	07/07/26 09:51
TOOK FB	6070730-02	Air	06/25/26 23:59	07/07/26 09:51



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK	Lab ID: 6061710-02	Sampled: 06/13/26 23:59
Temp Rec'd (°C): 1.0	Sample Volume: 1162.3 L	Received: 06/17/26 13:28
	Cartridge Pouch #: 26G1051	Analyzed: 06/19/26 22:47

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.27		0.0545
Acetaldehyde	1.52		0.0225
Acetone	0.694		0.0835
Propionaldehyde	0.257		0.178
Crotonaldehyde	ND	U	0.0163
Butyraldehyde	0.174		0.00964
2-Butanone	0.158		0.0363
Benzaldehyde	0.102		0.0373
Valeraldehyde	0.145		0.0104
Hexaldehyde	0.142		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6062603-01

Sampled: 06/19/26 23:59

Temp Rec'd (°C): 9.6

Sample Volume: 1162.2 L

Received: 06/26/26 10:35

Cartridge Pouch #: 26G1052

Analyzed: 07/06/26 14:46

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.26		0.0545
Acetaldehyde	1.55		0.0225
Acetone	1.08		0.0835
Propionaldehyde	0.314		0.178
Crotonaldehyde	ND	U	0.0163
Butyraldehyde	0.207		0.00964
2-Butanone	0.264		0.0363
Benzaldehyde	0.135		0.0373
Valeraldehyde	0.207		0.0104
Hexaldehyde	0.151		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6070730-01

Sampled: 06/25/26 23:59

Temp Rec'd (°C): 5.5

Sample Volume: 1162 L

Received: 07/07/26 09:51

Cartridge Pouch #: 26G1053

Analyzed: 07/13/26 16:37

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.85		0.0545
Acetaldehyde	1.60		0.0225
Acetone	0.619		0.0835
Propionaldehyde	0.241		0.178
Crotonaldehyde	ND	CE, U	0.0164
Butyraldehyde	0.186		0.00964
2-Butanone	0.0921		0.0363
Benzaldehyde	0.103		0.0373
Valeraldehyde	0.138		0.0104
Hexaldehyde	0.134		0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK FB

Lab ID: 6070730-02

Sampled: 06/25/26 23:59

Temp Rec'd (°C): 5.5

Sample Volume: 1162 L

Received: 07/07/26 09:51

Cartridge Pouch #: 26G1054

Analyzed: 07/13/26 16:53

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0380	U	0.0545
Acetaldehyde	0.0205	U	0.0225
Acetone	0.0560	U	0.0835
Propionaldehyde	ND	CE, U	0.178
Crotonaldehyde	ND	U	0.0164
Butyraldehyde	ND	U	0.00964
2-Butanone	ND	U	0.0363
Benzaldehyde	0.0346	U	0.0373
Valeraldehyde	ND	U	0.0104
Hexaldehyde	ND	U	0.0105



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606078 - B6F1806

Calibration Blank (2606078-CCB1)

Prepared & Analyzed: 06/19/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2606078-CCB2)

Prepared: 06/19/26 Analyzed: 06/20/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2606078-CCV1)

Prepared & Analyzed: 06/19/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.8	84.9-115.1
Propionaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Butyraldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1
2-Butanone	0.143	ug/ml	0.15000	95.7	84.9-115.1
Benzaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Valeraldehyde	0.145	ug/ml	0.15000	96.7	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.0	84.9-115.1

Calibration Check (2606078-CCV2)

Prepared & Analyzed: 06/19/26

Formaldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606078 - B6F1806

Calibration Check (2606078-CCV2) Contin

Prepared & Analyzed: 06/19/26

Acetone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	95.0	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.6	84.9-115.1
2-Butanone	0.141	ug/ml	0.15000	93.8	84.9-115.1
Benzaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Valeraldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1
Hexaldehyde	0.140	ug/ml	0.15000	93.6	84.9-115.1

Calibration Check (2606078-CCV3)

Prepared & Analyzed: 06/19/26

Formaldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.9	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.5	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.7	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	96.3	84.9-115.1
Benzaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1

Calibration Check (2606078-CCV4)

Prepared & Analyzed: 06/19/26

Formaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.0	84.9-115.1
Propionaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.9	84.9-115.1
2-Butanone	0.143	ug/ml	0.15000	95.5	84.9-115.1
Benzaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Valeraldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1

Calibration Check (2606078-CCV5)

Prepared: 06/19/26 Analyzed: 06/20/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.145	ug/ml	0.15000	96.6	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.3	84.9-115.1
Propionaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Crotonaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	93.8	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2606078 - B6F1806

Calibration Check (2606078-CCV5) Contin

Prepared: 06/19/26 Analyzed: 06/20/26

2-Butanone	0.144	ug/ml	0.15000	96.3	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.0	84.9-115.1
Hexaldehyde	0.141	ug/ml	0.15000	93.8	84.9-115.1

Batch 2607017 - B6G0202

Calibration Blank (2607017-CCB1)

Prepared & Analyzed: 07/06/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2607017-CCB2)

Prepared & Analyzed: 07/06/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2607017-CCV1)

Prepared & Analyzed: 07/06/26

Formaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Acetaldehyde	0.142	ug/ml	0.15000	95.0	84.9-115.1
Acetone	0.144	ug/ml	0.15000	95.8	84.9-115.1
Propionaldehyde	0.144	ug/ml	0.15000	95.8	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	97.8	84.9-115.1
Butyraldehyde	0.139	ug/ml	0.15000	93.0	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	96.0	84.9-115.1
Benzaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Valeraldehyde	0.141	ug/ml	0.15000	94.1	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2607017 - B6G0202

Calibration Check (2607017-CCV1) Contin

Prepared & Analyzed: 07/06/26

Hexaldehyde	0.138	ug/ml	0.15000	92.0	84.9-115.1
-------------	-------	-------	---------	------	------------

Calibration Check (2607017-CCV2)

Prepared & Analyzed: 07/06/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Acetone	0.143	ug/ml	0.15000	95.5	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
Crotonaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.4	84.9-115.1
2-Butanone	0.141	ug/ml	0.15000	94.2	84.9-115.1
Benzaldehyde	0.145	ug/ml	0.15000	96.8	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.3	84.9-115.1

Calibration Check (2607017-CCV3)

Prepared & Analyzed: 07/06/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
Acetaldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
Acetone	0.145	ug/ml	0.15000	96.9	84.9-115.1
Propionaldehyde	0.143	ug/ml	0.15000	95.6	84.9-115.1
Crotonaldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
Butyraldehyde	0.139	ug/ml	0.15000	92.5	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	95.8	84.9-115.1
Benzaldehyde	0.146	ug/ml	0.15000	97.2	84.9-115.1
Valeraldehyde	0.144	ug/ml	0.15000	96.0	84.9-115.1
Hexaldehyde	0.141	ug/ml	0.15000	93.7	84.9-115.1

Batch 2607045 - B6G0907

Calibration Blank (2607045-CCB1)

Prepared & Analyzed: 07/13/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2607045-CCB2)

Prepared: 07/13/26 Analyzed: 07/14/26

Formaldehyde	0.00	ug/ml			U
--------------	------	-------	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2607045 - B6G0907

Calibration Blank (2607045-CCB2) Contin

Prepared: 07/13/26 Analyzed: 07/14/26

Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2607045-CCV1)

Prepared & Analyzed: 07/13/26

Formaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Acetaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Acetone	0.153	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
2-Butanone	0.154	ug/ml	0.15000	103	84.9-115.1
Benzaldehyde	0.160	ug/ml	0.15000	107	84.9-115.1
Valeraldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1

Calibration Check (2607045-CCV2)

Prepared & Analyzed: 07/13/26

Formaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Acetaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Acetone	0.153	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Butyraldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
2-Butanone	0.152	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.5	84.9-115.1

Calibration Check (2607045-CCV3)

Prepared & Analyzed: 07/13/26

Formaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Acetaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Acetone	0.153	ug/ml	0.15000	102	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2607045 - B6G0907

Calibration Check (2607045-CCV3) Contin

Prepared & Analyzed: 07/13/26

Butyraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
2-Butanone	0.157	ug/ml	0.15000	105	84.9-115.1
Benzaldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1

Calibration Check (2607045-CCV4)

Prepared & Analyzed: 07/13/26

Formaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Acetaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Acetone	0.155	ug/ml	0.15000	103	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.157	ug/ml	0.15000	104	84.9-115.1
Butyraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
2-Butanone	0.156	ug/ml	0.15000	104	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1

Calibration Check (2607045-CCV5)

Prepared: 07/13/26 Analyzed: 07/14/26

Formaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Acetaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Acetone	0.154	ug/ml	0.15000	103	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Butyraldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
2-Butanone	0.158	ug/ml	0.15000	105	84.9-115.1
Benzaldehyde	0.159	ug/ml	0.15000	106	84.9-115.1
Valeraldehyde	0.158	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1

Batch B6F1806 - TO-11A extraction

Blank (B6F1806-BLK1)

Prepared: 06/18/26 Analyzed: 06/19/26

Formaldehyde	ND	ug/m ³ Air	U
Acetaldehyde	ND	ug/m ³ Air	U
Acetone	ND	ug/m ³ Air	U
Propionaldehyde	ND	ug/m ³ Air	CE, U
Crotonaldehyde	ND	ug/m ³ Air	U
Butyraldehyde	ND	ug/m ³ Air	U
2-Butanone	ND	ug/m ³ Air	U
Benzaldehyde	ND	ug/m ³ Air	U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6F1806 - TO-11A extraction

Blank (B6F1806-BLK1) Continued

Prepared: 06/18/26 Analyzed: 06/19/26

Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6F1806-BS1)

Prepared: 06/18/26 Analyzed: 06/19/26

Formaldehyde	0.279	ug/ml	0.30000		93.0	79.9-120.1			
Acetaldehyde	0.290	ug/ml	0.30000		96.7	79.9-120.1			
Acetone	0.295	ug/ml	0.30000		98.4	69.9-130.1			
Propionaldehyde	0.277	ug/ml	0.30000		92.3	69.9-130.1			
Crotonaldehyde	0.267	ug/ml	0.30000		89.0	69.9-130.1			
Butyraldehyde	0.248	ug/ml	0.30000		82.7	69.9-130.1			
2-Butanone	0.230	ug/ml	0.30000		76.5	69.9-130.1			
Benzaldehyde	0.303	ug/ml	0.30000		101	69.9-130.1			
Valeraldehyde	0.254	ug/ml	0.30000		84.8	69.9-130.1			
Hexaldehyde	0.244	ug/ml	0.30000		81.5	69.9-130.1			

LCS Dup (B6F1806-BSD1)

Prepared: 06/18/26 Analyzed: 06/19/26

Formaldehyde	0.275	ug/ml	0.30000		91.5	79.9-120.1	1.65	20	
Acetaldehyde	0.286	ug/ml	0.30000		95.4	79.9-120.1	1.38	20	
Acetone	0.289	ug/ml	0.30000		96.4	69.9-130.1	2.00	20	
Propionaldehyde	0.277	ug/ml	0.30000		92.4	69.9-130.1	0.113	20	
Crotonaldehyde	0.272	ug/ml	0.30000		90.5	69.9-130.1	1.66	20	
Butyraldehyde	0.244	ug/ml	0.30000		81.4	69.9-130.1	1.64	20	
2-Butanone	0.233	ug/ml	0.30000		77.5	69.9-130.1	1.30	20	
Benzaldehyde	0.297	ug/ml	0.30000		99.0	69.9-130.1	1.98	20	
Valeraldehyde	0.245	ug/ml	0.30000		81.8	69.9-130.1	3.68	20	
Hexaldehyde	0.242	ug/ml	0.30000		80.7	69.9-130.1	0.925	20	

Batch B6G0202 - TO-11A extraction

Blank (B6G0202-BLK1)

Prepared: 07/02/26 Analyzed: 07/06/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0262	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							CE, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6G0202-BS1)

Prepared: 07/02/26 Analyzed: 07/06/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6G0202 - TO-11A extraction

LCS (B6G0202-BS1) Continued

Prepared: 07/02/26 Analyzed: 07/06/26

Formaldehyde	0.278	ug/ml	0.30000	92.8	79.9-120.1
Acetaldehyde	0.287	ug/ml	0.30000	95.7	79.9-120.1
Acetone	0.289	ug/ml	0.30000	96.3	69.9-130.1
Propionaldehyde	0.256	ug/ml	0.30000	85.2	69.9-130.1
Crotonaldehyde	0.275	ug/ml	0.30000	91.7	69.9-130.1
Butyraldehyde	0.247	ug/ml	0.30000	82.4	69.9-130.1
2-Butanone	0.228	ug/ml	0.30000	75.9	69.9-130.1
Benzaldehyde	0.298	ug/ml	0.30000	99.3	69.9-130.1
Valeraldehyde	0.251	ug/ml	0.30000	83.8	69.9-130.1
Hexaldehyde	0.244	ug/ml	0.30000	81.4	69.9-130.1

LCS Dup (B6G0202-BSD1)

Prepared: 07/02/26 Analyzed: 07/06/26

Formaldehyde	0.278	ug/ml	0.30000	92.6	79.9-120.1	0.279	20
Acetaldehyde	0.289	ug/ml	0.30000	96.3	79.9-120.1	0.603	20
Acetone	0.295	ug/ml	0.30000	98.2	69.9-130.1	1.94	20
Propionaldehyde	0.267	ug/ml	0.30000	89.1	69.9-130.1	4.51	20
Crotonaldehyde	0.272	ug/ml	0.30000	90.8	69.9-130.1	1.00	20
Butyraldehyde	0.246	ug/ml	0.30000	81.9	69.9-130.1	0.615	20
2-Butanone	0.226	ug/ml	0.30000	75.5	69.9-130.1	0.524	20
Benzaldehyde	0.303	ug/ml	0.30000	101	69.9-130.1	1.56	20
Valeraldehyde	0.248	ug/ml	0.30000	82.7	69.9-130.1	1.29	20
Hexaldehyde	0.243	ug/ml	0.30000	81.1	69.9-130.1	0.329	20

Batch B6G0903 - TO-11A extraction

Blank (B6G0903-BLK1)

Prepared: 07/09/26 Analyzed: 07/13/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0550	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					CE, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	0.0547	ug/m ³ Air					
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6G0903-BS1)

Prepared: 07/09/26 Analyzed: 07/13/26

Formaldehyde	0.292	ug/ml	0.30000	97.2	79.9-120.1
Acetaldehyde	0.307	ug/ml	0.30000	102	79.9-120.1
Acetone	0.309	ug/ml	0.30000	103	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6G0903 - TO-11A extraction

LCS (B6G0903-BS1) Continued

Prepared: 07/09/26 Analyzed: 07/13/26

Propionaldehyde	0.285	ug/ml	0.30000	95.0	69.9-130.1
Crotonaldehyde	0.293	ug/ml	0.30000	97.7	69.9-130.1
Butyraldehyde	0.263	ug/ml	0.30000	87.8	69.9-130.1
2-Butanone	0.253	ug/ml	0.30000	84.5	69.9-130.1
Benzaldehyde	0.316	ug/ml	0.30000	105	69.9-130.1
Valeraldehyde	0.262	ug/ml	0.30000	87.3	69.9-130.1
Hexaldehyde	0.253	ug/ml	0.30000	84.3	69.9-130.1

LCS Dup (B6G0903-BSD1)

Prepared: 07/09/26 Analyzed: 07/13/26

Formaldehyde	0.287	ug/ml	0.30000	95.8	79.9-120.1	1.44	20
Acetaldehyde	0.302	ug/ml	0.30000	101	79.9-120.1	1.63	20
Acetone	0.300	ug/ml	0.30000	100	69.9-130.1	2.80	20
Propionaldehyde	0.278	ug/ml	0.30000	92.5	69.9-130.1	2.60	20
Crotonaldehyde	0.283	ug/ml	0.30000	94.4	69.9-130.1	3.40	20
Butyraldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1	3.46	20
2-Butanone	0.244	ug/ml	0.30000	81.2	69.9-130.1	3.92	20
Benzaldehyde	0.316	ug/ml	0.30000	105	69.9-130.1	0.177	20
Valeraldehyde	0.260	ug/ml	0.30000	86.5	69.9-130.1	0.906	20
Hexaldehyde	0.248	ug/ml	0.30000	82.7	69.9-130.1	1.89	20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 08/14/26 09:02

SUBMITTED: 06/17/26 to 07/07/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
S-Rec	Sample received by laboratory outside of initial extraction hold time - 14 days -
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.