



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

May 14, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 03/18/26 11:31 through 04/02/26 09:55.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6031916-01	Air	03/06/26 23:59	03/19/26 11:18
TOOK	6031916-02	Air	03/09/26 23:59	03/19/26 11:18
TOOK	6033039-01	Air	03/21/26 23:59	03/30/26 14:30
TOOK	6040319-01	Air	03/27/26 23:59	04/02/26 09:55



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6031916-01

Sampled: 03/06/26 23:59

Pressure @ Receipt: 3.0

Canister #: 35158

Received: 03/19/26 11:18

Comments: Makeup

Analyzed: 03/25/26 01:35

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.428	2.12		0.0237
Chloromethane	0.542	1.12		0.0213
Dichlorotetrafluoroethane	0.0154	0.11		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00610	0.01	U	0.0215
Ethylene oxide	0.0826	0.15		0.0276
Bromomethane	0.0119	0.05		0.0104
Chloroethane	0.00830	0.02	U	0.0136
Acetonitrile	15.1	25.40	E	0.184
Acrolein	0.292	0.67	CF, LK	0.0977
Trichlorofluoromethane	0.240	1.35	LK	0.00950
Acrylonitrile	0.0109	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.0874	0.30		0.0489
Carbon Disulfide	0.00670	0.02	U	0.0473
Trichlorotrifluoroethane	0.0610	0.47		0.0210
trans-1,2-Dichloroethylene	ND	ND	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	0.00360	0.01	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0183	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.00980	0.04	U	0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.190	0.61		0.0158
Carbon Tetrachloride	0.0741	0.47		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00170	0.01	U	0.00889
Trichloroethylene	0.00380	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0183	0.08	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6031916-01

Sampled: 03/06/26 23:59

Pressure @ Receipt: 3.0

Canister #: 35158

Received: 03/19/26 11:18

Comments: Makeup

Analyzed: 03/25/26 01:35

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.362	1.37		0.0380
Dibromochloromethane	7.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0534	0.25		0.0258
Tetrachloroethylene	0.00580	0.04	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0320	0.14		0.0130
m,p-Xylene	0.0888	0.39		0.0418
Bromoform	0.00220	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0339	0.15		0.0235
1,3,5-Trimethylbenzene	0.00800	0.04	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00130	0.01	U	0.0287
o-Dichlorobenzene	2.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	1.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6031916-02

Sampled: 03/09/26 23:59

Pressure @ Receipt: 4.0

Canister #: 39936

Received: 03/19/26 11:18

Comments:

Analyzed: 03/25/26 02:38

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.419	2.08		0.0237
Chloromethane	0.540	1.12		0.0213
Dichlorotetrafluoroethane	0.0156	0.11		0.0112
Vinyl chloride	6.00E-4	0.00	U	0.0140
1,3-Butadiene	0.00540	0.01	U	0.0215
Ethylene oxide	0.117	0.21		0.0276
Bromomethane	0.00660	0.03	U	0.0104
Chloroethane	0.0147	0.04		0.0136
Acetonitrile	15.1	25.40	CF, E, LK	0.184
Acrolein	0.249	0.57	CF, LK	0.0977
Trichlorofluoromethane	0.240	1.35	LK	0.00950
Acrylonitrile	0.00980	0.02	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.113	0.39		0.0489
Carbon Disulfide	0.0195	0.06	U	0.0473
Trichlorotrifluoroethane	0.0648	0.50		0.0210
trans-1,2-Dichloroethylene	9.00E-4	0.00	U	0.0171
1,1-Dichloroethane	7.00E-4	0.00	U	0.00655
Methyl tert-Butyl Ether	0.00370	0.01	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0208	0.10		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0133	0.05		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.401	1.28		0.0158
Carbon Tetrachloride	0.0755	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	9.00E-4	0.01	U	0.00889
Trichloroethylene	ND	ND	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0211	0.09	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6031916-02

Sampled: 03/09/26 23:59

Pressure @ Receipt: 4.0

Canister #: 39936

Received: 03/19/26 11:18

Comments:

Analyzed: 03/25/26 02:38

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.415	1.57		0.0380
Dibromochloromethane	6.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0742	0.35		0.0258
Tetrachloroethylene	0.00810	0.06	U	0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0437	0.19		0.0130
m,p-Xylene	0.142	0.62		0.0418
Bromoform	0.00170	0.02	U	0.0166
Styrene	ND	ND	CE, U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0597	0.26		0.0235
1,3,5-Trimethylbenzene	0.0115	0.06	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	3.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00270	0.02	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	5.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	ND	ND	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6033039-01

Sampled: 03/21/26 23:59

Pressure @ Receipt: 4.0

Canister #: 41626

Received: 03/30/26 14:30

Comments:

Analyzed: 04/15/26 03:59

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.527	2.61		0.0237
Chloromethane	0.549	1.14		0.0213
Dichlorotetrafluoroethane	0.0191	0.13		0.0112
Vinyl chloride	8.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0297	0.07	CF, LL	0.0215
Ethylene oxide	0.107	0.19	CF, LK	0.0276
Bromomethane	0.00870	0.03	U	0.0104
Chloroethane	0.0155	0.04		0.0136
Acetonitrile	169	284.00	E	0.184
Acrolein	0.612	1.41	CF, LK	0.0977
Trichlorofluoromethane	0.227	1.28		0.00950
Acrylonitrile	0.0241	0.05		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.173	0.60		0.0489
Carbon Disulfide	0.0140	0.04	U	0.0473
Trichlorotrifluoroethane	0.0757	0.58		0.0210
trans-1,2-Dichloroethylene	0.0480	0.19		0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	2.00E-4	0.00	U	0.00905
Chloroform	0.0293	0.14		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0147	0.06		0.0116
1,1,1-Trichloroethane	0.00150	0.01	U	0.00952
Benzene	0.580	1.86		0.0158
Carbon Tetrachloride	0.0761	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00290	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0344	0.14	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6033039-01

Sampled: 03/21/26 23:59

Pressure @ Receipt: 4.0

Canister #: 41626

Received: 03/30/26 14:30

Comments:

Analyzed: 04/15/26 03:59

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	1.00	3.78		0.0380
Dibromochloromethane	8.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.163	0.76		0.0258
Tetrachloroethylene	0.0360	0.25		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.119	0.52		0.0130
m,p-Xylene	0.395	1.72		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	ND	ND	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.168	0.73		0.0235
1,3,5-Trimethylbenzene	0.0341	0.17		0.0316
1,2,4-Trimethylbenzene	ND	ND	U	0.0320
m-Dichlorobenzene	7.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00520	0.03	U	0.0287
o-Dichlorobenzene	8.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00220	0.02	QB-01, U	0.0792
Hexachloro-1,3-butadiene	3.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6040319-01

Sampled: 03/27/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43389

Received: 04/02/26 09:55

Comments:

Analyzed: 04/21/26 03:51

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.511	2.53		0.0237
Chloromethane	0.491	1.02		0.0213
Dichlorotetrafluoroethane	0.0180	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.00530	0.01	CF, LL, U	0.0215
Ethylene oxide	0.0522	0.09	CF, LK	0.0276
Bromomethane	0.00790	0.03	U	0.0104
Chloroethane	0.00870	0.02	U	0.0136
Acetonitrile	85.3	144.00	E, QB-01, QB-03	0.184
Acrolein	0.170	0.39	CF, LK, QB-01	0.0977
Trichlorofluoromethane	0.234	1.32		0.00950
Acrylonitrile	0.0115	0.03	U	0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.101	0.35		0.0489
Carbon Disulfide	0.00920	0.03	U	0.0473
Trichlorotrifluoroethane	0.0770	0.59		0.0210
trans-1,2-Dichloroethylene	7.00E-4	0.00	U	0.0171
1,1-Dichloroethane	0.00100	0.00	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0161	0.08		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0186	0.08		0.0116
1,1,1-Trichloroethane	0.00110	0.01	U	0.00952
Benzene	0.112	0.36		0.0158
Carbon Tetrachloride	0.0765	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	5.00E-4	0.00	U	0.00889
Trichloroethylene	0.00450	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.00920	0.04	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6040319-01

Sampled: 03/27/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43389

Received: 04/02/26 09:55

Comments:

Analyzed: 04/21/26 03:51

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.136	0.51		0.0380
Dibromochloromethane	3.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0155	0.07	U	0.0258
Tetrachloroethylene	0.0119	0.08		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0160	0.07		0.0130
m,p-Xylene	0.0484	0.21		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	0.00740	0.03	U	0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0170	0.07	U	0.0235
1,3,5-Trimethylbenzene	0.00290	0.01	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	0.00120	0.01	U	0.0277
p-Dichlorobenzene	0.00160	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	3.00E-4	0.00	U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2410 - Summa Canister Prep

Blank (B6C2410-BLK1)

Prepared: 03/18/26 Analyzed: 03/24/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				LK, U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6C2410 - Summa Canister Prep

Blank (B6C2410-BLK1) Continued

Prepared: 03/18/26 Analyzed: 03/24/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D0807 - Summa Canister Prep

Blank (B6D0807-BLK1)

Prepared: 04/07/26 Analyzed: 04/14/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D0807 - Summa Canister Prep

Blank (B6D0807-BLK1) Continued

Prepared: 04/07/26 Analyzed: 04/14/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6D1703 - Summa Canister Prep

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D1703 - Summa Canister Prep

Blank (B6D1703-BLK1)

Prepared: 04/14/26 Analyzed: 04/20/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				QB-01, U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6D1703 - Summa Canister Prep

Blank (B6D1703-BLK1) Continued

Prepared: 04/14/26 Analyzed: 04/20/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV1)

Prepared & Analyzed: 03/23/26

Dichlorodifluoromethane	0.960	ppbv	-6.6	30.00	
Chloromethane	0.902	ppbv	-11.6	30.00	
Dichlorotetrafluoroethane	0.944	ppbv	-8.9	30.00	
Vinyl chloride	0.941	ppbv	-8.9	30.00	
1,3-Butadiene	0.847	ppbv	-18.1	30.00	
Ethylene oxide	1.09	ppbv	13.8	30.00	
Bromomethane	0.912	ppbv	-11.6	30.00	
Chloroethane	0.936	ppbv	-9.7	30.00	
Acetonitrile	0.880	ppbv	-15.2	30.00	QB-01
Acrolein	1.08	ppbv	5.9	30.00	QB-01
Trichlorofluoromethane	1.31	ppbv	29.4	30.00	
Acrylonitrile	1.13	ppbv	2.0	30.00	
1,1-Dichloroethene	1.05	ppbv	1.0	30.00	
Dichloromethane	0.975	ppbv	-5.9	30.00	
Carbon Disulfide	0.969	ppbv	-10.9	30.00	
Trichlorotrifluoroethane	0.948	ppbv	-7.9	30.00	
trans-1,2-Dichloroethylene	1.02	ppbv	-2.3	30.00	
1,1-Dichloroethane	1.10	ppbv	5.9	30.00	
Methyl tert-Butyl Ether	1.03	ppbv	-0.3	30.00	
Chloroprene	1.04	ppbv	-0.06	30.00	
cis-1,2-Dichloroethylene	1.07	ppbv	2.9	30.00	
Bromochloromethane	1.03	ppbv	0.1	30.00	
Chloroform	1.15	ppbv	9.9	30.00	
Ethyl tert-Butyl Ether	1.03	ppbv	-0.5	30.00	
1,2-Dichloroethane	1.18	ppbv	13.9	30.00	
1,1,1-Trichloroethane	1.16	ppbv	12.6	30.00	
Benzene	1.07	ppbv	2.6	30.00	
Carbon Tetrachloride	1.12	ppbv	9.2	30.00	
tert-Amyl Methyl Ether	1.04	ppbv	0.1	30.00	
1,2-Dichloropropane	1.10	ppbv	5.7	30.00	
Ethyl Acrylate	1.07	ppbv	2.2	30.00	
Bromodichloromethane	1.13	ppbv	9.0	30.00	
Trichloroethylene	1.04	ppbv	0.6	30.00	
Methyl Methacrylate	1.10	ppbv	5.9	30.00	
cis-1,3-Dichloropropene	1.07	ppbv	2.3	30.00	
Methyl Isobutyl Ketone	1.17	ppbv	12.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV1) Continued

Prepared & Analyzed: 03/23/26

trans-1,3-Dichloropropene	1.10	ppbv	4.7	30.00
1,1,2-Trichloroethane	1.09	ppbv	4.1	30.00
Toluene	1.10	ppbv	5.6	30.00
Dibromochloromethane	1.09	ppbv	5.2	30.00
1,2-Dibromoethane	1.09	ppbv	4.8	30.00
n-Octane	1.11	ppbv	6.5	30.00
Tetrachloroethylene	1.10	ppbv	4.9	30.00
Chlorobenzene	1.07	ppbv	1.7	30.00
Ethylbenzene	1.06	ppbv	1.7	30.00
m,p-Xylene	2.21	ppbv	6.0	30.00
Bromoform	1.07	ppbv	2.6	30.00
Styrene	1.15	ppbv	8.9	30.00
1,1,2,2-Tetrachloroethane	1.06	ppbv	0.9	30.00
o-Xylene	1.18	ppbv	11.7	30.00
1,3,5-Trimethylbenzene	1.11	ppbv	5.7	30.00
1,2,4-Trimethylbenzene	1.11	ppbv	5.3	30.00
m-Dichlorobenzene	1.09	ppbv	4.2	30.00
p-Dichlorobenzene	1.08	ppbv	3.5	30.00
o-Dichlorobenzene	1.10	ppbv	4.9	30.00
1,2,4-Trichlorobenzene	0.865	ppbv	-16.6	30.00
Hexachloro-1,3-butadiene	0.970	ppbv	-8.0	30.00

Calibration Check (2603073-CCV2)

Prepared: 03/23/26 Analyzed: 03/24/26

Dichlorodifluoromethane	0.891	ppbv	-13.3	30.00	
Chloromethane	1.08	ppbv	6.0	30.00	
Dichlorotetrafluoroethane	1.03	ppbv	-0.4	30.00	
Vinyl chloride	0.920	ppbv	-10.9	30.00	
1,3-Butadiene	0.762	ppbv	-26.3	30.00	
Ethylene oxide	1.15	ppbv	20.6	30.00	
Bromomethane	1.03	ppbv	-0.3	30.00	
Chloroethane	1.07	ppbv	3.5	30.00	
Acetonitrile	0.972	ppbv	-6.4	30.00	QB-01
Acrolein	1.12	ppbv	9.9	30.00	QB-01
Trichlorofluoromethane	1.27	ppbv	26.0	30.00	
Acrylonitrile	1.16	ppbv	5.2	30.00	
1,1-Dichloroethene	1.09	ppbv	4.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV2) Continued

Prepared: 03/23/26 Analyzed: 03/24/26

Dichloromethane	1.03	ppbv	-0.9	30.00
Carbon Disulfide	1.01	ppbv	-7.4	30.00
Trichlorotrifluoroethane	1.11	ppbv	7.7	30.00
trans-1,2-Dichloroethylene	0.988	ppbv	-5.0	30.00
1,1-Dichloroethane	1.12	ppbv	8.6	30.00
Methyl tert-Butyl Ether	0.997	ppbv	-3.8	30.00
Chloroprene	0.998	ppbv	-3.9	30.00
cis-1,2-Dichloroethylene	1.06	ppbv	2.4	30.00
Bromochloromethane	1.04	ppbv	1.5	30.00
Chloroform	1.19	ppbv	13.2	30.00
Ethyl tert-Butyl Ether	0.992	ppbv	-4.2	30.00
1,2-Dichloroethane	1.19	ppbv	14.9	30.00
1,1,1-Trichloroethane	1.18	ppbv	14.3	30.00
Benzene	1.08	ppbv	3.8	30.00
Carbon Tetrachloride	1.11	ppbv	8.7	30.00
tert-Amyl Methyl Ether	1.00	ppbv	-3.2	30.00
1,2-Dichloropropane	1.12	ppbv	8.0	30.00
Ethyl Acrylate	1.06	ppbv	1.5	30.00
Bromodichloromethane	1.13	ppbv	8.7	30.00
Trichloroethylene	1.02	ppbv	-1.0	30.00
Methyl Methacrylate	1.09	ppbv	5.0	30.00
cis-1,3-Dichloropropene	1.05	ppbv	0.7	30.00
Methyl Isobutyl Ketone	1.18	ppbv	13.6	30.00
trans-1,3-Dichloropropene	1.08	ppbv	3.2	30.00
1,1,2-Trichloroethane	1.08	ppbv	3.7	30.00
Toluene	1.11	ppbv	6.7	30.00
Dibromochloromethane	1.08	ppbv	4.4	30.00
1,2-Dibromoethane	1.08	ppbv	3.3	30.00
n-Octane	1.13	ppbv	8.5	30.00
Tetrachloroethylene	1.09	ppbv	4.4	30.00
Chlorobenzene	1.06	ppbv	0.4	30.00
Ethylbenzene	1.03	ppbv	-1.6	30.00
m,p-Xylene	2.15	ppbv	2.9	30.00
Bromoform	1.06	ppbv	1.3	30.00
Styrene	1.15	ppbv	8.7	30.00
1,1,2,2-Tetrachloroethane	1.08	ppbv	2.2	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603073

Calibration Check (2603073-CCV2) Continued

Prepared: 03/23/26 Analyzed: 03/24/26

o-Xylene	1.17	ppbv	10.5	30.00
1,3,5-Trimethylbenzene	1.09	ppbv	3.7	30.00
1,2,4-Trimethylbenzene	1.09	ppbv	3.3	30.00
m-Dichlorobenzene	1.06	ppbv	1.4	30.00
p-Dichlorobenzene	1.04	ppbv	-0.3	30.00
o-Dichlorobenzene	1.05	ppbv	0.8	30.00
1,2,4-Trichlorobenzene	0.836	ppbv	-19.4	30.00
Hexachloro-1,3-butadiene	0.947	ppbv	-10.1	30.00

Sequence 2603079

Calibration Check (2603079-CCV1)

Prepared & Analyzed: 03/24/26

Dichlorodifluoromethane	0.883	ppbv	-14.1	30.00
Chloromethane	1.05	ppbv	3.1	30.00
Dichlorotetrafluoroethane	1.02	ppbv	-1.9	30.00
Vinyl chloride	0.888	ppbv	-13.9	30.00
1,3-Butadiene	0.732	ppbv	-29.2	30.00
Ethylene oxide	1.15	ppbv	20.2	30.00
Bromomethane	0.958	ppbv	-7.1	30.00
Chloroethane	0.993	ppbv	-4.2	30.00
Acetonitrile	1.13	ppbv	8.6	30.00
Acrolein	1.17	ppbv	14.5	30.00
Trichlorofluoromethane	1.31	ppbv	29.4	30.00
Acrylonitrile	1.18	ppbv	6.8	30.00
1,1-Dichloroethene	1.05	ppbv	0.6	30.00
Dichloromethane	1.14	ppbv	9.7	30.00
Carbon Disulfide	0.989	ppbv	-9.1	30.00
Trichlorotrifluoroethane	0.931	ppbv	-9.6	30.00
trans-1,2-Dichloroethylene	1.02	ppbv	-2.3	30.00
1,1-Dichloroethane	1.11	ppbv	7.3	30.00
Methyl tert-Butyl Ether	1.06	ppbv	2.3	30.00
Chloroprene	1.08	ppbv	4.2	30.00
cis-1,2-Dichloroethylene	1.08	ppbv	4.0	30.00
Bromochloromethane	1.05	ppbv	1.8	30.00
Chloroform	1.15	ppbv	9.6	30.00
Ethyl tert-Butyl Ether	1.08	ppbv	3.8	30.00
1,2-Dichloroethane	1.14	ppbv	10.7	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603079

Calibration Check (2603079-CCV1) Continued

Prepared & Analyzed: 03/24/26

1,1,1-Trichloroethane	1.14	ppbv	10.8	30.00
Benzene	1.09	ppbv	5.2	30.00
Carbon Tetrachloride	1.09	ppbv	6.6	30.00
tert-Amyl Methyl Ether	1.06	ppbv	2.5	30.00
1,2-Dichloropropane	1.09	ppbv	5.2	30.00
Ethyl Acrylate	1.09	ppbv	4.0	30.00
Bromodichloromethane	1.09	ppbv	5.7	30.00
Trichloroethylene	1.05	ppbv	1.7	30.00
Methyl Methacrylate	1.10	ppbv	5.8	30.00
cis-1,3-Dichloropropene	1.07	ppbv	2.0	30.00
Methyl Isobutyl Ketone	1.17	ppbv	12.8	30.00
trans-1,3-Dichloropropene	1.07	ppbv	1.8	30.00
1,1,2-Trichloroethane	1.08	ppbv	2.8	30.00
Toluene	1.11	ppbv	6.5	30.00
Dibromochloromethane	1.07	ppbv	3.1	30.00
1,2-Dibromoethane	1.08	ppbv	3.4	30.00
n-Octane	1.11	ppbv	6.7	30.00
Tetrachloroethylene	1.08	ppbv	3.5	30.00
Chlorobenzene	1.07	ppbv	1.5	30.00
Ethylbenzene	1.06	ppbv	1.4	30.00
m,p-Xylene	2.19	ppbv	5.1	30.00
Bromoform	1.08	ppbv	3.5	30.00
Styrene	1.16	ppbv	10.1	30.00
1,1,2,2-Tetrachloroethane	1.19	ppbv	12.7	30.00
o-Xylene	1.19	ppbv	12.8	30.00
1,3,5-Trimethylbenzene	0.762	ppbv	-27.6	30.00
1,2,4-Trimethylbenzene	0.763	ppbv	-27.6	30.00
m-Dichlorobenzene	0.913	ppbv	-12.4	30.00
p-Dichlorobenzene	0.888	ppbv	-15.2	30.00
o-Dichlorobenzene	0.880	ppbv	-15.8	30.00
1,2,4-Trichlorobenzene	0.875	ppbv	-15.7	30.00
Hexachloro-1,3-butadiene	0.963	ppbv	-8.6	30.00

Calibration Check (2603079-CCV2)

Prepared: 03/24/26 Analyzed: 03/25/26

Dichlorodifluoromethane	0.857	ppbv	-16.6	30.00
Chloromethane	1.03	ppbv	0.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603079

Calibration Check (2603079-CCV2) Continued

Prepared: 03/24/26 Analyzed: 03/25/26

Dichlorotetrafluoroethane	0.987	ppbv	-4.7	30.00	
Vinyl chloride	0.893	ppbv	-13.5	30.00	
1,3-Butadiene	0.736	ppbv	-28.9	30.00	
Ethylene oxide	1.22	ppbv	28.0	30.00	
Bromomethane	0.999	ppbv	-3.2	30.00	
Chloroethane	1.00	ppbv	-3.2	30.00	
Acetonitrile	1.14	ppbv	10.1	30.00	
Acrolein	1.17	ppbv	15.1	30.00	
Trichlorofluoromethane	1.43	ppbv	41.2	30.00	LK
Acrylonitrile	1.18	ppbv	6.7	30.00	
1,1-Dichloroethene	1.11	ppbv	7.0	30.00	
Dichloromethane	1.14	ppbv	10.4	30.00	
Carbon Disulfide	1.03	ppbv	-5.8	30.00	
Trichlorotrifluoroethane	0.973	ppbv	-5.5	30.00	
trans-1,2-Dichloroethylene	1.02	ppbv	-1.9	30.00	
1,1-Dichloroethane	1.13	ppbv	9.3	30.00	
Methyl tert-Butyl Ether	1.07	ppbv	3.5	30.00	
Chloroprene	1.09	ppbv	5.4	30.00	
cis-1,2-Dichloroethylene	1.09	ppbv	5.5	30.00	
Bromochloromethane	1.07	ppbv	3.7	30.00	
Chloroform	1.19	ppbv	13.8	30.00	
Ethyl tert-Butyl Ether	1.08	ppbv	4.6	30.00	
1,2-Dichloroethane	1.19	ppbv	15.2	30.00	
1,1,1-Trichloroethane	1.20	ppbv	16.1	30.00	
Benzene	1.12	ppbv	7.4	30.00	
Carbon Tetrachloride	1.13	ppbv	10.8	30.00	
tert-Amyl Methyl Ether	1.08	ppbv	4.0	30.00	
1,2-Dichloropropane	1.10	ppbv	5.5	30.00	
Ethyl Acrylate	1.08	ppbv	3.7	30.00	
Bromodichloromethane	1.11	ppbv	6.9	30.00	
Trichloroethylene	1.05	ppbv	1.9	30.00	
Methyl Methacrylate	1.11	ppbv	7.0	30.00	
cis-1,3-Dichloropropene	1.07	ppbv	2.2	30.00	
Methyl Isobutyl Ketone	1.18	ppbv	13.1	30.00	
trans-1,3-Dichloropropene	1.08	ppbv	2.5	30.00	
1,1,2-Trichloroethane	1.07	ppbv	2.7	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2603079

Calibration Check (2603079-CCV2) Continued

Prepared: 03/24/26 Analyzed: 03/25/26

Toluene	1.12	ppbv	7.6	30.00
Dibromochloromethane	1.08	ppbv	4.6	30.00
1,2-Dibromoethane	1.08	ppbv	3.9	30.00
n-Octane	1.11	ppbv	7.1	30.00
Tetrachloroethylene	1.10	ppbv	5.1	30.00
Chlorobenzene	1.06	ppbv	1.2	30.00
Ethylbenzene	1.03	ppbv	-1.9	30.00
m,p-Xylene	2.15	ppbv	2.8	30.00
Bromoform	1.00	ppbv	-4.1	30.00
Styrene	0.977	ppbv	-7.7	30.00
1,1,2,2-Tetrachloroethane	1.16	ppbv	10.1	30.00
o-Xylene	0.889	ppbv	-15.8	30.00
1,3,5-Trimethylbenzene	0.829	ppbv	-21.2	30.00
1,2,4-Trimethylbenzene	0.826	ppbv	-21.7	30.00
m-Dichlorobenzene	0.943	ppbv	-9.5	30.00
p-Dichlorobenzene	0.920	ppbv	-12.1	30.00
o-Dichlorobenzene	0.916	ppbv	-12.4	30.00
1,2,4-Trichlorobenzene	0.935	ppbv	-9.9	30.00
Hexachloro-1,3-butadiene	0.940	ppbv	-10.8	30.00

Sequence 2604029

Calibration Check (2604029-CCV1)

Prepared & Analyzed: 04/14/26

Dichlorodifluoromethane	1.12	ppbv	9.2	30.00
Chloromethane	1.12	ppbv	9.4	30.00
Dichlorotetrafluoroethane	1.14	ppbv	10.0	30.00
Vinyl chloride	1.13	ppbv	9.2	30.00
1,3-Butadiene	1.14	ppbv	9.8	30.00
Ethylene oxide	0.951	ppbv	-0.6	30.00
Bromomethane	1.14	ppbv	10.6	30.00
Chloroethane	1.14	ppbv	9.9	30.00
Acetonitrile	1.09	ppbv	4.7	30.00
Acrolein	1.08	ppbv	6.2	30.00
Trichlorofluoromethane	1.12	ppbv	10.9	30.00
Acrylonitrile	1.13	ppbv	2.2	30.00
1,1-Dichloroethene	1.15	ppbv	10.3	30.00
Dichloromethane	1.14	ppbv	10.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604029

Calibration Check (2604029-CCV1) Continued

Prepared & Analyzed: 04/14/26

Carbon Disulfide	1.16	ppbv	6.9	30.00
Trichlorotrifluoroethane	1.08	ppbv	4.4	30.00
trans-1,2-Dichloroethylene	1.15	ppbv	10.4	30.00
1,1-Dichloroethane	1.11	ppbv	7.5	30.00
Methyl tert-Butyl Ether	1.12	ppbv	8.0	30.00
Chloroprene	1.12	ppbv	7.9	30.00
cis-1,2-Dichloroethylene	1.13	ppbv	8.7	30.00
Bromochloromethane	1.10	ppbv	6.9	30.00
Chloroform	1.11	ppbv	6.3	30.00
Ethyl tert-Butyl Ether	1.14	ppbv	10.1	30.00
1,2-Dichloroethane	1.07	ppbv	3.7	30.00
1,1,1-Trichloroethane	1.12	ppbv	8.3	30.00
Benzene	1.08	ppbv	4.3	30.00
Carbon Tetrachloride	1.15	ppbv	12.1	30.00
tert-Amyl Methyl Ether	1.13	ppbv	9.6	30.00
1,2-Dichloropropane	1.15	ppbv	10.8	30.00
Ethyl Acrylate	1.16	ppbv	11.2	30.00
Bromodichloromethane	1.15	ppbv	10.6	30.00
Trichloroethylene	1.10	ppbv	6.8	30.00
Methyl Methacrylate	1.16	ppbv	11.5	30.00
cis-1,3-Dichloropropene	1.19	ppbv	13.8	30.00
Methyl Isobutyl Ketone	1.15	ppbv	10.1	30.00
trans-1,3-Dichloropropene	1.21	ppbv	15.0	30.00
1,1,2-Trichloroethane	1.14	ppbv	9.4	30.00
Toluene	1.15	ppbv	11.0	30.00
Dibromochloromethane	1.17	ppbv	13.4	30.00
1,2-Dibromoethane	1.16	ppbv	11.2	30.00
n-Octane	1.16	ppbv	11.7	30.00
Tetrachloroethylene	1.21	ppbv	15.6	30.00
Chlorobenzene	1.17	ppbv	10.8	30.00
Ethylbenzene	1.13	ppbv	7.8	30.00
m,p-Xylene	2.28	ppbv	9.2	30.00
Bromoform	1.21	ppbv	15.3	30.00
Styrene	1.15	ppbv	9.0	30.00
1,1,2,2-Tetrachloroethane	1.17	ppbv	11.3	30.00
o-Xylene	1.16	ppbv	10.0	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604029

Calibration Check (2604029-CCV1) Continued

Prepared & Analyzed: 04/14/26

1,3,5-Trimethylbenzene	1.19	ppbv	12.8	30.00	QB-01
1,2,4-Trimethylbenzene	1.17	ppbv	10.8	30.00	
m-Dichlorobenzene	1.14	ppbv	9.5	30.00	
p-Dichlorobenzene	1.14	ppbv	9.2	30.00	
o-Dichlorobenzene	1.15	ppbv	9.6	30.00	
1,2,4-Trichlorobenzene	1.18	ppbv	13.6	30.00	
Hexachloro-1,3-butadiene	1.14	ppbv	8.5	30.00	

Calibration Check (2604029-CCV2)

Prepared: 04/14/26 Analyzed: 04/15/26

Dichlorodifluoromethane	1.19	ppbv	16.0	30.00
Chloromethane	1.08	ppbv	5.7	30.00
Dichlorotetrafluoroethane	1.16	ppbv	12.0	30.00
Vinyl chloride	1.04	ppbv	0.5	30.00
1,3-Butadiene	1.03	ppbv	-0.3	30.00
Ethylene oxide	0.990	ppbv	3.5	30.00
Bromomethane	1.10	ppbv	6.3	30.00
Chloroethane	1.06	ppbv	2.4	30.00
Acetonitrile	1.09	ppbv	5.2	30.00
Acrolein	1.06	ppbv	4.0	30.00
Trichlorofluoromethane	1.12	ppbv	11.4	30.00
Acrylonitrile	1.15	ppbv	4.0	30.00
1,1-Dichloroethene	1.12	ppbv	7.5	30.00
Dichloromethane	1.09	ppbv	5.6	30.00
Carbon Disulfide	1.07	ppbv	-1.5	30.00
Trichlorotrifluoroethane	1.09	ppbv	5.4	30.00
trans-1,2-Dichloroethylene	1.06	ppbv	1.7	30.00
1,1-Dichloroethane	1.13	ppbv	9.2	30.00
Methyl tert-Butyl Ether	1.11	ppbv	7.5	30.00
Chloroprene	1.13	ppbv	9.0	30.00
cis-1,2-Dichloroethylene	1.10	ppbv	6.5	30.00
Bromochloromethane	1.04	ppbv	1.3	30.00
Chloroform	1.16	ppbv	10.7	30.00
Ethyl tert-Butyl Ether	1.15	ppbv	10.7	30.00
1,2-Dichloroethane	1.15	ppbv	11.2	30.00
1,1,1-Trichloroethane	1.18	ppbv	14.2	30.00
Benzene	1.11	ppbv	6.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604029

Calibration Check (2604029-CCV2) Continued

Prepared: 04/14/26 Analyzed: 04/15/26

Carbon Tetrachloride	1.14	ppbv	11.6	30.00	
tert-Amyl Methyl Ether	1.16	ppbv	11.7	30.00	
1,2-Dichloropropane	1.12	ppbv	7.5	30.00	
Ethyl Acrylate	1.11	ppbv	6.4	30.00	
Bromodichloromethane	1.09	ppbv	5.4	30.00	
Trichloroethylene	1.08	ppbv	4.6	30.00	
Methyl Methacrylate	1.13	ppbv	8.2	30.00	
cis-1,3-Dichloropropene	1.17	ppbv	12.1	30.00	
Methyl Isobutyl Ketone	1.12	ppbv	7.4	30.00	
trans-1,3-Dichloropropene	1.20	ppbv	14.1	30.00	
1,1,2-Trichloroethane	1.02	ppbv	-2.7	30.00	
Toluene	1.11	ppbv	6.4	30.00	
Dibromochloromethane	1.08	ppbv	4.1	30.00	
1,2-Dibromoethane	1.03	ppbv	-1.6	30.00	
n-Octane	1.13	ppbv	8.7	30.00	
Tetrachloroethylene	1.19	ppbv	14.0	30.00	
Chlorobenzene	1.10	ppbv	4.1	30.00	
Ethylbenzene	1.03	ppbv	-1.9	30.00	
m,p-Xylene	2.11	ppbv	1.1	30.00	
Bromoform	1.24	ppbv	18.7	30.00	
Styrene	1.11	ppbv	4.5	30.00	
1,1,2,2-Tetrachloroethane	1.24	ppbv	17.4	30.00	
o-Xylene	1.15	ppbv	9.3	30.00	
1,3,5-Trimethylbenzene	1.16	ppbv	10.0	30.00	
1,2,4-Trimethylbenzene	1.14	ppbv	8.0	30.00	
m-Dichlorobenzene	1.05	ppbv	0.6	30.00	
p-Dichlorobenzene	1.05	ppbv	0.5	30.00	
o-Dichlorobenzene	1.05	ppbv	0.8	30.00	
1,2,4-Trichlorobenzene	1.13	ppbv	9.2	30.00	QB-01
Hexachloro-1,3-butadiene	1.19	ppbv	13.1	30.00	

Sequence 2604057

Calibration Check (2604057-CCV1)

Prepared & Analyzed: 04/20/26

Dichlorodifluoromethane	1.18	ppbv	14.5	30.00	
Chloromethane	1.12	ppbv	9.4	30.00	
Dichlorotetrafluoroethane	1.16	ppbv	12.1	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604057

Calibration Check (2604057-CCV1) Continued

Prepared & Analyzed: 04/20/26

Vinyl chloride	1.10	ppbv	6.6	30.00	
1,3-Butadiene	1.01	ppbv	-2.0	30.00	
Ethylene oxide	1.03	ppbv	7.9	30.00	
Bromomethane	1.16	ppbv	12.7	30.00	
Chloroethane	1.16	ppbv	11.9	30.00	
Acetonitrile	0.823	ppbv	-20.7	30.00	QB-01
Acrolein	1.09	ppbv	6.8	30.00	QB-01
Trichlorofluoromethane	1.21	ppbv	19.5	30.00	
Acrylonitrile	1.17	ppbv	6.2	30.00	
1,1-Dichloroethene	1.23	ppbv	18.6	30.00	
Dichloromethane	1.16	ppbv	12.1	30.00	
Carbon Disulfide	1.27	ppbv	17.0	30.00	
Trichlorotrifluoroethane	1.23	ppbv	19.6	30.00	
trans-1,2-Dichloroethylene	1.13	ppbv	8.6	30.00	
1,1-Dichloroethane	1.17	ppbv	12.8	30.00	
Methyl tert-Butyl Ether	1.08	ppbv	4.4	30.00	
Chloroprene	1.05	ppbv	1.4	30.00	
cis-1,2-Dichloroethylene	1.12	ppbv	8.5	30.00	
Bromochloromethane	1.18	ppbv	15.1	30.00	
Chloroform	1.21	ppbv	15.5	30.00	
Ethyl tert-Butyl Ether	1.02	ppbv	-1.2	30.00	
1,2-Dichloroethane	1.20	ppbv	15.7	30.00	
1,1,1-Trichloroethane	1.20	ppbv	16.8	30.00	
Benzene	1.12	ppbv	7.6	30.00	
Carbon Tetrachloride	1.21	ppbv	18.1	30.00	
tert-Amyl Methyl Ether	0.988	ppbv	-4.5	30.00	
1,2-Dichloropropane	1.03	ppbv	-0.6	30.00	
Ethyl Acrylate	0.992	ppbv	-5.2	30.00	
Bromodichloromethane	1.11	ppbv	7.2	30.00	
Trichloroethylene	1.09	ppbv	5.2	30.00	
Methyl Methacrylate	0.992	ppbv	-4.8	30.00	
cis-1,3-Dichloropropene	1.05	ppbv	1.0	30.00	
Methyl Isobutyl Ketone	1.02	ppbv	-2.3	30.00	
trans-1,3-Dichloropropene	1.12	ppbv	6.9	30.00	
1,1,2-Trichloroethane	1.12	ppbv	7.5	30.00	
Toluene	1.05	ppbv	1.4	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604057

Calibration Check (2604057-CCV1) Continued

Prepared & Analyzed: 04/20/26

Dibromochloromethane	1.13	ppbv	9.2	30.00
1,2-Dibromoethane	1.11	ppbv	6.1	30.00
n-Octane	0.985	ppbv	-5.1	30.00
Tetrachloroethylene	1.14	ppbv	9.0	30.00
Chlorobenzene	1.12	ppbv	6.9	30.00
Ethylbenzene	1.04	ppbv	-0.7	30.00
m,p-Xylene	2.16	ppbv	3.6	30.00
Bromoform	1.18	ppbv	12.7	30.00
Styrene	1.12	ppbv	6.3	30.00
1,1,2,2-Tetrachloroethane	1.15	ppbv	9.0	30.00
o-Xylene	1.06	ppbv	0.4	30.00
1,3,5-Trimethylbenzene	1.03	ppbv	-2.4	30.00
1,2,4-Trimethylbenzene	1.01	ppbv	-3.9	30.00
m-Dichlorobenzene	1.11	ppbv	7.0	30.00
p-Dichlorobenzene	1.10	ppbv	5.1	30.00
o-Dichlorobenzene	1.11	ppbv	6.0	30.00
1,2,4-Trichlorobenzene	1.03	ppbv	-0.7	30.00
Hexachloro-1,3-butadiene	1.08	ppbv	2.8	30.00

Calibration Check (2604057-CCV2)

Prepared: 04/20/26 Analyzed: 04/21/26

Dichlorodifluoromethane	1.19	ppbv	15.6	30.00
Chloromethane	1.12	ppbv	10.2	30.00
Dichlorotetrafluoroethane	1.18	ppbv	14.1	30.00
Vinyl chloride	1.06	ppbv	2.5	30.00
1,3-Butadiene	0.892	ppbv	-13.7	30.00
Ethylene oxide	1.02	ppbv	6.6	30.00
Bromomethane	1.18	ppbv	14.1	30.00
Chloroethane	1.18	ppbv	13.7	30.00
Acetonitrile	0.935	ppbv	-9.9	30.00
Acrolein	1.09	ppbv	7.2	30.00
Trichlorofluoromethane	1.20	ppbv	19.0	30.00
Acrylonitrile	1.15	ppbv	4.1	30.00
1,1-Dichloroethene	1.22	ppbv	16.8	30.00
Dichloromethane	1.18	ppbv	14.2	30.00
Carbon Disulfide	1.27	ppbv	16.5	30.00
Trichlorotrifluoroethane	1.25	ppbv	21.1	30.00

QB-01

QB-01

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604057

Calibration Check (2604057-CCV2) Continued

Prepared: 04/20/26 Analyzed: 04/21/26

trans-1,2-Dichloroethylene	1.13	ppbv	8.7	30.00
1,1-Dichloroethane	1.17	ppbv	12.7	30.00
Methyl tert-Butyl Ether	0.957	ppbv	-7.6	30.00
Chloroprene	0.972	ppbv	-6.4	30.00
cis-1,2-Dichloroethylene	1.11	ppbv	7.2	30.00
Bromochloromethane	1.18	ppbv	14.7	30.00
Chloroform	1.19	ppbv	13.2	30.00
Ethyl tert-Butyl Ether	0.891	ppbv	-14.0	30.00
1,2-Dichloroethane	1.18	ppbv	14.2	30.00
1,1,1-Trichloroethane	1.15	ppbv	11.3	30.00
Benzene	1.08	ppbv	3.6	30.00
Carbon Tetrachloride	1.17	ppbv	13.9	30.00
tert-Amyl Methyl Ether	0.838	ppbv	-18.9	30.00
1,2-Dichloropropane	1.16	ppbv	11.3	30.00
Ethyl Acrylate	0.942	ppbv	-9.9	30.00
Bromodichloromethane	1.11	ppbv	7.4	30.00
Trichloroethylene	1.08	ppbv	4.2	30.00
Methyl Methacrylate	0.941	ppbv	-9.7	30.00
cis-1,3-Dichloropropene	0.989	ppbv	-5.3	30.00
Methyl Isobutyl Ketone	0.974	ppbv	-6.4	30.00
trans-1,3-Dichloropropene	1.09	ppbv	4.2	30.00
1,1,2-Trichloroethane	1.16	ppbv	10.8	30.00
Toluene	1.02	ppbv	-2.4	30.00
Dibromochloromethane	1.12	ppbv	7.9	30.00
1,2-Dibromoethane	1.10	ppbv	5.8	30.00
n-Octane	0.959	ppbv	-7.6	30.00
Tetrachloroethylene	1.11	ppbv	6.7	30.00
Chlorobenzene	1.13	ppbv	7.3	30.00
Ethylbenzene	1.01	ppbv	-3.8	30.00
m,p-Xylene	2.14	ppbv	2.3	30.00
Bromoform	1.17	ppbv	11.8	30.00
Styrene	1.10	ppbv	3.7	30.00
1,1,2,2-Tetrachloroethane	1.16	ppbv	9.8	30.00
o-Xylene	0.993	ppbv	-5.9	30.00
1,3,5-Trimethylbenzene	0.926	ppbv	-12.0	30.00
1,2,4-Trimethylbenzene	0.913	ppbv	-13.4	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2604057

Calibration Check (2604057-CCV2) Continued

Prepared: 04/20/26 Analyzed: 04/21/26

m-Dichlorobenzene	1.10	ppbv	5.3	30.00
p-Dichlorobenzene	1.07	ppbv	2.7	30.00
o-Dichlorobenzene	1.08	ppbv	3.3	30.00
1,2,4-Trichlorobenzene	0.945	ppbv	-9.0	30.00
Hexachloro-1,3-butadiene	1.04	ppbv	-1.1	30.00



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 05/14/26 15:12

SUBMITTED: 03/19/26 to 04/02/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.