



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

July 15, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: TOOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 05/07/26 11:31 through 06/15/26 10:15.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
TOOK	6050707-01	Air	05/02/26 23:59	05/07/26 11:31
TOOK	6051509-01	Air	05/08/26 23:59	05/15/26 11:51
TOOK D1	6052820-01	Air	05/14/26 23:59	05/28/26 12:37
TOOK	6061505-01	Air	05/26/26 23:59	06/15/26 10:15



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6050707-01

Sampled: 05/02/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43358

Received: 05/07/26 11:31

Comments:

Analyzed: 05/19/26 15:04

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.462	2.29		0.0237
Chloromethane	0.458	0.95		0.0213
Dichlorotetrafluoroethane	0.0172	0.12		0.0112
Vinyl chloride	0.00140	0.00	U	0.0140
1,3-Butadiene	0.0160	0.04	U, CF, LL	0.0215
Ethylene oxide	0.0581	0.11	CF, LK	0.0276
Bromomethane	0.00880	0.03	U	0.0104
Chloroethane	0.0105	0.03	U	0.0136
Acetonitrile	33.5	56.40	E, QB-01	0.184
Acrolein	0.234	0.54	CF, LK	0.0977
Trichlorofluoromethane	0.207	1.17		0.00950
Acrylonitrile	ND	ND	U, CE	0.0116
1,1-Dichloroethene	6.00E-4	0.00	U	0.0158
Dichloromethane	0.133	0.46		0.0489
Carbon Disulfide	0.0141	0.04	U	0.0473
Trichlorotrifluoroethane	0.0628	0.48		0.0210
trans-1,2-Dichloroethylene	0.00200	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	7.00E-4	0.00	U	0.00905
Chloroform	0.0193	0.09		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0180	0.07		0.0116
1,1,1-Trichloroethane	0.00150	0.01	U	0.00952
Benzene	0.225	0.72		0.0158
Carbon Tetrachloride	0.0645	0.41		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00150	0.01	U	0.00889
Trichloroethylene	0.00900	0.05	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0209	0.09	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6050707-01

Sampled: 05/02/26 23:59

Pressure @ Receipt: 4.0

Canister #: 43358

Received: 05/07/26 11:31

Comments:

Analyzed: 05/19/26 15:04

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.703	2.65		0.0380
Dibromochloromethane	0.00110	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0823	0.39		0.0258
Tetrachloroethylene	0.0162	0.11		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0687	0.30		0.0130
m,p-Xylene	0.217	0.94		0.0418
Bromoform	0.00130	0.01	U	0.0166
Styrene	0.0279	0.12		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0800	0.35		0.0235
1,3,5-Trimethylbenzene	0.0137	0.07	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	0.00160	0.01	U	0.0277
p-Dichlorobenzene	0.00690	0.04	U	0.0287
o-Dichlorobenzene	0.00170	0.01	U	0.0364
1,2,4-Trichlorobenzene	0.00400	0.03	U, QB-01	0.0792
Hexachloro-1,3-butadiene	0.00110	0.01	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6051509-01

Sampled: 05/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 41636

Received: 05/15/26 11:51

Comments:

Analyzed: 05/23/26 03:22

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.460	2.28		0.0237
Chloromethane	0.514	1.06		0.0213
Dichlorotetrafluoroethane	0.0163	0.11		0.0112
Vinyl chloride	9.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0124	0.03	U, CF, LL	0.0215
Ethylene oxide	0.0729	0.13		0.0276
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.0151	0.04	CF, LK	0.0136
Acetonitrile	55.6	93.50	E, QB-03	0.184
Acrolein	0.251	0.58	CF, LK	0.0977
Trichlorofluoromethane	0.201	1.13		0.00950
Acrylonitrile	0.0145	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.320	1.11		0.0489
Carbon Disulfide	0.0144	0.04	U	0.0473
Trichlorotrifluoroethane	0.0577	0.44		0.0210
trans-1,2-Dichloroethylene	0.00810	0.03	U	0.0171
1,1-Dichloroethane	0.00130	0.01	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	ND	ND	U	0.00905
Chloroform	ND	ND	U	0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0172	0.07		0.0116
1,1,1-Trichloroethane	0.00100	0.01	U	0.00952
Benzene	0.722	2.31		0.0158
Carbon Tetrachloride	0.0699	0.44		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	ND	ND	U	0.00889
Trichloroethylene	0.00290	0.02	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0201	0.08	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6051509-01

Sampled: 05/08/26 23:59

Pressure @ Receipt: 4.0

Canister #: 41636

Received: 05/15/26 11:51

Comments:

Analyzed: 05/23/26 03:22

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.678	2.56		0.0380
Dibromochloromethane	5.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.134	0.63		0.0258
Tetrachloroethylene	0.0154	0.11		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0823	0.36		0.0130
m,p-Xylene	0.278	1.21		0.0418
Bromoform	9.00E-4	0.01	U	0.0166
Styrene	0.0485	0.21		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.104	0.45		0.0235
1,3,5-Trimethylbenzene	0.0161	0.08	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	U, CE	0.0320
m-Dichlorobenzene	4.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00260	0.02	U	0.0287
o-Dichlorobenzene	5.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	8.00E-4	0.01	U	0.0792
Hexachloro-1,3-butadiene	4.00E-4	0.00	U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK D1

Lab ID: 6052820-01

Sampled: 05/14/26 23:59

Pressure @ Receipt: 4.5

Canister #: 41598

Received: 05/28/26 12:37

Comments: D2 is invalid 6051414-01

Analyzed: 06/08/26 20:26

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.477	2.36		0.0237
Chloromethane	0.512	1.06		0.0213
Dichlorotetrafluoroethane	0.0170	0.12		0.0112
Vinyl chloride	8.00E-4	0.00	U	0.0140
1,3-Butadiene	0.0164	0.04	U	0.0215
Ethylene oxide	0.0612	0.11		0.0276
Bromomethane	0.00840	0.03	U	0.0104
Chloroethane	0.0135	0.04	U	0.0136
Acetonitrile	72.5	122.00	CF, E, LK, QB-01	0.184
Acrolein	0.248	0.57		0.0977
Trichlorofluoromethane	0.222	1.25		0.00950
Acrylonitrile	0.0137	0.03		0.0116
1,1-Dichloroethene	ND	ND	LJ, U	0.0158
Dichloromethane	0.176	0.61	CF, LK	0.0489
Carbon Disulfide	0.00890	0.03	U	0.0473
Trichlorotrifluoroethane	0.0618	0.48		0.0210
trans-1,2-Dichloroethylene	0.00100	0.00	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	3.00E-4	0.00	U	0.00905
Chloroform	0.0210	0.10		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0207	0.08		0.0116
1,1,1-Trichloroethane	9.00E-4	0.00	U	0.00952
Benzene	0.316	1.01		0.0158
Carbon Tetrachloride	0.0761	0.48		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	6.00E-4	0.00	U	0.00889
Trichloroethylene	0.00220	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0126	0.05	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
ATTN: Mr. Bryan Sims
PHONE: (405) 702-0100

FILE #: 3508.00
REPORTED: 07/15/26 12:42
SUBMITTED: 05/07/26 to 06/15/26
AQS SITE CODE: 40 143 0235
SITE CODE: TOOK

Description: TOOK D1	Lab ID: 6052820-01	Sampled: 05/14/26 23:59
Pressure @ Receipt: 4.5	Canister #: 41598	Received: 05/28/26 12:37
Comments: D2 is invalid 6051414-01		Analyzed: 06/08/26 20:26

Air Toxics by EPA Compendium Method TO-15

<u>Analyte</u>	<u>Results</u>		<u>Flag</u>	<u>MDL</u>
	<u>ppbv</u>	<u>ug/m³</u>		<u>ppbv</u>
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.527	1.99		0.0380
Dibromochloromethane	4.00E-4	0.00	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0480	0.23		0.0258
Tetrachloroethylene	0.0132	0.09		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0666	0.29		0.0130
m,p-Xylene	0.216	0.94		0.0418
Bromoform	7.00E-4	0.01	U	0.0166
Styrene	0.0274	0.12		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0763	0.33		0.0235
1,3,5-Trimethylbenzene	0.0137	0.07	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	2.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00200	0.01	U	0.0287
o-Dichlorobenzene	3.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	4.00E-4	0.00	QB-01, U	0.0792
Hexachloro-1,3-butadiene	2.00E-4	0.00	QB-01, U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6061505-01

Sampled: 05/26/26 23:59

Pressure @ Receipt: 4.0

Canister #: 42161

Received: 06/15/26 10:15

Comments:

Analyzed: 06/17/26 04:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
Dichlorodifluoromethane	0.477	2.36		0.0237
Chloromethane	0.550	1.14		0.0213
Dichlorotetrafluoroethane	0.0178	0.13		0.0112
Vinyl chloride	ND	ND	U	0.0140
1,3-Butadiene	0.0136	0.03	CF, LL, U	0.0215
Ethylene oxide	0.0865	0.16		0.0276
Bromomethane	0.00860	0.03	U	0.0104
Chloroethane	0.00900	0.02	U	0.0136
Acetonitrile	55.9	94.00	E	0.184
Acrolein	0.294	0.68	CF, LK	0.0977
Trichlorofluoromethane	0.207	1.17		0.00950
Acrylonitrile	0.0153	0.03		0.0116
1,1-Dichloroethene	ND	ND	U	0.0158
Dichloromethane	0.125	0.44		0.0489
Carbon Disulfide	0.0163	0.05	U	0.0473
Trichlorotrifluoroethane	0.0720	0.55		0.0210
trans-1,2-Dichloroethylene	0.00170	0.01	U	0.0171
1,1-Dichloroethane	ND	ND	U	0.00655
Methyl tert-Butyl Ether	ND	ND	U	0.0101
Chloroprene	ND	ND	U	0.0123
cis-1,2-Dichloroethylene	ND	ND	U	0.0172
Bromochloromethane	5.00E-4	0.00	U	0.00905
Chloroform	0.0301	0.15		0.0134
Ethyl tert-Butyl Ether	ND	ND	U	0.00893
1,2-Dichloroethane	0.0163	0.07		0.0116
1,1,1-Trichloroethane	0.00160	0.01	U	0.00952
Benzene	0.274	0.88		0.0158
Carbon Tetrachloride	0.0729	0.46		0.00750
tert-Amyl Methyl Ether	ND	ND	U	0.0126
1,2-Dichloropropane	ND	ND	U	0.0215
Ethyl Acrylate	ND	ND	U	0.00780
Bromodichloromethane	0.00170	0.01	U	0.00889
Trichloroethylene	0.00260	0.01	U	0.0170
Methyl Methacrylate	ND	ND	U	0.0412
cis-1,3-Dichloropropene	ND	ND	U	0.00753
Methyl Isobutyl Ketone	0.0270	0.11	U	0.0578

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Description: TOOK

Lab ID: 6061505-01

Sampled: 05/26/26 23:59

Pressure @ Receipt: 4.0

Canister #: 42161

Received: 06/15/26 10:15

Comments:

Analyzed: 06/17/26 04:19

Air Toxics by EPA Compendium Method TO-15

Analyte	Results		Flag	MDL
	ppbv	ug/m ³		ppbv
trans-1,3-Dichloropropene	ND	ND	U	0.00822
1,1,2-Trichloroethane	ND	ND	U	0.00749
Toluene	0.660	2.49		0.0380
Dibromochloromethane	8.00E-4	0.01	U	0.00884
1,2-Dibromoethane	ND	ND	U	0.00660
n-Octane	0.0425	0.20		0.0258
Tetrachloroethylene	0.0104	0.07		0.00860
Chlorobenzene	ND	ND	U	0.0137
Ethylbenzene	0.0760	0.33		0.0130
m,p-Xylene	0.240	1.04		0.0418
Bromoform	0.00120	0.01	U	0.0166
Styrene	0.0423	0.18		0.0273
1,1,2,2-Tetrachloroethane	ND	ND	U	0.0190
o-Xylene	0.0882	0.38		0.0235
1,3,5-Trimethylbenzene	0.0140	0.07	U	0.0316
1,2,4-Trimethylbenzene	ND	ND	CE, U	0.0320
m-Dichlorobenzene	7.00E-4	0.00	U	0.0277
p-Dichlorobenzene	0.00330	0.02	U	0.0287
o-Dichlorobenzene	8.00E-4	0.00	U	0.0364
1,2,4-Trichlorobenzene	0.00150	0.01	QB-01, QB-03, U	0.0792
Hexachloro-1,3-butadiene	4.00E-4	0.00	QB-03, U	0.0701



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1908 - Summa Canister Prep

Blank (B6E1908-BLK1)

Prepared: 05/06/26 Analyzed: 05/19/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E1908 - Summa Canister Prep

Blank (B6E1908-BLK1) Continued

Prepared: 05/06/26 Analyzed: 05/19/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6E2204 - Summa Canister Prep

Blank (B6E2204-BLK1)

Prepared: 05/06/26 Analyzed: 05/22/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6E2204 - Summa Canister Prep

Blank (B6E2204-BLK1) Continued

Prepared: 05/06/26 Analyzed: 05/22/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				U
Hexachloro-1,3-butadiene	ND	ppbv				U

Batch B6F0803 - Summa Canister Prep

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0803 - Summa Canister Prep

Blank (B6F0803-BLK1)

Prepared: 06/02/26 Analyzed: 06/08/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				QB-01, U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				LJ, U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U
Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F0803 - Summa Canister Prep

Blank (B6F0803-BLK1) Continued

Prepared: 06/02/26 Analyzed: 06/08/26

Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				QB-01, U

Batch B6F1604 - Summa Canister Prep

Blank (B6F1604-BLK1)

Prepared: 06/05/26 Analyzed: 06/16/26

Dichlorodifluoromethane	ND	ppbv				U
Chloromethane	ND	ppbv				U
Dichlorotetrafluoroethane	ND	ppbv				U
Vinyl chloride	ND	ppbv				U
1,3-Butadiene	ND	ppbv				U
Ethylene oxide	ND	ppbv				U
Bromomethane	ND	ppbv				U
Chloroethane	ND	ppbv				U
Acetonitrile	ND	ppbv				U
Acrolein	ND	ppbv				U
Trichlorofluoromethane	ND	ppbv				U
Acrylonitrile	ND	ppbv				U
1,1-Dichloroethene	ND	ppbv				U
Dichloromethane	ND	ppbv				U
Carbon Disulfide	ND	ppbv				U
Trichlorotrifluoroethane	ND	ppbv				U
trans-1,2-Dichloroethylene	ND	ppbv				U
1,1-Dichloroethane	ND	ppbv				U
Methyl tert-Butyl Ether	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	Source Result	RPD	RPD Limit	Notes
---------	--------	-------	---------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Batch B6F1604 - Summa Canister Prep

Blank (B6F1604-BLK1) Continued

Prepared: 06/05/26 Analyzed: 06/16/26

Chloroprene	ND	ppbv				U
cis-1,2-Dichloroethylene	ND	ppbv				U
Bromochloromethane	ND	ppbv				U
Chloroform	ND	ppbv				U
Ethyl tert-Butyl Ether	ND	ppbv				U
1,2-Dichloroethane	ND	ppbv				U
1,1,1-Trichloroethane	ND	ppbv				U
Benzene	ND	ppbv				U
Carbon Tetrachloride	ND	ppbv				U
tert-Amyl Methyl Ether	ND	ppbv				U
1,2-Dichloropropane	ND	ppbv				U
Ethyl Acrylate	ND	ppbv				U
Bromodichloromethane	ND	ppbv				U
Trichloroethylene	ND	ppbv				U
Methyl Methacrylate	ND	ppbv				U
cis-1,3-Dichloropropene	ND	ppbv				U
Methyl Isobutyl Ketone	ND	ppbv				U
trans-1,3-Dichloropropene	ND	ppbv				U
1,1,2-Trichloroethane	ND	ppbv				U
Toluene	ND	ppbv				U
Dibromochloromethane	ND	ppbv				U
1,2-Dibromoethane	ND	ppbv				U
n-Octane	ND	ppbv				U
Tetrachloroethylene	ND	ppbv				U
Chlorobenzene	ND	ppbv				U
Ethylbenzene	ND	ppbv				U
m,p-Xylene	ND	ppbv				U
Bromoform	ND	ppbv				U
Styrene	ND	ppbv				U
1,1,2,2-Tetrachloroethane	ND	ppbv				U
o-Xylene	ND	ppbv				U
1,3,5-Trimethylbenzene	ND	ppbv				U
1,2,4-Trimethylbenzene	ND	ppbv				U
m-Dichlorobenzene	ND	ppbv				U
p-Dichlorobenzene	ND	ppbv				U
o-Dichlorobenzene	ND	ppbv				U
1,2,4-Trichlorobenzene	ND	ppbv				QB-01, U
Hexachloro-1,3-butadiene	ND	ppbv				U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV1)

Prepared & Analyzed: 05/19/26

Dichlorodifluoromethane	1.07	ppbv	4.1	30.00	QB-01
Chloromethane	1.06	ppbv	3.7	30.00	
Dichlorotetrafluoroethane	1.11	ppbv	7.1	30.00	
Vinyl chloride	0.998	ppbv	-3.2	30.00	
1,3-Butadiene	0.858	ppbv	-17.0	30.00	
Ethylene oxide	0.966	ppbv	1.1	30.00	
Bromomethane	1.12	ppbv	8.8	30.00	
Chloroethane	1.16	ppbv	11.7	30.00	
Acetonitrile	1.12	ppbv	7.6	30.00	
Acrolein	1.06	ppbv	3.8	30.00	
Trichlorofluoromethane	1.08	ppbv	7.1	30.00	
Acrylonitrile	1.13	ppbv	2.2	30.00	
1,1-Dichloroethene	1.11	ppbv	6.9	30.00	
Dichloromethane	1.14	ppbv	9.7	30.00	
Carbon Disulfide	1.38	ppbv	26.6	30.00	
Trichlorotrifluoroethane	1.21	ppbv	17.9	30.00	
trans-1,2-Dichloroethylene	1.07	ppbv	2.5	30.00	
1,1-Dichloroethane	1.14	ppbv	10.2	30.00	
Methyl tert-Butyl Ether	0.981	ppbv	-5.3	30.00	
Chloroprene	0.927	ppbv	-10.7	30.00	
cis-1,2-Dichloroethylene	0.965	ppbv	-6.8	30.00	
Bromochloromethane	1.00	ppbv	-2.7	30.00	
Chloroform	1.12	ppbv	6.8	30.00	
Ethyl tert-Butyl Ether	0.980	ppbv	-5.5	30.00	
1,2-Dichloroethane	1.07	ppbv	3.5	30.00	
1,1,1-Trichloroethane	1.06	ppbv	2.5	30.00	
Benzene	1.03	ppbv	-0.8	30.00	
Carbon Tetrachloride	1.03	ppbv	0.2	30.00	
tert-Amyl Methyl Ether	0.917	ppbv	-11.3	30.00	
1,2-Dichloropropane	1.23	ppbv	18.9	30.00	
Ethyl Acrylate	1.01	ppbv	-3.6	30.00	
Bromodichloromethane	1.22	ppbv	18.2	30.00	
Trichloroethylene	1.15	ppbv	11.1	30.00	
Methyl Methacrylate	1.03	ppbv	-1.2	30.00	
cis-1,3-Dichloropropene	1.07	ppbv	2.3	30.00	
Methyl Isobutyl Ketone	1.14	ppbv	9.1	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV1) Continued

Prepared & Analyzed: 05/19/26

trans-1,3-Dichloropropene	1.16	ppbv	10.8	30.00	
1,1,2-Trichloroethane	1.21	ppbv	16.0	30.00	
Toluene	1.08	ppbv	3.4	30.00	
Dibromochloromethane	1.21	ppbv	17.5	30.00	
1,2-Dibromoethane	1.12	ppbv	7.0	30.00	
n-Octane	1.04	ppbv	0.2	30.00	
Tetrachloroethylene	1.14	ppbv	8.9	30.00	
Chlorobenzene	1.11	ppbv	5.5	30.00	
Ethylbenzene	1.00	ppbv	-4.2	30.00	
m,p-Xylene	2.08	ppbv	-0.4	30.00	
Bromoform	1.15	ppbv	9.5	30.00	
Styrene	1.01	ppbv	-4.6	30.00	
1,1,2,2-Tetrachloroethane	1.24	ppbv	17.6	30.00	
o-Xylene	0.967	ppbv	-8.4	30.00	
1,3,5-Trimethylbenzene	0.871	ppbv	-17.2	30.00	
1,2,4-Trimethylbenzene	0.850	ppbv	-19.3	30.00	
m-Dichlorobenzene	1.08	ppbv	3.2	30.00	
p-Dichlorobenzene	1.05	ppbv	0.2	30.00	
o-Dichlorobenzene	1.05	ppbv	0.7	30.00	
1,2,4-Trichlorobenzene	0.958	ppbv	-7.7	30.00	QB-01
Hexachloro-1,3-butadiene	1.12	ppbv	6.6	30.00	

Calibration Check (2605061-CCV2)

Prepared: 05/19/26 Analyzed: 05/20/26

Dichlorodifluoromethane	1.12	ppbv	8.6	30.00	
Chloromethane	1.08	ppbv	5.6	30.00	
Dichlorotetrafluoroethane	1.14	ppbv	10.5	30.00	
Vinyl chloride	0.978	ppbv	-5.2	30.00	
1,3-Butadiene	0.798	ppbv	-22.8	30.00	
Ethylene oxide	0.948	ppbv	-0.9	30.00	
Bromomethane	1.17	ppbv	13.8	30.00	
Chloroethane	1.19	ppbv	14.8	30.00	
Acetonitrile	1.13	ppbv	9.0	30.00	QB-01
Acrolein	1.08	ppbv	6.1	30.00	
Trichlorofluoromethane	1.11	ppbv	9.8	30.00	
Acrylonitrile	1.14	ppbv	3.4	30.00	
1,1-Dichloroethene	1.14	ppbv	9.5	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV2) Continued

Prepared: 05/19/26 Analyzed: 05/20/26

Dichloromethane	1.15	ppbv	10.8	30.00
Carbon Disulfide	1.22	ppbv	12.4	30.00
Trichlorotrifluoroethane	1.08	ppbv	4.4	30.00
trans-1,2-Dichloroethylene	1.06	ppbv	2.4	30.00
1,1-Dichloroethane	1.05	ppbv	1.0	30.00
Methyl tert-Butyl Ether	0.963	ppbv	-7.0	30.00
Chloroprene	0.779	ppbv	-24.9	30.00
cis-1,2-Dichloroethylene	0.903	ppbv	-12.9	30.00
Bromochloromethane	0.999	ppbv	-2.9	30.00
Chloroform	1.13	ppbv	7.9	30.00
Ethyl tert-Butyl Ether	0.749	ppbv	-27.7	30.00
1,2-Dichloroethane	1.08	ppbv	4.2	30.00
1,1,1-Trichloroethane	1.06	ppbv	3.2	30.00
Benzene	1.00	ppbv	-3.7	30.00
Carbon Tetrachloride	1.02	ppbv	0.02	30.00
tert-Amyl Methyl Ether	0.864	ppbv	-16.4	30.00
1,2-Dichloropropane	1.25	ppbv	20.1	30.00
Ethyl Acrylate	0.980	ppbv	-6.3	30.00
Bromodichloromethane	1.27	ppbv	22.3	30.00
Trichloroethylene	1.16	ppbv	12.0	30.00
Methyl Methacrylate	0.997	ppbv	-4.3	30.00
cis-1,3-Dichloropropene	1.04	ppbv	-0.8	30.00
Methyl Isobutyl Ketone	1.12	ppbv	7.4	30.00
trans-1,3-Dichloropropene	1.15	ppbv	10.0	30.00
1,1,2-Trichloroethane	1.24	ppbv	18.4	30.00
Toluene	1.05	ppbv	0.7	30.00
Dibromochloromethane	1.23	ppbv	19.4	30.00
1,2-Dibromoethane	1.11	ppbv	6.8	30.00
n-Octane	1.02	ppbv	-2.2	30.00
Tetrachloroethylene	1.16	ppbv	11.4	30.00
Chlorobenzene	1.12	ppbv	6.1	30.00
Ethylbenzene	0.966	ppbv	-7.6	30.00
m,p-Xylene	2.05	ppbv	-2.0	30.00
Bromoform	1.18	ppbv	12.6	30.00
Styrene	0.994	ppbv	-6.0	30.00
1,1,2,2-Tetrachloroethane	1.30	ppbv	22.9	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605061

Calibration Check (2605061-CCV2) Continued

Prepared: 05/19/26 Analyzed: 05/20/26

o-Xylene	0.934	ppbv	-11.5	30.00	QB-01
1,3,5-Trimethylbenzene	0.829	ppbv	-21.2	30.00	
1,2,4-Trimethylbenzene	0.810	ppbv	-23.2	30.00	
m-Dichlorobenzene	1.09	ppbv	4.6	30.00	
p-Dichlorobenzene	1.05	ppbv	0.7	30.00	
o-Dichlorobenzene	1.06	ppbv	1.7	30.00	
1,2,4-Trichlorobenzene	0.962	ppbv	-7.3	30.00	
Hexachloro-1,3-butadiene	1.13	ppbv	7.6	30.00	

Sequence 2605075

Calibration Check (2605075-CCV1)

Prepared & Analyzed: 05/22/26

Dichlorodifluoromethane	1.14	ppbv	11.3	30.00
Chloromethane	1.10	ppbv	8.1	30.00
Dichlorotetrafluoroethane	1.14	ppbv	10.3	30.00
Vinyl chloride	1.13	ppbv	9.1	30.00
1,3-Butadiene	1.09	ppbv	5.6	30.00
Ethylene oxide	0.988	ppbv	3.3	30.00
Bromomethane	1.13	ppbv	9.8	30.00
Chloroethane	1.13	ppbv	9.3	30.00
Acetonitrile	1.03	ppbv	-0.7	30.00
Acrolein	1.02	ppbv	-0.2	30.00
Trichlorofluoromethane	1.15	ppbv	13.6	30.00
Acrylonitrile	1.17	ppbv	6.0	30.00
1,1-Dichloroethene	1.06	ppbv	2.4	30.00
Dichloromethane	1.12	ppbv	8.5	30.00
Carbon Disulfide	0.949	ppbv	-12.8	30.00
Trichlorotrifluoroethane	1.03	ppbv	-0.4	30.00
trans-1,2-Dichloroethylene	1.16	ppbv	11.5	30.00
1,1-Dichloroethane	1.14	ppbv	10.3	30.00
Methyl tert-Butyl Ether	1.22	ppbv	17.4	30.00
Chloroprene	1.11	ppbv	7.0	30.00
cis-1,2-Dichloroethylene	1.14	ppbv	9.9	30.00
Bromochloromethane	1.15	ppbv	12.1	30.00
Chloroform	1.17	ppbv	11.7	30.00
Ethyl tert-Butyl Ether	1.21	ppbv	16.9	30.00
1,2-Dichloroethane	1.16	ppbv	11.7	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605075

Calibration Check (2605075-CCV1) Continued

Prepared & Analyzed: 05/22/26

1,1,1-Trichloroethane	1.18	ppbv	14.5	30.00
Benzene	1.14	ppbv	9.6	30.00
Carbon Tetrachloride	1.17	ppbv	13.8	30.00
tert-Amyl Methyl Ether	1.21	ppbv	17.3	30.00
1,2-Dichloropropane	1.12	ppbv	8.1	30.00
Ethyl Acrylate	1.09	ppbv	4.5	30.00
Bromodichloromethane	1.12	ppbv	7.7	30.00
Trichloroethylene	1.10	ppbv	6.6	30.00
Methyl Methacrylate	1.07	ppbv	2.8	30.00
cis-1,3-Dichloropropene	1.12	ppbv	6.9	30.00
Methyl Isobutyl Ketone	1.15	ppbv	10.1	30.00
trans-1,3-Dichloropropene	1.14	ppbv	8.7	30.00
1,1,2-Trichloroethane	1.14	ppbv	9.3	30.00
Toluene	1.13	ppbv	8.8	30.00
Dibromochloromethane	1.13	ppbv	9.4	30.00
1,2-Dibromoethane	1.14	ppbv	9.1	30.00
n-Octane	1.10	ppbv	5.6	30.00
Tetrachloroethylene	1.13	ppbv	8.3	30.00
Chlorobenzene	1.13	ppbv	7.7	30.00
Ethylbenzene	1.14	ppbv	8.6	30.00
m,p-Xylene	2.29	ppbv	9.7	30.00
Bromoform	1.14	ppbv	8.7	30.00
Styrene	1.15	ppbv	8.7	30.00
1,1,2,2-Tetrachloroethane	1.14	ppbv	8.0	30.00
o-Xylene	1.16	ppbv	9.7	30.00
1,3,5-Trimethylbenzene	1.14	ppbv	7.9	30.00
1,2,4-Trimethylbenzene	1.15	ppbv	9.3	30.00
m-Dichlorobenzene	1.14	ppbv	9.6	30.00
p-Dichlorobenzene	1.15	ppbv	9.8	30.00
o-Dichlorobenzene	1.15	ppbv	9.8	30.00
1,2,4-Trichlorobenzene	1.15	ppbv	10.4	30.00
Hexachloro-1,3-butadiene	1.13	ppbv	7.5	30.00

Calibration Check (2605075-CCV2)

Prepared: 05/22/26 Analyzed: 05/23/26

Dichlorodifluoromethane	1.11	ppbv	8.2	30.00
Chloromethane	1.09	ppbv	6.7	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605075

Calibration Check (2605075-CCV2) Continued

Prepared: 05/22/26 Analyzed: 05/23/26

Dichlorotetrafluoroethane	1.11	ppbv	6.8	30.00
Vinyl chloride	1.05	ppbv	1.7	30.00
1,3-Butadiene	0.984	ppbv	-4.9	30.00
Ethylene oxide	0.987	ppbv	3.2	30.00
Bromomethane	1.12	ppbv	9.0	30.00
Chloroethane	1.09	ppbv	5.5	30.00
Acetonitrile	1.07	ppbv	3.3	30.00
Acrolein	0.996	ppbv	-2.4	30.00
Trichlorofluoromethane	1.08	ppbv	6.6	30.00
Acrylonitrile	1.13	ppbv	2.7	30.00
1,1-Dichloroethene	1.06	ppbv	1.8	30.00
Dichloromethane	1.08	ppbv	4.5	30.00
Carbon Disulfide	0.948	ppbv	-12.9	30.00
Trichlorotrifluoroethane	0.978	ppbv	-5.1	30.00
trans-1,2-Dichloroethylene	1.07	ppbv	2.7	30.00
1,1-Dichloroethane	1.08	ppbv	4.0	30.00
Methyl tert-Butyl Ether	1.04	ppbv	0.8	30.00
Chloroprene	1.03	ppbv	-0.6	30.00
cis-1,2-Dichloroethylene	1.07	ppbv	3.1	30.00
Bromochloromethane	1.05	ppbv	1.9	30.00
Chloroform	1.13	ppbv	7.5	30.00
Ethyl tert-Butyl Ether	1.03	ppbv	-0.3	30.00
1,2-Dichloroethane	1.10	ppbv	6.6	30.00
1,1,1-Trichloroethane	1.10	ppbv	6.5	30.00
Benzene	1.08	ppbv	4.3	30.00
Carbon Tetrachloride	1.11	ppbv	8.3	30.00
tert-Amyl Methyl Ether	1.02	ppbv	-1.1	30.00
1,2-Dichloropropane	1.03	ppbv	-0.4	30.00
Ethyl Acrylate	1.06	ppbv	1.5	30.00
Bromodichloromethane	1.14	ppbv	10.1	30.00
Trichloroethylene	1.12	ppbv	8.2	30.00
Methyl Methacrylate	1.09	ppbv	4.7	30.00
cis-1,3-Dichloropropene	1.06	ppbv	1.4	30.00
Methyl Isobutyl Ketone	1.15	ppbv	10.6	30.00
trans-1,3-Dichloropropene	1.09	ppbv	3.4	30.00
1,1,2-Trichloroethane	1.14	ppbv	8.5	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2605075

Calibration Check (2605075-CCV2) Continued

Prepared: 05/22/26 Analyzed: 05/23/26

Toluene	1.12	ppbv	7.6	30.00
Dibromochloromethane	1.08	ppbv	4.9	30.00
1,2-Dibromoethane	1.11	ppbv	6.0	30.00
n-Octane	1.15	ppbv	10.8	30.00
Tetrachloroethylene	1.15	ppbv	9.9	30.00
Chlorobenzene	1.18	ppbv	11.8	30.00
Ethylbenzene	1.10	ppbv	5.5	30.00
m,p-Xylene	2.26	ppbv	8.5	30.00
Bromoform	1.14	ppbv	8.7	30.00
Styrene	1.16	ppbv	9.6	30.00
1,1,2,2-Tetrachloroethane	1.18	ppbv	12.1	30.00
o-Xylene	1.13	ppbv	6.8	30.00
1,3,5-Trimethylbenzene	1.08	ppbv	2.5	30.00
1,2,4-Trimethylbenzene	1.10	ppbv	4.2	30.00
m-Dichlorobenzene	1.09	ppbv	4.8	30.00
p-Dichlorobenzene	1.09	ppbv	4.4	30.00
o-Dichlorobenzene	1.09	ppbv	4.0	30.00
1,2,4-Trichlorobenzene	1.05	ppbv	1.6	30.00
Hexachloro-1,3-butadiene	1.11	ppbv	4.9	30.00

Sequence 2606028

Calibration Check (2606028-CCV1)

Prepared & Analyzed: 06/08/26

Dichlorodifluoromethane	1.17	ppbv	13.9	30.00
Chloromethane	1.08	ppbv	5.9	30.00
Dichlorotetrafluoroethane	1.15	ppbv	11.2	30.00
Vinyl chloride	1.11	ppbv	7.2	30.00
1,3-Butadiene	1.06	ppbv	2.6	30.00
Ethylene oxide	0.988	ppbv	3.3	30.00
Bromomethane	1.20	ppbv	16.0	30.00
Chloroethane	1.18	ppbv	13.9	30.00
Acetonitrile	1.05	ppbv	1.2	30.00
Acrolein	1.09	ppbv	7.3	30.00
Trichlorofluoromethane	1.21	ppbv	19.4	30.00
Acrylonitrile	1.14	ppbv	3.5	30.00
1,1-Dichloroethene	1.08	ppbv	4.3	30.00
Dichloromethane	1.12	ppbv	8.0	30.00

QB-01

LJ

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606028

Calibration Check (2606028-CCV1) Continued

Prepared & Analyzed: 06/08/26

Carbon Disulfide	1.03	ppbv	-5.3	30.00
Trichlorotrifluoroethane	1.06	ppbv	2.6	30.00
trans-1,2-Dichloroethylene	1.16	ppbv	11.3	30.00
1,1-Dichloroethane	1.16	ppbv	11.6	30.00
Methyl tert-Butyl Ether	1.21	ppbv	16.9	30.00
Chloroprene	1.04	ppbv	0.7	30.00
cis-1,2-Dichloroethylene	1.14	ppbv	9.7	30.00
Bromochloromethane	1.20	ppbv	16.4	30.00
Chloroform	1.18	ppbv	12.3	30.00
Ethyl tert-Butyl Ether	1.19	ppbv	14.7	30.00
1,2-Dichloroethane	1.19	ppbv	14.9	30.00
1,1,1-Trichloroethane	1.21	ppbv	17.9	30.00
Benzene	1.12	ppbv	7.6	30.00
Carbon Tetrachloride	1.23	ppbv	20.4	30.00
tert-Amyl Methyl Ether	1.21	ppbv	16.7	30.00
1,2-Dichloropropane	1.15	ppbv	11.1	30.00
Ethyl Acrylate	1.03	ppbv	-2.0	30.00
Bromodichloromethane	1.12	ppbv	8.5	30.00
Trichloroethylene	1.13	ppbv	9.1	30.00
Methyl Methacrylate	0.995	ppbv	-4.5	30.00
cis-1,3-Dichloropropene	1.06	ppbv	1.5	30.00
Methyl Isobutyl Ketone	1.08	ppbv	3.6	30.00
trans-1,3-Dichloropropene	1.13	ppbv	7.6	30.00
1,1,2-Trichloroethane	1.14	ppbv	9.0	30.00
Toluene	1.09	ppbv	4.6	30.00
Dibromochloromethane	1.14	ppbv	10.7	30.00
1,2-Dibromoethane	1.14	ppbv	8.8	30.00
n-Octane	1.03	ppbv	-0.6	30.00
Tetrachloroethylene	1.17	ppbv	12.1	30.00
Chlorobenzene	1.15	ppbv	9.5	30.00
Ethylbenzene	1.09	ppbv	4.4	30.00
m,p-Xylene	2.31	ppbv	10.4	30.00
Bromoform	1.15	ppbv	10.2	30.00
Styrene	1.16	ppbv	9.9	30.00
1,1,2,2-Tetrachloroethane	1.12	ppbv	6.5	30.00
o-Xylene	1.10	ppbv	4.1	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606028

Calibration Check (2606028-CCV1) Continued

Prepared & Analyzed: 06/08/26

1,3,5-Trimethylbenzene	1.08	ppbv	2.8	30.00	
1,2,4-Trimethylbenzene	1.11	ppbv	5.0	30.00	
m-Dichlorobenzene	1.19	ppbv	14.3	30.00	
p-Dichlorobenzene	1.19	ppbv	14.0	30.00	
o-Dichlorobenzene	1.20	ppbv	14.5	30.00	
1,2,4-Trichlorobenzene	1.19	ppbv	14.7	30.00	QB-01
Hexachloro-1,3-butadiene	1.26	ppbv	19.7	30.00	QB-01

Calibration Check (2606028-CCV2)

Prepared: 06/08/26 Analyzed: 06/09/26

Dichlorodifluoromethane	1.14	ppbv	10.4	30.00	
Chloromethane	1.07	ppbv	5.4	30.00	
Dichlorotetrafluoroethane	1.10	ppbv	6.6	30.00	
Vinyl chloride	1.02	ppbv	-1.4	30.00	
1,3-Butadiene	0.896	ppbv	-13.3	30.00	
Ethylene oxide	1.04	ppbv	8.4	30.00	
Bromomethane	1.17	ppbv	13.8	30.00	
Chloroethane	1.22	ppbv	17.7	30.00	
Acetonitrile	1.09	ppbv	4.8	30.00	QB-01
Acrolein	1.15	ppbv	12.5	30.00	
Trichlorofluoromethane	1.19	ppbv	18.1	30.00	
Acrylonitrile	1.13	ppbv	1.9	30.00	
1,1-Dichloroethene	1.13	ppbv	8.6	30.00	LJ
Dichloromethane	1.12	ppbv	7.8	30.00	
Carbon Disulfide	1.34	ppbv	22.8	30.00	
Trichlorotrifluoroethane	1.06	ppbv	2.6	30.00	
trans-1,2-Dichloroethylene	1.08	ppbv	3.5	30.00	
1,1-Dichloroethane	1.14	ppbv	9.8	30.00	
Methyl tert-Butyl Ether	1.12	ppbv	8.0	30.00	
Chloroprene	1.00	ppbv	-3.2	30.00	
cis-1,2-Dichloroethylene	1.10	ppbv	6.5	30.00	
Bromochloromethane	1.15	ppbv	12.2	30.00	
Chloroform	1.17	ppbv	11.5	30.00	
Ethyl tert-Butyl Ether	1.10	ppbv	6.3	30.00	
1,2-Dichloroethane	1.18	ppbv	14.5	30.00	
1,1,1-Trichloroethane	1.18	ppbv	14.8	30.00	
Benzene	1.08	ppbv	4.3	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606028

Calibration Check (2606028-CCV2) Continued

Prepared: 06/08/26 Analyzed: 06/09/26

Carbon Tetrachloride	1.18	ppbv	15.3	30.00	
tert-Amyl Methyl Ether	1.12	ppbv	8.0	30.00	
1,2-Dichloropropane	1.04	ppbv	0.4	30.00	
Ethyl Acrylate	1.02	ppbv	-3.0	30.00	
Bromodichloromethane	1.11	ppbv	7.5	30.00	
Trichloroethylene	1.08	ppbv	4.8	30.00	
Methyl Methacrylate	1.01	ppbv	-3.2	30.00	
cis-1,3-Dichloropropene	1.01	ppbv	-3.5	30.00	
Methyl Isobutyl Ketone	1.08	ppbv	3.9	30.00	
trans-1,3-Dichloropropene	1.10	ppbv	4.5	30.00	
1,1,2-Trichloroethane	1.11	ppbv	6.3	30.00	
Toluene	1.05	ppbv	1.0	30.00	
Dibromochloromethane	1.11	ppbv	7.7	30.00	
1,2-Dibromoethane	1.08	ppbv	3.3	30.00	
n-Octane	1.04	ppbv	0.1	30.00	
Tetrachloroethylene	1.10	ppbv	5.5	30.00	
Chlorobenzene	1.13	ppbv	7.6	30.00	
Ethylbenzene	1.04	ppbv	-0.4	30.00	
m,p-Xylene	2.24	ppbv	7.2	30.00	
Bromoform	1.13	ppbv	7.6	30.00	
Styrene	1.15	ppbv	8.8	30.00	
1,1,2,2-Tetrachloroethane	1.14	ppbv	8.5	30.00	
o-Xylene	1.06	ppbv	0.5	30.00	
1,3,5-Trimethylbenzene	1.01	ppbv	-3.7	30.00	
1,2,4-Trimethylbenzene	1.04	ppbv	-1.3	30.00	
m-Dichlorobenzene	1.14	ppbv	9.1	30.00	
p-Dichlorobenzene	1.13	ppbv	8.0	30.00	
o-Dichlorobenzene	1.13	ppbv	8.2	30.00	
1,2,4-Trichlorobenzene	1.08	ppbv	4.5	30.00	QB-01
Hexachloro-1,3-butadiene	1.20	ppbv	13.8	30.00	QB-01

Sequence 2606064

Calibration Check (2606064-CCV1)

Prepared & Analyzed: 06/16/26

Dichlorodifluoromethane	1.07	ppbv	4.2	30.00	
Chloromethane	1.09	ppbv	7.2	30.00	
Dichlorotetrafluoroethane	1.08	ppbv	4.6	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606064

Calibration Check (2606064-CCV1) Continued

Prepared & Analyzed: 06/16/26

Vinyl chloride	1.10	ppbv	6.3	30.00
1,3-Butadiene	1.11	ppbv	7.2	30.00
Ethylene oxide	0.938	ppbv	-1.9	30.00
Bromomethane	1.09	ppbv	5.9	30.00
Chloroethane	1.12	ppbv	7.9	30.00
Acetonitrile	1.16	ppbv	12.0	30.00
Acrolein	1.08	ppbv	5.6	30.00
Trichlorofluoromethane	1.04	ppbv	3.2	30.00
Acrylonitrile	1.20	ppbv	8.3	30.00
1,1-Dichloroethene	1.14	ppbv	9.4	30.00
Dichloromethane	1.14	ppbv	10.5	30.00
Carbon Disulfide	1.17	ppbv	7.7	30.00
Trichlorotrifluoroethane	1.08	ppbv	5.3	30.00
trans-1,2-Dichloroethylene	1.11	ppbv	7.0	30.00
1,1-Dichloroethane	1.13	ppbv	9.1	30.00
Methyl tert-Butyl Ether	1.12	ppbv	8.1	30.00
Chloroprene	1.13	ppbv	9.2	30.00
cis-1,2-Dichloroethylene	1.13	ppbv	9.2	30.00
Bromochloromethane	1.07	ppbv	3.9	30.00
Chloroform	1.12	ppbv	6.7	30.00
Ethyl tert-Butyl Ether	1.14	ppbv	9.9	30.00
1,2-Dichloroethane	1.09	ppbv	5.3	30.00
1,1,1-Trichloroethane	1.09	ppbv	5.5	30.00
Benzene	1.09	ppbv	5.0	30.00
Carbon Tetrachloride	1.08	ppbv	5.6	30.00
tert-Amyl Methyl Ether	1.12	ppbv	8.2	30.00
1,2-Dichloropropane	1.19	ppbv	14.6	30.00
Ethyl Acrylate	1.22	ppbv	16.2	30.00
Bromodichloromethane	1.17	ppbv	13.4	30.00
Trichloroethylene	1.11	ppbv	7.8	30.00
Methyl Methacrylate	1.22	ppbv	17.0	30.00
cis-1,3-Dichloropropene	1.21	ppbv	15.9	30.00
Methyl Isobutyl Ketone	1.16	ppbv	11.2	30.00
trans-1,3-Dichloropropene	1.20	ppbv	14.4	30.00
1,1,2-Trichloroethane	1.15	ppbv	10.4	30.00
Toluene	1.19	ppbv	14.8	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606064

Calibration Check (2606064-CCV1) Continued

Prepared & Analyzed: 06/16/26

Dibromochloromethane	1.17	ppbv	13.1	30.00	
1,2-Dibromoethane	1.18	ppbv	13.1	30.00	
n-Octane	1.20	ppbv	15.1	30.00	
Tetrachloroethylene	1.15	ppbv	10.6	30.00	
Chlorobenzene	1.14	ppbv	7.9	30.00	
Ethylbenzene	1.13	ppbv	8.4	30.00	
m,p-Xylene	2.33	ppbv	11.5	30.00	
Bromoform	1.11	ppbv	6.3	30.00	
Styrene	1.18	ppbv	11.2	30.00	
1,1,2,2-Tetrachloroethane	1.16	ppbv	9.8	30.00	
o-Xylene	1.15	ppbv	8.5	30.00	
1,3,5-Trimethylbenzene	1.10	ppbv	4.1	30.00	
1,2,4-Trimethylbenzene	1.10	ppbv	3.9	30.00	
m-Dichlorobenzene	1.11	ppbv	6.6	30.00	
p-Dichlorobenzene	1.11	ppbv	5.9	30.00	
o-Dichlorobenzene	1.11	ppbv	6.2	30.00	
1,2,4-Trichlorobenzene	1.02	ppbv	-1.9	30.00	QB-01
Hexachloro-1,3-butadiene	1.09	ppbv	3.7	30.00	

Calibration Check (2606064-CCV2)

Prepared: 06/16/26 Analyzed: 06/17/26

Dichlorodifluoromethane	1.13	ppbv	10.3	30.00	
Chloromethane	1.19	ppbv	16.7	30.00	
Dichlorotetrafluoroethane	1.16	ppbv	11.8	30.00	
Vinyl chloride	1.12	ppbv	8.7	30.00	
1,3-Butadiene	1.01	ppbv	-2.0	30.00	
Ethylene oxide	1.02	ppbv	6.5	30.00	
Bromomethane	1.13	ppbv	9.5	30.00	
Chloroethane	1.18	ppbv	13.8	30.00	
Acetonitrile	1.25	ppbv	20.3	30.00	
Acrolein	1.01	ppbv	-0.7	30.00	
Trichlorofluoromethane	1.06	ppbv	5.4	30.00	
Acrylonitrile	1.20	ppbv	8.7	30.00	
1,1-Dichloroethene	1.12	ppbv	8.0	30.00	
Dichloromethane	1.19	ppbv	14.7	30.00	
Carbon Disulfide	1.19	ppbv	9.8	30.00	
Trichlorotrifluoroethane	1.13	ppbv	9.8	30.00	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606064

Calibration Check (2606064-CCV2) Continued

Prepared: 06/16/26 Analyzed: 06/17/26

trans-1,2-Dichloroethylene	1.10	ppbv	5.7	30.00
1,1-Dichloroethane	1.16	ppbv	11.7	30.00
Methyl tert-Butyl Ether	1.05	ppbv	1.3	30.00
Chloroprene	1.07	ppbv	3.5	30.00
cis-1,2-Dichloroethylene	1.13	ppbv	9.4	30.00
Bromochloromethane	1.11	ppbv	7.9	30.00
Chloroform	1.15	ppbv	10.1	30.00
Ethyl tert-Butyl Ether	1.07	ppbv	3.3	30.00
1,2-Dichloroethane	1.12	ppbv	8.2	30.00
1,1,1-Trichloroethane	1.11	ppbv	7.6	30.00
Benzene	1.13	ppbv	8.5	30.00
Carbon Tetrachloride	1.09	ppbv	6.3	30.00
tert-Amyl Methyl Ether	1.05	ppbv	1.8	30.00
1,2-Dichloropropane	1.16	ppbv	12.2	30.00
Ethyl Acrylate	1.13	ppbv	8.1	30.00
Bromodichloromethane	1.13	ppbv	9.4	30.00
Trichloroethylene	1.09	ppbv	5.6	30.00
Methyl Methacrylate	1.12	ppbv	8.0	30.00
cis-1,3-Dichloropropene	1.02	ppbv	-2.7	30.00
Methyl Isobutyl Ketone	1.04	ppbv	-0.5	30.00
trans-1,3-Dichloropropene	1.11	ppbv	5.8	30.00
1,1,2-Trichloroethane	1.16	ppbv	10.9	30.00
Toluene	1.02	ppbv	-2.1	30.00
Dibromochloromethane	1.10	ppbv	6.4	30.00
1,2-Dibromoethane	1.09	ppbv	4.5	30.00
n-Octane	1.05	ppbv	0.8	30.00
Tetrachloroethylene	1.07	ppbv	2.2	30.00
Chlorobenzene	1.14	ppbv	8.8	30.00
Ethylbenzene	1.09	ppbv	4.1	30.00
m,p-Xylene	2.31	ppbv	10.5	30.00
Bromoform	1.16	ppbv	10.5	30.00
Styrene	1.17	ppbv	10.8	30.00
1,1,2,2-Tetrachloroethane	1.27	ppbv	20.5	30.00
o-Xylene	1.11	ppbv	4.7	30.00
1,3,5-Trimethylbenzene	1.04	ppbv	-0.9	30.00
1,2,4-Trimethylbenzene	1.06	ppbv	0.6	30.00

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Analyte	Result	Units	% Difference	Limit (%)	Notes
---------	--------	-------	--------------	-----------	-------

Air Toxics by EPA Compendium Method TO-15 - Quality Control

Sequence 2606064

Calibration Check (2606064-CCV2) Continued

Prepared: 06/16/26 Analyzed: 06/17/26

m-Dichlorobenzene	1.14	ppbv	9.3	30.00	
p-Dichlorobenzene	1.12	ppbv	7.1	30.00	
o-Dichlorobenzene	1.13	ppbv	8.1	30.00	
1,2,4-Trichlorobenzene	1.02	ppbv	-1.3	30.00	QB-01
Hexachloro-1,3-butadiene	1.11	ppbv	5.6	30.00	



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 07/15/26 12:42

SUBMITTED: 05/07/26 to 06/15/26

AQS SITE CODE: 40 143 0235

SITE CODE: TOOK

Notes and Definitions

U	Under Detection Limit
QB-03	Analyte exceeds sampling media lot blank criteria.
QB-01	Analyte exceeds method blank criteria
LL	Analyte identified; Reported value may be biased low.
LK	Analyte identified; Reported value may be biased high.
LJ	Identification of analyte is acceptable; reported value is an estimate.
E	The concentration for this analyte is an estimated value above the calibration range of the instrument.
CF	Canister bias: NATTS/UATMP data for compounds that have failed certification for the canister
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: This test is accredited under the 2016 TNI Standard; however the following analytes are not accredited: acetylene, bromodichloroethane, dichlorotetrafluoromethane, ethyl tert butyl ether, n-octane, tert amyl methyl ether, trichlorofluoroethane, and bromochloromethane.