

# CERTIFICATE of CALIBRATION for LI-COR SENSOR

Pyranometer  
Model Number: LI-200

Serial Number: **PY103193**

Calibration Date: **April 4, 2016**  
Manufacture Date: **April 4, 2016**

## Calibration Constant(s):

---

Output: 52.58 microamps per 1000 watts  $\text{m}^{-2}$

## For use with LI-COR handheld meters and loggers:

---

Multiplier: -19.02 watts  $\text{m}^{-2}$  per microamp

## If this is an LI-200-BL (3-wire bare leads):

---

Multiplier: 19.02 watts  $\text{m}^{-2}$  per microamp

## For use with LI-COR 2220 (147 ohm) millivolt Adapter:

---

Multiplier: -129.37 watts  $\text{m}^{-2}$  per millivolt

## If this is an SL or SMV sensor:

---

Multiplier: -100 watts  $\text{m}^{-2}$  per millivolt

**IMPORTANT:** Read the appropriate instruction manual (<https://www.licor.com/support/LI-200R/manuals.html>) before using this sensor

**IMPORTANT:** It is recommended that sensors be recalibrated every two years after field deployment.

Calibration Technician: John Brooks

Calibrations are traceable to the World Radiometric Reference through Secondary Standard Thermopile radiometers, which are calibrated annually by the National Renewable Energy Laboratory.



LI-COR Biosciences, Global Headquarters • 4647 Superior Street • Lincoln, NE 68504 USA  
Phone: +1-402-467-3576 • Fax: +1-402-467-2819 • Toll-free: 800-447-3576 (USA & Canada)  
[envsales@licor.com](mailto:envsales@licor.com) • [envsupport@licor.com](mailto:envsupport@licor.com) • [www.licor.com](http://www.licor.com)  
Manual: <https://www.licor.com/support/LI-200R/manuals.html>