

CERTIFICATE of CALIBRATION for LI-COR SENSOR

Pyranometer
Model Number: LI-200

Serial Number: **PY105686**

Calibration Date: **August 14, 2017**

Manufacture Date: **July 21, 2016**

Calibration Constant(s):

Output: 59.82 microamps per 1000 watts m^{-2}

For use with LI-COR handheld meters and loggers:

Multiplier: -16.72 watts m^{-2} per microamp

If this is an LI-200-BL (3-wire bare leads):

Multiplier: 16.72 watts m^{-2} per microamp

For use with LI-COR 2220 (147 ohm) millivolt Adapter:

Multiplier: -113.72 watts m^{-2} per millivolt

If this is an SL or SMV sensor:

Multiplier: -100 watts m^{-2} per millivolt

IMPORTANT: Read the appropriate instruction manual (<https://www.licor.com/support/LI-200R/manuals.html>) before using this sensor

IMPORTANT: It is recommended that sensors be recalibrated every two years after field deployment.

Calibration Technician: John Brooks

Calibrations are traceable to the World Radiometric Reference through Secondary Standard Thermopile radiometers, which are calibrated annually by the National Renewable Energy Laboratory.



LI-COR Biosciences, Global Headquarters • 4647 Superior Street • Lincoln, NE 68504 USA
Phone: +1-402-467-3576 • Fax: +1-402-467-2819 • Toll-free: 800-447-3576 (USA & Canada)
envsales@licor.com • envsupport@licor.com • www.licor.com
Manual: <https://www.licor.com/support/LI-200R/manuals.html>