

CERTIFICATE of CALIBRATION for LI-COR SENSOR

Pyranometer
Model Number: LI-200

Serial Number: **PY75566**

Calibration Date: **June 27, 2011**
Manufacture Date: **June 27, 2011**

Calibration Constant(s):

Output: 102.25 microamps per 1000 watts m^{-2}

For use with LI-COR handheld meters and loggers:

Multiplier: -9.78 watts m^{-2} per microamp

If this is an LI-200-BL (3-wire bare leads):

Multiplier: 9.78 watts m^{-2} per microamp

For use with LI-COR 2220 (147 ohm) millivolt Adapter:

Multiplier: -66.53 watts m^{-2} per millivolt

If this is an SL or SMV sensor:

Multiplier: -100 watts m^{-2} per millivolt

IMPORTANT: Read the appropriate instruction manual (<https://www.licor.com/support/LI-200R/manuals.html>) before using this sensor

IMPORTANT: It is recommended that sensors be recalibrated every two years after field deployment.

Calibration Technician: John Brooks

Calibrations are traceable to the World Radiometric Reference through Secondary Standard Thermopile radiometers, which are calibrated annually by the National Renewable Energy Laboratory.



LI-COR Biosciences, Global Headquarters • 4647 Superior Street • Lincoln, NE 68504 USA
Phone: +1-402-467-3576 • Fax: +1-402-467-2819 • Toll-free: 800-447-3576 (USA & Canada)
envsales@licor.com • envsupport@licor.com • www.licor.com
Manual: <https://www.licor.com/support/LI-200R/manuals.html>