

General

Name of Run	230615_TKN.run	Name of Analysis	TKN.ANL
Date of Report	6/15/2023	System	AA500
Run Start	6/15/2023 3:22:44 PM	Run Stop	4:44:30 PM
Operator Run/Rec. CLMR		Software Version	AACE 8.02 SP2
Comment			

Parameters

Channel	1		
Method	TKN	Dilution	
Unit	mg/L		
Calib. Fit	Linear		
Corr.Coeff.(r)	0.9962		
AUFS/Gain	0.14/74	Data	
Offset	-341	Field	
Sensitivity	0.1110	'#2	
Sample Limit 1	-		
Sample Limit 2	-		

Results

Pk	Cup	Type	Sample ID	Value	Data	Value
0	0	B	Initial Base	-0.097	0	
1	921	P	Primer	2.257	1	
2	921	D	Drift	2.258	1	
3	921	C	2.24	2.256	1	
4	901	C	1.68	1.711	1	
5	902	C	1.12	1.089	1	
6	903	C	0.56	0.498	1	
7	904	C	0.224	0.141	1	
8	905	C	Blank	0.128	1	
9	0	B	Baseline	-0.097	1	
10	921	H1	High	2.253	1	
11	904	L1	Low	0.142	1	
12	904	L1	Low	0.142	1	
13	906	R	Recovery std N02 (1)	0.052	1	
14	907	R	Recovery std N03 (1)	0.134	1	
15	908	QC1	CRM	0.074	1	
16	909	QC2	CRM	0.031	1	
17	1	S	1a	-0.052	1	
18	2	S	1	0.121	1	
19	3	S	3	0.163	1	
20	4	S	4	0.152	1	
21	5	S	5	0.137	1	
22	6	S	6	0.237	1	
23	7	S	20	0.128	1	
24	8	S	22	0.078	1	
25	9	S	24	0.103	1	
26	10	S	26	0.058	1	
27	11	K	26 spiked	0.222	1	
28	908	QC1	CRM	0.067	1	
29	909	QC2	CRM	0.073	1	
30	12	S	28	0.034	1	
31	13	S	30	0.035	1	
32	14	S	38	0.178	1	
33	15	S	39	0.084	1	
34	16	S	40	0.062	1	
35	17	S	41	0.143	1	
36	18	S	42	0.135	1	
37	19	S	43	0.066	1	
38	20	S	44	0.085	1	
39	21	K	44 spiked	0.297	1	
40	908	QC1	CRM	0.063	1	
41	909	QC2	CRM	0.068	1	
42	22	S	45	0.084	1	

43	23	S	46	0.054	1
44	24	S	47	0.029	1
45	25	S	48	0.031	1
46	26	S	50	-0.078	1
47	27	S	51	0.005	1
48	28	S	52	0.057	1
49	29	K	52 spiked	0.239	1
50	908	QC1	CRM	0.044	1
51	909	QC2	CRM	0.062	1
52	921	D	Drift	2.258	1
53	0	B	Final Base	-0.097	1

Corrections

Channel	1
Baseline	done
Drift	done
Carryover	done
%	0.21
TON Recovery Factor 1 %	100.00

Spike Recovery


26 spiked Added	0.135
26 spiked Found	0.164
Recovery %	121.698
44 spiked Added	0.133
44 spiked Found	0.212
Recovery %	159.621
52 spiked Added	0.135
52 spiked Found	0.182
Recovery %	134.812

Calibration

Channel 1	TKN
-----------	-----

Type	Observed	Calculated	Target	% Error
1C	57800	2.256	2.240	0.70
2C	45357	1.711	1.680	1.87
3C	31131	1.089	1.120	-2.77
4C	17633	0.498	0.560	-10.98
5C	9463	0.141	0.224	-37.02
6C	9174	0.128	0.000	-

Calibration Fit: Linear
 Correlation Coefficient: 0.9962
 Carryover (%): 0.2077
 Calibration Equation: $y = bx + a$
 y = Peak Height Digit. Units
 x = Concentration Unit mg/L
 a = 6.238143E+003
 b = 2.285871E+004



Legend

* ... Sample offscale
 + ... Result higher than sample limit
 - ... Result lower than sample limit
 P ... Standard passed
 F ... Standard failed
 N ... Value not calculated or not used
 R ... Resample after offscale
 M ... Peak marker moved manually
 d ... Pre-Diluted sample
 D ... Post-Diluted sample
 H ... Dual calibration used
 Z ... Report negative values as zero