

General

Name of Run	231116_Q2_TKN6R4.run	Name of Analysis	TKN.ANL
Date of Report	11/22/2023	System	AA500
Run Start	11/16/2023 3:56:47 PM	Run Stop	5:17:11 PM
Operator Run/Rec.	CM MR / CLMR	Software Version	AACE 8.02 SP2
Comment	Recalculate from Run231116_Q2_TKN6.run		

Parameters

Channel	1		
Method	TKN	Dilution	
Unit	mg/L		
Calib. Fit	Linear		
Corr.Coeff.(r)	0.9968		
AUFS/Gain	0.17/60	Data	
Offset	-1844	Field	
Sensitivity	0.1378	'#2	
Sample Limit 1	-		
Sample Limit 2	-		

Results

Pk	Cup	Type	Sample ID	Value	Data	Value
0	0	B	Initial Base	-4.98	0	
1	921	P	Primer	11.05	1	
2	921	D	Drift	10.69	1	
3	921	C	10.57	10.98	1	
4	901	C	7.93	7.59	1	
5	902	C	5.29	4.85	1	
6	903	C	2.64	2.87	1	
7	904	C	1.06	1.15	1	
8	905	C	Blank	0.06	1	
9	0	B	Baseline	-5.25	1	
10	921	H1	High	10.11	1	
11	904	L1	Low	1.21	1	
12	904	L1	Low	1.33	1	
13	905	S	Blank	0.28	1	
14	906	K	Blank spiked	8.86	1	
15	907	QC1	CRM [6]	4.78	1	
16	0	B	Baseline	-4.14	1	
17	1	S	552	-5.92	1	
18	2	S	376	1.62	1	
19	3	S	563	1.79	1	
20	4	S	618	1.50	1	
21	5	S	475	3.12	1	
22	6	K	475 spiked	10.37	1	
23	907	QC1	CRM [6]	4.49	1	
24	0	B	Baseline	-4.57	1	
25	7	S	83	1.07	1	
26	8	S	574	2.03	1	
27	9	S	541	1.66	1	
28	10	S	596	1.55	1	
29	11	S	585	2.94	1	
30	12	K	585 spiked	11.53	1	
31	907	QC1	CRM [6]	4.52	1	
32	0	B	Baseline	-4.98	1	
33	13	S	508	2.10	1	
34	14	S	519	1.99	1	
35	15	S	77	2.27	1	
36	16	S	530	2.73	1	
37	17	S	486	1.73	1	
38	18	K	486 spiked	9.20	1	
39	19	S	552	2.52	1	
40	907	QC1	CRM [6]	4.71	1	
41	0	B	Baseline	-4.79	1	
42	921	D	Drift	10.69	1	

Corrections

Channel	1
Baseline	none
Drift	done
Carryover	none
%	-1.40

Spike Recovery


Blank spiked Added	8.51
Blank spiked Found	8.58
Recovery %	100.88
475 spiked Added	8.47
475 spiked Found	7.25
Recovery %	85.57
585 spiked Added	8.47
585 spiked Found	8.59
Recovery %	101.41
486 spiked Added	8.49
486 spiked Found	7.47
Recovery %	88.00

Calibration

Channel 1	TKN
-----------	-----

Type	Observed	Calculated	Target	% Error
1C	57171	10.98	10.57	3.92
2C	45691	7.59	7.93	-4.32
3C	36423	4.85	5.29	-8.40
4C	29733	2.87	2.64	8.57
5C	23939	1.15	1.06	8.68
6C	20232	0.06	0.00	-

Calibration Fit: Linear
 Correlation Coefficient: 0.9968
 Carryover (%): no
 Calibration Equation: $y = bx + a$
 y = Peak Height Digit. Units
 x = Concentration Unit mg/L
 a = 2.004531E+004
 b = 3.380035E+003



Legend

* ... Sample offscale
 + ... Result higher than sample limit
 - ... Result lower than sample limit
 P ... Standard passed
 F ... Standard failed
 N ... Value not calculated or not used
 R ... Resample after offscale
 M ... Peak marker moved manually
 d ... Pre-Diluted sample
 D ... Post-Diluted sample
 H ... Dual calibration used
 Z ... Report negative values as zero