

General

Name of Run	231122_Q2_TKN_ReRunR1.RUN	Name of Analysis	TKN.ANL
Date of Report	11/22/2023	System	AA500
Run Start	11/22/2023 12:30:32 PM	Run Stop	2:43:11 PM
Operator Run/Rec.	CLMR / CLMR	Software Version	AACE 8.02 SP2
Comment	Recalculate from Run231122_Q2_TKN_ReRunR1.RUN		

Parameters

Channel	1	
Method	TKN	Dilution
Unit	mg/L	
Calib. Fit	Linear	
Corr.Coeff.(r)	0.9997	
AUFS/Gain	0.17/58	Data
Offset	-1319	Field
Sensitivity	0.1458	'#2
Sample Limit 1	-	
Sample Limit 2	-	

Results

Pk	Cup	Type	Sample ID	Value	Data	Value
0	0	B	Initial Base	-3.843	0	
1	921	P	Primer	10.637	1	
2	921	D	Drift	10.624	1	
3	921	C	10.57	10.713	1	
4	901	C	7.93	7.831	1	
5	902	C	5.29	5.149	1	
6	903	C	2.64	2.641	1	
7	904	C	1.06	1.080	1	
8	905	C	Blank	0.077	1	
9	0	B	Baseline	-3.843	1	
10	921	H1	High	10.718	1	
11	904	L1	Low	1.023	1	
12	904	L1	Low	1.026	1	
13	905	S	Blank	-0.124	1	
14	906	K	Blank spiked	8.736	1	
15	907	QC1	CRM [6]	5.686	1	
16	1	S	78	0.201	1	
17	2	S	76	-2.386	1	
18	3	S	77	2.977	1	
19	4	S	75	-2.649	1	
20	5	S	80	-1.987	1	
21	6	S	83	2.357	1	
22	7	S	84	2.215	1	
23	8	S	79	-1.048	1	
24	9	S	81	-2.280	1	
25	10	S	87	-1.428	1	
26	11	K	87 spiked	7.725	1	
27	907	QC1	CRM [6]	5.965	1	
28	0	B	Baseline	-3.843	1	
29	12	S	82	-1.610	1	
30	13	S	244	-2.638	1	
31	14	S	255	-2.413	1	
32	15	S	266	-2.296	1	
33	16	S	277	-2.609	1	
34	17	S	299	-2.650	1	
35	18	S	288	-0.157	1	
36	19	S	211	-0.851	1	
37	20	S	222	-2.498	1	
38	21	S	233	-1.856	1	
39	22	K	233 spiked	8.105	1	
40	907	QC1	CRM [6]	6.123	1	
41	0	B	Baseline	-3.843	1	
42	23	S	332	-3.044	1	

43	24	S	310	-3.277	1
44	25	S	398	-3.105	1
45	26	S	321	-3.267	1
46	27	S	585	3.878	1
47	28	K	585 spiked	12.499	1
48	29	S	387	-3.590	1
49	30	S	365	-3.422	1
50	31	S	464	-3.269	1
51	32	S	409	-2.489	1
52	33	S	453	-3.317	1
53	34	K	453 spiked	-0.892	1
54	907	QC1	CRM [6]	5.999	1
55	0	B	Baseline	-3.843	1
56	35	S	376	1.540	1
57	36	S	431	-2.521	1
58	37	S	442	-2.214	1
59	38	S	552	2.478	1
60	39	S	475	2.814	1
61	40	K	475 spiked	-0.143	1
62	41	S	541	-1.865	1
63	42	S	530	2.762	1
64	43	S	519	-1.914	1
65	44	S	508	2.650	1
66	45	S	354	-3.442	1
67	46	K	354 spiked	6.049	1
68	907	QC1	CRM [6]	5.965	1
69	0	B	Baseline	-3.843	1
70	47	S	563	2.477	1
71	48	S	574	-1.322	1
72	49	S	486	2.158	1
73	50	K	486 spiked	-1.024	1
74	51	S	618	2.969	1
75	52	S	596	2.262	1
76	53	S	420	-3.461	1
77	54	K	420 spiked	-0.977	1
78	907	QC1	CRM [6]	5.990	1
79	0	B	Baseline	-3.843	1
80	921	D	Drift	10.624	1
81	0	B	Final Base	-3.843	1

Corrections

Channel	1
Baseline	done
Drift	done
Carryover	done
%	0.73

Spike Recovery

Blank spiked Added	8.591
Blank spiked Found	8.860
Recovery %	103.132
87 spiked Added	8.608
87 spiked Found	9.153
Recovery %	106.339
233 spiked Added	8.613
233 spiked Found	9.961
Recovery %	115.643
585 spiked Added	8.539
585 spiked Found	8.621
Recovery %	100.960
453 spiked Added	8.632
453 spiked Found	2.424
Recovery %	28.085
475 spiked Added	8.553
475 spiked Found	-2.957
Recovery %	-34.575
354 spiked Added	1.992
354 spiked Found	9.491
Recovery %	476.366
486 spiked Added	8.561
486 spiked Found	-3.182
Recovery %	-37.163


420 spiked Added	8.634
420 spiked Found	2.484
Recovery %	28.766

Calibration

Channel 1 TKN

Type	Observed	Calculated	Target	% Error
1C	58743	10.713	10.570	1.35
2C	47716	7.831	7.930	-1.25
3C	37453	5.149	5.290	-2.67
4C	27857	2.641	2.640	0.02
5C	21887	1.080	1.060	1.91
6C	18047	0.077	0.000	-

Calibration Fit: Linear
 Correlation Coefficient: 0.9997
 Carryover (%): 0.7274
 Calibration Equation: $y = bx + a$
 y = Peak Height Digit. Units
 x = Concentration Unit mg/L
 a = $1.775363E+004$
 b = $3.826163E+003$



Legend

* ... Sample offscale
 + ... Result higher than sample limit
 - ... Result lower than sample limit
 P ... Standard passed
 F ... Standard failed
 N ... Value not calculated or not used
 R ... Resample after offscale
 M ... Peak marker moved manually
 d ... Pre-Diluted sample
 D ... Post-Diluted sample
 H ... Dual calibration used
 Z ... Report negative values as zero