

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	ESDC	1.0	Travis Courtney	

Date	Time	User	Sonde
April 5 th , 2024	1:15	Manuel A. Nieves Ortiz	Eureka

Parameter	Test	Standard		Measured		Pass/Fail	Note
Depth (± 0.4 m)	Check	0.00		-0.03		P	
	Calibration	0.00		-0.03		P	
	Calib. Check	0.00		0.00		P	
Temperature (± 0.1°C)	Check	23.63		23.65		P	
	Calibration						
	Calib. Check						
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50.00		50.41		F	
	Calibration	50.00		50.39		F	
	Calib. Check	50.00		50.01		P	
pH (± 0.1 units)	Check	9.0		9.04		P	
	Calibration	9.0	7.0	9.05	6.99	P	
	Calib. Check	7.0		6.99		P	
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100.0		100.4		P	
	Calibration						
	Calib. Check						

Notes: