

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	ESDC	1.0	Travis Courtney	

Date	Time	User	Sonde
May 5th, 2024	4:45 am	Manuel A. Nieves Ortiz	Eureka

Parameter	Test	Standard		Measured		Pass/Fail		Note
Depth (± 0.4 m)	Check	0.00		-0.05		P		
	Calibration							
	Calib. Check							
Temperature (± 0.1°C)	Check	26.74		26.74		P		
	Calibration							
	Calib. Check							
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50.00		49.89		P		
	Calibration							
	Calib. Check							
pH (± 0.1 units)	Check	10.00		10.23		F		
	Calibration	10.00	7.00	10.23	6.95	F	P	
	Calib. Check	7.00		7.00		P		
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100.0		99.4		P		
	Calibration							
	Calib. Check							

Notes: