

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	ESDC	1.0	Travis Courtney	

Date	Time	User	Sonde
April 9 th , 2024	5:30	Manuel A. Nieves Ortiz	Eureka

Parameter	Test	Standard		Measured		Pass/Fail		Note
Depth (± 0.4 m)	Check	0.00		-0.01		P		
	Calibration							
	Calib. Check							
Temperature (± 0.1°C)	Check	23.73		23.71		P		
	Calibration							
	Calib. Check							
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50.00		50.29		F		
	Calibration	50.00		50.24		P		
	Calib. Check	50.00		50.03		P		
pH (± 0.1 units)	Check	10.00		10.20		F		
	Calibration	10.00	7.00	10.20	7.00	F	P	
	Calib. Check	7.00		7.00		P		
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100.0		99.9		P		
	Calibration							
	Calib. Check							

Notes: