

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	ESDC	1.0	Travis Courtney	

Date	Time	User	Sonde
April 10 th , 2024	5:30	Manuel A. Nieves Ortiz	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.00	-0.03	P	
	Calibration				
	Calib. Check				
Temperature (± 0.1°C)	Check	23.51	23.52	P	
	Calibration				
	Calib. Check				
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50.00	50.07	P	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	10.00	9.98	P	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100.0	99.8	P	
	Calibration				
	Calib. Check				

Notes: