

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
6/28/2024	4:55	Margaret Gordon (messie)	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.00	0.04	Pass	
	Calibration				
	Calib. Check				
Temperature (± 0.1°C) sp. Cond.	Check	50.00 mS	49.09	Fail	
	Calibration	50	49.02	RF 102	
	Calib. Check	50	50.00	Pass	
Sp. Conductivity Temp (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	26.45°C	26.45°C	Pass	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	10.01	9.98	Pass	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100 %	100.11 ^{mg}	Pass	
	Atmospheric Pressure		99.81	mm Hg	
	Calibration				
	Calib. Check				

Notes: