

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Aug/28/2024	4:59	M. Olmeda J. Collazo I. Santos	Eureka 10Lanta

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.0	0.01	Pass	
	Calibration				
	Calib. Check				
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check	± 0.1	24.63	Pass	
	Calibration				
	Calib. Check				
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check	± 0.25	29.500 29.500	Pass	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	± 0.1	10.05	Pass	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check	$\pm 1.5\%$	29.1000 29.1000	Pass	
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check				

Notes: