

|                                                                                                                                                                         |                 |         |                 |                                                                                     |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|---------|-----------------|-------------------------------------------------------------------------------------|
| <b>Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab</b><br><b>University of Puerto Rico – Mayagüez Campus</b><br><b>Department of Marine Sciences</b> |                 |         |                 |  |
| Document Title                                                                                                                                                          | Document Number | Versión | Author          |                                                                                     |
| Eureka Sonde Daily Check                                                                                                                                                | Appendix J      | 1.0     | Travis Courtney |                                                                                     |

| Date       | Time   | User              | Sonde  |
|------------|--------|-------------------|--------|
| 09/07/2024 | 4:50am | Johann y Victoria | Eureka |

| Parameter                                                                    | Test                 | Standard | Measured | Pass/Fail | Note |
|------------------------------------------------------------------------------|----------------------|----------|----------|-----------|------|
| Depth<br>( $\pm 0.4$ m)                                                      | Check                | 0.0      | -0.06    | Pass      |      |
|                                                                              | Calibration          | /        | /        | /         |      |
|                                                                              | Calib. Check         | /        | /        | /         |      |
| Temperature<br>( $\pm 0.1^{\circ}\text{C}$ )                                 | Check                | 29.96    | 29.97    | Pass      |      |
|                                                                              | Calibration          | /        | /        | /         |      |
|                                                                              | Calib. Check         | /        | /        | /         |      |
| Sp. Conductivity<br>( $\pm 0.5\%$ or $\pm 0.25$ mS/cm<br>for a 50 mS/cm std) | Check                | 50       | 50.50    | Fail      |      |
|                                                                              | Calibration          | 50       | 48.20    | Pass      |      |
|                                                                              | Calib. Check         | 50       | 50.01    | Pass      |      |
| pH<br>( $\pm 0.1$ units)                                                     | Check                | 10.01    | 9.99     | Pass      |      |
|                                                                              | Calibration          | /        | /        | /         |      |
|                                                                              | Calib. Check         | /        | /        | /         |      |
| Dissolved Oxygen<br>( $\pm 0.1$ mg/L or $\pm 1.5\%$ )                        | Check                | 100      | 99.7     | Pass      |      |
|                                                                              | Atmospheric Pressure |          |          | mm Hg     |      |
|                                                                              | Calibration          | /        | /        | /         |      |
|                                                                              | Calib. Check         | /        | /        | /         |      |

**Notes:**

