

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
8/5/24	4:45	SJS	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0	-0.05	P	
	Calibration				
	Calib. Check				
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check	26.46	26.48	P	
	Calibration				
	Calib. Check				
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration	50.00	52.41	F	
	Calib. Check	50.00	49.96	P	
pH (± 0.1 units)	Check	10.01	10.09	P	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check	100	99.5		
	Atmospheric Pressure		685.29	mm Hg	
	Calibration				
	Calib. Check				

Notes: