

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
09/16/24	5:06 am	Victorin	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.0	-0.07	Pass	
	Calibration	/	/	/	
	Calib. Check	/	/	/	
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check	24.85	24.82	Pass	
	Calibration	/	/		
	Calib. Check	/	/		
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50	49.89	Pass	
	Calibration	/	/		
	Calib. Check	/	/		
pH (± 0.1 units)	Check	10.01	10.12	Fail	
	Calibration	10.01	10.01	Pass	
	Calib. Check	10.01	10.08	Pass	
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check	100	99.6	Pass	
	Atmospheric Pressure		758.44	mm Hg	
	Calibration	/	/		
	Calib. Check	/	/		

Notes:

