

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Sep 16/24	4:30	M. Olmeda	Eureka Marita

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.0 ± 0.4	0.01	Pass	
	Calibration				
	Calib. Check				
Temperature (± 0.1°C)	Check	± 0.1	29.84	Pass	
	Calibration				
	Calib. Check				
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	± 0.25	50.05	Pass	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	± 0.1	10.00	Pass	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	± 1.5%	100.9%	Pass	
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check				

Notes:

