

General

Name of Run	240925_Q6_TKN3.run	Name of Analysis	TKN.ANL
Date of Report	9/25/2024	System	AA500
Run Start	9/25/2024 12:53:29 PM	Run Stop	2:35:33 PM
Operator Run/Rec.	CLMR	Software Version	AACE 8.02 SP2
Comment	2nd Batch		

Parameters

Channel	1		
Method	TKN	Dilution	
Unit	mg/L		
Calib. Fit	Linear		
Corr.Coeff.(r)	0.9984		
AUFS/Gain	0.14/69	Data	
Offset	-2595	Field	
Sensitivity	0.1147	'#2	
Sample Limit 1	-		
Sample Limit 2	-		

Results

Pk	Cup	Type	Sample ID	Value	Data	Value
0	0	B	Initial Base	0.116	0	
1	921	P	Primer	2.293	1	
2	921	D	Drift	2.275	1	
3	921	C	2.22	2.276	1	
4	901	C	1.66	1.578	1	
5	902	C	1.11	1.128	1	
6	903	C	0.55	0.523	1	
7	904	C	0.22	0.251	1	
8	905	C	Blank	0.004	1	
9	0	B	Baseline	0.116	1	
10	921	H1	High	2.273	1	
11	904	L1	Low	0.235	1	
12	904	L1	Low	0.235	1	
13	905	S	Blank	0.020	1	
14	906	K	Blank Spike	0.978	1	
15	908	QC1	CRM 9/24 [1.5 mg/L]	1.338	1	
16	0	B	Baseline	0.116	1	
17	1	S	556	0.021	1	
18	2	S	452	0.168	1	
19	3	S	426	-0.005	1	
20	4	S	504	0.020	1	
21	5	S	517	0.020	1	
22	6	S	530	-0.007	1	
23	7	S	595	0.099	1	
24	8	S	569	0.040	1	
25	9	S	400	0.027	1	
26	10	S	413	0.017	1	
27	11	S	543	0.020	1	
28	12	S	387	0.003	1	
29	13	K	387 spiked	1.094	1	
30	908	QC1	CRM 9/24 [1.5 mg/L]	1.347	1	
31	0	B	Baseline	0.116	1	
32	14	S	309	-0.020	1	
33	15	S	374	-0.008	1	
34	16	S	465	-0.041	1	
35	17	S	348	-0.064	1	
36	18	S	361	-0.011	1	
37	19	S	491	0.033	1	
38	20	S	231	-0.001	1	
39	21	S	439	-0.053	1	
40	22	S	478	-0.053	1	
41	23	S	608	0.218	1	
42	24	S	582	-0.046	1	

43	25	K	582 spiked	0.702	1
44	908	QC1	CRM 9/24 [1.5 mg/L]	1.326	1
45	0	B	Baseline	0.116	1
46	905	S	Blank	0.020	1
47	905	S	Blank	0.027	1
48	905	S	Blank	0.022	1
49	905	S	Blank	0.022	1
50	905	S	Blank	0.025	1
51	905	S	Blank	0.021	1
52	905	S	Blank	0.021	1
53	905	S	Blank	0.014	1
54	905	S	Blank	0.018	1
55	905	S	Blank	0.035	1
56	906	K	Blank spiked	0.942	1
57	908	QC1	CRM 9/24 [1.5 mg/L]	1.269	1
58	0	B	Baseline	0.116	1
59	921	D	Drift	2.264 *	1
60	0	B	Final Base	0.116	1

Corrections

Channel	1
Baseline	done
Drift	none
Carryover	done
%	0.53

Spike Recovery


Blank Spike Added	1.001
Blank Spike Found	0.958
Recovery %	95.704
387 spiked Added	1.001
387 spiked Found	1.091
Recovery %	109.017
582 spiked Added	1.001
582 spiked Found	0.748
Recovery %	74.730
Blank spiked Added	1.001
Blank spiked Found	0.907
Recovery %	90.636

Calibration

Channel 1	TKN
-----------	-----

Type	Observed	Calculated	Target	% Error
1C	56569	2.276	2.220	2.53
2C	39938	1.578	1.660	-4.94
3C	29211	1.128	1.110	1.58
4C	14801	0.523	0.550	-5.00
5C	8344	0.251	0.220	14.27
6C	2460	0.004	0.000	-

Calibration Fit: Linear
 Correlation Coefficient: 0.9984
 Carryover (%): 0.5264
 Calibration Equation: $y = bx + a$
 y = Peak Height Digit. Units
 x = Concentration Unit mg/L
 a = 2.356960E+003
 b = 2.381619E+004



Legend

* ... Sample offscale
+ ... Result higher than sample limit
- ... Result lower than sample limit
P ... Standard passed
F ... Standard failed
N ... Value not calculated or not used
R ... Resample after offscale
M ... Peak marker moved manually
d ... Pre-Diluted sample
D ... Post-Diluted sample
H ... Dual calibration used
Z ... Report negative values as zero