

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Dec-03-2024	5:56	M.O/mo V. Mendez	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check				
	Calibration	± 0.4	0.05	Pass	
	Calib. Check				
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check				
	Calibration	± 0.1	24.75 $^{\circ}\text{C}$	Pass	
	Calib. Check				
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration	± 0.25	50.05	Pass	
	Calib. Check				
pH (± 0.1 units)	Check				
	Calibration	± 0.1	10.01	Pass	
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check				
	Atmospheric Pressure			mm Hg	
	Calibration	$\pm 1.5\%$	99.9%	Pass	
	Calib. Check				

Notes: