

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
9/Enc/25	5:09 AM	J.S./Vic/Algh	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	0.00	0.00	Pass	
	Calibration				
	Calib. Check				
Temperature (± 0.1°C)	Check	24.77	24.78	Pass.	
	Calibration				
	Calib. Check				
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check	50.00	49.98	Pass	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	7.00	7.05	Pass.	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check	100.00	99.9		
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check				

Notes: