

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Nov/19/24	5:10	M. Olmeida / J. C. Lopez	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check				
	Calibration				
	Calib. Check	0.0 ± 0.4 m	-0.04 m	Pass	
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check				
	Calibration				
	Calib. Check	$\pm 0.1^{\circ}\text{C}$	24.29 $^{\circ}\text{C}$	Pass	
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration				
	Calib. Check	± 0.25	49.81 mS	Pass	
pH (± 0.1 units)	Check				
	Calibration				
	Calib. Check	± 0.1	9.98	Pass	
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check				
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check	$\pm 1.5\%$	99.6	Pass	

Notes: