

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Nov/20/24	9:41AM	M. Olmedo, J. Calles	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check				
	Calibration				
	Calib. Check	± 0.4m	-0.03m	Pass	
Temperature (± 0.1°C)	Check				
	Calibration				
	Calib. Check	± 0.1°C	24.73°C	Pass	
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration				
	Calib. Check	50.00±0.25	50.00	Pass	
pH (± 0.1 units)	Check				
	Calibration				
	Calib. Check	10.00±0.1	10.04	Pass	
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check				
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check	100±1.5%	99.4%	Pass	

Notes: