

General

Name of Run	250212_Q7_TKN4_Rerun.run	Name of Analysis	TKN.ANL
Date of Report	2/12/2025	System	AA500
Run Start	2/12/2025 4:31:20 PM	Run Stop	5:33:16 PM
Operator Run/Rec. CLMR		Software Version	AACE 8.02 SP2
Comment			

250207_Q7_TKN6.run

Parameters

Channel	1		
Method	TKN	Dilution	
Unit	mg/L		
Calib. Fit	Linear		
Corr.Coeff.(r)	0.9979		
AUFS/Gain	0.14/71	Data	
Offset	-2539	Field	
Sensitivity	0.1182	'#2	
Sample Limit 1	-		
Sample Limit 2	-		

Results

Pk	Cup	Type	Sample ID	Value	Data	Value
0	0	B	Initial Base	-0.40	0	
1	921	P	Primer	2.32	1	
2	921	D	Drift	2.24	1	
3	921	H1	High	2.23	1	
4	904	L1	Low	0.29	1	
5	904	L1	Low	0.29	1	
6	905	S	Blank	-0.13	1	
7	906	K	Blank Spike	0.88	1	
8	907	QC1	CRM [1.5 mg/L]	1.32	1	
9	0	B	Baseline	-0.40	1	
10	1	S	283	-0.40	1	
11	2	S	400	-0.04	1	
12	3	S	387	0.08	1	
13	4	S	478	0.10	1	
14	5	S	335	0.20	1	
15	6	S	322	0.07	1	
16	7	S	296	0.00	1	
17	8	S	309	-0.01	1	
18	1	S	283	0.02	1	
19	9	S	465	-0.06	1	
20	10	K	465 spiked	1.06	1	
21	907	QC1	CRM [1.5 mg/L]	1.36	1	
22	0	B	Baseline	-0.40	1	
23	905	S	Blank	-0.14	1	
24	905	S	Blank	0.01	1	
25	905	S	Blank	0.01	1	
26	905	S	Blank	-0.03	1	
27	905	S	Blank	-0.03	1	
28	905	S	Blank	-0.04	1	
29	906	K	Blank spiked	0.94	1	
30	907	QC1	CRM [1.5 mg/L]	1.41	1	
31	0	B	Baseline	-0.40	1	
32	921	D	Drift	2.24	1	
33	0	B	Final Base	-0.40	1	

Corrections

Channel	1
Baseline	done
Drift	done
Carryover	done
%	2.18

Spike Recovery


Blank Spike Added	1.00
Blank Spike Found	1.01
Recovery %	101.09
465 spiked Added	1.00
465 spiked Found	1.12
Recovery %	112.03
Blank spiked Added	1.00
Blank spiked Found	0.98
Recovery %	97.52

Calibration

Channel 1 TKN

Type	Observed	Calculated	Target	% Error
1C	57439	2.23	2.22	0.57
2C	46203	1.68	1.66	0.97
3C	33849	1.06	1.11	-4.14
4C	22533	0.50	0.55	-8.47
5C	18820	0.32	0.22	45.21
6C	11653	-0.04	0.00	-

Calibration Fit: Linear
Correlation Coefficient: 0.9979
Carryover (%): 2.1792
Calibration Equation: $y = bx + a$
y = Peak Height Digit. Units
x = Concentration Unit mg/L
a = 1.237165E+004
b = 2.018526E+004



Legend

* ... Sample offscale
+ ... Result higher than sample limit
- ... Result lower than sample limit
P ... Standard passed
F ... Standard failed
N ... Value not calculated or not used
R ... Resample after offscale
M ... Peak marker moved manually
d ... Pre-Diluted sample
D ... Post-Diluted sample
H ... Dual calibration used
Z ... Report negative values as zero