

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
Feb/20/25	5:48	MD/meda/ACHabrier	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check				
	Calibration	± 0.4 m	-0.01	Pass	
	Calib. Check				
Temperature ($\pm 0.1^{\circ}\text{C}$)	Check	$\pm 0.1^{\circ}\text{C}$	27.52 $^{\circ}\text{C}$	Pass	
	Calibration				
	Calib. Check				
Sp. Conductivity ($\pm 0.5\%$ or ± 0.25 mS/cm for a 50 mS/cm std)	Check	100 mS/cm	100.1	Pass	
	Calibration				
	Calib. Check				
pH (± 0.1 units)	Check	7.0	7.00	Pass	
	Calibration				
	Calib. Check				
Dissolved Oxygen (± 0.1 mg/L or $\pm 1.5\%$)	Check	$\pm 1.5\%$	99.9%	Pass	
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check				

Notes: