

|                                                                                                                                                                          |                 |         |                 |                                                                                     |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|---------|-----------------|-------------------------------------------------------------------------------------|
| <b>Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab</b><br><b>University of Puerto Rico – Mayagüez Campus</b><br><b>Department of Marine Sciences'</b> |                 |         |                 |  |
| Document Title                                                                                                                                                           | Document Number | Version | Author          |                                                                                     |
| Eureka Sonde Daily Check                                                                                                                                                 | Appendix J      | 1.0     | Travis Courtney |                                                                                     |

| Date     | Time | User   | Sonde  |
|----------|------|--------|--------|
| 02/21/25 | 5:00 | Molina | Eureka |

| Parameter                                                          | Test                 | Standard  | Measured    | Pass/Fail | Note |
|--------------------------------------------------------------------|----------------------|-----------|-------------|-----------|------|
| Depth<br>(± 0.4 m)                                                 | Check                | ± 0.4m    | -0.05m      | Pass      |      |
|                                                                    | Calibration          |           |             |           |      |
|                                                                    | Calib. Check         |           |             |           |      |
| Temperature<br>(± 0.1°C)                                           | Check                | ± 0.1°C   | 25.40°C     | Pass      |      |
|                                                                    | Calibration          |           |             |           |      |
|                                                                    | Calib. Check         |           |             |           |      |
| Sp. Conductivity<br>(± 0.5% or ± 0.25 mS/cm<br>for a 50 mS/cm std) | Check                | 100 mS/cm | 100.1 mS/cm | Pass      |      |
|                                                                    | Calibration          |           |             |           |      |
|                                                                    | Calib. Check         |           |             |           |      |
| pH<br>(± 0.1 units)                                                | Check                | 7.0       | 7.04        | Pass      |      |
|                                                                    | Calibration          |           |             |           |      |
|                                                                    | Calib. Check         |           |             |           |      |
| Dissolved Oxygen<br>(± 0.1 mg/L or ± 1.5 %)                        | Check                | 100       | 99.8%       | Pass      |      |
|                                                                    | Atmospheric Pressure |           |             | mm Hg     |      |
|                                                                    | Calibration          |           |             |           |      |
|                                                                    | Calib. Check         |           |             |           |      |

**Notes:**