

03/03/25
M. Olmeda, Am Chabrier, VS Méndez

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Versión	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
03/03/25	5:30am	M. Olmeda, A. Chabrier V. Méndez	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check	VSMT			
	Calibration				
	Calib. Check	0	-0.06	Pass	
Temperature (± 0.1°C)	Check				
	Calibration	28.92	28.99	Pass	
	Calib. Check				
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration				
	Calib. Check	100.00	100.00		
pH (± 0.1 units)	Check				
	Calibration				
	Calib. Check	7.0	7.03	Pass	
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check				
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check	VSMT 100	99.3	Pass	

Notes: