

Caribbean Coral Reef Institute (CCRI) Water Quality Monitoring Lab University of Puerto Rico – Mayagüez Campus Department of Marine Sciences				
Document Title	Document Number	Version	Author	
Eureka Sonde Daily Check	Appendix J	1.0	Travis Courtney	

Date	Time	User	Sonde
03/12/25	5:42am	M. Olmeda / V. Méndez	Eureka

Parameter	Test	Standard	Measured	Pass/Fail	Note
Depth (± 0.4 m)	Check				
	Calibration				
	Calib. Check	0	-0.02	Pass	
Temperature (± 0.1°C)	Check				
	Calibration				
	Calib. Check	25.43	25.36	Pass	
Sp. Conductivity (± 0.5% or ± 0.25 mS/cm for a 50 mS/cm std)	Check				
	Calibration				
	Calib. Check	100	100.3	Pass	
pH (± 0.1 units)	Check				
	Calibration				
	Calib. Check	7.0	7.08	Pass	
Dissolved Oxygen (± 0.1 mg/L or ± 1.5 %)	Check				
	Atmospheric Pressure			mm Hg	
	Calibration				
	Calib. Check	100	99.3	Pass	

Notes: