UGAIYA BIO-SCIENCES

TG12M Micro Hematocrit Centrifuge

Specification

Model	:	TG12M
Max Speed	:	12000rpm
Max RCF	:	14800xg
Max Capacity	:	36 Capillaries
Time Setting Range	:	1s to 99 min
Speed Accuracy	:	±30 rpm
Power Supply	:	AC220V±22V 50Hz 5A
Power Consumption	:	300W
Noise	:	≤ 62 dB(A)
Dimension(WxDxH)	:	320 x 370 x 235 (mm)
Package Dimension (WxDxH)	:	530 x 410 x 290 (mm)
Net Weight	:	19 kg

Product Advantages

- Microprocessor control, DC motor without carbon brush offering stable and quiet operation.
- Parameters can be modified during running.
- Automatic calculation the value of RCF, and a special key is set to observe the RCF value at any time.
- Time is shown in second when the time counts down less than 1 minutes.
- Accomodating self-diagnostic system tio detect the fault of over-speed, or the speed without signal.
- Designed to determine the microhematocrit, seperating traces of blood of solutions.
- Main machine is in cast steel shell, stainless steel centrifuge chamber & aluminium alloy rotors.

Max Speed : 12000 rpm
Max RCF : 14800xg
Capillaries : 36

Exclusive Distributor : PT Gravia Citra Lestari
GC Aesthetics Clinic Building 3rd & 4th Floor
Jl. Wolter Monginsidi No. 84 HH, Petogogan Kebayoran Baru, Jakarta Selatan, Indonesia – 12170
Website : www.graviacitralestari.id
Phone : (021) 7398967 | Email : sales@graviacitralestari.id

CDAKB ISO 9001:2015