UGAIYA BIO-SCIENCES

# TD-24K Gel Card Centrifuge

## Specification

| | | |
|---|---|---|
| Model | : | TD-24K |
| Max Speed | : | 2000 rpm |
| Max RCF | : | 1790xg |
| Microcolumn Gel Number | : | 24 Cards |
| Max Capacity | : | 12x20ml |
| Time Setting Range | : | 1 - 99 min |
| Speed Accuracy | : | ±30r/min |
| Power Supply | : | AC 220V±22V 50/60Hz |
| Power Consumption | : | 260W |
| Noise | : | ≤ 60dB(A) |
| Dimension(WxDxH) | : | 38 x 50 x 30 cm |
| Package Dimension (WxDxH) | : | 46,5 x 66 x 35 cm |
| Net Weight | : | 19 kg |

Rotor data special program for blood type test card

| | | | |
|---|---|---|---|
| Speed | : | 800rpm | 1500rpm |
| RCF | : | 75xg | 200xg |
| Duration | : | 0-2 minutes | 2-5 minutes |

Two steps of operations can be conducted through one key.
Note : Speed can be customized (max. speed is 2000rpm).

Blood type test rotor with 24 places (Microcolumn gel)

## Product Advantages

- Built in standard programs for various experiments such as blood type detection, blood serum, etc., so that clinical experimental detection and research are standardized.
- User-defined program can also be set to meet the actual needs.
- User-friendly design with soft-touch door closure automatically.
- Microprocessor control, DC brushless motor.
- Low noise and no carbon dust pollution.

LKPP ecatalogue

AKL 20102520259

**Exclusive Distributor : PT Gravia Citra Lestari**
**GC Aesthetics Clinic Building 3rd & 4th Floor**
Jl. Wolter Monginsidi No. 84 HH, Petogogan Kebayoran Baru, Jakarta Selatan, Indonesia – 12170
Website : www.graviacitralestari.id
Phone : (021) 7398967 | Email : sales@graviacitralestari.id

CDAKB ISO 9001:2015